

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 10/9/2025 5:49:00 PM

JOB DESCRIPTION

RED-HILL
PFAS: Halawa Shaft Viewing Pool

JOB NUMBER

380-174638-1

Eurofins Eaton Analytical Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
10/9/2025 5:49:00 PM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	15
QC Association Summary	26
Lab Chronicle	27
Certification Summary	28
Method Summary	29
Sample Summary	30
Chain of Custody	31
Receipt Checklists	32

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Qualifiers

LCMS

Qualifier	Qualifier Description
*5-	Isotope dilution analyte is outside acceptance limits, low biased.
F1	MS and/or MSD recovery exceeds control limits.
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-174638-1

Job ID: 380-174638-1

Eurofins Eaton Analytical Pomona

Job Narrative 380-174638-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 10/2/2025 10:06 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperatures of the 4 coolers at receipt time were 0.8°C, 3.2°C, 4.8°C and 5.6°C.

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-174638-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanesulfonic acid (PFHxS)	3.3		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.2		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.1		2.0	ng/L	1		537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.4		2.0	ng/L	1		537.1	Total/NA

Client Sample ID: HALAWA SHAFT VIEWING POOL BLANK

Lab Sample ID: 380-174638-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical Pomona



Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-174638-1

Date Collected: 09/30/25 09:45

Matrix: Drinking Water

Date Received: 10/02/25 10:06

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluorohexanesulfonic acid (PFHxS)	3.3		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluorooctanesulfonic acid (PFOS)	3.2		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:07	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	112		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C6 PFDA	104		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C5 PFHxA	114		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C4 PFHpA	108		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C8 PFOA	108		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C9 PFNA	106		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C7 PFUnA	102		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C2 PFDoA	103		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C4 PFBA	110		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C5 PFPeA	107		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C3 PFBS	120		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C3 PFHxS	113		50 - 200	10/06/25 07:46	10/07/25 10:07	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-174638-1

Date Collected: 09/30/25 09:45

Matrix: Drinking Water

Date Received: 10/02/25 10:06

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	110		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C2-4:2-FTS	114		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C2-6:2-FTS	101		50 - 200	10/06/25 07:46	10/07/25 10:07	1
13C2-8:2-FTS	95		50 - 200	10/06/25 07:46	10/07/25 10:07	1

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluorooctanesulfonic acid (PFOS)	3.1		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
N-methylperfluorooctanesulfonamideacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
N-ethylperfluorooctanesulfonamideacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluorohexanesulfonic acid (PFHxS)	3.4		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:25	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	109		70 - 130			10/06/25 01:42	10/06/25 18:25	1
13C2 PFHxA	121		70 - 130			10/06/25 01:42	10/06/25 18:25	1
13C2 PFDA	123		70 - 130			10/06/25 01:42	10/06/25 18:25	1
13C3-GenX	121		70 - 130			10/06/25 01:42	10/06/25 18:25	1

Client Sample ID: HALAWA SHAFT VIEWING POOL BLANK

Lab Sample ID: 380-174638-2

Date Collected: 09/30/25 09:45

Matrix: Water

Date Received: 10/02/25 10:06

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL BLANK

Lab Sample ID: 380-174638-2

Date Collected: 09/30/25 09:45

Matrix: Water

Date Received: 10/02/25 10:06

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		10/06/25 07:46	10/07/25 10:16	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	103		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C6 PFDA	85		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C5 PFHxA	105		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C4 PFHpA	101		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C8 PFOA	95		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C9 PFNA	89		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C7 PFUnA	82		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C2 PFDoA	85		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C4 PFBA	101		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C5 PFPeA	97		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C3 PFBS	108		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C3 PFHxS	102		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C8 PFOS	91		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C2-4:2-FTS	98		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C2-6:2-FTS	86		50 - 200	10/06/25 07:46	10/07/25 10:16	1
13C2-8:2-FTS	76		50 - 200	10/06/25 07:46	10/07/25 10:16	1

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL BLANK

Lab Sample ID: 380-174638-2

Date Collected: 09/30/25 09:45

Matrix: Water

Date Received: 10/02/25 10:06

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
N-methylperfluorooctanesulfonamideacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
N-ethylperfluorooctanesulfonamideacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		10/06/25 01:42	10/06/25 18:36	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	101		70 - 130			10/06/25 01:42	10/06/25 18:36	1
13C2 PFHxA	118		70 - 130			10/06/25 01:42	10/06/25 18:36	1
13C2 PFDA	121		70 - 130			10/06/25 01:42	10/06/25 18:36	1
13C3-GenX	116		70 - 130			10/06/25 01:42	10/06/25 18:36	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-174638-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.3		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.2		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.1		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.4		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

Client Sample ID: HALAWA SHAFT VIEWING POOL BLANK

Lab Sample ID: 380-174638-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

Surrogate Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-174638-1
 SDG: PFAS: Halawa Shaft Viewing Pool

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-174638-1	HALAWA SHAFT VIEWING POC	109	121	123	121

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFHxA = 13C2 PFHxA
 PFDA = 13C2 PFDA
 GenX = 13C3-GenX

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-174629-B-1-A MS	Matrix Spike	103	120	119	114
380-174629-C-1-A MSD	Matrix Spike Duplicate	103	116	120	115
380-174638-2	HALAWA SHAFT VIEWING POOL BLANK	101	118	121	116
LCS 380-177960/27-A	Lab Control Sample	97	116	117	115
MBL 380-177960/25-A	Method Blank	102	117	121	117
MRL 380-177960/26-A	Lab Control Sample	94	120	122	103

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFHxA = 13C2 PFHxA
 PFDA = 13C2 PFDA
 GenX = 13C3-GenX

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDoA (50-200)
380-174638-1	HALAWA SHAFT VIEWING POC	112	104	114	108	108	106	102	103

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-174638-1	HALAWA SHAFT VIEWING POC	110	107	120	113	110	114	101	95

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDoA = 13C2 PFDoA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDoA (50-200)
380-174633-E-1-A MS	Matrix Spike	33 *5-	30 *5-	35 *5-	35 *5-	34 *5-	31 *5-	37 *5-	53
380-174633-F-1-A MSD	Matrix Spike Duplicate	60	70	64	63	62	64	81	90
380-174638-2	HALAWA SHAFT VIEWING POOL BLANK	103	85	105	101	95	89	82	85
LCS 380-177966/22-A	Lab Control Sample	114	110	114	112	111	107	106	107
MBL 380-177966/20-A	Method Blank	111	113	115	116	112	111	109	108
MRL 380-177966/21-A	Lab Control Sample	111	106	115	111	108	110	106	106

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-174633-E-1-A MS	Matrix Spike	38 *5-	34 *5-	116	112	114	115	105	102
380-174633-F-1-A MSD	Matrix Spike Duplicate	64	60	115	113	111	104	103	97
380-174638-2	HALAWA SHAFT VIEWING POOL BLANK	101	97	108	102	91	98	86	76
LCS 380-177966/22-A	Lab Control Sample	115	110	120	115	114	114	109	104
MBL 380-177966/20-A	Method Blank	115	109	120	117	115	121	109	108
MRL 380-177966/21-A	Lab Control Sample	111	105	121	115	112	117	106	104

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDoA = 13C2 PFDoA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-177966/20-A
Matrix: Water
Analysis Batch: 178159

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 177966

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.25		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		10/06/25 07:46	10/07/25 06:27	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	111		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C6 PFDA	113		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C5 PFHxA	115		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C4 PFHpA	116		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C8 PFOA	112		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C9 PFNA	111		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C7 PFUnA	109		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C2 PFDoA	108		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C4 PFBA	115		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C5 PFPeA	109		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C3 PFBS	120		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C3 PFHxS	117		50 - 200	10/06/25 07:46	10/07/25 06:27	1

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-177966/20-A
Matrix: Water
Analysis Batch: 178159

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 177966

<i>Isotope Dilution</i>	<i>MBL %Recovery</i>	<i>MBL Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C8 PFOS	115		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C2-4:2-FTS	121		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C2-6:2-FTS	109		50 - 200	10/06/25 07:46	10/07/25 06:27	1
13C2-8:2-FTS	108		50 - 200	10/06/25 07:46	10/07/25 06:27	1

Lab Sample ID: LCS 380-177966/22-A
Matrix: Water
Analysis Batch: 178159

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 177966

<i>Analyte</i>	<i>Spike Added</i>	<i>LCS Result</i>	<i>LCS Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	60.2	54.7		ng/L		91	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	60.2	53.5		ng/L		89	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	60.2	56.8		ng/L		94	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	60.2	60.1		ng/L		100	70 - 130
Perfluorobutanesulfonic acid (PFBS)	60.2	55.5		ng/L		92	70 - 130
Perfluorodecanoic acid (PFDA)	60.2	56.9		ng/L		95	70 - 130
Perfluorododecanoic acid (PFDoA)	60.2	58.6		ng/L		97	70 - 130
Perfluoroheptanoic acid (PFHpA)	60.2	56.6		ng/L		94	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	60.2	57.5		ng/L		95	70 - 130
Perfluorohexanoic acid (PFHxA)	60.2	59.0		ng/L		98	70 - 130
Perfluorononanoic acid (PFNA)	60.2	57.7		ng/L		96	70 - 130
Perfluorooctanesulfonic acid (PFOS)	60.2	57.4		ng/L		95	70 - 130
Perfluorooctanoic acid (PFOA)	60.2	55.6		ng/L		92	70 - 130
Perfluoroundecanoic acid (PFUnA)	60.2	58.0		ng/L		96	70 - 130
Perfluorobutanoic acid (PFBA)	60.2	56.4		ng/L		94	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	60.2	62.4		ng/L		104	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	60.2	62.7		ng/L		104	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	60.2	58.2		ng/L		97	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	60.2	60.2		ng/L		100	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	60.2	57.3		ng/L		95	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	60.2	56.2		ng/L		93	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	60.2	56.4		ng/L		94	70 - 130
Perfluoropentanoic acid (PFPeA)	60.2	57.9		ng/L		96	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	60.2	57.7		ng/L		96	70 - 130

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-177966/22-A
Matrix: Water
Analysis Batch: 178159

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 177966

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	60.2	58.9		ng/L		98	70 - 130
LCS LCS							
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	114		50 - 200				
13C6 PFDA	110		50 - 200				
13C5 PFHxA	114		50 - 200				
13C4 PFHpA	112		50 - 200				
13C8 PFOA	111		50 - 200				
13C9 PFNA	107		50 - 200				
13C7 PFUnA	106		50 - 200				
13C2 PFDoA	107		50 - 200				
13C4 PFBA	115		50 - 200				
13C5 PFPeA	110		50 - 200				
13C3 PFBS	120		50 - 200				
13C3 PFHxS	115		50 - 200				
13C8 PFOS	114		50 - 200				
13C2-4:2-FTS	114		50 - 200				
13C2-6:2-FTS	109		50 - 200				
13C2-8:2-FTS	104		50 - 200				

Lab Sample ID: MRL 380-177966/21-A
Matrix: Water
Analysis Batch: 178159

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 177966

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.76	J	ng/L		88	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.01	1.75	J	ng/L		87	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.88	J	ng/L		94	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	2.03	J	ng/L		101	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	1.79	J	ng/L		89	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	1.92	J	ng/L		96	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	1.98	J	ng/L		99	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	1.96	J	ng/L		98	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	1.87	J	ng/L		93	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	2.04	J	ng/L		102	50 - 150
Perfluorononanoic acid (PFNA)	2.01	1.87	J	ng/L		93	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	1.99	J	ng/L		99	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	1.86	J	ng/L		92	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	1.87	J	ng/L		93	50 - 150
Perfluorobutanoic acid (PFBA)	2.01	1.96	J	ng/L		98	50 - 150

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-177966/21-A
Matrix: Water
Analysis Batch: 178159

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 177966

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.01	2.10	J	ng/L		105	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.01	2.28	J	ng/L		114	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.01	1.95	J	ng/L		97	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.01	2.13	J	ng/L		106	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.01	1.83	J	ng/L		91	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.01	1.89	J	ng/L		94	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.01	1.82	J	ng/L		91	50 - 150
Perfluoropentanoic acid (PFPeA)	2.01	2.00	J	ng/L		100	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.01	1.95	J	ng/L		97	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.01	1.90	J	ng/L		94	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	111		50 - 200
13C6 PFDA	106		50 - 200
13C5 PFHxA	115		50 - 200
13C4 PFHpA	111		50 - 200
13C8 PFOA	108		50 - 200
13C9 PFNA	110		50 - 200
13C7 PFUnA	106		50 - 200
13C2 PFDoA	106		50 - 200
13C4 PFBA	111		50 - 200
13C5 PFPeA	105		50 - 200
13C3 PFBS	121		50 - 200
13C3 PFHxS	115		50 - 200
13C8 PFOS	112		50 - 200
13C2-4:2-FTS	117		50 - 200
13C2-6:2-FTS	106		50 - 200
13C2-8:2-FTS	104		50 - 200

Lab Sample ID: 380-174633-E-1-A MS
Matrix: Water
Analysis Batch: 178159

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 177966

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.5	53.8		ng/L		89	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.5	52.4		ng/L		87	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.5	51.0	*5-	ng/L		84	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-174633-E-1-A MS
Matrix: Water
Analysis Batch: 178159

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 177966

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		60.5	57.9	*5-	ng/L		96	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	<2.0		60.5	58.4		ng/L		97	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		60.5	56.3	*5-	ng/L		93	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		60.5	57.8		ng/L		96	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		60.5	58.3	*5-	ng/L		96	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		60.5	57.9		ng/L		96	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		60.5	58.2	*5-	ng/L		96	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		60.5	57.8	*5-	ng/L		96	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		60.5	56.9		ng/L		94	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		60.5	55.3	*5-	ng/L		91	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		60.5	58.8	*5-	ng/L		97	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		60.5	55.9	*5-	ng/L		92	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.5	62.1		ng/L		103	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.5	61.3		ng/L		101	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.5	58.9		ng/L		97	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.5	52.9	*5-	ng/L		87	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.5	58.2		ng/L		96	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.5	51.8	*5-	ng/L		86	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.5	51.8	*5-	ng/L		86	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		60.5	57.4	*5-	ng/L		95	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.5	57.4		ng/L		95	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.5	58.5		ng/L		97	70 - 130

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	33	*5-	50 - 200
13C6 PFDA	30	*5-	50 - 200
13C5 PFHxA	35	*5-	50 - 200
13C4 PFHpA	35	*5-	50 - 200
13C8 PFOA	34	*5-	50 - 200
13C9 PFNA	31	*5-	50 - 200
13C7 PFUnA	37	*5-	50 - 200
13C2 PFDoA	53		50 - 200
13C4 PFBA	38	*5-	50 - 200
13C5 PFPeA	34	*5-	50 - 200
13C3 PFBS	116		50 - 200
13C3 PFHxS	112		50 - 200
13C8 PFOS	114		50 - 200

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-174633-E-1-A MS
Matrix: Water
Analysis Batch: 178159

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 177966

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>MS MS Qualifier</i>	<i>Limits</i>
13C2-4:2-FTS	115		50 - 200
13C2-6:2-FTS	105		50 - 200
13C2-8:2-FTS	102		50 - 200

Lab Sample ID: 380-174633-F-1-A MSD
Matrix: Water
Analysis Batch: 178159

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 177966

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.5	54.8		ng/L		91	70 - 130	2	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.5	54.7		ng/L		90	70 - 130	4	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.5	53.5		ng/L		88	70 - 130	5	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		60.5	60.6		ng/L		100	70 - 130	5	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		60.5	60.5		ng/L		100	70 - 130	3	30
Perfluorodecanoic acid (PFDA)	<2.0		60.5	58.9		ng/L		97	70 - 130	4	30
Perfluorododecanoic acid (PFDoA)	<2.0		60.5	59.3		ng/L		98	70 - 130	3	30
Perfluoroheptanoic acid (PFHpA)	<2.0		60.5	58.9		ng/L		97	70 - 130	1	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		60.5	57.0		ng/L		94	70 - 130	1	30
Perfluorohexanoic acid (PFHxA)	<2.0		60.5	57.5		ng/L		95	70 - 130	1	30
Perfluorononanoic acid (PFNA)	<2.0		60.5	58.6		ng/L		97	70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		60.5	58.1		ng/L		96	70 - 130	2	30
Perfluorooctanoic acid (PFOA)	<2.0		60.5	58.4		ng/L		97	70 - 130	6	30
Perfluoroundecanoic acid (PFUnA)	<2.0		60.5	58.4		ng/L		97	70 - 130	1	30
Perfluorobutanoic acid (PFBA)	<2.0		60.5	58.7		ng/L		97	70 - 130	5	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.5	63.7		ng/L		105	70 - 130	3	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.5	64.7		ng/L		107	70 - 130	5	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.5	58.6		ng/L		97	70 - 130	0	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.5	57.3		ng/L		95	70 - 130	8	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		60.5	57.7		ng/L		95	70 - 130	1	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.5	53.2		ng/L		88	70 - 130	3	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.5	53.0		ng/L		88	70 - 130	2	30
Perfluoropentanoic acid (PFPeA)	<2.0		60.5	62.6		ng/L		104	70 - 130	9	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.5	58.2		ng/L		96	70 - 130	1	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.5	59.1		ng/L		98	70 - 130	1	30

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

<i>Isotope Dilution</i>	<i>MSD</i>	<i>MSD</i>	<i>Limits</i>
	<i>%Recovery</i>	<i>Qualifier</i>	
13C3 HFPO-DA	60		50 - 200
13C6 PFDA	70		50 - 200
13C5 PFHxA	64		50 - 200
13C4 PFHpA	63		50 - 200
13C8 PFOA	62		50 - 200
13C9 PFNA	64		50 - 200
13C7 PFUnA	81		50 - 200
13C2 PFDoA	90		50 - 200
13C4 PFBA	64		50 - 200
13C5 PFPeA	60		50 - 200
13C3 PFBS	115		50 - 200
13C3 PFHxS	113		50 - 200
13C8 PFOS	111		50 - 200
13C2-4:2-FTS	104		50 - 200
13C2-6:2-FTS	103		50 - 200
13C2-8:2-FTS	97		50 - 200

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MBL 380-177960/25-A
Matrix: Water
Analysis Batch: 178050

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 177960

<i>Analyte</i>	<i>MBL</i>	<i>MBL</i>	<i>RL</i>	<i>Unit</i>	<i>D</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
	<i>Result</i>	<i>Qualifier</i>						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
Perfluorotridecanoic acid (PFTrDA)	<0.36		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		10/06/25 01:42	10/06/25 15:43	1
<i>Surrogate</i>	<i>MBL</i>	<i>MBL</i>	<i>Limits</i>			<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
	<i>%Recovery</i>	<i>Qualifier</i>						
d5-NEtFOSAA	102		70 - 130			10/06/25 01:42	10/06/25 15:43	1
13C2 PFHxA	117		70 - 130			10/06/25 01:42	10/06/25 15:43	1
13C2 PFDA	121		70 - 130			10/06/25 01:42	10/06/25 15:43	1

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: MBL 380-177960/25-A
Matrix: Water
Analysis Batch: 178050

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 177960

<i>Surrogate</i>	<i>MBL</i>	<i>MBL</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C3-GenX	117	Qualifier	70 - 130	10/06/25 01:42	10/06/25 15:43	1

Lab Sample ID: LCS 380-177960/27-A
Matrix: Water
Analysis Batch: 178050

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 177960

<i>Analyte</i>	<i>Spike Added</i>	<i>LCS Result</i>	<i>LCS Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	25.1	27.0	Qualifier	ng/L	-	108	70 - 130
Perfluorooctanesulfonic acid (PFOS)	25.1	24.8	Qualifier	ng/L	-	99	70 - 130
Perfluoroundecanoic acid (PFUnA)	25.1	28.4	Qualifier	ng/L	-	113	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	25.1	24.8	Qualifier	ng/L	-	99	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	25.1	24.8	Qualifier	ng/L	-	99	70 - 130
Perfluorohexanoic acid (PFHxA)	25.1	27.6	Qualifier	ng/L	-	110	70 - 130
Perfluorododecanoic acid (PFDoA)	25.1	28.5	Qualifier	ng/L	-	114	70 - 130
Perfluorooctanoic acid (PFOA)	25.1	26.2	Qualifier	ng/L	-	105	70 - 130
Perfluorodecanoic acid (PFDA)	25.1	28.8	Qualifier	ng/L	-	115	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	25.1	24.9	Qualifier	ng/L	-	100	70 - 130
Perfluorobutanesulfonic acid (PFBS)	25.1	24.9	Qualifier	ng/L	-	99	70 - 130
Perfluoroheptanoic acid (PFHpA)	25.1	28.2	Qualifier	ng/L	-	113	70 - 130
Perfluorononanoic acid (PFNA)	25.1	26.7	Qualifier	ng/L	-	107	70 - 130
Perfluorotetradecanoic acid (PFTA)	25.1	29.2	Qualifier	ng/L	-	117	70 - 130
Perfluorotridecanoic acid (PFTrDA)	25.1	31.8	Qualifier	ng/L	-	127	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	25.1	25.4	Qualifier	ng/L	-	101	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	25.1	25.1	Qualifier	ng/L	-	100	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	25.1	27.4	Qualifier	ng/L	-	109	70 - 130

<i>Surrogate</i>	<i>LCS %Recovery</i>	<i>LCS Qualifier</i>	<i>Limits</i>
d5-NEtFOSAA	97	Qualifier	70 - 130
13C2 PFHxA	116	Qualifier	70 - 130
13C2 PFDA	117	Qualifier	70 - 130
13C3-GenX	115	Qualifier	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: MRL 380-177960/26-A
Matrix: Water
Analysis Batch: 178050

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 177960

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	1.97	J	ng/L		98	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	2.14	J	ng/L		106	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	2.42	J	ng/L		120	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.01	1.90	J	ng/L		95	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.01	1.84	J	ng/L		92	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	2.05	J	ng/L		102	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	2.27	J	ng/L		113	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	2.09	J	ng/L		104	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	2.28	J	ng/L		114	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	2.01	J	ng/L		100	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	1.99	J	ng/L		99	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	2.22	J	ng/L		110	50 - 150
Perfluorononanoic acid (PFNA)	2.01	2.22	J	ng/L		110	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.01	2.46	J	ng/L		123	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.01	2.73	J	ng/L		136	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	2.01	2.06	J	ng/L		103	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	2.02	J	ng/L		100	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	2.09	J	ng/L		104	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	94		70 - 130
13C2 PFHxA	120		70 - 130
13C2 PFDA	122		70 - 130
13C3-GenX	103		70 - 130

Lab Sample ID: 380-174629-B-1-A MS
Matrix: Water
Analysis Batch: 178050

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 177960

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		25.2	27.3		ng/L		109	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		25.2	27.6		ng/L		103	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		25.2	29.0		ng/L		115	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.2	25.7		ng/L		102	70 - 130

Eurofins Eaton Analytical Pomona

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

LCMS

Prep Batch: 177960

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-174638-1	HALAWA SHAFT VIEWING POOL	Total/NA	Drinking Water	537.1 DW	
380-174638-2	HALAWA SHAFT VIEWING POOL BLANK	Total/NA	Water	537.1 DW	
MBL 380-177960/25-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-177960/27-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-177960/26-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-174629-B-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-174629-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

Prep Batch: 177966

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-174638-1	HALAWA SHAFT VIEWING POOL	Total/NA	Drinking Water	533	
380-174638-2	HALAWA SHAFT VIEWING POOL BLANK	Total/NA	Water	533	
MBL 380-177966/20-A	Method Blank	Total/NA	Water	533	
LCS 380-177966/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-177966/21-A	Lab Control Sample	Total/NA	Water	533	
380-174633-E-1-A MS	Matrix Spike	Total/NA	Water	533	
380-174633-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Analysis Batch: 178050

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-174638-1	HALAWA SHAFT VIEWING POOL	Total/NA	Drinking Water	537.1	177960
380-174638-2	HALAWA SHAFT VIEWING POOL BLANK	Total/NA	Water	537.1	177960
MBL 380-177960/25-A	Method Blank	Total/NA	Water	537.1	177960
LCS 380-177960/27-A	Lab Control Sample	Total/NA	Water	537.1	177960
MRL 380-177960/26-A	Lab Control Sample	Total/NA	Water	537.1	177960
380-174629-B-1-A MS	Matrix Spike	Total/NA	Water	537.1	177960
380-174629-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1	177960

Analysis Batch: 178159

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-174638-1	HALAWA SHAFT VIEWING POOL	Total/NA	Drinking Water	533	177966
380-174638-2	HALAWA SHAFT VIEWING POOL BLANK	Total/NA	Water	533	177966
MBL 380-177966/20-A	Method Blank	Total/NA	Water	533	177966
LCS 380-177966/22-A	Lab Control Sample	Total/NA	Water	533	177966
MRL 380-177966/21-A	Lab Control Sample	Total/NA	Water	533	177966
380-174633-E-1-A MS	Matrix Spike	Total/NA	Water	533	177966
380-174633-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	177966

Lab Chronicle

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-174638-1
 SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-174638-1

Date Collected: 09/30/25 09:45

Matrix: Drinking Water

Date Received: 10/02/25 10:06

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			177966	XTD8	EA POM	10/06/25 07:46
Total/NA	Analysis	533		1	178159	SZ9R	EA POM	10/07/25 10:07
Total/NA	Prep	537.1 DW			177960	G9MN	EA POM	10/06/25 01:42
Total/NA	Analysis	537.1		1	178050	M7ML	EA POM	10/06/25 18:25

Client Sample ID: HALAWA SHAFT VIEWING POOL BLANK

Lab Sample ID: 380-174638-2

Date Collected: 09/30/25 09:45

Matrix: Water

Date Received: 10/02/25 10:06

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			177966	XTD8	EA POM	10/06/25 07:46
Total/NA	Analysis	533		1	178159	SZ9R	EA POM	10/07/25 10:16
Total/NA	Prep	537.1 DW			177960	G9MN	EA POM	10/06/25 01:42
Total/NA	Analysis	537.1		1	178050	M7ML	EA POM	10/06/25 18:36

Laboratory References:

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Laboratory: Eurofins Eaton Analytical Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
537.1	Perfluorinated Alkyl Acids (LC/MS)	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-174638-1
SDG: PFAS: Halawa Shaft Viewing Pool

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-174638-1	HALAWA SHAFT VIEWING POOL	Drinking Water	09/30/25 09:45	10/02/25 10:06	Hawaii
380-174638-2	HALAWA SHAFT VIEWING POOL BLANK	Water	09/30/25 09:45	10/02/25 10:06	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-174638-1

SDG Number: PFAS: Halawa Shaft Viewing Pool

Login Number: 174638

List Number: 1

Creator: Sanchez, Joseph G

List Source: Eurofins Eaton Analytical Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

