

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

# ANALYTICAL REPORT

## PREPARED FOR

Attn: Mr. Erwin Kawata  
City & County of Honolulu  
630 South Beretania Street  
Public Service Bldg. Room 310  
Honolulu, Hawaii 96843

Generated 1/27/2025 4:19:23 PM

## JOB DESCRIPTION

RED-HILL  
Weekly PFAS

## JOB NUMBER

380-131733-1

# Eurofins Eaton Analytical Pomona

## Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

## Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

## Authorization



Generated  
1/27/2025 4:19:23 PM

Authorized for release by  
Rachelle Arada, Project Manager  
[Rachelle.Arada@et.eurofinsus.com](mailto:Rachelle.Arada@et.eurofinsus.com)  
(626)386-1106



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	3
Definitions/Glossary . . . . .	4
Case Narrative . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	7
Action Limit Summary . . . . .	11
Surrogate Summary . . . . .	12
Isotope Dilution Summary . . . . .	13
QC Sample Results . . . . .	14
QC Association Summary . . . . .	25
Lab Chronicle . . . . .	26
Certification Summary . . . . .	27
Method Summary . . . . .	28
Sample Summary . . . . .	29
Chain of Custody . . . . .	30
Receipt Checklists . . . . .	31

# Definitions/Glossary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Qualifiers

### LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

# Case Narrative

Client: City & County of Honolulu  
Project: RED-HILL

Job ID: 380-131733-1

**Job ID: 380-131733-1**

**Eurofins Eaton Analytical Pomona**

## Job Narrative 380-131733-1

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable.

- Matrix QC may not be reported if insufficient sample is provided or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

### Receipt

The samples were received on 1/23/2025 9:37 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 2.8°C.

### PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

# Detection Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

**Client Sample ID: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-131733-1**

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanesulfonic acid (PFHxS)	3.8		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.5		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.2		2.0	ng/L	1		537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.3		2.0	ng/L	1		537.1	Total/NA

**Client Sample ID: FB: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-131733-2**

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical Pomona



# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

**Client Sample ID: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-131733-1**

Date Collected: 01/21/25 10:00

Matrix: Water

Date Received: 01/23/25 09:37

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
<b>Perfluorohexanesulfonic acid (PFHxS)</b>	<b>3.8</b>		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>3.5</b>		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:26	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	80		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C6 PFDA	85		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C5 PFHxA	83		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C4 PFHpA	84		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C8 PFOA	85		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C9 PFNA	83		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C7 PFUnA	86		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C2 PFDoA	93		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C4 PFBA	95		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C5 PFPeA	90		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C3 PFBS	107		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C3 PFHxS	107		50 - 200	01/24/25 15:59	01/25/25 19:26	1

Eurofins Eaton Analytical Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

**Client Sample ID: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-131733-1**

Date Collected: 01/21/25 10:00

Matrix: Water

Date Received: 01/23/25 09:37

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)**

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	105		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C2-4:2-FTS	124		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C2-6:2-FTS	123		50 - 200	01/24/25 15:59	01/25/25 19:26	1
13C2-8:2-FTS	112		50 - 200	01/24/25 15:59	01/25/25 19:26	1

**Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>3.2</b>		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
N-methylperfluorooctanesulfonamide cetic acid (NMeFOSAA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
N-ethylperfluorooctanesulfonamide cetic acid (NEtFOSAA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
<b>Perfluorohexanesulfonic acid (PFHxS)</b>	<b>3.3</b>		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
11-Chloroeicosafuoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:04	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>			<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
d5-NEtFOSAA	93		70 - 130			01/24/25 02:15	01/24/25 15:04	1
13C2 PFHxA	101		70 - 130			01/24/25 02:15	01/24/25 15:04	1
13C2 PFDA	98		70 - 130			01/24/25 02:15	01/24/25 15:04	1
13C3-GenX	94		70 - 130			01/24/25 02:15	01/24/25 15:04	1

**Client Sample ID: FB: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-131733-2**

Date Collected: 01/21/25 10:00

Matrix: Water

Date Received: 01/23/25 09:37

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafuoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1

Eurofins Eaton Analytical Pomona



# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

**Client Sample ID: FB: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-131733-2**

Date Collected: 01/21/25 10:00

Matrix: Water

Date Received: 01/23/25 09:37

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		01/24/25 15:59	01/25/25 19:36	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	94		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C6 PFDA	100		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C5 PFHxA	98		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C4 PFHpA	107		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C8 PFOA	103		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C9 PFNA	101		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C7 PFUnA	96		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C2 PFDoA	98		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C4 PFBA	108		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C5 PFPeA	106		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C3 PFBS	106		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C3 PFHxS	108		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C8 PFOS	107		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C2-4:2-FTS	114		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C2-6:2-FTS	118		50 - 200	01/24/25 15:59	01/25/25 19:36	1
13C2-8:2-FTS	109		50 - 200	01/24/25 15:59	01/25/25 19:36	1

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

**Client Sample ID: FB: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-131733-2**

Date Collected: 01/21/25 10:00

Matrix: Water

Date Received: 01/23/25 09:37

**Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
N-methylperfluorooctanesulfonamideacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
N-ethylperfluorooctanesulfonamideacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/24/25 02:15	01/24/25 15:43	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>			<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
d5-NEtFOSAA	93		70 - 130			01/24/25 02:15	01/24/25 15:43	1
13C2 PFHxA	96		70 - 130			01/24/25 02:15	01/24/25 15:43	1
13C2 PFDA	102		70 - 130			01/24/25 02:15	01/24/25 15:43	1
13C3-GenX	100		70 - 130			01/24/25 02:15	01/24/25 15:43	1

# Action Limit Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

**Client Sample ID: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-131733-1**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.8		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.5		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.2		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.3		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

**Client Sample ID: FB: HALAWA SHAFT VIEWING POOL**

**Lab Sample ID: 380-131733-2**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

# Surrogate Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-131733-1  
 SDG: Weekly PFAS

**Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)**

**Matrix: Water**

**Prep Type: Total/NA**

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-131733-1	HALAWA SHAFT VIEWING POC	93	101	98	94
380-131733-1 MS	HALAWA SHAFT VIEWING POOL	97	97	102	91
380-131733-1 MSD	HALAWA SHAFT VIEWING POOL	97	91	97	89
380-131733-2	FB: HALAWA SHAFT VIEWING POOL	93	96	102	100
LCS 380-130172/22-A	Lab Control Sample	95	96	95	97
MBL 380-130172/20-A	Method Blank	100	97	102	91
MRL 380-130172/21-A	Lab Control Sample	103	100	100	93

**Surrogate Legend**

d5NEFOS = d5-NEtFOSAA

PFHxA = 13C2 PFHxA

PFDA = 13C2 PFDA

GenX = 13C3-GenX



# Isotope Dilution Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDoA (50-200)
380-131668-B-2-A MS	Matrix Spike	98	104	103	107	106	99	105	102
380-131668-C-2-A MSD	Matrix Spike Duplicate	100	101	104	108	100	103	99	109
380-131733-1	HALAWA SHAFT VIEWING POOL	80	85	83	84	85	83	86	93
380-131733-2	FB: HALAWA SHAFT VIEWING POOL	94	100	98	107	103	101	96	98
LCS 380-130338/23-A	Lab Control Sample	99	101	103	106	101	102	102	107
MBL 380-130338/21-A	Method Blank	100	106	104	107	106	106	107	115
MRL 380-130338/22-A	Lab Control Sample	100	101	103	105	103	100	100	106

  

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-131668-B-2-A MS	Matrix Spike	107	109	104	101	103	119	124	110
380-131668-C-2-A MSD	Matrix Spike Duplicate	106	113	105	105	103	118	124	108
380-131733-1	HALAWA SHAFT VIEWING POOL	95	90	107	107	105	124	123	112
380-131733-2	FB: HALAWA SHAFT VIEWING POOL	108	106	106	108	107	114	118	109
LCS 380-130338/23-A	Lab Control Sample	109	112	106	105	102	116	125	109
MBL 380-130338/21-A	Method Blank	108	109	108	106	108	122	123	113
MRL 380-130338/22-A	Lab Control Sample	106	109	107	105	107	119	124	105

**Surrogate Legend**

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDoA = 13C2 PFDoA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

**Lab Sample ID: MBL 380-130338/21-A**  
**Matrix: Water**  
**Analysis Batch: 130434**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 130338**

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.25		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		01/24/25 15:59	01/25/25 17:19	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	100		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C6 PFDA	106		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C5 PFHxA	104		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C4 PFHpA	107		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C8 PFOA	106		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C9 PFNA	106		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C7 PFUnA	107		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C2 PFDoA	115		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C4 PFBA	108		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C5 PFPeA	109		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C3 PFBS	108		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C3 PFHxS	106		50 - 200	01/24/25 15:59	01/25/25 17:19	1

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: MBL 380-130338/21-A**  
**Matrix: Water**  
**Analysis Batch: 130434**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 130338**

<i>Isotope Dilution</i>	<i>MBL %Recovery</i>	<i>MBL Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C8 PFOS	108		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C2-4:2-FTS	122		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C2-6:2-FTS	123		50 - 200	01/24/25 15:59	01/25/25 17:19	1
13C2-8:2-FTS	113		50 - 200	01/24/25 15:59	01/25/25 17:19	1

**Lab Sample ID: LCS 380-130338/23-A**  
**Matrix: Water**  
**Analysis Batch: 130434**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 130338**

<i>Analyte</i>	<i>Spike Added</i>	<i>LCS Result</i>	<i>LCS Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	60.1	63.1		ng/L		105	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	60.1	64.4		ng/L		107	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	60.1	59.3		ng/L		99	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	60.1	62.9		ng/L		105	70 - 130
Perfluorobutanesulfonic acid (PFBS)	60.1	65.1		ng/L		108	70 - 130
Perfluorodecanoic acid (PFDA)	60.1	63.6		ng/L		106	70 - 130
Perfluorododecanoic acid (PFDoA)	60.1	61.4		ng/L		102	70 - 130
Perfluoroheptanoic acid (PFHpA)	60.1	61.1		ng/L		102	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	60.1	60.9		ng/L		101	70 - 130
Perfluorohexanoic acid (PFHxA)	60.1	62.3		ng/L		104	70 - 130
Perfluorononanoic acid (PFNA)	60.1	60.4		ng/L		100	70 - 130
Perfluorooctanesulfonic acid (PFOS)	60.1	61.1		ng/L		102	70 - 130
Perfluorooctanoic acid (PFOA)	60.1	63.4		ng/L		105	70 - 130
Perfluoroundecanoic acid (PFUnA)	60.1	62.0		ng/L		103	70 - 130
Perfluorobutanoic acid (PFBA)	60.1	61.5		ng/L		102	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	60.1	66.6		ng/L		111	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	60.1	62.3		ng/L		104	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	60.1	62.5		ng/L		104	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	60.1	64.7		ng/L		108	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	60.1	60.4		ng/L		100	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	60.1	61.1		ng/L		102	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	60.1	61.9		ng/L		103	70 - 130
Perfluoropentanoic acid (PFPeA)	60.1	62.5		ng/L		104	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	60.1	61.7		ng/L		103	70 - 130

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: LCS 380-130338/23-A**  
**Matrix: Water**  
**Analysis Batch: 130434**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 130338**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	60.1	62.3		ng/L		104	70 - 130
<b>LCS LCS</b>							
<b>Isotope Dilution</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				
13C3 HFPO-DA	99		50 - 200				
13C6 PFDA	101		50 - 200				
13C5 PFHxA	103		50 - 200				
13C4 PFHpA	106		50 - 200				
13C8 PFOA	101		50 - 200				
13C9 PFNA	102		50 - 200				
13C7 PFUnA	102		50 - 200				
13C2 PFDoA	107		50 - 200				
13C4 PFBA	109		50 - 200				
13C5 PFPeA	112		50 - 200				
13C3 PFBS	106		50 - 200				
13C3 PFHxS	105		50 - 200				
13C8 PFOS	102		50 - 200				
13C2-4:2-FTS	116		50 - 200				
13C2-6:2-FTS	125		50 - 200				
13C2-8:2-FTS	109		50 - 200				

**Lab Sample ID: MRL 380-130338/22-A**  
**Matrix: Water**  
**Analysis Batch: 130434**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 130338**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	1.98	J	ng/L		99	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	2.12	J	ng/L		106	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.18	J	ng/L		109	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	2.16	J	ng/L		108	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	2.21	J	ng/L		110	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.28	J	ng/L		114	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.27	J	ng/L		113	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.28	J	ng/L		114	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.13	J	ng/L		106	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.12	J	ng/L		106	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.23	J	ng/L		111	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.05	J	ng/L		102	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.20	J	ng/L		110	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.13	J	ng/L		106	50 - 150
Perfluorobutanoic acid (PFBA)	2.00	2.38	J	ng/L		119	50 - 150

Eurofins Eaton Analytical Pomona



# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: MRL 380-130338/22-A**  
**Matrix: Water**  
**Analysis Batch: 130434**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 130338**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.00	2.27	J	ng/L		113	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.00	2.21	J	ng/L		110	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.00	2.34	J	ng/L		117	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.00	2.09	J	ng/L		104	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.00	2.05	J	ng/L		102	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.00	2.11	J	ng/L		105	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.00	2.12	J	ng/L		106	50 - 150
Perfluoropentanoic acid (PFPeA)	2.00	2.14	J	ng/L		107	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.00	2.04	J	ng/L		102	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.00	2.13	J	ng/L		106	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	100		50 - 200
13C6 PFDA	101		50 - 200
13C5 PFHxA	103		50 - 200
13C4 PFHpA	105		50 - 200
13C8 PFOA	103		50 - 200
13C9 PFNA	100		50 - 200
13C7 PFUnA	100		50 - 200
13C2 PFDoA	106		50 - 200
13C4 PFBA	106		50 - 200
13C5 PFPeA	109		50 - 200
13C3 PFBS	107		50 - 200
13C3 PFHxS	105		50 - 200
13C8 PFOS	107		50 - 200
13C2-4:2-FTS	119		50 - 200
13C2-6:2-FTS	124		50 - 200
13C2-8:2-FTS	105		50 - 200

**Lab Sample ID: 380-131668-B-2-A MS**  
**Matrix: Water**  
**Analysis Batch: 130434**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 130338**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.2	62.3		ng/L		103	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.2	60.1		ng/L		100	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.2	59.6		ng/L		99	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: 380-131668-B-2-A MS**  
**Matrix: Water**  
**Analysis Batch: 130434**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 130338**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		60.2	61.3		ng/L		102	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	6.7		60.2	66.9		ng/L		100	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		60.2	61.5		ng/L		102	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		60.2	61.3		ng/L		102	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		60.2	62.1		ng/L		101	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	5.9		60.2	69.9		ng/L		106	70 - 130
Perfluorohexanoic acid (PFHxA)	2.6		60.2	62.7		ng/L		100	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		60.2	65.7		ng/L		108	70 - 130
Perfluorooctanesulfonic acid (PFOS)	7.3		60.2	67.5		ng/L		100	70 - 130
Perfluorooctanoic acid (PFOA)	3.3		60.2	64.7		ng/L		102	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		60.2	58.6		ng/L		97	70 - 130
Perfluorobutanoic acid (PFBA)	5.9		60.2	67.4		ng/L		102	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.2	62.9		ng/L		104	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.2	62.5		ng/L		104	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.2	63.4		ng/L		105	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.2	63.7		ng/L		106	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.2	59.3		ng/L		98	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.2	61.7		ng/L		102	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.2	63.0		ng/L		105	70 - 130
Perfluoropentanoic acid (PFPeA)	3.0		60.2	66.7		ng/L		106	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.2	61.6		ng/L		102	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.2	64.6		ng/L		106	70 - 130

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	98		50 - 200
13C6 PFDA	104		50 - 200
13C5 PFHxA	103		50 - 200
13C4 PFHpA	107		50 - 200
13C8 PFOA	106		50 - 200
13C9 PFNA	99		50 - 200
13C7 PFUnA	105		50 - 200
13C2 PFDoA	102		50 - 200
13C4 PFBA	107		50 - 200
13C5 PFPeA	109		50 - 200
13C3 PFBS	104		50 - 200
13C3 PFHxS	101		50 - 200
13C8 PFOS	103		50 - 200

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: 380-131668-B-2-A MS**  
**Matrix: Water**  
**Analysis Batch: 130434**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 130338**

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>MS MS Qualifier</i>	<i>Limits</i>
13C2-4:2-FTS	119		50 - 200
13C2-6:2-FTS	124		50 - 200
13C2-8:2-FTS	110		50 - 200

**Lab Sample ID: 380-131668-C-2-A MSD**  
**Matrix: Water**  
**Analysis Batch: 130434**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**  
**Prep Batch: 130338**

<b>Analyte</b>	<b>Sample Result</b>	<b>Sample Qualifier</b>	<b>Spike Added</b>	<b>MSD Result</b>	<b>MSD Qualifier</b>	<b>Unit</b>	<b>D</b>	<b>%Rec</b>	<b>%Rec Limits</b>	<b>RPD</b>	<b>RPD Limit</b>
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.2	60.6		ng/L		101	70 - 130	3	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.2	62.6		ng/L		104	70 - 130	4	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.2	60.6		ng/L		101	70 - 130	2	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		60.2	63.9		ng/L		106	70 - 130	4	30
Perfluorobutanesulfonic acid (PFBS)	6.7		60.2	68.4		ng/L		102	70 - 130	2	30
Perfluorodecanoic acid (PFDA)	<2.0		60.2	63.4		ng/L		105	70 - 130	3	30
Perfluorododecanoic acid (PFDoA)	<2.0		60.2	60.3		ng/L		100	70 - 130	2	30
Perfluoroheptanoic acid (PFHpA)	<2.0		60.2	62.7		ng/L		102	70 - 130	1	30
Perfluorohexanesulfonic acid (PFHxS)	5.9		60.2	67.7		ng/L		103	70 - 130	3	30
Perfluorohexanoic acid (PFHxA)	2.6		60.2	64.7		ng/L		103	70 - 130	3	30
Perfluorononanoic acid (PFNA)	<2.0		60.2	60.6		ng/L		99	70 - 130	8	30
Perfluorooctanesulfonic acid (PFOS)	7.3		60.2	67.2		ng/L		99	70 - 130	0	30
Perfluorooctanoic acid (PFOA)	3.3		60.2	67.1		ng/L		106	70 - 130	4	30
Perfluoroundecanoic acid (PFUnA)	<2.0		60.2	61.4		ng/L		102	70 - 130	5	30
Perfluorobutanoic acid (PFBA)	5.9		60.2	68.4		ng/L		104	70 - 130	2	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.2	62.0		ng/L		103	70 - 130	1	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.2	62.9		ng/L		104	70 - 130	1	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.2	62.8		ng/L		104	70 - 130	1	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.2	67.1		ng/L		111	70 - 130	5	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.2	59.8		ng/L		99	70 - 130	1	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.2	64.5		ng/L		107	70 - 130	4	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.2	60.8		ng/L		101	70 - 130	4	30
Perfluoropentanoic acid (PFPeA)	3.0		60.2	63.3		ng/L		100	70 - 130	5	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.2	60.7		ng/L		101	70 - 130	1	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.2	61.7		ng/L		101	70 - 130	5	30

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

<i>Isotope Dilution</i>	<i>MSD</i>	<i>MSD</i>	<i>Limits</i>
	<i>%Recovery</i>	<i>Qualifier</i>	
13C3 HFPO-DA	100		50 - 200
13C6 PFDA	101		50 - 200
13C5 PFHxA	104		50 - 200
13C4 PFHpA	108		50 - 200
13C8 PFOA	100		50 - 200
13C9 PFNA	103		50 - 200
13C7 PFUnA	99		50 - 200
13C2 PFDoA	109		50 - 200
13C4 PFBA	106		50 - 200
13C5 PFPeA	113		50 - 200
13C3 PFBS	105		50 - 200
13C3 PFHxS	105		50 - 200
13C8 PFOS	103		50 - 200
13C2-4:2-FTS	118		50 - 200
13C2-6:2-FTS	124		50 - 200
13C2-8:2-FTS	108		50 - 200

## Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

**Lab Sample ID: MBL 380-130172/20-A**  
**Matrix: Water**  
**Analysis Batch: 130290**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 130172**

<i>Analyte</i>	<i>MBL</i>	<i>MBL</i>	<i>RL</i>	<i>Unit</i>	<i>D</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
	<i>Result</i>	<i>Qualifier</i>						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
Perfluorotridecanoic acid (PFTrDA)	<0.36		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		01/24/25 02:15	01/24/25 14:33	1
<i>Surrogate</i>	<i>MBL</i>	<i>MBL</i>	<i>Limits</i>			<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
	<i>%Recovery</i>	<i>Qualifier</i>						
d5-NEtFOSAA	100		70 - 130			01/24/25 02:15	01/24/25 14:33	1
13C2 PFHxA	97		70 - 130			01/24/25 02:15	01/24/25 14:33	1
13C2 PFDA	102		70 - 130			01/24/25 02:15	01/24/25 14:33	1

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

**Lab Sample ID: MBL 380-130172/20-A**  
**Matrix: Water**  
**Analysis Batch: 130290**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 130172**

<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C3-GenX	91		70 - 130	01/24/25 02:15	01/24/25 14:33	1

**Lab Sample ID: LCS 380-130172/22-A**  
**Matrix: Water**  
**Analysis Batch: 130290**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 130172**

<i>Analyte</i>	<i>Spike Added</i>	<i>LCS Result</i>	<i>LCS Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>Limits</i>
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	25.1	21.6		ng/L		86	70 - 130
Perfluorooctanesulfonic acid (PFOS)	25.1	23.1		ng/L		92	70 - 130
Perfluoroundecanoic acid (PFUnA)	25.1	25.1		ng/L		100	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	25.1	23.1		ng/L		92	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	25.1	23.2		ng/L		93	70 - 130
Perfluorohexanoic acid (PFHxA)	25.1	22.8		ng/L		91	70 - 130
Perfluorododecanoic acid (PFDoA)	25.1	26.2		ng/L		105	70 - 130
Perfluorooctanoic acid (PFOA)	25.1	22.8		ng/L		91	70 - 130
Perfluorodecanoic acid (PFDA)	25.1	23.1		ng/L		92	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	25.1	23.5		ng/L		94	70 - 130
Perfluorobutanesulfonic acid (PFBS)	25.1	22.5		ng/L		90	70 - 130
Perfluoroheptanoic acid (PFHpA)	25.1	24.8		ng/L		99	70 - 130
Perfluorononanoic acid (PFNA)	25.1	22.3		ng/L		89	70 - 130
Perfluorotetradecanoic acid (PFTA)	25.1	22.4		ng/L		89	70 - 130
Perfluorotridecanoic acid (PFTrDA)	25.1	22.8		ng/L		91	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	25.1	23.6		ng/L		94	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	25.1	22.4		ng/L		89	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	25.1	22.3		ng/L		89	70 - 130

  

<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>
d5-NEtFOSAA	95		70 - 130
13C2 PFHxA	96		70 - 130
13C2 PFDA	95		70 - 130
13C3-GenX	97		70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

**Lab Sample ID: MRL 380-130172/21-A**  
**Matrix: Water**  
**Analysis Batch: 130290**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 130172**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	1.85	J	ng/L		92	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	1.87	J	ng/L		93	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	2.01	J	ng/L		100	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.01	1.99	J	ng/L		99	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.01	1.97	J	ng/L		98	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	2.01	J	ng/L		100	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	2.13	J	ng/L		106	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	1.83	J	ng/L		91	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	1.87	J	ng/L		93	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	1.93	J	ng/L		96	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	1.74	J	ng/L		87	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	2.04	J	ng/L		102	50 - 150
Perfluorononanoic acid (PFNA)	2.01	1.89	J	ng/L		94	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.01	1.99	J	ng/L		99	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.01	1.89	J	ng/L		94	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	2.01	1.90	J	ng/L		95	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.72	J	ng/L		86	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.80	J	ng/L		90	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	103		70 - 130
13C2 PFHxA	100		70 - 130
13C2 PFDA	100		70 - 130
13C3-GenX	93		70 - 130

**Lab Sample ID: 380-131733-1 MS**  
**Matrix: Water**  
**Analysis Batch: 130290**

**Client Sample ID: HALAWA SHAFT VIEWING POOL**  
**Prep Type: Total/NA**  
**Prep Batch: 130172**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		25.1	21.5		ng/L		86	70 - 130
Perfluorooctanesulfonic acid (PFOS)	3.2		25.1	26.3		ng/L		92	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		25.1	24.6		ng/L		98	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.1	22.4		ng/L		89	70 - 130

Eurofins Eaton Analytical Pomona



# QC Sample Results

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-131733-1  
 SDG: Weekly PFAS

## Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: 380-131733-1 MSD

Client Sample ID: HALAWA SHAFT VIEWING POOL

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 130290

Prep Batch: 130172

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Perfluorohexanesulfonic acid (PFHxS)	3.3		25.2	25.1		ng/L		87	70 - 130	2	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		25.2	22.5		ng/L		89	70 - 130	0	30
Perfluoroheptanoic acid (PFHpA)	<2.0		25.2	23.3		ng/L		90	70 - 130	1	30
Perfluorononanoic acid (PFNA)	<2.0		25.2	22.8		ng/L		91	70 - 130	2	30
Perfluorotetradecanoic acid (PFTA)	<2.0		25.2	22.2		ng/L		88	70 - 130	5	30
Perfluorotridecanoic acid (PFTTrDA)	<2.0		25.2	22.6		ng/L		90	70 - 130	1	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		25.2	22.9		ng/L		91	70 - 130	1	30
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		25.2	21.6		ng/L		86	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		25.2	20.8		ng/L		83	70 - 130	3	30
<b>Surrogate</b>	<b>%Recovery</b>	<b>MSD Qualifier</b>		<b>Limits</b>							
d5-NEtFOSAA	97			70 - 130							
13C2 PFHxA	91			70 - 130							
13C2 PFDA	97			70 - 130							
13C3-GenX	89			70 - 130							



# QC Association Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## LCMS

### Prep Batch: 130172

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-131733-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1 DW	
380-131733-2	FB: HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1 DW	
MBL 380-130172/20-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-130172/22-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-130172/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-131733-1 MS	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1 DW	
380-131733-1 MSD	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1 DW	

### Analysis Batch: 130290

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-131733-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1	130172
380-131733-2	FB: HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1	130172
MBL 380-130172/20-A	Method Blank	Total/NA	Water	537.1	130172
LCS 380-130172/22-A	Lab Control Sample	Total/NA	Water	537.1	130172
MRL 380-130172/21-A	Lab Control Sample	Total/NA	Water	537.1	130172
380-131733-1 MS	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1	130172
380-131733-1 MSD	HALAWA SHAFT VIEWING POOL	Total/NA	Water	537.1	130172

### Prep Batch: 130338

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-131733-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	
380-131733-2	FB: HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	
MBL 380-130338/21-A	Method Blank	Total/NA	Water	533	
LCS 380-130338/23-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-130338/22-A	Lab Control Sample	Total/NA	Water	533	
380-131668-B-2-A MS	Matrix Spike	Total/NA	Water	533	
380-131668-C-2-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

### Analysis Batch: 130434

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-131733-1	HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	130338
380-131733-2	FB: HALAWA SHAFT VIEWING POOL	Total/NA	Water	533	130338
MBL 380-130338/21-A	Method Blank	Total/NA	Water	533	130338
LCS 380-130338/23-A	Lab Control Sample	Total/NA	Water	533	130338
MRL 380-130338/22-A	Lab Control Sample	Total/NA	Water	533	130338
380-131668-B-2-A MS	Matrix Spike	Total/NA	Water	533	130338
380-131668-C-2-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	130338

# Lab Chronicle

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Client Sample ID: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-131733-1

Date Collected: 01/21/25 10:00

Matrix: Water

Date Received: 01/23/25 09:37

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			130338	E9PK	EA POM	01/24/25 15:59
Total/NA	Analysis	533		1	130434	Y7BM	EA POM	01/25/25 19:26
Total/NA	Prep	537.1 DW			130172	G9MN	EA POM	01/24/25 02:15
Total/NA	Analysis	537.1		1	130290	Y5FM	EA POM	01/24/25 15:04

## Client Sample ID: FB: HALAWA SHAFT VIEWING POOL

Lab Sample ID: 380-131733-2

Date Collected: 01/21/25 10:00

Matrix: Water

Date Received: 01/23/25 09:37

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			130338	E9PK	EA POM	01/24/25 15:59
Total/NA	Analysis	533		1	130434	Y7BM	EA POM	01/25/25 19:36
Total/NA	Prep	537.1 DW			130172	G9MN	EA POM	01/24/25 02:15
Total/NA	Analysis	537.1		1	130290	Y5FM	EA POM	01/24/25 15:43

### Laboratory References:

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

# Accreditation/Certification Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

## Laboratory: Eurofins Eaton Analytical Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-25

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

# Method Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
537.1	Perfluorinated Alkyl Acids (LC/MS)	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

**Protocol References:**

EPA = US Environmental Protection Agency

**Laboratory References:**

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



# Sample Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-131733-1  
SDG: Weekly PFAS

---

<u>Lab Sample ID</u>	<u>Client Sample ID</u>	<u>Matrix</u>	<u>Collected</u>	<u>Received</u>
380-131733-1	HALAWA SHAFT VIEWING POOL	Water	01/21/25 10:00	01/23/25 09:37
380-131733-2	FB: HALAWA SHAFT VIEWING POOL	Water	01/21/25 10:00	01/23/25 09:37

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

**Eurofins Eaton Analytical Pomona**

941 Corporate Center Drive  
 Pomona, CA 91768-2642  
 Phone (626) 386-1100

**Chain of Custody Record**

**eurofins** | Environment Testing

<b>Client Information</b>		Sampler: Ryan Greer		Lab PM: Arada, Rachele		Carrier Tracking No(s):		COC No: 380-28005-2757.1																																																						
Client Contact: Dr. Ron Fenstermacher		Phone: 808-748-5840		E-Mail: Rachele.Arada@et.eurofins.com		State of Origin: HI		Page: Page 1 of 1																																																						
Company: City & County of Honolulu		PWSID:		<b>Analysis Requested</b>						Job #:																																																				
Address: 630 South Beretania Street Chemistry Lab		Due Date Requested:		<table border="1"> <tr> <td>Field Filtered Sample (Yes or No)</td> <td>Perform MS/MSD (Yes or No)</td> <td>625.1, 626.1, SIM</td> <td>8016B_GRO_LL - (MOD) GRO</td> <td>8016B_DRO_LL_CS - HNL Ranges: C10-C24/C24-C36/C8-C16</td> <td>625.2_PREC - (MOD) 626plus Plus TICs</td> <td>637.1_DW_PREC - 637.1 Full List</td> <td>633 - All Analytes</td> <td rowspan="5">Total Number of containers</td> </tr> <tr> <td>City: Honolulu</td> <td>TAT Requested (days):</td> <td colspan="7">Preservation Codes: R - NaThioSO4 RA - NaThio/HK Q - Na2SO3 QA - Na2SO3/ Y - Trizma I - NH4 Acetate</td> </tr> <tr> <td>State, Zip: HI, 96843</td> <td>Compliance Project: <input type="checkbox"/> Yes <input type="checkbox"/> No</td> <td colspan="7">380-131733 COC</td> </tr> <tr> <td>Phone: 808-748-5091(Tel)</td> <td>PO #: C20525101 exp 05312023</td> <td colspan="7">Other:</td> </tr> <tr> <td>Email: RFENSTEMACHER@hbws.org</td> <td>WO #:</td> <td colspan="7">Special Instructions/Note:</td> </tr> </table>						Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	625.1, 626.1, SIM	8016B_GRO_LL - (MOD) GRO	8016B_DRO_LL_CS - HNL Ranges: C10-C24/C24-C36/C8-C16	625.2_PREC - (MOD) 626plus Plus TICs	637.1_DW_PREC - 637.1 Full List	633 - All Analytes	Total Number of containers	City: Honolulu	TAT Requested (days):	Preservation Codes: R - NaThioSO4 RA - NaThio/HK Q - Na2SO3 QA - Na2SO3/ Y - Trizma I - NH4 Acetate							State, Zip: HI, 96843	Compliance Project: <input type="checkbox"/> Yes <input type="checkbox"/> No	380-131733 COC							Phone: 808-748-5091(Tel)	PO #: C20525101 exp 05312023	Other:							Email: RFENSTEMACHER@hbws.org	WO #:	Special Instructions/Note:							Project Name: RED-HILL/HBWS Sites Event Desc: RUSH Weekly Red Hill		Project #: 38001111		SSOW#:		Site: Hawaii	
Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	625.1, 626.1, SIM	8016B_GRO_LL - (MOD) GRO							8016B_DRO_LL_CS - HNL Ranges: C10-C24/C24-C36/C8-C16	625.2_PREC - (MOD) 626plus Plus TICs	637.1_DW_PREC - 637.1 Full List	633 - All Analytes	Total Number of containers																																																
City: Honolulu	TAT Requested (days):	Preservation Codes: R - NaThioSO4 RA - NaThio/HK Q - Na2SO3 QA - Na2SO3/ Y - Trizma I - NH4 Acetate																																																												
State, Zip: HI, 96843	Compliance Project: <input type="checkbox"/> Yes <input type="checkbox"/> No	380-131733 COC																																																												
Phone: 808-748-5091(Tel)	PO #: C20525101 exp 05312023	Other:																																																												
Email: RFENSTEMACHER@hbws.org	WO #:	Special Instructions/Note:																																																												
<b>Sample Identification</b>		Sample Date		Sample Time		Sample Type (C=Comp, G=grab)		Matrix (W=water, S=solid, O=waste/oil, BT=Tissue, A=Air)		Preservation Code:																																																				
Halawa Shaft Viewing Pool		1/21/25		1000		Water		Water		R RA Q QA Y I																																																				
FB: Halawa Shaft Viewing Pool		1/21/25		1000		Water		Water		1 1																																																				
<b>Possible Hazard Identification</b>		<input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological		<b>Sample Disposal ( A fee may be assessed if samples are retained longer than 1 month)</b>																																																										
Deliverable Requested: I, II, III, IV, Other (specify)		Special Instructions/QC Requirements:		<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months																																																										
Empty Kit Relinquished by:		Date:		Time:		Method of Shipment: <b>FED Ex 7716 0684 8140</b>																																																								
Relinquished: [Redacted]		Date/Time: 1/22/25 1300		Company: HBWS		Received by: <i>[Signature]</i>		Date/Time: 01/23/25 09:37		Company: EEAP																																																				
Relinquished by:		Date/Time:		Company:		Received by:		Date/Time:		Company:																																																				
Custody Seals Intact: <input type="checkbox"/> Yes <input type="checkbox"/> No		Custody Seal No.:		Cooler Temperature(s) °C and Other Remarks: <b>(75N) 2.8°-0.0°=2.8° GEL-FROZEN</b>																																																										

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17

## Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-131733-1

SDG Number: Weekly PFAS

**Login Number: 131733**

**List Source: Eurofins Eaton Analytical Pomona**

**List Number: 1**

**Creator: Edrosa, Rey**

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

