

ANALYTICAL REPORT

Eurofins Eaton Monrovia
750 Royal Oaks Drive
Suite 100
Monrovia, CA 91016
Tel: (626)386-1100

Laboratory Job ID: 380-16794-1
Client Project/Site: RED-HILL

For:
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 308
Honolulu, Hawaii 96843

Attn: Mr. Erwin Kawata



Authorized for release by:
10/18/2022 9:30:52 PM

Rachelle Arada, Manager of Project Management
(626)386-1106
Rachelle.Arada@et.eurofinsus.com

LINKS

Review your project
results through



Have a Question?



Visit us at:

www.eurofinsus.com/Env

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW,Water matrices)



Rachelle Arada
Manager of Project Management
10/18/2022 9:30:52 PM





Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	10
Surrogate Summary	11
QC Sample Results	14
QC Association Summary	29
Lab Chronicle	31
Certification Summary	32
Method Summary	34
Sample Summary	35
Subcontract Data	36
Chain of Custody	84
Receipt Checklists	88

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Qualifiers

GC/MS Semi VOA

Qualifier	Qualifier Description
^3-	Reporting Limit Check Standard is outside acceptance limits, low biased.
^3+	Reporting Limit Check Standard is outside acceptance limits, high biased
F1	MS and/or MSD recovery exceeds control limits.
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

GC/MS Semi VOA TICs

Qualifier	Qualifier Description
J	Indicates an Estimated Value for TICs
N	Presumptive evidence of material.
T	Result is a tentatively identified compound (TIC) and an estimated value.

Subcontract

Qualifier	Qualifier Description
U	This analyte was not detected.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Job ID: 380-16794-1

Laboratory: Eurofins Eaton Monrovia

Narrative

Job Narrative 380-16794-1

Comments

No additional comments.

Receipt

The samples were received on 8/16/2022 5:24 PM. Unless otherwise noted below, the samples arrived in good condition, and where required, properly preserved and on ice. The temperature of the cooler at receipt was 4.6° C.

GC/MS Semi VOA

Method 525.2: MRL was below acceptance limits for Caffeine. Caffeine is not a target analyte for this project. Data is not reported. Halawa Wells Pump 1 (380-16794-1) and (MRL 380-14908/2-A)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

Subcontract non-Sister

See attached subcontract report.

Organic Prep

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page



Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Client Sample ID: Halawa Wells Pump 1

Lab Sample ID: 380-16794-1

No Detections.

Client Sample ID: Halawa Wells Pump 1 TB

Lab Sample ID: 380-16794-2

No Detections.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Monrovia

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Client Sample ID: Halawa Wells Pump 1

Lab Sample ID: 380-16794-1

Date Collected: 08/15/22 10:20

Matrix: Water

Date Received: 08/16/22 17:24

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
2,4'-DDD	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
2,4'-DDE	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
2,4'-DDT	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
2,4-Dinitrotoluene	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
2,6-Dinitrotoluene	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
4,4'-DDD	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
4,4'-DDE	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
4,4'-DDT	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Acenaphthene	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Acenaphthylene	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Acetochlor	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Alachlor	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
alpha-BHC	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
alpha-Chlordane	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Anthracene	ND		0.020	ug/L		08/25/22 10:21	08/29/22 11:16	1
Atrazine	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Benz(a)anthracene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Benzo[a]pyrene	ND		0.020	ug/L		08/25/22 10:21	08/29/22 11:16	1
Benzo[b]fluoranthene	ND		0.020	ug/L		08/25/22 10:21	08/29/22 11:16	1
Benzo[g,h,i]perylene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Benzo[k]fluoranthene	ND		0.020	ug/L		08/25/22 10:21	08/29/22 11:16	1
beta-BHC	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Bromacil	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Butachlor	ND	^3+	0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Butylbenzylphthalate	ND		0.49	ug/L		08/25/22 10:21	08/29/22 11:16	1
Chlorobenzilate	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Chloroneb	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Chlorothalonil (Draconil, Bravo)	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Chlorpyrifos	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Chrysene	ND		0.020	ug/L		08/25/22 10:21	08/29/22 11:16	1
delta-BHC	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Di(2-ethylhexyl)adipate	ND		0.59	ug/L		08/25/22 10:21	08/29/22 11:16	1
Bis(2-ethylhexyl) phthalate	ND		0.59	ug/L		08/25/22 10:21	08/29/22 11:16	1
Diazinon (Qualitative)	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Dibenz(a,h)anthracene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Diclorvos (DDVP)	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Dieldrin	ND		0.20	ug/L		08/25/22 10:21	08/29/22 11:16	1
Diethylphthalate	ND		0.49	ug/L		08/25/22 10:21	08/29/22 11:16	1
Dimethoate	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Dimethylphthalate	ND		0.49	ug/L		08/25/22 10:21	08/29/22 11:16	1
Di-n-butyl phthalate	ND		0.98	ug/L		08/25/22 10:21	08/29/22 11:16	1
Di-n-octyl phthalate	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Endosulfan I (Alpha)	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Endosulfan II (Beta)	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Endosulfan sulfate	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Endrin	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Endrin aldehyde	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
EPTC	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Fluoranthene	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1

Eurofins Eaton Monrovia

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Client Sample ID: Halawa Wells Pump 1

Lab Sample ID: 380-16794-1

Date Collected: 08/15/22 10:20

Matrix: Water

Date Received: 08/16/22 17:24

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluorene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
gamma-Chlordane	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Heptachlor	ND		0.039	ug/L		08/25/22 10:21	08/29/22 11:16	1
Heptachlor epoxide (isomer B)	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Hexachlorobenzene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Hexachlorocyclopentadiene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Indeno[1,2,3-cd]pyrene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Isophorone	ND		0.49	ug/L		08/25/22 10:21	08/29/22 11:16	1
Lindane	ND		0.039	ug/L		08/25/22 10:21	08/29/22 11:16	1
Malathion	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Methoxychlor	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Metolachlor	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Metribuzin	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Molinate	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Naphthalene	ND		0.29	ug/L		08/25/22 10:21	08/29/22 11:16	1
Parathion	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Pendimethalin (Penoxaline)	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Total Permethrin (mixed isomers)	ND		0.20	ug/L		08/25/22 10:21	08/29/22 11:16	1
Phenanthrene	ND		0.039	ug/L		08/25/22 10:21	08/29/22 11:16	1
Propachlor	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Pyrene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Simazine	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Terbacil	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Terbuthylazine	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
Thiobencarb	ND		0.20	ug/L		08/25/22 10:21	08/29/22 11:16	1
trans-Nonachlor	ND		0.049	ug/L		08/25/22 10:21	08/29/22 11:16	1
Trifluralin	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
1-Methylnaphthalene	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1
2-Methylnaphthalene	ND		0.098	ug/L		08/25/22 10:21	08/29/22 11:16	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L				08/25/22 10:21	08/29/22 11:16	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	90		70 - 130	08/25/22 10:21	08/29/22 11:16	1
Triphenylphosphate	92		70 - 130	08/25/22 10:21	08/29/22 11:16	1
Perylene-d12	90		70 - 130	08/25/22 10:21	08/29/22 11:16	1

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
1-Methylphenanthrene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
2,3,5-Trimethylnaphthalene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
2,6-Dimethylnaphthalene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
2-Methylnaphthalene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Acenaphthene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Acenaphthylene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Anthracene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Benz[a]anthracene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Benzo[a]pyrene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1

Eurofins Eaton Monrovia

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Client Sample ID: Halawa Wells Pump 1

Lab Sample ID: 380-16794-1

Date Collected: 08/15/22 10:20

Matrix: Water

Date Received: 08/16/22 17:24

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i (Continued)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Benzo[b]fluoranthene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Benzo[e]pyrene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Benzo[g,h,i]perylene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Benzo[k]fluoranthene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Biphenyl	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Chrysene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Dibenz[a,h]anthracene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Dibenzo[a,i]pyrene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Dibenzothiophene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Disalicylidenepropanediamine	ND		0.1	0.05	µg/L		08/17/22 00:00	08/21/22 15:07	1
Fluoranthene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Fluorene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Indeno[1,2,3-cd]pyrene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Naphthalene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Perylene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Phenanthrene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1
Pyrene	ND		0.005	0.001	µg/L		08/17/22 00:00	08/21/22 15:07	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
(d10-Acenaphthene)	80		45 - 118	08/17/22 00:00	08/21/22 15:07	1
(d10-Phenanthrene)	89		56 - 123	08/17/22 00:00	08/21/22 15:07	1
(d12-Chrysene)	90		36 - 142	08/17/22 00:00	08/21/22 15:07	1
(d12-Perylene)	83		36 - 161	08/17/22 00:00	08/21/22 15:07	1
(d8-Naphthalene)	79		20 - 112	08/17/22 00:00	08/21/22 15:07	1

Method: 8015 Diesel LL (EAL) and Motor Oil - 8015 - TPH DRO/ORO

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
GASOLINE	ND	U	0.02		mg/L			08/23/22 15:25	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
BROMOFLUOROBENZENE	92		60 - 140		08/23/22 15:25	1

Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
DIESEL	ND	U	0.027		mg/L			08/22/22 23:09	1
MOTOR OIL	ND	U	0.053		mg/L			08/22/22 23:09	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
BROMOBENZENE	68		60 - 130		08/22/22 23:09	1
HEXACOSANE	81		60 - 130		08/22/22 23:09	1

Client Sample ID: Halawa Wells Pump 1 TB

Lab Sample ID: 380-16794-2

Date Collected: 08/15/22 10:20

Matrix: Water

Date Received: 08/16/22 17:24

Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
GASOLINE	ND	U	0.02		mg/L			08/23/22 16:01	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
BROMOFLUOROBENZENE	89		60 - 140		08/23/22 16:01	1

Eurofins Eaton Monrovia

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Client Sample ID: Halawa Wells Pump 1

Lab Sample ID: 380-16794-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Alachlor	ND		ug/L	2	0.049	525.2	Total/NA
Atrazine	ND		ug/L	3	0.049	525.2	Total/NA
Benzo[a]pyrene	ND		ug/L	0.2	0.020	525.2	Total/NA
Di(2-ethylhexyl)adipate	ND		ug/L	400	0.59	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	ND		ug/L	6	0.59	525.2	Total/NA
Endrin	ND		ug/L	2	0.098	525.2	Total/NA
Heptachlor	ND		ug/L	0.4	0.039	525.2	Total/NA
Heptachlor epoxide (isomer B)	ND		ug/L	0.2	0.049	525.2	Total/NA
Hexachlorobenzene	ND		ug/L	1	0.049	525.2	Total/NA
Hexachlorocyclopentadiene	ND		ug/L	50	0.049	525.2	Total/NA
Lindane	ND		ug/L	0.2	0.039	525.2	Total/NA
Methoxychlor	ND		ug/L	40	0.098	525.2	Total/NA
Simazine	ND		ug/L	4	0.049	525.2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	TPP (70-130)	PRY (70-130)
380-16460-T-1-A DU	Duplicate	90	91	89
380-16794-1	Halawa Wells Pump 1	90	92	90
380-16794-1 MS	Halawa Wells Pump 1	94	92	93
LCS 380-14908/3-A	Lab Control Sample	93	91	90
LCSD 380-14908/4-A	Lab Control Sample Dup	93	94	91
MB 380-14908/1-A	Method Blank	90	93	87
MRL 380-14908/2-A	Lab Control Sample	92	92	85

Surrogate Legend

2NMX = 2-Nitro-m-xylene

TPP = Triphenylphosphate

PRY = Perylene-d12

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i

Matrix: water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)				
		ANT (65-113)	CRY (60-139)	NPT (44-119)	PHN (80-111)	PRY (36-161)
99233-B1	Method Blank	92	92	88	93	88
99233-BS1	Lab Control Sample	87	84	81	91	86
99233-BS2	Lab Control Sample Dup	92	88	90	98	95

Surrogate Legend

ANT = (d10-Acenaphthene)

CRY = (d12-Chrysene)

NPT = (d8-Naphthalene)

PHN = (d10-Phenanthrene)

PRY = (d12-Perylene)

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)				
		ANT (45-118)	CRY (36-142)	NPT (20-112)	PHN (56-123)	PRY (36-161)
380-16794-1	Halawa Wells Pump 1	80	90	79	89	83

Surrogate Legend

ANT = (d10-Acenaphthene)

CRY = (d12-Chrysene)

NPT = (d8-Naphthalene)

PHN = (d10-Phenanthrene)

PRY = (d12-Perylene)

Method: 8015 Diesel LL (EAL) and Motor Oil - 8015 - TPH DRO/ORO

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)
		BFB (60-140)
380-16794-1	Halawa Wells Pump 1	92

Eurofins Eaton Monrovia

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Surrogate Legend

BFB = BROMOFLUOROBENZENE

Method: 8015 Diesel LL (EAL) and Motor Oil - 8015 - TPH DRO/ORO

Matrix: WATER

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

BFB

Lab Sample ID	Client Sample ID
22VGH7H09B	Method Blank

Surrogate Legend

BFB = BROMOFLUOROBENZENE

Method: 8015 Diesel LL (EAL) and Motor Oil - 8015 - TPH DRO/ORO

Matrix: WATER

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

BFB

Lab Sample ID	Client Sample ID	(70-130)
22VGH7H09C	LCD	118
22VGH7H09L	Lab Control Sample	118

Surrogate Legend

BFB = BROMOFLUOROBENZENE

Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

BB :XACOSAI

Lab Sample ID	Client Sample ID	(60-130)	(60-130)
380-16794-1	Halawa Wells Pump 1	68	81

Surrogate Legend

BB = BROMOBENZENE

HEXACOSANE = HEXACOSANE

Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

BFB

Lab Sample ID	Client Sample ID	(60-140)
380-16794-2	Halawa Wells Pump 1 TB	89

Surrogate Legend

BFB = BROMOFLUOROBENZENE

Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

Matrix: WATER

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

BB :XACOSAI

Lab Sample ID	Client Sample ID
22DSH034WB	Method Blank

Surrogate Legend

BB = BROMOBENZENE

HEXACOSANE = HEXACOSANE

Eurofins Eaton Monrovia

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

Matrix: WATER

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BB (60-130)	HEXACOSANE (60-130)
22DSH034WL	Lab Control Sample	75	93

Surrogate Legend

BB = BROMOBENZENE

HEXACOSANE = HEXACOSANE

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 380-14908/1-A
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 14908

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
2,4'-DDD	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
2,4'-DDE	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
2,4'-DDT	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
2,4-Dinitrotoluene	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
2,6-Dinitrotoluene	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
4,4'-DDD	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
4,4'-DDE	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
4,4'-DDT	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Acenaphthene	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Acenaphthylene	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Acetochlor	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Alachlor	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
alpha-BHC	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
alpha-Chlordane	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Anthracene	ND		0.020	ug/L		08/25/22 10:21	08/29/22 10:55	1
Atrazine	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Benz(a)anthracene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Benzo[a]pyrene	ND		0.020	ug/L		08/25/22 10:21	08/29/22 10:55	1
Benzo[b]fluoranthene	ND		0.020	ug/L		08/25/22 10:21	08/29/22 10:55	1
Benzo[g,h,i]perylene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Benzo[k]fluoranthene	ND		0.020	ug/L		08/25/22 10:21	08/29/22 10:55	1
beta-BHC	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Bromacil	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Butachlor	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Butylbenzylphthalate	ND		0.49	ug/L		08/25/22 10:21	08/29/22 10:55	1
Chlorobenzilate	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Chloroneb	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Chlorothalonil (Draconil, Bravo)	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Chlorpyrifos	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Chrysene	ND		0.020	ug/L		08/25/22 10:21	08/29/22 10:55	1
delta-BHC	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Di(2-ethylhexyl)adipate	ND		0.59	ug/L		08/25/22 10:21	08/29/22 10:55	1
Bis(2-ethylhexyl) phthalate	ND		0.59	ug/L		08/25/22 10:21	08/29/22 10:55	1
Diazinon (Qualitative)	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Dibenz(a,h)anthracene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Diclorvos (DDVP)	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Dieldrin	ND		0.20	ug/L		08/25/22 10:21	08/29/22 10:55	1
Diethylphthalate	ND		0.49	ug/L		08/25/22 10:21	08/29/22 10:55	1
Dimethoate	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Dimethylphthalate	ND		0.49	ug/L		08/25/22 10:21	08/29/22 10:55	1
Di-n-butyl phthalate	ND		0.99	ug/L		08/25/22 10:21	08/29/22 10:55	1
Di-n-octyl phthalate	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Endosulfan I (Alpha)	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Endosulfan II (Beta)	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Endosulfan sulfate	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Endrin	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Endrin aldehyde	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
EPTC	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MB 380-14908/1-A
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 14908

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluoranthene	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Fluorene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
gamma-Chlordane	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Heptachlor	ND		0.039	ug/L		08/25/22 10:21	08/29/22 10:55	1
Heptachlor epoxide (isomer B)	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Hexachlorobenzene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Hexachlorocyclopentadiene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Indeno[1,2,3-cd]pyrene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Isophorone	ND		0.49	ug/L		08/25/22 10:21	08/29/22 10:55	1
Lindane	ND		0.039	ug/L		08/25/22 10:21	08/29/22 10:55	1
Malathion	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Methoxychlor	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Metolachlor	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Metribuzin	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Molinate	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Naphthalene	ND		0.30	ug/L		08/25/22 10:21	08/29/22 10:55	1
Parathion	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Pendimethalin (Penoxaline)	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Total Permethrin (mixed isomers)	ND		0.20	ug/L		08/25/22 10:21	08/29/22 10:55	1
Phenanthrene	ND		0.039	ug/L		08/25/22 10:21	08/29/22 10:55	1
Propachlor	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Pyrene	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Simazine	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Terbacil	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Terbutylazine	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
Thiobencarb	ND		0.20	ug/L		08/25/22 10:21	08/29/22 10:55	1
trans-Nonachlor	ND		0.049	ug/L		08/25/22 10:21	08/29/22 10:55	1
Trifluralin	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
1-Methylnaphthalene	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1
2-Methylnaphthalene	ND		0.099	ug/L		08/25/22 10:21	08/29/22 10:55	1

<i>Tentatively Identified Compound</i>	MB Est. Result	MB Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
<i>n-Hexadecanoic acid</i>	0.632	T J N	ug/L		5.92	57-10-3	08/25/22 10:21	08/29/22 10:55	1
<i>9-Octadecenoic acid, (E)-</i>	0.832	T J N	ug/L		6.56	112-79-8	08/25/22 10:21	08/29/22 10:55	1
<i>Octadecanoic acid</i>	0.566	T J N	ug/L		6.63	57-11-4	08/25/22 10:21	08/29/22 10:55	1
<i>9-Octadecenamamide, (Z)-</i>	0.907	T J N	ug/L		7.67	301-02-0	08/25/22 10:21	08/29/22 10:55	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	90		70 - 130	08/25/22 10:21	08/29/22 10:55	1
Triphenylphosphate	93		70 - 130	08/25/22 10:21	08/29/22 10:55	1
Perylene-d12	87		70 - 130	08/25/22 10:21	08/29/22 10:55	1

Lab Sample ID: LCS 380-14908/3-A
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
2,4'-DDD	1.97	1.99		ug/L		101	70 - 130

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-14908/3-A
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
2,4'-DDE	1.97	2.08		ug/L		106	70 - 130
2,4'-DDT	1.97	2.08		ug/L		105	70 - 130
2,4-Dinitrotoluene	1.97	1.96		ug/L		99	70 - 130
2,6-Dinitrotoluene	1.97	1.87		ug/L		95	70 - 130
4,4'-DDD	1.97	2.11		ug/L		107	70 - 130
4,4'-DDE	1.97	1.86		ug/L		94	70 - 130
4,4'-DDT	1.97	2.04		ug/L		104	70 - 130
Acenaphthene	1.97	1.88		ug/L		95	70 - 130
Acenaphthylene	1.97	1.84		ug/L		93	70 - 130
Acetochlor	1.97	2.08		ug/L		105	70 - 130
Alachlor	1.97	2.03		ug/L		103	70 - 130
alpha-BHC	1.97	2.00		ug/L		101	70 - 130
alpha-Chlordane	1.97	1.75		ug/L		89	70 - 130
Anthracene	1.97	1.96		ug/L		99	70 - 130
Atrazine	1.97	2.15		ug/L		109	70 - 130
Benz(a)anthracene	1.97	2.04		ug/L		104	70 - 130
Benzo[a]pyrene	1.97	2.12		ug/L		108	70 - 130
Benzo[b]fluoranthene	1.97	2.22		ug/L		113	70 - 130
Benzo[g,h,i]perylene	1.97	1.95		ug/L		99	70 - 130
Benzo[k]fluoranthene	1.97	2.30		ug/L		117	70 - 130
beta-BHC	1.97	2.00		ug/L		102	70 - 130
Bromacil	1.97	2.15		ug/L		109	70 - 130
Butachlor	1.97	2.01		ug/L		102	70 - 130
Butylbenzylphthalate	1.97	2.08		ug/L		106	70 - 130
Caffeine	1.97	0.942		ug/L		48	45 - 137
Chlorobenzilate	1.97	1.95		ug/L		99	70 - 130
Chloroneb	1.97	2.03		ug/L		103	70 - 130
Chlorothalonil (Draconil, Bravo)	1.97	1.97		ug/L		100	70 - 130
Chlorpyrifos	1.97	2.06		ug/L		104	70 - 130
Chrysene	1.97	2.19		ug/L		111	70 - 130
delta-BHC	1.97	1.98		ug/L		100	70 - 130
Di(2-ethylhexyl)adipate	1.97	2.00		ug/L		101	70 - 130
Bis(2-ethylhexyl) phthalate	1.97	1.83		ug/L		93	70 - 130
Diazinon (Qualitative)	1.97	1.74		ug/L		88	15 - 132
Dibenz(a,h)anthracene	1.97	2.10		ug/L		107	70 - 130
Diclorvos (DDVP)	1.97	1.96		ug/L		100	70 - 130
Dieldrin	1.97	2.03		ug/L		103	70 - 130
Diethylphthalate	1.97	2.01		ug/L		102	70 - 130
Dimethoate	1.97	0.957		ug/L		49	35 - 100
Dimethylphthalate	1.97	2.02		ug/L		102	70 - 130
Di-n-butyl phthalate	3.94	4.10		ug/L		104	70 - 130
Di-n-octyl phthalate	1.97	1.74		ug/L		88	70 - 130
Endosulfan I (Alpha)	1.97	1.75		ug/L		89	70 - 130
Endosulfan II (Beta)	1.97	1.99		ug/L		101	70 - 130
Endosulfan sulfate	1.97	2.13		ug/L		108	70 - 130
Endrin	1.97	1.99		ug/L		101	70 - 130
Endrin aldehyde	1.97	1.60		ug/L		81	70 - 130
EPTC	1.97	1.99		ug/L		101	70 - 130
Fluoranthene	1.97	2.07		ug/L		105	70 - 130

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-14908/3-A
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Fluorene	1.97	2.02		ug/L		103	70 - 130
gamma-Chlordane	1.97	1.81		ug/L		92	70 - 130
Heptachlor	1.97	1.87		ug/L		95	70 - 130
Heptachlor epoxide (isomer B)	1.97	1.85		ug/L		94	70 - 130
Hexachlorobenzene	1.97	1.87		ug/L		95	70 - 130
Hexachlorocyclopentadiene	1.97	2.25		ug/L		114	70 - 130
Indeno[1,2,3-cd]pyrene	1.97	2.04		ug/L		104	70 - 130
Isophorone	1.97	1.88		ug/L		95	70 - 130
Lindane	1.97	1.95		ug/L		99	70 - 130
Malathion	1.97	2.20		ug/L		112	70 - 130
Methoxychlor	1.97	2.19		ug/L		111	70 - 130
Metolachlor	1.97	2.15		ug/L		109	70 - 130
Metribuzin	1.97	1.90		ug/L		97	70 - 130
Molinate	1.97	2.06		ug/L		104	70 - 130
Naphthalene	1.97	1.84		ug/L		94	70 - 130
Parathion	1.97	2.13		ug/L		108	70 - 130
Pendimethalin (Penoxaline)	1.97	2.10		ug/L		106	70 - 130
Phenanthrene	1.97	1.87		ug/L		95	70 - 130
Propachlor	1.97	2.10		ug/L		107	70 - 130
Pyrene	1.97	2.08		ug/L		106	70 - 130
Simazine	1.97	2.06		ug/L		105	70 - 130
Terbacil	1.97	1.94		ug/L		98	70 - 130
Terbutylazine	1.97	2.13		ug/L		108	70 - 130
Thiobencarb	1.97	1.98		ug/L		101	70 - 130
trans-Nonachlor	1.97	1.94		ug/L		99	70 - 130
Trifluralin	1.97	1.97		ug/L		100	70 - 130
1-Methylnaphthalene	1.97	1.87		ug/L		95	70 - 130
2-Methylnaphthalene	1.97	1.91		ug/L		97	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
2-Nitro-m-xylene	93		70 - 130
Triphenylphosphate	91		70 - 130
Perylene-d12	90		70 - 130

Lab Sample ID: LCSD 380-14908/4-A
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
2,4'-DDD	1.96	2.06		ug/L		105	70 - 130	3	20
2,4'-DDE	1.96	2.11		ug/L		107	70 - 130	1	20
2,4'-DDT	1.96	2.12		ug/L		108	70 - 130	2	20
2,4-Dinitrotoluene	1.96	1.96		ug/L		100	70 - 130	0	20
2,6-Dinitrotoluene	1.96	1.90		ug/L		97	70 - 130	1	20
4,4'-DDD	1.96	2.14		ug/L		109	70 - 130	2	20
4,4'-DDE	1.96	1.94		ug/L		99	70 - 130	4	20
4,4'-DDT	1.96	2.09		ug/L		106	70 - 130	2	20
Acenaphthene	1.96	1.89		ug/L		96	70 - 130	1	20

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCSD 380-14908/4-A
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec		RPD	RPD Limit
							Limits	RPD		
Acenaphthylene	1.96	1.84		ug/L		94	70 - 130	0	20	
Acetochlor	1.96	2.05		ug/L		104	70 - 130	1	20	
Alachlor	1.96	2.02		ug/L		103	70 - 130	1	20	
alpha-BHC	1.96	2.02		ug/L		103	70 - 130	1	20	
alpha-Chlordane	1.96	1.78		ug/L		91	70 - 130	2	20	
Anthracene	1.96	1.95		ug/L		99	70 - 130	0	20	
Atrazine	1.96	2.21		ug/L		113	70 - 130	3	20	
Benz(a)anthracene	1.96	2.07		ug/L		106	70 - 130	1	20	
Benzo[a]pyrene	1.96	2.14		ug/L		109	70 - 130	1	20	
Benzo[b]fluoranthene	1.96	2.23		ug/L		114	70 - 130	1	20	
Benzo[g,h,i]perylene	1.96	2.03		ug/L		103	70 - 130	4	20	
Benzo[k]fluoranthene	1.96	2.28		ug/L		116	70 - 130	1	20	
beta-BHC	1.96	2.02		ug/L		103	70 - 130	1	20	
Bromacil	1.96	1.99		ug/L		101	70 - 130	8	20	
Butachlor	1.96	2.06		ug/L		105	70 - 130	2	20	
Butylbenzylphthalate	1.96	2.12		ug/L		108	70 - 130	2	20	
Caffeine	1.96	0.982		ug/L		50	45 - 137	4	20	
Chlorobenzilate	1.96	2.00		ug/L		102	70 - 130	3	20	
Chloroneb	1.96	2.05		ug/L		104	70 - 130	1	20	
Chlorothalonil (Draconil, Bravo)	1.96	2.06		ug/L		105	70 - 130	4	20	
Chlorpyrifos	1.96	2.08		ug/L		106	70 - 130	1	20	
Chrysene	1.96	2.18		ug/L		111	70 - 130	0	20	
delta-BHC	1.96	1.97		ug/L		100	70 - 130	0	20	
Di(2-ethylhexyl)adipate	1.96	2.06		ug/L		105	70 - 130	3	20	
Bis(2-ethylhexyl) phthalate	1.96	1.87		ug/L		95	70 - 130	2	20	
Diazinon (Qualitative)	1.96	1.76		ug/L		90	15 - 132	1	20	
Dibenz(a,h)anthracene	1.96	2.15		ug/L		109	70 - 130	2	20	
Diclorvos (DDVP)	1.96	1.97		ug/L		100	70 - 130	0	20	
Dieldrin	1.96	2.02		ug/L		103	70 - 130	0	20	
Diethylphthalate	1.96	2.03		ug/L		103	70 - 130	1	20	
Dimethoate	1.96	0.977		ug/L		50	35 - 100	2	20	
Dimethylphthalate	1.96	2.02		ug/L		103	70 - 130	0	20	
Di-n-butyl phthalate	3.93	4.21		ug/L		107	70 - 130	3	20	
Di-n-octyl phthalate	1.96	1.76		ug/L		90	70 - 130	2	20	
Endosulfan I (Alpha)	1.96	1.73		ug/L		88	70 - 130	1	20	
Endosulfan II (Beta)	1.96	2.07		ug/L		105	70 - 130	4	20	
Endosulfan sulfate	1.96	2.14		ug/L		109	70 - 130	0	20	
Endrin	1.96	2.02		ug/L		103	70 - 130	1	20	
Endrin aldehyde	1.96	1.59		ug/L		81	70 - 130	0	20	
EPTC	1.96	2.00		ug/L		102	70 - 130	0	20	
Fluoranthene	1.96	2.08		ug/L		106	70 - 130	0	20	
Fluorene	1.96	2.02		ug/L		103	70 - 130	0	20	
gamma-Chlordane	1.96	1.84		ug/L		94	70 - 130	2	20	
Heptachlor	1.96	1.89		ug/L		96	70 - 130	1	20	
Heptachlor epoxide (isomer B)	1.96	1.83		ug/L		93	70 - 130	1	20	
Hexachlorobenzene	1.96	1.92		ug/L		98	70 - 130	3	20	
Hexachlorocyclopentadiene	1.96	2.30		ug/L		117	70 - 130	2	20	
Indeno[1,2,3-cd]pyrene	1.96	2.11		ug/L		107	70 - 130	3	20	
Isophorone	1.96	1.87		ug/L		95	70 - 130	0	20	

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCSD 380-14908/4-A
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Lindane	1.96	1.99		ug/L		101	70 - 130	2	20
Malathion	1.96	2.25		ug/L		114	70 - 130	2	20
Methoxychlor	1.96	2.24		ug/L		114	70 - 130	2	20
Metolachlor	1.96	2.18		ug/L		111	70 - 130	1	20
Metribuzin	1.96	1.91		ug/L		97	70 - 130	1	20
Molinate	1.96	2.05		ug/L		105	70 - 130	0	20
Naphthalene	1.96	1.86		ug/L		95	70 - 130	1	20
Parathion	1.96	2.19		ug/L		112	70 - 130	3	20
Pendimethalin (Penoxaline)	1.96	2.13		ug/L		109	70 - 130	2	20
Phenanthrene	1.96	1.85		ug/L		94	70 - 130	1	20
Propachlor	1.96	2.13		ug/L		108	70 - 130	1	20
Pyrene	1.96	2.11		ug/L		107	70 - 130	1	20
Simazine	1.96	2.09		ug/L		106	70 - 130	1	20
Terbacil	1.96	1.90		ug/L		97	70 - 130	2	20
Terbutylazine	1.96	2.20		ug/L		112	70 - 130	3	20
Thiobencarb	1.96	1.98		ug/L		101	70 - 130	0	20
trans-Nonachlor	1.96	1.94		ug/L		99	70 - 130	0	20
Trifluralin	1.96	2.02		ug/L		103	70 - 130	2	20
1-Methylnaphthalene	1.96	1.90		ug/L		97	70 - 130	1	20
2-Methylnaphthalene	1.96	1.94		ug/L		99	70 - 130	1	20

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
2-Nitro-m-xylene	93		70 - 130
Triphenylphosphate	94		70 - 130
Perylene-d12	91		70 - 130

Lab Sample ID: MRL 380-14908/2-A
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
2,4'-DDD	0.0987	0.137		ug/L		139	50 - 150
2,4'-DDE	0.0987	0.112		ug/L		114	50 - 150
2,4'-DDT	0.0987	0.121		ug/L		123	50 - 150
2,4-Dinitrotoluene	0.0987	0.117		ug/L		118	50 - 150
2,6-Dinitrotoluene	0.0987	0.0878	J	ug/L		89	50 - 150
4,4'-DDD	0.0987	0.0984	J	ug/L		100	50 - 150
4,4'-DDE	0.0987	0.0983	J	ug/L		100	50 - 150
4,4'-DDT	0.0987	0.125		ug/L		127	50 - 150
Acenaphthene	0.0987	0.0985	J	ug/L		100	50 - 150
Acenaphthylene	0.0987	0.0819	J	ug/L		83	50 - 150
Acetochlor	0.0493	0.0446	J	ug/L		90	50 - 150
Alachlor	0.0493	0.0615		ug/L		125	50 - 150
alpha-BHC	0.0987	0.106		ug/L		107	50 - 150
alpha-Chlordane	0.0493	0.0470	J	ug/L		95	50 - 150
Anthracene	0.0197	0.0210		ug/L		107	50 - 150
Atrazine	0.0493	0.0474	J	ug/L		96	50 - 150
Benz(a)anthracene	0.0493	0.0396	J	ug/L		80	50 - 150

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-14908/2-A
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Benzo[a]pyrene	0.0197	0.0206		ug/L		104	50 - 150
Benzo[b]fluoranthene	0.0197	0.0219		ug/L		111	50 - 150
Benzo[g,h,i]perylene	0.0493	0.0399	J	ug/L		81	50 - 150
Benzo[k]fluoranthene	0.0197	0.0209		ug/L		106	50 - 150
beta-BHC	0.0987	0.109		ug/L		110	50 - 150
Bromacil	0.0987	0.125		ug/L		127	50 - 150
Butachlor	0.0493	0.0786	^3+	ug/L		159	50 - 150
Butylbenzylphthalate	0.148	0.177	J	ug/L		120	50 - 150
Caffeine	0.0493	0.0222	J ^3-	ug/L		45	50 - 150
Chlorobenzilate	0.0987	0.125		ug/L		127	50 - 150
Chloroneb	0.0987	0.133		ug/L		135	50 - 150
Chlorothalonil (Draconil, Bravo)	0.0987	0.130		ug/L		132	50 - 150
Chlorpyrifos	0.0493	0.0500		ug/L		101	50 - 150
Chrysene	0.0197	0.0244		ug/L		124	50 - 150
delta-BHC	0.0987	0.131		ug/L		133	50 - 150
Di(2-ethylhexyl)adipate	0.296	0.324	J	ug/L		109	50 - 150
Bis(2-ethylhexyl) phthalate	0.592	0.598		ug/L		101	50 - 150
Diazinon (Qualitative)	0.0987	0.0868	J	ug/L		88	15 - 132
Dibenz(a,h)anthracene	0.0493	0.0458	J	ug/L		93	50 - 150
Diclorvos (DDVP)	0.0493	0.0499		ug/L		101	50 - 150
Dieldrin	0.0987	0.119	J	ug/L		121	50 - 150
Diethylphthalate	0.148	0.161	J	ug/L		109	50 - 150
Dimethoate	0.0987	0.0441	J	ug/L		45	35 - 100
Dimethylphthalate	0.296	0.305	J	ug/L		103	50 - 150
Di-n-butyl phthalate	0.296	0.320	J	ug/L		108	49 - 243
Di-n-octyl phthalate	0.0987	0.107		ug/L		108	50 - 150
Endosulfan I (Alpha)	0.0987	0.106		ug/L		107	50 - 150
Endosulfan II (Beta)	0.0987	0.103		ug/L		104	50 - 150
Endosulfan sulfate	0.0987	0.0987	J	ug/L		100	50 - 150
Endrin	0.0987	0.121		ug/L		123	50 - 150
Endrin aldehyde	0.0987	ND		ug/L		70	50 - 150
EPTC	0.0987	0.0956	J	ug/L		97	50 - 150
Fluoranthene	0.0493	0.0506	J	ug/L		103	50 - 150
Fluorene	0.0493	0.0502		ug/L		102	50 - 150
gamma-Chlordane	0.0493	0.0470	J	ug/L		95	50 - 150
Heptachlor	0.0395	0.0522		ug/L		132	50 - 150
Heptachlor epoxide (isomer B)	0.0493	0.0425	J	ug/L		86	50 - 150
Hexachlorobenzene	0.0493	0.0658		ug/L		133	50 - 150
Hexachlorocyclopentadiene	0.0493	0.0462	J	ug/L		94	50 - 150
Indeno[1,2,3-cd]pyrene	0.0493	0.0437	J	ug/L		89	50 - 150
Isophorone	0.0987	0.0976	J	ug/L		99	50 - 150
Lindane	0.0493	0.0463		ug/L		94	50 - 150
Malathion	0.0987	0.0977	J	ug/L		99	50 - 150
Methoxychlor	0.0987	0.0960	J	ug/L		97	50 - 150
Metolachlor	0.0493	0.0581		ug/L		118	50 - 150
Metribuzin	0.0493	0.0410	J	ug/L		83	50 - 150
Molinate	0.0987	0.105		ug/L		106	50 - 150
Naphthalene	0.0987	0.102	J	ug/L		103	50 - 150
Parathion	0.0987	0.128		ug/L		130	50 - 150

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-14908/2-A
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Pendimethalin (Penoxaline)	0.0987	0.134		ug/L		136	50 - 150
Phenanthrene	0.0197	0.0219	J	ug/L		111	50 - 150
Propachlor	0.0493	0.0532		ug/L		108	50 - 150
Pyrene	0.0493	0.0521		ug/L		106	50 - 150
Simazine	0.0493	0.0535		ug/L		109	50 - 150
Terbacil	0.0987	0.140		ug/L		142	50 - 150
Terbutylazine	0.0987	0.0947	J	ug/L		96	50 - 150
Thiobencarb	0.0987	0.109	J	ug/L		110	50 - 150
trans-Nonachlor	0.0493	0.0495		ug/L		100	50 - 150
Trifluralin	0.0987	0.124		ug/L		126	50 - 150
1-Methylnaphthalene	0.0987	0.101		ug/L		102	50 - 150
2-Methylnaphthalene	0.0987	0.0978	J	ug/L		99	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
2-Nitro-m-xylene	92		70 - 130
Triphenylphosphate	92		70 - 130
Perylene-d12	85		70 - 130

Lab Sample ID: 380-16794-1 MS
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Halawa Wells Pump 1
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
2,4'-DDD	ND		1.99	2.04		ug/L		103	70 - 130
2,4'-DDE	ND		1.99	2.06		ug/L		104	70 - 130
2,4'-DDT	ND		1.99	2.05		ug/L		103	70 - 130
2,4-Dinitrotoluene	ND		1.99	1.69		ug/L		85	70 - 130
2,6-Dinitrotoluene	ND		1.99	1.66		ug/L		84	70 - 130
4,4'-DDD	ND		1.99	2.07		ug/L		104	70 - 130
4,4'-DDE	ND		1.99	1.83		ug/L		92	70 - 130
4,4'-DDT	ND		1.99	1.99		ug/L		100	70 - 130
Acenaphthene	ND		1.99	1.93		ug/L		97	70 - 130
Acenaphthylene	ND		1.99	1.85		ug/L		93	70 - 130
Acetochlor	ND		1.99	2.05		ug/L		103	70 - 130
Alachlor	ND		1.99	2.01		ug/L		101	70 - 130
alpha-BHC	ND		1.99	2.01		ug/L		101	70 - 130
alpha-Chlordane	ND		1.99	1.77		ug/L		89	70 - 130
Anthracene	ND		1.99	1.97		ug/L		99	70 - 130
Atrazine	ND		1.99	2.12		ug/L		107	70 - 130
Benz(a)anthracene	ND		1.99	2.03		ug/L		102	70 - 130
Benzo[a]pyrene	ND		1.99	2.08		ug/L		105	70 - 130
Benzo[b]fluoranthene	ND		1.99	2.23		ug/L		112	70 - 130
Benzo[g,h,i]perylene	ND		1.99	2.03		ug/L		102	70 - 130
Benzo[k]fluoranthene	ND		1.99	2.14		ug/L		108	70 - 130
beta-BHC	ND		1.99	2.03		ug/L		102	70 - 130
Bromacil	ND		1.99	1.56		ug/L		75	70 - 130
Butachlor	ND	^3+	1.99	2.01		ug/L		101	70 - 130
Butylbenzylphthalate	ND		1.99	2.09		ug/L		105	70 - 130

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-16794-1 MS
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Halawa Wells Pump 1
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
Caffeine	ND	F1 ^3-	1.99	0.749	F1	ug/L		38	46 - 144
Chlorobenzilate	ND		1.99	1.97		ug/L		99	70 - 130
Chloroneb	ND		1.99	2.06		ug/L		104	70 - 130
Chlorothalonil (Draconil, Bravo)	ND		1.99	2.00		ug/L		101	70 - 130
Chlorpyrifos	ND		1.99	2.06		ug/L		104	70 - 130
Chrysene	ND		1.99	2.17		ug/L		109	70 - 130
delta-BHC	ND		1.99	1.98		ug/L		100	70 - 130
Di(2-ethylhexyl)adipate	ND		1.99	1.92		ug/L		97	70 - 130
Bis(2-ethylhexyl) phthalate	ND		1.99	1.73		ug/L		87	70 - 130
Diazinon (Qualitative)	ND		1.99	1.78		ug/L		89	15 - 132
Dibenz(a,h)anthracene	ND		1.99	2.13		ug/L		107	70 - 130
Diclorvos (DDVP)	ND		1.99	1.81		ug/L		91	70 - 130
Dieldrin	ND		1.99	2.04		ug/L		103	70 - 130
Diethylphthalate	ND		1.99	2.03		ug/L		102	70 - 130
Dimethoate	ND		1.99	0.716		ug/L		36	34 - 111
Dimethylphthalate	ND		1.99	1.96		ug/L		99	70 - 130
Di-n-butyl phthalate	ND		3.97	4.08		ug/L		103	70 - 130
Di-n-octyl phthalate	ND		1.99	1.66		ug/L		84	70 - 130
Endosulfan I (Alpha)	ND		1.99	1.79		ug/L		90	70 - 130
Endosulfan II (Beta)	ND		1.99	2.05		ug/L		103	70 - 130
Endosulfan sulfate	ND		1.99	2.17		ug/L		109	70 - 130
Endrin	ND		1.99	2.01		ug/L		101	70 - 130
Endrin aldehyde	ND		1.99	1.58		ug/L		80	70 - 130
EPTC	ND		1.99	2.05		ug/L		103	70 - 130
Fluoranthene	ND		1.99	2.06		ug/L		104	70 - 130
Fluorene	ND		1.99	2.05		ug/L		103	70 - 130
gamma-Chlordane	ND		1.99	1.81		ug/L		90	70 - 130
Heptachlor	ND		1.99	1.90		ug/L		96	70 - 130
Heptachlor epoxide (isomer B)	ND		1.99	1.78		ug/L		90	70 - 130
Hexachlorobenzene	ND		1.99	1.94		ug/L		98	70 - 130
Hexachlorocyclopentadiene	ND		1.99	2.33		ug/L		118	70 - 130
Indeno[1,2,3-cd]pyrene	ND		1.99	2.09		ug/L		105	70 - 130
Isophorone	ND		1.99	1.78		ug/L		90	70 - 130
Lindane	ND		1.99	2.02		ug/L		102	70 - 130
Malathion	ND		1.99	2.22		ug/L		112	70 - 130
Methoxychlor	ND		1.99	2.20		ug/L		111	70 - 130
Metolachlor	ND		1.99	2.15		ug/L		108	70 - 130
Metribuzin	ND		1.99	1.62		ug/L		81	70 - 130
Molinate	ND		1.99	2.07		ug/L		104	70 - 130
Naphthalene	ND		1.99	1.85		ug/L		93	70 - 130
Parathion	ND		1.99	2.18		ug/L		110	70 - 130
Pendimethalin (Penoxaline)	ND		1.99	2.12		ug/L		107	70 - 130
Phenanthrene	ND		1.99	1.87		ug/L		94	70 - 130
Propachlor	ND		1.99	2.13		ug/L		107	70 - 130
Pyrene	ND		1.99	2.10		ug/L		106	70 - 130
Simazine	ND		1.99	1.79		ug/L		90	70 - 130
Terbacil	ND		1.99	1.65		ug/L		83	70 - 130
Terbutylazine	ND		1.99	2.12		ug/L		107	70 - 130
Thiobencarb	ND		1.99	1.98		ug/L		100	70 - 130

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-16794-1 MS
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Halawa Wells Pump 1
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
trans-Nonachlor	ND		1.99	1.93		ug/L		97	70 - 130
Trifluralin	ND		1.99	1.99		ug/L		100	70 - 130
1-Methylnaphthalene	ND		1.99	1.91		ug/L		96	70 - 130
2-Methylnaphthalene	ND		1.99	1.93		ug/L		97	70 - 130

Surrogate	MS %Recovery	MS Qualifier	MS Limits
2-Nitro-m-xylene	94		70 - 130
Triphenylphosphate	92		70 - 130
Perylene-d12	93		70 - 130

Lab Sample ID: 380-16460-T-1-A DU
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Duplicate
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	Limit
2,4'-DDD	ND		ND		ug/L		NC	20
2,4'-DDE	ND		ND		ug/L		NC	20
2,4'-DDT	ND		ND		ug/L		NC	20
2,4-Dinitrotoluene	ND		ND		ug/L		NC	20
2,6-Dinitrotoluene	ND		ND		ug/L		NC	20
4,4'-DDD	ND		ND		ug/L		NC	20
4,4'-DDE	ND		ND		ug/L		NC	20
4,4'-DDT	ND		ND		ug/L		NC	20
Acenaphthene	ND		ND		ug/L		NC	20
Acenaphthylene	ND		ND		ug/L		NC	20
Acetochlor	ND		ND		ug/L		NC	20
Alachlor	ND		ND		ug/L		NC	20
alpha-BHC	ND		ND		ug/L		NC	20
alpha-Chlordane	ND		ND		ug/L		NC	20
Anthracene	ND		ND		ug/L		NC	20
Atrazine	ND		ND		ug/L		NC	20
Benz(a)anthracene	ND		ND		ug/L		NC	20
Benzo[a]pyrene	ND		ND		ug/L		NC	20
Benzo[b]fluoranthene	ND		ND		ug/L		NC	20
Benzo[g,h,i]perylene	ND		ND		ug/L		NC	20
Benzo[k]fluoranthene	ND		ND		ug/L		NC	20
beta-BHC	ND		ND		ug/L		NC	20
Bromacil	ND		ND		ug/L		NC	20
Butachlor	ND	^3+	ND		ug/L		NC	20
Butylbenzylphthalate	ND		ND		ug/L		NC	20
Chlorobenzilate	ND		ND		ug/L		NC	20
Chloroneb	ND		ND		ug/L		NC	20
Chlorothalonil (Draconil, Bravo)	ND		ND		ug/L		NC	20
Chlorpyrifos	ND		ND		ug/L		NC	20
Chrysene	ND		ND		ug/L		NC	20
delta-BHC	ND		ND		ug/L		NC	20
Di(2-ethylhexyl)adipate	ND		ND		ug/L		NC	20
Bis(2-ethylhexyl) phthalate	ND		ND		ug/L		NC	20

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-16460-T-1-A DU
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Duplicate
Prep Type: Total/NA
Prep Batch: 14908

Analyte	Sample	Sample	DU	DU	Unit	D	RPD	Limit
	Result	Qualifier	Result	Qualifier				
Diazinon (Qualitative)	ND		ND		ug/L		NC	20
Dibenz(a,h)anthracene	ND		ND		ug/L		NC	20
Diclorvos (DDVP)	ND		ND		ug/L		NC	20
Dieldrin	ND		ND		ug/L		NC	20
Diethylphthalate	ND		ND		ug/L		NC	20
Dimethoate	ND		ND		ug/L		NC	20
Dimethylphthalate	ND		ND		ug/L		NC	20
Di-n-butyl phthalate	ND		ND		ug/L		NC	20
Di-n-octyl phthalate	ND		ND		ug/L		NC	20
Endosulfan I (Alpha)	ND		ND		ug/L		NC	20
Endosulfan II (Beta)	ND		ND		ug/L		NC	20
Endosulfan sulfate	ND		ND		ug/L		NC	20
Endrin	ND		ND		ug/L		NC	20
Endrin aldehyde	ND		ND		ug/L		NC	20
EPTC	ND		ND		ug/L		NC	20
Fluoranthene	ND		ND		ug/L		NC	20
Fluorene	ND		ND		ug/L		NC	20
gamma-Chlordane	ND		ND		ug/L		NC	20
Heptachlor	ND		ND		ug/L		NC	20
Heptachlor epoxide (isomer B)	ND		ND		ug/L		NC	20
Hexachlorobenzene	ND		ND		ug/L		NC	20
Hexachlorocyclopentadiene	ND		ND		ug/L		NC	20
Indeno[1,2,3-cd]pyrene	ND		ND		ug/L		NC	20
Isophorone	ND		ND		ug/L		NC	20
Lindane	ND		ND		ug/L		NC	20
Malathion	ND		ND		ug/L		NC	20
Methoxychlor	ND		ND		ug/L		NC	20
Metolachlor	ND		ND		ug/L		NC	20
Metribuzin	ND		ND		ug/L		NC	20
Molinate	ND		ND		ug/L		NC	20
Naphthalene	ND		ND		ug/L		NC	20
Parathion	ND		ND		ug/L		NC	20
Pendimethalin (Penoxaline)	ND		ND		ug/L		NC	20
Total Permethrin (mixed isomers)	ND		ND		ug/L		NC	20
Phenanthrene	ND		ND		ug/L		NC	20
Propachlor	ND		ND		ug/L		NC	20
Pyrene	ND		ND		ug/L		NC	20
Simazine	ND		ND		ug/L		NC	20
Terbacil	ND		ND		ug/L		NC	20
Terbutylazine	ND		ND		ug/L		NC	20
Thiobencarb	ND		ND		ug/L		NC	20
trans-Nonachlor	ND		ND		ug/L		NC	20
Trifluralin	ND		ND		ug/L		NC	20
1-Methylnaphthalene	ND		ND		ug/L		NC	20
2-Methylnaphthalene	ND		ND		ug/L		NC	20

Surrogate	%Recovery	DU Qualifier	DU Limits
2-Nitro-m-xylene	90		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-16460-T-1-A DU
Matrix: Water
Analysis Batch: 15264

Client Sample ID: Duplicate
Prep Type: Total/NA
Prep Batch: 14908

Surrogate	%Recovery	DU DU Qualifier	Limits
Triphenylphosphate	91		70 - 130
Perylene-d12	89		70 - 130

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i

Lab Sample ID: 99233-B1
Matrix: water
Analysis Batch: O-38084

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: O-38084_P

Analyte	Blank Result	Blank Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
1-Methylphenanthrene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
2,3,5-Trimethylnaphthalene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
2,6-Dimethylnaphthalene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
2-Methylnaphthalene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Acenaphthene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Acenaphthylene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Anthracene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Benz[a]anthracene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Benzo[a]pyrene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Benzo[b]fluoranthene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Benzo[e]pyrene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Benzo[g,h,i]perylene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Benzo[k]fluoranthene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Biphenyl	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Chrysene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Dibenz[a,h]anthracene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Dibenzo[a,l]pyrene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Dibenzothiophene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Disalicylidenepropanediamine	ND		0.1	0.05	µg/L		08/15/22 00:00	08/20/22 16:04	1
Fluoranthene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Fluorene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Indeno[1,2,3-cd]pyrene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Naphthalene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Perylene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Phenanthrene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1
Pyrene	ND		0.005	0.001	µg/L		08/15/22 00:00	08/20/22 16:04	1

Surrogate	%Recovery	Blank Blank Qualifier	Limits	Prepared	Analyzed	Dil Fac
(d10-Acenaphthene)	92		65 - 113	08/15/22 00:00	08/20/22 16:04	1
(d10-Phenanthrene)	93		80 - 111	08/15/22 00:00	08/20/22 16:04	1
(d12-Chrysene)	92		60 - 139	08/15/22 00:00	08/20/22 16:04	1
(d12-Perylene)	88		36 - 161	08/15/22 00:00	08/20/22 16:04	1
(d8-Naphthalene)	88		44 - 119	08/15/22 00:00	08/20/22 16:04	1

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i (Continued)

Lab Sample ID: 99233-BS1
Matrix: water
Analysis Batch: O-38084

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: O-38084_P

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	0.5	0.414		µg/L		83	49 - 117
1-Methylphenanthrene	0.5	0.434		µg/L		87	66 - 127
2,3,5-Trimethylnaphthalene	0.5	0.427		µg/L		85	57 - 120
2,6-Dimethylnaphthalene	0.5	0.413		µg/L		83	54 - 117
2-Methylnaphthalene	0.5	0.41		µg/L		82	47 - 130
Acenaphthene	0.5	0.425		µg/L		85	53 - 131
Acenaphthylene	0.5	0.418		µg/L		84	43 - 140
Anthracene	0.5	0.449		µg/L		90	58 - 135
Benz[a]anthracene	0.5	0.365		µg/L		73	55 - 145
Benzo[a]pyrene	0.5	0.401		µg/L		80	51 - 143
Benzo[b]fluoranthene	0.5	0.446		µg/L		89	46 - 165
Benzo[e]pyrene	0.5	0.427		µg/L		85	42 - 152
Benzo[g,h,i]perylene	0.5	0.444		µg/L		89	63 - 133
Benzo[k]fluoranthene	0.5	0.426		µg/L		85	56 - 145
Biphenyl	0.5	0.422		µg/L		84	56 - 119
Chrysene	0.5	0.384		µg/L		77	56 - 141
Dibenz[a,h]anthracene	0.5	0.407		µg/L		81	55 - 150
Dibenzo[a,l]pyrene	0.5	0.499		µg/L		100	50 - 150
Dibenzothiophene	0.5	0.436		µg/L		87	75 - 113
Disalicylidenepropanediamine	25	18.1		µg/L		72	50 - 150
Fluoranthene	0.5	0.442		µg/L		88	60 - 146
Fluorene	0.5	0.432		µg/L		86	58 - 131
Indeno[1,2,3-cd]pyrene	0.5	0.415		µg/L		83	50 - 151
Naphthalene	0.5	0.402		µg/L		80	41 - 126
Perylene	0.5	0.4		µg/L		80	48 - 141
Phenanthrene	0.5	0.45		µg/L		90	67 - 127
Pyrene	0.5	0.425		µg/L		85	54 - 156

Surrogate	LCS %Recovery	LCS Qualifier	LCS Limits
(d10-Acenaphthene)	87		65 - 113
(d10-Phenanthrene)	91		80 - 111
(d12-Chrysene)	84		60 - 139
(d12-Perylene)	86		36 - 161
(d8-Naphthalene)	81		44 - 119

Lab Sample ID: 99233-BS2
Matrix: water
Analysis Batch: O-38084

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: O-38084_P

Analyte	Spike Added	LCS DUP Result	LCS DUP Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1-Methylnaphthalene	0.5	0.442		µg/L		88	49 - 117	6	30
1-Methylphenanthrene	0.5	0.443		µg/L		89	66 - 127	2	30
2,3,5-Trimethylnaphthalene	0.5	0.45		µg/L		90	57 - 120	6	30
2,6-Dimethylnaphthalene	0.5	0.44		µg/L		88	54 - 117	6	30
2-Methylnaphthalene	0.5	0.439		µg/L		88	47 - 130	7	30
Acenaphthene	0.5	0.449		µg/L		90	53 - 131	6	30
Acenaphthylene	0.5	0.449		µg/L		90	43 - 140	7	30
Anthracene	0.5	0.427		µg/L		85	58 - 135	6	30

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 625 PAH Physis LL (EAL) + TICs - EPA 625 Base/Neutral and Acid Organics i (Continued)

Lab Sample ID: 99233-BS2
Matrix: water
Analysis Batch: O-38084

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: O-38084_P

Analyte	Spike Added	LCS DUP Result	LCS DUP Qualifier	Unit	D	%Rec	%Rec		RPD	Limit
							Limits	RPD		
Benz[a]anthracene	0.5	0.387		µg/L		77	55 - 145	5	30	
Benzo[a]pyrene	0.5	0.447		µg/L		89	51 - 143	11	30	
Benzo[b]fluoranthene	0.5	0.472		µg/L		94	46 - 165	5	30	
Benzo[e]pyrene	0.5	0.453		µg/L		91	42 - 152	7	30	
Benzo[g,h,i]perylene	0.5	0.465		µg/L		93	63 - 133	4	30	
Benzo[k]fluoranthene	0.5	0.468		µg/L		94	56 - 145	10	30	
Biphenyl	0.5	0.455		µg/L		91	56 - 119	8	30	
Chrysene	0.5	0.4		µg/L		80	56 - 141	4	30	
Dibenz[a,h]anthracene	0.5	0.439		µg/L		88	55 - 150	8	30	
Dibenzo[a,l]pyrene	0.5	0.512		µg/L		102	50 - 150	2	30	
Dibenzothiophene	0.5	0.461		µg/L		92	75 - 113	6	30	
Disalicylidenepropanediamine	25	23.3		µg/L		93	50 - 150	25	30	
Fluoranthene	0.5	0.459		µg/L		92	60 - 146	4	30	
Fluorene	0.5	0.449		µg/L		90	58 - 131	5	30	
Indeno[1,2,3-cd]pyrene	0.5	0.442		µg/L		88	50 - 151	6	30	
Naphthalene	0.5	0.437		µg/L		87	41 - 126	8	30	
Perylene	0.5	0.428		µg/L		86	48 - 141	7	30	
Phenanthrene	0.5	0.47		µg/L		94	67 - 127	4	30	
Pyrene	0.5	0.443		µg/L		89	54 - 156	5	30	

Surrogate	LCS DUP		Limits
	%Recovery	Qualifier	
(d10-Acenaphthene)	92		65 - 113
(d10-Phenanthrene)	98		80 - 111
(d12-Chrysene)	88		60 - 139
(d12-Perylene)	95		36 - 161
(d8-Naphthalene)	90		44 - 119

Method: 8015 Diesel LL (EAL) and Motor Oil - 8015 - TPH DRO/ORO

Lab Sample ID: 22VGH7H09B
Matrix: WATER
Analysis Batch: 22VGH7H09

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB MB		RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier							
GASOLINE	ND	U	0.02		mg/L			08/23/22 13:39	1

Surrogate	MB MB		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
BROMOFLUOROBENZENE					08/23/22 13:39	1

Lab Sample ID: 22VGH7H09L
Matrix: WATER
Analysis Batch: 22VGH7H09

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec	
							Limits	RPD
GASOLINE	0.5	0.464		mg/L		93	60 - 130	

Surrogate	LCS LCS		Limits
	%Recovery	Qualifier	
BROMOFLUOROBENZENE	118		70 - 130

Eurofins Eaton Monrovia

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method: 8015 Gas (Purgeable) LL (EAL) - SW846 8015B Gasoline Range Organics

Lab Sample ID: 22DSH034WB
Matrix: WATER
Analysis Batch: 22DSH034W

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
DIESEL	ND	U	0.025		mg/L			08/22/22 17:35	1
MOTOR OIL	ND	U	0.05		mg/L			08/22/22 17:35	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
BROMOBENZENE					08/22/22 17:35	1
HEXACOSANE					08/22/22 17:35	1

Lab Sample ID: 22DSH034WL
Matrix: WATER
Analysis Batch: 22DSH034W

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
DIESEL	2.5	2.11		mg/L		84	50 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
BROMOBENZENE	75		60 - 130
HEXACOSANE	93		60 - 130

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

GC/MS Semi VOA

Prep Batch: 14908

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-16794-1	Halawa Wells Pump 1	Total/NA	Water	525.2	
MB 380-14908/1-A	Method Blank	Total/NA	Water	525.2	
LCS 380-14908/3-A	Lab Control Sample	Total/NA	Water	525.2	
LCSD 380-14908/4-A	Lab Control Sample Dup	Total/NA	Water	525.2	
MRL 380-14908/2-A	Lab Control Sample	Total/NA	Water	525.2	
380-16794-1 MS	Halawa Wells Pump 1	Total/NA	Water	525.2	
380-16460-T-1-A DU	Duplicate	Total/NA	Water	525.2	

Analysis Batch: 15264

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-16794-1	Halawa Wells Pump 1	Total/NA	Water	525.2	14908
MB 380-14908/1-A	Method Blank	Total/NA	Water	525.2	14908
LCS 380-14908/3-A	Lab Control Sample	Total/NA	Water	525.2	14908
LCSD 380-14908/4-A	Lab Control Sample Dup	Total/NA	Water	525.2	14908
MRL 380-14908/2-A	Lab Control Sample	Total/NA	Water	525.2	14908
380-16794-1 MS	Halawa Wells Pump 1	Total/NA	Water	525.2	14908
380-16460-T-1-A DU	Duplicate	Total/NA	Water	525.2	14908

Subcontract

Analysis Batch: O-38084

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-16794-1	Halawa Wells Pump 1	Total/NA	Water	625 PAH Physis LL (EAL) + TICs	O-38084_P
99233-B1	Method Blank	Total/NA	water	625 PAH Physis LL (EAL) + TICs	O-38084_P
99233-BS1	Lab Control Sample	Total/NA	water	625 PAH Physis LL (EAL) + TICs	O-38084_P
99233-BS2	Lab Control Sample Dup	Total/NA	water	625 PAH Physis LL (EAL) + TICs	O-38084_P

Analysis Batch: 22DSH034W

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-16794-1	Halawa Wells Pump 1	Total/NA	Water	8015 Gas (Purgeable) LL (EAL)	
22DSH034WB	Method Blank	Total/NA	WATER	8015 Gas (Purgeable) LL (EAL)	
22DSH034WL	Lab Control Sample	Total/NA	WATER	8015 Gas (Purgeable) LL (EAL)	

Analysis Batch: 22VGH7H09

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-16794-1	Halawa Wells Pump 1	Total/NA	Water	8015 Diesel LL (EAL) and Motor Oil	
380-16794-2	Halawa Wells Pump 1 TB	Total/NA	Water	8015 Gas (Purgeable) LL (EAL)	
22VGH7H09B	Method Blank	Total/NA	WATER	8015 Diesel LL (EAL) and Motor Oil	

Eurofins Eaton Monrovia

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Subcontract (Continued)

Analysis Batch: 22VGH7H09 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
22VGH7H09L	Lab Control Sample	Total/NA	WATER	8015 Diesel LL (EAL) and Motor Oil	

Prep Batch: O-38084_P

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-16794-1	Halawa Wells Pump 1	Total/NA	Water	EPA_625	
99233-B1	Method Blank	Total/NA	water	EPA_625	
99233-BS1	Lab Control Sample	Total/NA	water	EPA_625	
99233-BS2	Lab Control Sample Dup	Total/NA	water	EPA_625	



Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Client Sample ID: Halawa Wells Pump 1

Lab Sample ID: 380-16794-1

Date Collected: 08/15/22 10:20

Matrix: Water

Date Received: 08/16/22 17:24

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			14908	N8NE	EA MON	08/25/22 10:21
Total/NA	Analysis	525.2		1	15264	Q8LA	EA MON	08/29/22 11:16
Total/NA	Prep	EPA_625		1	O-38084_P			08/17/22 00:00
Total/NA	Analysis	625 PAH Physis LL (EAL) + TICs		1	O-38084	YC		08/21/22 15:07
Total/NA	Analysis	8015 Diesel LL (EAL) and Motor Oil		1	22VGH7H09	SCerva		08/23/22 15:25
Total/NA	Analysis	8015 Gas (Purgeable) LL (EAL)		1	22DSH034W	SDees		08/22/22 23:09

Client Sample ID: Halawa Wells Pump 1 TB

Lab Sample ID: 380-16794-2

Date Collected: 08/15/22 10:20

Matrix: Water

Date Received: 08/16/22 17:24

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015 Gas (Purgeable) LL (EAL)		1	22VGH7H09	SCerva		08/23/22 16:01

Laboratory References:

= Physis Environmental Laboratories, 1904 Wright Circle, Anaheim, CA 92806

EA MON = Eurofins Eaton Monrovia, 750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-16794-1

Laboratory: Eurofins Eaton Monrovia

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-23

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Water	1-Methylnaphthalene
525.2	525.2	Water	2,4'-DDD
525.2	525.2	Water	2,4'-DDE
525.2	525.2	Water	2,4'-DDT
525.2	525.2	Water	2,4-Dinitrotoluene
525.2	525.2	Water	2,6-Dinitrotoluene
525.2	525.2	Water	2-Methylnaphthalene
525.2	525.2	Water	4,4'-DDD
525.2	525.2	Water	4,4'-DDE
525.2	525.2	Water	4,4'-DDT
525.2	525.2	Water	Acenaphthene
525.2	525.2	Water	Acenaphthylene
525.2	525.2	Water	Acetochlor
525.2	525.2	Water	alpha-BHC
525.2	525.2	Water	alpha-Chlordane
525.2	525.2	Water	Anthracene
525.2	525.2	Water	Benz(a)anthracene
525.2	525.2	Water	Benzo[b]fluoranthene
525.2	525.2	Water	Benzo[g,h,i]perylene
525.2	525.2	Water	Benzo[k]fluoranthene
525.2	525.2	Water	beta-BHC
525.2	525.2	Water	Bromacil
525.2	525.2	Water	Butylbenzylphthalate
525.2	525.2	Water	Chlorobenzilate
525.2	525.2	Water	Chloroneb
525.2	525.2	Water	Chlorothalonil (Draconil, Bravo)
525.2	525.2	Water	Chlorpyrifos
525.2	525.2	Water	Chrysene
525.2	525.2	Water	delta-BHC
525.2	525.2	Water	Diazinon (Qualitative)
525.2	525.2	Water	Dibenz(a,h)anthracene
525.2	525.2	Water	Diclorvos (DDVP)
525.2	525.2	Water	Diethylphthalate
525.2	525.2	Water	Dimethoate
525.2	525.2	Water	Dimethylphthalate
525.2	525.2	Water	Di-n-butyl phthalate
525.2	525.2	Water	Di-n-octyl phthalate
525.2	525.2	Water	Endosulfan I (Alpha)
525.2	525.2	Water	Endosulfan II (Beta)
525.2	525.2	Water	Endosulfan sulfate
525.2	525.2	Water	Endrin aldehyde
525.2	525.2	Water	EPTC
525.2	525.2	Water	Fluoranthene
525.2	525.2	Water	Fluorene
525.2	525.2	Water	gamma-Chlordane

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Laboratory: Eurofins Eaton Monrovia (Continued)

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
-----------	---------	-----------------------	-----------------

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Water	Indeno[1,2,3-cd]pyrene
525.2	525.2	Water	Isophorone
525.2	525.2	Water	Malathion
525.2	525.2	Water	Molinate
525.2	525.2	Water	Naphthalene
525.2	525.2	Water	Parathion
525.2	525.2	Water	Pendimethalin (Penoxaline)
525.2	525.2	Water	Phenanthrene
525.2	525.2	Water	Pyrene
525.2	525.2	Water	Terbacil
525.2	525.2	Water	Terbutylazine
525.2	525.2	Water	Thiobencarb
525.2	525.2	Water	Total Permethrin (mixed isomers)
525.2	525.2	Water	trans-Nonachlor
525.2	525.2	Water	Trifluralin



Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Method	Method Description	Protocol	Laboratory
525.2	Semivolatile Organic Compounds (GC/MS)	EPA	EA MON
625	EPA 625 Base/Neutral and Acid Organics i	EPA	
8015	8015 - TPH DRO/ORO	EPA	
8015B	SW846 8015B Gasoline Range Organics	SW846	
525.2	Extraction of Semivolatile Compounds	EPA	EA MON

Protocol References:

EPA = US Environmental Protection Agency

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

= Physis Environmental Laboratories, 1904 Wright Circle, Anaheim, CA 92806

EA MON = Eurofins Eaton Monrovia, 750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016, TEL (626)386-1100



Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-16794-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
380-16794-1	Halawa Wells Pump 1	Water	08/15/22 10:20	08/16/22 17:24
380-16794-2	Halawa Wells Pump 1 TB	Water	08/15/22 10:20	08/16/22 17:24

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17



3051 Fujita Street
Torrance, CA 90505
Tel: (310)-618-8889

Date: 09-06-2022
EMAX Batch No.: 22H218

Attn: Jackie Contreras

Eurofins Eaton Analytical
750 Royal Oaks Dr., Suite 100
Monrovia, CA 91016-3629

Subject: Laboratory Report
Project: 380-16794


Enclosed is the Laboratory report for samples received on 08/17/22.
The data reported relate only to samples listed below :

Sample ID	Control #	Col Date	Matrix	Analysis
380-16794-1	H218-01	08/15/22	WATER	TPH GASOLINE TPH DIESEL & MOTOR OIL
380-16794-2	H218-02	08/15/22	WATER	TPH GASOLINE

The results are summarized on the following pages.

Please feel free to call if you have any questions concerning these results.

Sincerely yours,


Caspar J. Pang
Laboratory Director

This report is confidential and intended solely for the use of the individual or entity to whom it is addressed. This report shall not be reproduced except in full or without the written approval of EMAX.

EMAX certifies that results included in this report meets all TNI & DOD requirements unless noted in the Case Narrative.

NELAP Accredited Certificate Number CA002912022-22
ANAB Accredited DoD ELAP and ISO/IEC 17025 Certificate Number L2278 Testing
California ELAP Accredited Certificate Number 2672

Monrovia, CA (Suite 100)
Chain of Custody Record

750 Royal Oaks Drive Suite 100
 Monrovia, CA 91016
 Phone: 626-386-1100



Client Information (Sub Contract Lab)
 Client Contact: Shipping/Receiving
 Company: EMAX Laboratories Inc
 Address: 3051 Fujita Street, Torrance
 State, Zip: CA, 90505
 Phone: [Blank]
 Email: [Blank]
 Project Name: RED-HILL
 Site: Honolulu BWS Sites

Sampler: Frank, Debbie L
 Lab PM: Debbie, Frank@eurofins.com
 State of Origin: Hawaii
 Carrier Tracking No(s): [Blank]
 COC No: 380-17920.1
 Page: 1 of 1
 Job #: 380-16794-1

Due Date Requested: 8/30/2022
 TAT Requested (days): [Blank]
 Analysis Requested: [Blank]

Field Filtered Sample (Yes or No)
 Perform MS/MSD (Yes or No)
 SUB (8015 Gas (Purgeable) LL (EAL)/ 8015 Gas (Purgeable) LL (EAL)
 SUB (8015 Diesel LL (EAL) and Motor Oil)/ 8015 Diesel LL (EAL) and Motor Oil
 Total Number of containers: 1
 Special Instructions/Note: See Attached Instructions

Sample Identification - Client ID (Lab ID)	Sample Date	Sample Time	Sample Type (C=comp, G=grab)	MATRIX (W=water, S=solid, O=overwash, BT=tissue, A=air)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	Special Instructions/Note
Halawa Wells Pump 1 (380-16794-1)	8/15/22	10:20		Water	X	X	See Attached Instructions
Halawa Wells Pump 1 TB (380-16794-2)	8/15/22	10:20		Water	X		See Attached Instructions

Note: Since laboratory accreditations are subject to change, Eurofins Eaton Analytical, LLC places the ownership of method, analyte & accreditation compliance upon out subcontract laboratories. This sample shipment is forwarded under chain-of-custody. If the laboratory does not currently maintain accreditation in the State of Origin listed above for analysis/method being analyzed, the samples must be shipped back to the Eurofins Eaton Analytical, LLC laboratory or other instructions will be provided. Any changes to accreditation status should be brought to Eurofins Eaton Analytical, LLC attention immediately. If all requested accreditations are current to date, return the signed Chain of Custody attesting to said compliance to Eurofins Eaton Analytical, LLC.

Possible Hazard Identification
 Unconfirmed
 Deliverable Requested: I, II, III, IV, Other (specify)
 Primary Deliverable Rank: 2
 Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)
 Return To Client Disposal By Lab Archive For _____ Months

Empty Kit Relinquished by: [Blank] Date: [Blank]
 Relinquished by: GRETNER Date/Time: 08/17/2022 Company: EBA
 Relinquished by: ALB-T Date/Time: 8-17-22 13:50 Company: EMAX
 Relinquished by: [Blank] Date/Time: [Blank] Company: [Blank]
 Cooler Temperature(s) °C and Other Remarks: Temp. 2.8
 REPORT# D: 22H218
 Custody Seal No.: [Blank]
 Method of Shipment: [Blank]



Type of Delivery	Airbill / Tracking Number	ECN 22H218
<input type="checkbox"/> Fedex <input type="checkbox"/> UPS <input type="checkbox"/> GSO <input type="checkbox"/> Others		Recipient Maria Rivera
<input type="checkbox"/> EMAX Courier <input checked="" type="checkbox"/> Client Delivery		Date 08/17/22 Time 13:50

COC INSPECTION

<input checked="" type="checkbox"/> Client Name	<input type="checkbox"/> Client PM/FC	<input type="checkbox"/> Sampler Name	<input checked="" type="checkbox"/> Sampling Date/Time	<input checked="" type="checkbox"/> Sample ID	<input checked="" type="checkbox"/> Matrix
<input checked="" type="checkbox"/> Address	<input checked="" type="checkbox"/> Tel # / Fax #	<input type="checkbox"/> Courier Signature	<input checked="" type="checkbox"/> Analysis Required	<input type="checkbox"/> Preservative (if any)	<input checked="" type="checkbox"/> TAT
Safety Issues (if any)	<input type="checkbox"/> High concentrations expected	<input type="checkbox"/> From Superfund Site	<input type="checkbox"/> Rad screening required		

Note: _____

PACKAGING INSPECTION

Container	<input checked="" type="checkbox"/> Cooler	<input type="checkbox"/> Box	<input type="checkbox"/> Other
Condition	<input type="checkbox"/> Custody Seal	<input type="checkbox"/> Intact	<input type="checkbox"/> Damaged
Packaging	<input checked="" type="checkbox"/> Bubble Pack	<input type="checkbox"/> Styrofoam	<input type="checkbox"/> Popcorn
Temperatures (Cool, ≤5 °C but not frozen)	<input checked="" type="checkbox"/> Cooler 1 28 °C	<input type="checkbox"/> Cooler 2 _____ °C	<input type="checkbox"/> Cooler 3 _____ °C
	<input type="checkbox"/> Cooler 6 _____ °C	<input type="checkbox"/> Cooler 7 _____ °C	<input type="checkbox"/> Cooler 8 _____ °C
Thermometer:	A - S/N _____	B - S/N 210760237	C - S/N _____
			<input checked="" type="checkbox"/> D - S/N 210760272

Comments: Temperature is out of range. PM was informed IMMEDIATELY.

Note: _____

DISCREPANCIES

LabSampleID	LabSampleContainerID	Code	ClientSample Label ID / Information	Corrective Action
1	1-6	D10		P8
<i>[Large diagonal scribble across the table]</i>				

pH holding time requirement for water samples is 15 mins. Water samples for pH analysis are received beyond 15 minutes from sampling time.

NOTES/OBSERVATIONS:
 SAMPLE MATRIX IS DRINKING WATER? YES NO

- LEGEND:**
- | | | |
|-------------------------------------------------------------------------------------|-----------------------------------------------|----------------------------------------------------------------------------------------|
| Code Description-Sample Management | Code Description-Sample Management | Code Description-Sample Management |
| D1 Analysis is not indicated in _____ | D13 Out of Holding Time | R1 Proceed as indicated in <input type="checkbox"/> COC <input type="checkbox"/> Label |
| D2 Analysis mismatch COC vs label | D14 Bubble is >6mm | R2 Refer to attached instruction |
| D3 Sample ID mismatch COC vs label | D15 No trip blank in cooler | R3 Cancel the analysis |
| D4 Sample ID is not indicated in _____ | D16 Preservation not indicated in _____ | R4 Use vial with smallest bubble first |
| D5 Container -[improper] [leaking] [broken] | D17 Preservation mismatch COC vs label | R5 Log-in with latest sampling date and time+1 min |
| D6 Date/Time is not indicated in _____ | D18 Insufficient chemical preservative | R6 Adjust pH as necessary |
| D7 Date/Time mismatch COC vs label | D19 Insufficient Sample | R7 Filter and preserved as necessary |
| D8 Sample listed in COC is not received | D20 No filtration info for dissolved analysis | R8 _____ |
| D9 Sample received is not listed in COC | D21 No sample for moisture determination | R9 _____ |
| <input checked="" type="checkbox"/> D10 No initial/date on corrections in COC/label | D22 _____ | R10 _____ |
| D11 Container count mismatch COC vs received | D23 _____ | R11 _____ |
| D12 Container size mismatch COC vs received | D24 _____ | R12 _____ |

REVIEWS: Sample Labeling **Maria Rivera** SRF **[Signature]** Date **08/17/22**

PM **[Signature]** Date **8/19/22**

REPORTING CONVENTIONS

DATA QUALIFIERS:

Lab Qualifier	AFCEE Qualifier	Description
J	F	Indicates that the analyte is positively identified and the result is less than RL but greater than MDL.
N		Indicates presumptive evidence of a compound.
B	B	Indicates that the analyte is found in the associated method blank as well as in the sample at above QC level.
E	J	Indicates that the result is above the maximum calibration range or estimated value.
*	*	Out of QC limit.

Note: The above qualifiers are used to flag the results unless the project requires a different set of qualification criteria.

ACRONYMS AND ABBREVIATIONS:

CRDL	Contract Required Detection Limit
RL	Reporting Limit
MRL	Method Reporting Limit
PQL	Practical Quantitation Limit
MDL	Method Detection Limit
DO	Diluted out

DATES

The date and time information for leaching and preparation reflect the beginning date and time of the procedure unless the method, protocol, or project specifically requires otherwise.

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

380-16794

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

SDG#: 22H218



CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 380-16794

SDG : 22H218

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

A total of two(2) water samples were received on 08/17/22 to be analyzed for Total Petroleum Hydrocarbons by Purge and Trap in accordance with Method 5030B/8015B and project specific requirements.

Holding Time

Samples were analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. VGH7H09B - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of LCS/LCD was analyzed. VGH7H09L/VGH7H09C were within LCS limits. Refer to LCS summary form for details.

Matrix QC Sample

Matrix spike sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of MS/MSD was analyzed. Gasoline was within MS QC limits in H242-01M/H242-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogate was added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

Samples were analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

LAB CHRONICLE
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

Client : EUROFINS EATON ANALYTICAL
 Project : 380-16794
 SDG NO. : 22H218
 Instrument ID : H7

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	Analysis DateTime	Extraction DateTime	Sample Data FN	Calibration Data FN	Prep. Batch	Notes
									WATER
M8LKT1W	VGH7H09B	1	NA	08/23/2213:39	08/23/2213:39	AH23005A	AH23004A	22VGH7H09	Method Blank
LCS1W	VGH7H09L	1	NA	08/23/2214:14	08/23/2214:14	AH23006A	AH23004A	22VGH7H09	Lab Control Sample (LCS)
LCD1W	VGH7H09C	1	NA	08/23/2214:50	08/23/2214:50	AH23007A	AH23004A	22VGH7H09	LCS Duplicate
380-16794-1	H218-01	1	NA	08/23/2215:25	08/23/2215:25	AH23008A	AH23004A	22VGH7H09	Field Sample
380-16794-2	H218-02	1	NA	08/23/2216:01	08/23/2216:01	AH23009A	AH23004A	22VGH7H09	Field Sample

FN - Filename
 % Moist - Percent Moisture



1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

SAMPLE RESULTS

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINs EATON ANALYTICAL      Date Collected: 08/15/22 10:20
Project     : 380-16794                      Date Received: 08/17/22
Batch No.   : 22H218                          Date Extracted: 08/23/22 15:25
Sample ID   : 380-16794-1                    Date Analyzed: 08/23/22 15:25
Lab Samp ID: H218-01                          Dilution Factor: 1
Lab File ID: AH23008A                          Matrix: WATER
Ext Btch ID: 22VGH7H09                          % Moisture: NA
Calib. Ref.: AH23004A                          Instrument ID: H7
=====
  
```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
GASOLINE	ND	0.020	0.010	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromofluorobenzene	0.0368	0.0400	92	60-140

Notes:

Parameter H-C Range
Gasoline C6-C10
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 5ml Final Volume : 5ml
Prepared by : SCerva Analyzed by : SCerva

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 08/15/22 10:20
Project     : 380-16794                   Date Received: 08/17/22
Batch No.   : 22H218                       Date Extracted: 08/23/22 16:01
Sample ID   : 380-16794-2                 Date Analyzed: 08/23/22 16:01
Lab Samp ID: H218-02                       Dilution Factor: 1
Lab File ID: AH23009A                       Matrix: WATER
Ext Btch ID: 22VGH7H09                       % Moisture: NA
Calib. Ref.: AH23004A                       Instrument ID: H7
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)
GASOLINE	ND	0.020	0.010

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromofluorobenzene	0.0357	0.0400	89	60-140

Notes:
Parameter H-C Range
Gasoline C6-C10
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.
Sample Amount : 5ml Final Volume : 5ml
Prepared by : SCerva Analyzed by : SCerva

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

QC SUMMARIES

METHOD 5030B/8015B
TOTAL PETROLEUM HYDROCARBONS BY PURGE AND TRAP

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 08/23/22 13:39
Project     : 380-16794                 Date Received: 08/23/22
Batch No.   : 22H218                   Date Extracted: 08/23/22 13:39
Sample ID   : MBLK1W                   Date Analyzed: 08/23/22 13:39
Lab Samp ID: VGH7H09B                 Dilution Factor: 1
Lab File ID: AH23005A                 Matrix: WATER
Ext Btch ID: 22VGH7H09                % Moisture: NA
Calib. Ref.: AH23004A                 Instrument ID: H7
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)
GASOLINE	ND	0.020	0.010

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromofluorobenzene	0.0375	0.0400	94	60-140

Notes:

Parameter H-C Range
Gasoline C6-C10
Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.
Sample Amount : 5ml Final Volume : 5ml
Prepared by : SCerva Analyzed by : SCerva

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 380-16794
BATCH NO. : 22H218
METHOD : 5030B/8015B

MATRIX : WATER		% MOISTURE:NA
DILUTION FACTOR: 1	1	1
SAMPLE ID : MBLK1W	LCS1W	LCD1W
LAB SAMPLE ID : VGH7H09B	VGH7H09L	VGH7H09C
LAB FILE ID : AH23005A	AH23006A	AH23007A
DATE PREPARED : 08/23/22 13:39	08/23/22 14:14	08/23/22 14:50
DATE ANALYZED : 08/23/22 13:39	08/23/22 14:14	08/23/22 14:50
PREP BATCH : 22VGH7H09	22VGH7H09	22VGH7H09
CALIBRATION REF: AH23004A	AH23004A	AH23004A

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Gasoline	ND	0.500	0.464	93	0.500	0.503	101	8	60-130	30

SURROGATE PARAMETER	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	SpikeAmt (mg/L)	LCDResult (mg/L)	LCDRec (%)	QCLimit (%)
Bromofluorobenzene	0.0400	0.0473	118	0.0400	0.0471	118	70-130

MB: Method Blank sample LCS: Lab Control Sample LCD: Lab Control Sample Duplicate

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 380-17089
BATCH NO. : 22H242
METHOD : 5030B/8015B

```

=====
MATRIX      : WATER                               % MOISTURE:NA
DILUTION FACTOR: 1                               1
SAMPLE ID   : 380-17089-1                       380-17089-1MS
LAB SAMPLE ID : H242-01                         H242-01M
LAB FILE ID  : AH23010A                         AH23011A
DATE PREPARED : 08/23/22 16:36                 08/23/22 17:11
DATE ANALYZED : 08/23/22 16:36                 08/23/22 17:11
PREP BATCH   : 22VGH7H09                       22VGH7H09
CALIBRATION REF: AH23004A                      AH23004A
    
```

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QCLimit (%)	MaxRPD (%)
Gasoline	ND	0.500	0.432	86	0.500	0.464	93	7	50-130	30

SURROGATE PARAMETER	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QCLimit (%)
Bromofluorobenzene	0.0400	0.0434	109	0.0400	0.0465	116	60-140

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

LABORATORY REPORT FOR

EUROFINS EATON ANALYTICAL

380-16794

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

SDG#: 22H218

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

CASE NARRATIVE

Client : EUROFINS EATON ANALYTICAL

Project: 380-16794

SDG : 22H218

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

One(1) water sample was received on 08/17/22 to be analyzed for Total Petroleum Hydrocarbons by Extraction in accordance with Method 3520C/8015B and project specific requirements.

Holding Time

The sample was analyzed within the prescribed holding time.

Calibration

Multi-calibration points were generated to establish initial calibration (ICAL). ICAL was verified using a secondary source (ICV). Continuing calibration (CCV) verifications were carried out on a frequency specified by the project. All calibration requirements were within acceptance criteria. Refer to calibration summary forms of ICAL, ICV and CCV for details. MRL was analyzed as required by the project. Refer to MRL summary form for details.

Method Blank

Method blank was prepared and analyzed at the frequency required by the project. For this SDG, one(1) method blank was analyzed. DSH034WB - result was compliant to project requirement. Refer to sample result summary form for details.

Lab Control Sample

Lab control sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) LCS was analyzed. Percent recovery for Diesel was within LCS QC limits in DSH034WL. Refer to LCS summary form for details.

Matrix QC Sample

Matrix spike sample was prepared and analyzed at a frequency required by the project. For this SDG, one(1) set of MS/MSD was analyzed. Diesel was within MS QC limits in 22H180-01M/22H180-01S. Refer to Matrix QC summary form for details.

Surrogate

Surrogates were added on QC and field samples. All surrogate recoveries were within QC limits. Refer to sample result summary forms for details.

Sample Analysis

The sample was analyzed according to prescribed analytical procedures. Results were evaluated in accordance to project requirements. For this SDG, all quality control requirements were met.

LAB CHRONICLE
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client : EUROFINS EATON ANALYTICAL
Project : 380-16794
=====
SDG NO. : 22H218
Instrument ID : D5
=====

```

Client Sample ID	Laboratory Sample ID	Dilution Factor	% Moist	Analysis DateTime	Extraction DateTime	Sample Data FN	Calibration Data FN	Prep. Batch	Notes
									WATER
LCS1W	DSH034WL	1	NA	08/22/2217:16	08/20/2215:00	LH22010A	LH22004A	22DSH034W	Lab Control Sample (LCS)
MBLK1W	DSH034WB	1	NA	08/22/2217:35	08/20/2215:00	LH22011A	LH22004A	22DSH034W	Method Blank
380-16794-1	H218-01	1	NA	08/22/2223:09	08/20/2215:00	LH22029A	LH22024A	22DSH034W	Field Sample

FN - Filename
% Moist - Percent Moisture



1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

SAMPLE RESULTS

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 08/15/22 10:20
Project     : 380-16794                   Date Received: 08/17/22
Batch No.   : 22H218                       Date Extracted: 08/20/22 15:00
Sample ID   : 380-16794-1                 Date Analyzed: 08/22/22 23:09
Lab Samp ID: 22H218-01                    Dilution Factor: 1
Lab File ID: LH22029A                      Matrix: WATER
Ext Btch ID: 22DSH034W                     % Moisture: NA
Calib. Ref.: LH22024A                      Instrument ID: D5
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)	
Diesel	ND	0.027	0.013	
Motor Oil	ND	0.053	0.027	
SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.361	0.530	68	60-130
Hexacosane	0.107	0.132	81	60-130

Notes:

Parameter H-C Range
Diesel C10-C24
Motor Oil C24-C36

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 940ml Final Volume : 5ml
Prepared by : DLI Analyzed by : SDeeso

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

QC SUMMARIES

METHOD 3520C/8015B
TOTAL PETROLEUM HYDROCARBONS BY EXTRACTION

```

=====
Client      : EUROFINS EATON ANALYTICAL   Date Collected: 08/20/22 15:00
Project    : 380-16794                   Date Received: 08/20/22
Batch No.  : 22H218                       Date Extracted: 08/20/22 15:00
Sample ID  : MBLK1W                       Date Analyzed: 08/22/22 17:35
Lab Samp ID: DSH034WB                     Dilution Factor: 1
Lab File ID: LH22011A                     Matrix: WATER
Ext Btch ID: 22DSH034W                   % Moisture: NA
Calib. Ref.: LH22004A                    Instrument ID: D5
=====

```

PARAMETERS	RESULTS (mg/L)	RL (mg/L)	MDL (mg/L)
Diesel	ND	0.025	0.012
Motor Oil	ND	0.050	0.025

SURROGATE PARAMETERS	RESULT	SPK_AMT	%RECOVERY	QC LIMIT
Bromobenzene	0.380	0.500	76	60-130
Hexacosane	0.113	0.125	90	60-130

Notes:

Parameter H-C Range
Diesel C10-C24
Motor Oil C24-C36

Reported ND at RL quantitated per pattern recognition.

Detection limits are reported relative to sample result significant figures.

Sample Amount : 1000ml Final Volume : 5ml
Prepared by : DLi Analyzed by : SDeeso

EMAX QUALITY CONTROL DATA
LAB CONTROL SAMPLE ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 380-16794
BATCH NO. : 22H218
METHOD : 3520C/8015B

=====

MATRIX	: WATER	% MOISTURE:NA
DILUTION FACTOR:	1	1
SAMPLE ID	: MBLK1W	LCS1W
LAB SAMPLE ID	: DSH034WB	DSH034WL
LAB FILE ID	: LH22011A	LH22010A
DATE PREPARED	: 08/20/22 15:00	08/20/22 15:00
DATE ANALYZED	: 08/22/22 17:35	08/22/22 17:16
PREP BATCH	: 22DSH034W	22DSH034W
CALIBRATION REF:	LH22004A	LH22004A

ACCESSION:

PARAMETERS	MBResult (mg/L)	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
-----	-----	-----	-----	-----	-----
Diesel	ND	2.50	2.11	84	50-130

SURROGATE PARAMETERS	SpikeAmt (mg/L)	LCSResult (mg/L)	LCSRec (%)	QCLimit (%)
-----	-----	-----	-----	-----
Bromobenzene	0.500	0.377	75	60-130
Hexacosane	0.125	0.116	93	60-130

=====

MB: Method Blank sample LCS: Lab Control Sample

EMAX QUALITY CONTROL DATA
MS/MSD ANALYSIS

CLIENT : EUROFINS EATON ANALYTICAL
PROJECT : 380-16110
BATCH NO. : 22H180
METHOD : 3520C/8015B

```

=====
MATRIX      : WATER                                     % MOISTURE:NA
DILUTION FACTOR: 1                                     1
SAMPLE ID   : 380-16110-1                             380-16110-1MS
LAB SAMPLE ID : 22H180-01                             22H180-01S
LAB FILE ID  : LH22015A                               LH22017A
DATE PREPARED : 08/20/22 15:00                       08/20/22 15:00
DATE ANALYZED : 08/22/22 18:49                       08/22/22 19:26
PREP BATCH   : 22DSH034W                             22DSH034W
CALIBRATION REF: LH22004A                             LH22004A
    
```

ACCESSION:

PARAMETERS	PSResult (mg/L)	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	RPD (%)	QLLimit (%)	MaxRPD (%)
Diesel	ND	2.75	2.82	103	2.58	2.32	90	19	50-130	30

SURROGATE PARAMETERS	SpikeAmt (mg/L)	MSResult (mg/L)	MSRec (%)	SpikeAmt (mg/L)	MSDResult (mg/L)	MSDRec (%)	QLLimit (%)
Bromobenzene	0.550	0.461	84	0.515	0.356	69	60-130
Hexacosane	0.138	0.132	96	0.129	0.118	92	60-130

PS: Parent Sample MS: Matrix Spike MSD: Matrix Spike Duplicate

August 22, 2022

Debbie Frank
Eurofins Eaton Analytical
750 Royal Oaks Drive
Suite 100
Monrovia, CA 91016-

Project Name: RED-HILL Project # 38001111 Job # 380-16794-1
Physis Project ID: 1407003-274

Dear Debbie,


Enclosed are the analytical results for the sample submitted to PHYSIS Environmental Laboratories, Inc. (PHYSIS) on 8/17/2022. A total of 1 sample was received for analysis in accordance with the attached chain of custody (COC). Per the COC, the sample was analyzed for:

Organics
Polynuclear Aromatic Hydrocarbons by EPA 625.1
Disalicylidenepropanediamine by EPA 625.1
Dibenzo [a,l] Pyrene w/ PAHs by EPA 625.1

Analytical results in this report apply only to samples submitted to PHYSIS in accordance with the COC and are intended to be considered in their entirety.

Please feel free to contact me at any time with any questions. PHYSIS appreciates the opportunity to provide you with our analytical and support services.

Regards,


Misty Mercier
714 602-5320
Extension 202
mistymercier@physislabs.com

PROJECT SAMPLE LIST

Eurofins Eaton Analytical

PHYSIS Project ID: 1407003-274

RED-HILL Project # 38001111 Job # 380-16794-1

Total Samples: 1

PHYSIS ID	Sample ID	Description	Date	Time	Matrix	Sample Type
99234	Halawa Wells Pump 1	(380-16794-1)	8/15/2022	10:20	Samplewater	Not Specified

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

ABBREVIATIONS and ACRONYMS

QM	Quality Manual
QA	Quality Assurance
QC	Quality Control
MDL	method detection limit
RL	reporting limit
R1	project sample
R2	project sample replicate
MS1	matrix spike
MS2	matrix spike replicate
B1	procedural blank
B2	procedural blank replicate
BS1	blank spike
BS2	blank spike replicate
LCS1	laboratory control spike
LCS2	laboratory control spike replicate
LCM1	laboratory control material
LCM2	laboratory control material replicate
CRM1	certified reference material
CRM2	certified reference material replicate
RPD	relative percent difference
LMW	low molecular weight
HMW	high molecular weight

QUALITY ASSURANCE SUMMARY

LABORATORY BATCH: Physis' QM defines a laboratory batch as a group of 20 or fewer project samples of similar matrix, processed together under the same conditions and with the same reagents. QC samples are associated with each batch and were used to assess the validity of the sample analyses.

PROCEDURAL BLANK: Laboratory contamination introduced during method use is assessed through the preparation and analysis of procedural blanks is provided at a minimum frequency of one per batch.

ACCURACY: Accuracy of analytical measurements is the degree of closeness based on percent recovery calculations between measured values and the actual or true value and includes a combination of reproducibility error and systematic bias due to sampling and analytical operations. Accuracy of the project data was indicated by analysis of MS, BS, LCS, LCM, CRM, and/or surrogate spikes on a minimum frequency of one per batch. Physis' QM requires that 95% of the target compounds greater than 10 times the MDL be within the specified acceptance limits.

PRECISION: Precision is the agreement among a set of replicate measurements without assumption of knowledge of the true value and is based on RPD calculations between repeated values. Precision of the project data was determined by analysis of replicate MS₁/MS₂, BS₁/BS₂, LCS₁/LCS₂, LCM₁/LCM₂, CRM₁/CRM₂, surrogate spikes and/or replicate project sample analysis (R₁/R₂) on a minimum frequency of one per batch. Physis' QM requires that for 95% of the compounds greater than 10 times the MDL, the percent RPD should be within the specified acceptance range.

BLANK SPIKES: BS is the introduction of a known concentration of analyte into the procedural blank. BS demonstrates performance of the preparation and analytical methods on a clean matrix void of potential matrix related interferences. The BS is performed in laboratory deionized water, making these recoveries a better indicator of the efficiency of the laboratory method per se.

MATRIX SPIKES: MS is the introduction of a known concentration of analyte into a sample. MS samples demonstrate the effect a particular project sample matrix has on the accuracy of a measurement. Individually, MS samples also indicate the bias of analytical measurements due to chemical interferences inherent in the in the specific project sample spiked. Intrinsic target analyte concentration in the specific project sample can also significantly impact MS recovery.

CERTIFIED REFERENCE MATERIALS: CRMs are materials of various matrices for which analytical information has been determined and certified by a recognized authority. These are used to provide a quantitative assessment of the accuracy of an analytical method. CRMs provide evidence that the laboratory preparation and analysis produces results that are comparable to those obtained by an independent organization.

LABORATORY CONTROL MATERIAL: LCM is provided because a suitable natural seawater CRM is not available and can be used to indicate accuracy of the method. Physis' internal LCM is seawater collected at ~800 meters in the Southern California San Pedro Basin and can be used as a reference for background concentrations in clean, natural seawater for comparison to project samples.

LABORATORY CONTROL SPIKES: LCS is the introduction of a known concentration of analyte into Physis' LCM. LCS samples were employed to assess the effect the seawater matrix has on the accuracy of a measurement. LCS also indicate the bias of this method due to chemical interferences inherent in the in the seawater matrix. Intrinsic LCM concentration can also significantly impact LCS recovery.

SURROGATES: A surrogate is a pure analyte unlikely to be found in any project sample, behaves similarly to

the target analyte and most often used with organic analytical procedures. Surrogates are added in known concentration to all samples and are measured to indicate overall efficiency of the method including processing and analyses.

HOLDING TIME: Method recommended holding times are the length of time a project sample can be stored under specific conditions after collection and prior to analysis without significantly affecting the analyte's concentration. Holding times can be extended if preservation techniques are employed to reduce biodegradation, volatilization, oxidation, sorption, precipitation, and other physical and chemical processes.

SAMPLE STORAGE/RETENTION: In order to maintain chemical integrity prior to analysis, all samples submitted to Physis are refrigerated (liquids) or frozen (solids) upon receipt unless otherwise recommended by applicable methods. Solid samples are retained for 1 year from collection while liquid samples are retained until method recommended holding times elapse.

TOTAL/DISSOLVED FRACTION: In some instances, the results for the dissolved fraction may be higher than the total fraction for a particular analyte (e.g. trace metals). This is typically caused by the analytical variation for each result and indicates that the target analyte is primarily in the dissolved phase, within the sample.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

PHYSIS QUALIFIER CODES

CODE	DEFINITION
#	see Case Narrative
ND	analyte not detected at or above the MDL
B	analyte was detected in the procedural blank greater than 10 times the MDL
E	analyte concentration exceeds the upper limit of the linear calibration range, reported value is estimated
H	sample received and/or analyzed past the recommended holding time
J	analyte was detected at a concentration below the RL and above the MDL, reported value is estimated
N	insufficient sample, analysis could not be performed
M	analyte was outside the specified accuracy and/or precision acceptance limits due to matrix interference. The associated B/BS were within limits, therefore the sample data was reported without further clarification
SH	analyte concentration in the project sample exceeded the spike concentration, therefore accuracy and/or precision acceptance limits do not apply
SL	analyte results were lower than 10 times the MDL, therefore accuracy and/or precision acceptance limits do not apply
NH	project sample was heterogeneous and sample homogeneity could not be readily achieved using routine laboratory practices, therefore accuracy and/or precision acceptance limits do not apply
Q	analyte was outside the specified QAPP acceptance limits for precision and/or accuracy but within Physis derived acceptance limits, therefore the sample data was reported without further clarification
R	Physis' QM allows for 5% of the target compounds greater than 10 times the MDL to be outside the specified acceptance limits for precision and/or accuracy. This is often due to random error and does not indicate any significant problems with the analysis of these project samples

CASE NARRATIVE

QUALIFIER NOTES

In addition to the use of analyte specific Physis Qualifier Codes where applicable, the following were also noted.

ND

MDL is listed due to report format restrictions; it is not used in reporting. Analytical results reported are ND at the RL.

ANALYTICALS

REPORT

TERRA AURA
ENVIRONMENTAL LABORATORIES, INC.

Innovative Solutions for Nature

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

Base/Neutral Extractable Compounds

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Sample ID: 99234-R1	Halawa Wells Pump 1 (380-16794-1) Matrix: Samplewater						Sampled:	15-Aug-22 10:20	Received:	17-Aug-22	
Disalicylidenepropanediamine	EPA 625.1	µg/L	ND	1	0.05	0.1	Total		O-38084	17-Aug-22	21-Aug-22



Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Sample ID: 99234-R1	Halawa Wells Pump 1 (380-16794-1) Matrix: Samplewater						Sampled:	15-Aug-22 10:20	Received:	17-Aug-22	
(d10-Acenaphthene)	EPA 625.1	% Recovery	80	1			Total		O-38084	17-Aug-22	21-Aug-22
(d10-Phenanthrene)	EPA 625.1	% Recovery	89	1			Total		O-38084	17-Aug-22	21-Aug-22
(d12-Chrysene)	EPA 625.1	% Recovery	90	1			Total		O-38084	17-Aug-22	21-Aug-22
(d12-Perylene)	EPA 625.1	% Recovery	83	1			Total		O-38084	17-Aug-22	21-Aug-22
(d8-Naphthalene)	EPA 625.1	% Recovery	79	1			Total		O-38084	17-Aug-22	21-Aug-22
1-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
1-Methylphenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
2,3,5-Trimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
2,6-Dimethylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
2-Methylnaphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Acenaphthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Acenaphthylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Benz[a]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Benzo[a]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Benzo[b]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Benzo[e]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Benzo[g,h,i]perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Benzo[k]fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Biphenyl	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Chrysene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Dibenz[a,h]anthracene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Dibenzo[a,l]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Dibenzothiophene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22

Polynuclear Aromatic Hydrocarbons

ANALYTE	Method	Units	RESULT	DF	MDL	RL	Fraction	QA CODE	Batch ID	Date Processed	Date Analyzed
Fluoranthene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Fluorene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Indeno[1,2,3-cd]pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Naphthalene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Perylene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Phenanthrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22
Pyrene	EPA 625.1	µg/L	ND	1	0.001	0.005	Total		O-38084	17-Aug-22	21-Aug-22



QUALITY CONTROL REPORT

TERRA CONSULTING AURA ENVIRONMENTAL LABORATORIES, INC.

Innovative Solutions for Nature

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

Base/Neutral Extractable Compounds

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODEc
							LEVEL	RESULT	%	LIMITS	%	LIMITS	
Sample ID: 99233-B1		QAQC Procedural Blank			Matrix: BlankMatrix			Sampled:		Received:			
		Method: EPA 625.1			Batch ID: O-38084			Prepared: 15-Aug-22		Analyzed: 20-Aug-22			
Disalicylidenepropanediamine	Total	ND	1	0.05	0.1	µg/L							
Sample ID: 99233-BS1		QAQC Procedural Blank			Matrix: BlankMatrix			Sampled:		Received:			
		Method: EPA 625.1			Batch ID: O-38084			Prepared: 15-Aug-22		Analyzed: 20-Aug-22			
Disalicylidenepropanediamine	Total	18.1	1	0.05	0.1	µg/L	25	0	72	50 - 150%	PASS		
Sample ID: 99233-BS2		QAQC Procedural Blank			Matrix: BlankMatrix			Sampled:		Received:			
		Method: EPA 625.1			Batch ID: O-38084			Prepared: 15-Aug-22		Analyzed: 20-Aug-22			
Disalicylidenepropanediamine	Total	23.3	1	0.05	0.1	µg/L	25	0	93	50 - 150%	PASS	25	30 PASS

Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY	PRECISION	QA CODEc
							LEVEL	RESULT	% LIMITS	% LIMITS	
Sample ID: 99233-B1		QAQC Procedural Blank			Matrix: BlankMatrix		Sampled:		Received:		
		Method: EPA 625.1				Batch ID: O-38084	Prepared: 15-Aug-22		Analyzed: 20-Aug-22		
(d10-Acenaphthene)	Total	92	1			% Recovery	100	92	65 - 113%	PASS	
(d10-Phenanthrene)	Total	93	1			% Recovery	100	93	80 - 111%	PASS	
(d12-Chrysene)	Total	92	1			% Recovery	100	92	60 - 139%	PASS	
(d12-Perylene)	Total	88	1			% Recovery	100	88	36 - 161%	PASS	
(d8-Naphthalene)	Total	88	1			% Recovery	100	88	44 - 119%	PASS	
1-Methylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
1-Methylphenanthrene	Total	ND	1	0.001	0.005	µg/L					
2,3,5-Trimethylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
2,6-Dimethylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
2-Methylnaphthalene	Total	ND	1	0.001	0.005	µg/L					
Acenaphthene	Total	ND	1	0.001	0.005	µg/L					
Acenaphthylene	Total	ND	1	0.001	0.005	µg/L					
Anthracene	Total	ND	1	0.001	0.005	µg/L					
Benz[a]anthracene	Total	ND	1	0.001	0.005	µg/L					
Benzo[a]pyrene	Total	ND	1	0.001	0.005	µg/L					
Benzo[b]fluoranthene	Total	ND	1	0.001	0.005	µg/L					
Benzo[e]pyrene	Total	ND	1	0.001	0.005	µg/L					
Benzo[g,h,i]perylene	Total	ND	1	0.001	0.005	µg/L					
Benzo[k]fluoranthene	Total	ND	1	0.001	0.005	µg/L					
Biphenyl	Total	ND	1	0.001	0.005	µg/L					
Chrysene	Total	ND	1	0.001	0.005	µg/L					
Dibenz[a,h]anthracene	Total	ND	1	0.001	0.005	µg/L					
Dibenzo[a,l]pyrene	Total	ND	1	0.001	0.005	µg/L					
Dibenzothiophene	Total	ND	1	0.001	0.005	µg/L					

Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODEc
							LEVEL	RESULT	%	LIMITS	%	LIMITS	
Fluoranthene	Total	ND	1	0.001	0.005	µg/L							
Fluorene	Total	ND	1	0.001	0.005	µg/L							
Indeno[1,2,3-cd]pyrene	Total	ND	1	0.001	0.005	µg/L							
Naphthalene	Total	ND	1	0.001	0.005	µg/L							
Perylene	Total	ND	1	0.001	0.005	µg/L							
Phenanthrene	Total	ND	1	0.001	0.005	µg/L							
Pyrene	Total	ND	1	0.001	0.005	µg/L							



Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY	PRECISION	QA CODEc	
							LEVEL	RESULT	% LIMITS	% LIMITS		
Sample ID: 99233-BS1		QAQC Procedural Blank			Matrix: BlankMatrix			Sampled:		Received:		
Method: EPA 625.1		Batch ID: O-38084			Prepared: 15-Aug-22			Analyzed: 20-Aug-22				
(d10-Acenaphthene)	Total	87	1			% Recovery	100	0	87	65 - 113%	PASS	
(d10-Phenanthrene)	Total	91	1			% Recovery	100	0	91	80 - 111%	PASS	
(d12-Chrysene)	Total	84	1			% Recovery	100	0	84	60 - 139%	PASS	
(d12-Perylene)	Total	86	1			% Recovery	100	0	86	36 - 161%	PASS	
(d8-Naphthalene)	Total	81	1			% Recovery	100	0	81	44 - 119%	PASS	
1-Methylnaphthalene	Total	0.414	1	0.001	0.005	µg/L	0.5	0	83	49 - 117%	PASS	
1-Methylphenanthrene	Total	0.434	1	0.001	0.005	µg/L	0.5	0	87	66 - 127%	PASS	
2,3,5-Trimethylnaphthalene	Total	0.427	1	0.001	0.005	µg/L	0.5	0	85	57 - 120%	PASS	
2,6-Dimethylnaphthalene	Total	0.413	1	0.001	0.005	µg/L	0.5	0	83	54 - 117%	PASS	
2-Methylnaphthalene	Total	0.41	1	0.001	0.005	µg/L	0.5	0	82	47 - 130%	PASS	
Acenaphthene	Total	0.425	1	0.001	0.005	µg/L	0.5	0	85	53 - 131%	PASS	
Acenaphthylene	Total	0.418	1	0.001	0.005	µg/L	0.5	0	84	43 - 140%	PASS	
Anthracene	Total	0.449	1	0.001	0.005	µg/L	0.5	0	90	58 - 135%	PASS	
Benz[a]anthracene	Total	0.365	1	0.001	0.005	µg/L	0.5	0	73	55 - 145%	PASS	
Benzo[a]pyrene	Total	0.401	1	0.001	0.005	µg/L	0.5	0	80	51 - 143%	PASS	
Benzo[b]fluoranthene	Total	0.446	1	0.001	0.005	µg/L	0.5	0	89	46 - 165%	PASS	
Benzo[e]pyrene	Total	0.427	1	0.001	0.005	µg/L	0.5	0	85	42 - 152%	PASS	
Benzo[g,h,i]perylene	Total	0.444	1	0.001	0.005	µg/L	0.5	0	89	63 - 133%	PASS	
Benzo[k]fluoranthene	Total	0.426	1	0.001	0.005	µg/L	0.5	0	85	56 - 145%	PASS	
Biphenyl	Total	0.422	1	0.001	0.005	µg/L	0.5	0	84	56 - 119%	PASS	
Chrysene	Total	0.384	1	0.001	0.005	µg/L	0.5	0	77	56 - 141%	PASS	
Dibenz[a,h]anthracene	Total	0.407	1	0.001	0.005	µg/L	0.5	0	81	55 - 150%	PASS	
Dibenzo[a,l]pyrene	Total	0.499	1	0.001	0.005	µg/L	0.5	0	100	50 - 150%	PASS	
Dibenzothiophene	Total	0.436	1	0.001	0.005	µg/L	0.5	0	87	75 - 113%	PASS	

Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODEc
							LEVEL	RESULT	%	LIMITS	%	LIMITS	
Fluoranthene	Total	0.442	1	0.001	0.005	µg/L	0.5	0	88	60 - 146%	PASS		
Fluorene	Total	0.432	1	0.001	0.005	µg/L	0.5	0	86	58 - 131%	PASS		
Indeno[1,2,3-cd]pyrene	Total	0.415	1	0.001	0.005	µg/L	0.5	0	83	50 - 151%	PASS		
Naphthalene	Total	0.402	1	0.001	0.005	µg/L	0.5	0	80	41 - 126%	PASS		
Perylene	Total	0.4	1	0.001	0.005	µg/L	0.5	0	80	48 - 141%	PASS		
Phenanthrene	Total	0.45	1	0.001	0.005	µg/L	0.5	0	90	67 - 127%	PASS		
Pyrene	Total	0.425	1	0.001	0.005	µg/L	0.5	0	85	54 - 156%	PASS		

Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODEc	
							LEVEL	RESULT	%	LIMITS	%	LIMITS		
Sample ID: 99233-BS2		QAQC Procedural Blank			Matrix: BlankMatrix			Sampled:			Received:			
		Method: EPA 625.1			Batch ID: O-38084			Prepared: 15-Aug-22			Analyzed: 20-Aug-22			
(d10-Acenaphthene)	Total	92	1			% Recovery	100	0	92	65 - 113%	PASS	6	30	PASS
(d10-Phenanthrene)	Total	98	1			% Recovery	100	0	98	80 - 111%	PASS	7	30	PASS
(d12-Chrysene)	Total	88	1			% Recovery	100	0	88	60 - 139%	PASS	5	30	PASS
(d12-Perylene)	Total	95	1			% Recovery	100	0	95	36 - 161%	PASS	10	30	PASS
(d8-Naphthalene)	Total	90	1			% Recovery	100	0	90	44 - 119%	PASS	11	30	PASS
1-Methylnaphthalene	Total	0.442	1	0.001	0.005	µg/L	0.5	0	88	49 - 117%	PASS	6	30	PASS
1-Methylphenanthrene	Total	0.443	1	0.001	0.005	µg/L	0.5	0	89	66 - 127%	PASS	2	30	PASS
2,3,5-Trimethylnaphthalene	Total	0.45	1	0.001	0.005	µg/L	0.5	0	90	57 - 120%	PASS	6	30	PASS
2,6-Dimethylnaphthalene	Total	0.44	1	0.001	0.005	µg/L	0.5	0	88	54 - 117%	PASS	6	30	PASS
2-Methylnaphthalene	Total	0.439	1	0.001	0.005	µg/L	0.5	0	88	47 - 130%	PASS	7	30	PASS
Acenaphthene	Total	0.449	1	0.001	0.005	µg/L	0.5	0	90	53 - 131%	PASS	6	30	PASS
Acenaphthylene	Total	0.449	1	0.001	0.005	µg/L	0.5	0	90	43 - 140%	PASS	7	30	PASS
Anthracene	Total	0.427	1	0.001	0.005	µg/L	0.5	0	85	58 - 135%	PASS	6	30	PASS
Benz[a]anthracene	Total	0.387	1	0.001	0.005	µg/L	0.5	0	77	55 - 145%	PASS	5	30	PASS
Benzo[a]pyrene	Total	0.447	1	0.001	0.005	µg/L	0.5	0	89	51 - 143%	PASS	11	30	PASS
Benzo[b]fluoranthene	Total	0.472	1	0.001	0.005	µg/L	0.5	0	94	46 - 165%	PASS	5	30	PASS
Benzo[e]pyrene	Total	0.453	1	0.001	0.005	µg/L	0.5	0	91	42 - 152%	PASS	7	30	PASS
Benzo[g,h,i]perylene	Total	0.465	1	0.001	0.005	µg/L	0.5	0	93	63 - 133%	PASS	4	30	PASS
Benzo[k]fluoranthene	Total	0.468	1	0.001	0.005	µg/L	0.5	0	94	56 - 145%	PASS	10	30	PASS
Biphenyl	Total	0.455	1	0.001	0.005	µg/L	0.5	0	91	56 - 119%	PASS	8	30	PASS
Chrysene	Total	0.4	1	0.001	0.005	µg/L	0.5	0	80	56 - 141%	PASS	4	30	PASS
Dibenz[a,h]anthracene	Total	0.439	1	0.001	0.005	µg/L	0.5	0	88	55 - 150%	PASS	8	30	PASS
Dibenzo[a,l]pyrene	Total	0.512	1	0.001	0.005	µg/L	0.5	0	102	50 - 150%	PASS	2	30	PASS
Dibenzothiophene	Total	0.461	1	0.001	0.005	µg/L	0.5	0	92	75 - 113%	PASS	6	30	PASS

Polynuclear Aromatic Hydrocarbons

QUALITY CONTROL REPORT

ANALYTE	FRACTION	RESULT	DF	MDL	RL	UNITS	SPIKE	SOURCE	ACCURACY		PRECISION		QA CODEc	
							LEVEL	RESULT	%	LIMITS	%	LIMITS		
Fluoranthene	Total	0.459	1	0.001	0.005	µg/L	0.5	0	92	60 - 146%	PASS	4	30	PASS
Fluorene	Total	0.449	1	0.001	0.005	µg/L	0.5	0	90	58 - 131%	PASS	5	30	PASS
Indeno[1,2,3-cd]pyrene	Total	0.442	1	0.001	0.005	µg/L	0.5	0	88	50 - 151%	PASS	6	30	PASS
Naphthalene	Total	0.437	1	0.001	0.005	µg/L	0.5	0	87	41 - 126%	PASS	8	30	PASS
Perylene	Total	0.428	1	0.001	0.005	µg/L	0.5	0	86	48 - 141%	PASS	7	30	PASS
Phenanthrene	Total	0.47	1	0.001	0.005	µg/L	0.5	0	94	67 - 127%	PASS	4	30	PASS
Pyrene	Total	0.443	1	0.001	0.005	µg/L	0.5	0	89	54 - 156%	PASS	5	30	PASS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

PHYSIS

TENTATIVELY

IDENTIFIED COMPOUNDS

ENVIRONMENTAL LABORATORIES, INC.

Innovative Solutions for Nature

Sample ID: 99234

RT	Area Pct	Concentration (ng/L)	Library/ID	Cas Number	Match Qual
31.6351	5.5081	1111	Anthracene-D10-	1719-06-8	92
24.2293	3.1342	632	Diethyl Phthalate	84-66-2	98
42.1448	1.9797	399	Terephthalic acid, isobutyl butyl ester	1000323-56-2	95
14.3868	0.8342	168	Cyclohexane, 1,2,4,5-tetraethyl-, (1.alpha.,2.alpha.,4.alpha.,5.alpha.)-	61142-24-3	83
63.5225	0.6349	128	1,4-Benzenedicarboxylic acid, bis(2-ethylhexyl) ester	6422-86-2	96

Concentration estimated using the response for Anthracene-d10

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Sample ID: Lab Blank Batch O-38084

RT	Area Pct	Concentration (ng/L)	Library/ID	Cas Number	Match Qual
31.6482	4.7930	1111	Anthracene-D10-	1719-06-8	95
42.1583	1.5901	369	Terephthalic acid, isobutyl butyl ester	1000323-56-2	95
14.3915	0.6801	158	3-Hexene, 3-ethyl-2,5-dimethyl-	62338-08-3	83
63.5413	0.4564	106	1,4-Benzenedicarboxylic acid, bis(2-ethylhexyl) ester	6422-86-2	95

Concentration estimated using the response for Anthracene-d10

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

PERFORMANCE CHAIN OF CUSTODY

TERRA ENVIRONMENTAL LABORATORIES, INC. AURA

Innovative Solutions for Nature

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Monrovia, CA (Suite 100)
 750 Royal Oaks Drive Suite 100
 Monrovia, CA 91016
 Phone: 826-386-1100

Chain of Custody Record



Client Information (Sub Contract Lab)

Client Contact: **Frank, Debbie L** Lab Pk: **Frank, Debbie L**

Shipping/Receiving: **Debbie.Frank@eurofins.com** E-Mail: **Debbie.Frank@eurofins.com**

Company: **Physis Environmental Laboratories** Accreditations Required (See note): **State - Hawaii**

Address: **1904 Wright Circle,** Due Date Requested: **8/30/2022**

City: **Anaheim** TAT Requested (days):

State Zip: **CA, 92806**

Phone: PO #:

Email: W/C #:

Project Name: **RED-HILL** Project #: **38001111**

Site: **Honolulu BWS Sites** SSCW#:

Analysis Requested

Field Filtered Sample (Yes or No)

Perform MS/MSD (Yes or No)

SUB (625 PAH Physis LL (EAL) + TICs)/ 625 PAH Physis LL (EAL) + TICs

Sample Identification - Client ID (Lab ID)	Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	MATRIX (Water, Sewage, Groundwater, Bacteria, Anal)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	Total Number of containers	Special Instructions/Note:
Halawa Wells Pump 1 (380-16794-1)	8/15/22	10:20	Water			X	4	See Attached Instructions

Possible Hazard Identification

Deliverable Requested: I, II, III, IV, Other (specify) **Primary Deliverable Rank: 2**

Empty Kit Relinquished by: Date: Time: Method of Shipment:

Relinquished by: **GP** Date/Time: **08/17/2022** Company: **ESA** Received by: **AD** Date/Time: **8/17/22 12** Company: **ESA**

Relinquished by: **MS** Date/Time: **8/17/22 15:40** Company: Received by: **MS** Date/Time: **8/17/22 15:46** Company: **MSIS**

Custody Seals Intact: Yes No Custody Seal No.: Cooler Temperature(s) °C and Other Remarks:

Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)

Return To Client Disposal By Lab Archive For **Months**

Special Instructions/Note:

Preservation Codes:
 A - HCL
 B - NaOH
 C - Zn Acetate
 D - Nitric Acid
 E - NaHSO4
 F - MeOH
 G - Another
 H - Ascorbic Acid
 I - Ice
 J - DI Water
 K - EDTA
 L - EDA
 M - Hexane
 N - None
 O - AsHClO2
 P - Na2CO3
 Q - Na2SO3
 R - Na2S2O3
 S - H2SO4
 T - TSP Dodecahydrate
 U - Acetone
 V - MCAA
 W - pH 4-5
 Y - Trizma
 Z - other (specify)

Project Iteration ID: 1407003-274
 Client Name: Eurofins Eaton Analytical
 Project Name: RED-HILL Project # 38001111
 Job # 380-16794-1
 COC Page Number: 2 of 2
 Bottle Label Color: NA

Sample Receipt Summary

Receiving Info

1. Initials Received By: YK
2. Date Received: 8/17/22
3. Time Received: 1540
4. Client Name: Eurofins
5. Courier Information: (Please circle)
 - Client
 - UPS
 - Area Fast
 - DRS
 - FedEx
 - GSO/GLS
 - Ontrac
 - PAMS
 - PHYSIS Driver:
 - i. Start Time: _____
 - ii. End Time: _____
 - iii. Total Mileage: _____
 - iv. Number of Pickups: _____
6. Container Information: (Please put the # of containers or circle none)
 - 3 Cooler
 - ___ Styrofoam Cooler
 - ___ Boxes
 - None
 - ___ Carboy(s)
 - ___ Carboy Trash Can(s)
 - ___ Carboy Cap(s)
 - Other _____
7. What type of ice was used: (Please circle any that apply)
 - Wet Ice
 - Blue Ice
 - Dry Ice
 - Water
 - None
8. Randomly Selected Samples Temperature (°C): -0.1
 Used I/R Thermometer # 12

Inspection Info

1. Initials Inspected By: RGH

Sample Integrity Upon Receipt:

1. COC(s) included and completely filled out..... Yes / No
2. All sample containers arrived intact..... Yes / No
3. All samples listed on COC(s) are present..... Yes / No
4. Information on containers consistent with information on COC(s)..... Yes / No
5. Correct containers and volume for all analyses indicated..... Yes / No
6. All samples received within method holding time..... Yes / No
7. Correct preservation used for all analyses indicated..... Yes / No
8. Name of sampler included on COC(s)..... Yes / No

Notes:



Eaton Analytical

CHAIN OF CUSTODY RECORD

EUROFINS EATON ANALYTICAL USE ONLY:

750 Royal Oaks Drive, Suite 100
Monrovia, CA 91016-3629

Phone: 626 386 1100
Fax: 626 386 1101

800 566 LABS (800 566 5227)

LOGIN COMMENTS: _____

SAMPLES CHECKED AGAINST COC BY: MB

SAMPLE TEMP RECEIVED AT:

Colton / No. California / Arizona _____ °C (Compliance: 4 ± 2 °C)

Monrovia 7.6 °C (Compliance: 4 ± 2 °C)

SAMPLES LOGGED IN BY: _____

SAMPLES REC'D DAY OF COLLECTION? (check for yes)

CONDITION OF BLUE ICE: Frozen _____ Partially Frozen _____ Thawed _____ Wet Ice _____ No Ice _____

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: _____

TO BE COMPLETED BY SAMPLER:

COMPANY/AGENCY NAME: BWS HONOLULU		PROJECT CODE: Red Hill Special		COMPLIANCE SAMPLES <input type="checkbox"/> NON-COMPLIANCE SAMPLES <input checked="" type="checkbox"/>		REGULATION INVOLVED: _____							
EEA CLIENT CODE:		COC ID:		SAMPLE GROUP: Weekly_RED_HILL (2022)		Type of samples (circle one): ROUTINE <input type="checkbox"/> <u>SPECIAL</u> CONFIRMATION <input type="checkbox"/> (eg. SDWA, Phase V, NPDES, FDA,...)							
TAT requested: rush by adv notice only		STD ___ 1 wk ___ X ___ 3 day ___ 2 day ___ 1 day ___		SEE ATTACHED BOTTLE ORDER FOR ANALYSES <input type="checkbox"/> (check for yes), <u>OR</u>				list ANALYSES REQUIRED (enter number of bottles sent for each test for each sample)					
SAMPLE DATE	SAMPLE TIME	SAMPLE ID	CLIENT LAB ID	MATRIX *	FIELD DATA	FIELD DATA	525.2 PREC+ (MOD) 525plus TICs (3x1L)	625 PAH + MSMSD Volume (3x1L)	TPH8015 Diesel and Motor Oil C (3x1L)	Subcontract - 8015 Gas (Purgeable) (4x10mL)	8015 Gas - C TB (2x10mL)	SAMPLER COMMENTS	
08/15/22	1020	Halawa Wells Pump 1	331-023	RGW			2	4	2	4	2		
												Temp Blank: <u>1</u> °C	



* MATRIX TYPES: RSW = Raw Surface Water CFW = Chlor(am)inated Finished Water SEAW = Sea Water BW = Bottled Water SO = Soil O = Other - Please Identify
 RGW = Raw Ground Water FW = Other Finished Water WW = Waste Water SW = Storm Water SL = Sludge

SIGNATURE	PRINT NAME	COMPANY/TITLE	DATE	TIME
	Lesli Laanui	Honolulu Board of Water Supply	8/15/2022	1020
	Lesli Laanui	Honolulu Board of Water Supply	8/15/2022	1200
	G. RETNER	EEA	08/16/2022	10:15
RECEIVED BY:				

WEEKLY

Bottle Order Information

Bottle Order: RUSH RED-HILL WEEKLY
Bottle Order #:
Request From Client:
Date Order Posted:
Order Status: Ready To Process
Prepared By:
Deliver By Date:
Lab Project Number: 38001111
PWSID: HI0000331

Order Completion Information

Creator:
Filled by:
Sent Date:
Sent Via:
Tracking #:

Method	bottles/set	Bottle Type Description	
525.2 PREC + (MOD) 525plus TICS	2	- 1 L amber glass [45 mg sulfite + 2mL 6N HCL]	2
625 PAH + MS/MSD Volume	4	- 1 L amber glass [1 mL Thio 8%]	4
Subcontract 8015 Diesel LL (EAL) & Motor Oil	2	- 1 L amber glass [1 mL Thio 8%]	2
Subcontract - 8015 Gas (Purgeable)	4	- 40ml amber glass [1 drop Thio + HCl dropper]	4
8015 Gas_C TB	2	- 40ml amber glass vial [1 drop Thio + HCL]	2

Total bottles: 8 (1L) + 4 (40mL) + 2 (TB) = 14

Notes: **WEEKLY** Red Hill set for Halawa Wells Pump 2



Eaton Analytical

INTERNAL CHAIN OF CUSTODY RECORD

EEA Folder Number: _____

SAMPLE TEMP RECEIVED:

Note: If samples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.

SAMPLES REC'D DAY OF COLLECTION? Yes / No

IR Gun ID = 649A (Observation = 4.9 °C) (Corr. Factor -0.3 °C) (Final = 4.6 °C)

TYPE OF ICE: Real Synthetic No Ice CONDITION OF ICE: Frozen Partially Frozen Thawed N/A

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: _____

Compliance Acceptance Criteria:

- 1) Chemistry: >0, ≤ 6°C, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C, not frozen (can be ≥ 10°C if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants

1 = (Observation = _____ °C) (Corr. Factor _____ °C) (Final = _____ °C)	2 = (Observation = _____ °C) (Corr. Factor _____ °C) (Final = _____ °C)
3 = (Observation = _____ °C) (Corr. Factor _____ °C) (Final = _____ °C)	4 = (Observation = _____ °C) (Corr. Factor _____ °C) (Final = _____ °C)

4 Dioxin (1813 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (if received after 24 hrs of sample collection)

5) pH Check. Manufacturer: _____ Lot Number: _____ pH strip type: 0 - 14 or _____ Expiration Date _____ Results: _____

6) Chlorine check. Manufacturer: Sansafe. Lot No.: _____ Expiration Date: _____ Results: _____

7) VOA and Radon Headspace:

No Samples with Headspace:

Samples with Headspace (see below):

Headspace Documentation (use additional VOC and Radon Internal COFC for additional bottles)

Exempt from headspace concerns: Methods 815.4, HAA (8251, 852), 805, SPME, @OH, 832LCMS, 858, 838, Anatoxin, LCMS methods using 40 ml vials, International clients:

Samp ID	Bottle #	None/<8 mm	>8mm	Test	Samp ID	Bottle #	None/<8 mm	>8mm	Test	Samp ID	Bottle #	None/<8 mm	>8mm	Test	Samp ID	Bottle #	None/<8 mm	>8mm	Test	

Note Sample IDs which have dissimilar headspace (i.e. potential sampling errors): _____

SIGNATURE	PRINT NAME	COMPANY/TITLE	DATE	TIME
	GRETNER	Eurofins Eaton Analytical	08/16/2022	10:15
	Am. Boeck	Eurofins Eaton Analytical	8-16-22	1724

ORIGIN ID:HIKA (808) 748-5840
BWS CHEMLAB
HONOLULU BOARD OF WATER SUPPLY
630 S. BERETANIA ST.
CHEMICAL LABORATORY
HONOLULU, HI 96843
UNITED STATES US


SHIP DATE: 15AUG22
ACTWGT: 75.00 LB
CAD: 100205419/INET4490

BILL RECIPIENT

TO **C CHUCK**
EUROFINS EATON ANALYTICAL, INC
750 ROYAL OAKS DR
SUITE 100
MONROVIA CA 91016

581J2/F39D/F E2D

(626) 386-1178 REF:
INV: DEPT:
PO:




TUE - 16 AUG 10:30A
PRIORITY OVERNIGHT

TRK# 7776 6605 7330
0201

WZ WHPA

91016
CA-US BUR



After printing this label:

1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number. Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g. jewelry, precious metals, negotiable instruments and other items listed in our ServiceGuide. Written claims must be filed within strict time limits, see current FedEx Service Guide.



Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-16794-1

Login Number: 16794
List Number: 1
Creator: Segura, Ryan

List Source: Eurofins Eaton Monrovia

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	