



908 North Temperance Ave. ▽ Clovis, CA 93611 ▽ Phone 559-275-2175 ▽ Fax 559-275-4422

Certification Number: CA1312
NELAP Certification number: 05233CA
DoD-ELAP Certificate number: ADE-1410

Data Validatable Report

August 10, 2012

Environet, Inc.
650 Iwilei Road, #204
Honolulu, Hawaii 96817

Attn: Max Solmssen

Title: Report of Data: Case 68258

Project: LTM Red Hill/1022-024

Contract #: Prime contract # for DoD: N62742-08-D-1930, CTO HC21

Dear Mr. Solmssen:

Three water samples were received July 19, 2012, in good condition. Written results for the requested analyses are provided on this August 10, 2012.

Results in this report apply to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.

If you have any questions or require further information, please contact your APPL Project Manager, Cynthia Clark, at your convenience. Thank you for choosing APPL, Inc.

I certify that this data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. These test results meet all requirements of NELAC and DoD QSM. Release of the hard copy has been authorized by the Laboratory Manager or her designee, as verified by the following signature.

A handwritten signature in black ink, appearing to read 'Sharon Dehmlow', written in a cursive style.

Sharon Dehmlow, Laboratory Director
APPL, Inc.

SD/cm
Enclosure
cc: File

Number of pages in this report: 413

Data Validation Package
for
LTM Red Hill / 1022-024
SDG 68258

TABLE OF CONTENTS

LABORATORY NAME: APPL, Inc.

Sample Receipt Information	<u>4</u>
Case Narrative	<u>6</u>
Chain of Custody and ARF	<u>12</u>
Method 8270D SIM	<u>17</u>
QC Summary	<u>18</u>
Sample Data	<u>27</u>
Calibration Data	<u>36</u>
Raw Data	<u>60</u>
Method 8015B TPH-Diesel	<u>77</u>
QC Summary	<u>78</u>
Sample Data	<u>83</u>
Calibration Data	<u>90</u>
Raw Data	<u>126</u>
Method 8260B	<u>141</u>
QC Summary	<u>142</u>
Sample Data	<u>163</u>
Calibration Data	<u>191</u>
Raw Data	<u>327</u>

Method 6020	<u>372</u>
QC Summary	<u>373</u>
Sample Data	<u>376</u>
Calibration Data	<u>381</u>
Raw Data	<u>401</u>

SAMPLE RECEIPT INFORMATION

Sample receipt information

ARF: 68258

Project: Red Hill/1022-024

Sample Receipt Information:

The samples were received on July 19, 2012, at 4.0°C. The samples were assigned Analytical Request Form (ARF) number 68258. The sample numbers and requested analyses were compared to the chain of custody and email communications. The analyses requested was provided by the client. No other exception was encountered.

Sample Table

CLIENT ID	APPL ID	Matrix	Date Sampled	Date Received
ES081	AY65112	WATER	07/18/12	07/19/12
ES082	AY65113	WATER	07/18/12	07/19/12
ES085-TRIP BLANK	AY65114	WATER	07/18/12	07/19/12

Samples and blanks were screened for J-value responses between the detection limit (DL) and limit of quantitation (LOQ).

APPL's Laboratory control limits generated in house statistically do not meet the control limits listed in DoD QSM 4.2 for all analytes. Laboratory control spike recoveries for this project meet all control limits listed in the DoD QSM 4.2 except where noted. A copy of our in house generated control limits is available upon request. In addition, a copy of our LOQ control limits, established using 7 data points, are also available upon request.

Only the portion of the injection log relative to these samples is included. A full sequence log is available upon request.

Measurement uncertainty can be reported upon request.

CASE NARRATIVE

EPA Method 8270D SIM

Polynuclear Aromatic Hydrocarbons

Sample Preparation:

The water samples were extracted according to EPA method 3510C. All holding times were met.

Sample Analysis Information:

The samples were analyzed according to EPA method 8270D using a Hewlett Packard Gas Chromatograph with a mass spectrometer detector in selective ion monitoring mode.

Quality Control/Assurance

Calibrations:

Initial and continuing calibrations were performed according to the method. All calibration criteria were met.

Blanks:

No target analyte was detected above one-half the limit of quantitation (LOQ) in the method blank.

Spikes:

A Laboratory Control Spike (LCS) was used for quality assurance. All spike criteria were met.

No sample was designated by the client for MS/MSD analysis.

Surrogates

Surrogate recoveries are summarized on the forms 2&8. All surrogate recoveries were within the control limits.

Tuning:

The instrument was tuned using DFTPP. All method criteria were met.

Internal Standards

The internal standard area counts were compared to the mid-point of the initial calibration according to method 8270. All method criteria were met.

Summary:

No problem was encountered.

EPA Method 8015B

Total Petroleum Hydrocarbons – Diesel

Sample Preparation:

The water samples were extracted according to EPA method 3510C. The samples were extracted within holding time.

Sample Analysis Information:

The samples were analyzed according to the method using an Agilent Gas Chromatograph with a flame ionization detector.

Quality Control/Assurance

Calibrations:

Initial and continuing calibrations were performed according to the method. All calibration criteria were met.

Blanks:

No target analyte was detected above one-half the limit of quantitation (LOQ) in the method blank.

Spikes:

A Laboratory Control Spike (LCS) was used for quality assurance All acceptance criteria were met.

No sample was designated by the client for MS/MSD analysis.

Surrogates:

The surrogate recoveries are summarized on the form 2 & 8. . For the method blank, Ortho-Terphenyl recovered below the 57% lower control limit at 48.6%. The Octacosane surrogate was acceptable. All other surrogate recoveries were within the control limits.

Summary:

No other problem was encountered

EPA Method 8260B

Volatile Organic Analysis

Sample Preparation:

The water samples were purged according to EPA method 5030B. All holding times were met.

Sample Analysis Information:

The samples were analyzed according to the method using a Hewlett Packard Gas Chromatograph with a mass spectrometer detector. All holding times were met. The samples were received in unpreserved vials. The vials were analyzed within seven days of collection. Manual integrations were performed in accordance with APPL's SOP. All injections for gasoline were manually integrated due to the original integration not following the baseline. A summary of the manual integrations on the samples, blank and LCS is included in the QC Summary section of the report. Chromatograms from before and after the manual integrations are enclosed.

Quality Control/Assurance:

Calibrations:

Initial and continuing calibrations were performed according to the method. All system performance check compounds and calibration check compounds met DoD acceptance criteria. All calibration criteria were met except for CCV 0719T34.D, which recovered gasoline above the 120% upper control limit at 132%.

Blanks:

No target analyte was detected above one-half the limit of quantitation (LOQ) in the method blanks.

Spikes:

Lab control spikes (LCS) were used for quality assurance. A second source standard was used for the LCS. Gasoline recovered above the 125% upper control limit at 130%. All other LCS acceptance criteria were met.

No sample was designated by the client for MS/MSD analysis.

Surrogates:

Surrogate recoveries are summarized on Form 2 & 8. All surrogate recoveries were within the acceptance limits.

Tuning:

The instrument was tuned using BFB. All method criteria were met.

Internal Standards:

The internal standard area counts were compared to the mid-point of the initial calibration according to method 8260. All method criteria were met.

Summary:

The gasoline recoveries in the SS, CCV, and LCS were above their respective upper recovery limits because the initial calibration curve was made without the injection of surrogate. The samples could not be re-injected within holding time. The samples were re-injected outside of holding time with an initial calibration curve that contained surrogate and with acceptable SS, CCV, and LCS recoveries. Gasoline was not detected in the initial injections nor in the re-injections. No other problem was encountered. The data generated are acceptable.

EPA Method 6020

Dissolved Lead

Digestion Information:

The water samples were digested according to EPA method 3015. All holding times were met.

Analysis Information:

Samples:

The samples were analyzed for dissolved lead according to EPA method 6020 using an Agilent 7500CX ICP-MS.

Calibrations:

The initial and continuing calibrations were analyzed according to the DoD QSM. The initial calibration verification is prepared from a second source standard.

Blanks:

No metal was detected at or above one-half the LOQ in the method blank.

Spikes:

Laboratory Control Spike (LCS), Matrix Spikes (MS/MSD), Post-Digestion Spike (PDS), and Dilution Test (DT) were used for quality assurance. All LCS recoveries were within the acceptance limits.

Sample ES084 (APPL SDG 68268) was selected by the laboratory for QC analysis. The results are reported in APPL report #68268.

Summary:

No analytical exception is noted. The data generated are acceptable.

Abbreviations and Flags

FLAG	DESCRIPTION
#	Recovery or RPD outside control limits
*	Recovery or RPD outside control limits
B	Analyte detected in associated method blank
C1	Reason for correction: wrote incorrect response
C2	Reason for correction: calculated incorrectly
C3	Reason for correction: needs to be rechecked
C4	Reason for correction: data not usable
DO	Diluted out
E	Exceeds linear range
F	Estimated value
G1	Includes a wide range of hydrocarbons which does not match our gasoline standard
G10	Includes a match to hydrocarbon profiles within the range of mineral spirits
G11	Includes a match to hydrocarbon profiles within the range of JP-4
G12	Pattern does not match the gasoline standard; the carbon range for this sample is consistent with JP8
G13	Closely resembles the hydrocarbon profile of aviation gasoline
G14	Analyte concentration may be biased due to carry over
G2	Closely resembles the boiling point hydrocarbon profile consistent with weathered gasoline
G3	Includes higher boiling hydrocarbons
G4	Includes dominant peak(s) not indicative of petroleum hydrocarbons
G5	Is mainly dominant peak(s) not indicative of petroleum hydrocarbons
G6	Contains recognizable contaminant peak(s) which has been removed from quantitation
G7	Is mainly a match to hydrocarbons within the range of gasoline
G8	Closely resembles the boiling point hydrocarbon profile consistent with weathered gasoline
G9	Includes hydrocarbons within the range of kerosene
J	Estimated value
M	Matrix effect
MI1	Manual integration: integration does not follow baseline
MI2	Manual integration: non-target peak interference
MI3	Manual integration: to split a peak that was integrated as one peak by the computer
MI4	Manual integration: to integrate a split peak
MI5	Manual integration: the whole peak or part of the peak was not integrated
MI6	Manual integration: computer integrated wrong peak
MI7	Manual integration: other - explain
MDL	Method detection limit
ND	Not detected
NT	Non-target
Q	Acceptance criteria not met
T1 I	Includes wide range of hydrocarbons not indicative of diesel
T1 M	Is mainly wide range of hydrocarbons not necessarily indicative of diesel
T2 I	Includes lower boiling hydrocarbons, i.e. mineral spirits, kerosene, stoddard solvent, white gas
T2 M	Is mainly lower boiling hydrocarbons, i.e. mineral spirits, kerosene, stoddard solvent, white gas
T3 I	Includes higher boiling hydrocarbons, i.e. asphaltene, waster oil, motor oil, or weathered diesel fuel
T3 M	Is mainly higher boiling hydrocarbons, i.e. asphaltene, waster oil, motor oil, or weathered diesel fuel
T4 I	Includes dominant peak(s) not indicative of hydrocarbons
T4 M	Is mainly dominant peak(s) not indicative of hydrocarbons
T5	Contains recognizable contaminant peak(s) which has been removed from quantitation
T6	Is mainly a match to hydrocarbons within range of diesel fuel
T7	Closely resembles the boiling point hydrocarbon profile consistent with diesel fuel
T8	Includes a match to hydrocarbon profiles within range of diesel and kerosene fuel
T9 I	Includes non-diesel hydrocarbons within boiling point range of diesel fuel
T9 M	Is mainly non-diesel hydrocarbons within boiling point range of diesel fuel.
Y	Percent difference between primary and confirmation column > 40%

**CHAIN OF CUSTODY,
ARF, CRF, AND
CLIENT COMMUNICATION**

APPL - Analysis Request Form

68258

Client: Environet, Inc.
 Address: 650 Iwilei Rd, #204
Honolulu, HI 96817
 Attn: Max Solmssen
 Phone: 808-833-2225 Fax: 808-833-2231
 Job: LTM Red Hill / 1022-024
 PO #: 1022-024
 Chain of Custody (Y/N): Y # 36496
 RAD Screen (Y/N): Y pH (Y/N): Y
 Turn Around Type: 2 WEEKS

Received by: TBV 
 Date Received: 07/19/12 Time: 11:50
 Delivered by: FED EX
 Shuttle Custody Seals (Y/N): N Time Zone: -10
 Chest Temp(s): 4.0°C
 Color: VOA,J-PURBLK,R-ORYEL
 Samples Chilled until Placed in Refrig/Freezer: Y
 Project Manager: Cynthia Clark
 QC Report Type: DVP4/ADRDOD/HI
 Due Date: 08/02/12

Comments:

14 day TAT for Form 1s & 21 day TAT for full package;
 prelims to OSDas@, MSolmssen@ & VDupra@environetinc.com
 1 pdf on CD or FTP (no hard copy), possible hard copy to LDC
 Guidance: DOD QSM, DoD Forms, J flag to DL, U flag at LOD
 EDD ADR A1/A3 (ADR 8.3a unchecked) to OSDas@ VDupra@ & MSolmssen@environetinc.com
 metals 6020: report Lead with 0.5ug/L RL
 TPH-Diesel only; VOCs: include gasoline by 8260B
 No analysis was requested on COCs

<p><u>Sample Distribution:</u> GC: 2-SSIMHC12W, 2-STPETD2 Extractions: 2- SEP004S, 2- SEP011 VOA: 3-\$86RHBF Metals: 2-\$602D(Pb) Other: 2- M3015</p>	<p><u>Charges:</u></p>	<p><u>Invoice To:</u> same</p>
--	------------------------	--

Client ID	APPL ID	Sampled	Analyses Requested
1. ES081	AY65112W 	07/18/12 09:30	\$602D(Pb), \$86RHBF, \$SIMHC12W, \$TPETD2 -- Unpreserved VOA
2. ES082	AY65113W 	07/18/12 11:00	\$602D(Pb), \$86RHBF, \$SIMHC12W, \$TPETD2 -- Unpreserved VOA
3. ES085-TRIP BLANK	AY65114W 	07/18/12 07:00	\$86RHBF -- Unpreserved VOA

APPL Sample Receipt Form

ARF# 68258

Sample	Container Type	Count	pH
AY65112	⁶ PL 500mL - HNO3	1	1.7
	¹⁵ VOAs - NP	3	NA
	¹⁷ Amber Liter	4	NA
AY65113	⁶ PL 500mL - HNO3	1	1.7
	¹⁵ VOAs - NP	3	NA
	¹⁷ Amber Liter	4	NA
AY65114	¹⁵ VOAs - NP	3	NA

Sample Container Type Count pH



APPL, Inc.
908 N Temperance Ave
Clovis, CA 93611

Phone: (559) 275-2175
Fax: (559) 275-4422

CHAIN OF CUSTODY RECORD

C.O.C. 36496

68258

4.0°C

Report to: PLEASE PRINT	Invoice to: <u>A.P.</u> PLEASE PRINT
Company Name: <u>Environet, Inc.</u> Phone: <u>808-833-2225</u>	Company Name: <u>Environet, Inc.</u> Phone: <u>808-833-2225</u>
Address: <u>650 Luilei Road, Suite 209</u> <u>Honolulu, HI 96817</u> Fax: <u>808-833-2231</u>	Address: <u>650 Luilei Road, Suite 209</u> <u>Honolulu, HI 96817</u> Fax: <u>808-833-2231</u>
Attn: <u>Max Solmssen / msolmssen@environetinc.com</u>	Attn: <u>A.P.</u>

Project Name/Number	Sampler (Print)	Analysis Requested/Method Number										Date Shipped: 7/18/12	
		Purchase Order Number	Sampler (Signature)	No. of Containers	Matrix			TPH-G-PO (8260 B)	VOCs (8260 B)	TPH-DPO (8015 B)	PAHs (8270 c SIM)		Lead* (6020)
Aq	Sed.				Soil	Waybill No.: 876412433254							
Sample Identification	Location	Date Collected	Time Collected	Time Zone									Comments: * Lead
ES081	Red Hill	7/18/12	9:30	HE	8	X							Sample was field-filtered.
ES082	↓	↓	11:00	↓	8	↓							
ES085 - trip blank	N/A	↓	7:00	↓	3	↓							

Shuttle Temperature:	<input checked="" type="checkbox"/> Standard 2-3wk <input type="checkbox"/> One week <input type="checkbox"/> 24/48 Hrs. <input type="checkbox"/> Other			Turnaround Requested: Check one <input checked="" type="checkbox"/> Standard 2-3wk <input type="checkbox"/> One week <input type="checkbox"/> 24/48 Hrs. <input type="checkbox"/> Other				Sample Disposal: <input type="checkbox"/> Return to client <input checked="" type="checkbox"/> Disposal by Lab (30-day retention)			
Relinquished by sampler: <u>MS</u>	Date: <u>7/18/12</u>	Time: <u>2:00 pm</u>	Received by:		Relinquished by:	Date:	Time:	Received by:			
Relinquished by:	Date:	Time:	Received by:		Relinquished by:	Date: <u>7/19/12</u>	Time: <u>11:50</u>	Received at lab by: <u>Will Pleu</u>			

COOLER RECEIPT FORM

- 1) Project: Red Hill/1022-024 Date Received: 7/14/12
2) Coolers: Number of Coolers: 1
3) YES NO Were coolers and samples screened for radioactivity?
4) YES NO Were custody seals on outside of cooler? How many? Date on seal?
5) Name on seal?
6) YES NO NA Were custody seals unbroken and intact at the time of arrival?
7) YES NO Did the cooler come with a shipping slip (air bill, etc.)? Carrier name: FedEx
8) Shipping slip numbers: 1) 876412433254 2) 3)
9) YES NO NA Was the shipping slip scanned into the database?
10) YES NO NA If cooler belongs to APPL, has it been logged into the ice chest database?
11) Describe type of packing in cooler (bubble wrap, popcorn, type of ice, etc.): Bubble wrap, wet ice

- 12) YES NO NA For hand delivered samples was sufficient ice present to start the cooling process?
13) YES NO Was a temperature blank included in the cooler?
14) Serial number of certified NIST thermometer used: A39267 Correction factor: 0
15) Cooler temp(s): 1) 4.0°C 2) 3) 4) 5) 6) 7) 8)

Chain of custody:

- 16) YES NO Was a chain of custody received?
17) YES NO Were the custody papers signed in the appropriate places?
18) YES NO Was the project identifiable from custody papers?
19) YES NO Did the chain of custody include date and time of sampling?
20) YES NO Is location where sample was taken listed on the chain of custody?

Sample Labels:

- 21) YES NO Were container labels in good condition?
22) YES NO Was the client ID on the label?
23) YES NO Was the date of sampling on the label?
24) YES NO Was the time of sampling on the label?
25) YES NO Did all container labels agree with custody papers?

Sample Containers:

- 26) YES NO Were all containers sealed in separate bags?
27) YES NO Did all containers arrive unbroken?
28) YES NO Was there any leakage from samples?
29) YES NO Were any of the lids cracked or broken?
30) YES NO Were correct containers used for the tests indicated?
31) YES NO Was a sufficient amount of sample sent for tests indicated?
32) YES NO NA Were bubbles present in volatile samples? If yes, the following were received with air bubbles:
Larger than a pea:
Smaller than a pea:

Preservation & Hold time:

- 33) YES NO NA Was a sufficient amount of holding time remaining to analyze the samples?
34) YES NO NA Do the sample containers contain the same preservative as what is stated on the COC?
35) YES NO NA Was the pH taken of all non-VOA preserved samples and written on the sample container?
36) YES NO NA Was the pH of acid preserved non-VOA samples < 2 & sodium hydroxide preserved samples > 12?
37) YES NO NA Unpreserved VOA Vials received?
38) YES NO NA Are unpreserved VOA vials noted in the ADD TEST FIELD on the ARF?

EW
7/14/12
7/14/12

Lab notified if pH was not adequate:
Deficiencies: No analysis marked on COC

Signature of personnel receiving samples: Will Shuch Second reviewer:
Signature of project manager notified: Renee Date and Time of notification: 7-19-12
Name of client notified: Date and Time of notification:
Information given to client: by whom (Initials):

EPA METHOD 8270
Polynuclear Aromatic Hydrocarbons



EPA METHOD 8270
Polynuclear Aromatic Hydrocarbons
QC Summary

APPL, INC.

Method Blank EPA 8270D SIM

Blank Name/QCG: **120723W-65144 - 169459**
Batch ID: #SIMHC-120723A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	1-METHYLNAPHTHALENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	2-METHYLNAPHTHALENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	ACENAPHTHENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	ACENAPHTHYLENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	ANTHRACENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
BLANK	BENZO(A)ANTHRACENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
BLANK	BENZO(A)PYRENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	BENZO(B)FLUORANTHENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	BENZO(GHI)PERYLENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
BLANK	BENZO(K)FLUORANTHENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
BLANK	CHRYSENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
BLANK	DIBENZ(A,H)ANTHRACENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
BLANK	FLUORANTHENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
BLANK	FLUORENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	INDENO(1,2,3-CD)PYRENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
BLANK	NAPHTHALENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
BLANK	PHENANTHRENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
BLANK	PYRENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
BLANK	SURROGATE: 2-FLUORBIPHENY	56.8	50-110			%	07/23/12	07/24/12
BLANK	SURROGATE: NITROBENZENE-	51.8	40-110			%	07/23/12	07/24/12
BLANK	SURROGATE: TERPHENYL-D14 (59.6	50-135			%	07/23/12	07/24/12

Quant Method: SIMB.M
Run #: 0724L003
Instrument: Linus
Sequence: L120613
Initials: LF

Printed: 07/27/12 5:34:34 PM
GC SC-Blank-REG MDLs

Surrogate Recovery

Lab Name: APPL, Inc.
 Case No: 68258
 Matrix: WATER

SDG No: 68258
 Date Analyzed: 07/24/12
 Instrument: Linus

APPL ID.	Client Sample No.	SURROGATE: 2-FLUORBIPHENYL			SURROGATE: NITROBENZENE-D5		
		Limits	Result	Qualifier	Limits	Result	Qualifier
120723A-BLK	Blank	50-110	56.8		40-110	51.8	
120723A-LCS	Lab Control Spike	50-110	63.0		40-110	74.5	
AY65112	ES081	50-110	62.0		40-110	63.7	
AY65113	ES082	50-110	55.2		40-110	59.0	

Comments: Batch: #SIMHC-120723A

Surrogate Recovery

Lab Name: APPL, Inc.
Case No: 68258
Matrix: WATER

SDG No: 68258
Date Analyzed: 07/24/12
Instrument: Linus

APPL ID.	Client Sample No.	SURROGATE: TERPHENYL-D14 (S)			Limits	Result	Qualifier
		Limits	Result	Qualifier			
120723A-BLK	Blank	50-135	59.6				
120723A-LCS	Lab Control Spike	50-135	58.0				
AY65112	ES081	50-135	56.9				
AY65113	ES082	50-135	59.6				

Comments: Batch: #SIMHC-120723A

Laboratory Control Spike Recovery

EPA 8270D SIM

APPL ID: 120723W-65144 LCS - 169459
 Batch ID: #SIMHC-120723A

APPL Inc.
 908 North Temperance Avenue
 Clovis, CA 93611

Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits
1-METHYLNAPHTHALENE	4.00	2.22	55.5	45-105
2-METHYLNAPHTHALENE	4.00	2.18	54.5	45-105
ACENAPHTHENE	4.00	2.18	54.5	45-110
ACENAPHTHYLENE	4.00	2.33	58.3	50-105
ANTHRACENE	4.00	2.39	59.8	55-110
BENZO(A)ANTHRACENE	4.00	2.21	55.3	55-110
BENZO(A)PYRENE	4.00	2.24	56.0	55-110
BENZO(B)FLUORANTHENE	4.00	2.23	55.8	45-120
BENZO(GHI)PERYLENE	4.00	2.46	61.5	40-125
BENZO(K)FLUORANTHENE	4.00	2.57	64.3	45-125
CHRYSENE	4.00	2.67	66.8	55-110
DIBENZ(A,H)ANTHRACENE	4.00	2.29	57.3	40-125
FLUORANTHENE	4.00	3.03	75.8	55-115
FLUORENE	4.00	2.75	68.8	50-110
INDENO(1,2,3-CD)PYRENE	4.00	2.25	56.3	45-125
NAPHTHALENE	4.00	2.15	53.8	40-100
PHENANTHRENE	4.00	2.78	69.5	50-115
PYRENE	4.00	2.45	61.3	50-130

SURROGATE: 2-FLUORBIPHENYL (S)	2.00	1.26	63.0	50-110
SURROGATE: NITROBENZENE-D5 (S)	2.00	1.49	74.5	40-110
SURROGATE: TERPHENYL-D14 (S)	2.00	1.16	58.0	50-135

Comments: _____

<u>Primary</u>	<u>SPK</u>
Quant Method :	SIMB.M
Extraction Date :	07/23/12
Analysis Date :	07/24/12
Instrument :	Linus
Run :	0724L004
Initials :	LF

Printed: 07/27/12 5:34:43 PM
 APPL Standard LCS

8270D-SIM

Form 4

Blank Summary

Lab Name: APPL, Inc.

SDG No: 68258

Case No: 68258

Date Analyzed: 07/24/12

Matrix: WATER

Instrument: Linus

Blank ID: 120723A-BLK

Time Analyzed: 1850

APPL ID.	Client Sample No.	File ID.	Date Analyzed
120723A-BLK	Blank	0724L003	07/24/12 1850
120723A-LCS	Lab Control Spike	0724L004	07/24/12 1916
AY65112	ES081	0724L008	07/24/12 2100
AY65113	ES082	0724L009	07/24/12 2126

Comments: Batch: #SIMHC-120723A

Printed: 07/27/12 5:34:45 PM
Form 4, Blank Summary

Form 5
Tune Summary

Lab Name: APPL Inc.
 Case No: 68258
 Matrix: Water
 ID: SVTUNE 2-28-12

SDG No: 68258
 Date Analyzed: 07/24/12
 Instrument: Linus
 Time Analyzed: 18:05

	Client Sample No.	APPL ID.	File ID.	Date Analyzed
1	Blank	120723A BLK 1/1000	0724L003.D	07/24/12 18:50
2	Lab Control Spike	120723A LCS-1 1/1000	0724L004.D	07/24/12 19:16
3	ES081	AY65112W07 1/1030	0724L008.D	07/24/12 21:00
4	ES082	AY65113W06 1/1050	0724L009.D	07/24/12 21:26
5				
6				
7				
8				
9				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				

m/e

51 29.95 - 60% of mass 198	<u>56.9</u>
68 0 - 2.05% of mass 69	<u>0.1</u>
70 0 - 2% of mass 69	<u>0.8</u>
127 40 - 60% of mass 198	<u>54.7</u>
197 0 - 1% of mass 198	<u>0.0</u>
198 100 - 100% of mass 198	<u>100.0</u>
199 5 - 9% of mass 198	<u>7.1</u>
275 10 - 30% of mass 198	<u>23.7</u>
365 1 - 100% of mass 198	<u>3.1</u>
441 0.01 - 100% of mass 443	<u>76.8</u>
442 40 - 150% of mass 198	<u>72.0</u>
443 17 - 23% of mass 442	<u>20.2</u>

8A
INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: APPL Inc. Contract: Review
 Lab Code: _____ SDG No.: 68258
 Lab File ID (Standard): 0613L007.D Date Analyzed: 06/13/12
 Instrument ID: Linus Time Analyzed: 15:33
 GC Column: _____ ID: Heated Purge: (Y/N) _____

		Naphthalene-D8(IS)		Acenaphthene-D10(IS)		Phenanthrene-D10(IS)	
		AREA #	RT #	AREA #	RT #	AREA #	RT #
12 HOUR STD		2713	6.09	1189	8.10	2090	9.82
UPPER LIMIT		5426	6.59	2378	8.60	4180	10.32
LOWER LIMIT		1357	5.59	595	7.60	1045	9.32
SAMPLE							
NO.							
01	120723A BLK 1/1000	2273	6.07	1022	8.08	2049	9.82
02	120723A LCS-1 1/1000	2043	6.07	992	8.08	1998	9.82
03	AY65112W07 1/1030	2379	6.08	1160	8.08	2250	9.82
04	AY65113W06 1/1050	2502	6.07	1438	8.07	2203	9.81
05							
06							
07							
08							
09							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

AREA UPPER LIMIT = +100% of internal standard area.
 AREA LOWER LIMIT = -50% of internal standard area.
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

8A
INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: APPL Inc. Contract: Review
 Lab Code: _____ SDG No.: 68258
 Lab File ID (Standard): 0613L007.D Date Analyzed: 06/13/12
 Instrument ID: Linus Time Analyzed: 15:33
 GC Column: _____ ID: _____ Heated Purge: (Y/N) _____

		Chrysene-D12(IS)		Perylene-D12(IS)			
		AREA #	RT #	AREA #	RT #	AREA #	RT #
	12 HOUR STD	2430	12.91	2133	14.52		
	UPPER LIMIT	4860	13.41	4266	15.02		
	LOWER LIMIT	1215	12.41	1067	14.02		
	SAMPLE NO.						
01	120723A BLK 1/1000	2655	12.91	2331	14.54		
02	120723A LCS-1 1/1000	2829	12.90	2395	14.52		
03	AY65112W07 1/1030	2823	12.91	2347	14.52		
04	AY65113W06 1/1050	2920	12.90	2435	14.52		
05							
06							
07							
08							
09							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

AREA UPPER LIMIT = +100% of internal standard area.
 AREA LOWER LIMIT = -50% of internal standard area.
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

EPA METHOD 8270
Polynuclear Aromatic Hydrocarbons
Sample Data

EPA 8270D SIM

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

Attn: Max Solmssen
Project: LTM Red Hill / 1022-024

Sample ID: ES081

Sample Collection Date: 07/18/12

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 68258

APPL ID: AY65112

QCG: #SIMHC-120723A-169459

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
8270D-SIM	1-METHYLNAPHTHALENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	2-METHYLNAPHTHALENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	ACENAPHTHENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	ACENAPHTHYLENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	ANTHRACENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
8270D-SIM	BENZO(A)ANTHRACENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
8270D-SIM	BENZO(A)PYRENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	BENZO(B)FLUORANTHENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	BENZO(GHI)PERYLENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
8270D-SIM	BENZO(K)FLUORANTHENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
8270D-SIM	CHRYSENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
8270D-SIM	DIBENZ(A,H)ANTHRACENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
8270D-SIM	FLUORANTHENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
8270D-SIM	FLUORENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	INDENO(1,2,3-CD)PYRENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
8270D-SIM	NAPHTHALENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
8270D-SIM	PHENANTHRENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
8270D-SIM	PYRENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
8270D-SIM	SURROGATE: 2-FLUORBIPHENYL (S)	62.0	50-110			%	07/23/12	07/24/12
8270D-SIM	SURROGATE: NITROBENZENE-D5 (S)	63.7	40-110			%	07/23/12	07/24/12
8270D-SIM	SURROGATE: TERPHENYL-D14 (S)	56.9	50-135			%	07/23/12	07/24/12

Quant Method: SIMB.M
Run #: 0724L008
Instrument: Linus
Sequence: L120613
Dilution Factor: 1
Initials: LF

Printed: 07/27/12 5:34:48 PM
APPL-F1-SC-NoMC-REG MDLs

Data File : M:\LINUS\DATA\L120613\0724L008.D Vial: 8
 Acq On : 24 Jul 12 21:00 Operator: LF
 Sample : AY65112W07 1/1030 Inst : Linus
 Misc : Multiplr: 0.97

Quant Time: Jul 27 7:53 2012 Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jul 25 18:38:43 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8(IS)	6.08	136	2379	2.50000	ppb	-0.04
6) Acenaphthene-D10(IS)	8.08	164	1160	2.50000	ppb	-0.05
12) Phenanthrene-D10(IS)	9.82	188	2250	2.50000	ppb	-0.04
16) Chrysene-D12(IS)	12.91	240	2823	2.50000	ppb	0.01
22) Perylene-D12(IS)	14.52	264	2347	2.50000	ppb	-0.01
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.32	82	567	1.23739	ppb	-0.01
Spiked Amount	1.942		Recovery	=	63.706%	
7) Surrogate Recovery (FBP)	7.32	172	1346	1.20336	ppb	-0.05
Spiked Amount	1.942		Recovery	=	61.955%	
18) Surrogate Recovery (TPH)	11.69	244	1607	1.10473	ppb	-0.05
Spiked Amount	1.942		Recovery	=	56.908%	

Target Compounds Qvalue

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0724L008.D
Acq On : 24 Jul 12 21:00
Sample : AY65112W07 1/1030
Misc :

Vial: 8
Operator: LF
Inst : Linus
Multiplr: 0.97

Quant Time: Jul 27 7:53 2012

Quant Results File: SIMB.RES

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jul 25 18:38:43 2012
Response via : Initial Calibration



EPA 8270D SIM

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

Attn: Max Solmssen
Project: LTM Red Hill / 1022-024

Sample ID: ES082
Sample Collection Date: 07/18/12

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 68258
APPL ID: AY65113
QCG: #SIMHC-120723A-169459

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
8270D-SIM	1-METHYLNAPHTHALENE	4.7	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	2-METHYLNAPHTHALENE	0.88	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	ACENAPHTHENE	0.23	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	ACENAPHTHYLENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	ANTHRACENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
8270D-SIM	BENZO(A)ANTHRACENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
8270D-SIM	BENZO(A)PYRENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	BENZO(B)FLUORANTHENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	BENZO(GHI)PERYLENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
8270D-SIM	BENZO(K)FLUORANTHENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
8270D-SIM	CHRYSENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
8270D-SIM	DIBENZ(A,H)ANTHRACENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
8270D-SIM	FLUORANTHENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
8270D-SIM	FLUORENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
8270D-SIM	INDENO(1,2,3-CD)PYRENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
8270D-SIM	NAPHTHALENE	17	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
8270D-SIM	PHENANTHRENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
8270D-SIM	PYRENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
8270D-SIM	SURROGATE: 2-FLUORBIPHENYL (S)	55.2	50-110			%	07/23/12	07/24/12
8270D-SIM	SURROGATE: NITROBENZENE-D5 (S)	59.0	40-110			%	07/23/12	07/24/12
8270D-SIM	SURROGATE: TERPHENYL-D14 (S)	59.6	50-135			%	07/23/12	07/24/12

Quant Method: SIMB.M
Run #: 0724L009
Instrument: Linus
Sequence: L120613
Dilution Factor: 1
Initials: LF

Printed: 07/27/12 5:34:48 PM
APPL-F1-SC-NoMC-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L120613\0724L009.D Vial: 9
 Acq On : 24 Jul 12 21:26 Operator: LF
 Sample : AY65113W06 1/1050 Inst : Linus
 Misc : Multiplr: 0.95

Quant Time: Jul 27 7:53 2012 Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jul 25 18:38:43 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8 (IS)	6.07	136	2502	2.50000	ppb	-0.05
6) Acenaphthene-D10 (IS)	8.07	164	1438	2.50000	ppb	-0.06
12) Phenanthrene-D10 (IS)	9.81	188	2203	2.50000	ppb	-0.05
16) Chrysene-D12 (IS)	12.90	240	2920	2.50000	ppb	0.00
22) Perylene-D12 (IS)	14.52	264	2435	2.50000	ppb	-0.01
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.31	82	552	1.12362	ppb	-0.02
Spiked Amount	1.905		Recovery	=	59.010%	
7) Surrogate Recovery (FBP)	7.31	172	1485	1.05057	ppb	-0.06
Spiked Amount	1.905		Recovery	=	55.178%	
18) Surrogate Recovery (TPH)	11.69	244	1741	1.13505	ppb	-0.05
Spiked Amount	1.905		Recovery	=	59.588%	
Target Compounds						
3) Napthalene	6.09	128	29195	17.25489	ppb	97
4) 2-Methylnaphthalene	6.88	142	963	0.87824	ppb	97
5) 1-Methylnaphthalene	6.99	142	5221	4.73282	ppb	89
10) Acenaphthene	8.11	154	258	0.22536	ppb	83

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0724L009.D
Acq On : 24 Jul 12 21:26
Sample : AY65113W06 1/1050
Misc :


Vial: 9
Operator: LF
Inst : Linus
Multiplr: 0.95

Quant Time: Jul 27 7:53 2012

Quant Results File: SIMB.RES





Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jul 25 18:38:43 2012
Response via : Initial Calibration








#3
 Naphthalene
 Concen: 17.25489 ppb
 RT: 6.09 min Scan# 325
 Delta R.T. -0.05 min
 Lab File: 0724L009.D
 Acq: 24 Jul 12 21:26


Tgt Ion	Resp	Lower	Upper
128	100		
129	12.5	7.8	14.4
127	12.9	8.6	16.0



#4
 2-Methylnaphthalene
 Concen: 0.87824 ppb
 RT: 6.88 min Scan# 390
 Delta R.T. -0.05 min
 Lab File: 0724L009.D
 Acq: 24 Jul 12 21:26





Tgt Ion	Resp	Lower	Upper
142	100		
141	87.0	62.7	116.5








#5
 1-Methylnaphthalene
 Concen: 4.73282 ppb
 RT: 6.99 min Scan# 399
 Delta R.T. -0.06 min
 Lab File: 0724L009.D
 Acq: 24 Jul 12 21:26

Tgt Ion:142 Resp: 5221
 Ion Ratio Lower Upper
 142 100
 141 92.7 57.8 107.3



#10
 Acenaphthene
 Concen: 0.22536 ppb
 RT: 8.11 min Scan# 492
 Delta R.T. -0.06 min
 Lab File: 0724L009.D
 Acq: 24 Jul 12 21:26

Tgt Ion:154 Resp: 258
 Ion Ratio Lower Upper
 154 100
 153 116.1 68.8 127.8
 152 54.4 31.5 58.5



EPA METHOD 8270
Polynuclear Aromatic Hydrocarbons
Calibration Data

EPA 8270C SIM

Form 6
Initial Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: _____

SDG No: 66258
Initial Cal. Date: 06/13/12
Instrument: Linus

Initials: _____

0613L003.D 0613L004.D 0613L005.D 0613L006.D 0613L007.D 0613L008.D 0613L009.D 0613L010.D

	Compound	0.1	0.2	0.5	1	5	10	50	100			Avg	%RSD	
1	I	Napthalene-D8(IS)												
2	S	Surrogate Recovery (NBZ)	0.4582	0.4160	0.5318	0.4779	0.4460	0.4748	0.4769	0.4584		0.47	7.1	S
3	TM	Naphthalene	1.842	1.750	1.792	1.659	1.423	1.727	1.409	1.279		1.6	13	TM
4	TM	2-Methylnaphthalene	1.241	1.076	1.116	1.120	0.9307	1.112	0.9262	0.8257		1.0	13	TM
5	TM	1-Methylnaphthalene	1.126	1.172	1.203	1.088	0.8644	1.036	0.8585			1.0	13	TM
6	I	Acenaphthene-D10(IS)												
7	S	Surrogate Recovery (FBP)	2.582	2.805	2.664	2.529	2.150	2.143	1.969	1.882		2.3	15	S
8	TM	1,1'-Biphenyl	2.787	2.890	2.770	2.823	2.494	2.718	2.250	2.042		2.6	12	TM
9	TM	Acenaphthylene	3.955	4.033	3.713	3.520	3.060	3.526	2.830	2.701		3.4	15	TM
10	*TM	Acenaphthene	2.090	2.180	2.070	2.027	1.756	1.959	1.627	1.454		1.9	13	*TM
11	TM	Fluorene	2.398	2.371	2.439	2.352	2.050	2.300	1.873	1.659		2.2	13	TM
12	I	Phenanthrene-D10(IS)												
13	TM	Phenanthrene	2.047	1.950	2.033	1.897	1.652	1.874	1.503	1.377		1.8	14	TM
14	TM	Anthracene	2.130	1.841	1.997	1.890	1.692	1.793	1.496	1.348		1.8	14	TM
15	*TM	Fluoranthene	3.076	2.754	2.876	2.744	2.354	2.691	2.122	2.002		2.6	15	*TM
16	I	Chrysene-D12(IS)												
17	TM	Pyrene	2.479	2.491	2.445	2.361	2.151	2.307	1.879	1.969		2.3	10	TM
18	S	Surrogate Recovery (TPH)	1.440	1.456	1.389	1.283	1.203	1.197	0.9916	1.046		1.3	14	S
19	TM	Benz (a) anthracene	2.260	2.204	2.209	2.058	1.786	1.987	1.662	1.724		2.0	12	TM
20	TM	Chrysene	2.088	2.135	2.151	2.031	1.970	1.967	1.407	1.602		1.9	14	TM
21	TM	Indeno (1,2,3-cd) pyrene	2.365	2.214	2.159	2.037	1.899	2.069	1.653	1.810		2.0	11	TM
22	I	Perylene-D12(IS)												
23	TM	Benzo (b) fluoranthene	2.382	2.407	2.462	2.408	1.885	2.105	2.227	1.721		2.2	12	TM
24	TM	Benzo (k) fluoranthene	2.745	2.558	2.205	2.115	2.223	2.494	1.828	1.795		2.2	15	TM
25	*TM	Benzo (a) pyrene	2.358	2.547	2.297	2.164	1.908	2.189	1.901	1.547		2.1	15	*TM
26	TM	Dibenz (a,h) anthracene	2.206	2.196	2.054	1.889	1.755	1.968	1.762	1.529		1.9	12	TM
27	TM	Benzo (g,h,i) perylene	2.288	2.284	2.189	1.980	1.781	2.022	1.834	1.643		2.0	12	TM
28														
29														
30														
31														
32														
33														
34														
35														

Data File : M:\LINUS\DATA\L120613\0613L003.D Vial: 3
 Acq On : 13 Jun 12 13:51 Operator: LF
 Sample : 0.1ug/ml PAH 06-13-12 Inst : Linus
 Misc : Multiplr: 1.00

Quant Time: Jun 13 16:10 2012 Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jun 13 13:28:43 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8 (IS)	6.09	136	2619	2.50000	ppb	-0.02
6) Acenaphthene-D10 (IS)	8.10	164	1220	2.50000	ppb	-0.04
12) Phenanthrene-D10 (IS)	9.83	188	2113	2.50000	ppb	-0.02
16) Chrysene-D12 (IS)	12.91	240	2622	2.50000	ppb	-0.01
22) Perylene-D12 (IS)	14.52	264	2131	2.50000	ppb	-0.01
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.35	82	48	0.18668	ppb	0.01
Spiked Amount	2.000		Recovery	=	9.350%	
7) Surrogate Recovery (FBP)	7.34	172	126	0.16296	ppb	-0.04
Spiked Amount	2.000		Recovery	=	8.150%	
18) Surrogate Recovery (TPH)	11.70	244	151	0.18456	ppb	-0.03
Spiked Amount	2.000		Recovery	=	9.250%	
Target Compounds						
						Qvalue
3) Naphthalene	6.12	128	193	0.12913	ppb	97
4) 2-Methylnaphthalene	6.91	142	130	0.14464	ppb	90
5) 1-Methylnaphthalene	7.01	142	118	0.14074	ppb	84
8) 1,1'-Biphenyl	7.46	154	136	0.14114	ppb	# 86
9) Acenaphthylene	7.94	152	193	0.16464	ppb	99
10) Acenaphthene	8.13	154	102	0.14944	ppb	84
11) Fluorene	8.75	166	117	0.14146	ppb	95
13) Phenanthrene	9.86	178	173	0.13796	ppb	99
14) Anthracene	9.92	178	180	0.15900	ppb	94
15) Fluoranthene	11.24	202	260	0.16914	ppb	97
17) Pyrene	11.50	202	260	0.17208	ppb	95
19) Benz (a) anthracene	12.90	228	237	0.18310	ppb	98
20) Chrysene	12.94	228	219	0.16763	ppb	# 88
21) Indeno (1,2,3-cd) pyrene	16.02	276	248	0.09203	ppb	# 76
23) Benzo (b) fluoranthene	14.09	252	203	0.15062	ppb	# 83
24) Benzo (k) fluoranthene	14.11	252	234	0.20915	ppb	# 92
25) Benzo (a) pyrene	14.46	252	201	0.16795	ppb	# 93
26) Dibenz (a,h) anthracene	16.03	278	188	0.15446	ppb	# 76
27) Benzo (g,h,i) perylene	16.45	276	195	0.05934	ppb	90

(#) = qualifier out of range (m) = manual integration

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0613L003.D
Acq On : 13 Jun 12 13:51
Sample : 0.1ug/ml PAH 06-13-12
Misc :

Vial: 3
Operator: LF
Inst : Linus
Multiplr: 1.00

Quant Time: Jun 13 16:10 2012

Quant Results File: SIMB.RES

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jun 13 17:38:06 2012
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L120613\0613L004.D Vial: 4
 Acq On : 13 Jun 12 14:16 Operator: LF
 Sample : 0.2ug/ml PAH Inst : Linus
 Misc : Multiplr: 1.00

Quant Time: Jun 13 16:10 2012 Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jun 13 14:51:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8 (IS)	6.09	136	2614	2.50000	ppb	-0.02
6) Acenaphthene-D10 (IS)	8.10	164	1181	2.50000	ppb	-0.04
12) Phenanthrene-D10 (IS)	9.83	188	2179	2.50000	ppb	-0.02
16) Chrysene-D12 (IS)	12.91	240	2524	2.50000	ppb	-0.01
22) Perylene-D12 (IS)	14.52	264	2140	2.50000	ppb	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) Surrogate Recovery (NBZ)	5.33	82	87	0.19035	ppb	0.00
Spiked Amount 2.000			Recovery =	9.500%		
7) Surrogate Recovery (FBP)	7.34	172	265	0.20827	ppb	-0.04
Spiked Amount 2.000			Recovery =	10.400%		
18) Surrogate Recovery (TPH)	11.70	244	294	0.20112	ppb	-0.03
Spiked Amount 2.000			Recovery =	10.050%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	6.12	128	366	0.19487	ppb	98
4) 2-Methylnaphthalene	6.91	142	225	0.18576	ppb	98
5) 1-Methylnaphthalene	7.01	142	245	0.20393	ppb	86
8) 1,1'-Biphenyl	7.45	154	273	0.20362	ppb	98
9) Acenaphthylene	7.94	152	381	0.20195	ppb	99
10) Acenaphthene	8.13	154	206	0.20422	ppb	91
11) Fluorene	8.75	166	224	0.19888	ppb	96
13) Phenanthrene	9.86	178	340	0.19518	ppb	97
14) Anthracene	9.92	178	321	0.18548	ppb	96
15) Fluoranthene	11.24	202	480	0.18893	ppb	# 97
17) Pyrene	11.50	202	503	0.20049	ppb	91
19) Benz (a) anthracene	12.90	228	445	0.19750	ppb	98
20) Chrysene	12.94	228	431	0.20220	ppb	# 92
21) Indeno (1,2,3-cd) pyrene	16.01	276	447	0.19341	ppb	# 87
23) Benzo (b) fluoranthene	14.09	252	412	0.20307	ppb	# 83
24) Benzo (k) fluoranthene	14.11	252	438	0.19501	ppb	# 92
25) Benzo (a) pyrene	14.46	252	436	0.20968	ppb	99
26) Dibenz (a,h) anthracene	16.03	278	376	0.20161	ppb	# 93
27) Benzo (g,h,i) perylene	16.44	276	391	0.17158	ppb	89

(#) = qualifier out of range (m) = manual integration
 0613L004.D SIMB.M Thu Jul 05 14:10:47 2012

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0613L004.D
Acq On : 13 Jun 12 14:16
Sample : 0.2ug/ml PAH
Misc :

Vial: 4
Operator: LF
Inst : Linus
Multiplr: 1.00

Quant Time: Jun 13 16:10 2012

Quant Results File: SIMB.RES

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jun 13 17:38:06 2012
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L120613\0613L005.D Vial: 5
 Acq On : 13 Jun 12 14:41 Operator: LF
 Sample : 0.5ug/ml PAH Inst : Linus
 Misc : Multiplr: 1.00

Quant Time: Jun 13 15:40 2012 Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jun 13 14:51:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Naphthalene-D8 (IS)	6.09	136	2576	2.50000	ppb	-0.02
6) Acenaphthene-D10 (IS)	8.10	164	1220	2.50000	ppb	-0.04
12) Phenanthrene-D10 (IS)	9.83	188	2083	2.50000	ppb	-0.02
16) Chrysene-D12 (IS)	12.91	240	2571	2.50000	ppb	-0.01
22) Perylene-D12 (IS)	14.52	264	2220	2.50000	ppb	-0.01
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.33	82	274	0.60835	ppb	0.00
Spiked Amount	2.000		Recovery	=	30.400%	
7) Surrogate Recovery (FBP)	7.34	172	650	0.49453	ppb	-0.04
Spiked Amount	2.000		Recovery	=	24.750%	
18) Surrogate Recovery (TPH)	11.70	244	714	0.47952	ppb	-0.03
Spiked Amount	2.000		Recovery	=	24.000%	
Target Compounds						
3) Naphthalene	6.12	128	923	0.49869	ppb	Qvalue 100
4) 2-Methylnaphthalene	6.90	142	575	0.48172	ppb	100
5) 1-Methylnaphthalene	7.01	142	620	0.52369	ppb	94
8) 1,1'-Biphenyl	7.44	154	676	0.48807	ppb	# 94
9) Acenaphthylene	7.94	152	906	0.46486	ppb	99
10) Acenaphthene	8.13	154	505	0.48464	ppb	91
11) Fluorene	8.74	166	595	0.51139	ppb	99
13) Phenanthrene	9.86	178	847	0.50863	ppb	98
14) Anthracene	9.92	178	832	0.50291	ppb	96
15) Fluoranthene	11.23	202	1198	0.49327	ppb	# 86
17) Pyrene	11.50	202	1257	0.49186	ppb	# 89
19) Benz (a) anthracene	12.90	228	1136	0.49495	ppb	98
20) Chrysene	12.94	228	1106	0.50938	ppb	# 93
21) Indeno (1,2,3-cd) pyrene	16.00	276	1110	0.47150	ppb	# 92
23) Benzo (b) fluoranthene	14.08	252	1093	0.51930	ppb	# 90
24) Benzo (k) fluoranthene	14.11	252	979	0.42017	ppb	# 95
25) Benzo (a) pyrene	14.45	252	1020	0.47286	ppb	# 95
26) Dibenz (a,h) anthracene	16.03	278	912	0.47138	ppb	# 95
27) Benzo (g,h,i) perylene	16.44	276	972	0.41115	ppb	93

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0613L005.D
Acq On : 13 Jun 12 14:41
Sample : 0.5ug/ml PAH
Misc :

Vial: 5
Operator: LF
Inst : Linus
Multiplr: 1.00

Quant Time: Jun 13 15:40 2012

Quant Results File: SIMB.RES

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jun 13 17:38:06 2012
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L120613\0613L006.D
 Acq On : 13 Jun 12 15:07
 Sample : 1.0ug/ml PAH
 Misc :

Vial: 6
 Operator: LF
 Inst : Linus
 Multiplr: 1.00

Quant Time: Jun 13 15:40 2012

Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jun 13 14:51:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8 (IS)	6.09	136	2621	2.50000	ppb	-0.02
6) Acenaphthene-D10 (IS)	8.10	164	1201	2.50000	ppb	-0.04
12) Phenanthrene-D10 (IS)	9.83	188	2124	2.50000	ppb	-0.02
16) Chrysene-D12 (IS)	12.91	240	2585	2.50000	ppb	-0.01
22) Perylene-D12 (IS)	14.52	264	2229	2.50000	ppb	-0.01

System Monitoring Compounds

2) Surrogate Recovery (NBZ)	5.33	82	501	1.01960	ppb	0.00
Spiked Amount	2.000		Recovery	=	51.000%	
7) Surrogate Recovery (FBP)	7.34	172	1215	0.94245	ppb	-0.04
Spiked Amount	2.000		Recovery	=	47.100%	
18) Surrogate Recovery (TPH)	11.70	244	1327	0.89865	ppb	-0.03
Spiked Amount	2.000		Recovery	=	44.950%	

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	6.12	128	1739	0.92424	ppb	99
4) 2-Methylnaphthalene	6.90	142	1174	0.97858	ppb	98
5) 1-Methylnaphthalene	7.01	142	1141	0.93248	ppb	94
8) 1,1'-Biphenyl	7.45	154	1356	1.00249	ppb	# 91
9) Acenaphthylene	7.94	152	1691	0.90251	ppb	99
10) Acenaphthene	8.13	154	974	0.95935	ppb	89
11) Fluorene	8.74	166	1130	0.97914	ppb	98
13) Phenanthrene	9.86	178	1612	0.94390	ppb	99
14) Anthracene	9.92	178	1606	0.95018	ppb	98
15) Fluoranthene	11.23	202	2331	0.94550	ppb	# 88
17) Pyrene	11.50	202	2441	0.95516	ppb	# 88
19) Benz (a) anthracene	12.90	228	2128	0.92526	ppb	97
20) Chrysene	12.94	228	2100	0.95596	ppb	# 94
21) Indeno (1,2,3-cd) pyrene	15.99	276	2106	0.90696	ppb	# 82
23) Benzo (b) fluoranthene	14.08	252	2147	1.00305	ppb	# 88
24) Benzo (k) fluoranthene	14.11	252	1886	0.85148	ppb	# 94
25) Benzo (a) pyrene	14.45	252	1929	0.90706	ppb	# 95
26) Dibenz (a,h) anthracene	16.03	278	1684	0.88374	ppb	97
27) Benzo (g,h,i) perylene	16.44	276	1765	0.79039	ppb	95

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0613L006.D
Acq On : 13 Jun 12 15:07
Sample : 1.0ug/ml PAH
Misc :

Vial: 6
Operator: LF
Inst : Linus
Multiplr: 1.00

Quant Time: Jun 13 15:40 2012

Quant Results File: SIMB.RES

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jun 13 17:38:06 2012
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L120613\0613L007.D
 Acq On : 13 Jun 12 15:33
 Sample : 5.0ug/ml PAH
 Misc :

Vial: 7
 Operator: LF
 Inst : Linus
 Multiplr: 1.00

Quant Time: Jun 13 16:08 2012

Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jun 13 14:51:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8(IS)	6.09	136	2713	2.50000	ppb	-0.02
6) Acenaphthene-D10(IS)	8.10	164	1189	2.50000	ppb	-0.04
12) Phenanthrene-D10(IS)	9.82	188	2090	2.50000	ppb	-0.04
16) Chrysene-D12(IS)	12.91	240	2430	2.50000	ppb	-0.01
22) Perylene-D12(IS)	14.52	264	2133	2.50000	ppb	-0.01

System Monitoring Compounds

2) Surrogate Recovery (NBZ)	5.32	82	2420	4.73481	ppb	-0.01
Spiked Amount	2.000		Recovery	=	236.750%	
7) Surrogate Recovery (FBP)	7.34	172	5112	4.06377	ppb	-0.04
Spiked Amount	2.000		Recovery	=	203.200%	
18) Surrogate Recovery (TPH)	11.70	244	5848	4.32241	ppb	-0.03
Spiked Amount	2.000		Recovery	=	216.100%	

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	6.12	128	7720	4.04041	ppb	100
4) 2-Methylnaphthalene	6.90	142	5050	4.08854	ppb	95
5) 1-Methylnaphthalene	7.01	142	4690	3.76651	ppb	93
8) 1,1'-Biphenyl	7.45	154	5931	4.42630	ppb	# 89
9) Acenaphthylene	7.93	152	7276	4.02049	ppb	97
10) Acenaphthene	8.13	154	4176	4.19734	ppb	93
11) Fluorene	8.74	166	4875	4.28917	ppb	98
13) Phenanthrene	9.86	178	6907	4.16861	ppb	99
14) Anthracene	9.92	178	7071	4.30520	ppb	98
15) Fluoranthene	11.23	202	9839	4.11183	ppb	95
17) Pyrene	11.49	202	10454	4.40089	ppb	# 90
19) Benz (a) anthracene	12.90	228	8681	4.09173	ppb	96
20) Chrysene	12.94	228	9575	4.68837	ppb	# 96
21) Indeno (1,2,3-cd) pyrene	15.99	276	9227	4.32779	ppb	# 88
23) Benzo (b) fluoranthene	14.08	252	8043	3.92370	ppb	# 84
24) Benzo (k) fluoranthene	14.12	252	9483	4.64656	ppb	# 92
25) Benzo (a) pyrene	14.45	252	8141	4.09554	ppb	98
26) Dibenz (a,h) anthracene	16.02	278	7487	4.22884	ppb	# 91
27) Benzo (g,h,i) perylene	16.43	276	7598	3.75225	ppb	96

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0613L007.D
Acq On : 13 Jun 12 15:33
Sample : 5.0ug/ml PAH
Misc :

Vial: 7
Operator: LF
Inst : Linus
Multiplr: 1.00

Quant Time: Jun 13 16:08 2012

Quant Results File: SIMB.RES

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jun 13 17:38:06 2012
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L120613\0613L008.D
 Acq On : 13 Jun 12 15:59
 Sample : 10ug/ml PAH
 Misc :

Vial: 8
 Operator: LF
 Inst : Linus
 Multiplr: 1.00

Quant Time: Jun 13 17:35 2012

Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jun 13 14:51:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8 (IS)	6.09	136	2467	2.50000	ppb	-0.02
6) Acenaphthene-D10 (IS)	8.10	164	1136	2.50000	ppb	-0.04
12) Phenanthrene-D10 (IS)	9.82	188	2001	2.50000	ppb	-0.04
16) Chrysene-D12 (IS)	12.90	240	2373	2.50000	ppb	-0.02
22) Perylene-D12 (IS)	14.52	264	2033	2.50000	ppb	-0.01

System Monitoring Compounds

2) Surrogate Recovery (NBZ)	5.31	82	4685	10.18847	ppb	-0.02
Spiked Amount	2.000					
Recovery				=	509.400%	
7) Surrogate Recovery (FBP)	7.34	172	9738	8.41759	ppb	-0.04
Spiked Amount	2.000					
Recovery				=	420.900%	
18) Surrogate Recovery (TPH)	11.70	244	11363	8.84002	ppb	-0.03
Spiked Amount	2.000					
Recovery				=	442.000%	

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	6.12	128	17040	10.19897	ppb	99
4) 2-Methylnaphthalene	6.90	142	10976	10.14218	ppb	94
5) 1-Methylnaphthalene	7.01	142	10222	9.49636	ppb	94
8) 1,1'-Biphenyl	7.45	154	12349	9.87257	ppb #	88
9) Acenaphthylene	7.93	152	16024	9.64536	ppb	98
10) Acenaphthene	8.13	154	8901	9.67450	ppb	93
11) Fluorene	8.74	166	10449	9.90386	ppb	97
13) Phenanthrene	9.86	178	14996	9.77834	ppb	99
14) Anthracene	9.92	178	14348	9.38520	ppb	99
15) Fluoranthene	11.23	202	21536	9.74671	ppb	99
17) Pyrene	11.49	202	21902	9.67353	ppb	92
19) Benz (a) anthracene	12.89	228	18864	9.44825	ppb	97
20) Chrysene	12.94	228	18670	9.47946	ppb #	96
21) Indeno (1,2,3-cd) pyrene	15.99	276	19639	9.69329	ppb #	90
23) Benzo (b) fluoranthene	14.08	252	17117	9.11749	ppb #	86
24) Benzo (k) fluoranthene	14.12	252	20282	10.52648	ppb #	92
25) Benzo (a) pyrene	14.45	252	17798	9.70662	ppb	99
26) Dibenz (a,h) anthracene	16.02	278	16005	9.74367	ppb #	94
27) Benzo (g,h,i) perylene	16.43	276	16439	9.60673	ppb	97

Data File : M:\LINUS\DATA\L120613\0613L009.D Vial: 9
 Acq On : 13 Jun 12 16:25 Operator: LF
 Sample : 50ug/ml PAH Inst : Linus
 Misc : Multiplr: 1.00

Quant Time: Jun 13 17:37 2012 Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jun 13 14:51:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8 (IS)	6.09	136	2323	2.50000	ppb	-0.02
6) Acenaphthene-D10 (IS)	8.10	164	1076	2.50000	ppb	-0.04
12) Phenanthrene-D10 (IS)	9.82	188	1906	2.50000	ppb	-0.04
16) Chrysene-D12 (IS)	12.91	240	2336	2.50000	ppb	-0.01
22) Perylene-D12 (IS)	14.52	264	1770	2.50000	ppb	-0.01
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.31	82	22158	51.01380	ppb	-0.02
Spiked Amount	2.000				Recovery = 2550.700%	
7) Surrogate Recovery (FBP)	7.34	172	42363	39.70801	ppb	-0.04
Spiked Amount	2.000				Recovery = 1985.400%	
18) Surrogate Recovery (TPH)	11.70	244	46329	37.33504	ppb	-0.03
Spiked Amount	2.000				Recovery = 1866.750%	
Target Compounds						
						Qvalue
3) Naphthalene	6.11	128	65485	41.48686	ppb	98
4) 2-Methylnaphthalene	6.90	142	43032	42.12800	ppb	92
5) 1-Methylnaphthalene	7.01	142	39886	39.68464	ppb	95
8) 1,1'-Biphenyl	7.45	154	48419	40.95469	ppb	# 87
9) Acenaphthylene	7.93	152	60904	38.93445	ppb	97
10) Acenaphthene	8.13	154	35017	40.40146	ppb	92
11) Fluorene	8.74	166	40304	40.39620	ppb	97
13) Phenanthrene	9.86	178	57308	39.37645	ppb	98
14) Anthracene	9.92	178	57012	39.55630	ppb	99
15) Fluoranthene	11.23	202	80905	38.60379	ppb	# 91
17) Pyrene	11.50	202	87777	39.59828	ppb	# 83
19) Benz (a) anthracene	12.90	228	77651	39.87510	ppb	99
20) Chrysene	12.94	228	65735	34.20150	ppb	# 92
21) Indeno (1,2,3-cd) pyrene	15.99	276	77220	38.91637	ppb	# 80
23) Benzo (b) fluoranthene	14.09	252	78843	48.95647	ppb	# 80
24) Benzo (k) fluoranthene	14.12	252	64724	38.24790	ppb	94
25) Benzo (a) pyrene	14.45	252	67281	42.35279	ppb	# 96
26) Dibenz (a,h) anthracene	16.03	278	62359	43.79148	ppb	97
27) Benzo (g,h,i) perylene	16.44	276	64939	43.87588	ppb	99

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0613L009.D
 Acq On : 13 Jun 12 16:25
 Sample : 50ug/ml PAH
 Misc :

Vial: 9
 Operator: LF
 Inst : Linus
 Multiplr: 1.00

Quant Time: Jun 13 17:37 2012

Quant Results File: SIMB.RES

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jun 13 17:38:06 2012
 Response via : Initial Calibration



Data File : M:\LINUS\DATA\L120613\0613L010.D Vial: 10
 Acq On : 13 Jun 12 16:51 Operator: LF
 Sample : 100ug/ml PAH Inst : Linus
 Misc : Multiplr: 1.00

Quant Time: Jun 13 17:37 2012 Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jun 13 14:51:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8 (IS)	6.09	136	2546	2.50000	ppb	-0.02
6) Acenaphthene-D10 (IS)	8.10	164	1146	2.50000	ppb	-0.04
12) Phenanthrene-D10 (IS)	9.82	188	2043	2.50000	ppb	-0.04
16) Chrysene-D12 (IS)	12.91	240	2154	2.50000	ppb	-0.01
22) Perylene-D12 (IS)	14.52	264	2023	2.50000	ppb	-0.01
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.31	82	46683	97.78012	ppb	-0.02
Spiked Amount	2.000				Recovery = 4889.000%	
7) Surrogate Recovery (FBP)	7.34	172	86281	78.23418	ppb	-0.04
Spiked Amount	2.000				Recovery = 3911.700%	
18) Surrogate Recovery (TPH)	11.70	244	90106	81.70547	ppb	-0.03
Spiked Amount	2.000				Recovery = 4085.250%	
Target Compounds						
3) Naphthalene	6.12	128	130271	77.17939	ppb	99
4) 2-Methylnaphthalene	6.90	142	84094	76.84481	ppb	94
5) 1-Methylnaphthalene	7.01	142	77537	72.52602	ppb	94
8) 1,1'-Biphenyl	7.45	154	93605	76.31079	ppb #	91
9) Acenaphthylene	7.94	152	123810	76.74039	ppb	99
10) Acenaphthene	8.13	154	66674	74.26410	ppb	89
11) Fluorene	8.74	166	76061	73.59790	ppb	99
13) Phenanthrene	9.86	178	112505	74.37620	ppb	97
14) Anthracene	9.92	178	110199	73.52547	ppb	97
15) Fluoranthene	11.23	202	163589	75.27303	ppb #	83
17) Pyrene	11.50	202	169609	85.52128	ppb #	90
19) Benz (a) anthracene	12.90	228	148541	85.18770	ppb	98
20) Chrysene	12.95	228	138030	81.56593	ppb	97
21) Indeno (1,2,3-cd) pyrene	16.02	276	155909	87.99871	ppb #	87
23) Benzo (b) fluoranthene	14.09	252	139278	76.65546	ppb #	85
24) Benzo (k) fluoranthene	14.13	252	145240	79.13503	ppb	89
25) Benzo (a) pyrene	14.47	252	125203	71.12137	ppb	96
26) Dibenz (a,h) anthracene	16.04	278	123729	78.09989	ppb #	94
27) Benzo (g,h,i) perylene	16.45	276	132960	80.72903	ppb #	89

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0613L010.D
Acq On : 13 Jun 12 16:51
Sample : 100ug/ml PAH
Misc :

Vial: 10
Operator: LF
Inst : Linus
Multiplr: 1.00

Quant Time: Jun 13 17:37 2012

Quant Results File: SIMB.RES

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jun 13 17:38:06 2012
Response via : Initial Calibration



EPA 8270C SIM

Form 7

Second Source Calibration

Lab Name: APPL, Inc.

SDG No: 68258

Case No: _____

Date Analyzed: 06/13/12

Matrix: _____

Instrument: Linus

Initial Cal. Date: 06/13/12

Data File: 0613L011.D

		Compound	MEAN	CCRF	%D	%Drift
1	I	Napthalene-D8(IS)	ISTD			I
2	TM	Napthalene	1.610	1.637	1.7	TM
3	TM	2-Methylnapthalene	1.043	1.049	0.54	TM
4	TM	1-Methylnapthalene	1.050	1.039	1.1	TM
5	I	Acenaphthene-D10(IS)	ISTD			I
6	TM	1,1'-Biphenyl	2.597	2.752	6.0	TM
7	TM	Acenaphthylene	3.417	3.382	1.0	TM
8	*TM	Acenaphthene	1.896	1.964	3.6	*TM
9	TM	Fluorene	2.180	2.312	6.0	TM
10	I	Phenanthrene-D10(IS)	ISTD			I
11	TM	Phenanthrene	1.792	1.916	6.9	TM
12	TM	Anthracene	1.773	1.884	6.2	TM
13	*TM	Fluoranthene	2.577	2.638	2.4	*TM
14	I	Chrysene-D12(IS)	ISTD			I
15	TM	Pyrene	2.260	2.408	6.6	TM
16	TM	Benz (a) anthracene	1.986	2.024	1.9	TM
17	TM	Chrysene	1.919	2.238	17	TM
18	TM	Indeno (1,2,3-cd) pyrene	2.025	2.112	4.3	TM
19	I	Perylene-D12(IS)	ISTD			I
20	TM	Benzo (b) fluoranthene	2.200	2.149	2.3	TM
21	TM	Benzo (k) fluoranthene	2.246	2.481	11	TM
22	*TM	Benzo (a) pyrene	2.114	2.200	4.1	*TM
23	TM	Dibenz (a,h) anthracene	1.920	2.027	5.6	TM
24	TM	Benzo (g,h,i) perylene	2.003	2.072	3.5	TM
25						
26						
27						
28						
29						
30						
31						
32						
33						
34						
35						
36						
37						
38						
39						
40						

Average

4.8

Data File : M:\LINUS\DATA\L120613\0613L011.D Vial: 11
 Acq On : 13 Jun 12 17:17 Operator: LF
 Sample : 5.0ug/ml SS PAH 06-13-12 Inst : Linus
 Misc : Multiplr: 1.00

Quant Time: Jun 13 17:38 2012 Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jun 13 17:38:06 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Naphthalene-D8 (IS)	6.09	136	2569	2.50000	ppb	-0.02
6) Acenaphthene-D10 (IS)	8.10	164	1144	2.50000	ppb	-0.04
12) Phenanthrene-D10 (IS)	9.82	188	1967	2.50000	ppb	-0.04
16) Chrysene-D12 (IS)	12.90	240	2262	2.50000	ppb	-0.02
22) Perylene-D12 (IS)	14.52	264	1992	2.50000	ppb	-0.01

System Monitoring Compounds

2) Surrogate Recovery (NBZ)	0.00	82	0d	0.00000	ppb	
Spiked Amount	2.000		Recovery	=	0.000%	
7) Surrogate Recovery (FBP)	0.00	172	0d	0.00000	ppb	
Spiked Amount	2.000		Recovery	=	0.000%	
18) Surrogate Recovery (TPH)	0.00	244	0d	0.00000	ppb	
Spiked Amount	2.000		Recovery	=	0.000%	

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	6.12	128	8410	5.08291	ppb	100
4) 2-Methylnaphthalene	6.90	142	5390	5.02676	ppb	95
5) 1-Methylnaphthalene	7.01	142	5336	4.94647	ppb	94
8) 1,1'-Biphenyl	7.45	154	6296	5.29864	ppb #	88
9) Acenaphthylene	7.93	152	7739	4.94910	ppb	97
10) Acenaphthene	8.13	154	4494	5.18102	ppb	93
11) Fluorene	8.74	166	5289	5.30164	ppb	98
13) Phenanthrene	9.86	178	7536	5.34571	ppb	99
14) Anthracene	9.92	178	7411	5.31149	ppb	98
15) Fluoranthene	11.23	202	10378	5.11798	ppb	96
17) Pyrene	11.49	202	10896	5.32816	ppb #	90
19) Benz (a) anthracene	12.90	228	9158	5.09566	ppb	96
20) Chrysene	12.94	228	10125	5.83187	ppb #	96
21) Indeno (1,2,3-cd) pyrene	15.99	276	9556	5.21433	ppb #	91
23) Benzo (b) fluoranthene	14.08	252	8563	4.88587	ppb #	84
24) Benzo (k) fluoranthene	14.12	252	9886	5.52530	ppb #	92
25) Benzo (a) pyrene	14.45	252	8766	5.20460	ppb	98
26) Dibenz (a,h) anthracene	16.02	278	8077	5.28014	ppb	95
27) Benzo (g,h,i) perylene	16.43	276	8254	5.17286	ppb	97

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0613L011.D
Acq On : 13 Jun 12 17:17
Sample : 5.0ug/ml SS PAH 06-13-12
Misc :

Vial: 11
Operator: LF
Inst : Linus
Multiplr: 1.00

Quant Time: Jun 13 17:38 2012

Quant Results File: SIMB.RES

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jun 13 17:38:06 2012
Response via : Initial Calibration



EPA 8270C SIM

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
 Case No: _____
 Matrix: _____

SDG No: 69258
 Date Analyzed: 07/24/12
 Instrument: Linus
 Initial Cal. Date: 06/13/12
 Data File: 0724L002.D

		Compound	MEAN	CCRF	%D	%Drift
1	I	Napthalene-D8(IS)	ISTD			I
2	S	Surrogate Recovery (NBZ)	0.4675	0.4917	5.2	S
3	TM	Naphthalene	1.610	1.635	1.6	TM
4	TM	2-Methylnaphthalene	1.043	1.021	2.1	TM
5	TM	1-Methylnaphthalene	1.050	1.009	3.8	TM
6	I	Acenaphthene-D10(IS)	ISTD			I
7	S	Surrogate Recovery (FBP)	2.340	2.514	7.4	S
8	TM	1,1'-Biphenyl	2.597	2.910	12	TM
9	TM	Acenaphthylene	3.417	3.777	11	TM
10	*TM	Acenaphthene	1.896	2.038	7.5	*TM
11	TM	Fluorene	2.180	2.387	9.5	TM
12	I	Phenanthrene-D10(IS)	ISTD			I
13	TM	Phenanthrene	1.792	2.108	18	TM
14	TM	Anthracene	1.773	2.054	16	TM
15	*TM	Fluoranthene	2.577	2.968	15	*TM
16	I	Chrysene-D12(IS)	ISTD			I
17	TM	Pyrene	2.260	2.410	6.6	TM
18	S	Surrogate Recovery (TPH)	1.251	1.410	13	S
19	TM	Benz (a) anthracene	1.986	1.933	2.7	TM
20	TM	Chrysene	1.919	1.988	3.6	TM
21	TM	Indeno (1,2,3-cd) pyrene	2.025	1.648	19	TM
22	I	Perylene-D12(IS)	ISTD			I
23	TM	Benzo (b) fluoranthene	2.200	2.040	7.3	TM
24	TM	Benzo (k) fluoranthene	2.246	2.293	2.1	TM
25	*TM	Benzo (a) pyrene	2.114	1.964	7.1	*TM
26	TM	Dibenz (a,h) anthracene	1.920	1.569	18	TM
27	TM	Benzo (g,h,i) perylene	2.003	1.675	16	TM
28						
29						
30						
31						
32						
33						
34						
35						
36						
37						
38						
39						
40						

Average

9.3

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L120613\0724L002.D Vial: 2
 Acq On : 24 Jul 12 18:24 Operator: LF
 Sample : 5.0ug/ml PAH 06-13-12 Inst : Linus
 Misc : Multiplr: 1.00

Quant Time: Jul 27 7:44 2012 Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Fri Jul 13 13:02:51 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8 (IS)	6.07	136	2955	2.50000	ppb	-0.05
6) Acenaphthene-D10 (IS)	8.08	164	1209	2.50000	ppb	-0.05
12) Phenanthrene-D10 (IS)	9.81	188	1981	2.50000	ppb	-0.05
16) Chrysene-D12 (IS)	12.90	240	2531	2.50000	ppb	0.00
22) Perylene-D12 (IS)	14.52	264	2136	2.50000	ppb	-0.01
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.30	82	2906	5.25890	ppb	-0.04
Spiked Amount	2.000		Recovery	=	262.950%	
7) Surrogate Recovery (FBP)	7.31	172	6080	5.37184	ppb	-0.06
Spiked Amount	2.000		Recovery	=	268.600%	
18) Surrogate Recovery (TPH)	11.69	244	7139	5.63813	ppb	-0.05
Spiked Amount	2.000		Recovery	=	281.900%	
Target Compounds						
3) Napthalene	6.09	128	9664	5.07785	ppb	99
4) 2-Methylnapthalene	6.89	142	6037	4.89471	ppb	90
5) 1-Methylnapthalene	7.00	142	5966	4.80805	ppb	97
8) 1,1'-Biphenyl	7.43	154	7037	5.60386	ppb #	86
9) Acenaphthylene	7.92	152	9132	5.52595	ppb	99
10) Acenaphthene	8.12	154	4929	5.37700	ppb	96
11) Fluorene	8.72	166	5771	5.47378	ppb	96
13) Phenanthrene	9.83	178	8351	5.88197	ppb	96
14) Anthracene	9.91	178	8138	5.79131	ppb	98
15) Fluoranthene	11.22	202	11760	5.75853	ppb	96
17) Pyrene	11.48	202	12200	5.33176	ppb #	88
19) Benz (a) anthracene	12.89	228	9785	4.86588	ppb	99
20) Chrysene	12.92	228	10065	5.18116	ppb #	93
21) Indeno (1,2,3-cd) pyrene	16.01	276	8340	4.06714	ppb #	79
23) Benzo (b) fluoranthene	14.08	252	8713	4.63630	ppb #	83
24) Benzo (k) fluoranthene	14.10	252	9795	5.10538	ppb	98
25) Benzo (a) pyrene	14.45	252	8389	4.64499	ppb	99
26) Dibenz (a,h) anthracene	16.03	278	6703	4.08651	ppb #	94
27) Benzo (g,h,i) perylene	16.45	276	7156	4.18239	ppb	94

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0724L002.D
Acq On : 24 Jul 12 18:24
Sample : 5.0ug/ml PAH 06-13-12
Misc :

Vial: 2
Operator: LF
Inst : Linus
Multiplr: 1.00

Quant Time: Jul 27 7:44 2012

Quant Results File: SIMB.RES

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jul 25 18:38:43 2012
Response via : Initial Calibration



EPA METHOD 8270
Polynuclear Aromatic Hydrocarbons
Raw Data

Method Blank EPA 8270D SIM

Blank Name/QCG: 120723W-65144 - 169459
Batch ID: #SIMHC-120723A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	1-METHYLNAPHTHALENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	2-METHYLNAPHTHALENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	ACENAPHTHENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	ACENAPHTHYLENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	ANTHRACENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
BLANK	BENZO(A)ANTHRACENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
BLANK	BENZO(A)PYRENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	BENZO(B)FLUORANTHENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	BENZO(GHI)PERYLENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
BLANK	BENZO(K)FLUORANTHENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
BLANK	CHRYSENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
BLANK	DIBENZ(A,H)ANTHRACENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
BLANK	FLUORANTHENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
BLANK	FLUORENE	0.12 U	0.2	0.12	0.06	ug/L	07/23/12	07/24/12
BLANK	INDENO(1,2,3-CD)PYRENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
BLANK	NAPHTHALENE	0.10 U	0.2	0.10	0.05	ug/L	07/23/12	07/24/12
BLANK	PHENANTHRENE	0.14 U	0.2	0.14	0.07	ug/L	07/23/12	07/24/12
BLANK	PYRENE	0.16 U	0.2	0.16	0.08	ug/L	07/23/12	07/24/12
BLANK	SURROGATE: 2-FLUORBIPHENY	56.8	50-110			%	07/23/12	07/24/12
BLANK	SURROGATE: NITROBENZENE-	51.8	40-110			%	07/23/12	07/24/12
BLANK	SURROGATE: TERPHENYL-D14 (59.6	50-135			%	07/23/12	07/24/12

Quant Method: SIMB.M
Run #: 0724L003
Instrument: Linus
Sequence: L120613
Initials: LF

Printed: 07/27/12 5:34:51 PM
GC SC-Blank-REG MDLs

Quantitation Report (QT Reviewed)

Data File : M:\LINUS\DATA\L120613\0724L003.D Vial: 3
 Acq On : 24 Jul 12 18:50 Operator: LF
 Sample : 120723A BLK 1/1000 Inst : Linus
 Misc : Multiplr: 1.00

Quant Time: Jul 27 7:45 2012 Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jul 25 18:38:43 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8 (IS)	6.07	136	2273	2.50000	ppb	-0.05
6) Acenaphthene-D10 (IS)	8.08	164	1022	2.50000	ppb	-0.05
12) Phenanthrene-D10 (IS)	9.82	188	2049	2.50000	ppb	-0.04
16) Chrysene-D12 (IS)	12.91	240	2655	2.50000	ppb	0.01
22) Perylene-D12 (IS)	14.54	264	2331	2.50000	ppb	0.00
System Monitoring Compounds						
2) Surrogate Recovery (NBZ)	5.32	82	440	1.03517	ppb	-0.01
Spiked Amount	2.000		Recovery	=	51.750%	
7) Surrogate Recovery (FBP)	7.32	172	1086	1.13508	ppb	-0.05
Spiked Amount	2.000		Recovery	=	56.750%	
18) Surrogate Recovery (TPH)	11.69	244	1583	1.19181	ppb	-0.05
Spiked Amount	2.000		Recovery	=	59.600%	

Target Compounds Qvalue

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0724L003.D
Acq On : 24 Jul 12 18:50
Sample : 120723A BLK 1/1000
Misc :

Vial: 3
Operator: LF
Inst : Linus
Multiplr: 1.00

Quant Time: Jul 27 7:45 2012

Quant Results File: SIMB.RES

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jul 25 18:38:43 2012
Response via : Initial Calibration



Laboratory Control Spike Recovery

EPA 8270D SIM

APPL ID: 120723W-65144 LCS - 169459

Batch ID: #SIMHC-120723A

APPL Inc.

908 North Temperance Avenue
Clovis, CA 93611

Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits
1-METHYLNAPHTHALENE	4.00	2.22	55.5	45-105
2-METHYLNAPHTHALENE	4.00	2.18	54.5	45-105
ACENAPHTHENE	4.00	2.18	54.5	45-110
ACENAPHTHYLENE	4.00	2.33	58.3	50-105
ANTHRACENE	4.00	2.39	59.8	55-110
BENZO(A)ANTHRACENE	4.00	2.21	55.3	55-110
BENZO(A)PYRENE	4.00	2.24	56.0	55-110
BENZO(B)FLUORANTHENE	4.00	2.23	55.8	45-120
BENZO(GHI)PERYLENE	4.00	2.46	61.5	40-125
BENZO(K)FLUORANTHENE	4.00	2.57	64.3	45-125
CHRYSENE	4.00	2.67	66.8	55-110
DIBENZ(A,H)ANTHRACENE	4.00	2.29	57.3	40-125
FLUORANTHENE	4.00	3.03	75.8	55-115
FLUORENE	4.00	2.75	68.8	50-110
INDENO(1,2,3-CD)PYRENE	4.00	2.25	56.3	45-125
NAPHTHALENE	4.00	2.15	53.8	40-100
PHENANTHRENE	4.00	2.78	69.5	50-115
PYRENE	4.00	2.45	61.3	50-130

SURROGATE: 2-FLUORBIPHENYL (S)	2.00	1.26	63.0	50-110
SURROGATE: NITROBENZENE-D5 (S)	2.00	1.49	74.5	40-110
SURROGATE: TERPHENYL-D14 (S)	2.00	1.16	58.0	50-135

Comments: _____

<u>Primary</u>	<u>SPK</u>
Quant Method :	SIMB.M
Extraction Date :	07/23/12
Analysis Date :	07/24/12
Instrument :	Linus
Run :	0724L004
Initials :	LF

Printed: 07/27/12 5:34:53 PM

APPL Standard LCS

Data File : M:\LINUS\DATA\L120613\0724L004.D Vial: 4
 Acq On : 24 Jul 12 19:16 Operator: LF
 Sample : 120723A LCS-1 1/1000 Inst : Linus
 Misc : Multiplr: 1.00

Quant Time: Jul 27 7:48 2012 Quant Results File: SIMB.RES

Quant Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C
 Last Update : Wed Jul 25 18:38:43 2012
 Response via : Initial Calibration
 DataAcq Meth : 87SIMAQ

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Napthalene-D8 (IS)	6.07	136	2043 ✓	2.50000	ppb ✓	-0.05
6) Acenaphthene-D10 (IS)	8.08	164	992	2.50000	ppb	-0.05
12) Phenanthrene-D10 (IS)	9.82	188	1998	2.50000	ppb	-0.04
16) Chrysene-D12 (IS)	12.90	240	2829	2.50000	ppb	0.00
22) Perylene-D12 (IS)	14.52	264	2395	2.50000	ppb	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) Surrogate Recovery (NBZ)	5.31	82	570	1.49198	ppb	-0.02
Spiked Amount	2.000		Recovery	=	74.600%	
7) Surrogate Recovery (FBP)	7.32	172	1171	1.26093	ppb	-0.05
Spiked Amount	2.000		Recovery	=	63.050%	
18) Surrogate Recovery (TPH)	11.69	244	1642	1.16019	ppb	-0.05
Spiked Amount	2.000		Recovery	=	58.000%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Napthalene	6.09	128	2835 ✓	2.15459 ✓	ppb	99
4) 2-Methylnaphthalene	6.89	142	1857	2.17775	ppb	91
5) 1-Methylnaphthalene	7.00	142	1904	2.21943	ppb	97
8) 1,1'-Biphenyl	7.43	154	2358	2.28854	ppb	89
9) Acenaphthylene	7.92	152	3161	2.33120	ppb	97
10) Acenaphthene	8.12	154	1637	2.17643	ppb	97
11) Fluorene	8.72	166	2378	2.74892	ppb	99
13) Phenanthrene	9.85	178	3979	2.77873	ppb	99
14) Anthracene	9.91	178	3390	2.39193	ppb	99
15) Fluoranthene	11.22	202	6232	3.02567	ppb #	90
17) Pyrene	11.49	202	6273	2.45271	ppb #	90
19) Benz (a) anthracene	12.89	228	4978	2.21470	ppb	97
20) Chrysene	12.94	228	5791	2.66702	ppb	99
21) Indeno (1,2,3-cd) pyrene	16.02	276	5164	2.25304	ppb	79
23) Benzo (b) fluoranthene	14.08	252	4702	2.23142	ppb	84
24) Benzo (k) fluoranthene	14.12	252	5525	2.56833	ppb #	93
25) Benzo (a) pyrene	14.45	252	4531	2.23751	ppb	97
26) Dibenz (a,h) anthracene	16.03	278	4217	2.29289	ppb	90
27) Benzo (g,h,i) perylene	16.45	276	4720	2.46032	ppb	92

$\frac{2835 \times 2.5}{2043 \times 1.66} = 2.15$
 WFB/11/12

Quantitation Report


Data File : M:\LINUS\DATA\L120613\0724L004.D
Acq On : 24 Jul 12 19:16
Sample : 120723A LCS-1 1/1000
Misc :

Vial: 4
Operator: LF
Inst : Linus
Multiplr: 1.00

Quant Time: Jul 27 7:48 2012

Quant Results File: SIMB.RES



Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
Title : EPA 8270C
Last Update : Wed Jul 25 18:38:43 2012
Response via : Initial Calibration



Data File : M:\LINUS\DATA\L120613\0613L001.D
 Acq On : 13 Jun 12 13:07
 Sample : SVTUNE 2-28-12
 Misc :

Vial: 1
 Operator: LF
 Inst : Linus
 Multiplr: 1.00

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C



Spectrum Information: Average of 6.939 to 6.958 min.



Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	43.8	164628	PASS
68	69	0.00	2	0.0	0	PASS
70	69	0.00	2	0.5	681	PASS
127	198	40	60	48.3	181462	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	375479	PASS
199	198	5	9	7.2	27064	PASS
275	198	10	30	22.8	85713	PASS
365	198	1	100	2.5	9250	PASS
441	443	0.01	100	75.8	30917	PASS
442	198	40	150	53.9	202264	PASS
443	442	17	23	20.2	40782	PASS

DFTPP

Data File : M:\LINUS\DATA\L120613\0724L001.D
 Acq On : 24 Jul 12 18:05
 Sample : SVTUNE 2-28-12
 Misc :

Vial: 1
 Operator: LF
 Inst : Linus
 Multiplr: 1.00

Method : M:\LINUS\DATA\L120613\SIMB.M (RTE Integrator)
 Title : EPA 8270C



Spectrum Information: Average of 6.921 to 6.939 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	30	60	56.9	159505	PASS
68	69	0.00	2	0.1	140	PASS
70	69	0.00	2	0.8	952	PASS
127	198	40	60	54.7	153315	PASS
197	198	0.00	1	0.0	0	PASS
198	198	100	100	100.0	280338	PASS
199	198	5	9	7.1	20025	PASS
275	198	10	30	23.7	66402	PASS
365	198	1	100	3.1	8553	PASS
441	443	0.01	100	76.8	31366	PASS
442	198	40	150	72.0	201931	PASS
443	442	17	23	20.2	40841	PASS

VF 10/18/12

Part #: 10001 Laboratory Use Only - See MSDS
 Lot #: 042910 Exp: 042913 Storage 0 °C
 CLP Semi-Volatiles Base/Neutrals Mix #1
 14 components
 2000 ug/mL in methy
ABSOLUTE STANDARD
 CLP Semi-Volatiles Base/Neutrals Mix #1
 Lot #: 042910 - 28440
 Rec: 3/8/11 MFR exp. 4/29/2013

exp 10/18/12

VF 10/18/12

Part #: 10001 Laboratory Use Only - See MSDS
 Lot #: 042910 Exp: 042913 Storage 0 °C
 CLP Semi-Volatiles Base/Neutrals Mix #1
 14 components
 2000 ug/mL in m
ABSOLUTE STANDAR
 CLP Semi-Volatiles Base/Neutrals Mix #1
 Lot #: 042910 - 29085
 Rec: 8/4/11 MFR exp. 04/29/13

exp 10/18/12

VF 7/31/12

Part #: 10002 Laboratory Use Only - See MSDS
 Lot #: 073109 Exp: 073112 Storage 4 °C
 CLP Semi-Volatiles Base/Neutrals Mix #2
 14 components
 2000 ug/mL in methyle
ABSOLUTE STANDARDS
 CLP Semi-Volatiles Base/Neutrals Mix #2
 Lot #: 073109 - 28446
 Rec: 3/8/11 MFR exp. 7/31/2012

exp 7/31/12

VF 7/31/12

Part #: 10002 Laboratory Use Only - See MSDS
 Lot #: 073109 Exp: 073112 Storage 4 °C
 CLP Semi-Volatiles Base/Neutrals Mix #2
 14 components
 2000 ug/mL in met
ABSOLUTE STANDAR
 CLP Semi-Volatiles Base Neutrals Mix #2
 Lot #: 073109 - 29090
 Rec: 8/4/11 MFR exp. 07/31/12

exp 7/31/12

VF 10/15/12

Part #: 10004 Laboratory Use Only - See MSDS
 Lot #: 101509 Exp: 101514 Storage 4 °C
 CLP Semi-Volatiles Toxic Substances #1
 4 components
 2000 ug/mL in methyl
ABSOLUTE STANDARD
 CLP Semi-Volatiles Toxic Substances #1
 Lot #: 101509 - 28453
 Rec: 3/8/11 MFR exp. 10/15/201

exp 10/15/12

VF 10/15/12

Part #: 10004 Laboratory Use Only - See MSDS
 Lot #: 101509 Exp: 101514 Storage 4 °C
 CLP Semi-Volatiles Toxic Substances #1
 4 components
 2000 ug/mL in met
ABSOLUTE STANDAR
 CLP Semi-Volatiles Toxic Substances #1
 Lot #: 101509 - 29095
 Rec: 8/4/11 MFR exp. 10/15/14

exp 10/15/12

VF 10/18/12

Part #: 10005 Laboratory Use Only - See MSDS
 Lot #: 061209 Exp: 061214 Storage 4 °C
 CLP Semi-Volatiles Toxic Substances #2
 8 components
 2000 ug/mL in methy
ABSOLUTE STANDARD
 CLP Semi-Volatiles Toxic Substances #2
 Lot #: 061209 - 28458
 Rec: 3/8/11 MFR exp. 6/12/2014

exp 10/18/12

VF 10/18/12

Part #: 10005 Laboratory Use Only - See MSDS
 Lot #: 121208 Exp: 121213 Storage 4 °C
 CLP Semi-Volatiles Toxic Substances #2
 8 components
 2000 ug/mL in met
ABSOLUTE STANDAR
 CLP Semi-Volatiles Toxic Substances #2
 Lot #: 121208 - 29100
 Rec: 8/4/11 MFR exp. 12/12/13

exp 10/18/12

up 2/25/12

8270D PAH SIM Solution,
200 mg/L, 1 ml
110780-01
Lot # Storage Expiry
170253 S-10 Degrees C 3/3/13
Solv: Methylene Chloride
3270D PAH SIM
Lot # 170253-28478
Rec 3/10/11 MFR exp 3/3/2013

up 2/25/13

up 2/25/12

8270D PAH SIM Solution,
Second Source, 200 mg/L, 1 ml
110780-01-85
Lot # Storage Expiry
170254 S-10 Degrees C 3/3/13
Solv: Methylene Chloride
8270D PAH SIM (SS)
Lot # 170256-28490
Rec 3/10/11 MFR exp 3/3/2013

up 2/25/13

up 2/25/12

8270 BN:A (200:400)
Surrogate Solution, 1 ml
118004-17
Lot # Storage Expiry
167802 S-10 Degrees C 4/9/13
Solv: Methylene Chloride
8270 BN:A (200:400) Surrogate Solution
Lot #: 167802-29314
Rec: 8/8/11 MFR exp 01/09/13

up 2/25/13

up 2/25/12

Method 8270 Internal
Standard Solution, 2,000
mg/L, 1 ml
118001-42
Lot # Storage Expiry
167766 S-10 Degrees C 4/20/13
Solv: Methylene Chloride
8270 Internal Standard
Lot #. 167766-28151
Rec 1/20/11 MFR exp. 04/20/13

up 2/25/13

up 2/25/12

PREP DATE:	02-25-12					
SIM Semivolatiles Int. Std. Mix	125 ug/ml					
Exp:	08-25-12					
Supplier	ID #	µg/mL	Lot #	Code	Exp. Date	µL
O2SI	Int. Std	2000	167766-28151	02/25/12	02-25-13	100
EM Science	MeCl2		47186			1500
						1600

up 2/25/12

PREP DATE:	02-25-12													
8270 SIM STANDARD CURVE														
		Conc.		Date	CODE:	0.10	0.20	0.50	1.00	5.00	10.00	50.00	100.00	
Supplier	ID #	µg/mL	Lot #	Code	Exp. Date	µL	µL	µL	µL	µL	µL	µL	µL	
	8270D PAH SIM	200	170253-28478	02/25/12	02-25-13	0	0	0	0	5	5	25	50	
	5 0ug/mL	5		02/25/12		0	0	10	20	0	0	0	0	
	1 0ug/mL	1		02/25/12		10	20	0	0	0	0	0	0	
	Surrogate Stock	VAR	167802-29314	02/25/12	01-09-13	0	0	0	0	5	5	25	50	
EM Science	Methylene Chloride		47186			90	80	90	80	190	90	50	0	
					Final Vol	100	100	100	100	200	100	100	100	

VF 2/27/12

PREP DATE:	02-25-12								
SIM 8270 Second Source (5µg/mL)									
Exp:	03-10-12								
		Conc.		Date		CODE:			
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date		µL		
	8270D PAH SIM (SS)	200	170256-28490	02/25/12	02-25-13		5		
	MeCl2		Lot#47186				195		
				Final Volume			200		

VF 2/28/12

GCM-160-1
 Lot CH-2137
 Exp 07/31/2013
 Semi-Volatiles GC/MS Tuning Standard
 4 analyte(s) at 1000 µg/mL in dichloromethane
 250 Smith St, No Kingstown, RI 02852 USA



on 2/28/12

PREP DATE:	02-28-12								
SV Tune Mix 50ug/ml									
Exp:	02-28-13								
		Conc.		Date		CODE:			B
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date		µL		
U Scientific	GCM-150	1000	CH-2137	02/28/12	07-31-13		1000		
EM Science	MeCl2		47080				19000		
				Final Vol			20000		

VF 2/28/12

VF 2/29/12

PREP DATE:	02-29-12													
8270 SIM STANDARD CURVE														
		Conc.		Date		CODE:	A	A	C	D	E	F	G	H
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date		µL							
	8270D PAH SIM	200	170253-28478	02/25/12	02-25-13		0	0	0	0	5	5	25	50
	5.0ug/mL	5		02/29/12			0	0	10	20	0	0	0	0
	1.0ug/mL	1		02/29/12			10	20	0	0	0	0	0	0
	Surrogate Stock	VAR	167802-29314	02/25/12	01-09-13		0	0	0	0	5	5	25	50
EM Science	Methylene Chloride		47186				90	80	90	80	190	90	50	0
				Final Vol.			100	100	100	100	200	100	100	100

VF 2/29/12

PREP DATE:	02-29-12								
SIM 8270 Second Source (5µg/mL)									
Exp:	03-14-12								
		Conc.		Date		CODE:			
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date		µL		
	8270D PAH SIM (SS)	200	170256-28490	02/25/12	02-25-13		5		
	MeCl2		Lot#47186				195		
				Final Volume			200		

VF 3/18/12

PREP DATE:	03-18-12													
8270 STANDARD CURVE														
		Conc.		Date			5	10	20	40	50	60	80	100
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date		µL							
	8270T Stock	200		02/13/12	07-31-12		5	5	10	20	25	30	40	50
	Surrogate Stock	VAR	167802-29314	02/25/12	01-09-13		5	5	10	20	25	30	40	50
EM Science	Methylene Chloride		47186				190	90	80	60	50	40	20	0
				Final Vol.			200	100	100	100	100	100	100	100

VF 3/18/12

PREP DATE:	03-18-12								
8270 Second Source (SS) 50ug/mL									
		Conc.		Date		CODE:			50
Supplier	ID #	µg/mL	Lot #	Code	Exp.Date		µL		
	8270C SS	200		10/11/11	10-11-12		25		
EM Science	Methylene Chloride		47186				75		

VF 5/1/12

Part #: 10001 Laboratory Use Only - See MSDS
 Lot #: 042910 Exp: 042913 Storage 0 °C
 **CLP Semi-Volatiles Base/Neutrals Mix #1**
 14 components CLP Semi-Volatiles Base/Neutrals Mix #1
 2000 ug/mL in meth Lot #: 042910 - 29081
 Rec: 8/4/11 MFR exp. 04/29/13
ABSOLUTE STANDARD

exp 4/29/13

VF 5/1/12

Part #: 10002 Laboratory Use Only - See MSDS
 Lot #: 073109 Exp: 073112 Storage 4 °C
 **CLP Semi-Volatiles Base/Neutrals Mix #2**
 14 components CLP Semi-Volatiles Base Neutrals Mix #2
 2000 ug/mL in meth Lot #: 073109 - 29086
 Rec: 8/4/11 MFR exp. 07/31/12
ABSOLUTE STANDARD

exp 7/31/12

VF 5/1/12

Part #: 10004 Laboratory Use Only - See MSDS
 Lot #: 101509 Exp: 101514 Storage 4 °C
 **CLP Semi-Volatiles Toxic Substances #1**
 4 components CLP Semi-Volatiles Toxic Substances #1
 2000 ug/mL in meth Lot #: 101509 - 29091
 Rec: 8/4/11 MFR exp. 10/15/14
ABSOLUTE STANDARD

exp 10/15/14

VF 5/1/12

Part #: 10005 Laboratory Use Only - See MSDS
 Lot #: 121208 Exp: 121213 Storage 4 °C
 **CLP Semi-Volatiles Toxic Substances #2**
 8 components CLP Semi-Volatiles Toxic Substances #2
 2000 ug/mL in meth Lot #: 121208 - 29097
 Rec: 8/4/11 MFR exp 12/12/13
ABSOLUTE STANDARD

exp 12/12/13

VF 5/1/12

Part #: 10006 Laboratory Use Only - See MSDS
 Lot #: 071211 Exp: 071214 Storage 4 °C
 **CLP Semi-Volatiles - Benzidines**
 2 components CLP Semi-Volatiles - Benzidines
 2000 ug/mL in meth Lot #: 071211 - 29102
 Rec: 8/4/11 MFR exp. 07/12/14
ABSOLUTE STANDARD

exp 7/12/14

VF 5/1/12

Part #: 10007 Laboratory Use Only - See MSDS
 Lot #: 100909 Exp: 100914 Storage 4 °C
 **CLP Semi-Volatiles - PAH Standard**
 17 components CLP Semi-Volatiles - PAH Mix
 2000 ug/mL in meth Lot #: 100909 - 29107
 Rec: 8/4/11 MFR exp. 10/09/14
ABSOLUTE STANDARD

exp 10/09/14

VF 5/1/12

Part #: 10018 Laboratory Use Only - See MSDS
 Lot #: 062111 Exp: 062116 Storage 4 °C
 **EPA Method 8270A - Analytes Mix #8**
 13 components - PEPA Method 8270A - Analytes Mix #8
 2000 ug/mL in meth Lot #: 062111 - 29112
 Rec: 8/4/11 MFR exp. 06/21/16
ABSOLUTE STANDARD

exp 6/21/16

VF 5/1/12

Part #: 70023 Laboratory Use Only - See MSDS
 Lot #: 031611 Exp: 031616 Storage 4 °C
 **Atrazine**
 Atrazine
 1000 ug/mL in ac Lot #: 031611 - 29117 72
 Rec: 8/4/11 MFR exp. 03/16/16
ABSOLUTE STANDARD

exp 3/16/16

1F5/11/12

Part #: **82705** Laboratory Use Only - See MSDS
 Lot #: **041911** Exp: **041914** Storage **4 °C**
 **EPA Method 8270A** **EPA Method 82/UA - Mix #18
4 components Lot #: 041911 - 29122
2000 ug/mL in acet Rec: 8/4/11 MFR exp. 04/19/14
ABSOLUTE STANDARD

exp 4/19/14

1F5/11/12

Part #: **94552** Laboratory Use Only - See MSDS
 Lot #: **030411** Exp: **030414** Storage **4 °C**
 **Semi-Volatile Standard**
11 components Semi-Volatile Standard
Varied ug/mL in n Lot #: 030411 - 29127
 Rec: 8/4/11 MFR exp. 03/04/14
ABSOLUTE STANDARD

exp 3/4/14

1F5/11/12

PREP DATE:	05-01-12														
8270C Stock/Spike Standard															
Exp:	07-31-12														
		Conc.		Date	CODE:	P									
Supplier	ID #	µg/mL	Lot #	Code	Exp. Date	µL									
Absolute	10001	2000	042910-29081	05/01/12	04-29-13	1000									
Absolute	10002	2000	073109-29086	05/01/12	07-31-12	1000									
Absolute	10004	2000	101509-29091	05/01/12	10-15-14	1000									
Absolute	10005	2000	121208-29097	05/01/12	12-12-13	1000									
Absolute	10006	2000	071211-29102	05/01/12	07-12-14	1000									
Absolute	10007	2000	100909-29107	05/01/12	10-09-14	1000									
Absolute	10018	2000	062111-29112	05/01/12	06-21-16	1000									
Absolute	70023	1000	031611-29117	05/01/12	03-16-16	1000									
Absolute	82705	2000	041911-29122	05/01/12	04-19-14	1000									
Absolute	94552	2000	030411-29127	05/01/12	03-14-14	1000									
					Final Vol	10000									

1F5/4/12

PREP DATE:	05-04-12														
8270 STANDARD CURVE															
		Conc.		Date		5	10	20	40	50	60	80	100		
Supplier	ID #	µg/mL	Lot #	Code	Exp. Date	µL	µL	µL							
	8270T Stock	200		05/01/12	07-31-12	5	5	10	20	25	30	40	50		
	Surrogate Stock	VAR	167802-29314	02/25/12	01-09-13	5	5	10	20	25	30	40	50		
EM Science	Methylene Chloride		47186			190	90	80	60	50	40	20	0		
					Final Vol.	200	100	100	100	100	100	100	100		

1F5/4/12

PREP DATE:	05-04-12														
8270 Second Source (SS) 50ug/mL															
		Conc.		Date	CODE:										
Supplier	ID #	µg/mL	Lot #	Code	Exp. Date	µL									
	8270C SS	200		10/11/11	10-11-12	25									
EM science	Methylene Chloride		47186			75									
					Final Vol.	100									

1F5/11/12

Part #: **10001** Laboratory Use Only - See MSDS
 Lot #: **042910** Exp: **042913** Storage **0 °C**
 **CLP Semi-Volatiles Base/Neutrals Mix #1**
14 components CLP Semi-Volatiles Base/Neutrals Mix #1
2000 ug/mL in me Lot #: 042910 - 29082
 Rec: 8/4/11 MFR exp. 04/29/13
ABSOLUTE STANDARD

exp 4/29/13

1F5/11/12

Part #: **10002** Laboratory Use Only - See MSDS
 Lot #: **073109** Exp: **073112** Storage **4 °C**
 **CLP Semi-Volatiles Base/Neutrals Mix #2**
14 components CLP Semi-Volatiles Base Neutrals Mix #2
2000 ug/mL in me Lot #: 073109 - 29087
 Rec: 8/4/11 MFR exp. 07/31/12
ABSOLUTE STANDARD

exp 7/31/12

Organic Extraction Worksheet

Method	SIM Separatory Funnel Extra 3510C	Extraction Set	120723A	Extraction Method	SEP004S	Units	mL
Spiked ID 1	SIM Spike 170745-30363	Surrogate ID 1	8270 SIM Surrogate 188684-30653				
Spiked ID 2		Surrogate ID 2					
Spiked ID 3		Surrogate ID 3					
Spiked ID 4		Surrogate ID 4					
Spiked ID 5		Surrogate ID 5					
Spiked ID 6		Sufficient Vol for Matrix QC:		YES			
Spiked ID 7		Ext. Start Time:		07/23/12 16:30			
Spiked ID 8		Ext. End Time:		07/24/12 15:13			
		GC Requires Extract By:		08/01/12 0:00			
		pH1	2	07/23/12 4:45:00 PM	Water Bath Temp Criteria		78,80,78 °
		pH2	14	7/24/12 10:55:00 AM			
		pH3					

Spiked By: DL

Date 07/23/12

Witnessed By: JM

Date 07/23/12

Sample	Sample Container	Spike Amount	Spike ID	Surrogate Amount	Surrogate ID	Extract Amount	Final Volume	pH	Extract Date/Time	Comments
1	120723A BIK			0.025	1	1000	1	2/1	07/23/12 16:30	
						equip E-WB7,78				
2	120723A LCS-1	0.025	1	0.025	1	1000	1	2/1	07/23/12 16:30	
						equip E-WB7,78				
3	AY65041 AY65041W07			0.025	1	1050	1	2/1	07/23/12 16:30	68248-2 WEEK RUSH -- Amber Liter
						equip E-WB7,78				
4	AY65043 AY65043W05			0.025	1	1060	1	2/1	07/23/12 16:30	68248-2 WEEK RUSH -- Amber Liter
						equip E-WB7,78				
5	AY65044 AY65044W04			0.025	1	1060	1	2/1	07/23/12 16:30	68248-2 WEEK RUSH -- Amber Liter
						equip E-WB7,78				
6	AY65112 AY65112W07			0.025	1	1030	1	2/1	07/23/12 16:30	68258-2 WEEK RUSH -- Amber Liter
						equip E-WB7,78				
7	AY65113 AY65113W06			0.025	1	1050	1	2/1	07/23/12 16:30	68258-2 WEEK RUSH -- Amber Liter
						equip E-WB7,78				
8	AY65144 MS-1 AY65144W09	0.025	1	0.025	1	1050	1	2/1	07/23/12 16:30	68266-2 WEEK RUSH -- Amber Liter
						equip E-WB7,78				
9	AY65144 MSD-1 AY65144W10	0.025	1	0.025	1	1060	1	2/1	07/23/12 16:30	68266-2 WEEK RUSH -- Amber Liter
						equip E-WB7,78				
10	AY65144 AY65144W12			0.025	1	1040	1	2/1	07/23/12 16:30	68266-2 WEEK RUSH -- Amber Liter
						equip E-WB6,80				
11	AY65145 AY65145W03			0.025	1	1060	1	2/1	07/23/12 16:30	68266-2 WEEK RUSH -- Amber Liter
						equip E-WB6,80				
12	AY65146 AY65146W07			0.025	1	1050	1	2/1	07/23/12 16:30	68266-2 WEEK RUSH -- Amber Liter
						equip E-WB6,80				
13	AY65147 AY65147W05			0.025	1	1040	1	2/1	07/23/12 16:30	68266-2 WEEK RUSH -- Amber Liter
						equip E-WB6,80				

Solvent and Lot#	
MC	EMD52104
Na2SO4	2351C512
10N NaOH	07/06/12
I+1 Acid	06/27/12
A. Na2SO4	06/28/12

Extraction COC Transfer	
Extraction lab employee Initials	DRA
GC analyst's initials	LE
Date	7/24/12
Time	12:00
Refrigerator	10007

Technician's Initials	
Scanned By	JM
Sample Preparation	JM
Extraction	JM/DL/GH
Concentration	IC
Modified	07/24/12 3:52:10 PM

Reviewed By: DRA 74 Date 07/24/12

Organic Extraction Worksheet

Method	SIM Separatory Funnel Extra 3510C	Extraction Set	120723A	Extraction Method	SEP004S	Units	mL	
Spiked ID 1	SIM Spike 170745-30363	Surrogate ID 1	8270 SIM Surrogate 188684-30653					
Spiked ID 2		Surrogate ID 2						
Spiked ID 3		Surrogate ID 3						
Spiked ID 4		Surrogate ID 4						
Spiked ID 5		Surrogate ID 5						
Spiked ID 6		Sufficient Vol for Matrix QC:		YES				
Spiked ID 7		Ext. Start Time:		07/23/12 16:30				
Spiked ID 8		Ext. End Time:		07/24/12 15:13				
		GC Requires Extract By:		08/01/12 0:00				
		pH1	2	07/23/12 4:45:00 PM	Water Bath Temp Criteria			78,80,78 °
		pH2	14	7/24/12 10:55:00 AM				
		pH3						

Spiked By: DL

Date 07/23/12

Witnessed By: JM

Date 07/23/12

Sample	Sample Container	Spike Amount	Spike ID	Surrogate Amount	Surrogate ID	Extract Amount	Final Volume	pH	Extract Date/Time	Comments
14	AY65148 			0.025	1	1050 equip E-WB5,78	1	2/1	07/23/12 16:30	68266-2 WEEK RUSH -- Amber Liter
15	AY65149 			0.025	1	1060 equip E-WB5,78	1	2/1	07/23/12 16:30	68266-2 WEEK RUSH -- Amber Liter
16	AY65150 			0.025	1	1060 equip E-WB5,78	1	2/1	07/23/12 16:30	68266-2 WEEK RUSH -- Amber Liter
17	AY65151 			0.025	1	1060 equip E-WB5,78	1	2/1	07/23/12 16:30	68266-2 WEEK RUSH -- Amber Liter

DRA 7/24/12

Solvent and Lot#	
MC	EMD52104
Na2SO4	2351CS12
10N NaOH	07/06/12
I+1 Acid	06/27/12
A. Na2SO4	06/28/12

Extraction COC Transfer	
Extraction lab employee Initials	DRA
GC analyst's initials	<i>JK</i>
Date	7/24/12
Time	17:00
Refrigerator	<i>Hobart</i>

Technician's Initials	
Scanned By	JM
Sample Preparation	JM
Extraction	JM/DL/GH
Concentration	IC
Modified	07/24/12 3:52:10 PM

Reviewed By: DRA 75 Date 07/24/12

Injection Log

Directory: M:\LINUS\DATA\L120613\

Line	Vial	FileName	Multiplier	SampleName	Misc Info	Injected
1	1	0613L001.D	1	SVTUNE 2-28-12		13 Jun 12 13:07
2	3	0613L003.D	1	0.1ug/ml PAH 06-13-12		13 Jun 12 13:51
3	4	0613L004.D	1	0.2ug/ml PAH		13 Jun 12 14:16
4	5	0613L005.D	1	0.5ug/ml PAH		13 Jun 12 14:41
5	6	0613L006.D	1	1.0ug/ml PAH		13 Jun 12 15:07
6	7	0613L007.D	1	5.0ug/ml PAH		13 Jun 12 15:33
7	8	0613L008.D	1	10ug/ml PAH		13 Jun 12 15:59
8	9	0613L009.D	1	50ug/ml PAH		13 Jun 12 16:25
9	10	0613L010.D	1	100ug/ml PAH		13 Jun 12 16:51
10	11	0613L011.D	1	5.0ug/ml SS PAH 06-13-12		13 Jun 12 17:17
11	1	0724L001.D	1	SVTUNE 2-28-12		24 Jul 12 18:05
12	2	0724L002.D	1	5.0ug/ml PAH 06-13-12		24 Jul 12 18:24
13	3	0724L003.D	1	120723A BLK 1/1000		24 Jul 12 18:50
14	4	0724L004.D	1	120723A LCS-1 1/1000		24 Jul 12 19:16
15	8	0724L008.D	0.97087	AY65112W07 1/1030		24 Jul 12 21:00
16	9	0724L009.D	0.95238	AY65113W06 1/1050		24 Jul 12 21:26

**EPA 8015B
Total Petroleum Hydrocarbons**

**EPA 8015B
Total Petroleum Hydrocarbons -
QC Summary**

Method Blank

TPH Diesel Water

Blank Name/QCG: **120723W-65041 - 169578**
Batch ID: #TPETD-120723A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	DIESEL FUEL	80.8 U	150	80.8	40.4	ug/L	07/23/12	07/31/12
BLANK	SURROGATE: OCTACOSANE (S)	40.6	28-142			%	07/23/12	07/31/12
BLANK	SURROGATE: ORTHO-TERPHEN	48.6 #	57-132			%	07/23/12	07/31/12

= Recovery (or RPD) is outside QC limits.

Quant Method: TPH0719.M
Run #: 731013
Instrument: Apollo
Sequence: 120731
Initials: SD

Printed: 08/02/12 6:05:52 PM
GC SC-Blank-REG MDLs

Surrogate Recovery

Lab Name: APPL, Inc.
 Case No: 68258
 Matrix: WATER

SDG No: 68258
 Date Analyzed: 07/31/12
 Instrument: Apollo

APPL ID.	Client Sample No.	SURROGATE: OCTACOSANE (S)			SURROGATE: ORTHO-TERPHENYL (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
120723A-BLK	Blank	28-142	40.6		57-132	48.6	#
120723A-LCS	Lab Control Spike	28-142	58.5		57-132	91.3	
AY65112	ES081	28-142	58.9		57-132	70.6	
AY65113	ES082	28-142	54.6		57-132	77.5	

Comments: Batch: #TPETD-120723A
 # = Recovery outside of Control Limits on Sample.

Laboratory Control Spike Recovery

TPH Diesel Water

APPL ID: 120723W-65041 LCS - 169578

Batch ID: #TPETD-120723A

APPL Inc.

908 North Temperance Avenue

Clovis, CA 93611

Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits
DIESEL FUEL	2000	1440	72.0	61-143
SURROGATE: OCTACOSANE (S)	150	87.7	58.5	28-142
SURROGATE: ORTHO-TERPHENYL (S)	150	137	91.3	57-132

Comments: _____

<u>Primary</u>	<u>SPK</u>
Quant Method :	TPH0719.M
Extraction Date :	07/23/12
Analysis Date :	07/31/12
Instrument :	Apollo
Run :	731014
Initials :	SD

Printed: 08/02/12 6:05:47 PM

APPL Standard LCS

EPA 8015B-e

Form 4

Blank Summary

Lab Name: APPL, Inc.

SDG No: 68258

Case No: 68258

Date Analyzed: 07/31/12

Matrix: WATER

Instrument: Apollo

Blank ID: 120723A-BLK

Time Analyzed: 1439

APPL ID.	Client Sample No.	File ID.	Date Analyzed
120723A-BLK	Blank	731013	07/31/12 1439
120723A-LCS	Lab Control Spike	731014	07/31/12 1503
AY65112	ES081	731019	07/31/12 1703
AY65113	ES082	731022	07/31/12 1816

Comments: Batch: #TPETD-120723A

**EPA 8015B
Total Petroleum Hydrocarbons -
Sample Data**

TPH Diesel Water

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

Attn: Max Solmssen
Project: LTM Red Hill / 1022-024

Sample ID: ES081

Sample Collection Date: 07/18/12

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 68258

APPL ID: AY65112

QCG: #TPETD-120723A-169578

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8015B-	DIESEL FUEL	80.8 U	150	80.8	40.4	ug/L	07/23/12	07/31/12
EPA 8015B-	SURROGATE: OCTACOSANE (S)	58.9	28-142			%	07/23/12	07/31/12
EPA 8015B-	SURROGATE: ORTHO-TERPHENYL (S)	70.6	57-132			%	07/23/12	07/31/12

Quant Method: TPH0719.M
Run #: 731019
Instrument: Apollo
Sequence: 120731
Dilution Factor: 1
Initials: SD

Printed: 08/02/12 6:05:49 PM
APPL-F1-SC-NoMC-REG MDLs

Data File : G:\APOLLO\DATA\120731\731019.D Vial: 19
 Acq On : 7-31-12 17:03:40 Operator: LAC
 Sample : AY65112W04 5/1030 Inst : Apollo
 Misc : Water Multiplr: 4.85
 IntFile : events.e
 Quant Time: Aug 1 16:32 2012 Quant Results File: TPH0719.RES

Method : G:\APOLLO\DATA\120731\TPH0719.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Thu Aug 02 17:43:25 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A



Compound	R.T.	Response	Conc Units
System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.80	29825661	102.736 ppb
Surrogate Spike 145.631		Recovery =	70.55%
6) SC Octacosane(S)	11.37	26606823	85.697 ppb
Surrogate Spike 145.631		Recovery =	58.85%

Target Compounds

Quantitation Report

Data File: G:\APOLLO\DATA\120731\731019.D

Sample : AY65112W04 5/1030



TPH Diesel Water

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

Attn: Max Solmssen
Project: LTM Red Hill / 1022-024

Sample ID: ES082
Sample Collection Date: 07/18/12

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 68258
APPL ID: AY65113
QCG: #TPETD-120723A-169578

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8015B-	DIESEL FUEL	1700 ++	150	80.8	40.4	ug/L	07/23/12	07/31/12
EPA 8015B-	SURROGATE: OCTACOSANE (S)	54.6	28-142			%	07/23/12	07/31/12
EPA 8015B-	SURROGATE: ORTHO-TERPHENYL (S)	77.5	57-132			%	07/23/12	07/31/12

++(T2M) The analyst has noted that the chromatogram of this sample is mainly lower boiling hydrocarbons.

Quant Method: TPH0719.M
Run #: 731022
Instrument: Apollo
Sequence: 120731
Dilution Factor: 1
Initials: SD

Printed: 08/02/12 6:05:50 PM
APPL-F1-SC-NoMC-REG MDLs

Data File : G:\APOLLO\DATA\120731\731022.D Vial: 22
 Acq On : 7-31-12 18:16:17 Operator: LAC
 Sample : AY65113W05 5/1030 Inst : Apollo
 Misc : Water Multiplr: 4.85
 IntFile : events.e
 Quant Time: Aug 1 16:38 2012 Quant Results File: TPH0719.RES

Method : G:\APOLLO\DATA\120731\TPH0719.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Thu Aug 02 17:43:25 2012
 Response via : Multiple Level Calibration




Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units
System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.80	32771315	112.882 ppb
Surrogate Spike 145.631		Recovery =	77.51%
6) SC Octacosane(S)	11.36	24693342	79.534 ppb
Surrogate Spike 145.631		Recovery =	54.61%
Target Compounds			
1) HATM Diesel (C10-C28)	8.60	377168158	1666.008 ppb

Quantitation Report

Data File: G:\APOLLO\DATA\120731\731022.D

Sample : AY65113W05 5/1030



**EPA 8015B
Total Petroleum Hydrocarbons -
Calibration Data**

TPH Extractables
TPH0719

Form 6
Initial Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: _____

SDG No: 68258
Initial Cal. Date: 06/22/2012 and 7/19/12
Instrument: Apollo Initials: sd

Surrogate	.	622004.D	622005.D	622006.D	622007.D	622008.D
DRO	622009.D	622010.D	622011.D	622012.D	622013.D	622014.D
MO	719003.D	719004.D	719005.D	719006.D	719007.D	719008.D

	Compound	1	2	3	4	5	6		Avg	%RSD	
1	HATM Diesel (C10-C28)	642703	509920	531557	542684	530047	540036		549491	8.6	HATM
2	HBTM Motor Oil (C18-C36)	415224	409753	447761	467949	423444	430885		432503	5.1	HBTM
3	SC Ortho-Terphenyl(S)	.	700048	705066	717492	699409	701217		704646	1.1	SC
4	SC Octacosane(S)	.	754341	750395	766254	747028	749884		753580	1.0	SC
5											
6											
7											
8											
9											
10											
11											
12											
13											
14											
15											
16											
17											
18											
19											
20											
21											
22											
23											
24											
25											
26											
27											
28											
29											
30											
31											
32											
33											

* Not Used

0.475552

Data File : G:\APOLLO\DATA\120622\622004.D Vial: 4
 Acq On : 6-22-12 18:22:29 Operator: LAC
 Sample : TCH SURROGATE 100/1000 Inst : Apollo
 Misc : Mix(c) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:37 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A



Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.84	7000476	2.493 ppb
Surrogate Spike 30.000		Recovery =	8.31%
6) SC Octacosane(S)	11.46	7543411	3.161 ppb
Surrogate Spike 30.000		Recovery =	10.54%
Target Compounds			

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622004.D

Sample : TCH SURROGATE 100/1000



Data File : G:\APOLLO\DATA\120622\622005.D Vial: 5
 Acq On : 6-22-12 18:46:55 Operator: LAC
 Sample : TCH SURROGATE 400/1000 Inst : Apollo
 Misc : Mix(c) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:37 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units



System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.84	28202647	10.113 ppb
Surrogate Spike 30.000		Recovery =	33.71%
6) SC Octacosane(S)	11.47	30015782	12.394 ppb
Surrogate Spike 30.000		Recovery =	41.31%

Target Compounds

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622005.D

Sample : TCH SURROGATE 400/1000



Data File : G:\APOLLO\DATA\120622\622006.D Vial: 6
 Acq On : 6-22-12 19:10:46 Operator: LAC
 Sample : TCH SURROGATE 600/1000 Inst : Apollo
 Misc : Mix(c) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:38 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units



System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.84	43049549	15.420 ppb
Surrogate Spike 30.000		Recovery =	51.40%
6) SC Octacosane(S)	11.48	45975259	18.583 ppb
Surrogate Spike 30.000		Recovery =	61.94%

Target Compounds

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622006.D

Sample : TCH SURROGATE 600/1000



Data File : G:\APOLLO\DATA\120622\622007.D Vial: 7
 Acq On : 6-22-12 19:34:47 Operator: LAC
 Sample : TCH SURROGATE 800/1000 Inst : Apollo
 Misc : Mix(c) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:38 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.85	55952695	19.926 ppb
Surrogate Spike 30.000		Recovery =	66.42%
6) SC Octacosane(S)	11.48	59762243	23.528 ppb
Surrogate Spike 30.000		Recovery =	78.43%



Target Compounds

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622007.D

Sample : TCH SURROGATE 800/1000

622007.D\FID1B



Data File : G:\APOLLO\DATA\120622\622008.D Vial: 8
 Acq On : 6-22-12 19:58:49 Operator: LAC
 Sample : TCH SURROGATE 1000/1000 Inst : Apollo
 Misc : Mix(c) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:39 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units



System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.85	70121711	24.864 ppb
Surrogate Spike 30.000		Recovery =	82.88%
6) SC Octacosane(S)	11.48	74988351	28.844 ppb
Surrogate Spike 30.000		Recovery =	96.15%

Target Compounds

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622008.D

Sample : TCH SURROGATE 1000/1000



Data File : G:\APOLLO\DATA\120622\622009.D Vial: 9
 Acq On : 6-22-12 20:22:56 Operator: LAC
 Sample : DIESEL 10/1000 6/22/12 Inst : Apollo
 Misc : Mix(A) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:08 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units




System Monitoring Compounds			
3) SA Not Used(S)	8.85	1100828	0.688 ppb
Surrogate Spike 30.000		Recovery =	2.29%
5) SA Not Used2(S)	11.46	755848	0.635 ppb
Surrogate Spike 30.000		Recovery =	2.12%
Target Compounds			
1) HATM Diesel (C10-C28)	8.60	12854065	11.749 ppb

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622009.D

Sample : DIESEL 10/1000 6/22/12

622009.D\FID1B



Data File : G:\APOLLO\DATA\120622\622010.D Vial: 10
 Acq On : 6-22-12 20:47:06 Operator: LAC
 Sample : DIESEL 100/1000 Inst : Apollo
 Misc : Mix(A) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:05 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A




Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
3) SA Not Used(S)	8.84	8996588	5.622 ppb
Surrogate Spike 30.000		Recovery =	18.74%
5) SA Not Used2(S)	11.46	7054012	5.925 ppb
Surrogate Spike 30.000		Recovery =	19.75%
Target Compounds			
1) HATM Diesel (C10-C28)	8.60	101984030	93.220 ppb

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622010.D

Sample : DIESEL 100/1000



Data File : G:\APOLLO\DATA\120622\622011.D Vial: 11
 Acq On : 6-22-12 21:11:13 Operator: LAC
 Sample : DIESEL 400/1000 Inst : Apollo
 Misc : Mix(A) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:05 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units



System Monitoring Compounds			
3) SA Not Used(S)	8.84	31783742	19.863 ppb
Surrogate Spike 30.000		Recovery =	66.21%
5) SA Not Used2(S)	11.47	28563798	23.990 ppb
Surrogate Spike 30.000		Recovery =	79.97%
Target Compounds			
1) HATM Diesel (C10-C28)	8.60	425245865	388.700 ppb

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622011.D

Sample : DIESEL 400/1000

622011.D\FID1B



Data File : G:\APOLLO\DATA\120622\622012.D Vial: 12
 Acq On : 6-22-12 21:35:18 Operator: LAC
 Sample : DIESEL 600/1000 Inst : Apollo
 Misc : Mix(A) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:05 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units



System Monitoring Compounds			
3) SA Not Used(S)	8.84	48229746	30.140 ppb
Surrogate Spike 30.000		Recovery =	100.47%
5) SA Not Used2(S)	11.47	43434321	36.480 ppb
Surrogate Spike 30.000		Recovery =	121.60%
Target Compounds			
1) HATM Diesel (C10-C28)	8.60	651220989	595.255 ppb

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622012.D

Sample : DIESEL 600/1000

622012.D\FID1B



Data File : G:\APOLLO\DATA\120622\622013.D Vial: 13
 Acq On : 6-22-12 21:59:20 Operator: LAC
 Sample : DIESEL 800/1000 Inst : Apollo
 Misc : Mix(A) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:05 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A



Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
3) SA Not Used(S)	8.85	76202842	47.622 ppb
Surrogate Spike 30.000		Recovery =	158.74%
5) SA Not Used2(S)	11.48	57498014	48.292 ppb
Surrogate Spike 30.000		Recovery =	160.97%
Target Compounds			
1) HATM Diesel (C10-C28)	8.60	848074829	775.192 ppb

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622013.D
Sample : DIESEL 800/1000

622013.D\FID1B



Data File : G:\APOLLO\DATA\120622\622014.D Vial: 14
 Acq On : 6-22-12 22:23:21 Operator: LAC
 Sample : DIESEL 1000/1000 Inst : Apollo
 Misc : Mix(A) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:05 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A



Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
3) SA Not Used(S)	8.85	80954970	50.591 ppb
Surrogate Spike 30.000		Recovery =	168.64%
5) SA Not Used2(S)	11.48	71709415	60.228 ppb
Surrogate Spike 30.000		Recovery =	200.76%
Target Compounds			
1) HATM Diesel (C10-C28)	8.60	1080072891	987.252 ppb

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622014.D
Sample : DIESEL 1000/1000

622014.D\FID1B



TPH Extractables
TPH622

Form 7
Second Source

Lab Name: APPL, Inc.
Case No: _____
Matrix: _____

SDG No: 68258
Date Analyzed: 06/22/12
Instrument: Apollo
Initial Cal. Date: 06/22/12
Data File: 622015.D

	Compound	MEAN	CCRF	%D	%Drift
1	HATM Diesel (C10-C28)	549491	516614	6.0	HATM
2					
3					
4					
5					
6					
7					
8					
9					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					
23					
24					
25					
26					
27					
28					
29					
30					
31					
32					
33					
34					
35					
36					
37					
38					
39					
40	Average			6.0	

Data File : G:\APOLLO\DATA\120622\622015.D Vial: 15
 Acq On : 6-22-12 22:47:20 Operator: LAC
 Sample : DIESEL 2ND SRC 6/22/12 Inst : Apollo
 Misc : Mix(A) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 25 9:28 2012 Quant Results File: TPH0622.RES

Method : G:\APOLLO\DATA\120622\TPH0622.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Mon Jun 25 09:48:29 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

Target Compounds



1) HATM Diesel (C10-C28)	8.60	413291584	376.067 ppb
--------------------------	------	-----------	-------------

Quantitation Report

Data File: G:\APOLLO\DATA\120622\622015.D

Sample : DIESEL 2ND SRC 6/22/12

622015.D\FID1B



TPH Extractables
TPH0719

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: _____

SDG No: 68258
Date Analyzed: 07/31/12
Instrument: Apollo
Initial Cal. Date: 07/31/12
Data File: 731002.D, 003.d

		Compound	MEAN	CCRF	%D	%Drift
1	HATM	Diesel (C10-C28)	549491	494592	10	HATM
2	HBTM	Motor Oil (C18-C36)	432503	381666	12	HBTM
3						
4						
5						
6						
7						
8						
9						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						
23						
24						
25						
26						
27						
28						
29						
30						
31						
32						
33						
34						
35						
36						
37						
38						
39						
40		Average			11.0	

Data File : G:\APOLLO\DATA\120731\731002.D Vial: 2
 Acq On : 7-31-12 10:15:07 Operator: LAC
 Sample : DIESEL 400ppm 7/31/12 Inst : Apollo
 Misc : Mix(A) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Aug 2 17:32 2012 Quant Results File: TPH0719.RES

Method : G:\APOLLO\DATA\120731\TPH0719.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Thu Aug 02 17:43:25 2012
 Response via : Multiple Level Calibration



Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units
System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.80	34641855	24.581 ppb
Surrogate Spike 30.000		Recovery =	81.94%
6) SC Octacosane(S)	11.37	23218499	15.405 ppb
Surrogate Spike 30.000		Recovery =	51.35%
Target Compounds			
1) HATM Diesel (C10-C28)	8.60	395673584	360.036 ppb

Quantitation Report

Data File: G:\APOLLO\DATA\120731\731002.D

Sample : DIESEL 400ppm 7/31/12



TPH Extractables
TPH0719

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: _____

SDG No: 68258
Date Analyzed: 07/31/12
Instrument: Apollo
Initial Cal. Date: 07/31/12
Data File: 731020.D, 021.d

		Compound	MEAN	CCRF	%D	%Drift
1	HATM	Diesel (C10-C28)	549491	509123	7.3	HATM
2	HBTM	Motor Oil (C18-C36)	432503	394077	8.9	HBTM
3						
4						
5						
6						
7						
8						
9						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						
23						
24						
25						
26						
27						
28						
29						
30						
31						
32						
33						
34						
35						
36						
37						
38						
39						
40		Average			8.1	

Data File : G:\APOLLO\DATA\120731\731020.D Vial: 20
 Acq On : 7-31-12 17:28:05 Operator: LAC
 Sample : DIESEL 400ppm 7/30/12 Inst : Apollo
 Misc : Mix(A) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Aug 2 17:33 2012 Quant Results File: TPH0719.RES

Method : G:\APOLLO\DATA\120731\TPH0719.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Thu Aug 02 17:43:25 2012
 Response via : Multiple Level Calibration



Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units
System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.80	38819888	27.546 ppb
Surrogate Spike 30.000		Recovery =	91.82%
6) SC Octacosane(S)	11.36	23947613	15.889 ppb
Surrogate Spike 30.000		Recovery =	52.96%
Target Compounds			
1) HATM Diesel (C10-C28)	8.60	407298733	370.614 ppb

Quantitation Report

Data File: G:\APOLLO\DATA\120731\731020.D

Sample : DIESEL 400ppm 7/30/12



TPH Extractables
TPH0719

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: _____

SDG No: 68258
Date Analyzed: 07/31/12
Instrument: Apollo
Initial Cal. Date: 07/31/12
Data File: 731032.D, 033.d

		Compound	MEAN	CCRF	%D	%Drift
1	HATM	Diesel (C10-C28)	549491	522051	5.0	HATM
2	HBTM	Motor Oil (C18-C36)	432503	356513	18	HBTM
3						
4						
5						
6						
7						
8						
9						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						
23						
24						
25						
26						
27						
28						
29						
30						
31						
32						
33						
34						
35						
36						
37						
38						
39						
40						

Average

11.5

Data File : G:\APOLLO\DATA\120731\731032.D Vial: 32
 Acq On : 7-31-12 22:20:07 Operator: LAC
 Sample : DIESEL 400ppm 7/30/12 Inst : Apollo
 Misc : Mix(A) Multiplr: 1.00
 IntFile : events.e
 Quant Time: Aug 2 17:35 2012 Quant Results File: TPH0719.RES

Method : G:\APOLLO\DATA\120731\TPH0719.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Thu Aug 02 17:43:25 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A



Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.80	36556809	25.940 ppb
Surrogate Spike 30.000		Recovery =	86.47%
6) SC Octacosane(S)	11.36	23773019	15.773 ppb
Surrogate Spike 30.000		Recovery =	52.58%
Target Compounds			
1) HATM Diesel (C10-C28)	8.60	417641191	380.025 ppb

Quantitation Report

Data File: G:\APOLLO\DATA\120731\731032.D

Sample : DIESEL 400ppm 7/30/12



**EPA 8015B
Total Petroleum Hydrocarbons -
Raw Data**

Method Blank

TPH Diesel Water

Blank Name/QCG: **120723W-65041 - 169578**
Batch ID: #TPETD-120723A

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	DIESEL FUEL	80.8 U	150	80.8	40.4	ug/L	07/23/12	07/31/12
BLANK	SURROGATE: OCTACOSANE (S)	40.6	28-142			%	07/23/12	07/31/12
BLANK	SURROGATE: ORTHO-TERPHEN	48.6 #	57-132			%	07/23/12	07/31/12

= Recovery (or RPD) is outside QC limits.

Quant Method: TPH0719.M
Run #: 731013
Instrument: Apollo
Sequence: 120731
Initials: SD

Printed: 08/02/12 6:05:52 PM
GC SC-Blank-REG MDLs

Data File : G:\APOLLO\DATA\120731\731013.D Vial: 13
 Acq On : 7-31-12 14:39:54 Operator: LAC
 Sample : 120723A BLK 5/1000 Inst : Apollo
 Misc : Water Multiplr: 5.00
 IntFile : events.e
 Quant Time: Aug 1 15:55 2012 Quant Results File: TPH0719.RES

Method : G:\APOLLO\DATA\120731\TPH0719.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Thu Aug 02 17:43:25 2012
 Response via : Multiple Level Calibration

Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units



System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.81	20557348	72.935 ppb
Surrogate Spike 150.000		Recovery =	48.62%
6) SC Octacosane(S)	11.36	18363331	60.920 ppb
Surrogate Spike 150.000		Recovery =	40.61%

Target Compounds

Quantitation Report

Data File: G:\APOLLO\DATA\120731\731013.D

Sample : 120723A BLK 5/1000



Laboratory Control Spike Recovery

TPH Diesel Water

APPL ID: 120723W-65041 LCS - 169578

Batch ID: #TPETD-120723A

APPL Inc.

908 North Temperance Avenue

Clovis, CA 93611

Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits
DIESEL FUEL	2000	1440	72.0	61-143
SURROGATE: OCTACOSANE (S)	150	87.7	58.5	28-142
SURROGATE: ORTHO-TERPHENYL (S)	150	137	91.3	57-132

Comments: _____

<u>Primary</u>	<u>SPK</u>
Quant Method :	TPH0719.M
Extraction Date :	07/23/12
Analysis Date :	07/31/12
Instrument :	Apollo
Run :	731014
Initials :	SD

Printed: 08/02/12 6:05:48 PM

APPL Standard LCS

Data File : G:\APOLLO\DATA\120731\731014.D Vial: 14
 Acq On : 7-31-12 15:03:52 Operator: LAC
 Sample : 120723A LCS-1 5/1000 Inst : Apollo
 Misc : Water Multiplr: 5.00
 IntFile : events.e
 Quant Time: Aug 1 15:55 2012 Quant Results File: TPH0719.RES

Method : G:\APOLLO\DATA\120731\TPH0719.M (Chemstation Integrator)
 Title : Diesel
 Last Update : Thu Aug 02 17:43:25 2012
 Response via : Multiple Level Calibration



Volume Inj. : 2UL
 Signal Phase : DB-5
 Signal Info : FID02A

Compound	R.T.	Response	Conc Units
System Monitoring Compounds			
4) SC Ortho-Terphenyl(S)	8.80	38496498	136.581 ppb
Surrogate Spike 150.000		Recovery =	91.05%
6) SC Octacosane(S)	11.36	26424152	87.662 ppb
Surrogate Spike 150.000		Recovery =	58.44%
Target Compounds			
1) HATM Diesel (C10-C28)	8.60	317227952	1443.280 ppb

Quantitation Report

Data File: G:\APOLLO\DATA\120731\731014.D

Sample : 120723A LCS-1 5/1000



STANDARD

CONC DATE ALIQUOT VOLUME CONC LOT# INITIALS

INITIAL SOURCE .FINAL FINAL SOL. EN. 005
CONC DATE ALIQUOT VOLUME CONC LOT# INITIALS

PCB Soil Spike

AR1016 1,000 mg/L 02SI 1250 mL 25 mL 50% Aceton CM

AR1260 Aroclor 1016 + 1260 Solution, 1,000 mg/L, 1 ml #022912B 6-21-12

130011-03
Lot # 163759 Storage < Ambient Expiry 9/14/13
Solvent: Hexane
Aroclor 1016 + 1260 op. 6-21-12
Lot #: 163759 - 29969 ex. 6-21-13
Rec: 11/10/11 MFR exp. 09/14/13
CM 6-21-12

AND LOT: 163759-29971

16355 op. 2-4-12

6-21-12 ex. 2-4-13

OCL Soil Surrogate

DECA 5,000 mg/L 02SI 1 mL 250 mL 20% Acetone CM

DBC

TCMX

Pesticide Surrogate Solution, 5,000 mg/L, 1 ml

02si

Cat. No: 130070-02

Lot No: 154164

Pesticide Surrogate Soln, 5000mg/L

Lot #: 154164 - 29418

Rec: 8/26/11 MFR exp. 12/19/12

CM 6-21-12

Exp: 12/19/2012

Storage: < Ambient

Solvent Tol.: Hex. 1:1

ption For Research Use Only

Opened: 6-21-12

ex. 6-21-13

DIESEL CAL STD.

STD.	INITIAL CONC.	SOURCE DATE	ALIQUOT	FINAL VOL.	FINAL CONC.	SOLVENT / LOT#
DIESEL FUEL #2	50,000 ug/mL	O2SI CAT#011598-03 LOT#183767-30909 OP:6/22/12 EXP:6/22/13	1mL		1000ug/mL	MC LOT# 51306
O-TERPHENYL OCTACOSANE	600 ug/mL	O2SI CAT#110316-05 LOT#183766-30661 OP:3/5/12EXP:3/5/13	4160 µL	50mL	50ug/mL	

Diesel Fuel #2 Composite,
50,000 mg/L, 1 ml
011593-03
Lot # 183767 Storage < -10 Degrees C Expiry 2/1/16
Solvent: Methylene Chloride

Diesel Fuel #2 Composite op. 6-22-12

Lot #: 183767 - 30909 ex. 6-22-13

Rec: 5/30/12 MFR exp. 02/11/16
CM 6-22-12

DIESEL SECOND SOURCE

STANDARD	INITIAL CONC.	SOURCE DATE	ALIQUOT	FINAL VOL.	FINAL CONC.	SOLVENT / LOT#
DIESEL FUEL #2	50,000 ug/mL	O2SI CAT#011598-03 LOT#167769-29398 OP:6/22/12 EXP:6/22/13	200uL	10mL	1000ug/mL	MC #51306

CM 6-22-12

006
STANDARD

INITIAL SOURCE FINAL FINAL SOLVENT DATE
CONC DATE ALIQUOT VOLUME CONC LOT# INITIALS

MOTOR OIL CAL STD						
STD.	INITIAL CONC.	SOURCE DATE	ALIQUOT	FINAL VOL	FINAL CONC.	SOLVENT / LOT#
MOTOR OIL	50,000 ug/mL		1mL	50mL	1000ug/mL	MC LOT# 51306

Motor Oil Composite
50,000 mg/L, 1 ml
116390-02
Lot# 183768 Storage Expiry
≤ -10 Degrees C 1/8/15
Solvent: Methylene Chloride
Motor oil composite op. 6-22-12
Lot #: 183768 - 30232 CA. 6-22-12 CM
Rec: 1/10/12 MFR exp. 01/08/15
CA 6-22-12

CM
6-22-12
ex. 12-22-12

THC SURR CAL STD						
STD.	INITIAL CONC.	SOURCE DATE	ALIQUOT	FINAL VOL	FINAL CONC.	SOLVENT LOT#
O-TERPHENYL OCTACOSANE	600 ug/mL	O2SI CAT#110316-05 LOT#183766-30661 OP:6/12/12 EXP:6/12/13	834 µL	10mL	50ug/mL	MC LOT# 51306

CM
6-22-12
ex. 12-22-12

TCH SURROGATE CURVE										
STD	[µg/mL]	LOT #	DATE	EXP. DATE	µL	µL	µL	µL	µL	µL
THC SURR	50		06/22/12	12/22/12	50	100	400	600	800	1000
MC		51306			950	900	600	400	200	NA
					Final VOL.	1000	1000	1,000	1000	1000

CM
6-22-12
ex. 12-22-12

DIESEL CURVE										
STD	[µg/mL]	LOT #	DATE	EXP. DATE	µL	µL	µL	µL	µL	µL
DIESEL	1000		06/22/12	12/22/12	10	100	400	600	800	1000
MC		51306			990	900	600	400	200	NA
					Final VOL.	1000	1000	1,000	1000	1000

MOTOR OIL CURVE										
STD	[µg/mL]	LOT #	DATE	EXP. DATE	µL	µL	µL	µL	µL	µL
MOTOR OIL	1000		06/22/12	12/22/12	50	100	400	600	800	1000
MC		51306			950	900	600	400	200	NA
					Final VOL.	1000	1000	1,000	1000	1000

DIESEL 2ND SOURCE						
STD	Init. Conc	Source	Aliquot	Final Vol.	Final Conc.	Solvent
DIESEL 2ND SRC	1000µg/ml	O2SI	400µL	1 mL	400 µg/mL	MC
		Prep:	06/22/12			51306
		Exp:	12/22/12			

CM 6-22-12

LCMS STANDARD PREP LOG# _____

039

STANDARD	INITIAL CONC	SOURCE DATE	ALIQOT VOLUME	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE/INITIALS
MOTOR OIL STD	2000 $\mu\text{g}/\text{mL}$	025E M.O. STD prep. 7-19-12	250 μL	1 mL	500 $\mu\text{g}/\text{mL}$	MC # 51306	7-31-12 ex. 1-19-13 CAL
DIESEL STD	1000 $\mu\text{g}/\text{mL}$	Diesel STD prep. 6-22-12	400 μL	1 mL	400 $\mu\text{g}/\text{mL}$	MC # 51306	7-31-12 ex. 1-19-13 CAL

OCL
Second
Source

OCL Second Source					
Compounds	Conc in mix	Conc in stock	Aliquot	Source stock	Final Vol
a-BHC	10 $\mu\text{g}/\text{mL}$	100 $\mu\text{g}/\text{mL}$	100 μL	OCL 2nd Src Stk	10 μL
b-BHC			Prep: 06/23/11	Hexane	
d-BHC			Exp: 06/23/12	# 0010090	
g-BHC			Prep: 7/30/12	082610B	
aldrin			12/12/12		
heptachlor					LH 8/3/12
heptachlor-epoxide isomer B					
a-chlordane					
g-chlordane					
pp-DDD					
pp-DDE					
pp-DDT					
dieldrin					
endrin					
endrin aldehyde					
endrin ketone					
endosulfan I					
endosulfan II					
endosulfan sulfate					
methoxychlor					

LH
8/1/12
exp: 12/12/12

LH 8/1/12

LH 8/1/12

OCL
Curve

OCL CALIBRATION CURVE					
Compound	Conc. In Mix	Conc. Of Stock	Aliquot	stock source	Final Vol.
Various	1A: 0.0025 $\mu\text{g}/\text{mL}$	10 $\mu\text{g}/\text{mL}$	2.5 μL	OCL Stock	10 μL
Analytes	1 - 0.005 $\mu\text{g}/\text{mL}$	10 $\mu\text{g}/\text{mL}$	5 μL	prep: 2/13/12	10 μL
	2 - 0.050 $\mu\text{g}/\text{mL}$	10 $\mu\text{g}/\text{mL}$	250 μL	exp: 1/2/12	50 μL
	3 - 0.100 $\mu\text{g}/\text{mL}$	10 $\mu\text{g}/\text{mL}$	500 μL	Prep: 7/30/12	50 μL
	4 - 0.150 $\mu\text{g}/\text{mL}$	10 $\mu\text{g}/\text{mL}$	375 μL	7/30/12	25 μL
	5 - 0.200 $\mu\text{g}/\text{mL}$	10 $\mu\text{g}/\text{mL}$	200 μL	LH 8/3/12	10 μL
	6 - 0.250 $\mu\text{g}/\text{mL}$	10 $\mu\text{g}/\text{mL}$	250 μL		10 μL
	1B - 0.001 $\mu\text{g}/\text{mL}$	0.005 $\mu\text{g}/\text{mL}$	1000 μL	Lvl 1	5 μL
				prep: 2/4/12	8/1/12
				exp: 8/4/12	2/1/13
Solvent:	Hexane	Lot: 040744A	LH 8/3/12		LH 8/3/12

LH
8/1/12
exp 2/1/13

LC/MS STANDARD PREP LOG#

020
STANDARD

INITIAL CONC	SOURCE DATE	ALLOQUOT	FINAL VOLUME	FINAL CONC	SOLVENT LOT#	DATE INITIALS
--------------	-------------	----------	--------------	------------	--------------	---------------

AR 1248 CALIBRATION CURVE

AR1248 Prep: 3/26/12
 Exp: 9/26/12

7/18/12
[Signature]

LEVELS ID	initial conc.	final conc. (ug/ml)	Aliquot (uL)	Solvent	Final Vol. Solvent (ml)
LEVEL 10	1ug/ml	0.010	10 µL		1.0
LEVEL 50		0.050	50 µL	HEXANE	1.0
LEVEL 100		0.100	100 µL	EM SCIENCE	1.0
LEVEL 250		0.250	250 µL	LOT #082612B	1.0
LEVEL 1000		1.000	1000 µL		1.0

Diesel Spike

© B&D only, no human consumption. Made in the USA.
 Diesel Fuel #2 Composite,
 50,000 mg/L, 1 ml
 011598-03
 Lot # Storage Expiry
 183767 ≤ 10 Degrees C 2/11/16
 Solv: Methylene Chloride
 Diesel Fuel #2 Composite
 Lot #: 183767 - 30901
 Rec: 5/30/12 MFR exp. 02/11/16

7/18/12

DDK
 OP: 7/18/12
 EX: 7/18/12

STANDARD

INITIAL CONC

SOURCE DATE

ALIQOT VOLUME

FINAL VOLUME

FINAL CONC

SOLVENT/ LOTS

DATE/ INITIALS

015

DATE/ INITIALS

THC Surrogate (Gave to Extractions)

CM 7-6-12	O-Tetraphenyl	600 mg/L	025E	N/A	25 ml	600 mg/ml	N/A	CM
7-22-12	Octacosane		CAT: 11036-05					7-9-12
			LOT: 188683-30664 thru 668					ex. 7-9-12
			Op. 7-9-12					
			ex. 7-9-13					

CM 7-6-12								
ex. 7-28-12	13-DBP	100 mg/mL	1,3 DBP STK	35 ml	10 ml	Method		CM
			prep. 5-14-12			0.35 mg/ml		7-9-12
			ex. 5-14-13					ex. 10-9-12

OP FAMPHUR CURVE												
PREP:	07/09/12	EXP:	07/28/12			IA	1	2	3	4	5	6
SUPPLIER	ID#	[µg/mL]	LOT #	DATE	EXP. DATE	µL						
	OP/FAMPHUR S	5		07/09/12	07/28/12	2	10	50	200	500	700	1000
VWR	Hexane		082610B			998	990	950	800	500	300	NA
					Final VOL.	1000	1000	1000	1000	1000	1000	1000
OP 2ND SRC												
PREP:	07/09/12	5		DATE	EXP. DATE	500						
EXP:	09/23/12	Hexane Lot	082610B	05/11/12	09/23/12	1000						

CM 7-9-12

PREP DATE: 07/09/12											
OPC CURVE											
EXP:	10/06/12					1	2	3	4	5	6
SUPPLIER	ID#	[µg/mL]	LOT #	DATE	EXP. DATE	µL	µL	µL	µL	µL	µL
	OPC STD	5		06/19/12	10/06/12	10	50	200	500	700	1000
	Hexane		082610B			990	950	800	500	300	NA
					Final VOL.	1000	1000	1000	1000	1000	1000

CM 7-9-12

CM
7-6-12
ex. 7-20-12

CM
7-9-12
ex. 7-28-12

CM
7-9-12
ex. 10-6-12

Organic Extraction Worksheet

Method	THC Separatory Funnel Extraction 3510C	Extraction Set	120723A	Extraction Method	SEP011	Units	mL	
Spiked ID 1	Diesel Ampule 183767-30901	Surrogate ID 1	THC Surrogate 188683-30667					
Spiked ID 2	Motor Oil Ampule 183768-30234	Surrogate ID 2						
Spiked ID 3		Surrogate ID 3						
Spiked ID 4		Surrogate ID 4						
Spiked ID 5		Surrogate ID 5						
Spiked ID 6		Sufficient Vol for Matrix QC:		YES				
Spiked ID 7		Ext. Start Time:		07/23/12 14:15				
Spiked ID 8		Ext. End Time:		07/24/12 10:27				
		GC Requires Extract By:		08/01/12 0:00				
		pH1			Water Bath Temp Criteria			78,76,80 °
		pH2						
		pH3						

Spiked By: IC

Date 07/23/12

Witnessed By: JM

Date 07/23/12

Sample	Sample Container	Spike Amount	Spike ID	Surrogate Amount	Surrogate ID	Extract Amount	Final Volume	pH	Extract Date/Time	Comments
1	120723A Blk			0.250	1	1000	5	7	07/23/12 14:15	
						equip				
						E-WB7,78				
2	120723A LCS-1	0.040	1	0.250	1	1000	5	7	07/23/12 14:15	
						equip				
						E-WB7,78				
3	120723A LCS-2	0.040	2	0.250	1	1000	5	7	07/23/12 14:15	
						equip				
						E-WB7,78				
4	AY65041 AY65041W05			0.250	1	1050	5	7	07/23/12 14:15	68248-2 WEEK RUSH -- Amber Liter
						equip				
						E-WB7,78				
5	AY65043 AY65043W06			0.250	1	1050	5	7	07/23/12 14:15	68248-2 WEEK RUSH -- Amber Liter
						equip				
						E-WB7,78				
6	AY65044 AY65044W07			0.250	1	1040	5	7	07/23/12 14:15	68248-2 WEEK RUSH -- Amber Liter
						equip				
						E-WB7,78				
7	AY65112 AY65112W04			0.250	1	1030	5	7	07/23/12 14:15	68258-2 WEEK RUSH -- Amber Liter
						equip				
						E-WB7,78				
8	AY65113 AY65113W05			0.250	1	1030	5	7	07/23/12 14:15	68258-2 WEEK RUSH -- Amber Liter
						equip				
						E-WB7,78				
9	AY65144 MS-1 AY65144W15	0.040	1	0.250	1	1000	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
						equip				
						E-WB7,78				
0	AY65144 MSD-1 AY65144W07	0.040	1	0.250	1	1020	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
						equip				
						E-WB5,76				
1	AY65144 MS-2 AY65144W17	0.040	2	0.250	1	1010	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
						equip				
						E-WB5,76				
2	AY65144 MSD-2 AY65144W19	0.040	2	0.250	1	1050	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
						equip				
						E-WB5,76				
3	AY65144 AY65144W05			0.250	1	1010	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
						equip				
						E-WB5,76				

vent and Lot#	
	EMD52104
SO4	2351C512

Extraction COC Transfer	
Extraction lab employee Initials	DRA
GC analyst's initials	LH
Date	7/30/12
Time	1200
Refrigerator	Hobart

Technician's Initials	
Scanned By	JM
Sample Preparation	JM
Extraction	JM/IC
Concentration	IC
Modified	07/25/12 1:07:07 PM

Reviewed By: DRA Date 07/25/12

Organic Extraction Worksheet


Method	THC Separatory Funnel Extraction 3510C	Extraction Set	120723A	Extraction Method	SEP011	Units	mL
Spiked ID 1	Diesel Ampule 183767-30901	Surrogate ID 1	THC Surrogate 188683-30667				
Spiked ID 2	Motor Oil Ampule 183768-30234	Surrogate ID 2					
Spiked ID 3		Surrogate ID 3					
Spiked ID 4		Surrogate ID 4					
Spiked ID 5		Surrogate ID 5					
Spiked ID 6		Sufficient Vol for Matrix QC:		YES			
Spiked ID 7		Ext. Start Time:		07/23/12 14:15			
Spiked ID 8		Ext. End Time:		07/24/12 10:27			
		GC Requires Extract By:		08/01/12 0:00			
		pH1		Water Bath Temp Criteria		78,76,80 °	
		pH2					
		pH3					

Spiked By: IC

Date 07/23/12

Witnessed By: JM

Date 07/23/12

Sample	Sample Container	Spike Amount	Spike ID	Surrogate Amount	Surrogate ID	Extract Amount	Final Volume	pH	Extract Date/Time	Comments
14	AY65145 			0.250	1	1010 equip E-WB5,76	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
15	AY65146 			0.250	1	1020 equip E-WB5,76	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
16	AY65147 			0.250	1	1060 equip E-WB5,76	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
17	AY65148 			0.250	1	1050 equip E-WB5,76	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
18	AY65149 			0.250	1	1050 equip E-WB6,80	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
19	AY65150 			0.250	1	1070 equip E-WB6,80	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
20	AY65151 			0.250	1	1070 equip E-WB6,80	5	7	07/23/12 14:15	68266-2 WEEK RUSH -- Amber Liter
21	AY65169 			0.250	1	1050 equip E-WB6,80	5	7	07/23/12 14:15	68269-2 WEEK RUSH -- Amber Liter
22	AY65170 			0.250	1	1050 equip E-WB6,80	5	7	07/23/12 14:15	68269-2 WEEK RUSH -- Amber Liter

DRA 7/25/12

Solvent and Lot#	
AC	EMD52104
2a2SO4	2351C512

Extraction COC Transfer	
Extraction lab employee Initials	DRA
GC analyst's initials	
Date	
Time	
Refrigerator	

Technician's Initials	
Scanned By	JM
Sample Preparation	JM
Extraction	JM/IC
Concentration	IC
Modified	07/25/12 1:07:07 PM

Reviewed By: DRA

Date 07/25/12

Injection Log

Directory: G:\APOLLO\DATA\120622\

Line	Vial	FileName	Multiplier	SampleName	Misc Info	Injected
1	4	622004.D	1	TCH SURROGATE 100/1000	Mix(c)	6-22-12 18:22:29
2	5	622005.D	1	TCH SURROGATE 400/1000	Mix(c)	6-22-12 18:46:55
3	6	622006.D	1	TCH SURROGATE 600/1000	Mix(c)	6-22-12 19:10:46
4	7	622007.D	1	TCH SURROGATE 800/1000	Mix(c)	6-22-12 19:34:47
5	8	622008.D	1	TCH SURROGATE 1000/1000	Mix(c)	6-22-12 19:58:49
6	9	622009.D	1	DIESEL 10/1000 6/22/12	Mix(A)	6-22-12 20:22:56
7	10	622010.D	1	DIESEL 100/1000	Mix(A)	6-22-12 20:47:06
8	11	622011.D	1	DIESEL 400/1000	Mix(A)	6-22-12 21:11:13
9	12	622012.D	1	DIESEL 600/1000	Mix(A)	6-22-12 21:35:18
10	13	622013.D	1	DIESEL 800/1000	Mix(A)	6-22-12 21:59:20
11	14	622014.D	1	DIESEL 1000/1000	Mix(A)	6-22-12 22:23:21
12	15	622015.D	1	DIESEL 2ND SRC 6/22/12	Mix(A)	6-22-12 22:47:20
13	2	731002.D	1	DIESEL 400ppm 7/31/12	Mix(A)	7-31-12 10:15:07
14	13	731013.D	5	120723A BLK 5/1000	Water	7-31-12 14:39:54
15	14	731014.D	5	120723A LCS-1 5/1000	Water	7-31-12 15:03:52
16	19	731019.D	4.85437	AY65112W04 5/1030	Water	7-31-12 17:03:40
17	20	731020.D	1	DIESEL 400ppm 7/30/12	Mix(A)	7-31-12 17:28:05
18	22	731022.D	4.85437	AY65113W05 5/1030	Water	7-31-12 18:16:17
19	32	731032.D	1	DIESEL 400ppm 7/30/12	Mix(A)	7-31-12 22:20:07

EPA METHOD 8260B
Volatile Organic Compounds

APPL, INC.

**EPA METHOD 8260B
Volatile Organic Compounds
QC Summary**

Method Blank

EPA 8260B VOCs + Gas Water

Blank Name/QCG: **120719W-65112 - 169441**
 Batch ID: #86RHB-120719AT1

APPL Inc.
 908 North Temperance Avenue
 Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	1,1,1,2-TETRACHLOROETHANE	0.26 U	1.0	0.26	0.13	ug/L	07/20/12	07/20/12
BLANK	1,1,1-TRICHLOROETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
BLANK	1,1,2,2-TETRACHLOROETHANE	0.20 U	1.0	0.20	0.10	ug/L	07/20/12	07/20/12
BLANK	1,1,2-TRICHLOROETHANE	0.40 U	1.0	0.40	0.20	ug/L	07/20/12	07/20/12
BLANK	1,1-DICHLOROETHANE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	1,1-DICHLOROETHENE	0.60 U	1.0	0.60	0.30	ug/L	07/20/12	07/20/12
BLANK	1,2,3-TRICHLOROPROPANE	0.78 U	2.0	0.78	0.39	ug/L	07/20/12	07/20/12
BLANK	1,2,4-TRICHLOROBENZENE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
BLANK	1,2-DIBROMO-3-CHLOROPROPA	1.52 U	2.0	1.52	0.76	ug/L	07/20/12	07/20/12
BLANK	1,2-DIBROMOETHANE	0.40 U	1.0	0.40	0.20	ug/L	07/20/12	07/20/12
BLANK	1,2-DICHLOROBENZENE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
BLANK	1,2-DICHLOROETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
BLANK	1,2-DICHLOROPROPANE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
BLANK	1,3-DICHLOROBENZENE	0.22 U	1.0	0.22	0.11	ug/L	07/20/12	07/20/12
BLANK	1,3-DICHLOROPROPENE, TOTAL	0.36 U	1.0	0.36	0.18	ug/L	07/20/12	07/20/12
BLANK	1,4-DICHLOROBENZENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	2-BUTANONE	1.20 U	10.0	1.20	0.60	ug/L	07/20/12	07/20/12
BLANK	4-METHYL-2-PENTANONE	3.80 U	10.0	3.80	1.90	ug/L	07/20/12	07/20/12
BLANK	ACETONE	1.90 U	10.0	1.90	0.95	ug/L	07/20/12	07/20/12
BLANK	BENZENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
BLANK	BROMODICHLOROMETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
BLANK	BROMOFORM	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
BLANK	BROMOMETHANE	0.48 U	2.0	0.48	0.24	ug/L	07/20/12	07/20/12
BLANK	CARBON TETRACHLORIDE	0.20 U	1.0	0.20	0.10	ug/L	07/20/12	07/20/12
BLANK	CHLOROBENZENE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
BLANK	CHLORODIBROMOMETHANE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	CHLOROETHANE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
BLANK	CHLOROFORM	0.14 U	1.0	0.14	0.07	ug/L	07/20/12	07/20/12
BLANK	CHLOROMETHANE	0.84 U	1.0	0.84	0.42	ug/L	07/20/12	07/20/12
BLANK	CIS-1,2-DICHLOROETHENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
BLANK	ETHYLBENZENE	0.46 U	1.0	0.46	0.23	ug/L	07/20/12	07/20/12
BLANK	GASOLINE	12.12 U	20.0	12.12	6.06	ug/L	07/20/12	07/20/12
BLANK	HEXACHLOROBUTADIENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	METHYL TERT-BUTYL ETHER	0.52 U	1.0	0.52	0.26	ug/L	07/20/12	07/20/12

Quant Method: TALLW.M
 Run #: 0719T38
 Instrument: Thor
 Sequence: T120719
 Initials: ARS

Printed: 07/31/12 9:50:32 AM
 GC SC-Blank-REG MDLs

Method Blank
EPA 8260B VOCs + Gas Water

Blank Name/QCG: **120719W-65112 - 169441**
 Batch ID: #86RHB-120719AT1

APPL Inc.
 908 North Temperance Avenue
 Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	METHYLENE CHLORIDE	0.70 U	5.0	0.70	0.35	ug/L	07/20/12	07/20/12
BLANK	STYRENE	0.50 U	1.0	0.50	0.25	ug/L	07/20/12	07/20/12
BLANK	TETRACHLOROETHENE	0.48 U	1.0	0.48	0.24	ug/L	07/20/12	07/20/12
BLANK	TOLUENE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
BLANK	TRANS-1,2-DICHLOROETHENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	TRICHLOROETHENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
BLANK	VINYL CHLORIDE	0.46 U	1.0	0.46	0.23	ug/L	07/20/12	07/20/12
BLANK	XYLENES (TOTAL)	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	SURROGATE: 1,2-DICHLOROET	102	70-120			%	07/20/12	07/20/12
BLANK	SURROGATE: 4-BROMOFLUORO	101	75-120			%	07/20/12	07/20/12
BLANK	SURROGATE: DIBROMOFLUOR	100	85-115			%	07/20/12	07/20/12
BLANK	SURROGATE: TOLUENE-D8 (S)	99.7	85-120			%	07/20/12	07/20/12

Quant Method: TALLW.M
 Run #: 0719T38
 Instrument: Thor
 Sequence: T120719
 Initials: ARS

Printed: 07/31/12 9:50:32 AM
 GC SC-Blank-REG MDLs

Method Blank
EPA 8260B VOCS + GAS WATER

Blank Name/QCG: **120719W-65113 - 169517**
Batch ID: #86RHB-120719AT2

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	GASOLINE	12.12 U	20.0	12.12	6.06	ug/L	07/20/12	07/20/12
BLANK	SURROGATE: 1,2-DICHLOROET	102	70-120			%	07/20/12	07/20/12
BLANK	SURROGATE: 4-BROMOFLUORO	101	75-120			%	07/20/12	07/20/12
BLANK	SURROGATE: DIBROMOFLUOR	100	85-115			%	07/20/12	07/20/12
BLANK	SURROGATE: TOLUENE-D8 (S)	99.7	85-120			%	07/20/12	07/20/12

Quant Method: TALLW.M
Run #: 0719T38
Instrument: Thor
Sequence: T120719
Initials: ARS

Printed: 07/31/12 1:07:17 PM
GC SC-Blank-REG MDLs

Surrogate Recovery

Lab Name: APPL, Inc.
 Case No: 68258
 Matrix: WATER

SDG No: 68258
 Date Analyzed: 07/19/12
 Instrument: Thor

APPL ID.	Client Sample No.	SURROGATE: 1,2-DICHLOROETHANE-D4 (S)			SURROGATE: 4-BROMOFLUOROBENZENE (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
120719AT1-LCS	Lab Control Spike	70-120	97.8		75-120	102	
120719AT1-BLK	Blank	70-120	102		75-120	101	
AY65114	ES085-TRIP BLANK	70-120	99.9		75-120	99.3	
AY65112	ES081	70-120	98.7		75-120	98.5	
AY65113	ES082	70-120	99.0		75-120	100	

Comments: Batch: #86RHB-120719AT

Surrogate Recovery

Lab Name: APPL, Inc.
 Case No: 68258
 Matrix: WATER

SDG No: 68258
 Date Analyzed: 07/19/12
 Instrument: Thor

APPL ID.	Client Sample No.	SURROGATE: DIBROMOFLUOROMETHANE (S)			SURROGATE: TOLUENE-D8 (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
120719AT1-LCS	Lab Control Spike	85-115	98.2		85-120	98.0	
120719AT1-BLK	Blank	85-115	100		85-120	99.7	
AY65114	ES085-TRIP BLANK	85-115	98.6		85-120	99.2	
AY65112	ES081	85-115	98.8		85-120	98.8	
AY65113	ES082	85-115	99.3		85-120	98.7	

Comments: Batch: #86RHB-120719AT

Surrogate Recovery

Lab Name: APPL, Inc.
 Case No: 68258
 Matrix: WATER

SDG No: 68258
 Date Analyzed: 07/19/12
 Instrument: Thor

APPL ID.	Client Sample No.	SURROGATE: 1,2-DICHLOROETHANE-D4 (S)			SURROGATE: 4-BROMOFLUOROBENZENE (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
120719AT2-LCS	Lab Control Spike	70-120	97.8		75-120	102	
120719AT2-BLK	Blank	70-120	102		75-120	101	
AY65113	ES082	70-120	99.0		75-120	100	

Comments: Batch: #86RHB-120719AT

Surrogate Recovery

Lab Name: APPL, Inc.
 Case No: 68258
 Matrix: WATER

SDG No: 68258
 Date Analyzed: 07/19/12
 Instrument: Thor

APPL ID.	Client Sample No.	SURROGATE: DIBROMOFLUOROMETHANE (S)			SURROGATE: TOLUENE-D8 (S)		
		Limits	Result	Qualifier	Limits	Result	Qualifier
120719AT2-LCS	Lab Control Spike	85-115	98.2		85-120	98.0	
120719AT2-BLK	Blank	85-115	100		85-120	99.7	
AY65113	ES082	85-115	99.3		85-120	98.7	

Comments: Batch: #86RHB-120719AT

Laboratory Control Spike Recovery

EPA 8260B VOCs + Gas Water

APPL ID: 120719W-65112 LCS - 169441

Batch ID: #86RHB-120719AT1

APPL Inc.

908 North Temperance Avenue

Clovis, CA 93611

Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits
1,1,1,2-TETRACHLOROETHANE	10.00	9.79	97.9	80-130
1,1,1-TRICHLOROETHANE	10.00	9.62	96.2	65-130
1,1,2,2-TETRACHLOROETHANE	10.00	9.22	92.2	65-130
1,1,2-TRICHLOROETHANE	10.00	9.61	96.1	75-125
1,1-DICHLOROETHANE	10.00	10.1	101	70-135
1,1-DICHLOROETHENE	10.00	9.64	96.4	70-130
1,2,3-TRICHLOROPROPANE	10.00	10.3	103	75-125
1,2,4-TRICHLOROBENZENE	10.00	9.98	99.8	65-135
1,2-DIBROMO-3-CHLOROPROPANE	10.00	10.1	101	50-130
1,2-DIBROMOETHANE	10.00	9.79	97.9	70-130
1,2-DICHLOROBENZENE	10.00	9.83	98.3	70-120
1,2-DICHLOROETHANE	10.00	9.76	97.6	70-130
1,2-DICHLOROPROPANE	10.00	10.1	101	75-125
1,3-DICHLOROBENZENE	10.00	10.2	102	75-125
1,3-DICHLOROPROPENE, TOTAL	20.0	18.7	93.5	70-130
1,4-DICHLOROBENZENE	10.00	9.82	98.2	75-125
2-BUTANONE	10.00	10.3	103	30-150
4-METHYL-2-PENTANONE	10.00	10.2	102	60-135
ACETONE	10.00	11.8	118	40-140
BENZENE	10.00	9.47	94.7	80-120
BROMODICHLOROMETHANE	10.00	9.57	95.7	75-120
BROMOFORM	10.00	9.70	97.0	70-130
BROMOMETHANE	10.00	9.36	93.6	30-145
CARBON TETRACHLORIDE	10.00	10.0	100	65-140
CHLOROBENZENE	10.00	9.82	98.2	80-120
CHLORODIBROMOMETHANE	10.00	9.73	97.3	60-135
CHLOROETHANE	10.00	9.84	98.4	60-135
CHLOROFORM	10.00	9.60	96.0	65-135
CHLOROMETHANE	10.00	9.80	98.0	40-125
CIS-1,2-DICHLOROETHENE	10.00	9.99	99.9	70-125
ETHYLBENZENE	10.00	10.1	101	75-125

Comments: _____

<u>Primary</u>	<u>SPK</u>
Quant Method :	TALLW.M
Extraction Date :	07/19/12
Analysis Date :	07/19/12
Instrument :	Thor
Run :	0719T31
Initials :	ARS

Printed: 07/31/12 9:50:23 AM

APPL Standard LCS

Laboratory Control Spike Recovery

EPA 8260B VOCs + Gas Water

APPL ID: 120719W-65112 LCS - 169441
 Batch ID: #86RHB-120719AT1

APPL Inc.
 908 North Temperance Avenue
 Clovis, CA 93611

Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits
GASOLINE	300	290	96.7	75-125
HEXACHLOROBUTADIENE	10.00	9.23	92.3	50-140
METHYL TERT-BUTYL ETHER	10.00	9.48	94.8	65-125
METHYLENE CHLORIDE	10.00	9.45	94.5	55-140
STYRENE	10.00	10.4	104	65-135
TETRACHLOROETHENE	10.00	10.1	101	45-150
TOLUENE	10.00	10.1	101	75-120
TRANS-1,2-DICHLOROETHENE	10.00	9.79	97.9	60-140
TRICHLOROETHENE	10.00	10.6	106	70-125
VINYL CHLORIDE	10.00	10.1	101	50-145
XYLENES (TOTAL)	30.0	31.3	104	80-120
SURROGATE: 1,2-DICHLOROETHANE-D	33.6	32.9	97.8	70-120
SURROGATE: 4-BROMOFLUOROBENZE	29.5	30.2	102	75-120
SURROGATE: DIBROMOFLUOROMETH	31.9	31.3	98.2	85-115
SURROGATE: TOLUENE-D8 (S)	37.3	36.6	98.0	85-120

Comments: _____

<u>Primary</u>	<u>SPK</u>
Quant Method :	TALLW.M
Extraction Date :	07/19/12
Analysis Date :	07/19/12
Instrument :	Thor
Run :	0719T31
Initials :	ARS

Printed: 07/31/12 9:50:24 AM
 APPL Standard LCS

Laboratory Control Spike Recovery

EPA 8260B VOCS + GAS WATER

APPL ID: 120719W-65113 LCS - 169517
 Batch ID: #86RHB-120719AT2

APPL Inc.
 908 North Temperance Avenue
 Clovis, CA 93611

Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits
GASOLINE	300	389	130 #	75-125
SURROGATE: 1,2-DICHLOROETHANE-D	33.6	32.9	97.8	70-120
SURROGATE: 4-BROMOFLUOROBENZE	29.5	30.2	102	75-120
SURROGATE: DIBROMOFLUOROMETH	31.9	31.3	98.2	85-115
SURROGATE: TOLUENE-D8 (S)	37.3	36.6	98.0	85-120

= Recovery is outside QC limits.

Comments:

<u>Primary</u>	<u>SPK</u>
Quant Method :	TALLW.M
Extraction Date :	07/19/12
Analysis Date :	07/19/12
Instrument :	Thor
Run :	0719T31
Initials :	ARS

Printed: 07/31/12 1:07:01 PM
 APPL Standard LCS

EPA 8260B

Form 4

Blank Summary

Lab Name: APPL, Inc.

SDG No: 68258

Case No: 68258

Date Analyzed: 07/20/12

Matrix: WATER

Instrument: Thor

Blank ID: 120719AT1-BLK

Time Analyzed: 0218

<u>APPL ID.</u>	<u>Client Sample No.</u>	<u>File ID.</u>	<u>Date Analyzed</u>
120719AT1-LCS	Lab Control Spike	0719T31	07/19/12 2303
120719AT1-BLK	Blank	0719T38	07/20/12 0218
AY65114	ES085-TRIP BLANK	0719T42	07/20/12 0408
AY65112	ES081	0719T46	07/20/12 0559
AY65113	ES082	0719T47	07/20/12 0626

Comments: Batch: #86RHB-120719AT

Printed: 07/27/12 2:27:16 PM
Form 4, Blank Summary

EPA 8260B

Form 4

Blank Summary

Lab Name: APPL, Inc.
Case No: 68258
Matrix: WATER
Blank ID: 120719AT2-BLK

SDG No: 68258
Date Analyzed: 07/20/12
Instrument: Thor
Time Analyzed: 0218

APPL ID.	Client Sample No.	File ID.	Date Analyzed
120719AT2-LCS	Lab Control Spike	0719T31	07/19/12 2303
120719AT2-BLK	Blank	0719T38	07/20/12 0218
AY65113	ES082	0719T47	07/20/12 0626

Comments: Batch: #86RHB-120719AT

Printed: 07/31/12 1:06:51 PM
Form 4, Blank Summary

Form 5
Tune Summary

Lab Name: APPL Inc.
 Case No: 0719T28.D
 Matrix: Water
 ID: 5ng- BFB Std 07-16-12B

SDG No: 68258
 Date Analyzed: 07/19/12
 Instrument: Thor
 Time Analyzed: 21:40

	Client Sample No.	APPL ID.	File ID.	Date Analyzed
1		10ug/L Vol Std 07-19	0719T30.D	07/19/12 22:35
2	Lab Control Spike	120719A LCS-1WT (SS)	0719T31.D	07/19/12 23:03
3	Blank	120719A BLK-1WT	0719T38.D	07/20/12 2:18
4	ES085-TRIP BLANK	AY65114W01	0719T42.D	07/20/12 4:08
5	ES081	AY65112W01	0719T46.D	07/20/12 5:59
6	ES082	AY65113W01	0719T47.D	07/20/12 6:26
7				
8				
9				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				

m/e

50	14.9 - 40% of mass 95	<u>16.9</u>
75	30 - 60% of mass 95	<u>47.3</u>
95	100 - 100% of mass 95	<u>100.0</u>
96	5 - 9% of mass 95	<u>6.8</u>
173	0 - 2% of mass 174	<u>1.0</u>
174	50 - 100.49% of mass 95	<u>95.8</u>
175	5 - 9% of mass 174	<u>7.5</u>
176	95 - 101.49% of mass 174	<u>96.9</u>
177	5 - 9% of mass 176	<u>6.2</u>

Form 5
Tune Summary

Lab Name: APPL Inc.

SDG No: 68258

Case No: 0719T28.D

Date Analyzed: 07/19/12

Matrix: Water

Instrument: Thor

ID: 5ng- BFB Std 07-16-12B

Time Analyzed: 21:40

Client Sample No.	APPL ID.	File ID.	Date Analyzed	
1		CCV gas 300ug/L (SS)	0719T33.D	07/19/12 23:59
2		CCV gas 300ug/L	0719T34.D	07/20/12 0:27
3	Lab Control Spike	LCS gas 300ug/L	0719T35.D	07/20/12 0:54
4	Blank	120719A BLK-1WT	0719T38.D	07/20/12 2:18
5	ES082	AY65113W01	0719T47.D	07/20/12 6:26
6				
7				
8				
9				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				

m/e

50 15 - 40% of mass 95	<u>16.9</u>
75 30 - 60% of mass 95	<u>47.3</u>
95 100 - 100% of mass 95	<u>100.0</u>
96 5 - 9% of mass 95	<u>6.8</u>
173 0 - 2% of mass 174	<u>1.0</u>
174 50 - 100% of mass 95	<u>95.8</u>
175 5 - 9% of mass 174	<u>7.5</u>
176 95 - 101% of mass 174	<u>96.9</u>
177 5 - 9% of mass 176	<u>6.2</u>

Form 5
Tune Summary

Lab Name: APPL Inc.
 Case No: 0725T13.D
 Matrix: Water
 ID: 5ng- BFB STD 07-16-12B

SDG No: 68258
 Date Analyzed: 07/25/12
 Instrument: Thor
 Time Analyzed: 14:59

Client Sample No.	APPL ID.	File ID.	Date Analyzed
1	CCV gas 300ug/L	0725T14.D	07/25/12 15:27
2	Lab Control Spike	LCS gas 300ug/L (SS)	0725T15.D
3	Blank	120725A BLK-1WT	0725T20.D
4	ES085-TRIP BLANK	AY65114W02	0725T22.D
5	ES081	AY65112W02	07/25/12 19:37
6			
7			
8			
9			
10			
11			
12			
13			
14			
15			
16			
17			
18			
19			
20			
21			
22			

m/e

50 15 - 40% of mass 95	<u>16.9</u>
75 30 - 60% of mass 95	<u>46.8</u>
95 100 - 100% of mass 95	<u>100.0</u>
96 5 - 9% of mass 95	<u>6.7</u>
173 0 - 2% of mass 174	<u>0.3</u>
174 50 - 100% of mass 95	<u>94.5</u>
175 5 - 9% of mass 174	<u>7.4</u>
176 95 - 101% of mass 174	<u>97.6</u>
177 5 - 9% of mass 176	<u>6.1</u>

8A
INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: APPL Inc. Contract: Review
 Lab Code: _____ SDG No.: 68258
 Lab File ID (Standard): 0719T10.D Date Analyzed: 07/19/12
 Instrument ID: Thor Time Analyzed: 13:20
 GC Column: _____ ID: Heated Purge: (Y/N) _____

		Fluorobenzene (IS)		Chlorobenzene-D5 (IS)		1,4-Dichlorobenzene-D (IS)	
		AREA #	RT #	AREA #	RT #	AREA #	RT #
	12 HOUR STD	461760	6.74	382656	9.88	222464	12.20
	UPPER LIMIT	923520	7.24	765312	10.38	444928	12.70
	LOWER LIMIT	230880	6.24	191328	9.38	111232	11.70
	SAMPLE NO.						
01	10ug/L Vol Std 07-19-12	452736	6.73	376000	9.87	220224	12.20
02	120719A LCS-1WT (SS)	459584	6.73	371008	9.87	216768	12.20
03	120719A BLK-1WT	441792	6.73	355584	9.87	206976	12.20
04	AY65114W01	437504	6.73	353408	9.88	204928	12.20
05	AY65112W01	437824	6.73	354240	9.87	200896	12.20
06	AY65113W01	461440	6.72	370688	9.87	220608	12.20
07							
08							
09							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

AREA UPPER LIMIT = +100% of internal standard area.
 AREA LOWER LIMIT = -50% of internal standard area.
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

8A
INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: APPL Inc. Contract: Review
 Lab Code: _____ SDG No.: 68258
 Lab File ID (Standard): 0719T21.D Date Analyzed: 07/19/12
 Instrument ID: Thor Time Analyzed: 18:26
 GC Column: _____ ID: _____ Heated Purge: (Y/N) _____

	Fluorobenzene (IS)		Chlorobenzene-D5 (IS)		1,4-Dichlorobenzene-D (IS)		
	AREA #	RT #	AREA #	RT #	AREA #	RT #	
12 HOUR STD	908946	6.73	1020090	9.88	1167560	12.20	
UPPER LIMIT	1817892	7.23	2040180	10.38	2335120	12.70	
LOWER LIMIT	454473	6.23	510045	9.38	583780	11.70	
SAMPLE NO.							
01	CCV gas 300ug/L (SS)	931728	6.73	1044020	9.87	1183160	12.20
02	CCV gas 300ug/L	923126	6.73	1020890	9.87	1142880	12.20
03	LCS gas 300ug/L	943495	6.73	1050870	9.87	1213950	12.20
04	120719A BLK-1WT	913286	6.73	1028060	9.87	1144200	12.20
05	AY65113W01	945344	6.72	1059320	9.87	1222850	12.20
06							
07							
08							
09							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

AREA UPPER LIMIT = +100% of internal standard area.
 AREA LOWER LIMIT = -50% of internal standard area.
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

8A
INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: APPL Inc. Contract: Review
 Lab Code: _____ SDG No.: 68258
 Lab File ID (Standard): 0725T07.D Date Analyzed: 07/25/12
 Instrument ID: Thor Time Analyzed: 12:13
 GC Column: _____ ID: Heated Purge: (Y/N) _____

	Fluorobenzene (IS)		Chlorobenzene-D5 (IS)		1,4-Dichlorobenzene-D (IS)		
	AREA #	RT #	AREA #	RT #	AREA #	RT #	
12 HOUR STD	782981	6.73	897407	9.87	996199	12.20	
UPPER LIMIT	1565962	7.23	1794814	10.37	1992398	12.70	
LOWER LIMIT	391491	6.23	448704	9.37	498100	11.70	
SAMPLE NO.							
01	CCV gas 300ug/L	819782	6.73	915724	9.87	1043660	12.20
02	LCS gas 300ug/L (SS)	788179	6.73	879850	9.88	1024200	12.20
03	120725A BLK-1WT	787932	6.73	886149	9.87	982900	12.20
04	AY65114W02	780377	6.73	884072	9.87	988550	12.20
05	AY65112W02	775499	6.73	885186	9.88	1012070	12.20
06							
07							
08							
09							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

AREA UPPER LIMIT = +100% of internal standard area.
 AREA LOWER LIMIT = -50% of internal standard area.
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

Manual Integration Summary

ARF: 68258

APPL ID	Client ID	Method	Analyte	Type	Comment
AY65112	Blank	EPA 8260B	GASOLINE	Blank	(MI1) Integration does not follow baseline.
AY65112	LCS	EPA 8260B	GASOLINE	LCS	(MI1) Integration does not follow baseline.
AY65112	ES081	EPA 8260B	GASOLINE	Parent	(MI1) Integration does not follow baseline.
AY65114	ES085-TRIP BLANK	EPA 8260B	GASOLINE	Parent	(MI1) Integration does not follow baseline.

Manual Integration Summary

ARF: 68258

APPL ID	Client ID	Method	Analyte	Type	Comment
AY65113	Blank	EPA 8260B	GASOLINE	Blank	(MI1) Integration does not follow baseline.
AY65113	LCS	EPA 8260B	GASOLINE	LCS	(MI1) Integration does not follow baseline.
AY65113	ES082	EPA 8260B	GASOLINE	Parent	(MI1) Integration does not follow baseline.

EPA METHOD 8260B
Volatile Organic Compounds
Sample Data

EPA 8260B VOCs + Gas Water

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Max Solmssen

Project: LTM Red Hill / 1022-024

ARF: 68258

Sample ID: ES081

APPL ID: AY65112

Sample Collection Date: 07/18/12

QCG: #86RHB-120719AT1-169441

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8260B	1,1,1,2-TETRACHLOROETHANE	0.26 U	1.0	0.26	0.13	ug/L	07/20/12	07/20/12
EPA 8260B	1,1,1-TRICHLOROETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	1,1,2,2-TETRACHLOROETHANE	0.20 U	1.0	0.20	0.10	ug/L	07/20/12	07/20/12
EPA 8260B	1,1,2-TRICHLOROETHANE	0.40 U	1.0	0.40	0.20	ug/L	07/20/12	07/20/12
EPA 8260B	1,1-DICHLOROETHANE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	1,1-DICHLOROETHENE	0.60 U	1.0	0.60	0.30	ug/L	07/20/12	07/20/12
EPA 8260B	1,2,3-TRICHLOROPROPANE	0.78 U	2.0	0.78	0.39	ug/L	07/20/12	07/20/12
EPA 8260B	1,2,4-TRICHLOROBENZENE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DIBROMO-3-CHLOROPROPANE	1.52 U	2.0	1.52	0.76	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DIBROMOETHANE	0.40 U	1.0	0.40	0.20	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DICHLOROBENZENE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DICHLOROETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DICHLOROPROPANE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
EPA 8260B	1,3-DICHLOROBENZENE	0.22 U	1.0	0.22	0.11	ug/L	07/20/12	07/20/12
EPA 8260B	1,3-DICHLOROPROPENE, TOTAL	0.36 U	1.0	0.36	0.18	ug/L	07/20/12	07/20/12
EPA 8260B	1,4-DICHLOROBENZENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	2-BUTANONE	1.20 U	10.0	1.20	0.60	ug/L	07/20/12	07/20/12
EPA 8260B	4-METHYL-2-PENTANONE	3.80 U	10.0	3.80	1.90	ug/L	07/20/12	07/20/12
EPA 8260B	ACETONE	1.90 U	10.0	1.90	0.95	ug/L	07/20/12	07/20/12
EPA 8260B	BENZENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
EPA 8260B	BROMODICHLOROMETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	BROMOFORM	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	BROMOMETHANE	0.48 U	2.0	0.48	0.24	ug/L	07/20/12	07/20/12
EPA 8260B	CARBON TETRACHLORIDE	0.20 U	1.0	0.20	0.10	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROBENZENE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
EPA 8260B	CHLORODIBROMOMETHANE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROETHANE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROFORM	0.14 U	1.0	0.14	0.07	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROMETHANE	0.84 U	1.0	0.84	0.42	ug/L	07/20/12	07/20/12
EPA 8260B	CIS-1,2-DICHLOROETHENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
EPA 8260B	ETHYLBENZENE	0.46 U	1.0	0.46	0.23	ug/L	07/20/12	07/20/12
EPA 8260B	GASOLINE	12.12 U MI1	20.0	12.12	6.06	ug/L	07/20/12	07/20/12
EPA 8260B	HEXACHLOROBUTADIENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	METHYL TERT-BUTYL ETHER	0.52 U	1.0	0.52	0.26	ug/L	07/20/12	07/20/12

U = The analyte was analyzed for, but not detected. The associated numerical value is at or below the MDL.

(MI1) Manual integration: Integration does not follow baseline.

Quant Method: TALLW.M
Run #: 0719T46
Instrument: Thor
Sequence: T120719
Dilution Factor: 1
Initials: ARS

Printed: 07/31/12 9:50:45 AM
APPL-F1-SC-NoMC-REG MDLs

EPA 8260B VOCs + Gas Water

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

Attn: Max Solmssen

Project: LTM Red Hill / 1022-024

Sample ID: ES081

Sample Collection Date: 07/18/12

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 68258

APPL ID: AY65112

QCG: #86RHB-120719AT1-169441

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8260B	METHYLENE CHLORIDE	0.70 U	5.0	0.70	0.35	ug/L	07/20/12	07/20/12
EPA 8260B	STYRENE	0.50 U	1.0	0.50	0.25	ug/L	07/20/12	07/20/12
EPA 8260B	TETRACHLOROETHENE	0.48 U	1.0	0.48	0.24	ug/L	07/20/12	07/20/12
EPA 8260B	TOLUENE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
EPA 8260B	TRANS-1,2-DICHLOROETHENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	TRICHLOROETHENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
EPA 8260B	VINYL CHLORIDE	0.46 U	1.0	0.46	0.23	ug/L	07/20/12	07/20/12
EPA 8260B	XYLENES (TOTAL)	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	SURROGATE: 1,2-DICHLOROETHANE	98.7	70-120			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: 4-BROMOFLUOROBEN	98.5	75-120			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: DIBROMOFLUOROMET	98.8	85-115			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: TOLUENE-D8 (S)	98.8	85-120			%	07/20/12	07/20/12

U = The analyte was analyzed for, but not detected. The associated numerical value is at or below the MDL.
(MI1) Manual integration: Integration does not follow baseline.

Quant Method: TALLW.M
Run #: 0719T46
Instrument: Thor
Sequence: T120719
Dilution Factor: 1
Initials: ARS

*Printed: 07/31/12 9:50:45 AM
APPL-F1-SC-NoMC-REG MDLs*

Data File : M:\THOR\DATA\T120719\0719T46.D
Acq On : 20 Jul 12 5:59
Sample : AY65112W01
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 46
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 12:18 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 10:40:23 2012
Response via : Initial Calibration
DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) Fluorobenzene (IS)	6.73	96	437824	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.87	117	354240	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	200896	25.00000	ppb	0.00
System Monitoring Compounds						
31) Dibromofluoromethane(S)	5.94	111	215711	31.48437	ppb	0.00
Spiked Amount	31.881			Recovery =	98.754%	
36) 1,2-DCA-D4 (S)	6.32	65	211537	33.22239	ppb	0.00
Spiked Amount	33.647			Recovery =	98.738%	
56) Toluene-D8 (S)	8.43	98	772846	36.90355	ppb	0.00
Spiked Amount	37.345			Recovery =	98.819%	
64) 4-Bromofluorobenzene(S)	11.05	95	287895	29.06868	ppb	0.00
Spiked Amount	29.515			Recovery =	98.488%	

Target Compounds

Qvalue


Data File : M:\THOR\DATA\T120719\0719T46.D
Acq On : 20 Jul 12 5:59
Sample : AY65112W01
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 46
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 12:18 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 10:40:23 2012
Response via : Initial Calibration



Data File : M:\THOR\DATA\T120725\0725T23.D Vial: 22
 Acq On : 25 Jul 12 19:37 Operator: DG,RS,HW,ARS,SV
 Sample : AY65112W02 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 26 8:11 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	775499	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.88	TIC	885186	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1012066	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.43	TIC	10254338m	24.83425	ppb	ND 100

*No gasoline pattern detected.
 ARS 7/26/12*

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T23.D
Acq On : 25 Jul 12 19:37
Sample : AY65112W02
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 22
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 26 8:11 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T23.D
 Acq On : 25 Jul 12 19:37
 Sample : AY65112W02
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 8:10 2012

Vial: 22
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Multiple Level Calibration



TIC: 0725T23.D

(2) Gasoline (TMHB)

8.50min -56.2379ppb m

response 7514555



Ion	Exp%	Act%
TIC	100	100
0.00	0.00	1.30#
0.00	0.00	3.74#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T23.D
 Acq On : 25 Jul 12 19:37
 Sample : AY65112W02
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 8:11 2012

Vial: 22
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Multiple Level Calibration



TIC: 0725T23.D

(2) Gasoline (TMHB)

8.43min 24.8342ppb m

response 10254338

Ion	Exp%	Act%
TIC	100	100
0.00	0.00	0.95#
0.00	0.00	2.74#
0.00	0.00	0.00

EPA 8260B VOCs + Gas Water

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

Attn: Max Solmssen

Project: LTM Red Hill / 1022-024

Sample ID: ES082

Sample Collection Date: 07/18/12

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 68258

APPL ID: AY65113

QCG: #86RHB-120719AT1-169441

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8260B	1,1,1,2-TETRACHLOROETHANE	0.26 U	1.0	0.26	0.13	ug/L	07/20/12	07/20/12
EPA 8260B	1,1,1-TRICHLOROETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	1,1,2,2-TETRACHLOROETHANE	0.20 U	1.0	0.20	0.10	ug/L	07/20/12	07/20/12
EPA 8260B	1,1,2-TRICHLOROETHANE	0.40 U	1.0	0.40	0.20	ug/L	07/20/12	07/20/12
EPA 8260B	1,1-DICHLOROETHANE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	1,1-DICHLOROETHENE	0.60 U	1.0	0.60	0.30	ug/L	07/20/12	07/20/12
EPA 8260B	1,2,3-TRICHLOROPROPANE	0.78 U	2.0	0.78	0.39	ug/L	07/20/12	07/20/12
EPA 8260B	1,2,4-TRICHLOROBENZENE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DIBROMO-3-CHLOROPROPANE	1.52 U	2.0	1.52	0.76	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DIBROMOETHANE	0.40 U	1.0	0.40	0.20	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DICHLOROBENZENE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DICHLOROETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DICHLOROPROPANE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
EPA 8260B	1,3-DICHLOROBENZENE	0.22 U	1.0	0.22	0.11	ug/L	07/20/12	07/20/12
EPA 8260B	1,3-DICHLOROPROPENE, TOTAL	0.36 U	1.0	0.36	0.18	ug/L	07/20/12	07/20/12
EPA 8260B	1,4-DICHLOROBENZENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	2-BUTANONE	1.20 U	10.0	1.20	0.60	ug/L	07/20/12	07/20/12
EPA 8260B	4-METHYL-2-PENTANONE	3.80 U	10.0	3.80	1.90	ug/L	07/20/12	07/20/12
EPA 8260B	ACETONE	1.90 U	10.0	1.90	0.95	ug/L	07/20/12	07/20/12
EPA 8260B	BENZENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
EPA 8260B	BROMODICHLOROMETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	BROMOFORM	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	BROMOMETHANE	0.48 U	2.0	0.48	0.24	ug/L	07/20/12	07/20/12
EPA 8260B	CARBON TETRACHLORIDE	0.20 U	1.0	0.20	0.10	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROBENZENE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
EPA 8260B	CHLORODIBROMOMETHANE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROETHANE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROFORM	0.14 U	1.0	0.14	0.07	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROMETHANE	0.84 U	1.0	0.84	0.42	ug/L	07/20/12	07/20/12
EPA 8260B	CIS-1,2-DICHLOROETHENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
EPA 8260B	ETHYLBENZENE	0.46 U	1.0	0.46	0.23	ug/L	07/20/12	07/20/12
EPA 8260B	HEXACHLOROBUTADIENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	METHYL TERT-BUTYL ETHER	0.52 U	1.0	0.52	0.26	ug/L	07/20/12	07/20/12
EPA 8260B	METHYLENE CHLORIDE	0.70 U	5.0	0.70	0.35	ug/L	07/20/12	07/20/12

J = Estimated value.

Quant Method: TALLW.M
Run #: 0719T47
Instrument: Thor
Sequence: T120719
Dilution Factor: 1
Initials: ARS

Printed: 07/31/12 1:13:46 PM
APPL-F1-SC-NoMC-REG MDLs

EPA 8260B VOCs + Gas Water

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

Attn: Max Solmssen

Project: LTM Red Hill / 1022-024

Sample ID: ES082

Sample Collection Date: 07/18/12

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 68258

APPL ID: AY65113

QCG: #86RHB-120719AT1-169441

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8260B	STYRENE	0.50 U	1.0	0.50	0.25	ug/L	07/20/12	07/20/12
EPA 8260B	TETRACHLOROETHENE	0.48 U	1.0	0.48	0.24	ug/L	07/20/12	07/20/12
EPA 8260B	TOLUENE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
EPA 8260B	TRANS-1,2-DICHLOROETHENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	TRICHLOROETHENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
EPA 8260B	VINYL CHLORIDE	0.46 U	1.0	0.46	0.23	ug/L	07/20/12	07/20/12
EPA 8260B	XYLENES (TOTAL)	0.43 J	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	SURROGATE: 1,2-DICHLOROETHANE	99.0	70-120			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: 4-BROMOFLUOROBEN	100	75-120			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: DIBROMOFLUOROMET	99.3	85-115			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: TOLUENE-D8 (S)	98.7	85-120			%	07/20/12	07/20/12

J = Estimated value.

Quant Method: TALLW.M
Run #: 0719T47
Instrument: Thor
Sequence: T120719
Dilution Factor: 1
Initials: ARS

Printed: 07/31/12 1:13:46 PM
APPL-F1-SC-NoMC-REG MDLs

Data File : M:\THOR\DATA\T120719\0719T47.D
 Acq On : 20 Jul 12 6:26
 Sample : AY65113W01
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 47
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 24 12:21 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 10:40:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.72	96	461440	25.00000	ppb	-0.01
55) Chlorobenzene-D5 (IS)	9.87	117	370688	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	220608	25.00000	ppb	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
31) Dibromofluoromethane(S)	5.93	111	228631	31.66228	ppb	-0.01
Spiked Amount				31.881		
				Recovery =	99.312%	
36) 1,2-DCA-D4(S)	6.32	65	223642	33.32593	ppb	-0.01
Spiked Amount				33.647		
				Recovery =	99.047%	
56) Toluene-D8(S)	8.43	98	807883	36.86487	ppb	0.00
Spiked Amount				37.345		
				Recovery =	98.715%	
64) 4-Bromofluorobenzene(S)	11.05	95	305752	29.50187	ppb	0.00
Spiked Amount				29.515		
				Recovery =	99.955%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
45) Methyl Cyclohexane	7.35	83	1864	0.46377	ppb	N.T 85
62) o-Xylene	10.53	106	5109	0.43122	ppb	✓ 90
71) Isopropylbenzene	10.91	105	26568	0.92106	ppb	N.T 98
76) n-Propylbenzene	11.32	91	19942	0.53696	ppb	99
81) Tert-Butylbenzene	11.82	119	16941	0.69946	ppb	90
83) Sec-Butylbenzene	12.04	105	9972	0.30839	ppb	98
94) Naphthalene	14.43	128	952297	42.68779	ppb	99

Handwritten signature and date:
 8/9/12


Data File : M:\THOR\DATA\T120719\0719T47.D
Acq On : 20 Jul 12 6:26
Sample : AY65113W01
Misc : 10ml w/5ul of IS&S: 06-7-12


Vial: 47
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 12:21 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 10:40:23 2012
Response via : Initial Calibration









#45
 Methyl Cyclohexane
 Concen: 0.46377 ppb
 RT: 7.35 min Scan# 1917
 Delta R.T. -0.01 min
 Lab File: 0719T47.D
 Acq: 20 Jul 12 6:26

*N.T.
 ADD
 8/6/12*

Tgt Ion: 83 Resp: 1864




Ion	Ratio	Lower	Upper
83	100		
55	69.4	57.8	107.3



#62
 o-Xylene
 Concen: 0.43122 ppb
 RT: 10.53 min Scan# 2906
 Delta R.T. -0.00 min
 Lab File: 0719T47.D
 Acq: 20 Jul 12 6:26

Tgt Ion: 106 Resp: 5109

Ion	Ratio	Lower	Upper
106	100		
91	193.9	146.7	272.5



EPA 8260B VOCS + GAS WATER

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

Attn: Max Solmssen
Project: LTM Red Hill / 1022-024

Sample ID: ES082
Sample Collection Date: 07/18/12

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 68258
APPL ID: AY65113
QCG: #86RHB-120719AT2-169517

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8260B	GASOLINE	12.12 U MI1	20.0	12.12	6.06	ug/L	07/20/12	07/20/12
EPA 8260B	SURROGATE: 1,2-DICHLOROETHANE	99.0	70-120			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: 4-BROMOFLUOROBEN	100	75-120			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: DIBROMOFLUOROMET	99.3	85-115			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: TOLUENE-D8 (S)	98.7	85-120			%	07/20/12	07/20/12

U = The analyte was analyzed for, but not detected. The associated numerical value is at or below the MDL.
(MI1) Manual integration: Integration does not follow baseline.

Quant Method: TALLW.M
Run #: 0719T47
Instrument: Thor
Sequence: T120719
Dilution Factor: 1
Initials: ARS

Printed: 07/31/12 1:07:11 PM
APPL-F1-SC-NoMC-REG MDLs

Data File : M:\THOR\DATA\T120719\0719T47.D Vial: 47
 Acq On : 20 Jul 12 6:26 Operator: DG,RS,HW,ARS,SV
 Sample : AY65113W01 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 26 15:50 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.72	TIC	945344	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	1059323	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1222853	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.43	TIC	12330684m	144.07139	ppb	ND 100

*No gasoline pattern detected.
ARS 7/26/12*

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T47.D
Acq On : 20 Jul 12 6:26
Sample : AY65113W01
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 47
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 26 15:50 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T47.D
 Acq On : 20 Jul 12 6:26
 Sample : AY65113W01
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 15:40 2012

Vial: 47
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Single Level Calibration



TIC: 0719T47.D

(2) Gasoline (TMHB)

8.50min 72.8679ppb m

response 9430289



Ion	Exp%	Act%
TIC	100	100
0.00	0.60	1.27#
0.00	1.80	3.65#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T47.D
 Acq On : 20 Jul 12 6:26
 Sample : AY65113W01
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 15:50 2012

Vial: 47
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Single Level Calibration



TIC: 0719T47.D

(2) Gasoline (TMHB)		
8.43min	144.0714ppb m	
response	12330684	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.97#
0.00	1.80	2.79#
0.00	0.00	0.00

EPA 8260B VOCs + Gas Water

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Attn: Max Solmssen

Project: LTM Red Hill / 1022-024

Sample ID: **ES085-TRIP BLANK**

Sample Collection Date: 07/18/12

ARF: 68258

APPL ID: **AY65114**

QCG: #86RHB-120719AT1-169441

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8260B	1,1,1,2-TETRACHLOROETHANE	0.26 U	1.0	0.26	0.13	ug/L	07/20/12	07/20/12
EPA 8260B	1,1,1-TRICHLOROETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	1,1,2,2-TETRACHLOROETHANE	0.20 U	1.0	0.20	0.10	ug/L	07/20/12	07/20/12
EPA 8260B	1,1,2-TRICHLOROETHANE	0.40 U	1.0	0.40	0.20	ug/L	07/20/12	07/20/12
EPA 8260B	1,1-DICHLOROETHANE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	1,1-DICHLOROETHENE	0.60 U	1.0	0.60	0.30	ug/L	07/20/12	07/20/12
EPA 8260B	1,2,3-TRICHLOROPROPANE	0.78 U	2.0	0.78	0.39	ug/L	07/20/12	07/20/12
EPA 8260B	1,2,4-TRICHLOROBENZENE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DIBROMO-3-CHLOROPROPANE	1.52 U	2.0	1.52	0.76	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DIBROMOETHANE	0.40 U	1.0	0.40	0.20	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DICHLOROBENZENE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DICHLOROETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	1,2-DICHLOROPROPANE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
EPA 8260B	1,3-DICHLOROBENZENE	0.22 U	1.0	0.22	0.11	ug/L	07/20/12	07/20/12
EPA 8260B	1,3-DICHLOROPROPENE, TOTAL	0.36 U	1.0	0.36	0.18	ug/L	07/20/12	07/20/12
EPA 8260B	1,4-DICHLOROBENZENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	2-BUTANONE	1.20 U	10.0	1.20	0.60	ug/L	07/20/12	07/20/12
EPA 8260B	4-METHYL-2-PENTANONE	3.80 U	10.0	3.80	1.90	ug/L	07/20/12	07/20/12
EPA 8260B	ACETONE	1.90 U	10.0	1.90	0.95	ug/L	07/20/12	07/20/12
EPA 8260B	BENZENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
EPA 8260B	BROMODICHLOROMETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	BROMOFORM	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
EPA 8260B	BROMOMETHANE	0.48 U	2.0	0.48	0.24	ug/L	07/20/12	07/20/12
EPA 8260B	CARBON TETRACHLORIDE	0.20 U	1.0	0.20	0.10	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROBENZENE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
EPA 8260B	CHLORODIBROMOMETHANE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROETHANE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROFORM	0.14 U	1.0	0.14	0.07	ug/L	07/20/12	07/20/12
EPA 8260B	CHLOROMETHANE	0.84 U	1.0	0.84	0.42	ug/L	07/20/12	07/20/12
EPA 8260B	CIS-1,2-DICHLOROETHENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
EPA 8260B	ETHYLBENZENE	0.46 U	1.0	0.46	0.23	ug/L	07/20/12	07/20/12
EPA 8260B	GASOLINE	12.12 U MI1	20.0	12.12	6.06	ug/L	07/20/12	07/20/12
EPA 8260B	HEXACHLOROBUTADIENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	METHYL TERT-BUTYL ETHER	0.52 U	1.0	0.52	0.26	ug/L	07/20/12	07/20/12

J = Estimated value.

U = The analyte was analyzed for, but not detected. The associated numerical value is at or below the MDL.

(MI1) Manual integration: Integration does not follow baseline.

Quant Method: TALLW.M
Run #: 0719T42
Instrument: Thor
Sequence: T120719
Dilution Factor: 1
Initials: ARS

Printed: 07/31/12 9:50:46 AM
APPL-F1-SC-NoMC-REG MDLs

EPA 8260B VOCs + Gas Water

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

Attn: Max Solmssen

Project: LTM Red Hill / 1022-024

Sample ID: ES085-TRIP BLANK

Sample Collection Date: 07/18/12

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 68258

APPL ID: AY65114

QCG: #86RHB-120719AT1-169441

Method	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
EPA 8260B	METHYLENE CHLORIDE	0.80 J	5.0	0.70	0.35	ug/L	07/20/12	07/20/12
EPA 8260B	STYRENE	0.50 U	1.0	0.50	0.25	ug/L	07/20/12	07/20/12
EPA 8260B	TETRACHLOROETHENE	0.48 U	1.0	0.48	0.24	ug/L	07/20/12	07/20/12
EPA 8260B	TOLUENE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
EPA 8260B	TRANS-1,2-DICHLOROETHENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	TRICHLOROETHENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
EPA 8260B	VINYL CHLORIDE	0.46 U	1.0	0.46	0.23	ug/L	07/20/12	07/20/12
EPA 8260B	XYLENES (TOTAL)	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
EPA 8260B	SURROGATE: 1,2-DICHLOROETHANE	99.9	70-120			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: 4-BROMOFLUOROBEN	99.3	75-120			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: DIBROMOFLUOROMET	98.6	85-115			%	07/20/12	07/20/12
EPA 8260B	SURROGATE: TOLUENE-D8 (S)	99.2	85-120			%	07/20/12	07/20/12

J = Estimated value.

U = The analyte was analyzed for, but not detected. The associated numerical value is at or below the MDL.

(MI1) Manual integration: Integration does not follow baseline.

Quant Method: TALLW.M
Run #: 0719T42
Instrument: Thor
Sequence: T120719
Dilution Factor: 1
Initials: ARS

Printed: 07/31/12 9:50:46 AM
APPL-F1-SC-NoMC-REG MDLs

Data File : M:\THOR\DATA\T120719\0719T42.D
 Acq On : 20 Jul 12 4:08
 Sample : AY65114W01
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 42
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 24 12:13 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 10:40:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	96	437504	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.88	117	353408	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	204928	25.00000	ppb	0.00
System Monitoring Compounds						
31) Dibromofluoromethane(S)	5.95	111	215125	31.42180	ppb	0.00
Spiked Amount				31.881		
					Recovery =	98.560%
36) 1,2-DCA-D4(S)	6.33	65	213776	33.59859	ppb	0.00
Spiked Amount				33.647		
					Recovery =	99.858%
56) Toluene-D8(S)	8.43	98	774255	37.05787	ppb	0.00
Spiked Amount				37.345		
					Recovery =	99.232%
64) 4-Bromofluorobenzene(S)	11.05	95	289700	29.31979	ppb	0.00
Spiked Amount				29.515		
					Recovery =	99.338%
Target Compounds						Qvalue
18) Methylene chloride	3.45	84	2778	0.79676	ppb	99


Data File : M:\THOR\DATA\T120719\0719T42.D
Acq On : 20 Jul 12 4:08
Sample : AY65114W01
Misc : 10ml w/5ul of IS&S: 06-7-12


Vial: 42
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 12:13 2012


Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 10:40:23 2012
Response via : Initial Calibration







#18
 Methylene chloride
 Concen: 0.79676 ppb
 RT: 3.45 min Scan# 703
 Delta R.T. -0.00 min
 Lab File: 0719T42.D
 Acq: 20 Jul 12 4:08



Tgt Ion: 84 Resp: 2778
 Ion Ratio Lower Upper
 84 100
 86 66.7 47.5 88.3



Data File : M:\THOR\DATA\T120725\0725T22.D Vial: 21
 Acq On : 25 Jul 12 19:09 Operator: DG,RS,HW,ARS,SV
 Sample : AY65114W02 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 26 8:10 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	780377	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	884072	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	988550	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.43	TIC	10211299m	21.67195	ppb	ND 100

*No gasoline pattern detected.
 MRS 7/26/12*

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T22.D
Acq On : 25 Jul 12 19:09
Sample : AY65114W02
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 21
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 26 8:10 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T22.D
 Acq On : 25 Jul 12 19:09
 Sample : AY65114W02
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 8:09 2012

Vial: 21
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Multiple Level Calibration



TIC: 0725T22.D

(2) Gasoline (TMHB)

8.50min -55.2119ppb m

response 7596711



Ion	Exp%	Act%
TIC	100	100
0.00	0.00	1.25#
0.00	0.00	3.64#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T22.D
 Acq On : 25 Jul 12 19:09
 Sample : AY65114W02
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 8:10 2012

Vial: 21
 Operator: DG, RS, HW, ARS, SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Multiple Level Calibration



TIC: 0725T22.D

(2) Gasoline (TMHB)

8.43min 21.6719ppb m

response 10211299

Ion	Exp%	Act%
TIC	100	100
0.00	0.00	0.93#
0.00	0.00	2.71#
0.00	0.00	0.00

**EPA METHOD 8260B
Volatile Organic Compounds
Calibration Data**



VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 6
Initial Calibration

Lab Name: APPL, Inc.

SDG No: 68258

Case No:

Initial Cal. Date: 07/19/12

Matrix: Water

Instrument: Thor (TALLW.M)

Initials: _____

0719T05.D 0719T06.D 0719T07.D 0719T08.D 0719T09.D 0719T10.D 0719T11.D 0719T12.D 0719T13.D

	Compound	0.3	0.5	1	2	5	10	20	40	100		Avg	%RSD		r2	
1	I Fluorobenzene (IS)															
2	TM Dichlorodifluoromethane	0.1200	0.1240	0.1418	0.1356	0.1268	0.1115					0.13	8.6	TM		
3	TML Freon 114	0.1415	0.1046	0.1361	0.1772	0.1673	0.1621	0.1749	0.1896	0.1665		0.16	17	TML	0.997	
4	TM**L Chloromethane		0.4629	0.4159	0.3768	0.3605	0.2990	0.3105				0.37	17	TM**L	0.998	
5	TM* Vinyl chloride	0.5005	0.4852	0.4531	0.5101	0.5250	0.4769	0.4974	0.4968	0.5019		0.49	4.2	TM*		
6	TM Bromomethane		0.3766	0.3596	0.3226	0.3472	0.3041	0.2931	0.2684	0.2549		0.32	14	TM		
7	TM Chloroethane	0.2747	0.3114	0.2850	0.2613	0.2962	0.2719	0.2859	0.2914	0.2834		0.28	5.1	TM		
8	TMQ Dichlorofluoromethane	0.0090	0.0202	0.0126	0.0164	0.0152	0.0196	0.0254	0.0336	0.0648		0.02	70	TMQ	1.000	
9	TM Trichlorofluoromethane		0.0842	0.0966	0.1000	0.1198	0.1100					0.10	13	TM		
10	TMQ Acrolein													TMQ		
11	TML Acetone	0.3965	0.2944	0.1679	0.1468	0.1007	0.0866	0.0926	0.0801	0.0821		0.16	70	TML	0.999	
12	TM Freon-113	0.1778	0.1747	0.1931	0.2265	0.2246	0.2168	0.2084	0.2210	0.2060		0.21	9.5	TM		
13	TM* 1,1-DCE	0.3007	0.2729	0.2644	0.2720	0.2744	0.2696	0.2660	0.2835	0.2775		0.28	4.0	TM*		
14	TM t-Butanol		0.0075	0.0074	0.0072	0.0076	0.0077	0.0092	0.0102			0.01	14	TM		
15	TML Methyl Acetate	0.8726	0.6035	0.5101	0.4714	0.2802	0.2330	0.2248	0.2202	0.2132		0.40	57	TML	1.000	
16	TM Iodomethane	0.2492	0.2575	0.2660	0.2419	0.2622	0.2432	0.2408	0.2414	0.2418		0.25	4.0	TM		
17	TM Acrylonitrile	0.0874	0.0549	0.0692	0.0728	0.0933	0.0806	0.0840	0.0846	0.0838		0.08	15	TM		
18	TML Methylene chloride	0.3676	0.2647	0.1440	0.1292	0.1165	0.0964	0.0956	0.0949	0.0918		0.16	62	TML	1.000	
19	TML Carbon disulfide	0.0466	0.0453	0.0322	0.0318	0.0301	0.0270	0.0292	0.0278	0.0258		0.03	23	TML	0.999	
20	TM Methyl t-butyl ether (MtBE)	0.6119	0.5783	0.5229	0.5275	0.5662	0.5222	0.5070	0.4911	0.4631		0.53	8.6	TM		
21	TM Trans-1,2-DCE	0.2297	0.2354	0.1695	0.1842	0.1947	0.1734	0.1779	0.1766	0.1709		0.19	13	TM		
22	TM Diisopropyl Ether	0.1003	0.1361	0.1126	0.1164	0.1317	0.1212	0.1179	0.1198	0.1168		0.12	8.7	TM		
23	TM** 1,1-DCA	0.5526	0.4780	0.4682	0.4954	0.5506	0.5086	0.4958	0.5067	0.4843		0.50	5.9	TM**		
24	TM Vinyl Acetate	0.2861	0.3189	0.2551	0.2716	0.3128	0.2787	0.2776	0.2848	0.2788		0.28	6.9	TM		
25	TM Ethyl tert Butyl Ether	0.7356	0.6599	0.6680	0.6558	0.7593	0.6522	0.6460	0.6381	0.5738		0.67	8.2	TM		
26	TML MEK (2-Butanone)	0.2041	0.1795	0.1594	0.1391	0.1113	0.1135	0.1203	0.1216	0.1272		0.14	23	TML	1.000	
27	TM Cis-1,2-DCE	0.3407	0.3327	0.3111	0.3113	0.3441	0.3184	0.3199	0.3183	0.3119		0.32	4.0	TM		
28	TM 2,2-Dichloropropane		0.2047	0.2143	0.2086	0.2158	0.2037	0.1976	0.1966	0.1845		0.20	5.0	TM		
29	TM* Chloroform	0.7181	0.6349	0.6028	0.6230	0.6647	0.6037	0.5996	0.6037	0.5876		0.63	6.6	TM*		
30	TM Bromochloromethane	0.1510	0.1457	0.1513	0.1633	0.1810	0.1595	0.1543	0.1538	0.1561		0.16	6.5	TM		
31	S Dibromofluoromethane(S)	0.5051	0.3961	0.3766	0.3856	0.3635	0.3650	0.3693	0.3784	0.3815		0.39	11	S		
32	TM 1,1,1-TCA	0.4433	0.4047	0.3367	0.3737	0.3876	0.3695	0.3618	0.3671	0.3480		0.38	8.5	TM		
33	TM Cyclohexane	0.1075	0.0984	0.0973	0.1087	0.1023	0.1026	0.0982	0.1080	0.0976		0.10	4.6	TM		
34	TM 1,1-Dichloropropene	0.2952	0.2623	0.2578	0.2666	0.2963	0.2712	0.2714	0.2756	0.2672		0.27	4.9	TM		
35	TM 2,2,4-Trimethylpentane	0.4193	0.3897	0.3643	0.3992	0.4174	0.3920	0.3860	0.4075	0.3655		0.39	5.1	TM		

NT

ARS 7/27/12

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 6
Initial Calibration

Lab Name: APPL, Inc. _____
Case No: _____
Matrix: Water

SDG No: 68258
Initial Cal. Date: 07/19/12
Instrument: Thor (TALLW.M)

Initials: _____

		Compound	0.3	0.5	1	2	5	10	20	40	100		Avg	%RSD		r2
36	S	1,2-DCA-D4(S)	0.4628	0.4074	0.3542	0.3471	0.3446	0.3369	0.3408	0.3392	0.3393		0.36	12	S	
37	TM	Carbon Tetrachloride	0.3583	0.3585	0.3356	0.3521	0.3730	0.3393	0.3467	0.3628	0.3533		0.35	3.3	TM	
38	TM	Tert Amyl Methyl Ether	0.8104	0.7185	0.6881	0.6883	0.7716	0.7091	0.6860	0.6745	0.6278		0.71	7.6	TM	
39	TM	1,2-DCA	0.4427	0.3989	0.4073	0.4206	0.4411	0.4013	0.4023	0.3957	0.3873		0.41	4.8	TM	
40	TM	Benzene	1.330	1.229	1.037	1.104	1.170	1.075	1.059	1.066	1.029		1.1	9.1	TM	
41	TM	TCE	0.3220	0.3196	0.3000	0.3036	0.3233	0.2996	0.2973	0.2966	0.2829		0.30	4.5	TM	
42	TM	2-Pentanone	0.2622	0.2195	0.2350	0.2329	0.2443	0.2273	0.2432	0.2426	0.2555		0.24	5.6	TM	
43	TM*	1,2-Dichloropropane	0.3696	0.3784	0.3475	0.3772	0.4017	0.3593	0.3586	0.3516	0.3511		0.37	4.8	TM*	
44	TM	Bromodichloromethane	0.5464	0.5022	0.4808	0.4813	0.5587	0.4945	0.4955	0.4994	0.4996		0.51	5.4	TM	
45	TM	Methyl Cyclohexane	0.2160	0.2253	0.1988	0.2361	0.2203	0.2228	0.2114	0.2222	0.2069		0.22	5.1	TM	
46	TM	Dibromomethane	0.2224	0.1871	0.1941	0.1966	0.2119	0.1946	0.1959	0.1962	0.1934		0.20	5.5	TM	
47	TML	2-Chloroethyl vinyl ether			0.0023	0.0050	0.0079	0.0074	0.0077	0.0064	0.0063		0.01	32	TML	0.998
48	TM	MIBK (methyl isobutyl ketone)	0.1836	0.1952	0.1595	0.1697	0.1709	0.1619	0.1669	0.1687	0.1789		0.17	6.5	TM	
49	TM	1-Bromo-2-chloroethane	0.2615	0.2802	0.2197	0.2499	0.2843	0.2418	0.2534	0.2519	0.2500		0.25	7.6	TM	
50	TM	Cis-1,3-Dichloropropene	0.5288	0.5220	0.4508	0.4775	0.5235	0.4876	0.4899	0.5108	0.5199		0.50	5.3	TM	
51	TM*	Toluene	1.349	1.351	1.277	1.293	1.418	1.314	1.307	1.310	1.296		1.3	3.2	TM*	
52	TM	Trans-1,3-Dichloropropene	0.5060	0.4246	0.3998	0.3819	0.4704	0.4238	0.4402	0.4550	0.4756		0.44	8.8	TM	
53	TM	1,1,2-TCA	0.3231	0.2925	0.2917	0.2877	0.3215	0.2847	0.2839	0.2826	0.2852		0.29	5.4	TM	
54	TM	2-Hexanone	0.2109	0.1986	0.1812	0.1996	0.1958	0.1884	0.1957	0.2026	0.2106		0.20	4.8	TM	
55	I	Chlorobenzene-D5 (IS)														
56	S	Toluene-D8(S)	1.945	1.553	1.390	1.493	1.349	1.331	1.429	1.411	1.401		1.5	13	S	
57	TM	1,2-EDB	0.4293	0.3708	0.3376	0.3631	0.4033	0.3618	0.3677	0.3665	0.3733		0.37	7.1	TM	
58	TM	Tetrachloroethene	0.5273	0.3800	0.4081	0.4402	0.4287	0.4130	0.4140	0.4108	0.3923		0.42	10	TM	
59	TM	1-Chlorohexane		0.4233	0.5404	0.5484	0.5087	0.4910	0.5018	0.5163	0.5060		0.50	7.5	TM	
60	TM	1,1,1,2-Tetrachloroethane	0.5093	0.4750	0.4800	0.4853	0.5228	0.4867	0.5034	0.4895	0.5042		0.50	3.2	TM	
61	TM	m&p-Xylene	0.7775	0.7225	0.7109	0.7529	0.8468	0.7684	0.7958	0.7959	0.7812		0.77	5.3	TM	
62	TM	o-Xylene	0.8379	0.7723	0.6766	0.7783	0.8551	0.8049	0.8293	0.8221	0.8148		0.80	6.6	TM	
63	TM	Styrene	1.301	1.205	1.181	1.300	1.477	1.358	1.464	1.458	1.474		1.4	8.6	TM	
64	S	4-Bromofluorobenzene(S)	0.8941	0.6924	0.6735	0.7021	0.6390	0.6351	0.6830	0.6795	0.6919		0.70	11	S	
65	TM	1,3-Dichloropropane	0.6702	0.6903	0.6119	0.6806	0.6923	0.6397	0.6502	0.6458	0.6339		0.66	4.2	TM	
66	TM	Dibromochloromethane	0.5324	0.4622	0.4816	0.4777	0.5198	0.4771	0.5014	0.4950	0.5058		0.49	4.5	TM	
67	TM**	Chlorobenzene	1.394	1.309	1.286	1.325	1.349	1.240	1.255	1.250	1.223		1.3	4.4	TM**	
68	TM*	Ethylbenzene	2.124	2.073	1.840	2.023	2.142	1.972	2.058	2.044	2.014		2.0	4.4	TM*	
69	TM**	Bromoform	0.3594	0.3044	0.3153	0.3283	0.3636	0.3252	0.3395	0.3493	0.3641		0.34	6.4	TM**	
70	I	1,4-Dichlorobenzene-D (IS)														

ANLS7/27/12

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 6
Initial Calibration

Lab Name: APPL, Inc. _____
Case No: _____
Matrix: Water _____

SDG No: 68258
Initial Cal. Date: 07/19/12
Instrument: Thor (TALLW.M)

Initials: _____

		Compound	0.3	0.5	1	2	5	10	20	40	100		Avg	%RSD		r2
71	TM	Isopropylbenzene	3.467	3.050	3.049	3.087	3.489	3.289	3.342	3.375	3.271		3.3	5.2	TM	
72	TM**	1,1,2,2-Tetrachloroethane	0.8851	0.9679	0.8759	0.9162	0.9988	0.8592	0.8910	0.8858	0.8834		0.91	5.1	TM**	
73	TM	1,2,3-Trichloropropane	0.3011	0.2573	0.2216	0.2692	0.2788	0.2452	0.2476	0.2483	0.2478		0.26	8.9	TM	
74	TM	t-1,4-Dichloro-2-Butene	0.1233	0.1619	0.1587	0.1593	0.1955	0.1733	0.1868	0.1920	0.1997		0.17	14	TM	
75	TM	Bromobenzene	1.091	1.144	1.005	1.075	1.155	1.068	1.071	1.055	1.035		1.1	4.4	TM	
76	TM	n-Propylbenzene	4.174	3.908	3.893	4.133	4.515	4.215	4.378	4.400	4.261		4.2	5.0	TM	
77	TM	4-Ethyltoluene	3.403	3.468	3.298	3.466	3.887	3.743	3.772	3.801	3.689		3.6	5.7	TM	
78	TM	2-Chlorotoluene	3.081	2.980	2.812	2.959	3.223	3.008	3.013	3.014	2.922		3.0	3.7	TM	
79	TM	1,3,5-Trimethylbenzene	2.835	2.688	2.726	2.902	3.286	3.099	3.174	3.182	3.072		3.0	7.2	TM	
80	TM	4-Chlorotoluene	2.900	2.859	2.765	2.855	3.310	3.011	3.062	3.033	2.941		3.0	5.4	TM	
81	TM	Tert-Butylbenzene	2.860	2.656	2.519	2.623	2.937	2.735	2.783	2.826	2.764		2.7	4.7	TM	
82	TM	1,2,4-Trimethylbenzene	3.036	2.836	2.905	2.957	3.327	3.187	3.242	3.250	3.161		3.1	5.6	TM	
83	TM	Sec-Butylbenzene	3.394	3.341	3.380	3.572	4.054	3.748	3.858	3.877	3.756		3.7	6.9	TM	
84	TM	p-Isopropyltoluene	2.818	2.824	2.797	3.020	3.405	3.187	3.271	3.320	3.225		3.1	7.6	TM	
85	TM	Benzyl Chloride	1.053	0.8317	0.9028	0.8503	0.9797	0.8739	0.8908	0.9346	1.011		0.93	8.1	TM	
86	TM	1,3-DCB	2.012	2.075	1.942	2.040	2.212	2.027	2.054	2.010	1.970		2.0	3.8	TM	
87	TM	1,4-DCB	2.332	2.267	2.134	2.043	2.254	2.079	2.072	2.042	1.986		2.1	5.7	TM	
88	TM	n-Butylbenzene	2.593	2.637	2.640	2.648	2.950	2.837	2.897	2.936	2.840		2.8	5.2	TM	
89	TM	1,2-DCB	2.109	2.010	1.946	1.881	2.124	1.970	1.946	1.916	1.874		2.0	4.6	TM	
90	TM	Hexachloroethane	0.6385	0.6103	0.5276	0.5154	0.5816	0.5288	0.5569	0.5673	0.5792		0.57	7.1	TM	
91	TM	1,2-Dibromo-3-chloropropane	0.1427	0.1282	0.1718	0.1498	0.1896	0.1710	0.1873	0.1886	0.2003		0.17	14	TM	
92	TM	1,2,4-Trichlorobenzene	0.9309	0.8325	0.8167	0.8383	0.9761	0.9144	0.9363	0.9319	0.9714		0.91	6.7	TM	
93	TM	Hexachlorobutadiene	0.4199	0.3460	0.4009	0.3612	0.4008	0.3697	0.3737	0.3684	0.3634		0.38	6.3	TM	
94	TM	Naphthalene	2.301	2.300	2.079	2.264	2.715	2.596	2.749	2.843	2.906		2.5	12	TM	
95	TM	1,2,3-Trichlorobenzene	1.232	1.180	1.271	1.218	1.424	1.312	1.335	1.325	1.313		1.3	5.7	TM	
96																
97																
98																
99																
100																
101																
102																
103																
104																
105																

ARS 7/27/12

Data File : M:\THOR\DATA\T120719\0719T05.D
 Acq On : 19 Jul 12 11:01
 Sample : 0.3ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 5
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) Fluorobenzene (IS)	6.74	96	427072	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.88	117	343424	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	202048	25.00000	ppb	0.00

System Monitoring Compounds

31) Dibromofluoromethane(S)	5.95	111	5177	0.77464	ppb	0.00
Spiked Amount	29.744		Recovery	=	2.606%	
36) 1,2-DCA-D4(S)	6.33	65	4744	0.76381	ppb	0.00
Spiked Amount	29.083		Recovery	=	2.627%	
56) Toluene-D8(S)	8.43	98	16030	0.78954	ppb	0.00
Spiked Amount	30.231		Recovery	=	2.613%	
64) 4-Bromofluorobenzene(S)	11.05	95	7369	0.76748	ppb	0.00
Spiked Amount	28.321		Recovery	=	2.708%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.30	85	615	0.27612	ppb	95
3) Freon 114	1.42	85	725	-0.13478	ppb #	68
4) Chloromethane	1.45	50	2617	0.40957	ppb	87
5) Vinyl chloride	1.56	62	2565	0.30389	ppb	92
6) Bromomethane	1.87	94	2507	0.46469	ppb	98
7) Chloroethane	1.98	64	1408	0.28963	ppb	99
9) Trichlorofluoromethane	2.24	101	407	0.23329	ppb	93
11) Acetone	2.91	43	2032	0.21001	ppb #	82
12) Freon-113	2.85	101	911	0.25960	ppb #	62
13) 1,1-DCE	2.82	61	1541	0.32723	ppb #	78
14) t-Butanol	3.69	59	2213	15.97947	ppb	92
15) Methyl Acetate	3.35	43	4472	-0.20827	ppb #	84
16) Iodomethane	2.99	142	1277	0.29980	ppb #	77
17) Acrylonitrile	3.84	52	448	0.33215	ppb #	42
18) Methylene chloride	3.45	84	1884	0.26546	ppb	79
19) Carbon disulfide	3.07	76	239	-0.34303	ppb #	65
20) Methyl t-butyl ether (MtBE)	3.91	73	3136	0.34491	ppb #	79
21) Trans-1,2-DCE	3.87	96	1177	0.36216	ppb #	64
22) Diisopropyl Ether	4.71	59	514	0.25243	ppb #	40
23) 1,1-DCA	4.51	63	2832	0.32862	ppb #	79
24) Vinyl Acetate	4.71	87	1466	0.30118	ppb	75
25) Ethyl tert Butyl Ether	5.21	59	3770	0.33165	ppb	100
26) MEK (2-Butanone)	5.40	43	1046	0.82334	ppb	91
27) Cis-1,2-DCE	5.32	96	1746	0.31627	ppb	76
28) 2,2-Dichloropropane	5.32	77	1192	0.74979	ppb	93
29) Chloroform	5.76	83	3680	0.34387	ppb	87
30) Bromochloromethane	5.63	128	774	0.28796	ppb	75
32) 1,1,1-TCA	5.96	97	2272	0.35284	ppb	87
33) Cyclohexane	6.03	41	551	0.31531	ppb #	6
34) 1,1-Dichloropropene	6.16	75	1513	0.32355	ppb #	82
35) 2,2,4-Trimethylpentane	6.55	57	2149	0.31975	ppb	81
37) Carbon Tetrachloride	6.16	117	1836	0.30422	ppb	83
38) Tert Amyl Methyl Ether	6.59	73	4153	0.34325	ppb #	92
39) 1,2-DCA	6.42	62	2269	0.32331	ppb #	74
40) Benzene	6.40	78	6818	0.35570	ppb	94
41) TCE	7.14	95	1650	0.31670	ppb	90
42) 2-Pentanone	7.36	43	67186	16.36852	ppb	95
43) 1,2-Dichloropropane	7.37	63	1894	0.30283	ppb #	85
44) Bromodichloromethane	7.68	83	2800	0.32362	ppb	87
45) Methyl Cyclohexane	7.36	83	1107	0.29759	ppb #	41
46) Dibromomethane	7.50	93	1140	0.33509	ppb #	65

Data File : M:\THOR\DATA\T120719\0719T05.D
 Acq On : 19 Jul 12 11:01
 Sample : 0.3ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 5
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
48) MIBK (methyl isobutyl ket	8.33	43	941	0.31878	ppb	# 95
49) 1-Bromo-2-chloroethane	7.99	63	1340	0.30794	ppb	93
50) Cis-1,3-Dichloropropene	8.15	75	2710	0.31652	ppb	96
51) Toluene	8.50	91	6911	0.30559	ppb	97
52) Trans-1,3-Dichloropropene	8.72	75	2593	0.34348	ppb	94
53) 1,1,2-TCA	8.90	83	1656	0.32887	ppb	92
54) 2-Hexanone	9.18	43	1081	0.31934	ppb	# 88
57) 1,2-EDB	9.40	107	1769	0.34356	ppb	92
58) Tetrachloroethene	9.06	166	2173	0.37324	ppb	90
59) 1-Chlorohexane	9.90	91	3169	0.45728	ppb	98
60) 1,1,1,2-Tetrachloroethane	9.99	131	2099	0.30859	ppb	97
61) m&p-Xylene	10.14	106	6408	0.60392	ppb	98
62) o-Xylene	10.54	106	3453	0.31458	ppb	79
63) Styrene	10.55	104	5361	0.28746	ppb	90
65) 1,3-Dichloropropane	9.07	76	2762	0.30594	ppb	85
66) Dibromochloromethane	9.29	129	2194	0.32279	ppb	80
67) Chlorobenzene	9.91	112	5744	0.32352	ppb	86
68) Ethylbenzene	10.03	91	8755	0.31360	ppb	98
69) Bromoform	10.71	173	1481	0.31823	ppb	97
71) Isopropylbenzene	10.91	105	8406	0.31819	ppb	96
72) 1,1,2,2-Tetrachloroethane	11.19	83	2146	0.29274	ppb	# 84
73) 1,2,3-Trichloropropane	11.22	110	730	0.35086	ppb	# 56
74) t-1,4-Dichloro-2-Butene	11.24	53	299	0.21473	ppb	# 27
75) Bromobenzene	11.19	156	2644	0.30360	ppb	94
76) n-Propylbenzene	11.32	91	10120	0.29752	ppb	93
77) 4-Ethyltoluene	11.43	105	8252	0.28250	ppb	97
78) 2-Chlorotoluene	11.40	91	7471	0.30802	ppb	92
79) 1,3,5-Trimethylbenzene	11.50	105	6874	0.28388	ppb	93
80) 4-Chlorotoluene	11.50	91	7031	0.29285	ppb	85
81) Tert-Butylbenzene	11.82	119	6934	0.31259	ppb	87
82) 1,2,4-Trimethylbenzene	11.86	105	7361	0.29378	ppb	100
83) Sec-Butylbenzene	12.04	105	8228	0.27783	ppb	96
84) p-Isopropyltoluene	12.19	119	6833	0.27307	ppb	# 88
85) Benzyl Chloride	12.36	91	2552	0.34128	ppb	95
86) 1,3-DCB	12.13	146	4878	0.29617	ppb	90
87) 1,4-DCB	12.22	146	5654	0.32779	ppb	98
88) n-Butylbenzene	12.59	91	6287	0.28031	ppb	93
89) 1,2-DCB	12.59	146	5114	0.32036	ppb	94
90) Hexachloroethane	12.86	117	1548	0.33764	ppb	85
91) 1,2-Dibromo-3-chloropropan	13.36	157	346	0.25194	ppb	83
92) 1,2,4-Trichlorobenzene	14.20	180	2257	0.30845	ppb	85
93) Hexachlorobutadiene	14.38	223	1018	0.33304	ppb	93
94) Naphthalene	14.43	128	5580	0.27311	ppb	94
95) 1,2,3-Trichlorobenzene	14.68	180	2988	0.28660	ppb	98


Data File : M:\THOR\DATA\T120719\0719T05.D
Acq On : 19 Jul 12 11:01
Sample : 0.3ug/L Vol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 5
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 08:28:10 2012
Response via : Initial Calibration



Data File : M:\THOR\DATA\T120719\0719T06.D
 Acq On : 19 Jul 12 11:29
 Sample : 0.5ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 6
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) Fluorobenzene (IS)	6.73	96	440576	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.88	117	363776	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	205952	25.00000	ppb	0.00

System Monitoring Compounds

31) Dibromofluoromethane(S)	5.95	111	6980	1.01241	ppb	0.00
Spiked Amount	29.744		Recovery	=	3.402%	
36) 1,2-DCA-D4(S)	6.33	65	7179	1.12044	ppb	0.00
Spiked Amount	29.083		Recovery	=	3.851%	
56) Toluene-D8(S)	8.43	98	22596	1.05068	ppb	0.00
Spiked Amount	30.231		Recovery	=	3.477%	
64) 4-Bromofluorobenzene(S)	11.05	95	10075	0.99060	ppb	0.00
Spiked Amount	28.321		Recovery	=	3.499%	

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.30	85	1093	0.47570	ppb	85
3) Freon 114	1.42	85	922	-0.07628	ppb	90
4) Chloromethane	1.45	50	4079	0.61881	ppb	92
5) Vinyl chloride	1.56	62	4275	0.49095	ppb	97
6) Bromomethane	1.87	94	3318	0.59617	ppb	96
7) Chloroethane	1.98	64	2744	0.54714	ppb	97
8) Dichlorofluoromethane	2.18	67	178	0.48209	ppb	# 41
9) Trichlorofluoromethane	2.24	101	742	0.41227	ppb	86
11) Acetone	2.90	43	2594	0.55965	ppb	96
12) Freon-113	2.86	101	1539	0.42512	ppb	# 88
13) 1,1-DCE	2.82	61	2405	0.49505	ppb	96
14) t-Butanol	3.69	59	3316	23.21003	ppb	95
16) Iodomethane	2.98	142	2269	0.51637	ppb	# 76
17) Acrylonitrile	3.82	52	484	0.34784	ppb	# 38
18) Methylene chloride	3.46	84	2332	0.50714	ppb	95
20) Methyl t-butyl ether (MtBE)	3.91	73	5096	0.54331	ppb	# 93
21) Trans-1,2-DCE	3.86	96	2074	0.61860	ppb	# 74
22) Diisopropyl Ether	4.70	59	1199	0.57079	ppb	# 86
23) 1,1-DCA	4.51	63	4212	0.47377	ppb	# 92
24) Vinyl Acetate	4.70	87	2810	0.55961	ppb	62
25) Ethyl tert Butyl Ether	5.21	59	5815	0.49588	ppb	97
26) MEK (2-Butanone)	5.39	43	1582	1.04859	ppb	# 79
27) Cis-1,2-DCE	5.33	96	2932	0.51483	ppb	90
28) 2,2-Dichloropropane	5.32	77	1804	1.09997	ppb	93
29) Chloroform	5.76	83	5594	0.50670	ppb	93
30) Bromochloromethane	5.62	128	1284	0.46305	ppb	86
32) 1,1,1-TCA	5.96	97	3566	0.53682	ppb	84
33) Cyclohexane	6.03	41	867	0.48093	ppb	# 20
34) 1,1-Dichloropropene	6.16	75	2311	0.47905	ppb	# 87
35) 2,2,4-Trimethylpentane	6.55	57	3434	0.49528	ppb	94
37) Carbon Tetrachloride	6.17	117	3159	0.50739	ppb	79
38) Tert Amyl Methyl Ether	6.59	73	6331	0.50723	ppb	# 88
39) 1,2-DCA	6.42	62	3515	0.48550	ppb	# 91
40) Benzene	6.40	78	10831	0.54774	ppb	95
41) TCE	7.14	95	2816	0.52393	ppb	86
42) 2-Pentanone	7.36	43	96700	22.83691	ppb	100
43) 1,2-Dichloropropane	7.37	63	3334	0.51674	ppb	# 85
44) Bromodichloromethane	7.68	83	4425	0.49576	ppb	# 92
45) Methyl Cyclohexane	7.36	83	1985	0.51726	ppb	81
46) Dibromomethane	7.50	93	1649	0.46985	ppb	78
48) MIBK (methyl isobutyl ket)	8.33	43	1720	0.56481	ppb	# 91

Data File : M:\THOR\DATA\T120719\0719T06.D
 Acq On : 19 Jul 12 11:29
 Sample : 0.5ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 6
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
49) 1-Bromo-2-chloroethane	7.99	63	2469	0.55000	ppb	98
50) Cis-1,3-Dichloropropene	8.15	75	4600	0.52080	ppb	92
51) Toluene	8.50	91	11904	0.51023	ppb	98
52) Trans-1,3-Dichloropropene	8.72	75	3741	0.48036	ppb	84
53) 1,1,2-TCA	8.90	83	2577	0.49609	ppb	92
54) 2-Hexanone	9.18	43	1750	0.50113	ppb #	95
57) 1,2-EDB	9.40	107	2698	0.49467	ppb	98
58) Tetrachloroethene	9.05	166	2765	0.44835	ppb	85
59) 1-Chlorohexane	9.90	91	3080	0.41958	ppb #	69
60) 1,1,1,2-Tetrachloroethane	9.99	131	3456	0.47967	ppb	97
61) m&p-Xylene	10.14	106	10513	0.93536	ppb	95
62) o-Xylene	10.54	106	5619	0.48328	ppb	97
63) Styrene	10.55	104	8769	0.44389	ppb	95
65) 1,3-Dichloropropane	9.07	76	5022	0.52516	ppb	95
66) Dibromochloromethane	9.29	129	3363	0.46710	ppb	98
67) Chlorobenzene	9.91	112	9525	0.50646	ppb	95
68) Ethylbenzene	10.03	91	15081	0.50998	ppb	95
69) Bromoform	10.71	173	2215	0.44932	ppb	80
71) Isopropylbenzene	10.91	105	12562	0.46649	ppb	98
72) 1,1,2,2-Tetrachloroethane	11.19	83	3987	0.53357	ppb	83
73) 1,2,3-Trichloropropane	11.23	110	1060	0.49981	ppb	90
74) t-1,4-Dichloro-2-Butene	11.25	53	667	0.46994	ppb	84
75) Bromobenzene	11.20	156	4714	0.53102	ppb	86
76) n-Propylbenzene	11.32	91	16098	0.46430	ppb	98
77) 4-Ethyltoluene	11.43	105	14285	0.47977	ppb	98
78) 2-Chlorotoluene	11.39	91	12273	0.49640	ppb	94
79) 1,3,5-Trimethylbenzene	11.50	105	11071	0.44855	ppb	96
80) 4-Chlorotoluene	11.50	91	11777	0.48124	ppb	99
81) Tert-Butylbenzene	11.82	119	10940	0.48383	ppb	97
82) 1,2,4-Trimethylbenzene	11.86	105	11683	0.45743	ppb	95
83) Sec-Butylbenzene	12.04	105	13762	0.45588	ppb	96
84) p-Isopropyltoluene	12.19	119	11631	0.45600	ppb	98
85) Benzyl Chloride	12.36	91	3426	0.44948	ppb #	91
86) 1,3-DCB	12.14	146	8549	0.50922	ppb	92
87) 1,4-DCB	12.22	146	9338	0.53111	ppb	93
88) n-Butylbenzene	12.59	91	10860	0.47502	ppb	91
89) 1,2-DCB	12.59	146	8278	0.50874	ppb	90
90) Hexachloroethane	12.86	117	2514	0.53795	ppb #	49
91) 1,2-Dibromo-3-chloropropan	13.35	157	528	0.37717	ppb	90
92) 1,2,4-Trichlorobenzene	14.20	180	3429	0.45974	ppb	97
93) Hexachlorobutadiene	14.38	223	1425	0.45735	ppb	86
94) Naphthalene	14.43	128	9474	0.45490	ppb	99
95) 1,2,3-Trichlorobenzene	14.68	180	4860	0.45732	ppb	86

Data File : M:\THOR\DATA\T120719\0719T07.D
 Acq On : 19 Jul 12 11:57
 Sample : 1.0ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 7
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) Fluorobenzene (IS)	6.74	96	442240	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.88	117	361536	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	203840	25.00000	ppb	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev (Min)
31) Dibromofluoromethane(S)	5.95	111	13324	1.92530	ppb	0.00
Spiked Amount	29.744		Recovery	=	6.472%	
36) 1,2-DCA-D4(S)	6.33	65	12530	1.94822	ppb	0.00
Spiked Amount	29.083		Recovery	=	6.698%	
56) Toluene-D8(S)	8.43	98	40197	1.88068	ppb	0.00
Spiked Amount	30.231		Recovery	=	6.222%	
64) 4-Bromofluorobenzene(S)	11.05	95	19479	1.92710	ppb	0.00
Spiked Amount	28.321		Recovery	=	6.804%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.31	85	2509	1.08786	ppb	90
3) Freon 114	1.42	85	2408	0.42008	ppb	95
4) Chloromethane	1.45	50	7357	1.11191	ppb	88
5) Vinyl chloride	1.56	62	8016	0.91712	ppb	91
6) Bromomethane	1.87	94	6361	1.13863	ppb	87
7) Chloroethane	1.98	64	5042	1.00157	ppb	95
8) Dichlorofluoromethane	2.18	67	223	0.65944	ppb	# 41
9) Trichlorofluoromethane	2.24	101	1709	0.94597	ppb	98
11) Acetone	2.90	43	2970	0.81592	ppb	89
12) Freon-113	2.86	101	3415	0.93978	ppb	# 73
13) 1,1-DCE	2.83	61	4677	0.95909	ppb	93
14) t-Butanol	3.69	59	6579	45.87583	ppb	98
15) Methyl Acetate	3.35	43	9023	0.97185	ppb	97
16) Iodomethane	2.98	142	4706	1.06694	ppb	95
17) Acrylonitrile	3.83	52	1224	0.87635	ppb	# 55
18) Methylene chloride	3.46	84	2548	0.63556	ppb	95
19) Carbon disulfide	3.07	76	570	0.36158	ppb	# 65
20) Methyl t-butyl ether (MtBE)	3.91	73	9249	0.98236	ppb	# 88
21) Trans-1,2-DCE	3.87	96	2998	0.89083	ppb	94
22) Diisopropyl Ether	4.70	59	1992	0.94474	ppb	97
23) 1,1-DCA	4.51	63	8283	0.92818	ppb	# 90
24) Vinyl Acetate	4.70	87	4513	0.89537	ppb	98
25) Ethyl tert Butyl Ether	5.21	59	11816	1.00382	ppb	97
26) MEK (2-Butanone)	5.39	43	2819	1.59789	ppb	87
27) Cis-1,2-DCE	5.33	96	5504	0.96281	ppb	89
28) 2,2-Dichloropropane	5.32	77	3790	2.30222	ppb	89
29) Chloroform	5.76	83	10664	0.96229	ppb	98
30) Bromochloromethane	5.62	128	2677	0.96179	ppb	97
32) 1,1,1-TCA	5.96	97	5956	0.89324	ppb	87
33) Cyclohexane	6.03	41	1722	0.95160	ppb	# 36
34) 1,1-Dichloropropene	6.17	75	4561	0.94189	ppb	89
35) 2,2,4-Trimethylpentane	6.56	57	6445	0.92606	ppb	83
37) Carbon Tetrachloride	6.16	117	5937	0.95000	ppb	98
38) Tert Amyl Methyl Ether	6.59	73	12173	0.97161	ppb	# 90
39) 1,2-DCA	6.42	62	7205	0.99143	ppb	93
40) Benzene	6.40	78	18340	0.92399	ppb	97
41) TCE	7.15	95	5307	0.98367	ppb	92
42) 2-Pentanone	7.36	43	207854	48.90262	ppb	98
43) 1,2-Dichloropropane	7.37	63	6147	0.94914	ppb	99
44) Bromodichloromethane	7.68	83	8505	0.94929	ppb	94
45) Methyl Cyclohexane	7.36	83	3516	0.91277	ppb	# 49

(#) = qualifier out of range (m) = manual integration
 0719T07.D TALLW.M Fri Jul 20 08:29:33 2012

Data File : M:\THOR\DATA\T120719\0719T07.D
 Acq On : 19 Jul 12 11:57
 Sample : 1.0ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 7
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
46) Dibromomethane	7.49	93	3434	0.97477	ppb	85
47) 2-Chloroethyl vinyl ether	8.00	106	40	-0.85622	ppb	100
48) MIBK (methyl isobutyl ket	8.33	43	2821	0.92287	ppb	94
49) 1-Bromo-2-chloroethane	7.99	63	3886	0.86240	ppb	98
50) Cis-1,3-Dichloropropene	8.16	75	7974	0.89940	ppb	98
51) Toluene	8.50	91	22594	0.96478	ppb	99
52) Trans-1,3-Dichloropropene	8.72	75	7072	0.90466	ppb	90
53) 1,1,2-TCA	8.90	83	5160	0.98960	ppb	93
54) 2-Hexanone	9.18	43	3205	0.91433	ppb #	88
57) 1,2-EDB	9.40	107	4882	0.90064	ppb	93
58) Tetrachloroethene	9.06	166	5902	0.96294	ppb	92
59) 1-Chlorohexane	9.90	91	7815	1.07120	ppb	90
60) 1,1,1,2-Tetrachloroethane	9.99	131	6942	0.96947	ppb	97
61) m&p-Xylene	10.15	106	20562	1.84077	ppb	98
62) o-Xylene	10.54	106	9784	0.84671	ppb	82
63) Styrene	10.55	104	17077	0.86980	ppb	96
65) 1,3-Dichloropropane	9.07	76	8849	0.93108	ppb	100
66) Dibromochloromethane	9.29	129	6965	0.97340	ppb	81
67) Chlorobenzene	9.90	112	18604	0.99534	ppb	97
68) Ethylbenzene	10.03	91	26613	0.90552	ppb	97
69) Bromoform	10.71	173	4560	0.93074	ppb	93
71) Isopropylbenzene	10.91	105	24857	0.93263	ppb	98
72) 1,1,2,2-Tetrachloroethane	11.19	83	7142	0.96570	ppb	99
73) 1,2,3-Trichloropropane	11.22	110	1807	0.86086	ppb	97
74) t-1,4-Dichloro-2-Butene	11.25	53	1294	0.92115	ppb	84
75) Bromobenzene	11.19	156	8191	0.93226	ppb	95
76) n-Propylbenzene	11.32	91	31739	0.92491	ppb	98
77) 4-Ethyltoluene	11.43	105	26890	0.91247	ppb	98
78) 2-Chlorotoluene	11.39	91	22924	0.93681	ppb	96
79) 1,3,5-Trimethylbenzene	11.49	105	22226	0.90982	ppb	99
80) 4-Chlorotoluene	11.50	91	22548	0.93091	ppb	98
81) Tert-Butylbenzene	11.82	119	20536	0.91763	ppb	94
82) 1,2,4-Trimethylbenzene	11.86	105	23690	0.93716	ppb	93
83) Sec-Butylbenzene	12.04	105	27557	0.92232	ppb	97
84) p-Isopropyltoluene	12.19	119	22802	0.90322	ppb	99
85) Benzyl Chloride	12.35	91	7361	0.97575	ppb	95
86) 1,3-DCB	12.13	146	15833	0.95287	ppb	97
87) 1,4-DCB	12.22	146	17403	1.00007	ppb	95
88) n-Butylbenzene	12.59	91	21527	0.95135	ppb	90
89) 1,2-DCB	12.59	146	15870	0.98542	ppb	99
90) Hexachloroethane	12.85	117	4302	0.93008	ppb	87
91) 1,2-Dibromo-3-chloropropan	13.35	157	1401	1.01116	ppb	85
92) 1,2,4-Trichlorobenzene	14.20	180	6659	0.90204	ppb	90
93) Hexachlorobutadiene	14.39	223	3269	1.06005	ppb	89
94) Naphthalene	14.43	128	16948	0.82221	ppb	94
95) 1,2,3-Trichlorobenzene	14.68	180	10365	0.98545	ppb	94


Data File : M:\THOR\DATA\T120719\0719T07.D
Acq On : 19 Jul 12 11:57
Sample : 1.0ug/L Vol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 7
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 08:28:10 2012
Response via : Initial Calibration



Data File : M:\THOR\DATA\T120719\0719T08.D
 Acq On : 19 Jul 12 12:25
 Sample : 2.0ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 8
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) Fluorobenzene (IS)	6.73	96	436352	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.87	117	342912	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	204992	25.00000	ppb	0.00

System Monitoring Compounds

31) Dibromofluoromethane(S)	5.95	111	26923	3.94284	ppb	0.00
Spiked Amount	29.744		Recovery	=	13.256%	
36) 1,2-DCA-D4(S)	6.33	65	24230	3.81822	ppb	0.00
Spiked Amount	29.083		Recovery	=	13.128%	
56) Toluene-D8(S)	8.43	98	81925	4.04116	ppb	0.00
Spiked Amount	30.231		Recovery	=	13.367%	
64) 4-Bromofluorobenzene(S)	11.05	95	38521	4.01794	ppb	0.00
Spiked Amount	28.321		Recovery	=	14.187%	

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.30	85	4734	2.08028	ppb	100
3) Freon 114	1.41	85	6187	1.71327	ppb	100
4) Chloromethane	1.45	50	13155	2.01503	ppb	94
5) Vinyl chloride	1.56	62	17805	2.06457	ppb	100
6) Bromomethane	1.87	94	11262	2.04311	ppb	88
7) Chloroethane	1.97	64	9122	1.83649	ppb	94
8) Dichlorofluoromethane	2.18	67	573	2.01540	ppb	91
9) Trichlorofluoromethane	2.24	101	3492	1.95899	ppb	95
11) Acetone	2.90	43	5124	2.37170	ppb	95
12) Freon-113	2.86	101	7906	2.20502	ppb	91
13) 1,1-DCE	2.82	61	9495	1.97338	ppb	91
14) t-Butanol	3.69	59	9378	66.27585	ppb	96
15) Methyl Acetate	3.34	43	16454	3.02813	ppb	94
16) Iodomethane	2.98	142	8446	1.94071	ppb	98
17) Acrylonitrile	3.81	52	2540	1.84311	ppb	82
18) Methylene chloride	3.45	84	4510	1.88941	ppb	93
19) Carbon disulfide	3.06	76	1111	1.57628	ppb	# 87
20) Methyl t-butyl ether (MtBE)	3.90	73	18413	1.98209	ppb	93
21) Trans-1,2-DCE	3.86	96	6430	1.93640	ppb	99
22) Diisopropyl Ether	4.70	59	4063	1.95295	ppb	91
23) 1,1-DCA	4.51	63	17292	1.96386	ppb	95
24) Vinyl Acetate	4.70	87	9481	1.90640	ppb	92
25) Ethyl tert Butyl Ether	5.21	59	22892	1.97102	ppb	94
26) MEK (2-Butanone)	5.39	43	4855	2.53560	ppb	91
27) Cis-1,2-DCE	5.33	96	10866	1.92643	ppb	91
28) 2,2-Dichloropropane	5.32	77	7282	4.48311	ppb	100
29) Chloroform	5.76	83	21749	1.98906	ppb	98
30) Bromochloromethane	5.62	128	5699	2.07515	ppb	93
32) 1,1,1-TCA	5.96	97	13045	1.98279	ppb	100
33) Cyclohexane	6.04	41	3794	2.12491	ppb	# 44
34) 1,1-Dichloropropene	6.17	75	9305	1.94750	ppb	97
35) 2,2,4-Trimethylpentane	6.55	57	13935	2.02928	ppb	93
37) Carbon Tetrachloride	6.16	117	12292	1.99343	ppb	89
38) Tert Amyl Methyl Ether	6.59	73	24026	1.94355	ppb	97
39) 1,2-DCA	6.42	62	14684	2.04783	ppb	98
40) Benzene	6.40	78	38526	1.96717	ppb	99
41) TCE	7.14	95	10599	1.99108	ppb	93
42) 2-Pentanone	7.36	43	304878	72.69773	ppb	99
43) 1,2-Dichloropropane	7.37	63	13169	2.06083	ppb	98
44) Bromodichloromethane	7.68	83	16800	1.90044	ppb	95
45) Methyl Cyclohexane	7.36	83	8243	2.16879	ppb	82

Data File : M:\THOR\DATA\T120719\0719T08.D
 Acq On : 19 Jul 12 12:25
 Sample : 2.0ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 8
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
46) Dibromomethane	7.49	93	6864	1.97470	ppb	89
47) 2-Chloroethyl vinyl ether	7.99	106	173	0.37233	ppb #	100
48) MIBK (methyl isobutyl ket	8.33	43	5923	1.96382	ppb #	93
49) 1-Bromo-2-chloroethane	7.99	63	8723	1.96198	ppb	97
50) Cis-1,3-Dichloropropene	8.15	75	16667	1.90525	ppb	95
51) Toluene	8.50	91	45119	1.95261	ppb	99
52) Trans-1,3-Dichloropropene	8.73	75	13333	1.72859	ppb	98
53) 1,1,2-TCA	8.90	83	10044	1.95226	ppb	94
54) 2-Hexanone	9.18	43	6966	2.01409	ppb	94
57) 1,2-EDB	9.40	107	9962	1.93762	ppb	99
58) Tetrachloroethene	9.06	166	12075	2.07710	ppb	92
59) 1-Chlorohexane	9.90	91	15043	2.17393	ppb	99
60) 1,1,1,2-Tetrachloroethane	9.99	131	13314	1.96031	ppb	97
61) m&p-Xylene	10.14	106	41306	3.89866	ppb	100
62) o-Xylene	10.54	106	21351	1.94808	ppb	92
63) Styrene	10.55	104	35671	1.91554	ppb	97
65) 1,3-Dichloropropane	9.07	76	18670	2.07113	ppb	96
66) Dibromochloromethane	9.29	129	13106	1.93111	ppb	100
67) Chlorobenzene	9.90	112	36362	2.05107	ppb	99
68) Ethylbenzene	10.03	91	55504	1.99112	ppb	95
69) Bromoform	10.71	173	9006	1.93804	ppb	97
71) Isopropylbenzene	10.91	105	50633	1.88907	ppb	99
72) 1,1,2,2-Tetrachloroethane	11.19	83	15025	2.02018	ppb	91
73) 1,2,3-Trichloropropane	11.23	110	4415	2.09150	ppb	94
74) t-1,4-Dichloro-2-Butene	11.24	53	2613	1.84963	ppb	81
75) Bromobenzene	11.19	156	17628	1.99506	ppb	97
76) n-Propylbenzene	11.32	91	67771	1.96381	ppb	98
77) 4-Ethyltoluene	11.43	105	56841	1.91798	ppb	98
78) 2-Chlorotoluene	11.39	91	48533	1.97220	ppb	97
79) 1,3,5-Trimethylbenzene	11.50	105	47598	1.93748	ppb	100
80) 4-Chlorotoluene	11.50	91	46827	1.92242	ppb	99
81) Tert-Butylbenzene	11.82	119	43022	1.91160	ppb	96
82) 1,2,4-Trimethylbenzene	11.86	105	48500	1.90785	ppb	99
83) Sec-Butylbenzene	12.04	105	58577	1.94952	ppb	97
84) p-Isopropyltoluene	12.19	119	49518	1.95046	ppb	98
85) Benzyl Chloride	12.35	91	13945	1.83811	ppb	97
86) 1,3-DCB	12.14	146	33447	2.00162	ppb	99
87) 1,4-DCB	12.22	146	33507	1.91467	ppb	96
88) n-Butylbenzene	12.59	91	43428	1.90843	ppb	97
89) 1,2-DCB	12.59	146	30854	1.90506	ppb	96
90) Hexachloroethane	12.85	117	8452	1.81703	ppb	87
91) 1,2-Dibromo-3-chloropropan	13.35	157	2457	1.76335	ppb	96
92) 1,2,4-Trichlorobenzene	14.19	180	13747	1.85173	ppb	99
93) Hexachlorobutadiene	14.38	223	5924	1.91020	ppb	98
94) Naphthalene	14.43	128	37126	1.79099	ppb	99
95) 1,2,3-Trichlorobenzene	14.68	180	19978	1.88873	ppb	98

Data File : M:\THOR\DATA\T120719\0719T09.D Vial: 9
 Acq On : 19 Jul 12 12:53 Operator: DG,RS,HW,ARS,SV
 Sample : 5.0ug/L Vol Std 07-19-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012 Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	96	435456	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.87	117	363264	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	212352	25.00000	ppb	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
31) Dibromofluoromethane(S)	5.95	111	63312	9.29103	ppb	0.00
Spiked Amount	29.744		Recovery	=	31.237%	
36) 1,2-DCA-D4(S)	6.33	65	60027	9.47865	ppb	0.00
Spiked Amount	29.083		Recovery	=	32.593%	
56) Toluene-D8(S)	8.43	98	196082	9.13037	ppb	0.00
Spiked Amount	30.231		Recovery	=	30.201%	
64) 4-Bromofluorobenzene(S)	11.05	95	92855	9.14264	ppb	0.00
Spiked Amount	28.321		Recovery	=	32.283%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.30	85	11045	4.86354	ppb	99
3) Freon 114	1.41	85	14571	4.56836	ppb	98
4) Chloromethane	1.45	50	31396	4.81900	ppb	99
5) Vinyl chloride	1.56	62	45723	5.31271	ppb	98
6) Bromomethane	1.87	94	30238	5.49696	ppb	94
7) Chloroethane	1.97	64	25795	5.20387	ppb	97
8) Dichlorofluoromethane	2.17	67	1323	4.56544	ppb	86
9) Trichlorofluoromethane	2.24	101	10436	5.86657	ppb	95
11) Acetone	2.89	43	8768	4.96901	ppb	95
12) Freon-113	2.85	101	19563	5.46744	ppb	89
13) 1,1-DCE	2.82	61	23901	4.97764	ppb	92
14) t-Butanol	3.68	59	13164	93.22355	ppb	98
15) Methyl Acetate	3.34	43	24407	5.20751	ppb	97
16) Iodomethane	2.98	142	22834	5.25755	ppb	94
17) Acrylonitrile	3.81	52	8122	5.90572	ppb	96
18) Methylene chloride	3.45	84	10146	5.44308	ppb	94
19) Carbon disulfide	3.06	76	2620	4.92947	ppb	96
20) Methyl t-butyl ether (MtBE)	3.90	73	49307	5.31863	ppb	94
21) Trans-1,2-DCE	3.86	96	16955	5.11653	ppb	95
22) Diisopropyl Ether	4.70	59	11471	5.52507	ppb	# 86
23) 1,1-DCA	4.51	63	47950	5.45691	ppb	98
24) Vinyl Acetate	4.70	87	27238	5.48818	ppb	89
25) Ethyl tert Butyl Ether	5.21	59	66131	5.70565	ppb	100
26) MEK (2-Butanone)	5.38	43	9697	4.73433	ppb	96
27) Cis-1,2-DCE	5.32	96	29969	5.32411	ppb	99
28) 2,2-Dichloropropane	5.32	77	18795	11.59481	ppb	95
29) Chloroform	5.75	83	57887	5.30497	ppb	100
30) Bromochloromethane	5.62	128	15767	5.75298	ppb	100
32) 1,1,1-TCA	5.96	97	33756	5.14134	ppb	98
33) Cyclohexane	6.03	41	8909	4.99995	ppb	92
34) 1,1-Dichloropropene	6.17	75	25809	5.41283	ppb	95
35) 2,2,4-Trimethylpentane	6.55	57	36348	5.30407	ppb	98
37) Carbon Tetrachloride	6.16	117	32482	5.27854	ppb	92
38) Tert Amyl Methyl Ether	6.59	73	67201	5.44732	ppb	99
39) 1,2-DCA	6.42	62	38420	5.36908	ppb	99
40) Benzene	6.40	78	101885	5.21303	ppb	99
41) TCE	7.14	95	28157	5.30032	ppb	95
42) 2-Pentanone	7.36	43	425511	101.67128	ppb	99
43) 1,2-Dichloropropane	7.37	63	34984	5.48594	ppb	98
44) Bromodichloromethane	7.68	83	48662	5.51605	ppb	97
45) Methyl Cyclohexane	7.36	83	19188	5.05888	ppb	83

Data File : M:\THOR\DATA\T120719\0719T09.D
 Acq On : 19 Jul 12 12:53
 Sample : 5.0ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 9
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
46) Dibromomethane	7.49	93	18456	5.32052	ppb	99
47) 2-Chloroethyl vinyl ether	7.98	106	691	5.15121	ppb	100
48) MIBK (methyl isobutyl ket	8.33	43	14880	4.94374	ppb	94
49) 1-Bromo-2-chloroethane	7.99	63	24760	5.58047	ppb	100
50) Cis-1,3-Dichloropropene	8.15	75	45589	5.22214	ppb	96
51) Toluene	8.50	91	123530	5.35699	ppb	98
52) Trans-1,3-Dichloropropene	8.72	75	40971	5.32272	ppb	95
53) 1,1,2-TCA	8.90	83	27996	5.45280	ppb	95
54) 2-Hexanone	9.18	43	17051	4.94013	ppb	99
57) 1,2-EDB	9.40	107	29304	5.38033	ppb	97
58) Tetrachloroethene	9.05	166	31143	5.05699	ppb	95
59) 1-Chlorohexane	9.90	91	36955	5.04133	ppb	99
60) 1,1,1,2-Tetrachloroethane	9.99	131	37984	5.27931	ppb	95
61) m&p-Xylene	10.14	106	123042	10.96265	ppb	99
62) o-Xylene	10.54	106	62129	5.35109	ppb	94
63) Styrene	10.55	104	107306	5.43951	ppb	99
65) 1,3-Dichloropropane	9.07	76	50296	5.26692	ppb	99
66) Dibromochloromethane	9.29	129	37767	5.25303	ppb	94
67) Chlorobenzene	9.90	112	98026	5.21957	ppb	97
68) Ethylbenzene	10.03	91	155624	5.26999	ppb	98
69) Bromoform	10.71	173	26416	5.36609	ppb	93
71) Isopropylbenzene	10.91	105	148182	5.33691	ppb	99
72) 1,1,2,2-Tetrachloroethane	11.19	83	42420	5.50589	ppb	98
73) 1,2,3-Trichloropropane	11.23	110	11840	5.41452	ppb	94
74) t-1,4-Dichloro-2-Butene	11.25	53	8302	5.67296	ppb	93
75) Bromobenzene	11.19	156	49040	5.35777	ppb	99
76) n-Propylbenzene	11.32	91	191768	5.36430	ppb	100
77) 4-Ethyltoluene	11.43	105	165084	5.37733	ppb	97
78) 2-Chlorotoluene	11.39	91	136861	5.36875	ppb	100
79) 1,3,5-Trimethylbenzene	11.50	105	139561	5.48395	ppb	99
80) 4-Chlorotoluene	11.50	91	140582	5.57138	ppb	97
81) Tert-Butylbenzene	11.82	119	124728	5.34996	ppb	98
82) 1,2,4-Trimethylbenzene	11.86	105	141302	5.36577	ppb	99
83) Sec-Butylbenzene	12.04	105	172196	5.53229	ppb	99
84) p-Isopropyltoluene	12.19	119	144604	5.49839	ppb	99
85) Benzyl Chloride	12.35	91	41610	5.29458	ppb	97
86) 1,3-DCB	12.13	146	93935	5.42665	ppb	99
87) 1,4-DCB	12.22	146	95715	5.27981	ppb	96
88) n-Butylbenzene	12.59	91	125282	5.31467	ppb	98
89) 1,2-DCB	12.59	146	90224	5.37775	ppb	98
90) Hexachloroethane	12.86	117	24699	5.12581	ppb	98
91) 1,2-Dibromo-3-chloropropan	13.35	157	8054	5.57989	ppb	94
92) 1,2,4-Trichlorobenzene	14.20	180	41456	5.39062	ppb	97
93) Hexachlorobutadiene	14.38	223	17021	5.29823	ppb	85
94) Naphthalene	14.43	128	115311	5.36991	ppb	98
95) 1,2,3-Trichlorobenzene	14.68	180	60463	5.51807	ppb	98


Data File : M:\THOR\DATA\T120719\0719T09.D
Acq On : 19 Jul 12 12:53
Sample : 5.0ug/L Vol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 9
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 08:28:10 2012
Response via : Initial Calibration



Data File : M:\THOR\DATA\T120719\0719T10.D
 Acq On : 19 Jul 12 13:20
 Sample : 10ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 10
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) Fluorobenzene (IS)	6.74	96	461760	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.88	117	382656	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	222464	25.00000	ppb	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev (Min)
31) Dibromofluoromethane(S)	5.95	111	168520	23.32155	ppb	0.00
Spiked Amount 29.744			Recovery =	78.409%		
36) 1,2-DCA-D4 (S)	6.33	65	155567	23.16569	ppb	0.00
Spiked Amount 29.083			Recovery =	79.654%		
56) Toluene-D8 (S)	8.43	98	509225	22.50992	ppb	0.00
Spiked Amount 30.231			Recovery =	74.460%		
64) 4-Bromofluorobenzene(S)	11.05	95	243014	22.71494	ppb	0.00
Spiked Amount 28.321			Recovery =	80.206%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.30	85	20592	8.55092	ppb	100
3) Freon 114	1.41	85	29943	9.21523	ppb	100
4) Chloromethane	1.46	50	55224	7.99352	ppb	100
5) Vinyl chloride	1.57	62	88092	9.65263	ppb	100
6) Bromomethane	1.87	94	56164	9.62843	ppb	100
7) Chloroethane	1.97	64	50219	9.55403	ppb	100
8) Dichlorofluoromethane	2.18	67	3626	10.26166	ppb	100
9) Trichlorofluoromethane	2.24	101	20310	10.76684	ppb	100
11) Acetone	2.89	43	15999	9.46044	ppb	100
12) Freon-113	2.86	101	40039	10.55261	ppb	100
13) 1,1-DCE	2.83	61	49796	9.77980	ppb	100
14) t-Butanol	3.69	59	17712	118.28599	ppb	100
15) Methyl Acetate	3.34	43	43037	9.62218	ppb	100
16) Iodomethane	2.99	142	44928	9.75544	ppb	100
17) Acrylonitrile	3.81	52	14890	10.21016	ppb	100
18) Methylene chloride	3.45	84	17800	9.62295	ppb	100
19) Carbon disulfide	3.07	76	4992	9.56146	ppb	100
20) Methyl t-butyl ether (MtBE)	3.91	73	96445	9.81068	ppb	100
21) Trans-1,2-DCE	3.87	96	32035	9.11655	ppb	100
22) Diisopropyl Ether	4.71	59	22379	10.16494	ppb	100
23) 1,1-DCA	4.51	63	93949	10.08273	ppb	100
24) Vinyl Acetate	4.70	87	51479	9.78163	ppb	100
25) Ethyl tert Butyl Ether	5.21	59	120470	9.80182	ppb	100
26) MEK (2-Butanone)	5.38	43	20960	9.29722	ppb	100
27) Cis-1,2-DCE	5.33	96	58803	9.85150	ppb	100
28) 2,2-Dichloropropane	5.32	77	37619	21.88550	ppb	100
29) Chloroform	5.76	83	111509	9.63695	ppb	100
30) Bromochloromethane	5.62	128	29461	10.13722	ppb	100
32) 1,1,1-TCA	5.96	97	68253	9.80337	ppb	100
33) Cyclohexane	6.03	41	18945	10.02673	ppb	100
34) 1,1-Dichloropropene	6.17	75	50092	9.90716	ppb	100
35) 2,2,4-Trimethylpentane	6.55	57	72402	9.96339	ppb	100
37) Carbon Tetrachloride	6.17	117	62675	9.60491	ppb	100
38) Tert Amyl Methyl Ether	6.59	73	130972	10.01183	ppb	100
39) 1,2-DCA	6.42	62	74124	9.76853	ppb	100
40) Benzene	6.40	78	198603	9.58283	ppb	100
41) TCE	7.15	95	55341	9.82406	ppb	100
42) 2-Pentanone	7.36	43	524739	118.23847	ppb	100
43) 1,2-Dichloropropane	7.37	63	66363	9.81377	ppb	100
44) Bromodichloromethane	7.68	83	91332	9.76313	ppb	100
45) Methyl Cyclohexane	7.36	83	41159	10.23335	ppb	100

Data File : M:\THOR\DATA\T120719\0719T10.D
 Acq On : 19 Jul 12 13:20
 Sample : 10ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 10
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
46) Dibromomethane	7.49	93	35941	9.77089	ppb	100
47) 2-Chloroethyl vinyl ether	7.99	106	1370	10.69163	ppb	100
48) MIBK (methyl isobutyl ket	8.33	43	29904	9.36937	ppb	100
49) 1-Bromo-2-chloroethane	7.99	63	44656	9.49135	ppb	100
50) Cis-1,3-Dichloropropene	8.15	75	90066	9.72920	ppb	100
51) Toluene	8.50	91	242745	9.92720	ppb	100
52) Trans-1,3-Dichloropropene	8.73	75	78273	9.58952	ppb	100
53) 1,1,2-TCA	8.90	83	52576	9.65694	ppb	100
54) 2-Hexanone	9.18	43	34789	9.50513	ppb	100
57) 1,2-EDB	9.40	107	55383	9.65321	ppb	100
58) Tetrachloroethene	9.06	166	63218	9.74509	ppb	100
59) 1-Chlorohexane	9.90	91	75160	9.73357	ppb	100
60) 1,1,1,2-Tetrachloroethane	9.99	131	74500	9.82985	ppb	100
61) m&p-Xylene	10.15	106	235221	19.89538	ppb	100
62) o-Xylene	10.54	106	123202	10.07348	ppb	100
63) Styrene	10.55	104	207845	10.00206	ppb	100
65) 1,3-Dichloropropane	9.07	76	97910	9.73339	ppb	100
66) Dibromochloromethane	9.29	129	73026	9.64248	ppb	100
67) Chlorobenzene	9.90	112	189743	9.59121	ppb	100
68) Ethylbenzene	10.03	91	301792	9.70186	ppb	100
69) Bromoform	10.71	173	49779	9.59955	ppb	100
71) Isopropylbenzene	10.91	105	292683	10.06209	ppb	100
72) 1,1,2,2-Tetrachloroethane	11.19	83	76457	9.47264	ppb	100
73) 1,2,3-Trichloropropane	11.23	110	21819	9.52445	ppb	100
74) t-1,4-Dichloro-2-Butene	11.25	53	15421	10.05857	ppb	100
75) Bromobenzene	11.19	156	95023	9.90967	ppb	100
76) n-Propylbenzene	11.32	91	375107	10.01587	ppb	100
77) 4-Ethyltoluene	11.43	105	333095	10.35682	ppb	100
78) 2-Chlorotoluene	11.39	91	267654	10.02222	ppb	100
79) 1,3,5-Trimethylbenzene	11.50	105	275791	10.34442	ppb	100
80) 4-Chlorotoluene	11.50	91	267918	10.13519	ppb	100
81) Tert-Butylbenzene	11.82	119	243344	9.96331	ppb	100
82) 1,2,4-Trimethylbenzene	11.86	105	283593	10.27959	ppb	100
83) Sec-Butylbenzene	12.04	105	333541	10.22887	ppb	100
84) p-Isopropyltoluene	12.19	119	283601	10.29343	ppb	100
85) Benzyl Chloride	12.35	91	77761	9.44478	ppb	100
86) 1,3-DCB	12.13	146	180339	9.94467	ppb	100
87) 1,4-DCB	12.22	146	184984	9.74023	ppb	100
88) n-Butylbenzene	12.59	91	252451	10.22261	ppb	100
89) 1,2-DCB	12.59	146	175322	9.97497	ppb	100
90) Hexachloroethane	12.86	117	47057	9.32190	ppb	100
91) 1,2-Dibromo-3-chloropropan	13.35	157	15219	10.06460	ppb	100
92) 1,2,4-Trichlorobenzene	14.20	180	81368	10.09953	ppb	100
93) Hexachlorobutadiene	14.38	223	32894	9.77369	ppb	100
94) Naphthalene	14.43	128	230968	10.26703	ppb	100
95) 1,2,3-Trichlorobenzene	14.67	180	116755	10.17114	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T10.D
Acq On : 19 Jul 12 13:20
Sample : 10ug/L Vol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 10
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 08:28:10 2012
Response via : Initial Calibration



Data File : M:\THOR\DATA\T120719\0719T11.D
 Acq On : 19 Jul 12 13:48
 Sample : 20ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 11
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	96	450944	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.88	117	363136	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	216512	25.00000	ppb	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
31) Dibromofluoromethane(S)	5.95	111	266433	37.75615	ppb	0.00
Spiked Amount	29.744		Recovery	= 126.937%		
36) 1,2-DCA-D4(S)	6.33	65	245856	37.48887	ppb	0.00
Spiked Amount	29.083		Recovery	= 128.902%		
56) Toluene-D8(S)	8.43	98	830396	38.68020	ppb	0.00
Spiked Amount	30.231		Recovery	= 127.949%		
64) 4-Bromofluorobenzene(S)	11.05	95	396858	39.08900	ppb	0.00
Spiked Amount	28.321		Recovery	= 138.021%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.30	85	46664	19.84222	ppb	93
3) Freon 114	1.41	85	63081	20.32626	ppb	89
4) Chloromethane	1.45	50	112002	16.60083	ppb	96
5) Vinyl chloride	1.56	62	179429	20.13240	ppb	98
6) Bromomethane	1.86	94	105711	18.55715	ppb	99
7) Chloroethane	1.97	64	103142	20.09314	ppb	95
8) Dichlorofluoromethane	2.18	67	9181	20.87155	ppb	97
9) Trichlorofluoromethane	2.24	101	47356	25.70675	ppb	96
11) Acetone	2.89	43	33405	21.66341	ppb	94
12) Freon-113	2.85	101	75190	20.29226	ppb	97
13) 1,1-DCE	2.82	61	95955	19.29731	ppb	99
14) t-Butanol	3.69	59	24824	169.75836	ppb	100
15) Methyl Acetate	3.34	43	81096	19.91643	ppb	98
16) Iodomethane	2.98	142	86855	19.31159	ppb	99
17) Acrylonitrile	3.81	52	30307	21.28014	ppb	98
18) Methylene chloride	3.45	84	34488	20.02062	ppb	98
19) Carbon disulfide	3.06	76	10542	21.70326	ppb	# 88
20) Methyl t-butyl ether (MtBE)	3.90	73	182893	19.05066	ppb	99
21) Trans-1,2-DCE	3.87	96	64188	18.70481	ppb	97
22) Diisopropyl Ether	4.70	59	42535	19.78355	ppb	# 88
23) 1,1-DCA	4.51	63	178878	19.65788	ppb	98
24) Vinyl Acetate	4.70	87	100156	19.48731	ppb	96
25) Ethyl tert Butyl Ether	5.21	59	233058	19.41715	ppb	97
26) MEK (2-Butanone)	5.37	43	43408	19.33524	ppb	88
27) Cis-1,2-DCE	5.33	96	115419	19.80041	ppb	97
28) 2,2-Dichloropropane	5.32	77	71286	42.46656	ppb	98
29) Chloroform	5.76	83	216322	19.14362	ppb	99
30) Bromochloromethane	5.62	128	55667	19.61385	ppb	91
32) 1,1,1-TCA	5.96	97	130522	19.19690	ppb	97
33) Cyclohexane	6.03	41	35439	19.20613	ppb	98
34) 1,1-Dichloropropene	6.17	75	97918	19.83066	ppb	97
35) 2,2,4-Trimethylpentane	6.55	57	139234	19.61985	ppb	96
37) Carbon Tetrachloride	6.17	117	125056	19.62444	ppb	99
38) Tert Amyl Methyl Ether	6.59	73	247478	19.37158	ppb	99
39) 1,2-DCA	6.42	62	145135	19.58557	ppb	98
40) Benzene	6.40	78	382065	18.87726	ppb	98
41) TCE	7.14	95	107237	19.49316	ppb	98
42) 2-Pentanone	7.36	43	658133	151.85280	ppb	100
43) 1,2-Dichloropropane	7.37	63	129354	19.58769	ppb	97
44) Bromodichloromethane	7.68	83	178755	19.56672	ppb	98
45) Methyl Cyclohexane	7.36	83	76247	19.41196	ppb	99

Data File : M:\THOR\DATA\T120719\0719T11.D
 Acq On : 19 Jul 12 13:48
 Sample : 20ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 11
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
46) Dibromomethane	7.49	93	70661	19.67060	ppb	97
47) 2-Chloroethyl vinyl ether	7.99	106	2760	23.35204	ppb #	100
48) MIBK (methyl isobutyl ket	8.33	43	60201	19.31427	ppb	100
49) 1-Bromo-2-chloroethane	7.99	63	91400	19.89245	ppb	99
50) Cis-1,3-Dichloropropene	8.15	75	176747	19.55069	ppb	99
51) Toluene	8.50	91	471607	19.74924	ppb	98
52) Trans-1,3-Dichloropropene	8.72	75	158806	19.92258	ppb	97
53) 1,1,2-TCA	8.90	83	102413	19.26196	ppb	100
54) 2-Hexanone	9.18	43	70616	19.75664	ppb	98
57) 1,2-EDB	9.40	107	106822	19.61984	ppb	99
58) Tetrachloroethene	9.06	166	120268	19.53595	ppb	97
59) 1-Chlorohexane	9.90	91	145778	19.89376	ppb	99
60) 1,1,1,2-Tetrachloroethane	9.99	131	146253	20.33456	ppb	98
61) m&p-Xylene	10.15	106	462394	41.21236	ppb	99
62) o-Xylene	10.54	106	240916	20.75709	ppb	99
63) Styrene	10.55	104	425446	21.57415	ppb	98
65) 1,3-Dichloropropane	9.07	76	188875	19.78566	ppb	99
66) Dibromochloromethane	9.29	129	145665	20.26776	ppb	100
67) Chlorobenzene	9.90	112	364549	19.41792	ppb	98
68) Ethylbenzene	10.03	91	598003	20.25768	ppb	98
69) Bromoform	10.71	173	98619	20.04032	ppb	96
71) Isopropylbenzene	10.91	105	578914	20.44949	ppb	98
72) 1,1,2,2-Tetrachloroethane	11.19	83	154333	19.64672	ppb	98
73) 1,2,3-Trichloropropane	11.23	110	42893	19.23842	ppb	91
74) t-1,4-Dichloro-2-Butene	11.25	53	32354	21.68350	ppb	96
75) Bromobenzene	11.19	156	185530	19.88027	ppb	98
76) n-Propylbenzene	11.32	91	758387	20.80665	ppb	98
77) 4-Ethyltoluene	11.43	105	653339	20.87252	ppb	98
78) 2-Chlorotoluene	11.39	91	521845	20.07749	ppb	100
79) 1,3,5-Trimethylbenzene	11.50	105	549841	21.19049	ppb	98
80) 4-Chlorotoluene	11.50	91	530306	20.61267	ppb	99
81) Tert-Butylbenzene	11.82	119	482018	20.27796	ppb	98
82) 1,2,4-Trimethylbenzene	11.86	105	561538	20.91400	ppb	99
83) Sec-Butylbenzene	12.04	105	668260	21.05726	ppb	99
84) p-Isopropyltoluene	12.19	119	566536	21.12796	ppb	99
85) Benzyl Chloride	12.35	91	154299	19.25622	ppb	98
86) 1,3-DCB	12.14	146	355716	20.15495	ppb	99
87) 1,4-DCB	12.22	146	358848	19.41437	ppb	100
88) n-Butylbenzene	12.59	91	501731	20.87533	ppb	99
89) 1,2-DCB	12.59	146	337069	19.70479	ppb	99
90) Hexachloroethane	12.86	117	96458	19.63343	ppb	96
91) 1,2-Dibromo-3-chloropropan	13.36	157	32448	22.04834	ppb	94
92) 1,2,4-Trichlorobenzene	14.20	180	162176	20.68292	ppb	100
93) Hexachlorobutadiene	14.38	223	64729	19.76145	ppb	95
94) Naphthalene	14.43	128	476108	21.74583	ppb	99
95) 1,2,3-Trichlorobenzene	14.68	180	231177	20.69267	ppb	99

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T11.D
Acq On : 19 Jul 12 13:48
Sample : 20ug/L Vol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 11
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 08:28:10 2012
Response via : Initial Calibration



Data File : M:\THOR\DATA\T120719\0719T12.D Vial: 12
 Acq On : 19 Jul 12 14:16 Operator: DG,RS,HW,ARS,SV
 Sample : 40ug/L Vol Std 07-19-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012 Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	96	450048	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.88	117	369920	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	219712	25.00000	ppb	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
31) Dibromofluoromethane(S)	5.95	111	544884	77.36909	ppb	0.00
Spiked Amount	29.744		Recovery	= 260.117%		
36) 1,2-DCA-D4(S)	6.33	65	488560	74.64543	ppb	0.00
Spiked Amount	29.083		Recovery	= 256.659%		
56) Toluene-D8(S)	8.43	98	1669961	76.36095	ppb	0.00
Spiked Amount	30.231		Recovery	= 252.593%		
64) 4-Bromofluorobenzene(S)	11.05	95	804405	77.77781	ppb	0.00
Spiked Amount	28.321		Recovery	= 274.630%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.30	85	101464	43.22988	ppb	97
3) Freon 114	1.42	85	136520	44.52891	ppb	88
4) Chloromethane	1.46	50	282736	41.99030	ppb	99
5) Vinyl chloride	1.57	62	357763	40.22185	ppb	100
6) Bromomethane	1.86	94	193264	33.99428	ppb	99
7) Chloroethane	1.97	64	209796	40.95183	ppb	98
8) Dichlorofluoromethane	2.18	67	24179	39.62174	ppb	96
9) Trichlorofluoromethane	2.24	101	112595	61.24281	ppb	99
11) Acetone	2.89	43	57659	38.38775	ppb	99
12) Freon-113	2.85	101	159138	43.03364	ppb	95
13) 1,1-DCE	2.82	61	204122	41.13228	ppb	99
14) t-Butanol	3.69	59	32184	220.52773	ppb	100
15) Methyl Acetate	3.34	43	158595	40.42076	ppb	96
16) Iodomethane	2.98	142	173847	38.73060	ppb	98
17) Acrylonitrile	3.81	52	60943	42.87649	ppb	91
18) Methylene chloride	3.45	84	68312	40.66407	ppb	93
19) Carbon disulfide	3.06	76	20048	42.15606	ppb	# 85
20) Methyl t-butyl ether (MtBE)	3.90	73	353652	36.91075	ppb	98
21) Trans-1,2-DCE	3.87	96	127159	37.12876	ppb	95
22) Diisopropyl Ether	4.70	59	86276	40.20793	ppb	95
23) 1,1-DCA	4.51	63	364882	40.17871	ppb	98
24) Vinyl Acetate	4.70	87	205079	39.98158	ppb	95
25) Ethyl tert Butyl Ether	5.21	59	459486	38.35814	ppb	98
26) MEK (2-Butanone)	5.38	43	87533	38.72047	ppb	94
27) Cis-1,2-DCE	5.33	96	229166	39.39224	ppb	97
28) 2,2-Dichloropropane	5.32	77	141557	84.49635	ppb	96
29) Chloroform	5.76	83	434710	38.54666	ppb	98
30) Bromochloromethane	5.62	128	110740	39.09610	ppb	91
32) 1,1,1-TCA	5.96	97	264324	38.95361	ppb	96
33) Cyclohexane	6.04	41	77803	42.24920	ppb	96
34) 1,1-Dichloropropene	6.17	75	198474	40.27560	ppb	100
35) 2,2,4-Trimethylpentane	6.55	57	293410	41.42752	ppb	94
37) Carbon Tetrachloride	6.17	117	261231	41.07535	ppb	96
38) Tert Amyl Methyl Ether	6.59	73	485700	38.09434	ppb	97
39) 1,2-DCA	6.42	62	284928	38.52680	ppb	99
40) Benzene	6.40	78	767359	37.98954	ppb	99
41) TCE	7.15	95	213589	38.90274	ppb	97
42) 2-Pentanone	7.36	43	764190	176.67466	ppb	98
43) 1,2-Dichloropropane	7.37	63	253205	38.41842	ppb	97
44) Bromodichloromethane	7.68	83	359604	39.44102	ppb	99
45) Methyl Cyclohexane	7.36	83	159998	40.81549	ppb	97

Data File : M:\THOR\DATA\T120719\0719T12.D
 Acq On : 19 Jul 12 14:16
 Sample : 40ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 12
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
46) Dibromomethane	7.49	93	141296	39.41227	ppb	93
47) 2-Chloroethyl vinyl ether	7.99	106	4618	39.97505	ppb #	100
48) MIBK (methyl isobutyl ket	8.33	43	121497	39.05746	ppb	99
49) 1-Bromo-2-chloroethane	7.99	63	181376	39.55356	ppb	99
50) Cis-1,3-Dichloropropene	8.15	75	367817	40.76670	ppb	100
51) Toluene	8.50	91	942978	39.56722	ppb	100
52) Trans-1,3-Dichloropropene	8.72	75	327606	41.18075	ppb	97
53) 1,1,2-TCA	8.90	83	203529	38.35620	ppb	97
54) 2-Hexanone	9.18	43	145904	40.90166	ppb	99
57) 1,2-EDB	9.40	107	216913	39.10946	ppb	98
58) Tetrachloroethene	9.06	166	243143	38.77105	ppb	95
59) 1-Chlorohexane	9.90	91	305567	40.93481	ppb	97
60) 1,1,1,2-Tetrachloroethane	9.99	131	289716	39.54248	ppb	95
61) m&p-Xylene	10.15	106	942114	82.42904	ppb	98
62) o-Xylene	10.54	106	486606	41.15663	ppb	98
63) Styrene	10.55	104	862890	42.95425	ppb	100
65) 1,3-Dichloropropane	9.07	76	382242	39.30755	ppb	98
66) Dibromochloromethane	9.29	129	292949	40.01326	ppb	96
67) Chlorobenzene	9.90	112	739958	38.69148	ppb	99
68) Ethylbenzene	10.03	91	1209652	40.22613	ppb	98
69) Bromoform	10.71	173	206749	41.24287	ppb	99
71) Isopropylbenzene	10.91	105	1186391	41.29757	ppb	98
72) 1,1,2,2-Tetrachloroethane	11.19	83	311389	39.06275	ppb	100
73) 1,2,3-Trichloropropane	11.23	110	87283	38.57810	ppb	93
74) t-1,4-Dichloro-2-Butene	11.25	53	67511	44.58657	ppb	97
75) Bromobenzene	11.19	156	370849	39.15918	ppb	99
76) n-Propylbenzene	11.32	91	1546930	41.82252	ppb	99
77) 4-Ethyltoluene	11.43	105	1336329	42.07052	ppb	99
78) 2-Chlorotoluene	11.39	91	1059468	40.16835	ppb	99
79) 1,3,5-Trimethylbenzene	11.50	105	1118597	42.48207	ppb	99
80) 4-Chlorotoluene	11.50	91	1066136	40.83649	ppb	100
81) Tert-Butylbenzene	11.82	119	993558	41.18911	ppb	99
82) 1,2,4-Trimethylbenzene	11.86	105	1142659	41.93753	ppb	99
83) Sec-Butylbenzene	12.04	105	1362846	42.31861	ppb	100
84) p-Isopropyltoluene	12.19	119	1167081	42.89031	ppb	99
85) Benzyl Chloride	12.35	91	328559	40.40634	ppb	98
86) 1,3-DCB	12.13	146	706591	39.45252	ppb	99
87) 1,4-DCB	12.22	146	717680	38.26236	ppb	100
88) n-Butylbenzene	12.59	91	1032004	42.31282	ppb	99
89) 1,2-DCB	12.59	146	673414	38.79389	ppb	98
90) Hexachloroethane	12.86	117	199424	40.00032	ppb	95
91) 1,2-Dibromo-3-chloropropan	13.35	157	66284	44.38384	ppb	95
92) 1,2,4-Trichlorobenzene	14.20	180	327616	41.17358	ppb	100
93) Hexachlorobutadiene	14.38	223	129523	38.96681	ppb	98
94) Naphthalene	14.43	128	999454	44.98436	ppb	99
95) 1,2,3-Trichlorobenzene	14.68	180	465877	41.09332	ppb	98

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T12.D
Acq On : 19 Jul 12 14:16
Sample : 40ug/L Vol Std 07-19-12
Misc : 10ml w/Sul of IS&S: 06-7-12

Vial: 12
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 08:28:10 2012
Response via : Initial Calibration



Data File : M:\THOR\DATA\T120719\0719T13.D
 Acq On : 19 Jul 12 14:44
 Sample : 100ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 13
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) Fluorobenzene (IS)	6.73	96	444096	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.88	117	369984	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	225280	25.00000	ppb	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev (Min)
31) Dibromofluoromethane(S)	5.94	111	677704	97.51815	ppb	0.00
Spiked Amount 29.744			Recovery = 327.859%			
36) 1,2-DCA-D4(S)	6.33	65	602641	93.30952	ppb	0.00
Spiked Amount 29.083			Recovery = 320.837%			
56) Toluene-D8(S)	8.43	98	2073207	94.78345	ppb	0.00
Spiked Amount 30.231			Recovery = 313.531%			
64) 4-Bromofluorobenzene(S)	11.05	95	1023987	98.99204	ppb	0.00
Spiked Amount 28.321			Recovery = 349.536%			

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.29	85	254656	109.95320	ppb	99
3) Freon 114	1.41	85	295808	98.23896	ppb	92
4) Chloromethane	1.45	50	771844	116.16609	ppb	98
5) Vinyl chloride	1.56	62	891545	101.57617	ppb	98
6) Bromomethane	1.85	94	452818	80.71617	ppb	98
7) Chloroethane	1.95	64	503433	99.58633	ppb	94
8) Dichlorofluoromethane	2.18	67	115020	100.01762	ppb	99
9) Trichlorofluoromethane	2.23	101	328219	180.91796	ppb	99
11) Acetone	2.89	43	145827	100.36210	ppb	98
12) Freon-113	2.84	101	365975	100.29230	ppb	97
13) 1,1-DCE	2.81	61	492964	100.66770	ppb	98
14) t-Butanol	3.70	59	53864	374.02770	ppb	99
15) Methyl Acetate	3.33	43	378645	99.85965	ppb	99
16) Iodomethane	2.97	142	429518	96.97290	ppb	97
17) Acrylonitrile	3.80	52	148837	106.11781	ppb	92
18) Methylene chloride	3.45	84	163136	99.75173	ppb	96
19) Carbon disulfide	3.05	76	45848	98.86363	ppb	# 88
20) Methyl t-butyl ether (MtBE)	3.90	73	822710	87.01727	ppb	98
21) Trans-1,2-DCE	3.86	96	303532	89.81519	ppb	95
22) Diisopropyl Ether	4.70	59	207477	97.98816	ppb	93
23) 1,1-DCA	4.50	63	860226	95.99267	ppb	97
24) Vinyl Acetate	4.70	87	495299	97.85616	ppb	96
25) Ethyl tert Butyl Ether	5.21	59	1019255	86.22835	ppb	99
26) MEK (2-Butanone)	5.37	43	225877	100.70732	ppb	92
27) Cis-1,2-DCE	5.32	96	554128	96.52785	ppb	96
28) 2,2-Dichloropropane	5.32	77	327819	198.30000	ppb	99
29) Chloroform	5.75	83	1043860	93.80183	ppb	98
30) Bromochloromethane	5.62	128	277342	99.22624	ppb	93
32) 1,1,1-TCA	5.96	97	618230	92.33007	ppb	94
33) Cyclohexane	6.03	41	173334	95.38672	ppb	97
34) 1,1-Dichloropropene	6.16	75	474643	97.60846	ppb	99
35) 2,2,4-Trimethylpentane	6.55	57	649315	92.90765	ppb	94
37) Carbon Tetrachloride	6.16	117	627649	100.01275	ppb	97
38) Tert Amyl Methyl Ether	6.59	73	1115219	88.64096	ppb	96
39) 1,2-DCA	6.42	62	688055	94.28291	ppb	98
40) Benzene	6.40	78	1827390	91.68086	ppb	99
41) TCE	7.14	95	502537	92.75799	ppb	98
42) 2-Pentanone	7.36	43	907754	212.67824	ppb	98
43) 1,2-Dichloropropane	7.37	63	623762	95.91093	ppb	97
44) Bromodichloromethane	7.68	83	887397	98.63330	ppb	99
45) Methyl Cyclohexane	7.36	83	367578	95.02589	ppb	97

Data File : M:\THOR\DATA\T120719\0719T13.D
 Acq On : 19 Jul 12 14:44
 Sample : 100ug/L Vol Std 07-19-12
 Misc : 10ml w/Sul of IS&S: 06-7-12

Vial: 13
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 8:00 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 07:59:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
46) Dibromomethane	7.49	93	343569	97.11750	ppb	95
47) 2-Chloroethyl vinyl ether	7.99	106	11121	99.31396	ppb #	100
48) MIBK (methyl isobutyl ket	8.33	43	317753	103.51662	ppb	98
49) 1-Bromo-2-chloroethane	7.99	63	444096	98.14420	ppb	99
50) Cis-1,3-Dichloropropene	8.15	75	923494	103.72652	ppb	98
51) Toluene	8.50	91	2302514	97.90801	ppb	100
52) Trans-1,3-Dichloropropene	8.72	75	844862	107.62424	ppb	97
53) 1,1,2-TCA	8.90	83	506640	96.75885	ppb	98
54) 2-Hexanone	9.18	43	374170	106.29789	ppb	96
57) 1,2-EDB	9.40	107	552458	99.59106	ppb	98
58) Tetrachloroethene	9.06	166	580637	92.57110	ppb	96
59) 1-Chlorohexane	9.90	91	748840	100.29983	ppb	96
60) 1,1,1,2-Tetrachloroethane	9.99	131	746191	101.82778	ppb	98
61) m&p-Xylene	10.14	106	2312256	202.27283	ppb	97
62) o-Xylene	10.54	106	1205888	101.97512	ppb	97
63) Styrene	10.55	104	2181574	108.57892	ppb	97
65) 1,3-Dichloropropane	9.07	76	938122	96.45434	ppb	100
66) Dibromochloromethane	9.29	129	748578	102.22894	ppb	99
67) Chlorobenzene	9.90	112	1810618	94.65857	ppb	99
68) Ethylbenzene	10.03	91	2980271	99.08969	ppb	98
69) Bromoform	10.71	173	538782	107.45915	ppb	100
71) Isopropylbenzene	10.91	105	2947712	100.07206	ppb	98
72) 1,1,2,2-Tetrachloroethane	11.19	83	796018	97.38982	ppb	99
73) 1,2,3-Trichloropropane	11.23	110	223287	96.25110	ppb	93
74) t-1,4-Dichloro-2-Butene	11.25	53	179952	115.90904	ppb	99
75) Bromobenzene	11.19	156	932826	96.06567	ppb	100
76) n-Propylbenzene	11.32	91	3839951	101.25032	ppb	100
77) 4-Ethyltoluene	11.43	105	3324604	102.07879	ppb	99
78) 2-Chlorotoluene	11.40	91	2632771	97.35098	ppb	100
79) 1,3,5-Trimethylbenzene	11.50	105	2768017	102.52551	ppb	98
80) 4-Chlorotoluene	11.50	91	2649819	98.98815	ppb	100
81) Tert-Butylbenzene	11.82	119	2490617	100.69949	ppb	99
82) 1,2,4-Trimethylbenzene	11.86	105	2848266	101.95249	ppb	99
83) Sec-Butylbenzene	12.04	105	3384214	102.48812	ppb	100
84) p-Isopropyltoluene	12.19	119	2906241	104.16479	ppb	99
85) Benzyl Chloride	12.35	91	910665	109.22596	ppb	96
86) 1,3-DCB	12.14	146	1775349	96.67663	ppb	99
87) 1,4-DCB	12.22	146	1789528	93.04875	ppb	100
88) n-Butylbenzene	12.59	91	2558982	102.32670	ppb	99
89) 1,2-DCB	12.59	146	1688312	94.85606	ppb	99
90) Hexachloroethane	12.86	117	521928	102.10049	ppb	95
91) 1,2-Dibromo-3-chloropropan	13.35	157	180474	117.85879	ppb	94
92) 1,2,4-Trichlorobenzene	14.20	180	875328	107.28908	ppb	100
93) Hexachlorobutadiene	14.38	223	327441	96.07540	ppb	95
94) Naphthalene	14.43	128	2618767	114.95471	ppb	99
95) 1,2,3-Trichlorobenzene	14.68	180	1182785	101.75059	ppb	98

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T13.D
Acq On : 19 Jul 12 14:44
Sample : 100ug/L Vol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12


Vial: 13
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 20 8:00 2012


Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 08:28:10 2012
Response via : Initial Calibration







Method Name: M:\THOR\DATA\T120719\TALLW.M
Calibration Table Last Updated: Fri Jul 20 10:40:23 2012




Method Name: M:\THOR\DATA\T120719\TALLW.M
Calibration Table Last Updated: Fri Jul 20 10:40:23 2012




Method Name: M:\THOR\DATA\T120719\TALLW.M
Calibration Table Last Updated: Fri Jul 20 10:40:23 2012




Method Name: M:\THOR\DATA\T120719\TALLW.M
Calibration Table Last Updated: Fri Jul 20 10:40:23 2012




Method Name: M:\THOR\DATA\T120719\TALLW.M
Calibration Table Last Updated: Fri Jul 20 10:40:23 2012




Method Name: M:\THOR\DATA\T120719\TALLW.M
Calibration Table Last Updated: Fri Jul 20 10:40:23 2012



Method Name: M:\THOR\DATA\T120719\TALLW.M
Calibration Table Last Updated: Fri Jul 20 10:40:23 2012



Method Name: M:\THOR\DATA\T120719\TALLW.M
Calibration Table Last Updated: Fri Jul 20 10:40:23 2012



Method Name: M:\THOR\DATA\T120719\TALLW.M
Calibration Table Last Updated: Fri Jul 20 10:40:23 2012

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 7

Second Source Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: Water

SDG No: 68258
Date Analyzed: 07/19/12
Instrument: Thor
Initial Cal. Date: 07/19/12
Data File: 0719T31.D

		Compound	MEAN	CCRF	%D	%Drift	
1	I	Fluorobenzene (IS)	ISTD			I	
2	TM	Dichlorodifluoromethane	0.1266	0.1014	20	TM	
3	TML	Freon 114	0.1578	0.1581	0.22	TML	10
4	TM**L	Chloromethane	0.3709	0.3090	17	TM**L	2.0
5	TM*	Vinyl chloride	0.4941	0.4993	1.1	TM*	
6	TM	Bromomethane	0.3158	0.2956	6.4	TM	
7	TM	Chloroethane	0.2846	0.2799	1.6	TM	
8	TMQ	Dichlorofluoromethane	0.0241	0.0169	30	TMQ	9.1
9	TM	Trichlorofluoromethane	0.1021	0.1035	1.3	TM	
10	TMQ	Acrolein	0.0000	0.0065	0.00	TMQ	
11	TML	Acetone	0.1608	0.1059	34	TML	18
12	TM	Freon-113	0.2054	0.2048	0.31	TM	
13	TM*	1,1-DCE	0.2757	0.2657	3.6	TM*	
14	TM	t-Butanol	0.0081	0.0083	2.3	TM	
15	TML	Methyl Acetate	0.4032	0.2447	39	TML	1.8
16	TM	Iodomethane	0.2493	0.2358	5.4	TM	
17	TM	Acrylonitrile	0.0790	0.0808	2.3	TM	
18	TML	Methylene chloride	0.1556	0.0948	39	TML	5.5
19	TML	Carbon disulfide	0.0329	0.0300	8.9	TML	7.0
20	TM	Methyl t-butyl ether (MtBE)	0.5322	0.5046	5.2	TM	
21	TM	Trans-1,2-DCE	0.1902	0.1862	2.1	TM	
22	TM	Diisopropyl Ether	0.1192	0.1196	0.38	TM	
23	TM**	1,1-DCA	0.5045	0.5081	0.73	TM**	
24	TM	Vinyl Acetate	0.2849	0.2762	3.1	TM	
25	TM	Ethyl tert Butyl Ether	0.6654	0.6504	2.3	TM	
26	TML	MEK (2-Butanone)	0.1418	0.1260	11	TML	2.9
27	TM	Cis-1,2-DCE	0.3232	0.3228	0.12	TM	
28	TM	2,2-Dichloropropane	0.2032	0.1629	20	TM	
29	TM*	Chloroform	0.6265	0.6014	4.0	TM*	
30	TM	Bromochloromethane	0.1573	0.1601	1.8	TM	
31	S	Dibromofluoromethane(S)	0.3912	0.3840	1.8	S	
32	TM	1,1,1-TCA	0.3769	0.3627	3.8	TM	
33	TM	Cyclohexane	0.1023	0.1023	0.01	TM	
34	TM	1,1-Dichloropropene	0.2737	0.2734	0.13	TM	
35	TM	2,2,4-Trimethylpentane	0.3934	0.3395	14	TM	
36	S	1,2-DCA-D4(S)	0.3636	0.3559	2.1	S	
37	TM	Carbon Tetrachloride	0.3533	0.3549	0.46	TM	
38	TM	Tert Amyl Methyl Ether	0.7083	0.6971	1.6	TM	
39	TM	1,2-DCA	0.4108	0.4011	2.4	TM	
40	TM	Benzene	1.122	1.062	5.3	TM	

Average

7.6

AR5 7/27/12

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 7

Second Source Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: Water

SDG No: 68258
Date Analyzed: 07/19/12
Instrument: Thor
Cal. Date: 07/19/12
Data File: 0719T31.D

		Compound	MEAN	CCRF	%D	%Drift
41	TM	TCE	0.3050	0.3245	6.4	TM
42	TM	2-Pentanone	0.2403	0.2373	1.2	TM
43	TM*	1,2-Dichloropropane	0.3661	0.3693	0.88	TM*
44	TM	Bromodichloromethane	0.5065	0.4847	4.3	TM
45	TM	Methyl Cyclohexane	0.2178	0.2085	4.3	TM
46	TM	Dibromomethane	0.1991	0.1999	0.37	TM
47	TML	2-Chloroethyl vinyl ether	0.0061	0.0063	2.4	TML 11
48	TM	MIBK (methyl isobutyl ketone)	0.1728	0.1759	1.8	TM
49	TM	1-Bromo-2-chloroethane	0.2547	0.2495	2.1	TM
50	TM	Cis-1,3-Dichloropropene	0.5012	0.4724	5.7	TM
51	TM*	Toluene	1.324	1.341	1.3	TM*
52	TM	Trans-1,3-Dichloropropene	0.4419	0.4104	7.1	TM
53	TM	1,1,2-TCA	0.2948	0.2833	3.9	TM
54	TM	2-Hexanone	0.1982	0.2000	0.95	TM
55	I	Chlorobenzene-D5 (IS)	ISTD			I
56	S	Toluene-D8(S)	1.478	1.447	2.1	S
57	TM	1,2-EDB	0.3748	0.3669	2.1	TM
58	TM	Tetrachloroethene	0.4238	0.4269	0.73	TM
59	TM	1-Chlorohexane	0.5045	0.5027	0.36	TM
60	TM	1,1,1,2-Tetrachloroethane	0.4952	0.4849	2.1	TM
61	TM	m&p-Xylene	0.7724	0.8080	4.6	TM
62	TM	o-Xylene	0.7990	0.8295	3.8	TM
63	TM	Styrene	1.358	1.406	3.5	TM
64	S	4-Bromofluorobenzene(S)	0.6990	0.7156	2.4	S
65	TM	1,3-Dichloropropane	0.6572	0.6637	0.99	TM
66	TM	Dibromochloromethane	0.4948	0.4812	2.7	TM
67	TM**	Chlorobenzene	1.292	1.269	1.8	TM**
68	TM*	Ethylbenzene	2.032	2.056	1.2	TM*
69	TM**	Bromoform	0.3388	0.3286	3.0	TM**
70	I	1,4-Dichlorobenzene-D (IS)	ISTD			I
71	TM	Isopropylbenzene	3.269	3.409	4.3	TM
72	TM**	1,1,2,2-Tetrachloroethane	0.9070	0.8367	7.8	TM**
73	TM	1,2,3-Trichloropropane	0.2574	0.2664	3.5	TM
74	TM	t-1,4-Dichloro-2-Butene	0.1723	0.1823	5.8	TM
75	TM	Bromobenzene	1.078	1.079	0.15	TM
76	TM	n-Propylbenzene	4.209	4.445	5.6	TM
77	TM	4-Ethyltoluene	3.614	3.749	3.7	TM
78	TM	2-Chlorotoluene	3.001	3.080	2.6	TM
79	TM	1,3,5-Trimethylbenzene	2.996	3.186	6.3	TM
80	TM	4-Chlorotoluene	2.971	3.080	3.7	TM

Average

3.1

MS 7/27/12

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 7

Second Source Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: Water

SDG No: 68258
Date Analyzed: 07/19/12
Instrument: Thor
Cal. Date: 07/19/12
Data File: 0719T31.D

		Compound	MEAN	CCRF	%D	%Drift
81	TM	Tert-Butylbenzene	2.745	2.818	2.7	TM
82	TM	1,2,4-Trimethylbenzene	3.100	3.213	3.6	TM
83	TM	Sec-Butylbenzene	3.664	3.850	5.1	TM
84	TM	p-Isopropyltoluene	3.096	3.241	4.7	TM
85	TM	Benzyl Chloride	0.9252	0.6126	34	TM
86	TM	1,3-DCB	2.038	2.081	2.1	TM
87	TM	1,4-DCB	2.134	2.096	1.8	TM
88	TM	n-Butylbenzene	2.775	2.837	2.2	TM
89	TM	1,2-DCB	1.975	1.941	1.7	TM
90	TM	Hexachloroethane	0.5673	0.5516	2.8	TM
91	TM	1,2-Dibromo-3-chloropropane	0.1699	0.1722	1.4	TM
92	TM	1,2,4-Trichlorobenzene	0.9054	0.9040	0.15	TM
93	TM	Hexachlorobutadiene	0.3782	0.3490	7.7	TM
94	TM	Naphthalene	2.528	2.684	6.1	TM
95	TM	1,2,3-Trichlorobenzene	1.290	1.318	2.2	TM
96						
97						
98						
99						
100						
101						
102						
103						
104						
105						
106						
107						
108						
109						
110						
111						
112						
113						
114						
115						
116						
117						
118						
119						
120						

*NT

Average

5.2

Handwritten: 7/27/12

Data File : M:\THOR\DATA\T120719\0719T31.D
 Acq On : 19 Jul 12 23:03
 Sample : 120719A LCS-1WT (SS)
 Misc : 10ml w/Sul of IS&S: 06-7-12

Vial: 31
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 10:45 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 10:40:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	96	459584	25.00000 ppb	0.00
55) Chlorobenzene-D5 (IS)	9.87	117	371008	25.00000 ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	216768	25.00000 ppb	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc Units	Dev(Min)
31) Dibromofluoromethane(S)	5.95	111	225058	31.29333 ppb	0.00
Spiked Amount			31.881	Recovery = 98.155%	
36) 1,2-DCA-D4(S)	6.33	65	220138	32.93626 ppb	0.00
Spiked Amount			33.647	Recovery = 97.888%	
56) Toluene-D8(S)	8.43	98	802051	36.56718 ppb	0.00
Spiked Amount			37.345	Recovery = 97.917%	
64) 4-Bromofluorobenzene(S)	11.05	95	313456	30.21914 ppb	0.00
Spiked Amount			29.515	Recovery = 102.384%	

Target Compounds	R.T.	QIon	Response	Conc Units	Dev(Min)
2) Dichlorodifluoromethane	1.30	85	18648	8.01049 ppb	98
3) Freon 114	1.41	85	29065	8.97783 ppb	92
4) Chloromethane	1.45	50	56808	9.80339 ppb	99
5) Vinyl chloride	1.56	62	91788	10.10524 ppb	99
6) Bromomethane	1.87	94	54346	9.36087 ppb	98
7) Chloroethane	1.97	64	51463	9.83706 ppb	96
8) Dichlorofluoromethane	2.18	67	3106	9.09488 ppb	97
9) Trichlorofluoromethane	2.24	101	19028	10.13498 ppb	100
11) Acetone	2.88	43	19460	11.84185 ppb	98
12) Freon-113	2.85	101	37646	9.96889 ppb	94
13) 1,1-DCE	2.82	61	48838	9.63706 ppb	93
14) t-Butanol	3.69	59	19056	127.86417 ppb	98
15) Methyl Acetate	3.34	43	44993	10.18034 ppb	95
16) Iodomethane	2.98	142	43340	9.45518 ppb	97
17) Acrylonitrile	3.81	52	14853	10.23301 ppb	95
18) Methylene chloride	3.45	84	17424	9.44871 ppb	95
19) Carbon disulfide	3.06	76	5510	10.69990 ppb	# 86
20) Methyl t-butyl ether (MtBE)	3.90	73	92761	9.48061 ppb	98
21) Trans-1,2-DCE	3.87	96	34225	9.78590 ppb	97
22) Diisopropyl Ether	4.70	59	21995	10.03782 ppb	95
23) 1,1-DCA	4.51	63	93412	10.07257 ppb	98
24) Vinyl Acetate	4.70	87	50781	9.69469 ppb	95
25) Ethyl tert Butyl Ether	5.21	59	119561	9.77392 ppb	99
26) MEK (2-Butanone)	5.38	43	23166	10.28682 ppb	95
27) Cis-1,2-DCE	5.32	96	59336	9.98787 ppb	96
28) 2,2-Dichloropropane	5.32	77	29940	8.01402 ppb	99
29) Chloroform	5.75	83	110557	9.59991 ppb	94
30) Bromochloromethane	5.62	128	29433	10.17554 ppb	98
32) 1,1,1-TCA	5.96	97	66682	9.62307 ppb	96
33) Cyclohexane	6.03	41	18804	9.99923 ppb	94
34) 1,1-Dichloropropene	6.17	75	50257	9.98686 ppb	98
35) 2,2,4-Trimethylpentane	6.55	57	62413	8.62945 ppb	97
37) Carbon Tetrachloride	6.16	117	65247	10.04641 ppb	95
38) Tert Amyl Methyl Ether	6.59	73	128152	9.84264 ppb	100
39) 1,2-DCA	6.41	62	73737	9.76354 ppb	99
40) Benzene	6.40	78	195282	9.46720 ppb	97
41) TCE	7.14	95	59649	10.63894 ppb	98
42) 2-Pentanone	7.36	43	545318	123.45728 ppb	100
43) 1,2-Dichloropropane	7.37	63	67896	10.08801 ppb	96

Algorithm Check: (91788)(25) CI = 10.10522903 ✓
 (459584)(0.4941) Qvalue ARS 7/27/12

Data File : M:\THOR\DATA\T120719\0719T31.D
 Acq On : 19 Jul 12 23:03
 Sample : 120719A LCS-1WT (SS)
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 31
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 10:45 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 10:40:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
44) Bromodichloromethane	7.68	83	89098	9.56942	ppb	98
45) Methyl Cyclohexane	7.36	83	38326	9.57410	ppb	95
46) Dibromomethane	7.49	93	36747	10.03731	ppb	99
47) 2-Chloroethyl vinyl ether	7.99	106	1154	8.86120	ppb #	100
48) MIBK (methyl isobutyl ket	8.33	43	32328	10.17680	ppb	97
49) 1-Bromo-2-chloroethane	7.99	63	45864	9.79426	ppb	99
50) <u>Cis-1,3-Dichloropropene</u>	8.15	75	86842	<u>9.42535</u>	<u>ppb</u>	98
51) Toluene	8.50	91	246468	10.12718	ppb	98
52) <u>Trans-1,3-Dichloropropene</u>	8.72	75	75443	<u>9.28657</u>	<u>ppb</u>	98
53) 1,1,2-TCA	8.90	83	52073	9.60983	ppb	98
54) 2-Hexanone	9.18	43	36772	10.09450	ppb	96
57) 1,2-EDB	9.40	107	54442	9.78712	ppb	99
58) Tetrachloroethene	9.06	166	63354	10.07267	ppb	96
59) 1-Chlorohexane	9.90	91	74600	9.96437	ppb	98
60) 1,1,1,2-Tetrachloroethane	9.99	131	71965	9.79349	ppb	96
61) m&p-Xylene	10.14	106	239826	20.92173	ppb	98
62) o-Xylene	10.54	106	123104	10.38148	ppb	98
63) Styrene	10.55	104	208582	10.35266	ppb	98
65) 1,3-Dichloropropane	9.07	76	98494	10.09885	ppb	97
66) Dibromochloromethane	9.29	129	71411	9.72527	ppb	99
67) Chlorobenzene	9.90	112	188318	9.81804	ppb	99
68) Ethylbenzene	10.03	91	305101	10.11617	ppb	99
69) Bromoform	10.71	173	48764	9.69905	ppb	99
71) Isopropylbenzene	10.91	105	295625	10.43029	ppb	98
72) 1,1,2,2-Tetrachloroethane	11.19	83	72548	9.22452	ppb	99
73) 1,2,3-Trichloropropane	11.23	110	23096	10.34681	ppb	85
74) t-1,4-Dichloro-2-Butene	11.25	53	15810	10.58327	ppb	89
75) Bromobenzene	11.19	156	93573	10.01488	ppb	99
76) n-Propylbenzene	11.32	91	385440	10.56221	ppb	98
77) 4-Ethyltoluene	11.43	105	325068	10.37283	ppb	98
78) 2-Chlorotoluene	11.39	91	267062	10.26282	ppb	100
79) 1,3,5-Trimethylbenzene	11.50	105	276230	10.63314	ppb	99
80) 4-Chlorotoluene	11.50	91	267095	10.36956	ppb	100
81) Tert-Butylbenzene	11.82	119	244365	10.26802	ppb	99
82) 1,2,4-Trimethylbenzene	11.86	105	278601	10.36400	ppb	99
83) Sec-Butylbenzene	12.04	105	333801	10.50584	ppb	100
84) p-Isopropyltoluene	12.19	119	281061	10.46929	ppb	99
85) Benzyl Chloride	12.35	91	53118	6.62120	ppb	97
86) 1,3-DCB	12.13	146	180466	10.21317	ppb	99
87) 1,4-DCB	12.22	146	181734	9.82055	ppb	99
88) n-Butylbenzene	12.59	91	245949	10.22102	ppb	98
89) 1,2-DCB	12.59	146	168341	9.82946	ppb	97
90) Hexachloroethane	12.86	117	47831	9.72421	ppb	93
91) 1,2-Dibromo-3-chloropropan	13.35	157	14935	10.13631	ppb	95
92) 1,2,4-Trichlorobenzene	14.20	180	78384	9.98480	ppb	100
93) Hexachlorobutadiene	14.38	223	30261	9.22762	ppb	89
94) Naphthalene	14.43	128	232681	10.61496	ppb	98
95) 1,2,3-Trichlorobenzene	14.68	180	114268	10.21606	ppb	99

*1,3-dichloropropene, total:
18.71192 ppb*

ARS 7/27/12

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T31.D
Acq On : 19 Jul 12 23:03
Sample : 120719A LCS-1WT (SS)
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 31
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 20 10:45 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 10:40:23 2012
Response via : Initial Calibration



VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: Water

SDG No: 68258
Date Analyzed: 07/19/12
Instrument: Thor
Initial Cal. Date: 07/19/12
Data File: 0719T30.D

		Compound	MEAN	CCRF	%D	%Drift	
1	I	Fluorobenzene (IS)	ISTD			I	
2	TM	Dichlorodifluoromethane	0.1266	0.1166	7.9	TM	
3	TML	Freon 114	0.1578	0.1597	1.2	TML	9.3
4	TM**L	Chloromethane	0.3709	0.2977	20	TM**L	5.7
5	TM*	Vinyl chloride	0.4941	0.4702	4.8	TM*	
6	TM	Bromomethane	0.3158	0.2902	8.1	TM	
7	TM	Chloroethane	0.2846	0.2732	4.0	TM	
8	TMQ	Dichlorofluoromethane	0.0241	0.0166	31	TMQ	11
9	TM	Trichlorofluoromethane	0.1021	0.0978	4.2	TM	
10	TMQ	Acrolein	0.0000	0.0068	0.00	TMQ	
11	TML	Acetone	0.1608	0.0967	40	TML	7.0
12	TM	Freon-113	0.2054	0.1875	8.7	TM	
13	TM*	1,1-DCE	0.2757	0.2609	5.3	TM*	
14	TM	t-Butanol	0.0081	0.0087	7.1	TM	
15	TML	Methyl Acetate	0.4032	0.2359	41	TML	2.4
16	TM	Iodomethane	0.2493	0.2420	2.9	TM	
17	TM	Acrylonitrile	0.0790	0.0833	5.5	TM	
18	TML	Methylene chloride	0.1556	0.0889	43	TML	12
19	TML	Carbon disulfide	0.0329	0.0275	16	TML	2.7
20	TM	Methyl t-butyl ether (MtBE)	0.5322	0.5055	5.0	TM	
21	TM	Trans-1,2-DCE	0.1902	0.1709	10	TM	
22	TM	Diisopropyl Ether	0.1192	0.1231	3.2	TM	
23	TM**	1,1-DCA	0.5045	0.5008	0.73	TM**	
24	TM	Vinyl Acetate	0.2849	0.2716	4.7	TM	
25	TM	Ethyl tert Butyl Ether	0.6654	0.6458	2.9	TM	
26	TML	MEK (2-Butanone)	0.1418	0.1240	13	TML	1.3
27	TM	Cis-1,2-DCE	0.3232	0.3160	2.2	TM	
28	TM	2,2-Dichloropropane	0.2032	0.1621	20	TM	
29	TM*	Chloroform	0.6265	0.6125	2.2	TM*	
30	TM	Bromochloromethane	0.1573	0.1554	1.2	TM	
31	S	Dibromofluoromethane(S)	0.3912	0.3908	0.10	S	
32	TM	1,1,1-TCA	0.3769	0.3636	3.6	TM	
33	TM	Cyclohexane	0.1023	0.0982	4.0	TM	
34	TM	1,1-Dichloropropene	0.2737	0.2587	5.5	TM	
35	TM	2,2,4-Trimethylpentane	0.3934	0.3120	21	TM	*NT
36	S	1,2-DCA-D4(S)	0.3636	0.3700	1.8	S	
37	TM	Carbon Tetrachloride	0.3533	0.3368	4.7	TM	
38	TM	Tert Amyl Methyl Ether	0.7083	0.6804	3.9	TM	
39	TM	1,2-DCA	0.4108	0.3911	4.8	TM	
40	TM	Benzene	1.122	1.067	4.9	TM	

Average

9.5

MRS 7/27/12

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: Water

SDG No: 68258
Date Analyzed: 07/19/12
Instrument: Thor
Cal. Date: 07/19/12
Data File: 0719T30.D

		Compound	MEAN	CCRF	%D	%Drift
41	TM	TCE	0.3050	0.3112	2.1	TM
42	TM	2-Pentanone	0.2403	0.2399	0.15	TM
43	TM*	1,2-Dichloropropane	0.3661	0.3596	1.8	TM*
44	TM	Bromodichloromethane	0.5065	0.4910	3.1	TM
45	TM	Methyl Cyclohexane	0.2178	0.1937	11	TM
46	TM	Dibromomethane	0.1991	0.1997	0.25	TM
47	TML	2-Chloroethyl vinyl ether	0.0061	0.0058	5.7	TML 19
48	TM	MIBK (methyl isobutyl ketone)	0.1728	0.1756	1.6	TM
49	TM	1-Bromo-2-chloroethane	0.2547	0.2460	3.4	TM
50	TM	Cis-1,3-Dichloropropene	0.5012	0.4681	6.6	TM
51	TM*	Toluene	1.324	1.294	2.3	TM*
52	TM	Trans-1,3-Dichloropropene	0.4419	0.3995	9.6	TM
53	TM	1,1,2-TCA	0.2948	0.2755	6.5	TM
54	TM	2-Hexanone	0.1982	0.2041	3.0	TM
55	I	Chlorobenzene-D5 (IS)	ISTD			I
56	S	Toluene-D8(S)	1.478	1.440	2.6	S
57	TM	1,2-EDB	0.3748	0.3528	5.9	TM
58	TM	Tetrachloroethene	0.4238	0.3958	6.6	TM
59	TM	1-Chlorohexane	0.5045	0.4696	6.9	TM
60	TM	1,1,1,2-Tetrachloroethane	0.4952	0.4733	4.4	TM
61	TM	m&p-Xylene	0.7724	0.7473	3.3	TM
62	TM	o-Xylene	0.7990	0.7871	1.5	TM
63	TM	Styrene	1.358	1.344	1.0	TM
64	S	4-Bromofluorobenzene(S)	0.6990	0.6978	0.17	S
65	TM	1,3-Dichloropropane	0.6572	0.6315	3.9	TM
66	TM	Dibromochloromethane	0.4948	0.4681	5.4	TM
67	TM**	Chlorobenzene	1.292	1.221	5.5	TM**
68	TM*	Ethylbenzene	2.032	1.929	5.1	TM*
69	TM**	Bromoform	0.3388	0.3250	4.1	TM**
70	I	1,4-Dichlorobenzene-D (IS)	ISTD			I
71	TM	Isopropylbenzene	3.269	3.171	3.0	TM
72	TM**	1,1,2,2-Tetrachloroethane	0.9070	0.8329	8.2	TM**
73	TM	1,2,3-Trichloropropane	0.2574	0.2445	5.0	TM
74	TM	t-1,4-Dichloro-2-Butene	0.1723	0.1719	0.22	TM
75	TM	Bromobenzene	1.078	1.022	5.2	TM
76	TM	n-Propylbenzene	4.209	4.123	2.0	TM
77	TM	4-Ethyltoluene	3.614	3.563	1.4	TM
78	TM	2-Chlorotoluene	3.001	2.895	3.5	TM
79	TM	1,3,5-Trimethylbenzene	2.996	2.997	0.01	TM
80	TM	4-Chlorotoluene	2.971	2.935	1.2	TM

Average

3.8

MS 7/27/12

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: Water

SDG No: 68258
Date Analyzed: 07/19/12
Instrument: Thor
Cal. Date: 07/19/12
Data File: 0719T30.D

		Compound	MEAN	CCRF	%D	%Drift
81	TM	Tert-Butylbenzene	2.745	2.626	4.3	TM
82	TM	1,2,4-Trimethylbenzene	3.100	3.057	1.4	TM
83	TM	Sec-Butylbenzene	3.664	3.593	2.0	TM
84	TM	p-Isopropyltoluene	3.096	3.026	2.3	TM
85	TM	Benzyl Chloride	0.9252	0.5995	35	TM
86	TM	1,3-DCB	2.038	1.945	4.5	TM
87	TM	1,4-DCB	2.134	1.972	7.6	TM
88	TM	n-Butylbenzene	2.775	2.582	7.0	TM
89	TM	1,2-DCB	1.975	1.872	5.2	TM
90	TM	Hexachloroethane	0.5673	0.5003	12	TM
91	TM	1,2-Dibromo-3-chloropropane	0.1699	0.1792	5.4	TM
92	TM	1,2,4-Trichlorobenzene	0.9054	0.8728	3.6	TM
93	TM	Hexachlorobutadiene	0.3782	0.3394	10	TM
94	TM	Naphthalene	2.528	2.547	0.74	TM
95	TM	1,2,3-Trichlorobenzene	1.290	1.249	3.1	TM
96						
97						
98						
99						
100						
101						
102						
103						
104						
105						
106						
107						
108						
109						
110						
111						
112						
113						
114						
115						
116						
117						
118						
119						
120						

*NT

Average

6.9

ARS 7/27/12

Data File : M:\THOR\DATA\T120719\0719T30.D
 Acq On : 19 Jul 12 22:35
 Sample : 10ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 30
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 10:45 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 10:40:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	96	452736	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.87	117	376000	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	220224	25.00000	ppb	0.00

System Monitoring Compounds

31) Dibromofluoromethane(S)	5.95	111	225646	31.84967	ppb	0.00
Spiked Amount	31.881		Recovery	=	99.902%	
36) 1,2-DCA-D4(S)	6.33	65	225427	34.23774	ppb	0.00
Spiked Amount	33.647		Recovery	=	101.757%	
56) Toluene-D8(S)	8.43	98	808613	36.37690	ppb	0.00
Spiked Amount	37.345		Recovery	=	97.408%	
64) 4-Bromofluorobenzene(S)	11.05	95	309746	29.46501	ppb	0.00
Spiked Amount	29.515		Recovery	=	99.830%	

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.30	85	21112	9.20611	ppb	98
3) Freon 114	1.41	85	28914	9.07008	ppb	93
4) Chloromethane	1.45	50	53915	9.42929	ppb	99
5) Vinyl chloride	1.56	62	85149	9.51612	ppb	100
6) Bromomethane	1.87	94	52548	9.18808	ppb	97
7) Chloroethane	1.97	64	49476	9.60029	ppb	90
8) Dichlorofluoromethane	2.18	67	2999	8.94765	ppb	# 79
9) Trichlorofluoromethane	2.24	101	17719	9.58052	ppb	99
11) Acetone	2.89	43	17505	10.70365	ppb	99
12) Freon-113	2.85	101	33955	9.12750	ppb	95
13) 1,1-DCE	2.82	61	47256	9.46594	ppb	97
14) t-Butanol	3.69	59	19648	133.83058	ppb	98
15) Methyl Acetate	3.34	43	42726	9.76130	ppb	99
16) Iodomethane	2.98	142	43831	9.70694	ppb	98
17) Acrylonitrile	3.81	52	15078	10.54515	ppb	82
18) Methylene chloride	3.45	84	16095	8.80124	ppb	99
19) Carbon disulfide	3.06	76	4973	9.72911	ppb	# 82
20) Methyl t-butyl ether (MtBE)	3.90	73	91548	9.49816	ppb	98
21) Trans-1,2-DCE	3.87	96	30943	8.98131	ppb	90
22) Diisopropyl Ether	4.71	59	22285	10.32400	ppb	99
23) 1,1-DCA	4.51	63	90691	9.92708	ppb	98
24) Vinyl Acetate	4.71	87	49188	9.53261	ppb	93
25) Ethyl tert Butyl Ether	5.21	59	116957	9.70566	ppb	98
26) MEK (2-Butanone)	5.38	43	22460	10.12955	ppb	91
27) Cis-1,2-DCE	5.33	96	57221	9.77754	ppb	95
28) 2,2-Dichloropropane	5.32	77	29359	7.97737	ppb	98
29) Chloroform	5.76	83	110917	9.77685	ppb	96
30) Bromochloromethane	5.62	128	28139	9.87532	ppb	92
32) 1,1,1-TCA	5.96	97	65837	9.64484	ppb	99
33) Cyclohexane	6.04	41	17788	9.60203	ppb	98
34) 1,1-Dichloropropene	6.17	75	46858	9.45227	ppb	97
35) 2,2,4-Trimethylpentane	6.55	57	56506	7.93090	ppb	93
37) Carbon Tetrachloride	6.17	117	60992	9.53330	ppb	96
38) Tert Amyl Methyl Ether	6.59	73	123225	9.60738	ppb	98
39) 1,2-DCA	6.42	62	70817	9.51873	ppb	98
40) Benzene	6.40	78	193154	9.50568	ppb	99
41) TCE	7.15	95	56364	10.20509	ppb	97
42) 2-Pentanone	7.36	43	543080	124.81033	ppb	99
43) 1,2-Dichloropropane	7.37	63	65114	9.82099	ppb	97

(#) = qualifier out of range (m) = manual integration

Data File : M:\THOR\DATA\T120719\0719T30.D
 Acq On : 19 Jul 12 22:35
 Sample : 10ug/L Vol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 30
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 10:45 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 10:40:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
44) Bromodichloromethane	7.68	83	88920	9.69476	ppb	99
45) Methyl Cyclohexane	7.36	83	35085	8.89704	ppb	99
46) Dibromomethane	7.49	93	36156	10.02526	ppb	91
47) 2-Chloroethyl vinyl ether	7.99	106	1046	8.05600	ppb	100
48) MIBK (methyl isobutyl ket	8.33	43	31800	10.16200	ppb	99
49) 1-Bromo-2-chloroethane	7.99	63	44552	9.65799	ppb	98
50) Cis-1,3-Dichloropropene	8.15	75	84767	9.33930	ppb	99
51) Toluene	8.50	91	234345	9.77470	ppb	99
52) Trans-1,3-Dichloropropene	8.72	75	72356	9.04130	ppb	99
53) 1,1,2-TCA	8.90	83	49884	9.34511	ppb	96
54) 2-Hexanone	9.18	43	36953	10.29763	ppb	92
57) 1,2-EDB	9.40	107	53068	9.41345	ppb	98
58) Tetrachloroethene	9.06	166	59525	9.33824	ppb	95
59) 1-Chlorohexane	9.90	91	70621	9.30765	ppb	98
60) 1,1,1,2-Tetrachloroethane	9.99	131	71180	9.55805	ppb	98
61) m&p-Xylene	10.14	106	224782	19.34899	ppb	97
62) o-Xylene	10.54	106	118374	9.85006	ppb	97
63) Styrene	10.55	104	202135	9.89948	ppb	99
65) 1,3-Dichloropropane	9.07	76	94972	9.60845	ppb	100
66) Dibromochloromethane	9.29	129	70401	9.46043	ppb	99
67) Chlorobenzene	9.90	112	183635	9.44678	ppb	98
68) Ethylbenzene	10.03	91	290081	9.49046	ppb	99
69) Bromoform	10.71	173	48885	9.59403	ppb	93
71) Isopropylbenzene	10.91	105	279290	9.69932	ppb	97
72) 1,1,2,2-Tetrachloroethane	11.19	83	73373	9.18301	ppb	99
73) 1,2,3-Trichloropropane	11.23	110	21535	9.49610	ppb	91
74) t-1,4-Dichloro-2-Butene	11.25	53	15144	9.97836	ppb	95
75) Bromobenzene	11.19	156	89995	9.48078	ppb	98
76) n-Propylbenzene	11.32	91	363226	9.79728	ppb	99
77) 4-Ethyltoluene	11.43	105	313892	9.85902	ppb	98
78) 2-Chlorotoluene	11.39	91	254998	9.64544	ppb	99
79) 1,3,5-Trimethylbenzene	11.50	105	263962	10.00144	ppb	100
80) 4-Chlorotoluene	11.50	91	258569	9.88101	ppb	100
81) Tert-Butylbenzene	11.82	119	231316	9.56718	ppb	98
82) 1,2,4-Trimethylbenzene	11.86	105	269333	9.86200	ppb	99
83) Sec-Butylbenzene	12.04	105	316488	9.80462	ppb	99
84) p-Isopropyltoluene	12.19	119	266591	9.77446	ppb	99
85) Benzyl Chloride	12.35	91	52811	6.47962	ppb	100
86) 1,3-DCB	12.13	146	171365	9.54592	ppb	99
87) 1,4-DCB	12.22	146	173724	9.24038	ppb	98
88) n-Butylbenzene	12.59	91	227452	9.30400	ppb	99
89) 1,2-DCB	12.59	146	164890	9.47686	ppb	97
90) Hexachloroethane	12.86	117	44069	8.81878	ppb	96
91) 1,2-Dibromo-3-chloropropan	13.35	157	15783	10.54375	ppb	90
92) 1,2,4-Trichlorobenzene	14.20	180	76888	9.64054	ppb	96
93) Hexachlorobutadiene	14.38	223	29896	8.97326	ppb	94
94) Naphthalene	14.43	128	224347	10.07414	ppb	99
95) 1,2,3-Trichlorobenzene	14.68	180	110057	9.68516	ppb	98

(#) = qualifier out of range (m) = manual integration
 0719T30.D TALLW.M Fri Jul 20 10:53:27 2012

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 6
Initial Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: Water

SDG No: 68258
Initial Cal. Date: 07/19/12
Instrument: Thor (TGAS.M)

Initials: _____

0719T18.D 0719T19.D 0719T20.D 0719T21.D 0719T22.D 0719T23.D 0719T24.D

	Compound	20	50	100	300	600	800	1000				Avg	%RSD		r2
1	I Fluorobenzene (IS)														
2	TMHBL Gasoline	10.2	4.491	2.752	1.655	1.327	1.282	1.266				3.3	100	TMHBL	0.999
3	I Chlorobenzene-D5 (IS)														
4	I 1,4-Dichlorobenzene-D (IS)														
5															
6															
7															
8															
9															
10															
11															
12															
13															
14															
15															
16															
17															
18															
19															
20															
21															
22															
23															
24															
25															
26															
27															
28															
29															
30															
31															
32															
33															
34															
35															

MRS 7/26/12

Data File : M:\THOR\DATA\T120719\0719T17.D Vial: 17
 Acq On : 19 Jul 12 16:35 Operator: DG,RS,HW,ARS,SV
 Sample : VOC MIX MARKER Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 26 15:56 2012 Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 10:40:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.72	96	449600	25.00000	ppb	-0.01
55) Chlorobenzene-D5 (IS)	9.87	117	364480	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	212096	25.00000	ppb	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
31) Dibromofluoromethane(S)	6.05	111	829	0.11783	ppb	0.10
Spiked Amount	31.881		Recovery	=	0.370%	
36) 1,2-DCA-D4(S)	6.32	65	1100	0.16823	ppb	-0.01
Spiked Amount	33.647		Recovery	=	0.499%	
56) Toluene-D8(S)	8.43	98	4329	0.20090	ppb	0.00
Spiked Amount	37.345		Recovery	=	0.538%	
64) 4-Bromofluorobenzene(S)	11.05	95	3105	0.30470	ppb	0.00
Spiked Amount	29.515		Recovery	=	1.033%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.26	85	175	0.07684	ppb	# 43
3) Freon 114	1.39	85	73	-0.36210	ppb	89
4) Chloromethane	1.42	50	408	-0.35112	ppb	90
11) Acetone	2.87	43	1560	-0.18868	ppb	# 82
14) t-Butanol	3.67	59	137	0.93967	ppb	# 72
15) Methyl Acetate	3.49	43	392147	102.18783	ppb	# 52
18) Methylene chloride	3.41	84	352	-0.72918	ppb	84
19) Carbon disulfide	3.03	76	116	-0.63444	ppb	# 65
26) MEK (2-Butanone)	5.37	43	934	0.74996	ppb	# 46
34) 1,1-Dichloropropene	6.04	75	26930	5.47025	ppb	# 49
35) 2,2,4-Trimethylpentane	6.53	57	915	0.12932	ppb	# 80
37) Carbon Tetrachloride	6.04	117	36940	5.81415	ppb	# 14
38) Tert Amyl Methyl Ether	6.72	73	10154	0.79719	ppb	# 27
39) 1,2-DCA	6.39	62	8679	1.17471	ppb	# 74
40) Benzene	6.39	78	1047885	51.92920	ppb	98
45) Methyl Cyclohexane	7.35	83	655	0.16726	ppb	76
51) Toluene	8.49	91	1111309	46.67683	ppb	100
58) Tetrachloroethene	9.06	166	449	0.07267	ppb	# 39
59) 1-Chlorohexane	10.03	91	1239543	168.53186	ppb	# 17
61) m&p-Xylene	10.14	106	972156	86.32704	ppb	99
62) o-Xylene	10.53	106	481637	41.34437	ppb	95
63) Styrene	10.54	104	24726	1.24922	ppb	# 1
68) Ethylbenzene	10.03	91	1239770	41.84303	ppb	99
79) 1,3,5-Trimethylbenzene	11.49	105	5894	0.23188	ppb	98
80) 4-Chlorotoluene	11.50	91	1859	0.07376	ppb	# 80
81) Tert-Butylbenzene	11.86	119	127152	5.46052	ppb	# 70
82) 1,2,4-Trimethylbenzene	11.86	105	1005826	38.24110	ppb	99
83) Sec-Butylbenzene	11.86	105	976689	31.41682	ppb	# 55
88) n-Butylbenzene	12.59	91	2153	0.09144	ppb	91
92) 1,2,4-Trichlorobenzene	14.20	180	890	0.11587	ppb	# 76
93) Hexachlorobutadiene	14.39	223	372	0.11593	ppb	# 54
94) Naphthalene	14.43	128	825392	38.48401	ppb	100
95) 1,2,3-Trichlorobenzene	14.67	180	875	0.07995	ppb	96

ARS 7/26/12

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T17.D
Acq On : 19 Jul 12 16:35
Sample : VOC MIX MARKER
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 17
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 26 15:56 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 10:40:23 2012
Response via : Initial Calibration



Data File : M:\THOR\DATA\T120719\0719T18.D Vial: 18
 Acq On : 19 Jul 12 17:02 Operator: DG,RS,HW,ARS,SV
 Sample : 20ug/LVol Std 07-19-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 24 13:30 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	863157	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.88	TIC	980787	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1107225	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	9.88	TIC	7045985m	24.82016	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T18.D
Acq On : 19 Jul 12 17:02
Sample : 20ug/LVol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 18
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 13:30 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T18.D
 Acq On : 19 Jul 12 17:02
 Sample : 20ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:23 2012

Vial: 18
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T18.D



(2) Gasoline (TMHB)		
8.50min	-49.5217ppb m	
response	4305439	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	2.48#
0.00	1.80	7.19#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T18.D
 Acq On : 19 Jul 12 17:02
 Sample : 20ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:30 2012

Vial: 18
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T18.D

(2) Gasoline (TMHB)		
9.88min	24.8202ppb m	
response	7045985	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	1.52#
0.00	1.80	4.39#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120719\0719T19.D Vial: 19
 Acq On : 19 Jul 12 17:30 Operator: DG,RS,HW,ARS,SV
 Sample : 50ug/LVol Std 07-19-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 24 13:29 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	931554	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.88	TIC	1035987	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1151281	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	9.88	TIC	8367373m	43.99965	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T19.D
Acq On : 19 Jul 12 17:30
Sample : 50ug/LVol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 19
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 13:29 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T19.D
 Acq On : 19 Jul 12 17:30
 Sample : 50ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:23 2012

Vial: 19
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T19.D



(2) Gasoline (TMHB)		
8.50min	-28.6219ppb m	
response	5478107	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	2.05#
0.00	1.80	5.90#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T19.D
 Acq On : 19 Jul 12 17:30
 Sample : 50ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:29 2012

Vial: 19
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T19.D

(2) Gasoline (TMHB)		
9.88min	43.9996ppb m	
response	8367373	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	1.34#
0.00	1.80	3.86#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120719\0719T20.D Vial: 20
 Acq On : 19 Jul 12 17:58 Operator: DG,RS,HW,ARS,SV
 Sample : 100ug/LVol Std 07-19-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 24 13:28 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	913405	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.88	TIC	1040970	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1159278	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	9.88	TIC	10053011m	91.38877	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T20.D
Acq On : 19 Jul 12 17:58
Sample : 100ug/LVol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 20
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 13:28 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T20.D
 Acq On : 19 Jul 12 17:58
 Sample : 100ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:23 2012

Vial: 20
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T20.D



(2) Gasoline (TMHB)		
8.50min	18.9965ppb m	
response	7228978	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	1.56#
0.00	1.80	4.52#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T20.D
 Acq On : 19 Jul 12 17:58
 Sample : 100ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:28 2012

Vial: 20
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T20.D

(2) Gasoline (TMHB)		
9.88min	91.3888ppb m	
response	10053011	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	1.12#
0.00	1.80	3.25#
0.00	0.00	0.00

Quantitation Report (QT Reviewed)

Data File : M:\THOR\DATA\T120719\0719T21.D Vial: 21
 Acq On : 19 Jul 12 18:26 Operator: DG,RS,HW,ARS,SV
 Sample : 300ug/LVol Std 07-19-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 24 13:27 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	908946	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.88	TIC	1020089	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1167558	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	18049652m	298.64746	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T21.D
Acq On : 19 Jul 12 18:26
Sample : 300ug/LVol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 21
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 13:27 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T21.D
 Acq On : 19 Jul 12 18:26
 Sample : 300ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:25 2012

Vial: 21
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T21.D



(2) Gasoline (TMHB)		
8.50min	216.2049ppb m	
response	14849259	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.75#
0.00	1.80	2.19#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T21.D
 Acq On : 19 Jul 12 18:26
 Sample : 300ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:27 2012

Vial: 21
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T21.D

(2) Gasoline (TMHB)		
8.50min	298.6475ppb m	
response	18049652	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.61#
0.00	1.80	1.80#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120719\0719T22.D Vial: 22
 Acq On : 19 Jul 12 18:54 Operator: DG,RS,HW,ARS,SV
 Sample : 600ug/LVol Std 07-19-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 24 13:25 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	961687	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.88	TIC	1052445	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1221581	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	30636947m	579.61570	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T22.D
Acq On : 19 Jul 12 18:54
Sample : 600ug/LVol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 22
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 13:25 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T22.D
 Acq On : 19 Jul 12 18:54
 Sample : 600ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:23 2012

Vial: 22
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T22.D



(2) Gasoline (TMHB)		
8.50min	476.3926ppb m	
response	26397348	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.44#
0.00	1.80	1.29#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T22.D
 Acq On : 19 Jul 12 18:54
 Sample : 600ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:25 2012

Vial: 22
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T22.D

(2) Gasoline (TMHB)
 8.50min 579.6157ppb m
 response 30636947

Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.38#
0.00	1.80	1.11#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120719\0719T23.D Vial: 23
 Acq On : 19 Jul 12 19:21 Operator: DG,RS,HW,ARS,SV
 Sample : 800ug/LVol Std 07-19-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 24 13:24 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	935153	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	1058737	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1213057	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	38356495m	794.06441	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T23.D
Acq On : 19 Jul 12 19:21
Sample : 800ug/LVol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 23
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 13:24 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T23.D
Acq On : 19 Jul 12 19:21
Sample : 800ug/LVol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12
Quant Time: Jul 24 13:23 2012

Vial: 23
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00
Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 12:10:59 2012
Response via : Single Level Calibration



TIC: 0719T23.D

(2) Gasoline (TMHB)
8.50min 688.9101ppb m
response 34156743



Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.35#
0.00	1.80	0.99#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T23.D
 Acq On : 19 Jul 12 19:21
 Sample : 800ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:24 2012

Vial: 23
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T23.D

(2) Gasoline (TMHB)		
8.50min	794.0644ppb m	
response	38356495	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.31#
0.00	1.80	0.88#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120719\0719T24.D Vial: 24
 Acq On : 19 Jul 12 19:49 Operator: DG,RS,HW,ARS,SV
 Sample : 1000ug/LVol Std 07-19-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 24 13:24 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	909093	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	1037932	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1187621	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	46029569m	1019.22195	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T24.D
Acq On : 19 Jul 12 19:49
Sample : 1000ug/LVol Std 07-19-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 24
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 13:24 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T24.D
 Acq On : 19 Jul 12 19:49
 Sample : 1000ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:23 2012

Vial: 24
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration



TIC: 0719T24.D



(2) Gasoline (TMHB)		
8.50min	905.1239ppb m	
response	41599602	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.27#
0.00	1.80	0.80#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T24.D
 Acq On : 19 Jul 12 19:49
 Sample : 1000ug/LVol Std 07-19-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:24 2012

Vial: 24
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 12:10:59 2012
 Response via : Single Level Calibration




TIC: 0719T24.D

(2) Gasoline (TMHB)		
8.50min	1019.2219ppb m	
response	46029569	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.24#
0.00	1.80	0.72#
0.00	0.00	0.00

Gasoline

Response Ratio



Resp Ratio = 1.08e+000 * Amt + 6.84e+000
Coef of Det (r^2) = 0.999 Curve Fit: Linear

Method Name: M:\THOR\DATA\T120719\TGAS.M
Calibration Table Last Updated: Tue Jul 24 13:31:25 2012

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 7

Second Source Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: Water

SDG No: 68258
Date Analyzed: 07/19/12
Instrument: Thor
Initial Cal. Date: 07/19/12
Data File: 0719T33.D

	Compound	MEAN	CCRF	%D	%Drift	
1	Fluorobenzene (IS)	ISTD				
2	TMHB Gasoline	3.282	1.857	43	TMHBL	19
3	Chlorobenzene-D5 (IS)	ISTD				
4	1,4-Dichlorobenzene-D (IS)	ISTD				
5						
6						
7						
8						
9						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						
23						
24						
25						
26						
27						
28						
29						
30						
31						
32						
33						
34						
35						
36						
37						
38						
39						
40						

Average

43.0

RR: 7/31/12

Data File : M:\THOR\DATA\T120719\0719T33.D Vial: 33
 Acq On : 19 Jul 12 23:59 Operator: DG,RS,HW,ARS,SV
 Sample : CCV gas 300ug/L (SS) Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 24 13:33 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	931728	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	1044016	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1183160	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	20760218m	358.46091	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T33.D
Acq On : 19 Jul 12 23:59
Sample : CCV gas 300ug/L (SS)
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 33
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 13:33 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T33.D
 Acq On : 19 Jul 12 23:59
 Sample : CCV gas 300ug/L (SS)
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:33 2012

Vial: 33
 Operator: DG, RS, HW, ARS, SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Single Level Calibration



TIC: 0719T33.D



(2) Gasoline (TMHB)		
8.50min	280.7311ppb m	
response	17639584	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.65#
0.00	1.80	1.90#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T33.D
 Acq On : 19 Jul 12 23:59
 Sample : CCV gas 300ug/L (SS)
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:33 2012

Vial: 33
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Single Level Calibration



TIC: 0719T33.D

(2) Gasoline (TMHB)		
8.50min	358.4609ppb m	
response	20760218	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.55#
0.00	1.80	1.61#
0.00	0.00	0.00

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: Water

SDG No: 68258
Date Analyzed: 07/20/12
Instrument: Thor
Initial Cal. Date: 07/19/12
Data File: 0719T34.D

	Compound	MEAN	CCRF	%D	%Drift	
1	Fluorobenzene (IS)	ISTD				
2	TMHB Gasoline	3.282	1.987	39	TMHBL	32 *
3	Chlorobenzene-D5 (IS)	ISTD				
4	1,4-Dichlorobenzene-D (IS)	ISTD				
5						
6						
7						
8						
9						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						
21						
22						
23						
24						
25						
26						
27						
28						
29						
30						
31						
32						
33						
34						
35						
36						
37						
38						
39						
40						

Average

39.0

AKS 7/31/12

Data File : M:\THOR\DATA\T120719\0719T34.D Vial: 34
 Acq On : 20 Jul 12 00:27 Operator: DG,RS,HW,ARS,SV
 Sample : CCV gas 300ug/L Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 24 13:36 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	923126	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	1020885	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1142879	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	22015447m	394.83647	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T34.D
Acq On : 20 Jul 12 00:27
Sample : CCV gas 300ug/L
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 34
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 13:36 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T34.D
 Acq On : 20 Jul 12 00:27
 Sample : CCV gas 300ug/L
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:33 2012

Vial: 34
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Single Level Calibration



TIC: 0719T34.D



(2) Gasoline (TMHB)		
8.50min	320.1392ppb m	
response	19044247	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.58#
0.00	1.80	1.68#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T34.D
 Acq On : 20 Jul 12 00:27
 Sample : CCV gas 300ug/L
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:36 2012

Vial: 34
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Single Level Calibration



TIC: 0719T34.D

(2) Gasoline (TMHB)		
8.50min	394.8365ppb m	
response	22015447	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.50#
0.00	1.80	1.45#
0.00	0.00	0.00

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 6
Initial Calibration

Lab Name: APPL, Inc. _____

SDG No: 68258

Case No: _____

Initial Cal. Date: 07/25/12

Matrix: Water

Instrument: Thor (TGAS.M)

Initials: _____

0725T04.D 0725T05.D 0725T06.D 0725T07.D 0725T08.D 0725T09.D 0725T10.D

	Compound	20	50	100	300	600	800	1000				Avg	%RSD		r2
1	I Fluorobenzene (IS)	ISTD													
2	TMHBL Gasoline	16.5	7.205	4.047	2.093	1.605	1.465	1.393				4.9	113	TMHBL	1.000
3	I Chlorobenzene-D5 (IS)	ISTD													
4	I 1,4-Dichlorobenzene-D (IS)	ISTD													
5															
6															
7															
8															
9															
10															
11															
12															
13															
14															
15															
16															
17															
18															
19															
20															
21															
22															
23															
24															
25															
26															
27															
28															
29															
30															
31															
32															
33															
34															
35															

ARS 7/26/12

Data File : M:\THOR\DATA\T120725\0725T03.D
 Acq On : 25 Jul 12 10:22
 Sample : VOC MIX MARKER
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 2
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 26 8:59 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 10:40:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	96	383424	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.87	117	310848	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	187136	25.00000	ppb	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
31) Dibromofluoromethane(S)	5.95	111	196549	32.75773	ppb	0.00
Spiked Amount				31.881		
				Recovery	= 102.750%	
36) 1,2-DCA-D4(S)	6.33	65	189874	34.05104	ppb	0.00
Spiked Amount				33.647		
				Recovery	= 101.202%	
56) Toluene-D8(S)	8.43	98	687242	37.39680	ppb	0.00
Spiked Amount				37.345		
				Recovery	= 100.140%	
64) 4-Bromofluorobenzene(S)	11.05	95	268751	30.92365	ppb	0.00
Spiked Amount				29.515		
				Recovery	= 104.773%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Chloromethane	1.45	50	159	-0.39190	ppb	# 74
6) Bromomethane	1.78	94	376	0.07763	ppb	# 3
11) Acetone	2.90	43	3396	1.47860	ppb	98
14) t-Butanol	3.69	59	126	1.01338	ppb	# 72
15) Methyl Acetate	3.34	43	3113	-0.48779	ppb	93
18) Methylene chloride	3.45	84	326	-0.71073	ppb	84
23) 1,1-DCA	4.34	63	775	0.10017	ppb	# 1
26) MEK (2-Butanone)	5.39	43	1036	0.87321	ppb	# 46
34) 1,1-Dichloropropene	6.05	75	22005	5.24130	ppb	# 48
35) 2,2,4-Trimethylpentane	6.55	57	913	0.15131	ppb	91
37) Carbon Tetrachloride	6.05	117	28709	5.29852	ppb	# 14
38) Tert Amyl Methyl Ether	6.73	73	8830	0.81289	ppb	# 29
39) 1,2-DCA	6.40	62	6268	0.99480	ppb	# 74
40) Benzene	6.40	78	769435	44.71126	ppb	98
48) MIBK (methyl isobutyl ket	8.43	43	1645	0.62070	ppb	# 1
51) Toluene	8.50	91	828486	40.80362	ppb	100
58) Tetrachloroethene	9.06	166	842	0.15978	ppb	84
59) 1-Chlorohexane	10.03	91	895259	142.72325	ppb	# 17
61) m&p-Xylene	10.14	106	710590	73.98703	ppb	98
62) o-Xylene	10.54	106	355718	35.80371	ppb	99
63) Styrene	10.54	104	17860	1.05802	ppb	# 1
68) Ethylbenzene	10.03	91	895459	35.43670	ppb	99
79) 1,3,5-Trimethylbenzene	11.50	105	3503	0.15620	ppb	89
81) Tert-Butylbenzene	11.86	119	92293	4.49215	ppb	# 73
82) 1,2,4-Trimethylbenzene	11.86	105	731223	31.50884	ppb	99
83) Sec-Butylbenzene	11.86	105	709314	25.85946	ppb	# 55
94) Naphthalene	14.43	128	598073	31.60454	ppb	99

ARS 7/26/12

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T03.D
Acq On : 25 Jul 12 10:22
Sample : VOC MIX MARKER
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 2
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 26 8:59 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 10:40:23 2012
Response via : Initial Calibration



Data File : M:\THOR\DATA\T120725\0725T04.D Vial: 3
 Acq On : 25 Jul 12 10:50 Operator: DG,RS,HW,ARS,SV
 Sample : 20ug/L Vol Std 07-25-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 25 15:59 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	757122	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	882358	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	975664	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.43	TIC	10003915m	-268.75372	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T04.D
Acq On : 25 Jul 12 10:50
Sample : 20ug/L Vol Std 07-25-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 3
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 25 15:59 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T04.D
 Acq On : 25 Jul 12 10:50
 Sample : 20ug/L Vol Std 07-25-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:53 2012

Vial: 3
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Multiple Level Calibration



TIC: 0725T04.D

(2) Gasoline (TMHB)

8.50min -376.6351ppb m

response 7759068



Ion	Exp%	Act%
TIC	100	100
0.00	0.50	1.20#
0.00	1.40	3.55#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T04.D
 Acq On : 25 Jul 12 10:50
 Sample : 20ug/L Vol Std 07-25-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:59 2012

Vial: 3
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Multiple Level Calibration



TIC: 0725T04.D

(2) Gasoline (TMHB)

8.43min -268.7537ppb m

response 10003915

Ion	Exp%	Act%
TIC	100	100
0.00	0.50	0.93#
0.00	1.40	2.76#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120725\0725T05.D Vial: 4
 Acq On : 25 Jul 12 11:17 Operator: DG,RS,HW,ARS,SV
 Sample : 50ug/L Vol Std 07-25-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 25 15:58 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	757407	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	877869	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	954185	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.43	TIC	10913490m	-225.23930	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T05.D
Acq On : 25 Jul 12 11:17
Sample : 50ug/L Vol Std 07-25-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 4
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 25 15:58 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T05.D
 Acq On : 25 Jul 12 11:17
 Sample : 50ug/L Vol Std 07-25-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:53 2012

Vial: 4
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Multiple Level Calibration



TIC: 0725T05.D

(2) Gasoline (TMHB)

8.50min -333.5537ppb m

response 8658785



Ion	Exp%	Act%
TIC	100	100
0.00	0.50	1.05#
0.00	1.40	3.08#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T05.D
 Acq On : 25 Jul 12 11:17
 Sample : 50ug/L Vol Std 07-25-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:58 2012

Vial: 4
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Multiple Level Calibration



TIC: 0725T05.D

(2) Gasoline (TMHB)

8.43min -225.2393ppb m

response 10913490

Ion	Exp%	Act%
TIC	100	100
0.00	0.50	0.83#
0.00	1.40	2.44#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120725\0725T06.D Vial: 5
 Acq On : 25 Jul 12 11:45 Operator: DG,RS,HW,ARS,SV
 Sample : 100ug/L Vol Std 07-25-12 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 25 15:58 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	774747	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	873528	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	976201	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.43	TIC	12540540m	-160.56049	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T06.D
Acq On : 25 Jul 12 11:45
Sample : 100ug/L Vol Std 07-25-12
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 5
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 25 15:58 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T06.D
 Acq On : 25 Jul 12 11:45
 Sample : 100ug/L Vol Std 07-25-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:58 2012

Vial: 5
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Multiple Level Calibration



TIC: 0725T06.D

(2) Gasoline (TMHB)

8.43min -160.5605ppb m

response 12540540



Ion	Exp%	Act%
TIC	100	100
0.00	0.50	0.76#
0.00	1.40	2.17#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T06.D
 Acq On : 25 Jul 12 11:45
 Sample : 100ug/L Vol Std 07-25-12
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:53 2012

Vial: 5
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Multiple Level Calibration



TIC: 0725T06.D

(2) Gasoline (TMHB)

8.50min -268.9292ppb m

response 10233059

Ion	Exp%	Act%
TIC	100	100
0.00	0.50	0.93#
0.00	1.40	2.66#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120725\0725T07.D Vial: 6
 Acq On : 25 Jul 12 12:13 Operator: DG,RS,HW,ARS,SV
 Sample : 300ug/L Vol Std 07-25-13 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 25 15:50 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	782981	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	897407	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	996199	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	19663639m	410.65057	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T07.D
Acq On : 25 Jul 12 12:13
Sample : 300ug/L Vol Std 07-25-13
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 6
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 25 15:50 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T07.D
 Acq On : 25 Jul 12 12:13
 Sample : 300ug/L Vol Std 07-25-13
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:49 2012

Vial: 6
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Single Level Calibration



TIC: 0725T07.D

(2) Gasoline (TMHB)

8.50min 339.0063ppb m

response 17146776



Ion	Exp%	Act%
TIC	100	100
0.00	0.00	0.55#
0.00	0.00	1.63#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T07.D
 Acq On : 25 Jul 12 12:13
 Sample : 300ug/L Vol Std 07-25-13
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:50 2012

Vial: 6
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Single Level Calibration



TIC: 0725T07.D

(2) Gasoline (TMHB)

8.50min 410.6506ppb m

response 19663639

Ion	Exp%	Act%
TIC	100	100
0.00	0.00	0.48#
0.00	0.00	1.42#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120725\0725T08.D Vial: 7
 Acq On : 25 Jul 12 12:41 Operator: DG,RS,HW,ARS,SV
 Sample : 600ug/L Vol Std 07-25-14 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 25 15:56 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	782399	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	890063	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	996015	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	30141216m	652.19460	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T08.D
Acq On : 25 Jul 12 12:41
Sample : 600ug/L Vol Std 07-25-14
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 7
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 25 15:56 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T08.D
 Acq On : 25 Jul 12 12:41
 Sample : 600ug/L Vol Std 07-25-14
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:53 2012

Vial: 7
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Multiple Level Calibration



TIC: 0725T08.D



(2) Gasoline (TMHB)		
8.50min	500.4974ppb m	
response	26879245	
Ion	Exp%	Act%
TIC	100	100
0.00	0.50	0.36#
0.00	1.40	1.04#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T08.D
 Acq On : 25 Jul 12 12:41
 Sample : 600ug/L Vol Std 07-25-14
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:56 2012

Vial: 7
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Multiple Level Calibration



TIC: 0725T08.D

(2) Gasoline (TMHB)

8.50min 652.1946ppb m

response 30141216

Ion	Exp%	Act%
TIC	100	100
0.00	0.50	0.32#
0.00	1.40	0.93#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120725\0725T09.D Vial: 8
 Acq On : 25 Jul 12 13:08 Operator: DG,RS,HW,ARS,SV
 Sample : 800ug/L Vol Std 07-25-15 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 25 15:55 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	788221	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.88	TIC	883861	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1013991	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	36946726m	955.99215	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T09.D
Acq On : 25 Jul 12 13:08
Sample : 800ug/L Vol Std 07-25-15
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 8
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 25 15:55 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T09.D
 Acq On : 25 Jul 12 13:08
 Sample : 800ug/L Vol Std 07-25-15
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:53 2012

Vial: 8
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Multiple Level Calibration



TIC: 0725T09.D

(2) Gasoline (TMHB)

8.50min 790.6203ppb m

response 33364245



Ion	Exp%	Act%
TIC	100	100
0.00	0.50	0.30#
0.00	1.40	0.86#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T09.D
Acq On : 25 Jul 12 13:08
Sample : 800ug/L Vol Std 07-25-15
Misc : 10ml w/5ul of IS&S: 06-7-12
Quant Time: Jul 25 15:55 2012

Vial: 8
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00
Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 08:14:32 2012
Response via : Multiple Level Calibration



TIC: 0725T09.D

(2) Gasoline (TMHB)

8.50min 955.9921ppb m

response 36946726

Ion	Exp%	Act%
TIC	100	100
0.00	0.50	0.27#
0.00	1.40	0.77#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120725\0725T10.D Vial: 9
 Acq On : 25 Jul 12 13:36 Operator: DG,RS,HW,ARS,SV
 Sample : 1000ug/L Vol Std 07-25-16 Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 25 16:00 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	808332	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	927489	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1069004	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	45050186m	1278.31907	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T10.D
Acq On : 25 Jul 12 13:36
Sample : 1000ug/L Vol Std 07-25-16
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 9
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 25 16:00 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T10.D
 Acq On : 25 Jul 12 13:36
 Sample : 1000ug/L Vol Std 07-25-16
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 15:53 2012

Vial: 9
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Multiple Level Calibration



TIC: 0725T10.D

(2) Gasoline (TMHB)

8.50min 1108.4543ppb m

response 41276485



Ion	Exp%	Act%
TIC	100	100
0.00	0.50	0.25#
0.00	1.40	0.73#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T10.D
 Acq On : 25 Jul 12 13:36
 Sample : 1000ug/L Vol Std 07-25-16
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 25 16:00 2012

Vial: 9
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 08:14:32 2012
 Response via : Multiple Level Calibration




TIC: 0725T10.D

(2) Gasoline (TMHB)

8.50min 1278.3191ppb m

response 45050186

Ion	Exp%	Act%
TIC	100	100
0.00	0.50	0.23#
0.00	1.40	0.67#
0.00	0.00	0.00



Method Name: M:\THOR\DATA\T120725\TGAS.M
Calibration Table Last Updated: Wed Jul 25 16:07:29 2012

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 7

Second Source Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: Water

SDG No: 69258
Date Analyzed: 07/25/12
Instrument: Thor
Initial Cal. Date: 07/25/12
Data File: 0725T15.D

		Compound	MEAN	CCRF	%D	%Drift	
1	I	Fluorobenzene (IS)	ISTD			I	
2	TMHB	Gasoline	4.903	2.065	58	TMHBL	3.3
3	I	Chlorobenzene-D5 (IS)	ISTD			I	
4	I	1,4-Dichlorobenzene-D (IS)	ISTD			I	
5							
6							
7							
8							
9							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							
23							
24							
25							
26							
27							
28							
29							
30							
31							
32							
33							
34							
35							
36							
37							
38							
39							
40		Average			58.0		

MCS 7/31/12

Data File : M:\THOR\DATA\T120725\0725T15.D Vial: 14
 Acq On : 25 Jul 12 15:55 Operator: DG,RS,HW,ARS,SV
 Sample : LCS gas 300ug/L (SS) Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 26 8:23 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	788179	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.88	TIC	879850	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1024196	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	19535277m	290.16403	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T15.D
Acq On : 25 Jul 12 15:55
Sample : LCS gas 300ug/L (SS)
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 14
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 26 8:23 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T15.D
 Acq On : 25 Jul 12 15:55
 Sample : LCS gas 300ug/L (SS)
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 8:23 2012

Vial: 14
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Multiple Level Calibration



TIC: 0725T15.D

(2) Gasoline (TMHB)

8.50min 216.4348ppb m

response 17002901



Ion	Exp%	Act%
TIC	100	100
0.00	0.00	0.58#
0.00	0.00	1.69#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T15.D
 Acq On : 25 Jul 12 15:55
 Sample : LCS gas 300ug/L (SS)
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 8:23 2012

Vial: 14
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Multiple Level Calibration



TIC: 0725T15.D

(2) Gasoline (TMHB)

8.50min 290.1640ppb m
 response 19535277

Ion	Exp%	Act%
TIC	100	100
0.00	0.00	0.51#
0.00	0.00	1.47#
0.00	0.00	0.00

VOLATILE ORGANIC ANALYSIS BY
EPA METHOD 8260B

Form 7

Continuing Calibration

Lab Name: APPL, Inc.
Case No: _____
Matrix: Water

SDG No: 68258
Date Analyzed: 07/25/12
Instrument: Thor
Initial Cal. Date: 07/25/12
Data File: 0725T14.D

	Compound	MEAN	CCRF	%D	%Drift
1	Fluorobenzene (IS)	ISTD			
2	TMHB Gasoline	4.903	1.970	60	TMHBL 12
3	Chlorobenzene-D5 (IS)	ISTD			
4	1,4-Dichlorobenzene-D (IS)	ISTD			
5					
6					
7					
8					
9					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					
23					
24					
25					
26					
27					
28					
29					
30					
31					
32					
33					
34					
35					
36					
37					
38					
39					
40	Average			60.0	

APPL 7/31/12

Data File : M:\THOR\DATA\T120725\0725T14.D Vial: 13
 Acq On : 25 Jul 12 15:27 Operator: DG,RS,HW,ARS,SV
 Sample : CCV gas 300ug/L Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 26 8:22 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	819782	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	915724	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1043658	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	19379947m	263.88990	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T14.D
Acq On : 25 Jul 12 15:27
Sample : CCV gas 300ug/L
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 13
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 26 8:22 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T14.D
 Acq On : 25 Jul 12 15:27
 Sample : CCV gas 300ug/L
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 8:20 2012

Vial: 13
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Multiple Level Calibration



TIC: 0725T14.D



(2) Gasoline (TMHB)		
8.50min	191.4806ppb m	
response	16793186	
Ion	Exp%	Act%
TIC	100	100
0.00	0.00	0.61#
0.00	0.00	1.76#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T14.D
Acq On : 25 Jul 12 15:27
Sample : CCV gas 300ug/L
Misc : 10ml w/5ul of IS&S: 06-7-12
Quant Time: Jul 26 8:22 2012

Vial: 13
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00
Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Multiple Level Calibration



TIC: 0725T14.D

(2) Gasoline (TMHB)		
8.50min	263.8899ppb m	
response	19379947	
Ion	Exp%	Act%
TIC	100	100
0.00	0.00	0.53#
0.00	0.00	1.53#
0.00	0.00	0.00

**EPA METHOD 8260B
Volatile Organic Compounds
Raw Data**

APPL, INC.

Method Blank
EPA 8260B VOCs + Gas Water

Blank Name/QCG: **120719W-65112 - 169441**
Batch ID: #86RHB-120719AT1

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	1,1,1,2-TETRACHLOROETHANE	0.26 U	1.0	0.26	0.13	ug/L	07/20/12	07/20/12
BLANK	1,1,1-TRICHLOROETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
BLANK	1,1,2,2-TETRACHLOROETHANE	0.20 U	1.0	0.20	0.10	ug/L	07/20/12	07/20/12
BLANK	1,1,2-TRICHLOROETHANE	0.40 U	1.0	0.40	0.20	ug/L	07/20/12	07/20/12
BLANK	1,1-DICHLOROETHANE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	1,1-DICHLOROETHENE	0.60 U	1.0	0.60	0.30	ug/L	07/20/12	07/20/12
BLANK	1,2,3-TRICHLOROPROPANE	0.78 U	2.0	0.78	0.39	ug/L	07/20/12	07/20/12
BLANK	1,2,4-TRICHLOROBENZENE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
BLANK	1,2-DIBROMO-3-CHLOROPROPA	1.52 U	2.0	1.52	0.76	ug/L	07/20/12	07/20/12
BLANK	1,2-DIBROMOETHANE	0.40 U	1.0	0.40	0.20	ug/L	07/20/12	07/20/12
BLANK	1,2-DICHLOROBENZENE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
BLANK	1,2-DICHLOROETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
BLANK	1,2-DICHLOROPROPANE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
BLANK	1,3-DICHLOROBENZENE	0.22 U	1.0	0.22	0.11	ug/L	07/20/12	07/20/12
BLANK	1,3-DICHLOROPROPENE, TOTAL	0.36 U	1.0	0.36	0.18	ug/L	07/20/12	07/20/12
BLANK	1,4-DICHLOROBENZENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	2-BUTANONE	1.20 U	10.0	1.20	0.60	ug/L	07/20/12	07/20/12
BLANK	4-METHYL-2-PENTANONE	3.80 U	10.0	3.80	1.90	ug/L	07/20/12	07/20/12
BLANK	ACETONE	1.90 U	10.0	1.90	0.95	ug/L	07/20/12	07/20/12
BLANK	BENZENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
BLANK	BROMODICHLOROMETHANE	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
BLANK	BROMOFORM	0.28 U	1.0	0.28	0.14	ug/L	07/20/12	07/20/12
BLANK	BROMOMETHANE	0.48 U	2.0	0.48	0.24	ug/L	07/20/12	07/20/12
BLANK	CARBON TETRACHLORIDE	0.20 U	1.0	0.20	0.10	ug/L	07/20/12	07/20/12
BLANK	CHLOROBENZENE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
BLANK	CHLORODIBROMOMETHANE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	CHLOROETHANE	0.42 U	1.0	0.42	0.21	ug/L	07/20/12	07/20/12
BLANK	CHLOROFORM	0.14 U	1.0	0.14	0.07	ug/L	07/20/12	07/20/12
BLANK	CHLOROMETHANE	0.84 U	1.0	0.84	0.42	ug/L	07/20/12	07/20/12
BLANK	CIS-1,2-DICHLOROETHENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
BLANK	ETHYLBENZENE	0.46 U	1.0	0.46	0.23	ug/L	07/20/12	07/20/12
BLANK	GASOLINE	12.12 U	20.0	12.12	6.06	ug/L	07/20/12	07/20/12
BLANK	HEXACHLOROBUTADIENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	METHYL TERT-BUTYL ETHER	0.52 U	1.0	0.52	0.26	ug/L	07/20/12	07/20/12

Quant Method: TALLW.M
Run #: 0719T38
Instrument: Thor
Sequence: T120719
Initials: ARS

Printed: 07/31/12 9:50:34 AM
GC SC-Blank-REG MDLs

Method Blank
EPA 8260B VOCs + Gas Water

Blank Name/QCG: **120719W-65112 - 169441**
 Batch ID: #86RHB-120719AT1

APPL Inc.
 908 North Temperance Avenue
 Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	METHYLENE CHLORIDE	0.70 U	5.0	0.70	0.35	ug/L	07/20/12	07/20/12
BLANK	STYRENE	0.50 U	1.0	0.50	0.25	ug/L	07/20/12	07/20/12
BLANK	TETRACHLOROETHENE	0.48 U	1.0	0.48	0.24	ug/L	07/20/12	07/20/12
BLANK	TOLUENE	0.34 U	1.0	0.34	0.17	ug/L	07/20/12	07/20/12
BLANK	TRANS-1,2-DICHLOROETHENE	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	TRICHLOROETHENE	0.32 U	1.0	0.32	0.16	ug/L	07/20/12	07/20/12
BLANK	VINYL CHLORIDE	0.46 U	1.0	0.46	0.23	ug/L	07/20/12	07/20/12
BLANK	XYLENES (TOTAL)	0.38 U	1.0	0.38	0.19	ug/L	07/20/12	07/20/12
BLANK	SURROGATE: 1,2-DICHLOROET	102	70-120			%	07/20/12	07/20/12
BLANK	SURROGATE: 4-BROMOFLUORO	101	75-120			%	07/20/12	07/20/12
BLANK	SURROGATE: DIBROMOFLUOR	100	85-115			%	07/20/12	07/20/12
BLANK	SURROGATE: TOLUENE-D8 (S)	99.7	85-120			%	07/20/12	07/20/12

Quant Method: TALLW.M
 Run #: 0719T38
 Instrument: Thor
 Sequence: T120719
 Initials: ARS

Printed: 07/31/12 9:50:34 AM
 GC SC-Blank-REG MDLs

Data File : M:\THOR\DATA\T120719\0719T38.D
Acq On : 20 Jul 12 2:18
Sample : 120719A BLK-1WT
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 38
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 23 10:44 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 10:40:23 2012
Response via : Initial Calibration
DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	96	441792	25.00000	ppb	0.00
55) Chlorobenzene-D5 (IS)	9.87	117	355584	25.00000	ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	206976	25.00000	ppb	0.00

System Monitoring Compounds

31) Dibromofluoromethane(S)	5.95	111	220986	31.96459	ppb	0.00
Spiked Amount	31.881					
				Recovery	=	100.263%
36) 1,2-DCA-D4(S)	6.33	65	221504	34.47528	ppb	0.00
Spiked Amount	33.647					
				Recovery	=	102.462%
56) Toluene-D8(S)	8.43	98	782720	37.23377	ppb	0.00
Spiked Amount	37.345					
				Recovery	=	99.703%
64) 4-Bromofluorobenzene(S)	11.05	95	294956	29.66906	ppb	0.00
Spiked Amount	29.515					
				Recovery	=	100.521%

Target Compounds

Qvalue


Data File : M:\THOR\DATA\T120719\0719T38.D
Acq On : 20 Jul 12 2:18
Sample : 120719A BLK-1WT
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 38
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 23 10:44 2012

Quant Results File: TALLW.RES

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Fri Jul 20 10:40:23 2012
Response via : Initial Calibration



Data File : M:\THOR\DATA\T120725\0725T20.D Vial: 19
 Acq On : 25 Jul 12 18:14 Operator: DG,RS,HW,ARS,SV
 Sample : 120725A BLK-1WT Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 26 8:07 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	787932	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	886149	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	982900	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.43	TIC	10210921m	18.78182	ppb	ND 100

*No gasoline pattern detected.
ARS 7/26/12*

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T20.D
Acq On : 25 Jul 12 18:14
Sample : 120725A BLK-1WT
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 19
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 26 8:07 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T20.D
 Acq On : 25 Jul 12 18:14
 Sample : 120725A BLK-1WT
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 7:57 2012

Vial: 19
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Multiple Level Calibration



TIC: 0725T20.D

(2) Gasoline (TMHB)

8.50min -58.7487ppb m

response 7548819



Ion	Exp%	Act%
TIC	100	100
0.00	0.00	1.25#
0.00	0.00	3.68#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T20.D
 Acq On : 25 Jul 12 18:14
 Sample : 120725A BLK-1WT
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 8:07 2012

Vial: 19
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Multiple Level Calibration



TIC: 0725T20.D

(2) Gasoline (TMHB)

8.43min 18.7818ppb m

response 10210921

Ion	Exp%	Act%
TIC	100	100
0.00	0.00	0.92#
0.00	0.00	2.72#
0.00	0.00	0.00

Method Blank
EPA 8260B VOCS + GAS WATER

Blank Name/QCG: **120719W-65113 - 169517**
Batch ID: #86RHB-120719AT2

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Sample Type	Analyte	Result	LOQ	LOD	DL	Units	Extraction Date	Analysis Date
BLANK	GASOLINE	12.12 U	20.0	12.12	6.06	ug/L	07/20/12	07/20/12
BLANK	SURROGATE: 1,2-DICHLOROET	102	70-120			%	07/20/12	07/20/12
BLANK	SURROGATE: 4-BROMOFLUORO	101	75-120			%	07/20/12	07/20/12
BLANK	SURROGATE: DIBROMOFLUOR	100	85-115			%	07/20/12	07/20/12
BLANK	SURROGATE: TOLUENE-D8 (S)	99.7	85-120			%	07/20/12	07/20/12

Quant Method: TALLW.M
Run #: 0719T38
Instrument: Thor
Sequence: T120719
Initials: ARS

Printed: 07/31/12 1:07:19 PM
GC SC-Blank-REG MDLs

Data File : M:\THOR\DATA\T120719\0719T38.D Vial: 38
 Acq On : 20 Jul 12 2:18 Operator: DG,RS,HW,ARS,SV
 Sample : 120719A BLK-1WT Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 26 15:38 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	913286	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	1028057	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1144195	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.43	TIC	11231950m	126.77689	ppb	ND 100

*No gasoline pattern detected.
 ARS 7/26/12*

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T38.D
Acq On : 20 Jul 12 2:18
Sample : 120719A BLK-1WT
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 38
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 26 15:38 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T36.D
 Acq On : 20 Jul 12 1:22
 Sample : 120719A BLK-1WT
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 15:37 2012

Vial: 36
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Single Level Calibration



TIC: 0719T36.D

(2) Gasoline (TMHB)

8.50min 63.1186ppb m

response 8713545



Ion	Exp%	Act%
TIC	100	100
0.00	0.60	1.30#
0.00	1.80	3.61#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T38.D
 Acq On : 20 Jul 12 2:18
 Sample : 120719A BLK-1WT
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 15:38 2012

Vial: 38
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Single Level Calibration



TIC: 0719T38.D

(2) Gasoline (TMHB)

8.43min 126.7769ppb m

response 11231950

Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.98#
0.00	1.80	2.87#
0.00	0.00	0.00

Laboratory Control Spike Recovery

EPA 8260B VOCs + Gas Water

APPL ID: 120719W-65112 LCS - 169441
 Batch ID: #86RHB-120719AT1

APPL Inc.
 908 North Temperance Avenue
 Clovis, CA 93611

Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits
1,1,1,2-TETRACHLOROETHANE	10.00	9.79	97.9	80-130
1,1,1-TRICHLOROETHANE	10.00	9.62	96.2	65-130
1,1,2,2-TETRACHLOROETHANE	10.00	9.22	92.2	65-130
1,1,2-TRICHLOROETHANE	10.00	9.61	96.1	75-125
1,1-DICHLOROETHANE	10.00	10.1	101	70-135
1,1-DICHLOROETHENE	10.00	9.64	96.4	70-130
1,2,3-TRICHLOROPROPANE	10.00	10.3	103	75-125
1,2,4-TRICHLOROBENZENE	10.00	9.98	99.8	65-135
1,2-DIBROMO-3-CHLOROPROPANE	10.00	10.1	101	50-130
1,2-DIBROMOETHANE	10.00	9.79	97.9	70-130
1,2-DICHLOROBENZENE	10.00	9.83	98.3	70-120
1,2-DICHLOROETHANE	10.00	9.76	97.6	70-130
1,2-DICHLOROPROPANE	10.00	10.1	101	75-125
1,3-DICHLOROBENZENE	10.00	10.2	102	75-125
1,3-DICHLOROPROPENE, TOTAL	20.0	18.7	93.5	70-130
1,4-DICHLOROBENZENE	10.00	9.82	98.2	75-125
2-BUTANONE	10.00	10.3	103	30-150
4-METHYL-2-PENTANONE	10.00	10.2	102	60-135
ACETONE	10.00	11.8	118	40-140
BENZENE	10.00	9.47	94.7	80-120
BROMODICHLOROMETHANE	10.00	9.57	95.7	75-120
BROMOFORM	10.00	9.70	97.0	70-130
BROMOMETHANE	10.00	9.36	93.6	30-145
CARBON TETRACHLORIDE	10.00	10.0	100	65-140
CHLOROBENZENE	10.00	9.82	98.2	80-120
CHLORODIBROMOMETHANE	10.00	9.73	97.3	60-135
CHLOROETHANE	10.00	9.84	98.4	60-135
CHLOROFORM	10.00	9.60	96.0	65-135
CHLOROMETHANE	10.00	9.80	98.0	40-125
CIS-1,2-DICHLOROETHENE	10.00	9.99	99.9	70-125
ETHYLBENZENE	10.00	10.1	101	75-125

Comments: _____

<u>Primary</u>	<u>SPK</u>
Quant Method :	TALLW.M
Extraction Date :	07/19/12
Analysis Date :	07/19/12
Instrument :	Thor
Run :	0719T31
Initials :	ARS

Printed: 07/31/12 9:50:26 AM
 APPL Standard LCS

Laboratory Control Spike Recovery

EPA 8260B VOCs + Gas Water

APPL ID: 120719W-65112 LCS - 169441

Batch ID: #86RHB-120719AT1

APPL Inc.

908 North Temperance Avenue

Clovis, CA 93611

Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits
GASOLINE	300	290	96.7	75-125
HEXACHLOROBUTADIENE	10.00	9.23	92.3	50-140
METHYL TERT-BUTYL ETHER	10.00	9.48	94.8	65-125
METHYLENE CHLORIDE	10.00	9.45	94.5	55-140
STYRENE	10.00	10.4	104	65-135
TETRACHLOROETHENE	10.00	10.1	101	45-150
TOLUENE	10.00	10.1	101	75-120
TRANS-1,2-DICHLOROETHENE	10.00	9.79	97.9	60-140
TRICHLOROETHENE	10.00	10.6	106	70-125
VINYL CHLORIDE	10.00	10.1	101	50-145
XYLENES (TOTAL)	30.0	31.3	104	80-120
SURROGATE: 1,2-DICHLOROETHANE-D	33.6	32.9	97.8	70-120
SURROGATE: 4-BROMOFLUOROBENZE	29.5	30.2	102	75-120
SURROGATE: DIBROMOFLUOROMETH	31.9	31.3	98.2	85-115
SURROGATE: TOLUENE-D8 (S)	37.3	36.6	98.0	85-120

Comments: _____

<u>Primary</u>	<u>SPK</u>
Quant Method :	TALLW.M
Extraction Date :	07/19/12
Analysis Date :	07/19/12
Instrument :	Thor
Run :	0719T31
Initials :	ARS

Printed: 07/31/12 9:50:26 AM

APPL Standard LCS

Data File : M:\THOR\DATA\T120719\0719T31.D
 Acq On : 19 Jul 12 23:03
 Sample : 120719A LCS-1WT (SS)
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 31
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 10:45 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 10:40:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	96	459584	25.00000 ppb	0.00
55) Chlorobenzene-D5 (IS)	9.87	117	371008	25.00000 ppb	0.00
70) 1,4-Dichlorobenzene-D (IS)	12.20	152	216768	25.00000 ppb	0.00
System Monitoring Compounds					
31) Dibromofluoromethane(S)	5.95	111	225058	31.29333 ppb	0.00
Spiked Amount	31.881		Recovery	= 98.155%	
36) 1,2-DCA-D4(S)	6.33	65	220138	32.93626 ppb	0.00
Spiked Amount	33.647		Recovery	= 97.888%	
56) Toluene-D8(S)	8.43	98	802051	36.56718 ppb	0.00
Spiked Amount	37.345		Recovery	= 97.917%	
64) 4-Bromofluorobenzene(S)	11.05	95	313456	30.21914 ppb	0.00
Spiked Amount	29.515		Recovery	= 102.384%	
Target Compounds					
2) Dichlorodifluoromethane	1.30	85	18648	8.01049 ppb	98
3) Freon 114	1.41	85	29065	8.97783 ppb	92
4) Chloromethane	1.45	50	56808	9.80339 ppb	99
5) Vinyl chloride	1.56	62	91788	10.10524 ppb	99
6) Bromomethane	1.87	94	54346	9.36087 ppb	98
7) Chloroethane	1.97	64	51463	9.83706 ppb	96
8) Dichlorofluoromethane	2.18	67	3106	9.09488 ppb	97
9) Trichlorofluoromethane	2.24	101	19028	10.13498 ppb	100
11) Acetone	2.88	43	19460	11.84185 ppb	98
12) Freon-113	2.85	101	37646	9.96889 ppb	94
13) 1,1-DCE	2.82	61	48838	9.63706 ppb	93
14) t-Butanol	3.69	59	19056	127.86417 ppb	98
15) Methyl Acetate	3.34	43	44993	10.18034 ppb	95
16) Iodomethane	2.98	142	43340	9.45518 ppb	97
17) Acrylonitrile	3.81	52	14853	10.23301 ppb	95
18) Methylene chloride	3.45	84	17424	9.44871 ppb	95
19) Carbon disulfide	3.06	76	5510	10.69990 ppb	# 86
20) Methyl t-butyl ether (MtBE)	3.90	73	92761	9.48061 ppb	98
21) Trans-1,2-DCE	3.87	96	34225	9.78590 ppb	97
22) Diisopropyl Ether	4.70	59	21995	10.03782 ppb	95
23) 1,1-DCA	4.51	63	93412	10.07257 ppb	98
24) Vinyl Acetate	4.70	87	50781	9.69469 ppb	95
25) Ethyl tert Butyl Ether	5.21	59	119561	9.77392 ppb	99
26) MEK (2-Butanone)	5.38	43	23166	10.28682 ppb	95
27) Cis-1,2-DCE	5.32	96	59336	9.98787 ppb	96
28) 2,2-Dichloropropane	5.32	77	29940	8.01402 ppb	99
29) Chloroform	5.75	83	110557	9.59991 ppb	94
30) Bromochloromethane	5.62	128	29433	10.17554 ppb	98
32) 1,1,1-TCA	5.96	97	66682	9.62307 ppb	96
33) Cyclohexane	6.03	41	18804	9.99923 ppb	94
34) 1,1-Dichloropropene	6.17	75	50257	9.98686 ppb	98
35) 2,2,4-Trimethylpentane	6.55	57	62413	8.62945 ppb	97
37) Carbon Tetrachloride	6.16	117	65247	10.04641 ppb	95
38) Tert Amyl Methyl Ether	6.59	73	128152	9.84264 ppb	100
39) 1,2-DCA	6.41	62	73737	9.76354 ppb	99
40) Benzene	6.40	78	195282	9.46720 ppb	97
41) TCE	7.14	95	59649	10.63894 ppb	98
42) 2-Pentanone	7.36	43	545318	123.45728 ppb	100
43) 1,2-Dichloropropane	7.37	63	67896	10.08801 ppb	96

Algorithm Check: (91788)(25) (459584)(0.4941) Qvalue = 10.10522903 ✓
 ARS 7/27/12

Data File : M:\THOR\DATA\T120719\0719T31.D
 Acq On : 19 Jul 12 23:03
 Sample : 120719A LCS-1WT (SS)
 Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 31
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Quant Time: Jul 20 10:45 2012

Quant Results File: TALLW.RES

Quant Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Fri Jul 20 10:40:23 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Compound	R.T.	QIon	Response	Conc	Unit	Qvalue
44) Bromodichloromethane	7.68	83	89098	9.56942	ppb	98
45) Methyl Cyclohexane	7.36	83	38326	9.57410	ppb	95
46) Dibromomethane	7.49	93	36747	10.03731	ppb	99
47) 2-Chloroethyl vinyl ether	7.99	106	1154	8.86120	ppb	# 100
48) MIBK (methyl isobutyl ket	8.33	43	32328	10.17680	ppb	97
49) 1-Bromo-2-chloroethane	7.99	63	45864	9.79426	ppb	99
50) <u>Cis-1,3-Dichloropropene</u>	8.15	75	86842	9.42535	ppb	98
51) Toluene	8.50	91	246468	10.12718	ppb	98
52) <u>Trans-1,3-Dichloropropene</u>	8.72	75	75443	9.28657	ppb	98
53) 1,1,2-TCA	8.90	83	52073	9.60983	ppb	98
54) 2-Hexanone	9.18	43	36772	10.09450	ppb	96
57) 1,2-EDB	9.40	107	54442	9.78712	ppb	99
58) Tetrachloroethene	9.06	166	63354	10.07267	ppb	96
59) 1-Chlorohexane	9.90	91	74600	9.96437	ppb	98
60) 1,1,1,2-Tetrachloroethane	9.99	131	71965	9.79349	ppb	96
61) m&p-Xylene	10.14	106	239826	20.92173	ppb	98
62) o-Xylene	10.54	106	123104	10.38148	ppb	98
63) Styrene	10.55	104	208582	10.35266	ppb	98
65) 1,3-Dichloropropane	9.07	76	98494	10.09885	ppb	97
66) Dibromochloromethane	9.29	129	71411	9.72527	ppb	99
67) Chlorobenzene	9.90	112	188318	9.81804	ppb	99
68) Ethylbenzene	10.03	91	305101	10.11617	ppb	99
69) Bromoform	10.71	173	48764	9.69905	ppb	99
71) Isopropylbenzene	10.91	105	295625	10.43029	ppb	98
72) 1,1,2,2-Tetrachloroethane	11.19	83	72548	9.22452	ppb	99
73) 1,2,3-Trichloropropane	11.23	110	23096	10.34681	ppb	85
74) t-1,4-Dichloro-2-Butene	11.25	53	15810	10.58327	ppb	89
75) Bromobenzene	11.19	156	93573	10.01488	ppb	99
76) n-Propylbenzene	11.32	91	385440	10.56221	ppb	98
77) 4-Ethyltoluene	11.43	105	325068	10.37283	ppb	98
78) 2-Chlorotoluene	11.39	91	267062	10.26282	ppb	100
79) 1,3,5-Trimethylbenzene	11.50	105	276230	10.63314	ppb	99
80) 4-Chlorotoluene	11.50	91	267095	10.36956	ppb	100
81) Tert-Butylbenzene	11.82	119	244365	10.26802	ppb	99
82) 1,2,4-Trimethylbenzene	11.86	105	278601	10.36400	ppb	99
83) Sec-Butylbenzene	12.04	105	333801	10.50584	ppb	100
84) p-Isopropyltoluene	12.19	119	281061	10.46929	ppb	99
85) Benzyl Chloride	12.35	91	53118	6.62120	ppb	97
86) 1,3-DCB	12.13	146	180466	10.21317	ppb	99
87) 1,4-DCB	12.22	146	181734	9.82055	ppb	99
88) n-Butylbenzene	12.59	91	245949	10.22102	ppb	98
89) 1,2-DCB	12.59	146	168341	9.82946	ppb	97
90) Hexachloroethane	12.86	117	47831	9.72421	ppb	93
91) 1,2-Dibromo-3-chloropropan	13.35	157	14935	10.13631	ppb	95
92) 1,2,4-Trichlorobenzene	14.20	180	78384	9.98480	ppb	100
93) Hexachlorobutadiene	14.38	223	30261	9.22762	ppb	89
94) Naphthalene	14.43	128	232681	10.61496	ppb	98
95) 1,2,3-Trichlorobenzene	14.68	180	114268	10.21606	ppb	99

*1,3-dichloropropene, total!
18.71192 ppb*

ARS 7/27/12

Data File : M:\THOR\DATA\T120725\0725T15.D Vial: 14
 Acq On : 25 Jul 12 15:55 Operator: DG,RS,HW,ARS,SV
 Sample : LCS gas 300ug/L (SS) Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 26 8:23 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	788179	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.88	TIC	879850	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1024196	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.50	TIC	19535277m	290.16403	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120725\0725T15.D
Acq On : 25 Jul 12 15:55
Sample : LCS gas 300ug/L (SS)
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 14
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 26 8:23 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Wed Jul 25 16:07:29 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T15.D
 Acq On : 25 Jul 12 15:55
 Sample : LCS gas 300ug/L (SS)
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 8:23 2012

Vial: 14
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Multiple Level Calibration



TIC: 0725T15.D

(2) Gasoline (TMHB)

8.50min 216.4348ppb m

response 17002901



Ion	Exp%	Act%
TIC	100	100
0.00	0.00	0.58#
0.00	0.00	1.69#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120725\0725T15.D
 Acq On : 25 Jul 12 15:55
 Sample : LCS gas 300ug/L (SS)
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 26 8:23 2012

Vial: 14
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Wed Jul 25 16:07:29 2012
 Response via : Multiple Level Calibration



TIC: 0725T15.D

(2) Gasoline (TMHB)

8.50min 290.1640ppb m

response 19535277

Ion	Exp%	Act%
TIC	100	100
0.00	0.00	0.51#
0.00	0.00	1.47#
0.00	0.00	0.00

Laboratory Control Spike Recovery

EPA 8260B VOCS + GAS WATER

APPL ID: 120719W-65113 LCS - 169517
 Batch ID: #86RHB-120719AT2

APPL Inc.
 908 North Temperance Avenue
 Clovis, CA 93611

Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits
GASOLINE	300	389	130 #	75-125
SURROGATE: 1,2-DICHLOROETHANE-D	33.6	32.9	97.8	70-120
SURROGATE: 4-BROMOFLUOROBENZE	29.5	30.2	102	75-120
SURROGATE: DIBROMOFLUOROMETH	31.9	31.3	98.2	85-115
SURROGATE: TOLUENE-D8 (S)	37.3	36.6	98.0	85-120

= Recovery is outside QC limits.

Comments:

<u>Primary</u>	<u>SPK</u>
Quant Method :	TALLW.M
Extraction Date :	07/19/12
Analysis Date :	07/19/12
Instrument :	Thor
Run :	0719T31
Initials :	ARS

Printed: 07/31/12 1:07:00 PM
 APPL Standard LCS

Data File : M:\THOR\DATA\T120719\0719T35.D Vial: 35
 Acq On : 20 Jul 12 00:54 Operator: DG,RS,HW,ARS,SV
 Sample : LCS gas 300ug/L Inst : Thor
 Misc : 10ml w/5ul of IS&S: 06-7-12 Multiplr: 1.00

Quant Time: Jul 24 13:37 2012 Quant Results File: TGAS.RES

Quant Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Initial Calibration
 DataAcq Meth : 8260_BETA

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene (IS)	6.73	TIC	943495	25.00000	ppb	0.00
3) Chlorobenzene-D5 (IS)	9.87	TIC	1050866	25.00000	ppb	0.00
4) 1,4-Dichlorobenzene-D (IS)	12.20	TIC	1213947	25.00000	ppb	0.00

System Monitoring Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Gasoline	8.43	TIC	22281600m	389.43424	ppb	100

Quantitation Report


Data File : M:\THOR\DATA\T120719\0719T35.D
Acq On : 20 Jul 12 00:54
Sample : LCS gas 300ug/L
Misc : 10ml w/5ul of IS&S: 06-7-12

Vial: 35
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Quant Time: Jul 24 13:37 2012

Quant Results File: TGAS.RES

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B
Last Update : Tue Jul 24 13:31:25 2012
Response via : Initial Calibration





Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T35.D
 Acq On : 20 Jul 12 00:54
 Sample : LCS gas 300ug/L
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:33 2012

Vial: 35
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Single Level Calibration



TIC: 0719T35.D



(2) Gasoline (TMHB)		
8.50min	313.6512ppb m	
response	19200699	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.62#
0.00	1.80	1.77#
0.00	0.00	0.00

Quantitation Report

Data File : M:\THOR\DATA\T120719\0719T35.D
 Acq On : 20 Jul 12 00:54
 Sample : LCS gas 300ug/L
 Misc : 10ml w/5ul of IS&S: 06-7-12
 Quant Time: Jul 24 13:37 2012

Vial: 35
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00
 Quant Results File: temp.res

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B
 Last Update : Tue Jul 24 13:31:25 2012
 Response via : Single Level Calibration





TIC: 0719T35.D

(2) Gasoline (TMHB)		
8.43min	389.4342ppb m	
response	22281600	
Ion	Exp%	Act%
TIC	100	100
0.00	0.60	0.53#
0.00	1.80	1.52#
0.00	0.00	0.00

Data File : M:\THOR\DATA\T120719\0719T01T.D
 Acq On : 19 Jul 12 9:15
 Sample : 5ng- BFB STD 07-16-12B
 Misc : 2ul

Vial: 1
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B





AutoFind: Scans 554, 555, 556; Background Corrected with Scan 544

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	13331	PASS
75	95	30	60	46.4	35536	PASS
95	95	100	100	100.0	76600	PASS
96	95	5	9	6.7	5096	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	96.0	73547	PASS
175	174	5	9	7.2	5311	PASS
176	174	95	101	99.3	73019	PASS
177	176	5	9	7.0	5141	PASS

Data File : M:\THOR\DATA\T120719\0719T28.D
 Acq On : 19 Jul 12 21:40
 Sample : 5ng- BFB Std 07-16-12B
 Misc : 2uL

Vial: 28
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Method : M:\THOR\DATA\T120719\TALLW.M (RTE Integrator)
 Title : METHOD 8260B




AutoFind: Scans 3065, 3066, 3067; Background Corrected with Scan 3051

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.9	31552	PASS
75	95	30	60	47.3	88245	PASS
95	95	100	100	100.0	186709	PASS
96	95	5	9	6.8	12716	PASS
173	174	0.00	2	1.0	1785	PASS
174	95	50	100	95.8	178816	PASS
175	174	5	9	7.5	13428	PASS
176	174	95	101	96.9	173248	PASS
177	176	5	9	6.2	10814	PASS

Data File : M:\THOR\DATA\T120719\0719T15.D
 Acq On : 19 Jul 12 15:39
 Sample : 5ng-BFB STD 07-16-12B
 Misc : 2uL

Vial: 15
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
 Title : METHOD 8260B



Spectrum Information: Average of 11.045 to 11.051 min.


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	32016	PASS
75	95	30	60	47.7	88957	PASS
95	95	100	100	100.0	186347	PASS
96	95	5	9	6.8	12727	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	96.2	179243	PASS
175	174	5	9	7.2	12837	PASS
176	174	95	101	97.5	174763	PASS
177	176	5	9	6.6	11511	PASS

BFB

Data File : M:\THOR\DATA\T120719\0719T28.D
Acq On : 19 Jul 12 21:40
Sample : 5ng- BFB Std 07-16-12B
Misc : 2uL

Vial: 28
Operator: DG,RS,HW,ARS,SV
Inst : Thor
Multiplr: 1.00

Method : M:\THOR\DATA\T120719\TGAS.M (RTE Integrator)
Title : METHOD 8260B




Spectrum Information: Average of 11.045 to 11.052 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.9	31552	PASS
75	95	30	60	47.3	88245	PASS
95	95	100	100	100.0	186709	PASS
96	95	5	9	6.8	12716	PASS
173	174	0.00	2	1.0	1785	PASS
174	95	50	100	95.8	178816	PASS
175	174	5	9	7.5	13428	PASS
176	174	95	101	96.9	173248	PASS
177	176	5	9	6.2	10814	PASS

Data File : M:\THOR\DATA\T120725\0725T01T.D
 Acq On : 25 Jul 12 9:32
 Sample : 5ng- BFB STD 07-16-12B
 Misc : 2ul

Vial: 1
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B





Spectrum Information: Average of 1.760 to 1.767 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.1	26589	PASS
75	95	30	60	49.1	72077	PASS
95	95	100	100	100.0	146837	PASS
96	95	5	9	7.2	10518	PASS
173	174	0.00	2	0.4	583	PASS
174	95	50	100	95.3	139968	PASS
175	174	5	9	8.0	11175	PASS
176	174	95	101	99.8	139627	PASS
177	176	5	9	6.3	8859	PASS

Data File : M:\THOR\DATA\T120725\0725T13.D
 Acq On : 25 Jul 12 14:59
 Sample : 5ng- BFB STD 07-16-12B
 Misc : 2uL

Vial: 12
 Operator: DG,RS,HW,ARS,SV
 Inst : Thor
 Multiplr: 1.00

Method : M:\THOR\DATA\T120725\TGAS.M (RTE Integrator)
 Title : METHOD 8260B



Spectrum Information: Average of 11.045 to 11.051 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.9	27811	PASS
75	95	30	60	46.8	76795	PASS
95	95	100	100	100.0	164224	PASS
96	95	5	9	6.7	11042	PASS
173	174	0.00	2	0.3	404	PASS
174	95	50	100	94.5	155157	PASS
175	174	5	9	7.4	11517	PASS
176	174	95	101	97.6	151509	PASS
177	176	5	9	6.1	9224	PASS

048

GC/MS STANDARD PREPARATION BOOK # _____ PAGE # _____

6/10/12 RS

Volatile Standard Curve Preparation for 5mL Purge (8260 sol)-SWEETPEA

Expiration Date: 06/09/12		50µg/mL Vol Std #9		50µg/mL Vol Std #7		50µg/mL Vol Std #8		50µg/mL Vol Std #10		50µg/mL Vol Std #12	
Date	Conc.	06-02-12Z	06-02-12AD	06-02-12V	06-02-12X	06-02-12AC	06-02-12AA	06-02-12W	06-02-12Y	06-02-12AB	06-02-12B
Code	µg/L	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12
06-08-12A	2	2	2	n/a	n/a	n/a	2	n/a	2	n/a	n/a
06-08-12B	5	5	5	n/a	n/a	n/a	5	n/a	5	n/a	n/a
06-08-12C	10	10	10	n/a	n/a	n/a	10	n/a	10	n/a	n/a
06-08-12D	20	20	20	n/a	n/a	n/a	20	n/a	20	n/a	n/a
06-08-12E	50	n/a	n/a	5	5	5	n/a	5	n/a	5	n/a
06-08-12F	100	n/a	n/a	10	10	10	n/a	10	n/a	10	n/a
06-08-12G	200	n/a	n/a	20	20	20	n/a	20	n/a	20	n/a

6/11/12 RS

250µg/mL TBA	Final Vol
06-02-12AE	w/P&T H2O
Exp:06-09-12	mL
1	5
2	5
3	5
4	5
5	5
6	5
7	5

Date	Conc.	Code	Date	EXP:
06-11-12A	25µg/ml BFB STD	02SI	020135-03	4-Bromofluorobenzene
EXP:07-11-12	ug/ml	Lot#	CODE	Date
06-11-12A	25µg/ml BFB STD	02SI	020135-03	4-Bromofluorobenzene
EXP:07-11-12	ug/ml	Lot#	CODE	Date
J&T Baker	Purge & Trap MeOH	K14E06-00626	06/11/12	09/28/12
06-11-12B	25µg/ml BFB STD	02SI	020135-03	4-Bromofluorobenzene
EXP:07-11-12	ug/ml	Lot#	CODE	Date
J&T Baker	Purge & Trap MeOH	K14E06-00626	06/11/12	09/28/12
06-11-12C	25µg/ml BFB STD	02SI	020135-03	4-Bromofluorobenzene
EXP:07-11-12	ug/ml	Lot#	CODE	Date
J&T Baker	Purge & Trap MeOH	K14E06-00626	06/11/12	09/28/12

6/11/12 RS

6/11/12 RS

Date	Conc.
Code	µg/L
06-11-12J	0.3
06-11-12K	0.5
06-11-12L	1
06-11-12M	2
06-11-12N	5
06-11-12O	10
06-11-12P	20
06-11-12Q	40
06-11-12R	100

6/11/12 RS

D-

Method 8260 Internal Standard Solution, 2,000 mg/L, 1 ml
 Lot # 120302-03
 Storage 5-10 Degrees C
 Expiry 11/18/12
 Solv: P/T Methanol
 solutions

Method 8260 Internal Standard
 Lot #: 166255 - 29275
 Rec: 8/5/11 MFR exp. 11/18/12

RS

6/11/12 RS

E-

Fluorobenzene Solution, 2,000 mg/L, 1 ml
 Lot # 020132-02
 Storage 5-10 Degrees C
 Expiry 2/13/14
 Solv: P/T Methanol
 Fluorobenzene
 Lot #: 169170 - 28869
 Rec: 5/25/11 MFR exp. 02/13/14

RS

Date	Conc.
Code	µg/L
06-11-12R	2
06-11-12S	5
06-11-12T	10
06-11-12U	20
06-11-12V	50
06-11-12W	100
06-11-12X	200

#2	50µg/mL Vol Std #12
	06-02-12AB
	Exp:06-09-12
	n/a
	n/a
	n/a
	n/a
	5
	10
	20

6/11/12
RS
F-

Method 8260B Surrogate
Solution, 2,000 mg/L, 1 ml

120002-01
Lot# Storage Expiry
185763 < 10 Degrees C 2/19/15

Sub: P/T Methanol

Method 8260B Surrogate
Lot #: 185763 - 30467
Rec: 2/20/12 MFR exp. 02/19/15

RS

mL TBA	Final Vol
-12AE	w/P&T H2O
09-12	mL
1	5
2	5
3	5
4	5
5	5
6	5
7	5

6/11/12
RS

Thor						
50ug/ml 8260 Internal Standard						
Supplier	ID #	Conc.	Lot #	Date	Exp.	
ug/ml				Code	Date	uL
O2SI	120302-03	2000	166255-29275	06-11-12D	12/13/12	375
O2SI	020132-02	2000	169170-28869	06-11-12E	12/13/12	375
J.T Baker			Purge & Trap MeOH	K14E06-00626	06/11/12	14250

50ug/ml 8260B Surrogate-Thor						
Supplier	ID #	Conc.	Lot #	Date	Exp.	
ug/ml				Code	Date	uL
O2SI	8260B Surr	2000	178653-30467	06-11-12F	12/13/12	375
J.T Baker			Purge & Trap MeOH	K14E06-00626	06/11/12	14625

Volatile Standard Curve Preparation for 10mL Purge (8260 water)-THOR

Expiration Date: 06/12/12										
Date	Conc.	5µg/mL Vol Std #9	5µg/mL Surr	50µg/mL Vol Std #7	50µg/mL Vol Std #8	50µg/mL Surr	5µg/mL Vol Std #10	50µg/mL Vol Std #1	50µg/mL Vol Std #2	5µg/mL Vol Std #12
Code	µg/L	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12
06-11-12I	0.3	3	6	n/a	n/a	n/a	3	n/a	n/a	3
06-11-12J	0.5	5	10	n/a	n/a	n/a	5	n/a	n/a	5
06-11-12K	1	10	20	n/a	n/a	n/a	10	n/a	n/a	10
06-11-12L	2	20	40	n/a	n/a	n/a	20	n/a	n/a	20
06-11-12M	5	n/a	n/a	5	5	10	n/a	5	5	n/a
06-11-12N	10	n/a	n/a	10	10	25	n/a	10	10	n/a
06-11-12O	20	n/a	n/a	20	20	40	n/a	20	20	n/a
06-11-12P	40	n/a	n/a	40	40	80	n/a	40	40	n/a
06-11-12Q	100	n/a	n/a	100	100	100	n/a	100	100	n/a

6/11/12
RS

250µg/mL TAPD	Final Vol
06-02-12AE	w/P&T H2O
Exp:06-09-12	mL
3	50
5	50
10	50
15	50
20	50
25	50
30	50
35	50
40	50

Volatile Standard Curve Preparation for 5mL Purge (8260 soil)-SWEETPEA

Expiration Date: 06/12/12										
Date	Conc.	5µg/mL Vol Std #9	5µg/mL Surr	50µg/mL Vol Std #7	50µg/mL Vol Std #8	50µg/mL Surr	5µg/mL Vol Std #10	50µg/mL Vol Std #1	50µg/mL Vol Std #2	5µg/mL Vol Std #12
Code	µg/L	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12	Exp:06-09-12
06-11-12R	2	2	2	n/a	n/a	n/a	2	n/a	n/a	2
06-11-12S	5	5	5	n/a	n/a	n/a	5	n/a	n/a	5
06-11-12T	10	10	10	n/a	n/a	n/a	10	n/a	n/a	10
06-11-12U	20	20	20	n/a	n/a	n/a	20	n/a	n/a	20
06-11-12V	50	n/a	n/a	5	5	5	n/a	5	n/a	5
06-11-12W	100	n/a	n/a	10	10	10	n/a	10	n/a	10
06-11-12X	200	n/a	n/a	20	20	20	n/a	20	n/a	20

6/12/12
RS

250µg/mL TBA	Final Vol
06-02-12AE	w/P&T H2O
Exp:06-09-12	mL
1	5
2	5
3	5
4	5
5	5
6	5
7	5

Volatile Standard Curve Preparation for 5mL Purge (8260 soil)-SWEETPEA

Exp. #	Expiration Date: 07/12/12									
	50µg/mL Vol Std #9	5µg/mL Sur	50µg/mL Vol Std #7	50µg/mL Vol Std #8	50µg/mL Sur	50µg/mL Vol Std #10	50µg/mL Vol Std #11	50µg/mL Vol Std #2	50µg/mL Vol Std #12	
07-05-12I	07-05-12I	07-05-12M	07-05-12E	07-05-12G	07-05-12L	07-05-12J	07-05-12F	07-05-12H	07-05-12K	
Conc.	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	
µg/L	2	2	n/a	n/a	n/a	2	n/a	2	n/a	
µg/L	5	5	n/a	n/a	n/a	5	n/a	5	n/a	
µg/L	10	10	n/a	n/a	n/a	10	n/a	10	n/a	
µg/L	20	20	n/a	n/a	n/a	20	n/a	20	n/a	
µg/L	n/a	n/a	5	5	5	n/a	5	n/a	5	
µg/L	n/a	n/a	10	10	10	n/a	10	n/a	10	

250µg/mL TBA	Final Vol w/P&T H2O
07-05-12N	mL
Exp:07-12-12	
1	5
2	5
3	5
4	5
5	5
6	5

CHICO

Date	Conc.	Date	Exp.
µg/ml	Lot #	Code	µL
02SI	122450-02	524 Fortification Sol	1000
J&T Baker		Purge & Trap MeOH	176776-29295
			06-07-12A
			10/10/12
			12/22/13
			200
			1800

Volatile Standard Curve Preparation for 10mL Purge (524 water)-CHICO

Date	Expiration Date: 07/13/12								Final Vol w/P&T H2O
	50µg/mL Vol Std #9	5µg/mL Sur	50µg/mL Vol Std #12	50µg/mL Vol Std #7	50µg/mL Vol Std #8	50µg/mL Vol Std #2	250µg/mL TAPD		
07-12-12B	0.2	2	2	n/a	n/a	n/a	2	50	
07-12-12C	0.5	5	5	n/a	n/a	n/a	5	50	
07-12-12D	1	10	10	n/a	n/a	n/a	10	50	
07-12-12E	2	20	20	n/a	n/a	n/a	15	50	
07-12-12F	5	n/a	n/a	5	5	5	20	50	
07-12-12G	10	n/a	n/a	10	10	10	25	50	
07-12-12H	20	n/a	n/a	20	20	20	30	50	
07-12-12I	40	n/a	n/a	40	40	40	35	50	
07-02-12H	100	n/a	n/a	100	100	100	40	50	

4-Bromofluorobenzene Solution, 2,500 mg/L, 1 ml

020135-03
 Lot# Storage Expiry
 163173 5-18 Degree 8/24/13
 Solv: R/T Methanol

4-Bromofluorobenzene
 Lot #: 163173 - 29063
 Rec: 8/1/11 MFR exp. 08/24/13

Date	Conc.	Date	EXP:
µg/ml	Lot#	CODE	µl
02SI	020135-03	4-Bromofluorobenzene	2500
J&T Baker		Purge & Trap MeOH	163173-29063
			07-16-12A
			12/11/12
			20
			1980
07-16-12C			
25µg/ml BFB STD			
EXP:08-16-12			
02SI	020135-03	4-Bromofluorobenzene	2500
J&T Baker		Purge & Trap MeOH	163173-29063
			07-16-12A
			12/11/12
			20
			1980
07-16-12D			
25µg/ml BFB STD			
EXP:08-16-12			
02SI	020135-03	4-Bromofluorobenzene	2500
J&T Baker		Purge & Trap MeOH	163173-29063
			07-16-12A
			12/11/12
			20
			1980
07-16-12E			
25µg/ml BFB STD			
EXP:08-16-12			
02SI	020135-03	4-Bromofluorobenzene	2500
J&T Baker		Purge & Trap MeOH	163173-29063
			07-16-12A
			12/11/12
			20
			1980

072

7/17/12 RS

Volatile Standard Curve Preparation for 5mL Purge (8260 soln)-SWEETPEA

Expiration Date:		07/18/12									
Date	Conc.	5µg/mL Vol Std #9	5µg/mL Surr	50µg/mL Vol Std #7	50µg/mL Vol Std #8	50µg/mL Surr	5µg/mL Vol Std #10	50µg/mL Vol Std #1	5µg/mL Vol Std #2	50µg/mL Vol Std #3	
Code	µg/L	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	Exp:07-12-12	
07-17-12A	2	2	2	n/a	n/a	n/a	2	n/a	2	n/a	
07-17-12B	5	5	5	n/a	n/a	n/a	5	n/a	5	n/a	
07-17-12C	10	10	10	n/a	n/a	n/a	10	n/a	10	n/a	
07-17-12D	20	20	20	n/a	n/a	n/a	20	n/a	20	n/a	
07-17-12E	50	n/a	n/a	5	5	5	n/a	5	n/a	5	
07-17-12F	100	n/a	n/a	10	10	10	n/a	10	n/a	10	
07-17-12G	200	n/a	n/a	20	20	20	n/a	20	n/a	20	

7/18/12 RS

- Thor 524 curve on pg. 74 RS 7/18/12 RS.

250µg/mL TBA	50µg/mL Vol Std #1
07-05-12N	Exp:07-12-12
1	Exp:07-12-12
2	Exp:07-12-12
3	Exp:07-12-12
4	Exp:07-12-12
5	Exp:07-12-12
6	Exp:07-12-12
7	Exp:07-12-12

9/18/12 A- RS

Method 8260 Gases, 2,000 mg/L, 2 X 0.6 ml

120016-03
Lot# Storage Expiry
180013 ≤ -10 Degrees C 10/17/14

Solv: P/T Methanol
Method 8260 Gases
Lot #: 180013 - 29760
Rec: 10/24/11 MFR exp. 10/17/14

RS

9/18/12 B- RS

Hexachloroethane Solution, 1000 mg/L, 1 ml

020049-02
Lot# Storage Expiry
176700 ≤ -10 Degrees C 7/31/13

Solv: P/T Methanol
Hexachloroethane
Lot #: 176700 - 30724
Rec: 5/9/12 MFR exp. 07/31/13

RS

7/18/12 C- RS

Benzyl Chloride Solution, 1000 mg/L, 1 ml

020228-02
Lot# Storage Expiry
176701 ≤ -10 Degrees C 7/31/13

Solv: P/T Methanol
Benzyl Chloride
Lot #: 176701 - 31019
Rec: 6/19/12 MFR exp. 07/31/13

RS

7/18/12 D- RS

n-Hexane Solution, 1,000 mg/L, 1 ml

020620-02
Lot# Storage Expiry
176773 ≤ -10 Degrees C 7/30/16

Solv: P/T Methanol
n-Hexane Solution
Lot #: 176773 - 31024
Rec: 6/19/12 MFR exp. 07/30/16

RS

7/18/12 RS

11/12
RS

Heptane Solution, 1000 mg/L, 1 ml
 820546-02
 Lot# Storage Expiry
 169174 -5-10 Degrees C 2/18/14
 Solv: P/T Methanol
 Heptane Solution
 Lot #: 169174 - 31039
 Rec: 6/19/12 MFR exp. 02/18/14

RS

11/12
RS

VOC Mix 4-3, 2,000 mg/L, 1 ml
 120166-01
 Lot# Storage Expiry
 185760 ≤ 6 Degrees C 2/14/14
 Solv: P/T Methanol
 VOC Mix 4-3, 2000mg/L
 Lot #: 185760 - 30739
 Rec: 5/9/12 MFR exp. 02/14/14

RS

11/8/12
RS

Method 8260 Gases (Secand Source), 2,000 mg/L, 2 X 0.6 ml
 120016-03-88
 Lot# Storage Expiry
 187974 ≤ -10 Degrees C 4/8/15
 Solv: P/T Methanol
 Method 8260 Gases (SS)
 Lot #: 187974 - 31061
 Rec: 6/19/12 MFR exp. 04/08/15

RS

Supplier	ID #	ID	ug/ml	Lot #	Code	Date	Exp.	ul
07-18-12H								
50ug/ml Vol Work Std #7								
Exp:07/25/12								
Conc. Date Exp.								
Supplier	ID #	ID	ug/ml	Lot #	Code	Date	Exp.	ul
02SI	120016-03	Gas Mix	2000	180013-29760	07-18-12A	07/25/12		100
02SI	020049-02	HEXACHLOROETHANE	1000	176700-30724	07-18-12B	08/08/12		200
02SI	020228-02	Benzyl Chloride	1000	176701-31019	07-18-12C	08/08/12		200
J&T Brand		Purge & Trap MeOH		K14E06-00640	07/18/12	10/08/12		3500
07-18-12I								
50ug/ml Vol Work Std #1								
Exp:07/25/12								
Supplier	ID #	ID	ug/ml	Lot #	Code	Date	Exp.	ul
02SI	020145-02-02	2-CEVE	2000	176770-29827	06-19-12D	08/08/12		50
J&T Brand		Purge & Trap MeOH		K14E06-00640	07/18/12	10/08/12		1950
07-18-12J								
50ug/ml Vol Work Std #8								
Exp:07/25/12								
Supplier	ID #	ID	ug/ml	Lot #	Code	Date	Exp.	ul
02SI	122039-02	Volatile Mix, 20-29	2000	180114-29786	06-19-12E	08/08/12		100
02SI	120023-03	VOC'S-54 COMP	2000	176392-29207	06-19-12F	08/08/12		100
02SI	020232-02	Vinyl Acetate	2000	189764-30727	06-19-12G	05/13/12		100
02SI	020620-02	n-Hexane	1000	176773-31024	07-18-12D	08/08/12		200
02SI	020546-02	Heptane	1000	169174-31039	07-18-12E	08/08/12		200
J&T Brand		Purge & Trap MeOH		K14E06-00640	07/18/12	10/08/12		3300
07-18-12K								
50ug/ml Vol Work Std #2								
Exp:07/25/12								
Supplier	ID #	ID	ug/ml	Lot #	Code	Date	Exp.	ul
02SI	121020-05	HSL'S-Ketone Solution	2000	163375-27145	06-19-12J	08/08/12		100
J&T Brand		Purge & Trap MeOH		K14E06-00640	07/18/12	10/08/12		3900

074

GCMS STANDARD PREPARATION BOOK # _____ PAGE # _____

7/18/12
RS

07-18-12L		Exp: 07/25/12					
50ug/ml Vol Work Std #9		SOURCES		Lot		APPL Code	
50ug/ml Vol Work Std #7		50ug/ml Vol Work Std #8		J&T Brand		APPL Exp Date	
07-18-12M		Exp: 07/25/12					
50ug/ml Vol Work Std #1		50ug/ml Vol Work Std #2		J&T Brand		APPL Exp Date	
07-18-12N		Exp: 07/25/12					
50ug/ml Vol Work Std #2		J&T Brand				ul	
07-18-12O		50ug/ml 8260 Surrogate		Conc.		Date	
Exp: 07/25/12		Exp: 07/25/12		Date		Exp: 07/25/12	
02SI 120002-01		8260B Surr Solution		Lot #		Code	
J&T Brand		Purge & Trap MeOH		K14E06-00640		APPL Exp Date	
07-18-12P		Exp: 07/25/12					
5.0ug/ml 8260 Surrogate		50ug/ml 8260 Surrogate		Lot		APPL Code	
J&T Brand		Purge & Trap MeOH		06/18/12		APPL Exp Date	
07-18-12Q		250ug/ml TBA/IBA/Acetonitrile/Cyclohexanone/Acrolein/2-P					
Exp: 07/25/12		Supplier		ID #		Conc.	
02SI 120166-01		02SI 020229-09		Volat Mix 4-3		Acrolein	
J&T Brand		Purge & Trap MeOH		K14E06-00640		Date	

7/18/12
RS

07-18-12R		50ug/ml VOC std#5		Exp: 07/25/12			
Supplier		ID #		ID		Conc.	
02SI 120016-03-SS		02SI 020145-02-02-SS		8260 Gases(SS)		2-CEVE	
J&T Brand		Purge & Trap MeOH		K14E06-00640		Date	
07-18-12S		50ug/ml VOC Std#6		Exp: 07/25/12			
Supplier		ID #		ID		Conc.	
02SI 120023-03-SS		02SI 120296-01		VOC'S 54 COMP.		Custom 8260 Solution	
02SI 020232-02-SS		02SI 020620-02-SS		Vinyl Acetate(SS)		n-HEXANE	
02SI 020049-02-SS		02SI 020546-02-SS		HEXACHLOROETHANE		Heptane(SS)	
J&T Brand		Purge & Trap MeOH		K14E06-00640		Date	
07-18-12T		250ug/ml TBA/IBA/Acetonitrile/Cyclohexanone/Acrolein/2-P		Exp: 07/25/12			
Supplier		ID #		ID		Conc.	
02SI 120166-01-SS		02SI 020229-09-SS		VOC Mix 4-3 (SS)		Acrolein SOLUTION (SS)	
J&T Brand		Purge & Trap MeOH		K14E06-00640		Date	

7/18/12
RS

Volatile Standard Curve Preparation for 10mL Purge (524 water)-THOR									
Date	Conc. ug/L	07-05-12I		07-05-12K		07-05-12E		07-05-12G	
		Exp: 07-12-12							
07-17-12A	0.2	2	2	n/a	n/a	n/a	n/a	n/a	2
07-17-12B	0.5	5	5	n/a	n/a	n/a	n/a	n/a	5
07-17-12C	1	10	10	n/a	n/a	n/a	n/a	n/a	10
07-17-12D	2	20	20	n/a	n/a	n/a	n/a	n/a	20
07-17-12E	5	n/a	n/a	5	5	10	10	25	25
07-17-12F	10	n/a	n/a	10	10	40	40	35	35
07-17-12G	40	n/a	n/a	40	40	100	100	40	40
07-17-12H	100	n/a	n/a	100	100	366	100	100	100

7/18/12
RS

Volatile Standard Curve		Expiration Date	
Date	Conc. ug/L	Date	Conc. ug/L
07-18-12A	0.2	07-18-12A	0.2
07-18-12B	0.5	07-18-12B	0.5
07-18-12C	1	07-18-12C	1
07-18-12D	2	07-18-12D	2
07-18-12E	5	07-18-12E	5
07-18-12F	10	07-18-12F	10
07-18-12G	20	07-18-12G	20
07-18-12H	40	07-18-12H	40
07-18-12I	100	07-18-12I	100

7/18/12
RS

Volatile Standard Curve P		Expiration Date	
Date	Conc. ug/L	Date	Conc. ug/L
07-18-12A	0.2	07-18-12A	0.2
07-18-12B	0.5	07-18-12B	0.5
07-18-12C	1	07-18-12C	1
07-18-12D	2	07-18-12D	2
07-18-12E	5	07-18-12E	5
07-18-12F	10	07-18-12F	10
07-18-12G	20	07-18-12G	20
07-18-12H	40	07-18-12H	40
07-18-12I	100	07-18-12I	100

Volatile Standard Curve P		Expiration Date	
Date	Conc. ug/L	Date	Conc. ug/L
07-23-12A	0.2	07-23-12A	0.2
07-23-12B	0.5	07-23-12B	0.5
07-23-12C	1	07-23-12C	1
07-23-12D	2	07-23-12D	2
07-23-12E	5	07-23-12E	5
07-23-12F	10	07-23-12F	10
07-23-12G	20	07-23-12G	20
07-23-12H	40	07-23-12H	40
07-23-12I	100	07-23-12I	100

07/19/12A							
2000ug/ml Gasoline							
Supplier	ID #	Conc.	Lot #	Date	APPL	Exp.	
Supelco	LB82077	Gasoline	20,000	LB82077-29979	Code	Date	uL
J&T Brand		Purge & Trap MeOH		K08E01-00640	01-26-12A	02/01/14	200
				07/18/12	08/02/13	1800	

07/19/12B							
2000ug/ml Unleaded Gasoline							
Supplier	ID #	Conc.	Lot #	Date	APPL	Exp.	
Restek	30205	Unleaded Gasoline	50,000	A081012-29980	Code	Date	uL
J&T Brand		Purge & Trap MeOH		K08E01-00640	01-26-12B	02/01/14	80
				07/18/12	08/02/13	1920	

Volatile Standard Curve Preparation for 10mL Purge (8260 water)-THOR

Expiration Date: 07/20/12										
Conc.	5ug/mL Vol Std #9	5ug/mL Surr	50ug/mL Vol Std #7	50ug/mL Vol Std #8	50ug/mL Surr	5ug/mL Vol Std #10	50ug/mL Vol Std #1	50ug/mL Vol Std #2	5ug/mL Vol Std #12	50ug/mL Vol Std #12N
Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:02-26-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12
0.3	3	6	n/a	n/a	n/a	3	n/a	n/a	n/a	3
0.5	5	10	n/a	n/a	n/a	5	n/a	n/a	n/a	5
1	10	20	n/a	n/a	n/a	10	n/a	n/a	n/a	10
2	20	40	n/a	n/a	n/a	20	n/a	n/a	n/a	20
5	n/a	n/a	5	5	10	n/a	5	5	n/a	n/a
10	n/a	n/a	10	10	25	n/a	10	10	n/a	n/a
20	n/a	n/a	20	20	40	n/a	20	20	n/a	n/a
40	n/a	n/a	40	40	80	n/a	40	40	n/a	n/a
100	n/a	n/a	100	100	100	n/a	100	100	n/a	n/a

Gasoline Curve Preparation for 100mL Purge (water)-THOR

Expiration Date: 07/20/12			
Date	Conc.	50ug/mL Gasoline	Final Vol
Code	ug/L	Exp:01-03-13	w/P&T H2O
07-19-12L	20	1	100
07-19-12M	50	2.5	100
07-19-12N	100	5	100
07-19-12O	300	15	100
07-19-12P	600	30	100
07-19-12Q	800	40	100
07-19-12R	1000	50	100

250ug/mL TAPD	Final Vol
07-18-12Q	w/P&T H2O
Exp:07-25-12	mL
3	50
5	50
10	50
15	50
20	50
25	50
30	50
35	50
40	50

Volatile Standard Curve Preparation for 5mL Purge (8260 soil)-SWEETPEA

Expiration Date: 07/20/12										
Conc.	5ug/mL Vol Std #9	5ug/mL Surr	50ug/mL Vol Std #7	50ug/mL Vol Std #8	50ug/mL Surr	5ug/mL Vol Std #10	50ug/mL Vol Std #1	5ug/mL Vol Std #2	50ug/mL Vol Std #12	50ug/mL Vol Std #12N
Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12
2	2	2	n/a	n/a	n/a	2	n/a	2	n/a	n/a
5	5	5	n/a	n/a	n/a	5	n/a	5	n/a	n/a
10	10	10	n/a	n/a	n/a	10	n/a	10	n/a	n/a
20	20	20	n/a	n/a	n/a	20	n/a	20	n/a	n/a
50	n/a	n/a	5	5	5	n/a	5	n/a	n/a	5
100	n/a	n/a	10	10	10	n/a	10	n/a	n/a	10
200	n/a	n/a	20	20	20	n/a	20	n/a	n/a	20

250ug/mL TBA	Final Vol
07-18-12Q	w/P&T H2O
Exp:07-25-12	mL
1	5
2	5
3	5
4	5
5	5
6	5
7	5

Volatile Standard Curve Preparation for 10mL Purge (8260 water)-NEO

Expiration Date: 07/24/12										
Conc.	5ug/mL Vol Std #9	5ug/mL Surr	50ug/mL Vol Std #7	50ug/mL Vol Std #8	50ug/mL Surr	5ug/mL Vol Std #10	50ug/mL Vol Std #1	50ug/mL Vol Std #2	5ug/mL Vol Std #12	50ug/mL Vol Std #12N
Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12	Exp:07-25-12
0.3	3	6	n/a	n/a	n/a	3	n/a	n/a	n/a	3
0.5	5	10	n/a	n/a	n/a	5	n/a	n/a	n/a	5
10	10	20	n/a	n/a	n/a	10	n/a	n/a	n/a	10
5	n/a	n/a	5	5	10	n/a	5	5	n/a	n/a
10	n/a	n/a	10	10	25	n/a	10	10	n/a	n/a
20	n/a	n/a	20	20	40	n/a	20	20	n/a	n/a
40	n/a	n/a	40	40	80	n/a	40	40	n/a	n/a
100	n/a	n/a	100	100	100	n/a	100	100	n/a	n/a
200	n/a	n/a	200	200	125	n/a	200	200	n/a	n/a

250ug/mL TAPD	Final Vol
07-18-12Q	w/P&T H2O
Exp:07-25-12	mL
3	50
5	50
10	50
20	50
25	50
30	50
35	50
40	50
45	50

076

7/24/12
RS

Neo 524						
07-24-12A						
10ug/ml Neo-524 Internal Standard w/ Surrogate				Conc.	Date	Exp
				ug/ml	Lot #	Code
02SI	122450-02	524 Fortification Sol		1000	176776-29295	06-07-12A
J.T. Baker		Purge & Trap MeOH		K08E01-00645	07/20/12	12/15/12

7/24/12
RS

Volatile Standard Curve Preparation for 10mL Purge (524 water)-NEO						
Expiration Date: 07/25/12						
Date	Conc.	5ug/mL Vol Std #9	5ug/mL Vol Std #12	50ug/mL Vol Std #7	50ug/mL Vol Std #8	50ug/mL Vol Std #2
07-24-12B	0.2	2	2	n/a	n/a	n/a
07-24-12C	0.5	5	5	n/a	n/a	n/a
07-24-12D	1	10	10	n/a	n/a	n/a
07-24-12E	2	20	20	n/a	n/a	n/a
07-24-12F	5	n/a	n/a	5	5	5
07-24-12G	10	n/a	n/a	10	10	10
07-24-12H	40	n/a	n/a	40	40	40

7/28/12
RS

Volatile Standard Curve Preparation for 5mL Purge (8260 soil)-SWEETPEA											
Expiration Date: 07/25/12											
Date	Conc.	5ug/mL Vol Std #9	5ug/mL Surr	50ug/mL Vol Std #7	50ug/mL Vol Std #8	50ug/mL Surr	5ug/mL Vol Std #10	50ug/mL Vol Std #11	5ug/mL Vol Std #12	50ug/mL Vol Std #13	50ug/mL Vol Std #14
07-24-12I	2	2	2	n/a	n/a	n/a	2	n/a	n/a	n/a	n/a
07-24-12J	5	5	5	n/a	n/a	n/a	5	n/a	n/a	n/a	n/a
07-24-12K	10	10	10	n/a	n/a	n/a	10	n/a	n/a	n/a	n/a
07-24-12L	20	20	20	n/a	n/a	n/a	20	n/a	n/a	n/a	n/a
07-24-12M	50	n/a	n/a	5	5	5	n/a	5	n/a	n/a	n/a
07-24-12N	100	n/a	n/a	10	10	10	n/a	10	n/a	n/a	n/a
07-24-12O	200	n/a	n/a	20	20	20	n/a	20	n/a	n/a	n/a

7/24/12
RS

Gasoline Curve Preparation for 100mL Purge (water)-THOR			
Expiration Date: 07/25/12			
Date	Conc.	50ug/mL Gasoline	Final Vol
07-24-12P	20	1	100
07-24-12Q	100	5	100
07-24-12R	300	15	100
07-24-12S	600	30	100
07-24-12T	800	40	100

7/25/12
RS

Gasoline Curve Preparation for 100mL Purge (water)-THOR			
Expiration Date: 07/26/12			
Date	Conc.	50ug/mL Gasoline	Final Vol
07-25-12A	20	1	100
07-25-12B	50	2.5	100
07-25-12C	100	5	100
07-25-12D	300	15	100
07-25-12E	600	30	100
07-25-12F	800	40	100
07-25-12G	1000	50	100

Custom VOC Mix, 16-4, 100
mg/L, 4 x 1 ml

122725-03-4PAK

Lot # Storage Expiry
181120 ≤ -10 Degrees C 11/6/13

Solv: P/T Methanol

Custom VOC Mix 16-4

Lot #: 181120 - 30032

Rec: 11/16/11 MFR exp. 11/06/13

Injection Log

Directory: MATHORADATA\120719

Line	Vial	FileName	Multiplier	SampleName	Misc Info	Injected
1	1	0719T01T.D	1	5ng- BFB STD 07-16-12B	2ul	07/19/2012 09:15
2	5	0719T05.D	1	0.3ug/L Vol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 11:01
3	6	0719T06.D	1	0.5ug/L Vol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 11:29
4	7	0719T07.D	1	1.0ug/L Vol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 11:57
5	8	0719T08.D	1	2.0ug/L Vol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 12:25
6	9	0719T09.D	1	5.0ug/L Vol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 12:53
7	10	0719T10.D	1	10ug/L Vol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 13:20
8	11	0719T11.D	1	20ug/L Vol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 13:48
9	12	0719T12.D	1	40ug/L Vol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 14:16
10	13	0719T13.D	1	100ug/L Vol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 14:44
11	28	0719T28.D	1	5ng- BFB Std 07-16-12B	2uL	07/19/2012 21:40
12	30	0719T30.D	1	10ug/L Vol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 22:35
13	31	0719T31.D	1	120719A LCS-1WT (SS)	10ml w/5ul of IS&S: 06-7	07/19/2012 23:03
14	38	0719T38.D	1	120719A BLK-1WT	10ml w/5ul of IS&S: 06-7	07/20/2012 02:18
15	42	0719T42.D	1	AY65114W01	10ml w/5ul of IS&S: 06-7	07/20/2012 04:08
16	46	0719T46.D	1	AY65112W01	10ml w/5ul of IS&S: 06-7	07/20/2012 05:59
17	47	0719T47.D	1	AY65113W01	10ml w/5ul of IS&S: 06-7	07/20/2012 06:26

Injection Log

Directory: M:\THOR\DATA\T120719

Line	Vial	FileName	Multiplier	SampleName	Misc Info	Injected
1	15	0719T15.D	1	5ng-BFB STD 07-16-12B	2uL	07/19/2012 15:39
2	17	0719T17.D	1	VOC MIX MARKER	10ml w/5ul of IS&S: 06-7	07/19/2012 16:35
3	18	0719T18.D	1	20ug/LVol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 17:02
4	19	0719T19.D	1	50ug/LVol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 17:30
5	20	0719T20.D	1	100ug/LVol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 17:58
6	21	0719T21.D	1	300ug/LVol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 18:26
7	22	0719T22.D	1	600ug/LVol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 18:54
8	23	0719T23.D	1	800ug/LVol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 19:21
9	24	0719T24.D	1	1000ug/LVol Std 07-19-12	10ml w/5ul of IS&S: 06-7	07/19/2012 19:49
10	28	0719T28.D	1	5ng- BFB Std 07-16-12B	2uL	07/19/2012 21:40
11	33	0719T33.D	1	CCV gas 300ug/L (SS)	10ml w/5ul of IS&S: 06-7	07/19/2012 23:59
12	34	0719T34.D	1	CCV gas 300ug/L	10ml w/5ul of IS&S: 06-7	07/20/2012 00:27
13	35	0719T35.D	1	LCS gas 300ug/L	10ml w/5ul of IS&S: 06-7	07/20/2012 00:54
14	38	0719T38.D	1	120719A BLK-1WT	10ml w/5ul of IS&S: 06-7	07/20/2012 02:18
15	47	0719T47.D	1	AY65113W01	10ml w/5ul of IS&S: 06-7	07/20/2012 06:26

Injection Log

Directory: MATHOR\DATA\T120725

Line	Vial	FileName	Multiplier	SampleName	Misc Info	Injected
1	1	0725T01T.D	1	5ng- BFB STD 07-16-12B	2ul	07/25/2012 09:32
2	2	0725T03.D	1	VOC MIX MARKER	10ml w/5ul of IS&S: 06-7	07/25/2012 10:22
3	3	0725T04.D	1	20ug/L Vol Std 07-25-12	10ml w/5ul of IS&S: 06-7	07/25/2012 10:50
4	4	0725T05.D	1	50ug/L Vol Std 07-25-12	10ml w/5ul of IS&S: 06-7	07/25/2012 11:17
5	5	0725T06.D	1	100ug/L Vol Std 07-25-12	10ml w/5ul of IS&S: 06-7	07/25/2012 11:45
6	6	0725T07.D	1	300ug/L Vol Std 07-25-13	10ml w/5ul of IS&S: 06-7	07/25/2012 12:13
7	7	0725T08.D	1	600ug/L Vol Std 07-25-14	10ml w/5ul of IS&S: 06-7	07/25/2012 12:41
8	8	0725T09.D	1	800ug/L Vol Std 07-25-15	10ml w/5ul of IS&S: 06-7	07/25/2012 13:08
9	9	0725T10.D	1	1000ug/L Vol Std 07-25-16	10ml w/5ul of IS&S: 06-7	07/25/2012 13:36
10	12	0725T13.D	1	5ng- BFB STD 07-16-12B	2uL	07/25/2012 14:59
11	13	0725T14.D	1	CCV gas 300ug/L	10ml w/5ul of IS&S: 06-7	07/25/2012 15:27
12	14	0725T15.D	1	LCS gas 300ug/L (SS)	10ml w/5ul of IS&S: 06-7	07/25/2012 15:55
13	19	0725T20.D	1	120725A BLK-1WT	10ml w/5ul of IS&S: 06-7	07/25/2012 18:14
14	21	0725T22.D	1	AY65114W02	10ml w/5ul of IS&S: 06-7	07/25/2012 19:09
15	22	0725T23.D	1	AY65112W02	10ml w/5ul of IS&S: 06-7	07/25/2012 19:37

METALS

APPL, INC.

METALS
QC Summary

APPL, INC.

METALS BLANK

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Analyte	Result	LOQ	LOD	DL	Units	Prep Date	Analysis Date	QC Group
6020	LEAD (PB) (DISSOL	0.22 U	0.5	0.22	0.11	ug/L	07/23/12	07/23/12	#602D-120723A-AY65167

Laboratory Control Spike Recovery
METALS

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits	Extract Date	Analysis Date	QC Group
6020	LEAD (PB) (DISSOLVED)	50.0	45.5	91.0	80-120	07/23/12	07/23/12	#602D-120723A-AY65167

Comments: _____

METALS
Sample Data

APPL, INC.

Metals Analysis

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

Attn: Max Solmssen
Project: LTM Red Hill / 1022-024

Sample ID: ES081

Sample Collection Date: 07/18/12

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 68258

APPL ID: AY65112

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
6020	LEAD (PB) (DISSOLVED)	0.22U	0.5	0.22	0.11	ug/L	1	07/23/12	07/23/12

Sample QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\029SMPL.D\029SMPL.D#
 Date Acquired: Jul 23 2012 02:07 pm
 Operator: NBS
 Sample Name: AY65112W08
 Misc Info: 120723A-3015
 Vial Number: 3202
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: Sample
 Prep Dil Factor: 1.11
 Total Dil Factor: 1.11

QC Elements

Element	Conc.	Corr. Conc.	RSD(%)	High Limit	Flag
7 (Li)	----- ug/l	#VALUE!	-----	0	
9 Be	0.00 ug/l	0.00	121.58	1000	
11 B	183.20 ug/l	203.54	1.35	1000	
23 Na	82810.00 ug/l	92001.91	0.90	25000	>Cal
24 Mg	25950.00 ug/l	28830.45	0.39	50000	
27 Al	10.32 ug/l	11.47	3.11	20000	
39 K	3233.00 ug/l	3591.86	0.90	20000	
44 Ca	21940.00 ug/l	24375.34	0.95	50000	
47 Ti	0.95 ug/l	1.06	4.20	1000	
51 V	18.78 ug/l	20.86	0.50	1000	
52 Cr	1.08 ug/l	1.20	1.86	1000	
55 Mn	30.96 ug/l	34.40	0.80	1000	
56 Fe	14.46 ug/l	16.07	1.21	20000	
59 Co	0.51 ug/l	0.56	0.36	1000	
60 Ni	3.53 ug/l	3.92	2.51	1000	
63 Cu	1.08 ug/l	1.20	1.92	1000	
65 Cu	1.12 ug/l	1.24	2.83	1000	
66 Zn	11.82 ug/l	13.13	1.67	1000	
75 As	0.88 ug/l	0.97	2.62	1000	
78 Se	0.14 ug/l	0.16	6.00	1000	
78 Se	0.64 ug/l	0.71	18.08	1000	
88 Sr	140.60 ug/l	156.21	0.59	1000	
88 Sr	135.90 ug/l	150.98	1.29	1000	
95 Mo	11.66 ug/l	12.95	0.69	1000	
106 (Cd)	----- ug/l	#VALUE!	-----	#####	
107 Ag	0.02 ug/l	0.02	3.84	500	
108 (Cd)	----- ug/l	#VALUE!	-----	#####	
111 Cd	0.06 ug/l	0.07	25.23	1000	
118 Sn	0.11 ug/l	0.12	7.82	#####	
118 Sn	0.12 ug/l	0.13	10.08	#####	
118 Sn	0.12 ug/l	0.13	3.24	1000	
121 Sb	0.12 ug/l	0.14	2.40	1000	
137 Ba	11.60 ug/l	12.89	0.74	1000	
205 Tl	0.08 ug/l	0.09	1.85	1000	
206 (Pb)	----- ug/l	#VALUE!	-----	#####	
207 (Pb)	----- ug/l	#VALUE!	-----	#####	
208 Pb	0.09 ug/l	0.10	3.59	1000	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-45225.24	11.37	-41328.95	109.4	70 - 120	IS Fai
45 Sc	3148469.80	0.29	3008024.30	104.7	70 - 120	
45 Sc	449359.94	0.48	423303.94	106.2	70 - 120	
45 Sc	9289287.00	1.01	8607281.00	107.9	70 - 120	
72 Ge	782338.50	1.59	774468.63	101.0	70 - 120	
72 Ge	284030.81	0.61	282128.91	100.7	70 - 120	
72 Ge	1928524.60	0.26	1882554.90	102.4	70 - 120	
115 In	5452920.50	0.27	5556751.00	98.1	70 - 120	
115 In	2960486.30	0.42	3029632.80	97.7	70 - 120	
115 In	12080720.00	0.71	12097256.00	99.9	70 - 120	
159 Tb	16492033.00	0.96	16269544.00	101.4	70 - 120	
165 Ho	16168885.00	1.08	15819307.00	102.2	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

1 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Fail
 ISTD: Fail

Metals Analysis

Environet, Inc.
650 Iwilei Rd, #204
Honolulu, HI 96817

Attn: Max Solmssen
Project: LTM Red Hill / 1022-024

Sample ID: ES082

Sample Collection Date: 07/18/12

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

ARF: 68258

APPL ID: AY65113

Method	Analyte	Result	LOQ	LOD	DL	Units	DF	Prep Date	Analysis Date
6020	LEAD (PB) (DISSOLVED)	0.42J	0.5	0.22	0.11	ug/L	1	07/23/12	07/23/12

J = Estimated value.

Sample QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\030SMPL.D\030SMPL.D#
 Date Acquired: Jul 23 2012 02:14 pm
 Operator: NBS
 Sample Name: AY65113W08
 Misc Info: 120723A-3015
 Vial Number: 3203
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: Sample
 Prep Dil Factor: 1.11
 Total Dil Factor: 1.11

QC Elements

Element	Conc.	Corr. Conc.	RSD(%)	High Limit	Flag
7 (Li)	----- ug/l	#VALUE!	-----	0	
9 Be	0.01 ug/l	0.02	30.62	1000	
11 B	109.20 ug/l	121.32	0.88	1000	
23 Na	60620.00 ug/l	67348.82	1.06	25000	>Cal
24 Mg	26470.00 ug/l	29408.17	0.39	50000	
27 Al	14.72 ug/l	16.35	4.62	20000	
39 K	2583.00 ug/l	2869.71	1.75	20000	
44 Ca	14340.00 ug/l	15931.74	0.56	50000	
47 Ti	1.61 ug/l	1.78	7.05	1000	
51 V	1.15 ug/l	1.28	0.48	1000	
52 Cr	0.65 ug/l	0.72	2.39	1000	
55 Mn	1561.00 ug/l	1734.27	0.81	1000	>Cal
56 Fe	1413.00 ug/l	1569.84	0.70	20000	
59 Co	0.32 ug/l	0.35	4.84	1000	
60 Ni	0.36 ug/l	0.40	1.64	1000	
63 Cu	0.32 ug/l	0.36	3.24	1000	
65 Cu	0.34 ug/l	0.37	2.54	1000	
66 Zn	10.49 ug/l	11.65	2.03	1000	
75 As	0.09 ug/l	0.10	6.25	1000	
78 Se	0.02 ug/l	0.03	19.52	1000	
78 Se	0.45 ug/l	0.50	29.48	1000	
88 Sr	121.10 ug/l	134.54	0.95	1000	
88 Sr	117.60 ug/l	130.65	0.37	1000	
95 Mo	0.15 ug/l	0.16	6.58	1000	
106 (Cd)	----- ug/l	#VALUE!	-----	#####	
107 Ag	0.00 ug/l	0.00	35.83	500	
108 (Cd)	----- ug/l	#VALUE!	-----	#####	
111 Cd	0.12 ug/l	0.14	14.77	1000	
118 Sn	0.10 ug/l	0.11	2.17	#####	
118 Sn	0.11 ug/l	0.12	16.89	#####	
118 Sn	0.10 ug/l	0.11	10.32	1000	
121 Sb	0.05 ug/l	0.06	5.36	1000	
137 Ba	24.42 ug/l	27.13	0.60	1000	
205 Tl	0.07 ug/l	0.08	2.92	1000	
206 (Pb)	----- ug/l	#VALUE!	-----	#####	
207 (Pb)	----- ug/l	#VALUE!	-----	#####	
208 Pb	0.38 ug/l	0.42	1.21	1000	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-47696.10	12.14	-41328.95	115.4	70 - 120	IS Fai
45 Sc	3209263.00	1.32	3008024.30	106.7	70 - 120	
45 Sc	452598.53	0.38	423303.94	106.9	70 - 120	
45 Sc	9421175.00	0.67	8607281.00	109.5	70 - 120	
72 Ge	796992.69	0.24	774468.63	102.9	70 - 120	
72 Ge	287613.75	0.82	282128.91	101.9	70 - 120	
72 Ge	1967507.10	0.45	1882554.90	104.5	70 - 120	
115 In	5581677.50	0.87	5556751.00	100.4	70 - 120	
115 In	2986463.00	0.38	3029632.80	98.6	70 - 120	
115 In	12343514.00	0.73	12097256.00	102.0	70 - 120	
159 Tb	16930864.00	0.64	16269544.00	104.1	70 - 120	
165 Ho	16544507.00	0.24	15819307.00	104.6	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

2 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Fail
 ISTD: Fail

**METALS
Calibration Data**

APPL, INC.

A.P.P.L. INC.

2A

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: A.P.P.L. INC. Contract: Environet, Inc.ARF No: 68258 SDG: 68258Initial Calibration Source: CPIContinuing Calibration Source: Environmental ExpressAnalysis Date: 07/23/12 Concentration Units: ug/L

Analyte	Initial Calibration			Continuing Calibration						M
	True	Found 11:45	%R(1)	True CCV1	Found 12:05	%R(1)	True CCV1	Found 13:32	%R(1)	
Lead (Pb)	100	99.7	99.7	50	52.39	105	50	51.97	104	P

(1) Control Limits: Metals 90-110

ILM02.0

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: A.P.P.L. INC. Contract: Environet, Inc.ARF No: 68258 SDG: 68258Initial Calibration Source: CPIContinuing Calibration Source: Environmental ExpressAnalysis Date: 07/23/12 Concentration Units: ug/L

Analyte	Initial Calibration			Continuing Calibration						M
	True	Found 11:45	%R(1)	True CCV1	Found 15:07	%R(1)	True	Found	%R(1)	
Lead (Pb)	100	99.7	99.7	50	51.71	103				P

A.P.P.L. INC.

3

BLANKS

Lab Name: A.P.P.L. INC.

Contract: Environet, Inc.

ARF No.: 68258

SDG: 68258

Preparation Blank Matrix (soil/water): water

Preparation Blank Concentration Units (ug/L or mg/kg): ug/L

Analysis Date: 07/23/12

Analyte	Initial Calibration Blank (ug/L) C	Continuing Calibration Blank (ug/L)						Preparation Blank C	M
		1 C	2 C	3 C	12:12	13:45	15:20		
Lead (Pb)	11:58 .50 U	.50 U	.50 U	.50 U	.50 U	.50 U	12:58 .50 U	P	

ICP INTERFERENCE CHECK SAMPLE

Lab Name: A.P.P.L. INC.
 ARF No.: 68258
 ICP ID Number: Optimus

Contract: Environet, Inc.
 SDG: 68258
 ICS Source: Environmental Express

Analysis Date: 07/23/12

Concentration Units: ug/L

Analyte	True		Initial Found		
	Sol A	Sol AB	Sol A 12:18	Sol AB 12:25	%R(1)
Lead (Pb)		500	0.4092	437.6	87.5

(1) Control Limits: Metals 80-120

Calibration Blank QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\004CAL
 Date Acquired: Jul 23 2012 11:12 am
 Operator: NBS
 Sample Name: Calibration Blank
 Misc Info:
 Vial Number: 1102
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:09 am
 Sample Type: CalBlk
 Total Dil Factor: 1.00

QC&ISTD Elements

Element	CPS Mean	SD	RSD(%)
6 Li	-41328.94 A	2428.00	5.87
7 (Li)	4659754.00 A	18000.00	0.39
9 Be	44.45 P	5.09	11.46
11 B	13122.80 P	299.70	2.28
23 Na	64072.43 P	449.30	0.70
24 Mg	208.90 P	51.68	24.74
27 Al	91.11 P	10.71	11.75
39 K	42023.77 P	406.10	0.97
44 Ca	376.38 P	20.70	5.50
45 Sc	3008024.00 A	6696.00	0.22
45 Sc	423303.91 A	3421.00	0.81
45 Sc	8607281.00 A	46940.00	0.55
47 Ti	1.78 P	1.54	86.62
51 V	48.45 P	8.57	17.69
52 Cr	424.90 P	14.87	3.50
55 Mn	212.89 P	30.82	14.48
56 Fe	3307.53 P	84.69	2.56
59 Co	83.56 P	8.04	9.62
60 Ni	109.34 P	9.33	8.54
63 Cu	295.56 P	2.04	0.69
65 Cu	135.11 P	17.40	12.88
66 Zn	310.23 P	10.10	3.26
72 Ge	774468.63 A	4129.00	0.53
72 Ge	282128.91 A	1085.00	0.38
72 Ge	1882555.00 A	1994.00	0.11
75 As	27.11 P	4.68	17.27
78 Se	22.89 P	3.27	14.30
78 Se	148.78 P	4.86	3.26
88 Sr	168.90 P	24.12	14.28
88 Sr	598.92 P	5.09	0.85
95 Mo	93.34 P	14.53	15.57
106 (Cd)	3.33 P	3.33	99.99
107 Ag	126.67 P	8.82	6.96
108 (Cd)	5.56 P	5.09	91.65
111 Cd	13.08 P	16.84	128.75
115 In	5556751.00 A	25450.00	0.46
115 In	3029633.00 A	2589.00	0.09
115 In	12097260.00 A	3381.00	0.03
118 Sn	187.79 P	62.04	33.04
118 Sn	96.67 P	8.82	9.12
118 Sn	352.24 P	45.38	12.88
121 Sb	131.12 P	15.75	12.01
137 Ba	72.23 P	27.76	38.44
159 Tb	16269540.00 A	108400.00	0.67
165 Ho	15819310.00 A	42930.00	0.27
205 Tl	195.56 P	65.18	33.33
206 (Pb)	530.03 P	72.19	13.62
207 (Pb)	423.36 P	56.67	13.39
208 Pb	1940.13 P	171.40	8.83

Calibration Standard QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\005CALB.D\005CALB.D#
 Date Acquired: Jul 23 2012 11:18 am
 Operator: NBS
 Sample Name: 120723 Standard 1
 Misc Info:
 Vial Number: 1103
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:16 am
 Sample Type: CalStd
 Total Dil Factor: 1.00

QC&ISTD Elements

Element	CPS Mean	SD	RSD(%)	Cal Coef
6 Li	-41816.46 A	2968.00	7.10	0.0000
7 (Li)	4828315.00 A	24030.00	0.50	0.0000
9 Be	431.13 P	39.49	9.16	0.0000
11 B	13550.97 P	211.20	1.56	0.0000
23 Na	68743.03 P	482.40	0.70	0.0000
24 Mg	1320.10 P	78.40	5.94	0.0000
27 Al	323.35 P	17.64	5.46	0.0000
39 K	42278.87 P	294.60	0.70	0.0000
44 Ca	453.33 P	18.07	3.99	0.0000
45 Sc	2965625.00 A	8375.00	0.28	0.0000
45 Sc	409348.19 A	5583.00	1.36	0.0000
45 Sc	8658112.00 A	37130.00	0.43	0.0000
47 Ti	16.44 P	2.78	16.88	0.0000
51 V	457.35 P	18.04	3.94	0.0000
52 Cr	881.37 P	20.83	2.36	0.0000
55 Mn	432.46 P	14.69	3.40	0.0000
56 Fe	10304.63 P	194.30	1.89	0.0000
59 Co	556.02 P	39.26	7.06	0.0000
60 Ni	221.78 P	29.82	13.45	0.0000
63 Cu	848.93 P	36.62	4.31	0.0000
65 Cu	429.35 P	5.81	1.35	0.0000
66 Zn	473.79 P	16.88	3.56	0.0000
72 Ge	782088.38 A	7482.00	0.96	0.0000
72 Ge	280806.41 A	3743.00	1.33	0.0000
72 Ge	1881152.00 A	8788.00	0.47	0.0000
75 As	89.00 P	5.24	5.89	0.0000
78 Se	46.89 P	2.87	6.13	0.0000
78 Se	154.67 P	3.71	2.40	0.0000
88 Sr	594.48 P	28.74	4.83	0.0000
88 Sr	4045.16 P	130.50	3.23	0.0000
95 Mo	706.71 P	10.00	1.42	0.0000
106 (Cd)	47.78 P	7.70	16.11	0.0000
107 Ag	994.52 P	69.32	6.97	0.0000
108 (Cd)	28.89 P	10.18	35.24	0.0000
111 Cd	397.49 P	77.04	19.38	0.0000
115 In	5449510.00 A	78530.00	1.44	0.0000
115 In	2939285.00 A	18850.00	0.64	0.0000
115 In	11960780.00 A	79640.00	0.67	0.0000
118 Sn	1004.52 P	97.90	9.75	0.0000
118 Sn	555.59 P	7.70	1.39	0.0000
118 Sn	2132.46 P	102.20	4.79	0.0000
121 Sb	1841.30 P	47.65	2.59	0.0000
137 Ba	583.37 P	41.64	7.14	0.0000
159 Tb	16219180.00 A	173500.00	1.07	0.0000
165 Ho	15690520.00 A	6789.00	0.04	0.0000
205 Tl	3149.38 P	120.10	3.81	0.0000
206 (Pb)	1167.88 P	66.20	5.67	0.0000
207 (Pb)	974.52 P	68.35	7.01	0.0000
208 Pb	4646.06 P	99.59	2.14	0.0000

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-41816.47	7.10	-41328.95	101.2	70 -	120 IS Fail
45 Sc	2965624.50	0.28	3008024.30	98.6	70 -	120
45 Sc	409348.25	1.36	423303.94	96.7	70 -	120
45 Sc	8658112.00	0.43	8607281.00	100.6	70 -	120
72 Ge	782088.44	0.96	774468.63	101.0	70 -	120
72 Ge	280806.38	1.33	282128.91	99.5	70 -	120
72 Ge	1881152.00	0.47	1882554.90	99.9	70 -	120
115 In	5449510.50	1.44	5556751.00	98.1	70 -	120
115 In	2939285.30	0.64	3029632.80	97.0	70 -	120
115 In	11960782.00	0.67	12097256.00	98.9	70 -	120
159 Tb	16219185.00	1.07	16269544.00	99.7	70 -	120
165 Ho	15690520.00	0.04	15819307.00	99.2	70 -	120

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

--- :Element Failures --- :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Pass
 ISTD: Fail

Calibration Standard QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\006CALB.D\006CALB.D#
 Date Acquired: Jul 23 2012 11:25 am
 Operator: NBS
 Sample Name: 120723 Standard 2
 Misc Info:
 Vial Number: 1104
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:22 am
 Sample Type: CalStd
 Total Dil Factor: 1.00

QC&ISTD Elements

Element	CPS Mean	SD	RSD(%)	Cal Coef
6 Li	-38271.70 A	2707.00	7.07	0.0000
7 (Li)	4793284.00 A	21050.00	0.44	1.0000
9 Be	4235.18 P	94.10	2.22	1.0000
11 B	14857.80 P	552.40	3.72	1.0000
23 Na	75362.73 P	100.00	0.13	1.0000
24 Mg	10653.06 P	135.80	1.27	1.0000
27 Al	1959.08 P	90.10	4.60	1.0000
39 K	48852.40 P	393.50	0.81	1.0000
44 Ca	1103.45 P	35.31	3.20	1.0000
45 Sc	2946623.00 A	16940.00	0.57	0.0000
45 Sc	412843.69 A	2177.00	0.53	0.0000
45 Sc	8594454.00 A	44450.00	0.52	0.0000
47 Ti	96.00 P	12.22	12.73	1.0000
51 V	2942.12 P	46.06	1.57	1.0000
52 Cr	3650.28 P	70.82	1.94	1.0000
55 Mn	2454.02 P	50.72	2.07	1.0000
56 Fe	62723.29 P	271.00	0.43	1.0000
59 Co	4763.95 P	121.10	2.54	1.0000
60 Ni	1299.19 P	41.19	3.17	1.0000
63 Cu	3529.81 P	57.76	1.64	1.0000
65 Cu	1788.58 P	45.70	2.56	1.0000
66 Zn	1012.94 P	61.69	6.09	1.0000
72 Ge	775596.38 A	1679.00	0.22	0.0000
72 Ge	280488.00 A	3782.00	1.35	0.0000
72 Ge	1879230.00 A	21100.00	1.12	0.0000
75 As	564.46 P	12.53	2.22	1.0000
78 Se	248.34 P	9.84	3.96	1.0000
78 Se	213.11 P	13.02	6.11	1.0000
88 Sr	4516.43 P	56.70	1.26	1.0000
88 Sr	32554.49 P	334.70	1.03	1.0000
95 Mo	6087.07 P	193.60	3.18	1.0000
106 (Cd)	324.46 P	6.94	2.14	1.0000
107 Ag	7810.19 P	109.40	1.40	1.0000
108 (Cd)	241.12 P	50.15	20.80	1.0000
111 Cd	3325.03 P	144.20	4.34	1.0000
115 In	5421674.00 A	15610.00	0.29	0.0000
115 In	2910753.00 A	26240.00	0.90	0.0000
115 In	11770410.00 A	19480.00	0.17	0.0000
118 Sn	4565.35 P	94.32	2.07	1.0000
118 Sn	2638.12 P	42.87	1.63	1.0000
118 Sn	9902.71 P	120.40	1.22	1.0000
121 Sb	13280.07 P	116.80	0.88	1.0000
137 Ba	5013.31 P	58.37	1.16	1.0000
159 Tb	16000270.00 A	160400.00	1.00	0.0000
165 Ho	15559700.00 A	126400.00	0.81	0.0000
205 Tl	28171.76 P	356.80	1.27	1.0000
206 (Pb)	9680.50 P	330.70	3.42	1.0000
207 (Pb)	8091.67 P	204.00	2.52	1.0000
208 Pb	38229.67 P	528.20	1.38	1.0000

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-38271.70	7.07	-41328.95	92.6	70 -	120 IS Fail
45 Sc	2946623.30	0.57	3008024.30	98.0	70 -	120
45 Sc	412843.72	0.53	423303.94	97.5	70 -	120
45 Sc	8594454.00	0.52	8607281.00	99.9	70 -	120
72 Ge	775596.38	0.22	774468.63	100.1	70 -	120
72 Ge	280488.03	1.35	282128.91	99.4	70 -	120
72 Ge	1879229.80	1.12	1882554.90	99.8	70 -	120
115 In	5421673.50	0.29	5556751.00	97.6	70 -	120
115 In	2910753.50	0.90	3029632.80	96.1	70 -	120
115 In	11770411.00	0.17	12097256.00	97.3	70 -	120
159 Tb	16000275.00	1.00	16269544.00	98.3	70 -	120
165 Ho	15559705.00	0.81	15819307.00	98.4	70 -	120

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

--- :Element Failures --- :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Pass
 ISTD: Fail

Calibration Standard QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\007CALB.D\007CALB.D#
 Date Acquired: Jul 23 2012 11:32 am
 Operator: NBS
 Sample Name: 120723 Standard 3
 Misc Info:
 Vial Number: 1105
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:29 am
 Sample Type: CalStd
 Total Dil Factor: 1.00

QC&ISTD Elements

Element	CPS Mean	SD	RSD(%)	Cal Coef
6 Li	-40141.68 A	5393.00	13.44	0.0000
7 (Li)	4772095.00 A	27750.00	0.58	0.4044
9 Be	208836.70 P	1267.00	0.61	1.0000
11 B	137641.50 P	1350.00	0.98	0.9952
23 Na	561413.63 P	2113.00	0.38	0.8916
24 Mg	524168.31 P	2475.00	0.47	1.0000
27 Al	92849.74 P	461.20	0.50	0.9997
39 K	344469.09 P	2025.00	0.59	0.9939
44 Ca	38040.14 P	664.10	1.75	0.9998
45 Sc	2929570.00 A	29350.00	1.00	0.0000
45 Sc	413552.91 A	3122.00	0.75	0.0000
45 Sc	8628072.00 A	89100.00	1.03	0.0000
47 Ti	4933.33 P	62.99	1.28	0.9984
51 V	135471.50 P	1339.00	0.99	0.9991
52 Cr	158083.50 P	1018.00	0.64	0.9990
55 Mn	112729.30 P	106.60	0.09	1.0000
56 Fe	2648855.00 A	17400.00	0.66	0.9998
59 Co	230427.80 P	180.70	0.08	1.0000
60 Ni	57533.91 P	431.00	0.75	1.0000
63 Cu	155413.50 P	641.40	0.41	0.9973
65 Cu	75685.46 P	891.90	1.18	0.9968
66 Zn	32393.10 P	111.10	0.34	0.9902
72 Ge	768266.81 A	21970.00	2.86	0.0000
72 Ge	277329.91 A	2067.00	0.75	0.0000
72 Ge	1886908.00 A	20910.00	1.11	0.0000
75 As	25744.79 P	105.60	0.41	0.9999
78 Se	10980.96 P	103.00	0.94	1.0000
78 Se	2890.63 P	48.87	1.69	0.9990
88 Sr	209820.50 P	803.00	0.38	1.0000
88 Sr	1454092.00 A	8453.00	0.58	1.0000
95 Mo	293486.50 P	604.00	0.21	1.0000
106 (Cd)	15022.92 P	191.90	1.28	0.9993
107 Ag	377162.50 P	3424.00	0.91	0.9999
108 (Cd)	10984.61 P	170.90	1.56	1.0000
111 Cd	165458.50 P	465.90	0.28	0.9999
115 In	5327447.00 A	55260.00	1.04	0.0000
115 In	2910250.00 A	24200.00	0.83	0.0000
115 In	11944630.00 A	109500.00	0.92	0.0000
118 Sn	198230.09 P	1792.00	0.90	0.9964
118 Sn	115238.30 P	443.30	0.38	0.9969
118 Sn	450003.50 P	3055.00	0.68	0.9966
121 Sb	649268.31 P	6236.00	0.96	0.9996
137 Ba	238687.59 P	2297.00	0.96	1.0000
159 Tb	16069320.00 A	147900.00	0.92	0.0000
165 Ho	15575810.00 A	135900.00	0.87	0.0000
205 Tl	1268132.00 A	1504.00	0.12	1.0000
206 (Pb)	468874.31 P	1209.00	0.26	0.9996
207 (Pb)	395048.69 P	2701.00	0.68	0.9997
208 Pb	1851604.00 P	7982.00	0.43	0.9997

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-40141.69	13.44	-41328.95	97.1	70 -	120 IS Fail
45 Sc	2929570.30	1.00	3008024.30	97.4	70 -	120
45 Sc	413552.91	0.75	423303.94	97.7	70 -	120
45 Sc	8628072.00	1.03	8607281.00	100.2	70 -	120
72 Ge	768266.88	2.86	774468.63	99.2	70 -	120
72 Ge	277329.88	0.75	282128.91	98.3	70 -	120
72 Ge	1886908.30	1.11	1882554.90	100.2	70 -	120
115 In	5327447.00	1.04	5556751.00	95.9	70 -	120
115 In	2910250.50	0.83	3029632.80	96.1	70 -	120
115 In	11944625.00	0.92	12097256.00	98.7	70 -	120
159 Tb	16069324.00	0.92	16269544.00	98.8	70 -	120
165 Ho	15575815.00	0.87	15819307.00	98.5	70 -	120

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

--- :Element Failures --- :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Pass
 ISTD: Fail

Calibration Standard QC Report

Data File: C:\ICPCHEM\1\DATA\12023k00.B\008CAL5.D\008CAL5.D#
 Date Acquired: Jul 23 2012 11:39 am
 Operator: NBS
 Sample Name: 120723 Standard 4
 Misc Info:
 Vial Number: 1106
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:36 am
 Sample Type: CalStd
 Total Dil Factor: 1.00

QC&ISTD Elements

Element	CPS Mean	SD	RSD(%)	Cal Coef
6 Li	-38098.84 A	8974.00	23.56	0.0000
7 (Li)	4836046.00 A	91920.00	1.90	0.0872
9 Be	423835.19 P	6357.00	1.50	1.0000
11 B	267890.91 P	3612.00	1.35	1.0000
23 Na	1070874.00 A	4310.00	0.40	1.0000
24 Mg	1031127.00 A	10470.00	1.02	1.0000
27 Al	188079.59 P	454.70	0.24	1.0000
39 K	660046.19 P	3327.00	0.50	1.0000
44 Ca	76449.13 P	280.60	0.37	1.0000
45 Sc	3012870.00 A	36160.00	1.20	0.0000
45 Sc	415503.00 A	2302.00	0.55	0.0000
45 Sc	8724734.00 A	61090.00	0.70	0.0000
47 Ti	10057.35 P	118.60	1.18	1.0000
51 V	274992.41 P	1358.00	0.49	1.0000
52 Cr	318432.59 P	2283.00	0.72	1.0000
55 Mn	229116.20 P	2629.00	1.15	1.0000
56 Fe	5265441.00 A	60300.00	1.15	1.0000
59 Co	463151.41 P	3143.00	0.68	1.0000
60 Ni	115989.80 P	858.10	0.74	1.0000
63 Cu	312560.19 P	284.40	0.09	1.0000
65 Cu	153215.50 P	1440.00	0.94	1.0000
66 Zn	65080.04 P	448.40	0.69	1.0000
72 Ge	784051.63 A	4382.00	0.56	0.0000
72 Ge	278662.41 A	4547.00	1.63	0.0000
72 Ge	1912165.00 A	7636.00	0.40	0.0000
75 As	52232.86 P	125.80	0.24	1.0000
78 Se	22351.78 P	86.89	0.39	1.0000
78 Se	5691.24 P	66.14	1.16	1.0000
88 Sr	432452.91 P	383.60	0.09	1.0000
88 Sr	2910274.00 A	5311.00	0.18	1.0000
95 Mo	594285.13 P	5554.00	0.93	1.0000
106 (Cd)	30253.12 P	587.80	1.94	1.0000
107 Ag	759463.69 P	6407.00	0.84	1.0000
108 (Cd)	22110.18 P	292.70	1.32	1.0000
111 Cd	329599.69 P	1169.00	0.35	1.0000
115 In	5454744.00 A	67430.00	1.24	0.0000
115 In	2935754.00 A	7842.00	0.27	0.0000
115 In	11935970.00 A	72980.00	0.61	0.0000
118 Sn	405178.81 P	4698.00	1.16	1.0000
118 Sn	234336.80 P	592.90	0.25	1.0000
118 Sn	907635.88 P	8976.00	0.99	1.0000
121 Sb	1204220.00 A	13910.00	1.16	1.0000
137 Ba	475384.59 P	3682.00	0.77	1.0000
159 Tb	16151070.00 A	201800.00	1.25	0.0000
165 Ho	15582170.00 A	199800.00	1.28	0.0000
205 Tl	2472780.00 A	25800.00	1.04	1.0000
206 (Pb)	935228.19 P	12790.00	1.37	1.0000
207 (Pb)	786432.88 P	10080.00	1.28	1.0000
208 Pb	3497400.00 A	34560.00	0.99	1.0000

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-38098.85	23.56	-41328.95	92.2	70 -	120 IS Fail
45 Sc	3012869.50	1.20	3008024.30	100.2	70 -	120
45 Sc	415502.97	0.55	423303.94	98.2	70 -	120
45 Sc	8724734.00	0.70	8607281.00	101.4	70 -	120
72 Ge	784051.56	0.56	774468.63	101.2	70 -	120
72 Ge	278662.41	1.63	282128.91	98.8	70 -	120
72 Ge	1912165.40	0.40	1882554.90	101.6	70 -	120
115 In	5454744.00	1.24	5556751.00	98.2	70 -	120
115 In	2935754.00	0.27	3029632.80	96.9	70 -	120
115 In	11935967.00	0.61	12097256.00	98.7	70 -	120
159 Tb	16151070.00	1.25	16269544.00	99.3	70 -	120
165 Ho	15582175.00	1.28	15819307.00	98.5	70 -	120

ISTD Ref File : C:\ICPCHEM\1\DATA\12023k00.B\004CALB.D\004CALB.D#

--- :Element Failures --- :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Pass
 ISTD: Fail

QCS QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\009_QCS.D\009_QCS.D#
 Date Acquired: Jul 23 2012 11:45 am
 Operator: NBS
 Sample Name: ICV 120723
 Misc Info:
 Vial Number: 1107
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: QCS
 Total Dil Factor: 1.00

QC Elements

Element	Conc.	RSD(%)	Expected	QC Range(%)	Flag
7 (Li)	----- ug/l	-----	100.00	90 - 110	
9 Be	100.60 ug/l	0.64	100.00	90 - 110	
11 B	101.80 ug/l	1.22	100.00	90 - 110	
23 Na	2398.00 ug/l	0.47	2500.00	90 - 110	
24 Mg	2487.00 ug/l	0.77	2500.00	90 - 110	
27 Al	2446.00 ug/l	0.17	2500.00	90 - 110	
39 K	2316.00 ug/l	0.48	2500.00	90 - 110	
44 Ca	2406.00 ug/l	0.84	2500.00	90 - 110	
47 Ti	97.92 ug/l	1.55	100.00	90 - 110	
51 V	101.80 ug/l	0.59	100.00	90 - 110	
52 Cr	101.70 ug/l	0.35	100.00	90 - 110	
55 Mn	101.10 ug/l	0.14	100.00	90 - 110	
56 Fe	2430.00 ug/l	0.38	2500.00	90 - 110	
59 Co	99.14 ug/l	0.73	100.00	90 - 110	
60 Ni	101.20 ug/l	0.59	100.00	90 - 110	
63 Cu	99.40 ug/l	0.49	100.00	90 - 110	
65 Cu	99.41 ug/l	0.52	100.00	90 - 110	
66 Zn	101.50 ug/l	1.04	100.00	90 - 110	
75 As	99.15 ug/l	1.09	100.00	90 - 110	
78 Se	99.79 ug/l	0.71	100.00	90 - 110	
78 Se	100.20 ug/l	1.67	100.00	90 - 110	
88 Sr	98.23 ug/l	1.15	100.00	90 - 110	
88 Sr	97.55 ug/l	0.36	100.00	90 - 110	
95 Mo	99.68 ug/l	1.27	100.00	90 - 110	
106 (Cd)	----- ug/l	-----	100.00	90 - 110	
107 Ag	51.17 ug/l	1.02	50.00	90 - 110	
108 (Cd)	----- ug/l	-----	100.00	90 - 110	
111 Cd	99.11 ug/l	0.54	100.00	90 - 110	
118 Sn	48.25 ug/l	5.45	50.00	90 - 110	
118 Sn	47.94 ug/l	8.36	50.00	90 - 110	
118 Sn	50.31 ug/l	5.64	50.00	90 - 110	
121 Sb	102.00 ug/l	1.63	100.00	90 - 110	
137 Ba	98.36 ug/l	0.60	100.00	90 - 110	
205 Tl	98.39 ug/l	0.49	100.00	90 - 110	
206 (Pb)	----- ug/l	-----	100.00	90 - 110	
207 (Pb)	----- ug/l	-----	100.00	90 - 110	
208 Pb	99.70 ug/l	0.09	100.00	90 - 110	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-42030.20	5.53	-41328.95	101.7	70 - 120	IS Fail
45 Sc	3002840.00	1.08	3008024.30	99.8	70 - 120	
45 Sc	417571.03	0.15	423303.94	98.6	70 - 120	
45 Sc	8771717.00	1.09	8607281.00	101.9	70 - 120	
72 Ge	784054.69	1.30	774468.63	101.2	70 - 120	
72 Ge	283383.13	0.95	282128.91	100.4	70 - 120	
72 Ge	1906694.80	0.29	1882554.90	101.3	70 - 120	
115 In	5475908.00	1.09	5556751.00	98.5	70 - 120	
115 In	2925404.50	1.07	3029632.80	96.6	70 - 120	
115 In	12051751.00	0.51	12097256.00	99.6	70 - 120	
159 Tb	16249860.00	0.32	16269544.00	99.9	70 - 120	
165 Ho	15777454.00	0.86	15819307.00	99.7	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

0 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Pass
 ISTD: Fail

CCB QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\011_CCB.D\011_CCB.D#
 Date Acquired: Jul 23 2012 11:58 am
 Operator: NBS
 Sample Name: ICB 120723
 Misc Info:
 Vial Number: 1102
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: CCB
 Total Dil Factor: 1.00

QC Elements

Element	Conc.	RSD(%)	High Limit	Flag
7 (Li)	----- ug/l	-----	#####	
9 Be	0.01 ug/l	62.17	0.12	
11 B	-0.06 ug/l	103.17	15.00	
23 Na	-8.87 ug/l	28.81	77.10	
24 Mg	-0.02 ug/l	266.60	7.50	
27 Al	0.35 ug/l	64.01	3.96	
39 K	2.17 ug/l	108.79	19.20	
44 Ca	-7.98 ug/l	38.97	90.00	
47 Ti	0.00 ug/l	285.52	0.78	
51 V	0.01 ug/l	77.78	0.21	
52 Cr	-0.02 ug/l	56.40	0.12	
55 Mn	-0.01 ug/l	30.35	0.18	
56 Fe	0.09 ug/l	39.91	40.80	
59 Co	0.01 ug/l	50.79	0.09	
60 Ni	0.00 ug/l	335.93	0.48	
63 Cu	-0.02 ug/l	3.48	0.39	
65 Cu	-0.01 ug/l	79.37	0.39	
66 Zn	0.01 ug/l	415.60	6.90	
75 As	0.01 ug/l	79.94	0.27	
78 Se	0.01 ug/l	150.15	0.30	
78 Se	0.08 ug/l	243.02	0.30	
88 Sr	0.00 ug/l	251.58	0.03	
88 Sr	0.00 ug/l	18.53	0.03	
95 Mo	0.05 ug/l	9.94	0.21	
106 (Cd)	----- ug/l	-----	#####	
107 Ag	0.00 ug/l	100.46	0.09	
108 (Cd)	----- ug/l	-----	#####	
111 Cd	0.01 ug/l	54.35	0.06	
118 Sn	0.02 ug/l	75.03	#####	
118 Sn	0.02 ug/l	36.24	#####	
118 Sn	0.02 ug/l	34.99	0.30	
121 Sb	0.03 ug/l	9.49	0.03	
137 Ba	0.00 ug/l	110.58	0.12	
205 Tl	0.01 ug/l	34.92	0.03	
206 (Pb)	----- ug/l	-----	#####	
207 (Pb)	----- ug/l	-----	#####	
208 Pb	-0.01 ug/l	13.26	0.33	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-42072.89	9.05	-41328.95	101.8	70 - 120	IS Fai
45 Sc	3095275.50	0.67	3008024.30	102.9	70 - 120	
45 Sc	425403.59	1.36	423303.94	100.5	70 - 120	
45 Sc	8639370.00	1.70	8607281.00	100.4	70 - 120	
72 Ge	799887.75	0.92	774468.63	103.3	70 - 120	
72 Ge	283411.25	0.73	282128.91	100.5	70 - 120	
72 Ge	1897265.50	1.53	1882554.90	100.8	70 - 120	
115 In	5586231.50	0.97	5556751.00	100.5	70 - 120	
115 In	3015473.30	1.15	3029632.80	99.5	70 - 120	
115 In	12146847.00	1.16	12097256.00	100.4	70 - 120	
159 Tb	16155302.00	0.85	16269544.00	99.3	70 - 120	
165 Ho	15737558.00	0.71	15819307.00	99.5	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

0 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Pass
 ISTD: Fail

CCV QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\012_CCV.D\012_CCV.D#
 Date Acquired: Jul 23 2012 12:05 pm
 Operator: NBS
 Sample Name: CCV 120723
 Misc Info:
 Vial Number: 1105
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: CCV
 Total Dil Factor: 1.00

QC Elements

Element	Conc.	RSD(%)	Expected	QC Range(%)	Flag
7 (Li)	----- ug/l	-----	50.00	90 - 110	
9 Be	49.60 ug/l	0.79	50.00	90 - 110	
11 B	51.10 ug/l	0.98	50.00	90 - 110	
23 Na	1238.00 ug/l	0.52	1250.00	90 - 110	
24 Mg	2526.00 ug/l	0.49	2500.00	90 - 110	
27 Al	990.40 ug/l	1.09	1000.00	90 - 110	
39 K	995.20 ug/l	0.67	1000.00	90 - 110	
44 Ca	2466.00 ug/l	0.42	2500.00	90 - 110	
47 Ti	48.94 ug/l	2.27	50.00	90 - 110	
51 V	49.00 ug/l	0.21	50.00	90 - 110	
52 Cr	49.04 ug/l	0.15	50.00	90 - 110	
55 Mn	48.94 ug/l	0.22	50.00	90 - 110	
56 Fe	993.80 ug/l	0.50	1000.00	90 - 110	
59 Co	49.21 ug/l	0.92	50.00	90 - 110	
60 Ni	49.66 ug/l	0.93	50.00	90 - 110	
63 Cu	49.23 ug/l	0.47	50.00	90 - 110	
65 Cu	48.96 ug/l	0.26	50.00	90 - 110	
66 Zn	49.71 ug/l	0.73	50.00	90 - 110	
75 As	49.67 ug/l	0.20	50.00	90 - 110	
78 Se	50.01 ug/l	1.21	50.00	90 - 110	
78 Se	49.30 ug/l	1.29	50.00	90 - 110	
88 Sr	50.00 ug/l	0.67	50.00	90 - 110	
88 Sr	49.73 ug/l	0.36	50.00	90 - 110	
95 Mo	49.04 ug/l	0.81	50.00	90 - 110	
106 (Cd)	----- ug/l	-----	50.00	90 - 110	
107 Ag	24.88 ug/l	1.17	25.00	90 - 110	
108 (Cd)	----- ug/l	-----	50.00	90 - 110	
111 Cd	49.56 ug/l	1.10	50.00	90 - 110	
118 Sn	49.77 ug/l	1.17	---	##### - #####	
118 Sn	49.48 ug/l	0.73	---	##### - #####	
118 Sn	49.50 ug/l	0.78	50.00	90 - 110	
121 Sb	52.78 ug/l	0.89	50.00	90 - 110	
137 Ba	49.50 ug/l	0.92	50.00	90 - 110	
205 Tl	51.14 ug/l	0.85	50.00	90 - 110	
206 (Pb)	----- ug/l	-----	50.00	90 - 110	
207 (Pb)	----- ug/l	-----	50.00	90 - 110	
208 Pb	52.39 ug/l	0.44	50.00	90 - 110	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-36962.72	9.13	-41328.95	89.4	70 - 120	IS Fail
45 Sc	3048428.00	0.52	3008024.30	101.3	70 - 120	
45 Sc	425742.41	0.55	423303.94	100.6	70 - 120	
45 Sc	8837992.00	0.79	8607281.00	102.7	70 - 120	
72 Ge	786063.00	1.03	774468.63	101.5	70 - 120	
72 Ge	283188.31	1.08	282128.91	100.4	70 - 120	
72 Ge	1919384.60	0.76	1882554.90	102.0	70 - 120	
115 In	5483121.50	0.63	5556751.00	98.7	70 - 120	
115 In	2961335.00	0.42	3029632.80	97.7	70 - 120	
115 In	12216251.00	0.54	12097256.00	101.0	70 - 120	
159 Tb	16140742.00	0.69	16269544.00	99.2	70 - 120	
165 Ho	15761536.00	0.35	15819307.00	99.6	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

0 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Pass
 ISTD: Fail

CCB QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\013_CCB.D\013_CCB.D#
 Date Acquired: Jul 23 2012 12:12 pm
 Operator: NBS
 Sample Name: CCB 120723
 Misc Info:
 Vial Number: 1102
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: CCB
 Total Dil Factor: 1.00

QC Elements	Element	Conc.	RSD(%)	High Limit	Flag
	7 (Li)	----- ug/l	-----	#####	
	9 Be	0.01 ug/l	28.98	0.12	
	11 B	-0.01 ug/l	2028.40	15.00	
	23 Na	-9.93 ug/l	20.37	77.10	
	24 Mg	0.41 ug/l	44.14	7.50	
	27 Al	0.58 ug/l	23.47	3.96	
	39 K	2.52 ug/l	104.72	19.20	
	44 Ca	-7.69 ug/l	34.34	90.00	
	47 Ti	0.02 ug/l	43.51	0.78	
	51 V	0.01 ug/l	46.98	0.21	
	52 Cr	-0.02 ug/l	32.90	0.12	
	55 Mn	-0.01 ug/l	25.20	0.18	
	56 Fe	0.13 ug/l	28.25	40.80	
	59 Co	0.02 ug/l	14.80	0.09	
	60 Ni	-0.01 ug/l	143.54	0.48	
	63 Cu	-0.02 ug/l	29.44	0.39	
	65 Cu	-0.02 ug/l	45.72	0.39	
	66 Zn	0.03 ug/l	109.32	6.90	
	75 As	0.00 ug/l	13.81	0.27	
	78 Se	0.00 ug/l	1935.10	0.30	
	78 Se	0.36 ug/l	43.44	0.30	Fail
	88 Sr	0.02 ug/l	61.41	0.03	
	88 Sr	0.01 ug/l	24.52	0.03	
	95 Mo	0.08 ug/l	10.75	0.21	
	106 (Cd)	----- ug/l	-----	#####	
	107 Ag	0.00 ug/l	33.45	0.09	
	108 (Cd)	----- ug/l	-----	#####	
	111 Cd	0.01 ug/l	112.87	0.06	
	118 Sn	0.08 ug/l	13.90	#####	
	118 Sn	0.08 ug/l	11.98	#####	
	118 Sn	0.04 ug/l	14.86	0.30	
	121 Sb	0.09 ug/l	3.60	0.03	Fail
	137 Ba	0.01 ug/l	23.15	0.12	
	205 Tl	0.01 ug/l	16.07	0.03	
	206 (Pb)	----- ug/l	-----	#####	
	207 (Pb)	----- ug/l	-----	#####	
	208 Pb	-0.01 ug/l	14.05	0.33	

ISTD Elements	Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
	6 Li	-41365.11	4.11	-41328.95	100.1	70 - 120	IS Fai
	45 Sc	3109463.00	0.74	3008024.30	103.4	70 - 120	
	45 Sc	436584.66	0.71	423303.94	103.1	70 - 120	
	45 Sc	8702714.00	0.34	8607281.00	101.1	70 - 120	
	72 Ge	794512.25	1.02	774468.63	102.6	70 - 120	
	72 Ge	289162.22	1.20	282128.91	102.5	70 - 120	
	72 Ge	1910565.90	0.18	1882554.90	101.5	70 - 120	
	115 In	5651735.50	0.41	5556751.00	101.7	70 - 120	
	115 In	3029258.30	0.13	3029632.80	100.0	70 - 120	
	115 In	12209930.00	0.36	12097256.00	100.9	70 - 120	
	159 Tb	16410115.00	0.53	16269544.00	100.9	70 - 120	
	165 Ho	15792125.00	0.38	15819307.00	99.8	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

2 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Fail
 ISTD: Fail

Sample QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\014SMPL.D\014SMPL.D#
 Date Acquired: Jul 23 2012 12:18 pm
 Operator: NBS
 Sample Name: ICSA 120723
 Misc Info:
 Vial Number: 2102
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: Sample
 Prep Dil Factor: 1.00
 Total Dil Factor: 1.00

QC Elements

Element	Conc.	Corr. Conc.	RSD(%)	High Limit	Flag
7 (Li)	----- ug/l	#VALUE!	-----	0	
9 Be	0.01 ug/l	0.01	16.22	1000	
11 B	1.24 ug/l	1.24	1.37	1000	
23 Na	89610.00 ug/l	89610.00	0.85	25000	>Cal
24 Mg	89070.00 ug/l	89070.00	0.41	50000	>Cal
27 Al	88750.00 ug/l	88750.00	0.72	20000	>Cal
39 K	87880.00 ug/l	87880.00	0.71	20000	>Cal
44 Ca	92120.00 ug/l	92120.00	0.78	50000	>Cal
47 Ti	1718.00 ug/l	1718.00	0.37	1000	>Cal
51 V	0.10 ug/l	0.10	8.87	1000	
52 Cr	1.58 ug/l	1.58	15.00	1000	
55 Mn	5.76 ug/l	5.76	1.34	1000	
56 Fe	90820.00 ug/l	90820.00	0.70	20000	>Cal
59 Co	1.94 ug/l	1.94	1.83	1000	
60 Ni	1.97 ug/l	1.97	2.48	1000	
63 Cu	0.75 ug/l	0.75	1.21	1000	
65 Cu	0.79 ug/l	0.79	3.08	1000	
66 Zn	1.40 ug/l	1.40	1.13	1000	
75 As	0.28 ug/l	0.28	2.03	1000	
78 Se	0.13 ug/l	0.13	8.50	1000	
78 Se	0.77 ug/l	0.77	9.40	1000	
88 Sr	1.32 ug/l	1.32	5.56	1000	
88 Sr	1.37 ug/l	1.37	1.25	1000	
95 Mo	1875.00 ug/l	1875.00	1.20	1000	>Cal
106 (Cd)	----- ug/l	#VALUE!	-----	#####	
107 Ag	0.07 ug/l	0.07	5.89	500	
108 (Cd)	----- ug/l	#VALUE!	-----	#####	
111 Cd	0.88 ug/l	0.88	1.42	1000	
118 Sn	0.21 ug/l	0.21	1.97	#####	
118 Sn	0.22 ug/l	0.22	18.96	#####	
118 Sn	0.23 ug/l	0.23	1.31	1000	
121 Sb	1.18 ug/l	1.18	1.23	1000	
137 Ba	2.58 ug/l	2.58	5.35	1000	
205 Tl	0.08 ug/l	0.08	2.29	1000	
206 (Pb)	----- ug/l	#VALUE!	-----	#####	
207 (Pb)	----- ug/l	#VALUE!	-----	#####	
208 Pb	0.41 ug/l	0.41	1.41	1000	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-37598.05	8.88	-41328.95	91.0	70 - 120	IS Fai
45 Sc	3054391.50	0.51	3008024.30	101.5	70 - 120	
45 Sc	436433.34	0.64	423303.94	103.1	70 - 120	
45 Sc	8890298.00	0.21	8607281.00	103.3	70 - 120	
72 Ge	769231.75	0.38	774468.63	99.3	70 - 120	
72 Ge	281075.50	0.23	282128.91	99.6	70 - 120	
72 Ge	1923643.30	0.95	1882554.90	102.2	70 - 120	
115 In	5221821.50	1.14	5556751.00	94.0	70 - 120	
115 In	2776903.00	0.79	3029632.80	91.7	70 - 120	
115 In	11448044.00	0.61	12097256.00	94.6	70 - 120	
159 Tb	16053537.00	0.92	16269544.00	98.7	70 - 120	
165 Ho	15595900.00	0.15	15819307.00	98.6	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

8 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Fail
 ISTD: Fail

ICS-AB QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\015ICSB.D\015ICSB.D#
 Date Acquired: Jul 23 2012 12:25 pm
 Acq. Method: 62A0723A.M
 Operator: NBS
 Sample Name: ICSAB 120723
 Misc Info:
 Vial Number: 2103
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal. Update: Jul 23 2012 11:42 am
 Sample Type: ICSAB
 Dilution Factor: 1.00

Data Results:
 Analytes: Pass
 ISTD: Fail

QC Elements

Element	IS Ref	Tune	Conc. ppb	RSD(%)	Expected	%Recovery	QC Range(%)	Flag
7 (Li)	---	3	-----	---	---	---	---	---
9 Be	45	3	241.10	0.75	250	96.4	80 - 120	
11 B	45	3	0.86	13.58	---	---	---	
23 Na	45	2	90140.00	1.67	---	---	---	
24 Mg	45	2	89040.00	1.39	---	---	---	
27 Al	45	2	88440.00	0.98	---	---	---	
39 K	45	2	88130.00	1.12	---	---	---	
44 Ca	45	2	91870.00	1.53	---	---	---	
47 Ti	45	2	1696.00	1.02	2000	84.8	80 - 120	
51 V	45	2	250.80	1.85	250	100.3	80 - 120	
52 Cr	45	2	238.10	1.41	250	95.2	80 - 120	
55 Mn	45	2	243.80	1.56	250	97.5	80 - 120	
56 Fe	45	2	89920.00	1.53	---	---	---	
59 Co	45	2	212.60	1.90	250	85.0	80 - 120	
60 Ni	45	2	449.60	1.55	500	89.9	80 - 120	
63 Cu	45	2	219.60	1.42	250	87.8	80 - 120	
65 Cu	45	2	219.10	1.07	250	87.6	80 - 120	
66 Zn	115	2	488.00	0.51	500	97.6	80 - 120	
75 As	115	2	263.10	0.67	250	105.2	80 - 120	
78 Se	115	1	254.90	1.16	250	102.0	80 - 120	
78 Se	115	2	253.80	0.63	250	101.5	80 - 120	
88 Sr	115	2	1.28	2.80	---	---	---	
88 Sr	115	3	1.41	0.61	---	---	---	
95 Mo	115	3	2133.00	1.20	2000	106.7	80 - 120	
106 (Cd)	---	3	-----	---	---	---	---	
107 Ag	115	3	479.00	7.77	500	95.8	80 - 120	
108 (Cd)	---	3	-----	---	---	---	---	
111 Cd	115	3	450.60	0.21	500	90.1	80 - 120	
118 Sn	115	1	0.21	7.95	---	---	---	
118 Sn	115	2	0.22	9.45	---	---	---	
118 Sn	115	3	0.23	3.94	---	---	---	
121 Sb	115	3	249.30	1.10	250	99.7	80 - 120	
137 Ba	115	3	241.40	0.52	250	96.6	80 - 120	
205 Tl	159	3	227.20	0.05	250	90.9	80 - 120	
206 (Pb)	---	3	-----	---	---	---	---	
207 (Pb)	---	3	-----	---	---	---	---	
208 Pb	159	3	437.60	0.77	500	87.5	80 - 120	

ISTD Elements

Element	Tune	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	3	-35443	14.51	-41329	85.8	70 - 120	IS Fail
45 Sc	1	3173175	2.09	3008024	105.5	70 - 120	
45 Sc	2	445878	1.47	423304	105.3	70 - 120	
45 Sc	3	8959869	1.06	8607281	104.1	70 - 120	
72 Ge	1	787580	1.18	774469	101.7	70 - 120	
72 Ge	2	286166	0.66	282129	101.4	70 - 120	
72 Ge	3	1957866	0.62	1882555	104.0	70 - 120	
115 In	1	5428667	1.26	5556751	97.7	70 - 120	
115 In	2	2857644	0.53	3029633	94.3	70 - 120	
115 In	3	11548093	0.63	12097256	95.5	70 - 120	
159 Tb	3	15976363	0.50	16269544	98.2	70 - 120	
165 Ho	3	15533992	0.96	15819307	98.2	70 - 120	

Tune File# 1 c:\icpchem\1\7500\h2_hmi.u
 Tune File# 2 c:\icpchem\1\7500\he_hmi.u
 Tune File# 3 c:\icpchem\1\7500\ng_hmi.u

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

0 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

CCV QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\024_CCV.D\024_CCV.D#
 Date Acquired: Jul 23 2012 01:32 pm
 Operator: NBS
 Sample Name: CCV 120723
 Misc Info:
 Vial Number: 1105
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: CCV
 Total Dil Factor: 1.00

QC Elements

Element	Conc.	RSD(%)	Expected	QC Range(%)	Flag
7 (Li)	----- ug/l	-----	50.00	90 - 110	
9 Be	49.44 ug/l	0.67	50.00	90 - 110	
11 B	49.03 ug/l	0.99	50.00	90 - 110	
23 Na	1218.00 ug/l	2.36	1250.00	90 - 110	
24 Mg	2498.00 ug/l	1.95	2500.00	90 - 110	
27 Al	977.70 ug/l	2.12	1000.00	90 - 110	
39 K	985.10 ug/l	2.42	1000.00	90 - 110	
44 Ca	2430.00 ug/l	1.38	2500.00	90 - 110	
47 Ti	47.89 ug/l	1.44	50.00	90 - 110	
51 V	48.79 ug/l	2.43	50.00	90 - 110	
52 Cr	49.04 ug/l	1.97	50.00	90 - 110	
55 Mn	48.80 ug/l	1.39	50.00	90 - 110	
56 Fe	973.20 ug/l	2.18	1000.00	90 - 110	
59 Co	48.59 ug/l	1.75	50.00	90 - 110	
60 Ni	48.69 ug/l	1.92	50.00	90 - 110	
63 Cu	48.38 ug/l	2.01	50.00	90 - 110	
65 Cu	48.32 ug/l	1.36	50.00	90 - 110	
66 Zn	50.19 ug/l	0.70	50.00	90 - 110	
75 As	49.86 ug/l	0.81	50.00	90 - 110	
78 Se	49.64 ug/l	1.15	50.00	90 - 110	
78 Se	50.19 ug/l	1.35	50.00	90 - 110	
88 Sr	50.53 ug/l	0.83	50.00	90 - 110	
88 Sr	50.45 ug/l	0.55	50.00	90 - 110	
95 Mo	50.00 ug/l	0.20	50.00	90 - 110	
106 (Cd)	----- ug/l	-----	50.00	90 - 110	
107 Ag	24.84 ug/l	0.45	25.00	90 - 110	
108 (Cd)	----- ug/l	-----	50.00	90 - 110	
111 Cd	49.60 ug/l	0.95	50.00	90 - 110	
118 Sn	49.74 ug/l	1.06	---	##### - #####	
118 Sn	49.48 ug/l	1.45	---	##### - #####	
118 Sn	49.65 ug/l	0.60	50.00	90 - 110	
121 Sb	53.03 ug/l	0.38	50.00	90 - 110	
137 Ba	49.68 ug/l	1.02	50.00	90 - 110	
205 Tl	50.98 ug/l	0.55	50.00	90 - 110	
206 (Pb)	----- ug/l	-----	50.00	90 - 110	
207 (Pb)	----- ug/l	-----	50.00	90 - 110	
208 Pb	51.97 ug/l	0.08	50.00	90 - 110	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-37208.29	23.54	-41328.95	90.0	70 - 120	IS Fail
45 Sc	3207763.50	0.47	3008024.30	106.6	70 - 120	
45 Sc	451396.47	1.87	423303.94	106.6	70 - 120	
45 Sc	9114044.00	0.70	8607281.00	105.9	70 - 120	
72 Ge	820047.81	0.42	774468.63	105.9	70 - 120	
72 Ge	296744.84	1.32	282128.91	105.2	70 - 120	
72 Ge	1965454.80	0.50	1882554.90	104.4	70 - 120	
115 In	5736711.50	0.79	5556751.00	103.2	70 - 120	
115 In	3076324.30	1.59	3029632.80	101.5	70 - 120	
115 In	12421903.00	0.30	12097256.00	102.7	70 - 120	
159 Tb	16660003.00	0.38	16269544.00	102.4	70 - 120	
165 Ho	16160466.00	0.80	15819307.00	102.2	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

0 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: **Pass**
ISTD: **Fail**

CCB QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\026_CCB.D\026_CCB.D#
 Date Acquired: Jul 23 2012 01:45 pm
 Operator: NBS
 Sample Name: CCB 120723
 Misc Info:
 Vial Number: 1102
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: CCB
 Total Dil Factor: 1.00

QC Elements

Element	Conc.	RSD(%)	High Limit	Flag
7 (Li)	----- ug/l	-----	#####	
9 Be	0.00 ug/l	68.10	0.12	
11 B	-1.57 ug/l	8.40	15.00	
23 Na	-19.39 ug/l	9.24	77.10	
24 Mg	-0.11 ug/l	32.48	7.50	
27 Al	0.65 ug/l	37.42	3.96	
39 K	-0.55 ug/l	422.51	19.20	
44 Ca	-13.14 ug/l	10.64	90.00	
47 Ti	0.05 ug/l	190.52	0.78	
51 V	0.00 ug/l	97.27	0.21	
52 Cr	-0.06 ug/l	3.59	0.12	
55 Mn	-0.02 ug/l	39.89	0.18	
56 Fe	0.03 ug/l	92.73	40.80	
59 Co	-0.01 ug/l	5.06	0.09	
60 Ni	-0.05 ug/l	20.94	0.48	
63 Cu	-0.03 ug/l	12.14	0.39	
65 Cu	-0.03 ug/l	27.56	0.39	
66 Zn	0.01 ug/l	522.05	6.90	
75 As	0.00 ug/l	201.72	0.27	
78 Se	-0.02 ug/l	28.35	0.30	
78 Se	0.30 ug/l	54.24	0.30	
88 Sr	0.01 ug/l	28.85	0.03	
88 Sr	0.00 ug/l	46.66	0.03	
95 Mo	0.04 ug/l	11.71	0.21	
106 (Cd)	----- ug/l	-----	#####	
107 Ag	0.01 ug/l	25.03	0.09	
108 (Cd)	----- ug/l	-----	#####	
111 Cd	0.00 ug/l	126.59	0.06	
118 Sn	0.02 ug/l	23.26	#####	
118 Sn	0.01 ug/l	84.96	#####	
118 Sn	0.01 ug/l	36.38	0.30	
121 Sb	0.03 ug/l	13.44	0.03	Fail
137 Ba	0.01 ug/l	47.38	0.12	
205 Tl	0.01 ug/l	3.62	0.03	
206 (Pb)	----- ug/l	-----	#####	
207 (Pb)	----- ug/l	-----	#####	
208 Pb	-0.02 ug/l	7.77	0.33	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec (%)	QC Range (%)	Flag
6 Li	-46601.30	1.93	-41328.95	112.8	70 - 120	IS Fai
45 Sc	3258366.00	1.08	3008024.30	108.3	70 - 120	
45 Sc	452200.25	0.37	423303.94	106.8	70 - 120	
45 Sc	8993459.00	0.85	8607281.00	104.5	70 - 120	
72 Ge	834886.63	1.35	774468.63	107.8	70 - 120	
72 Ge	300371.97	0.87	282128.91	106.5	70 - 120	
72 Ge	1964668.50	0.86	1882554.90	104.4	70 - 120	
115 In	5902867.00	0.62	5556751.00	106.2	70 - 120	
115 In	3136632.50	0.75	3029632.80	103.5	70 - 120	
115 In	12490374.00	0.82	12097256.00	103.2	70 - 120	
159 Tb	16903932.00	1.12	16269544.00	103.9	70 - 120	
165 Ho	16315825.00	1.13	15819307.00	103.1	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

1 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Fail
 ISTD: Fail

CCV QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\038_CCV.D\038_CCV.D#
 Date Acquired: Jul 23 2012 03:07 pm
 Operator: NBS
 Sample Name: CCV 120723
 Misc Info:
 Vial Number: 1105
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: CCV
 Total Dil Factor: 1.00

QC Elements

Element	Conc.	RSD(%)	Expected	QC Range(%)	Flag
7 (Li)	----- ug/l	-----	50.00	90 - 110	
9 Be	50.82 ug/l	0.70	50.00	90 - 110	
11 B	60.80 ug/l	0.41	50.00	90 - 110	Fail
23 Na	1289.00 ug/l	0.40	1250.00	90 - 110	
24 Mg	2570.00 ug/l	0.63	2500.00	90 - 110	
27 Al	1010.00 ug/l	0.91	1000.00	90 - 110	
39 K	1010.00 ug/l	0.97	1000.00	90 - 110	
44 Ca	2527.00 ug/l	0.75	2500.00	90 - 110	
47 Ti	48.23 ug/l	2.50	50.00	90 - 110	
51 V	48.21 ug/l	1.09	50.00	90 - 110	
52 Cr	47.74 ug/l	1.05	50.00	90 - 110	
55 Mn	48.54 ug/l	0.37	50.00	90 - 110	
56 Fe	970.00 ug/l	0.78	1000.00	90 - 110	
59 Co	47.62 ug/l	0.81	50.00	90 - 110	
60 Ni	47.79 ug/l	0.59	50.00	90 - 110	
63 Cu	47.35 ug/l	1.79	50.00	90 - 110	
65 Cu	47.08 ug/l	0.91	50.00	90 - 110	
66 Zn	51.38 ug/l	0.35	50.00	90 - 110	
75 As	51.04 ug/l	0.81	50.00	90 - 110	
78 Se	50.77 ug/l	0.44	50.00	90 - 110	
78 Se	51.66 ug/l	0.32	50.00	90 - 110	
88 Sr	51.25 ug/l	1.09	50.00	90 - 110	
88 Sr	50.81 ug/l	1.01	50.00	90 - 110	
95 Mo	50.10 ug/l	0.93	50.00	90 - 110	
106 (Cd)	----- ug/l	-----	50.00	90 - 110	
107 Ag	24.99 ug/l	1.82	25.00	90 - 110	
108 (Cd)	----- ug/l	-----	50.00	90 - 110	
111 Cd	49.92 ug/l	1.71	50.00	90 - 110	
118 Sn	49.95 ug/l	0.31	---	##### - #####	
118 Sn	49.60 ug/l	0.56	---	##### - #####	
118 Sn	50.19 ug/l	1.41	50.00	90 - 110	
121 Sb	53.47 ug/l	2.63	50.00	90 - 110	
137 Ba	49.86 ug/l	1.62	50.00	90 - 110	
205 Tl	50.30 ug/l	0.77	50.00	90 - 110	
206 (Pb)	----- ug/l	-----	50.00	90 - 110	
207 (Pb)	----- ug/l	-----	50.00	90 - 110	
208 Pb	51.71 ug/l	0.38	50.00	90 - 110	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-46384.73	10.88	-41328.95	112.2	70 - 120	IS Fail
45 Sc	3148572.80	0.41	3008024.30	104.7	70 - 120	
45 Sc	455889.47	1.12	423303.94	107.7	70 - 120	
45 Sc	9181851.00	0.79	8607281.00	106.7	70 - 120	
72 Ge	809349.19	0.46	774468.63	104.5	70 - 120	
72 Ge	291443.75	0.37	282128.91	103.3	70 - 120	
72 Ge	1986232.80	0.29	1882554.90	105.5	70 - 120	
115 In	5566992.00	0.59	5556751.00	100.2	70 - 120	
115 In	2998159.00	0.95	3029632.80	99.0	70 - 120	
115 In	12443523.00	1.27	12097256.00	102.9	70 - 120	
159 Tb	16734999.00	0.75	16269544.00	102.9	70 - 120	
165 Ho	16223907.00	0.28	15819307.00	102.6	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

1 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Fail
 ISTD: Fail

CCB QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\040_CCB.D\040_CCB.D#
 Date Acquired: Jul 23 2012 03:20 pm
 Operator: NBS
 Sample Name: CCB 120723
 Misc Info:
 Vial Number: 1102
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: CCB
 Total Dil Factor: 1.00

QC Elements

Element	Conc.	RSD(%)	High Limit	Flag
7 (Li)	----- ug/l	-----	#####	
9 Be	0.00 ug/l	628.28	0.12	
11 B	5.64 ug/l	2.47	15.00	
23 Na	6.28 ug/l	27.52	77.10	
24 Mg	0.40 ug/l	37.54	7.50	
27 Al	1.14 ug/l	18.54	3.96	
39 K	8.76 ug/l	45.49	19.20	
44 Ca	-12.54 ug/l	13.15	90.00	
47 Ti	0.02 ug/l	59.79	0.78	
51 V	0.01 ug/l	43.11	0.21	
52 Cr	-0.06 ug/l	3.00	0.12	
55 Mn	-0.03 ug/l	20.84	0.18	
56 Fe	0.03 ug/l	109.87	40.80	
59 Co	-0.01 ug/l	52.61	0.09	
60 Ni	-0.05 ug/l	20.44	0.48	
63 Cu	-0.03 ug/l	13.37	0.39	
65 Cu	-0.02 ug/l	31.00	0.39	
66 Zn	0.05 ug/l	67.78	6.90	
75 As	0.00 ug/l	140.00	0.27	
78 Se	-0.02 ug/l	104.66	0.30	
78 Se	0.46 ug/l	14.69	0.30	Fail
88 Sr	0.02 ug/l	51.01	0.03	
88 Sr	0.01 ug/l	6.84	0.03	
95 Mo	0.03 ug/l	27.08	0.21	
106 (Cd)	----- ug/l	-----	#####	
107 Ag	0.00 ug/l	53.78	0.09	
108 (Cd)	----- ug/l	-----	#####	
111 Cd	0.01 ug/l	63.57	0.06	
118 Sn	0.01 ug/l	44.20	#####	
118 Sn	0.02 ug/l	55.29	#####	
118 Sn	0.01 ug/l	8.30	0.30	
121 Sb	0.03 ug/l	16.15	0.03	Fail
137 Ba	0.01 ug/l	76.73	0.12	
205 Tl	0.01 ug/l	6.79	0.03	
206 (Pb)	----- ug/l	-----	#####	
207 (Pb)	----- ug/l	-----	#####	
208 Pb	-0.02 ug/l	2.57	0.33	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-45872.62	15.67	-41328.95	111.0	70 - 120	IS Fail
45 Sc	3225425.30	0.94	3008024.30	107.2	70 - 120	
45 Sc	455256.84	0.90	423303.94	107.5	70 - 120	
45 Sc	9025179.00	0.84	8607281.00	104.9	70 - 120	
72 Ge	829819.56	0.95	774468.63	107.1	70 - 120	
72 Ge	294491.41	1.41	282128.91	104.4	70 - 120	
72 Ge	1986792.00	0.34	1882554.90	105.5	70 - 120	
115 In	5757836.50	0.31	5556751.00	103.6	70 - 120	
115 In	3079085.80	1.03	3029632.80	101.6	70 - 120	
115 In	12604958.00	0.38	12097256.00	104.2	70 - 120	
159 Tb	16666065.00	1.06	16269544.00	102.4	70 - 120	
165 Ho	16233445.00	0.44	15819307.00	102.6	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

2 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Fail
 ISTD: Fail

**METALS
Raw Data**

APPL, INC.

METALS BLANK

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Analyte	Result	LOQ	LOD	DL	Units	Prep Date	Analysis Date	QC Group
6020	LEAD (PB) (DISSOL	0.22 U	0.5	0.22	0.11	ug/L	07/23/12	07/23/12	#602D-120723A-AY65167

Sample QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\019SMPL.D\019SMPL.D#
 Date Acquired: Jul 23 2012 12:58 pm
 Operator: NBS
 Sample Name: 120723A-3015-BLK
 Misc Info: 120723A-3015
 Vial Number: 3101
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: Sample
 Prep Dil Factor: 1.11
 Total Dil Factor: 1.11

QC Elements

Element	Conc.	Corr. Conc.	RSD(%)	High Limit	Flag
7 (Li)	----- ug/l	#VALUE!	-----	0	
9 Be	0.00 ug/l	0.00	101.03	1000	
11 B	-0.54 ug/l	-0.60	23.92	1000	
23 Na	-9.48 ug/l	-10.53	24.91	25000	
24 Mg	0.68 ug/l	0.75	15.39	50000	
27 Al	0.72 ug/l	0.80	27.58	20000	
39 K	3.46 ug/l	3.84	104.28	20000	
44 Ca	-9.56 ug/l	-10.62	23.29	50000	
47 Ti	0.11 ug/l	0.13	61.04	1000	
51 V	0.00 ug/l	0.00	187.34	1000	
52 Cr	0.09 ug/l	0.10	2.59	1000	
55 Mn	0.03 ug/l	0.04	24.16	1000	
56 Fe	2.25 ug/l	2.50	6.31	20000	
59 Co	0.25 ug/l	0.28	8.41	1000	
60 Ni	-0.03 ug/l	-0.03	21.25	1000	
63 Cu	0.06 ug/l	0.06	14.62	1000	
65 Cu	0.05 ug/l	0.05	41.40	1000	
66 Zn	0.25 ug/l	0.28	16.07	1000	
75 As	0.01 ug/l	0.01	13.09	1000	
78 Se	0.01 ug/l	0.01	54.14	1000	
78 Se	0.64 ug/l	0.71	8.48	1000	
88 Sr	0.00 ug/l	0.00	164.49	1000	
88 Sr	0.00 ug/l	0.00	197.18	1000	
95 Mo	0.23 ug/l	0.26	3.93	1000	
106 (Cd)	----- ug/l	#VALUE!	-----	#####	
107 Ag	0.82 ug/l	0.91	7.30	500	
108 (Cd)	----- ug/l	#VALUE!	-----	#####	
111 Cd	0.04 ug/l	0.04	6.12	1000	
118 Sn	0.13 ug/l	0.14	7.71	#####	
118 Sn	0.13 ug/l	0.14	3.28	#####	
118 Sn	0.11 ug/l	0.12	9.34	1000	
121 Sb	0.13 ug/l	0.14	5.45	1000	
137 Ba	0.00 ug/l	0.00	121.47	1000	
205 Tl	0.08 ug/l	0.09	3.11	1000	
206 (Pb)	----- ug/l	#VALUE!	-----	#####	
207 (Pb)	----- ug/l	#VALUE!	-----	#####	
208 Pb	-0.02 ug/l	-0.03	3.12	1000	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-41515.17	17.03	-41328.95	100.5	70 - 120	IS Fai
45 Sc	3134014.50	1.22	3008024.30	104.2	70 - 120	
45 Sc	431622.16	0.76	423303.94	102.0	70 - 120	
45 Sc	9041910.00	1.00	8607281.00	105.0	70 - 120	
72 Ge	790204.25	0.87	774468.63	102.0	70 - 120	
72 Ge	281787.47	2.04	282128.91	99.9	70 - 120	
72 Ge	1936014.10	0.17	1882554.90	102.8	70 - 120	
115 In	5556259.00	0.49	5556751.00	100.0	70 - 120	
115 In	2981638.30	0.15	3029632.80	98.4	70 - 120	
115 In	12374886.00	0.94	12097256.00	102.3	70 - 120	
159 Tb	16686016.00	0.57	16269544.00	102.6	70 - 120	
165 Ho	16112874.00	0.15	15819307.00	101.9	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#

0 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Pass
 ISTD: Fail

Laboratory Control Spike Recovery

METALS

APPL Inc.
908 North Temperance Avenue
Clovis, CA 93611

Method	Compound Name	Spike Level ug/L	SPK Result ug/L	SPK % Recovery	Recovery Limits	Extract Date	Analysis Date	QC Group
6020	LEAD (PB) (DISSOLVED)	50.0	45.5	91.0	80-120	07/23/12	07/23/12	#602D-120723A-AY65167

Comments:

Sample QC Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\021SMPL.D\021SMPL.D#
 Date Acquired: Jul 23 2012 01:11 pm
 Operator: NBS
 Sample Name: 120723A-3015-LCS
 Misc Info: 120723A-3015
 Vial Number: 3103
 Current Method: C:\ICPCHEM\1\METHODS\62A0723A.M
 Calibration File: C:\ICPCHEM\1\CALIB\62A0723A.C
 Last Cal Update: Jul 23 2012 11:42 am
 Sample Type: Sample
 Prep Dil Factor: 1.11
 Total Dil Factor: 1.11

QC Elements

Element	Conc.	Corr. Conc.	RSD(%)	High Limit	Flag
7 (Li)	ug/l	#VALUE!	-----	0	
9 Be	7.49 ug/l	8.32	1.00	1000	
11 B	37.46 ug/l	41.62	1.36	1000	
23 Na	3667.00 ug/l	4074.04	1.15	25000	
24 Mg	3760.00 ug/l	4177.36	1.30	50000	
27 Al	307.00 ug/l	341.08	1.70	20000	
39 K	767.90 ug/l	853.14	1.74	20000	
44 Ca	3952.00 ug/l	4390.67	2.20	50000	
47 Ti	47.17 ug/l	52.41	2.99	1000	
51 V	39.18 ug/l	43.53	1.78	1000	
52 Cr	38.95 ug/l	43.27	1.71	1000	
55 Mn	39.16 ug/l	43.51	1.76	1000	
56 Fe	172.60 ug/l	191.76	1.33	20000	
59 Co	38.06 ug/l	42.28	1.55	1000	
60 Ni	38.05 ug/l	42.27	1.84	1000	
63 Cu	37.17 ug/l	41.30	1.89	1000	
65 Cu	37.09 ug/l	41.21	2.15	1000	
66 Zn	75.57 ug/l	83.96	0.44	1000	
75 As	36.26 ug/l	40.28	0.94	1000	
78 Se	33.81 ug/l	37.56	0.78	1000	
78 Se	34.37 ug/l	38.19	1.94	1000	
88 Sr	39.64 ug/l	44.04	0.91	1000	
88 Sr	39.14 ug/l	43.48	1.36	1000	
95 Mo	46.06 ug/l	51.17	1.69	1000	
106 (Cd)	ug/l	#VALUE!	-----	#####	
107 Ag	17.93 ug/l	19.92	1.04	500	
108 (Cd)	ug/l	#VALUE!	-----	#####	
111 Cd	7.41 ug/l	8.23	1.74	1000	
118 Sn	47.18 ug/l	52.42	0.07	#####	
118 Sn	47.95 ug/l	53.27	1.12	#####	
118 Sn	47.76 ug/l	53.06	1.40	1000	
121 Sb	48.00 ug/l	53.33	1.91	1000	
137 Ba	38.21 ug/l	42.45	1.63	1000	
205 Tl	39.51 ug/l	43.90	0.49	1000	
206 (Pb)	ug/l	#VALUE!	-----	#####	
207 (Pb)	ug/l	#VALUE!	-----	#####	
208 Pb	40.96 ug/l	45.51	1.51	1000	

ISTD Elements

Element	CPS Mean	RSD(%)	Ref Value	Rec(%)	QC Range(%)	Flag
6 Li	-42417.97	6.61	-41328.95	102.6	70 - 120	IS Fai
45 Sc	3132146.50	0.85	3008024.30	104.1	70 - 120	
45 Sc	438590.66	1.66	423303.94	103.6	70 - 120	
45 Sc	8973542.00	0.43	8607281.00	104.3	70 - 120	
72 Ge	781391.50	0.41	774468.63	100.9	70 - 120	
72 Ge	283396.16	0.60	282128.91	100.4	70 - 120	
72 Ge	1916729.60	0.53	1882554.90	101.8	70 - 120	
115 In	5655188.50	0.79	5556751.00	101.8	70 - 120	
115 In	2993471.00	0.53	3029632.80	98.8	70 - 120	
115 In	12415622.00	1.62	12097256.00	102.6	70 - 120	
159 Tb	16697993.00	0.82	16269544.00	102.6	70 - 120	
165 Ho	16175574.00	0.28	15819307.00	102.3	70 - 120	

ISTD Ref File : C:\ICPCHEM\1\DATA\12G23k00.B\004CALB.D\004CALB.D#


0 :Element Failures 0 :Max. Number of Failures Allowed
 1 :ISTD Failures 0 :Max. Number of ISTD Failures Allowed

Data Results:

Analytes: Pass
 ISTD: Fail


Tune Report

Tune File : NG_HMI.u
 Comment : 120723



Integration Time: 0.1000 sec
 Sampling Period: 0.7200 sec
 n: 200
 Oxide: 156/140 0.740%
 Doubly Charged: 70/140 0.995%

m/z	Range	Count	Mean	RSD%	Background
7	20,000	16367.0	16532.4	2.08	2.20
89	50,000	37552.0	36668.9	2.15	2.80
205	50,000	26129.0	26508.7	2.48	7.40
156/140	2	0.770%	0.719%	7.34	
70/140	2	0.962%	0.970%	5.65	
140	50,000	38464.0	39095.6	2.26	4.80
59	50,000	21370.0	21531.4	2.38	2.70



m/z:	7	89	205
Height:	16,163	37,021	26,702
Axis:	7.00	88.95	204.95
W-50%:	0.60	0.60	0.50
W-10%:	0.6500	0.6500	0.6500

Integration Time: 0.1000 sec
 Acquisition Time: 22.7600 sec

Y axis : Linear

Tune Report

Tune File : NG_HMI.u
Comment : 120723

Tuning Parameters

===Plasma Condition===	===Ion Lenses===	===Q-Pole Parameters===
RF Power : 1600 W	Extract 1 : 0 V	AMU Gain : 128
RF Matching : 1.7 V	Extract 2 : -140 V	AMU Offset : 129
Smpl Depth : 8 mm	Omega Bias-ce : -24 V	Axis Gain : 0.9999
Torch-H : 0.2 mm	Omega Lens-ce : -0.4 V	Axis Offset : -0.05
Torch-V : -0.2 mm	Cell Entrance : -30 V	QP Bias : -3 V
Carrier Gas : 0.5 L/min	QP Focus : 5 V	
Makeup Gas : 0.5 L/min	Cell Exit : -30 V	===Detector Parameters===
Optional Gas : --- %		Discriminator : 8 mV
Nebulizer Pump : 0.1 rps	===Octopole Parameters===	Analog HV : 1720 V
Sample Pump : --- rps	OctP RF : 180 V	Pulse HV : 1350 V
S/C Temp : 2 degC	OctP Bias : -6 V	


===Reaction Cell===

Reaction Mode : OFF		
H2 Gas : 0 mL/min	He Gas : 0 mL/min	Optional Gas : --- %


200.8 QC Tune Report

Data File: C:\ICPCHEM\1\DATA\12G23k00.B\001TUNE.D
 Date Acquired: Jul 23 2012 10:53 am
 Acq. Method: TN200_8.M
 Operator: NBS
 Sample Name: 100ppb Tune sol
 Misc Info:
 Vial Number: 1303
 Current Method: C:\ICPCHEM\1\METHODS\TN200_8.M


Element	CPS Mean	Rep1	Rep2	Rep3	Rep4	Rep5	%RSD	Required	Flag
9 Be	789604	786311	783280	790871	795505	792052	0.90	5.00	
24 Mg	2414013	2388869	2415273	2402944	2433304	2429676	1.48	5.00	
59 Co	4416874	4413714	4443242	4412276	4404885	4410252	0.67	5.00	
115 In	22042886	22072398	22207138	21950680	22067872	21916340	0.77	5.00	
208 Pb	3431875	3417255	3453186	3434662	3446885	3407388	0.70	5.00	



9 Be
Mass Calib.
 Actual: 8.95
 Required: 8.90 - 9.10
 Flag:
Peak Width
 Actual: 0.60
 Required: 0.90
 Flag:



24 Mg
Mass Calib.
 Actual: 23.95
 Required: 23.90 - 24.10
 Flag:
Peak Width
 Actual: 0.60
 Required: 0.80
 Flag:



59 Co

Mass Calib.

Actual: 58.95

Required: 58.90 - 59.10


Flag:

Peak Width

Actual: 0.60

Required: 0.90

Flag:



115 In

Mass Calib.

Actual: 115.00

Required: 114.90 - 115.10


Flag:

Peak Width

Actual: 0.55

Required: 0.90

Flag:



208 Pb

Mass Calib.

Actual: 207.95

Required: 207.90 - 208.10

Flag:

Peak Width

Actual: 0.55

Required: 0.80

Flag:

Tune Result:

Pass

SQM 7/16/12
6029/6020A
(R)

ICP-MS STANDARDS 6020/6020A/3015/3051A
Today's Date: 07/18/12
Expires: 07/23/12
Prep 1% HNO3/1.0%HCL
20 mL HNO3 / 2000 mL DI Water
Lot #L08023
20mL HCL / 2000mL DI Water
Lot #51305
Expires: 07/23/12
Internal Standard Mix: Prep 07/12/2012
Standard 4
Amount STD Manufacturer Lot #
50 uL CCV-A ABS STDS 012512-30308
50 uL CCV-B ABS STDS 021312-30337
50 uL CCV-C ABS STDS 012512-30307
Prepared in 100 mL of 1% HNO3/1.0% HCL 07/17/12
Standard 3 07/23/12
Amount STD Manufacturer Lot #
25 uL CCV-A ABS STDS 012512-30308
25 uL CCV-B ABS STDS 021312-30337
25 uL CCV-C ABS STDS 012512-30307
Prepared in 100 mL of 1% HNO3/1.0% HCL 07/17/12
Intermediate-Sb 07/23/12
100 uL of Sb STD (CPI 12A011-30298) in 10 mL of 1% HNO3/1.0% HCL
ICV-Sb 07/23/12
100 uL of Intermediate-Sb in 10 mL of 1% HNO3/1.0% HCL
SQM 7/16/12

Standard 2 07/23/12
Amount STD
500 uL Standard 4 07/17/12
Prepared in 50 mL of 1% HNO3/1.0% HCL 07/17/12
Standard 1 07/23/12
Amount STD
50 uL Standard 4 07/17/12
Prepared in 50 mL of 1% HNO3/1.0% HCL 07/17/12
ICP-MS ICV 07/23/12
Amount STD
50 uL QCS ICV A CPI 11C184-30811
50 uL QCS ICV B CPI 11C184-30812
Prepared in 50 mL of 1% HNO3/1.0% HCL 07/18/12
ICSA Prep: 07/23/12
1 mL ICSA CPI 12E134
Prepared in 5 mL of 1% HNO3/1.0% HCL 07/18/12
ICSAB Prep: 07/23/12
1mL ICSA CPI 12E134
0.025mL INT O2SI 1032370-30265
Prepared in 5 mL of 1% HNO3/1.0% HCL 07/18/12
ICP-LDR 07/23/12
Amount STD
50 uL CCV-A ABS STDS 012512-30308
50 uL CCV-B ABS STDS 021312-30337
50 uL CCV-C ABS STDS 012512-30307
Prepared in 10 mL of 1% HNO3/1.0% HCL 07/18/12

Hg WORKING STANDARD

1ml X 10ug/ml Hg STOCK STD. (07/13/12RJS)/200ml 1% HNO3 Lot#L02030
1ml X 10ug/ml Hg STOCK ICV (07/13/12RJS)/200ml 1% HNO3 Lot#L02030
Final concentration is 50 ug/L. Expires 7/16/12

.02030

SQM 7/17/12
6029/6020A
(A)

ICP-MS STANDARDS 6020/6020A/3015/3051A
Today's Date: 07/17/12
Expires: 07/24/12
Prep 1% HNO3/1.0%HCL
20 mL HNO3 / 2000 mL DI Water
Lot #L08023
20mL HCL / 2000mL DI Water
Lot #51305
Expires: 07/24/12
Internal Standard Mix: Prep 07/17/2012
Standard 4
Amount STD Manufacturer Lot #
50 uL CCV-A ABS STDS 012512-30308
50 uL CCV-B ABS STDS 021312-30337
50 uL CCV-C ABS STDS 012512-30307
Prepared in 100 mL of 1% HNO3/1.0% HCL 07/17/12
Standard 3 07/24/12
Amount STD Manufacturer Lot #
25 uL CCV-A ABS STDS 012512-30308
25 uL CCV-B ABS STDS 021312-30337
25 uL CCV-C ABS STDS 012512-30307
Prepared in 100 mL of 1% HNO3/1.0% HCL 07/17/12
Intermediate-Sb 07/24/12
100 uL of Sb STD (CPI 12A011-30298) in 10 mL of 1% HNO3/1.0% HCL
ICV-Sb 07/24/12
100 uL of Intermediate-Sb in 10 mL of 1% HNO3/1.0% HCL

Standard 2 07/24/12
Amount STD
500 uL Standard 4 07/17/12
Prepared in 50 mL of 1% HNO3/1.0% HCL 07/17/12
Standard 1 07/24/12
Amount STD
50 uL Standard 4 07/17/12
Prepared in 50 mL of 1% HNO3/1.0% HCL 07/17/12
ICP-MS ICV 07/24/12
Amount STD
50 uL QCS ICV A CPI 11C184-30811
50 uL QCS ICV B CPI 11C184-30812
Prepared in 50 mL of 1% HNO3/1.0% HCL 07/17/12
ICSA Prep: 07/24/12
1 mL ICSA CPI 12E134
Prepared in 5 mL of 1% HNO3/1.0% HCL 07/17/12
ICSAB Prep: 07/24/12
1mL ICSA CPI 12E134
0.025mL INT O2SI 1032370-30265
Prepared in 5 mL of 1% HNO3/1.0% HCL 07/17/12
ICP-LDR 07/24/12
Amount STD
50 uL CCV-A ABS STDS 012512-30308
50 uL CCV-B ABS STDS 021312-30337
50 uL CCV-C ABS STDS 012512-30307
Prepared in 10 mL of 1% HNO3/1.0% HCL 07/17/12

Internal Standard Concentration

Amt	STD	Element	Vendor	Lot#	Final Conc. in Std	Expires
500uL	1000 ug/mL	Li	CPI	10L079-27839	5000 ug/L	06/10/12
500uL	1000 ug/mL	In	CPI	10J155-28574	5000 ug/L	09/25/12
500uL	1000 ug/mL	Ho	CPI	10A107-28576	5000 ug/L	09/25/12
500uL	1000 ug/mL	Tb	CPI	11B054-28575	5000 ug/L	09/25/12
500uL	1000 ug/mL	Sc	o2si	1024073-28527	5000 ug/L	08/18/12
500uL	1000 ug/mL	Ge	Environmental Express	116011-29381	5000 ug/L	02/08/13

Prep: 07/17/12 NBS Prep in - 1%HNO3/1.0%HCL: Lot #L08023/51305 in 100mL
Expires: 08/16/12

058

Metals Standards Log Book # 35 Page # 059

NBS 07/23/12

NBS 07/23/12
6520/6020A
Ⓟ

ICP-MS STANDARDS 6020/6020A/3015/3051A
Today's Date: 07/23/12

Expires: 07/30/12
Prep 1% HNO3/1.0% HCL
20 mL HNO3 / 2000 mL DI Water
Lot # 08023
20mL HCL / 2000mL DI Water
Lot # 51305
Expires: 07/30/12

Internal Standard Mix: Prep 07/17/2012

Standard 4	STD	Manufacturer	Lot #
50 uL	CCV-A	ABS STDS	012512-30306
50 uL	CCV-B	ABS STDS	021312-30337
50 uL	CCV-C	ABS STDS	012512-30307

Prepared in 100 mL of 1% HNO3/1.0% HCL 07/23/12

Standard 3	STD	Manufacturer	Lot #
25 uL	CCV-A	ABS STDS	012512-30306
25 uL	CCV-B	ABS STDS	021312-30337
25 uL	CCV-C	ABS STDS	012512-30307

Prepared in 100 mL of 1% HNO3/1.0% HCL 07/23/12

Intermediate-Sb 07/30/12
100 uL of Sb STD (CPI 12A011-30288) in 10 mL of 1% HNO3/1.0% HCL
ICV-Sb 07/30/12
100 uL of Intermediate-Sb in 10 mL of 1% HNO3/1.0% HCL

Standard 2 07/30/12
Amount STD 07/23/12
500 uL Standard 4 07/23/12
Prepared in 50 mL of 1% HNO3/1.0% HCL

Standard 1 07/30/12
Amount STD 07/23/12
50 uL Standard 4 07/23/12
Prepared in 50 mL of 1% HNO3/1.0% HCL

ICP-MS ICV 07/30/12
Amount STD 11C184-30811
50 uL QCS ICV A CPI 11C184-30812
50 uL QCS ICV B CPI 07/23/12
Prepared in 50 mL of 1% HNO3/1.0% HCL

ICSA Prep: 07/30/12
1 mL ICSA CPI 12E134
Prepared in 5 mL of 1% HNO3/1.0% HCL 07/23/12

ICSAB Prep: 07/30/12
1 mL ICSA CPI 12E134
0.025mL INT O2SI 1032370-30285
Prepared in 5 mL of 1% HNO3/1.0% HCL 07/23/12

ICP-LDR 07/30/12
Amount STD 012512-30306
50 uL CCV-A ABS STDS 021312-30337
50 uL CCV-B ABS STDS 012512-30307
50 uL CCV-C ABS STDS 07/23/12
Prepared in 10 mL of 1% HNO3/1.0% HCL

030
030
07/20/12
07/20/12
07/20/12
07/20/12
184-30611
184-30612
07/20/12
12E134
07/20/12
12E134
1370-30265
07/20/12
12-30306
12-30337
12-30307
07/20/12

*B.C. 7-23-12
6010B-C
Ⓟ
VSA Sign PV
B.C. 7-23-12

1% HNO3 / 5% HCl BLK					6010B / 6010C ICSA				
AMT.	REAGENT	MANUFACTURER	LOT	OPEN DATE	AMT.	STD	MANUFACTURER	LOT	EXP DATE
100 mL	HCL	EMD	51258	07/13/12	1 mL	Al	CPI	11J015-30092	05/28/13
20 mL	HNO3	JT BAKER	L10023	07/12/12	1 mL	Ca	CPI	11J031-29989	05/14/13
Prepared in 2000 mL DI Water					1 mL	Mg	CPI	11K178-30093	05/28/13
STD 1 / LDL 6010B/6010C					1 mL	Fe	O2SI	1030787-30616	05/17/13
AMT.	STD	MANUFACTURER	LOT	EXP DATE	Prepared in 50 ml 1% HNO3/5% HCl				
0.5 mL	6010 LDL	ABSOLUTE	091409-25205	09/14/12	6010B/6010C ICSAB				
Prepared in 50 ml 1% HNO3/5% HCl					1 mL	Al	CPI	11J015-30092	05/28/13
STD 3 / HDL 6010B/6010C					1 mL	Ca	CPI	11J031-29989	05/14/13
1 ML	CCV-A	ABSOLUTE	012512-30306	01/25/15	1 mL	Mg	CPI	11K178-30093	05/28/13
1 ML	CCV-B	ABSOLUTE	021312-30339	02/13/15	1 mL	Fe	O2SI	1030787-30616	05/17/13
1 ML	CCV-C	ABSOLUTE	012512-30307	01/25/15	0.5 mL	INT SPECIAL MIX	O2SI	1032370-30265	02/01/13
Prepared in 100 ml 1% HNO3 / 5% HCl					Prepared in 50 ml 1% HNO3/5% HCl				
STD 2 / CCV1 6010B/6010C/6010C					6010B/6010C ICV				
AMT.	STD	PREP DATE	EXP DATE	0.5 ML	QCS ICV A	CPI	12C184-30611	09/20/13	
25 mL	STD 3	Today	1 week	0.5 ML	QCS ICV B	CPI	12C184-30612	09/20/13	
25 mL	1% HNO3/5% HCl	Today	1 week	Prepared in 50ml 1% HNO3/5% HCl					
CCV2 6010B/6010C					YTTRIUM INTERNAL STANDARD				
AMT.	STD	PREP DATE	EXP DATE	2.0 mL	Yttrium	O2SI	1035051-30566	09/28/13	
15 mL	STD 3	Today	1 week	Prepared in 2000 ml 1% HNO3/5% HCl					
25 mL	1% HNO3/5% HCl	Today	1 week						

1/13
EXP DATE
0092 05/28/13
9989 05/14/13
0093 05/28/13
10616 05/17/13
0092 05/28/13
9989 05/14/13
0093 05/28/13
10616 05/17/13
10265 02/01/13
10611 09/20/13
10612 09/20/13
10566 09/28/13

*B.C. 7-23-12
200.7
Ⓟ
VSA Sign PV
B.C. 7-23-12

2% HNO3 / 2% HCl BLK					200.7 ICV				
AMOUNT	REAGENT	MANUFACTURER	LOT	OPEN DATE	AMOUNT	STD	MANUFACTURER	LOT	EXP DATE
40 mL	HCL	BDH	51258	07/13/12	0.25 ML	QCS ICV A	CPI	12C184-30611	09/20/13
40 mL	HNO3	JT BAKER	L10023	07/12/12	0.25 ML	QCS ICV B	CPI	12C184-30612	09/20/13
Prepared in 2000 mL DI Water					Prepared in 50ml 2% HNO3/2% HCl				
STD 1 / LDL 200.7					200.7 ICSA				
AMOUNT	STD	MANUFACTURER	LOT	EXP DATE	0.5 mL	Al	CPI	11J015-30092	05/28/13
0.250 mL	200.7 LDL	O2SI	1028857-29667	11/01/12	0.5 mL	Ca	CPI	11J031-29989	05/14/13
Prepared in 50 ml 2% HNO3/2% HCl					0.5 mL	Mg	CPI	11K178-30093	05/28/13
STD 3 / HDL 200.7					0.5 mL	Fe	O2SI	1030787-30616	05/17/13
0.5 mL	CCV-A	ABSOLUTE	012512-30306	01/25/15	Prepared in 50 ml 2% HNO3/2% HCl				
0.5 mL	CCV-B	ABSOLUTE	021312-30339	02/13/15	200.7 ICSAB				
0.5 mL	CCV-C	ABSOLUTE	012512-30307	01/25/15	0.5 mL	Al	CPI	11J015-30092	05/28/13
Prepared in 100 ml 2% HNO3/2% HCl					0.5 mL	Ca	CPI	11J031-29989	05/14/13
STD 2 / CCV1 200.7					0.5 mL	Mg	CPI	11K178-30093	05/28/13
AMOUNT	STD	PREP DATE	EXP DATE	0.5 mL	Fe	O2SI	1030787-30616	05/17/13	
25 mL	STD 3	TODAY	1 WEEK	0.25 mL	INT SPECIAL MIX	O2SI	1032370-30265	2/1/13	
25 mL	2% HNO3/2% HCl	TODAY	1 WEEK	Prepared in 50 ml 2% HNO3/2% HCl					
CCV2 200.7									
15 mL	STD 3	TODAY	1 WEEK						
25 mL	2% HNO3/2% HCl	TODAY	1 WEEK						

030
030

VSA 7-23-12

Metals Digestion Worksheet

Method Name 3015 Digestion

Prep Method M3015

Set 120723A

Units mL

Spikes	
Spiked ID 1	LCSW LOT# 1036660-30911
Spiked ID 2	LCSW LOT# 1036821-30981
Spiked ID 3	
Spiked ID 4	
Spiked By	NM Date: 07/23/12 9:00:00 AM
Witnessed By	LO Date: 07/23/12 9:00:00 AM

Starting Temp:	20 c
Ending Temp:	170 c
Temperature Type:	Microwave
Sufficient Vol for Matrix QC:	Yes
End Date/Time	07/23/12 10:00

Sample	Sample Container	Spike Amount	Spike ID	Digested Amount	Final Volume	Start Date/Time	Comments
1	120723A Blk			45mL	50mL	07/23/12 9:00	equip: Venus
2	120723A LCS	90uL	1+2	45mL	50mL	07/23/12 9:00	equip: Venus
3	AY64692 AY64692W06			45mL	50mL	07/23/12 9:00	equip: Venus
4	AY65112 AY65112W08			45mL	50mL	07/23/12 9:00	equip: Venus
5	AY65113 AY65113W08			45mL	50mL	07/23/12 9:00	equip: Venus
6	AY65144 AY65144W23			45mL	50mL	07/23/12 9:00	equip: Venus
7	AY65144 MS AY65144W24	90uL	1+2	45mL	50mL	07/23/12 9:00	equip: Venus
8	AY65144 MSD AY65144W24	90uL	1+2	45mL	50mL	07/23/12 9:00	equip: Venus
9	AY65145 AY65145W08			45mL	50mL	07/23/12 9:00	equip: Venus
10	AY65146 AY65146W08			45mL	50mL	07/23/12 9:00	equip: Venus
11	AY65147 AY65147W08			45mL	50mL	07/23/12 9:00	equip: Venus
12	AY65148 AY65148W08			45mL	50mL	07/23/12 9:00	equip: Venus
13	AY65149 AY65149W08			45mL	50mL	07/23/12 9:00	equip: Venus
14	AY65150 AY65150W08			45mL	50mL	07/23/12 9:00	equip: Venus
15	AY65151 AY65151W08			45mL	50mL	07/23/12 9:00	equip: Venus
16	AY65166 AY65166W08			45mL	50mL	07/23/12 9:00	equip: Venus
17	AY65167 AY65167W15			45mL	50mL	07/23/12 9:00	equip: Venus
18	AY65167 MS AY65167W16	90uL	1+2	45mL	50mL	07/23/12 9:00	equip: Venus
19	AY65167 MSD AY65167W16	90uL	1+2	45mL	50mL	07/23/12 9:00	equip: Venus

Solvent and Lot#
HNO3 J.T.B L10023 0229

Sample COC Transfer	
Sample prep employee Initials	nm
Analyst's initials	JA
Date	7-23-12
Time	10:00
Moved to	metals

Technician's Initials	
Scanned By	lo
Sample Preparation	nm
Digestion	nm
Bring up to volume	nm
Modified	07/23/12 8:25:49 AM

Reviewed By: *JA*

Date: 7-23-12

6020/200.8 Injection Log

Directory: K:\ICP-MS Optimus\raw data output csv\

RunID	Injected		Sample Name	Misc Info	FileName	Multiplier
1	23 Jul 2012	11:12	Calibration Blank		120723Arev	1.
2	23 Jul 2012	11:18	120723 Standard 1		120723Arev	1.
3	23 Jul 2012	11:25	120723 Standard 2		120723Arev	1.
4	23 Jul 2012	11:32	120723 Standard 3		120723Arev	1.
5	23 Jul 2012	11:39	120723 Standard 4		120723Arev	1.
6	23 Jul 2012	11:45	ICV 120723		120723Arev	1.
8	23 Jul 2012	11:58	ICB 120723		120723Arev	1.
9	23 Jul 2012	12:05	CCV 120723		120723Arev	1.
10	23 Jul 2012	12:12	CCB 120723		120723Arev	1.
11	23 Jul 2012	12:18	ICSA 120723		120723Arev	1.
12	23 Jul 2012	12:25	ICSAB 120723		120723Arev	1.
13	23 Jul 2012	12:58	120723A-3015-BLK		120723Arev	1.
15	23 Jul 2012	13:11	120723A-3015-LCS		120723Arev	1.
18	23 Jul 2012	13:32	CCV 120723		120723Arev	1.
19	23 Jul 2012	13:45	CCB 120723		120723Arev	1.
21	23 Jul 2012	14:07	AY65112W08		120723Arev	1.
22	23 Jul 2012	14:14	AY65113W08		120723Arev	1.
30	23 Jul 2012	15:07	CCV 120723		120723Arev	1.
31	23 Jul 2012	15:20	CCB 120723		120723Arev	1.