

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 4/21/2026 5:19:31 PM

JOB DESCRIPTION

RED-HILL
PFAS: Ka'amilo Wells P2

JOB NUMBER

380-208399-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
4/21/2026 5:19:31 PM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	14
QC Association Summary	25
Lab Chronicle	26
Certification Summary	27
Method Summary	28
Sample Summary	29
Chain of Custody	30
Receipt Checklists	31

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-208399-1

Job ID: 380-208399-1

Eurofins Pomona

Job Narrative 380-208399-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 4/15/2026 10:10 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 2.3°C.

PFAS

EPA 537.1 and EPA 533 are two distinct methods for the analysis of PFAS in drinking water. The analyses are conducted on differing instrumentation, with calibrations, extraction solvents and sample preservatives being dissimilar among the two methods. Therefore it is probable and not unexpected to see the methods having slight variations in analytical results: Ka'amilo Wells P2 (331-600-WL085) (380-208399-1). (XWB4)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)
PWSID Number: HI0000331

Lab Sample ID: 380-208399-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanesulfonic acid (PFBS)	3.2		2.0	ng/L	1		533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.8		2.0	ng/L	1		533	Total/NA
Perfluorohexanoic acid (PFHxA)	4.2		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	4.9		2.0	ng/L	1		533	Total/NA
Perfluorooctanoic acid (PFOA)	4.1		2.0	ng/L	1		533	Total/NA
Perfluoropentanoic acid (PFPeA)	4.1		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	5.0		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanoic acid (PFHxA)	4.1		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	4.4		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	4.2		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorobutanesulfonic acid (PFBS)	3.2		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluoroheptanoic acid (PFHpA)	2.1		2.0	ng/L	1		EPA 537.1 V2	Total/NA

Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)
PWSID Number: HI0000331

Lab Sample ID: 380-208399-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)

Lab Sample ID: 380-208399-1

Date Collected: 04/13/26 13:01

Matrix: Water

Date Received: 04/15/26 10:10

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluorobutanesulfonic acid (PFBS)	3.2		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluorohexanesulfonic acid (PFHxS)	3.8		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluorohexanoic acid (PFHxA)	4.2		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluorooctanesulfonic acid (PFOS)	4.9		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluorooctanoic acid (PFOA)	4.1		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluoropentanoic acid (PFPeA)	4.1		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 15:53	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	114		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C6 PFDA	115		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C5 PFHxA	104		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C4 PFHpA	110		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C8 PFOA	112		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C9 PFNA	108		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C7 PFUnA	113		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C2 PFDoA	124		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C4 PFBA	112		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C5 PFPeA	115		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C3 PFBS	113		50 - 200	04/17/26 15:20	04/18/26 15:53	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)

Lab Sample ID: 380-208399-1

Date Collected: 04/13/26 13:01

Matrix: Water

Date Received: 04/15/26 10:10

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 PFHxS	115		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C8 PFOS	115		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C2-4:2-FTS	90		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C2-6:2-FTS	88		50 - 200	04/17/26 15:20	04/18/26 15:53	1
13C2-8:2-FTS	99		50 - 200	04/17/26 15:20	04/18/26 15:53	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluorooctanesulfonic acid (PFOS)	5.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluorohexanoic acid (PFHxA)	4.1		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluorooctanoic acid (PFOA)	4.4		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluorohexanesulfonic acid (PFHxS)	4.2		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluorobutanesulfonic acid (PFBS)	3.2		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluoroheptanoic acid (PFHpA)	2.1		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:07	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	109		70 - 130	04/16/26 00:28	04/16/26 15:07	1
13C2 PFHxA	98		70 - 130	04/16/26 00:28	04/16/26 15:07	1
13C2 PFDA	114		70 - 130	04/16/26 00:28	04/16/26 15:07	1
13C3-GenX	105		70 - 130	04/16/26 00:28	04/16/26 15:07	1

Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)

Lab Sample ID: 380-208399-2

Date Collected: 04/13/26 13:01

Matrix: Water

Date Received: 04/15/26 10:10

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)

Lab Sample ID: 380-208399-2

Date Collected: 04/13/26 13:01

Matrix: Water

Date Received: 04/15/26 10:10

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		04/17/26 15:20	04/18/26 16:02	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	112		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C6 PFDA	114		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C5 PFHxA	108		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C4 PFHpA	110		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C8 PFOA	111		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C9 PFNA	106		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C7 PFUnA	111		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C2 PFDoA	121		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C4 PFBA	113		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C5 PFPeA	112		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C3 PFBS	109		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C3 PFHxS	112		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C8 PFOS	109		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C2-4:2-FTS	86		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C2-6:2-FTS	84		50 - 200	04/17/26 15:20	04/18/26 16:02	1
13C2-8:2-FTS	91		50 - 200	04/17/26 15:20	04/18/26 16:02	1

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)

Lab Sample ID: 380-208399-2

Date Collected: 04/13/26 13:01

Matrix: Water

Date Received: 04/15/26 10:10

PWSID Number: HI0000331

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		04/16/26 00:28	04/16/26 15:17	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	102		70 - 130			04/16/26 00:28	04/16/26 15:17	1
13C2 PFHxA	103		70 - 130			04/16/26 00:28	04/16/26 15:17	1
13C2 PFDA	108		70 - 130			04/16/26 00:28	04/16/26 15:17	1
13C3-GenX	102		70 - 130			04/16/26 00:28	04/16/26 15:17	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)

Lab Sample ID: 380-208399-1

PWSID Number: HI0000331

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.8		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	4.9		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	4.1		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	5.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	4.4		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	4.2		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)

Lab Sample ID: 380-208399-2

PWSID Number: HI0000331

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-208399-1
 SDG: PFAS: Ka'amilo Wells P2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-208380-B-1-A MS	Matrix Spike	99	110	116	111
380-208380-C-1-A MSD	Matrix Spike Duplicate	99	98	107	99
380-208399-1	Ka'amilo Wells P2 (331-600-WL085)	109	98	114	105
380-208399-2	FB: Ka'amilo Wells P2 (331-600-WL085)	102	103	108	102
LCS 380-220337/21-A	Lab Control Sample	101	107	113	102
MBL 380-220337/19-A	Method Blank	108	100	106	96
MRL 380-220337/20-A	Lab Control Sample	100	101	110	96

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFHxA = 13C2 PFHxA
 PFDA = 13C2 PFDA
 GenX = 13C3-GenX



Isotope Dilution Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-208399-1
 SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-208397-E-1-A MS	Matrix Spike	115	115	106	105	107	106	119	128
380-208397-F-1-A MSD	Matrix Spike Duplicate	122	114	102	110	110	107	127	127
380-208399-1	Ka'amilo Wells P2 (331-600-WL085)	114	115	104	110	112	108	113	124
380-208399-2	FB: Ka'amilo Wells P2 (331-600-WL085)	112	114	108	110	111	106	111	121
LCS 380-220944/22-A	Lab Control Sample	128	119	113	115	107	110	124	131
MBL 380-220944/20-A	Method Blank	122	119	120	117	119	115	123	131
MRL 380-220944/21-A	Lab Control Sample	106	110	104	101	105	105	115	119

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-208397-E-1-A MS	Matrix Spike	108	109	104	109	104	78	81	86
380-208397-F-1-A MSD	Matrix Spike Duplicate	107	110	108	108	106	77	81	88
380-208399-1	Ka'amilo Wells P2 (331-600-WL085)	112	115	113	115	115	90	88	99
380-208399-2	FB: Ka'amilo Wells P2 (331-600-WL085)	113	112	109	112	109	86	84	91
LCS 380-220944/22-A	Lab Control Sample	111	111	113	115	113	85	87	93
MBL 380-220944/20-A	Method Blank	122	120	125	122	120	99	94	103
MRL 380-220944/21-A	Lab Control Sample	110	111	111	113	113	88	88	95

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDaA = 13C2 PFDaA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-220944/20-A
Matrix: Water
Analysis Batch: 221069

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 220944

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.25		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		04/17/26 15:20	04/18/26 12:15	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	122		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C6 PFDA	119		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C5 PFHxA	120		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C4 PFHpA	117		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C8 PFOA	119		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C9 PFNA	115		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C7 PFUnA	123		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C2 PFDoA	131		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C4 PFBA	122		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C5 PFPeA	120		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C3 PFBS	125		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C3 PFHxS	122		50 - 200	04/17/26 15:20	04/18/26 12:15	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-220944/20-A
Matrix: Water
Analysis Batch: 221069

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 220944

<i>Isotope Dilution</i>	<i>MBL %Recovery</i>	<i>MBL Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C8 PFOS	120		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C2-4:2-FTS	99		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C2-6:2-FTS	94		50 - 200	04/17/26 15:20	04/18/26 12:15	1
13C2-8:2-FTS	103		50 - 200	04/17/26 15:20	04/18/26 12:15	1

Lab Sample ID: LCS 380-220944/22-A
Matrix: Water
Analysis Batch: 221069

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 220944

<i>Analyte</i>	<i>Spike Added</i>	<i>LCS Result</i>	<i>LCS Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	120	118		ng/L		98	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	120	124		ng/L		103	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	120	125		ng/L		104	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	120	112		ng/L		93	70 - 130
Perfluorobutanesulfonic acid (PFBS)	120	119		ng/L		99	70 - 130
Perfluorodecanoic acid (PFDA)	120	115		ng/L		95	70 - 130
Perfluorododecanoic acid (PFDoA)	120	117		ng/L		97	70 - 130
Perfluoroheptanoic acid (PFHpA)	120	110		ng/L		91	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	120	116		ng/L		97	70 - 130
Perfluorohexanoic acid (PFHxA)	120	119		ng/L		99	70 - 130
Perfluorononanoic acid (PFNA)	120	113		ng/L		94	70 - 130
Perfluorooctanesulfonic acid (PFOS)	120	114		ng/L		95	70 - 130
Perfluorooctanoic acid (PFOA)	120	118		ng/L		98	70 - 130
Perfluoroundecanoic acid (PFUnA)	120	118		ng/L		98	70 - 130
Perfluorobutanoic acid (PFBA)	120	114		ng/L		95	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	120	115		ng/L		96	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	120	117		ng/L		97	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	120	115		ng/L		96	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	120	122		ng/L		101	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	120	124		ng/L		103	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	120	120		ng/L		100	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	120	119		ng/L		99	70 - 130
Perfluoropentanoic acid (PFPeA)	120	117		ng/L		97	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	120	118		ng/L		98	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-220944/22-A
Matrix: Water
Analysis Batch: 221069

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 220944

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	120	116		ng/L		96	70 - 130
LCS LCS							
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	128		50 - 200				
13C6 PFDA	119		50 - 200				
13C5 PFHxA	113		50 - 200				
13C4 PFHpA	115		50 - 200				
13C8 PFOA	107		50 - 200				
13C9 PFNA	110		50 - 200				
13C7 PFUnA	124		50 - 200				
13C2 PFDoA	131		50 - 200				
13C4 PFBA	111		50 - 200				
13C5 PFPeA	111		50 - 200				
13C3 PFBS	113		50 - 200				
13C3 PFHxS	115		50 - 200				
13C8 PFOS	113		50 - 200				
13C2-4:2-FTS	85		50 - 200				
13C2-6:2-FTS	87		50 - 200				
13C2-8:2-FTS	93		50 - 200				

Lab Sample ID: MRL 380-220944/21-A
Matrix: Water
Analysis Batch: 221069

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 220944

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	2.01	J	ng/L		100	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	2.09	J	ng/L		104	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.30	J	ng/L		115	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	2.13	J	ng/L		106	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	2.16	J	ng/L		108	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.17	J	ng/L		108	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.16	J	ng/L		108	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.22	J	ng/L		111	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.02	J	ng/L		101	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.25	J	ng/L		112	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.08	J	ng/L		104	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.16	J	ng/L		108	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.21	J	ng/L		110	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.11	J	ng/L		105	50 - 150
Perfluorobutanoic acid (PFBA)	2.00	2.20	J	ng/L		110	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-220944/21-A
Matrix: Water
Analysis Batch: 221069

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 220944

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.00	2.20	J	ng/L		110	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.00	2.36	J	ng/L		118	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.00	2.41	J	ng/L		120	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.00	2.19	J	ng/L		109	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.00	2.12	J	ng/L		106	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.00	2.21	J	ng/L		110	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.00	2.16	J	ng/L		108	50 - 150
Perfluoropentanoic acid (PFPeA)	2.00	2.08	J	ng/L		104	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.00	2.05	J	ng/L		102	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.00	2.14	J	ng/L		107	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	106		50 - 200
13C6 PFDA	110		50 - 200
13C5 PFHxA	104		50 - 200
13C4 PFHpA	101		50 - 200
13C8 PFOA	105		50 - 200
13C9 PFNA	105		50 - 200
13C7 PFUnA	115		50 - 200
13C2 PFDoA	119		50 - 200
13C4 PFBA	110		50 - 200
13C5 PFPeA	111		50 - 200
13C3 PFBS	111		50 - 200
13C3 PFHxS	113		50 - 200
13C8 PFOS	113		50 - 200
13C2-4:2-FTS	88		50 - 200
13C2-6:2-FTS	88		50 - 200
13C2-8:2-FTS	95		50 - 200

Lab Sample ID: 380-208397-E-1-A MS
Matrix: Water
Analysis Batch: 221069

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 220944

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.2	58.7		ng/L		98	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.2	63.4		ng/L		105	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.2	65.7		ng/L		109	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-208397-E-1-A MS
Matrix: Water
Analysis Batch: 221069

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 220944

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		60.2	60.3		ng/L		100	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	3.5		60.2	65.6		ng/L		103	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		60.2	59.1		ng/L		98	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		60.2	57.6		ng/L		96	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		60.2	62.3		ng/L		101	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	3.7		60.2	60.8		ng/L		95	70 - 130
Perfluorohexanoic acid (PFHxA)	4.3		60.2	63.6		ng/L		99	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		60.2	60.2		ng/L		100	70 - 130
Perfluorooctanesulfonic acid (PFOS)	4.9		60.2	65.8		ng/L		101	70 - 130
Perfluorooctanoic acid (PFOA)	4.2		60.2	64.1		ng/L		99	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		60.2	59.0		ng/L		98	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		60.2	61.1		ng/L		98	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.2	62.2		ng/L		103	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.2	60.5		ng/L		101	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.2	61.0		ng/L		101	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.2	58.0		ng/L		96	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.2	62.8		ng/L		104	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.2	62.6		ng/L		104	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.2	61.9		ng/L		103	70 - 130
Perfluoropentanoic acid (PFPeA)	4.3		60.2	63.8		ng/L		99	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.2	61.0		ng/L		101	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.2	58.3		ng/L		96	70 - 130

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	115		50 - 200
13C6 PFDA	115		50 - 200
13C5 PFHxA	106		50 - 200
13C4 PFHpA	105		50 - 200
13C8 PFOA	107		50 - 200
13C9 PFNA	106		50 - 200
13C7 PFUnA	119		50 - 200
13C2 PFDoA	128		50 - 200
13C4 PFBA	108		50 - 200
13C5 PFPeA	109		50 - 200
13C3 PFBS	104		50 - 200
13C3 PFHxS	109		50 - 200
13C8 PFOS	104		50 - 200

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-208397-E-1-A MS
Matrix: Water
Analysis Batch: 221069

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 220944

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>MS MS Qualifier</i>	<i>Limits</i>
13C2-4:2-FTS	78		50 - 200
13C2-6:2-FTS	81		50 - 200
13C2-8:2-FTS	86		50 - 200

Lab Sample ID: 380-208397-F-1-A MSD
Matrix: Water
Analysis Batch: 221069

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 220944

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.1	58.2		ng/L		97	70 - 130	1	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.1	62.8		ng/L		104	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.1	65.6		ng/L		109	70 - 130	0	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		60.1	56.3		ng/L		94	70 - 130	7	30
Perfluorobutanesulfonic acid (PFBS)	3.5		60.1	61.4		ng/L		96	70 - 130	7	30
Perfluorodecanoic acid (PFDA)	<2.0		60.1	61.7		ng/L		103	70 - 130	4	30
Perfluorododecanoic acid (PFDoA)	<2.0		60.1	60.1		ng/L		100	70 - 130	4	30
Perfluoroheptanoic acid (PFHpA)	<2.0		60.1	61.3		ng/L		99	70 - 130	2	30
Perfluorohexanesulfonic acid (PFHxS)	3.7		60.1	62.1		ng/L		97	70 - 130	2	30
Perfluorohexanoic acid (PFHxA)	4.3		60.1	67.6		ng/L		105	70 - 130	6	30
Perfluorononanoic acid (PFNA)	<2.0		60.1	63.0		ng/L		105	70 - 130	5	30
Perfluorooctanesulfonic acid (PFOS)	4.9		60.1	64.0		ng/L		98	70 - 130	3	30
Perfluorooctanoic acid (PFOA)	4.2		60.1	63.6		ng/L		99	70 - 130	1	30
Perfluoroundecanoic acid (PFUnA)	<2.0		60.1	56.8		ng/L		94	70 - 130	4	30
Perfluorobutanoic acid (PFBA)	<2.0		60.1	62.4		ng/L		101	70 - 130	2	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.1	61.4		ng/L		102	70 - 130	1	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.1	63.8		ng/L		106	70 - 130	5	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.1	65.4		ng/L		109	70 - 130	7	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.1	60.8		ng/L		101	70 - 130	5	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		60.1	62.4		ng/L		104	70 - 130	1	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.1	64.6		ng/L		107	70 - 130	3	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.1	61.2		ng/L		102	70 - 130	1	30
Perfluoropentanoic acid (PFPeA)	4.3		60.1	66.3		ng/L		103	70 - 130	4	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.1	62.1		ng/L		103	70 - 130	2	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.1	60.9		ng/L		100	70 - 130	4	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

<i>Isotope Dilution</i>	<i>MSD</i>	<i>MSD</i>	<i>Limits</i>
<i>%Recovery</i>	<i>Qualifier</i>		
13C3 HFPO-DA	122		50 - 200
13C6 PFDA	114		50 - 200
13C5 PFHxA	102		50 - 200
13C4 PFHpA	110		50 - 200
13C8 PFOA	110		50 - 200
13C9 PFNA	107		50 - 200
13C7 PFUnA	127		50 - 200
13C2 PFDoA	127		50 - 200
13C4 PFBA	107		50 - 200
13C5 PFPeA	110		50 - 200
13C3 PFBS	108		50 - 200
13C3 PFHxS	108		50 - 200
13C8 PFOS	106		50 - 200
13C2-4:2-FTS	77		50 - 200
13C2-6:2-FTS	81		50 - 200
13C2-8:2-FTS	88		50 - 200

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-220337/19-A
Matrix: Water
Analysis Batch: 220443

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 220337

<i>Analyte</i>	<i>MBL</i>	<i>MBL</i>	<i>RL</i>	<i>Unit</i>	<i>D</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
	<i>Result</i>	<i>Qualifier</i>						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
Perfluorotridecanoic acid (PFTrDA)	<0.36		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		04/16/26 00:28	04/16/26 12:33	1
<i>Surrogate</i>	<i>MBL</i>	<i>MBL</i>	<i>Limits</i>			<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
	<i>%Recovery</i>	<i>Qualifier</i>						
d5-NEtFOSAA	108		70 - 130			04/16/26 00:28	04/16/26 12:33	1
13C2 PFHxA	100		70 - 130			04/16/26 00:28	04/16/26 12:33	1
13C2 PFDA	106		70 - 130			04/16/26 00:28	04/16/26 12:33	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MBL 380-220337/19-A
Matrix: Water
Analysis Batch: 220443

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 220337

<i>Surrogate</i>	<i>MBL</i>	<i>MBL</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C3-GenX	96	Qualifier	70 - 130	04/16/26 00:28	04/16/26 12:33	1

Lab Sample ID: LCS 380-220337/21-A
Matrix: Water
Analysis Batch: 220443

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 220337

<i>Analyte</i>	<i>Spike</i>	<i>LCS</i>	<i>LCS</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec</i>	<i>Limits</i>
	<i>Added</i>	<i>Result</i>	<i>Qualifier</i>					
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	50.0	50.6		ng/L		101		70 - 130
Perfluorooctanesulfonic acid (PFOS)	50.0	52.2		ng/L		104		70 - 130
Perfluoroundecanoic acid (PFUnA)	50.0	58.3		ng/L		117		70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	50.0	52.5		ng/L		105		70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	50.0	50.2		ng/L		100		70 - 130
Perfluorohexanoic acid (PFHxA)	50.0	52.1		ng/L		104		70 - 130
Perfluorododecanoic acid (PFDoA)	50.0	54.4		ng/L		109		70 - 130
Perfluorooctanoic acid (PFOA)	50.0	53.7		ng/L		107		70 - 130
Perfluorodecanoic acid (PFDA)	50.0	53.1		ng/L		106		70 - 130
Perfluorohexanesulfonic acid (PFHxS)	50.0	50.2		ng/L		100		70 - 130
Perfluorobutanesulfonic acid (PFBS)	50.0	50.7		ng/L		101		70 - 130
Perfluoroheptanoic acid (PFHpA)	50.0	52.6		ng/L		105		70 - 130
Perfluorononanoic acid (PFNA)	50.0	56.7		ng/L		113		70 - 130
Perfluorotetradecanoic acid (PFTA)	50.0	48.2		ng/L		96		70 - 130
Perfluorotridecanoic acid (PFTrDA)	50.0	56.2		ng/L		112		70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	50.0	53.4		ng/L		107		70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	50.0	51.2		ng/L		102		70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	50.0	50.7		ng/L		101		70 - 130

<i>Surrogate</i>	<i>LCS</i>	<i>LCS</i>	<i>Limits</i>
	<i>%Recovery</i>	<i>Qualifier</i>	
d5-NEtFOSAA	101		70 - 130
13C2 PFHxA	107		70 - 130
13C2 PFDA	113		70 - 130
13C3-GenX	102		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MRL 380-220337/20-A
Matrix: Water
Analysis Batch: 220443

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 220337

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	2.00	J	ng/L		100	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.13	J	ng/L		106	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.35	J	ng/L		118	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	2.10	J	ng/L		105	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	2.08	J	ng/L		104	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.12	J	ng/L		106	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.15	J	ng/L		107	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.21	J	ng/L		110	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.29	J	ng/L		115	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.06	J	ng/L		103	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.99	J	ng/L		100	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.21	J	ng/L		110	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.39	J	ng/L		120	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.00	1.98	J	ng/L		99	50 - 150
Perfluorotridecanoic acid (PFTTrDA)	2.00	2.21	J	ng/L		110	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	2.12	J	ng/L		106	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	2.07	J	ng/L		104	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.10	J	ng/L		105	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	100		70 - 130
13C2 PFHxA	101		70 - 130
13C2 PFDA	110		70 - 130
13C3-GenX	96		70 - 130

Lab Sample ID: 380-208380-B-1-A MS
Matrix: Water
Analysis Batch: 220443

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 220337

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		50.3	49.3		ng/L		98	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		50.3	51.9		ng/L		103	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		50.3	56.2		ng/L		112	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		50.3	55.4		ng/L		110	70 - 130

Eurofins Pomona

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

LCMS

Prep Batch: 220337

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-208399-1	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	537.1 DW	
380-208399-2	FB: Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	537.1 DW	
MBL 380-220337/19-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-220337/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-220337/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-208380-B-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-208380-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

Analysis Batch: 220443

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-208399-1	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	EPA 537.1 V2	220337
380-208399-2	FB: Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	EPA 537.1 V2	220337
MBL 380-220337/19-A	Method Blank	Total/NA	Water	EPA 537.1 V2	220337
LCS 380-220337/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	220337
MRL 380-220337/20-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	220337
380-208380-B-1-A MS	Matrix Spike	Total/NA	Water	EPA 537.1 V2	220337
380-208380-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	EPA 537.1 V2	220337

Prep Batch: 220944

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-208399-1	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	
380-208399-2	FB: Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	
MBL 380-220944/20-A	Method Blank	Total/NA	Water	533	
LCS 380-220944/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-220944/21-A	Lab Control Sample	Total/NA	Water	533	
380-208397-E-1-A MS	Matrix Spike	Total/NA	Water	533	
380-208397-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Analysis Batch: 221069

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-208399-1	Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	220944
380-208399-2	FB: Ka'amilo Wells P2 (331-600-WL085)	Total/NA	Water	533	220944
MBL 380-220944/20-A	Method Blank	Total/NA	Water	533	220944
LCS 380-220944/22-A	Lab Control Sample	Total/NA	Water	533	220944
MRL 380-220944/21-A	Lab Control Sample	Total/NA	Water	533	220944
380-208397-E-1-A MS	Matrix Spike	Total/NA	Water	533	220944
380-208397-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	220944

Lab Chronicle

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-208399-1
 SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: Ka'amilo Wells P2 (331-600-WL085)

Lab Sample ID: 380-208399-1

Date Collected: 04/13/26 13:01

Matrix: Water

Date Received: 04/15/26 10:10

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			220944	N8NE	EA POM	04/17/26 15:20
Total/NA	Analysis	533		1	221069	Y5FM	EA POM	04/18/26 15:53
Total/NA	Prep	537.1 DW			220337	G9MN	EA POM	04/16/26 00:28
Total/NA	Analysis	EPA 537.1 V2		1	220443	M7ML	EA POM	04/16/26 15:07

Client Sample ID: FB: Ka'amilo Wells P2 (331-600-WL085)

Lab Sample ID: 380-208399-2

Date Collected: 04/13/26 13:01

Matrix: Water

Date Received: 04/15/26 10:10

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			220944	N8NE	EA POM	04/17/26 15:20
Total/NA	Analysis	533		1	221069	Y5FM	EA POM	04/18/26 16:02
Total/NA	Prep	537.1 DW			220337	G9MN	EA POM	04/16/26 00:28
Total/NA	Analysis	EPA 537.1 V2		1	220443	M7ML	EA POM	04/16/26 15:17

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-208399-1
SDG: PFAS: Ka'amilo Wells P2


Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-208399-1	Ka'amilo Wells P2 (331-600-WL085)	Water	04/13/26 13:01	04/15/26 10:10	HI0000331
380-208399-2	FB: Ka'amilo Wells P2 (331-600-WL085)	Water	04/13/26 13:01	04/15/26 10:10	HI0000331

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Monrovia, CA (Suite 100)

750 Royal Oaks Drive Suite 100
 Monrovia, CA 91016
 Phone (626) 386-1100

Chain of Custody Record

Client Information		Sampler bailey		Lab PM. Lopez, Maria		Carrier Tracking No(s)		CDC No:					
Client Contact: kirk iwamoto		Phone: +1 808 748 5840		E-Mail: Maria.Lopez@et.euronisus.com		State of Origin:		Page:					
Company: City & County of Honolulu		PWSID:		Analysis Requested						Job #:			
Address: 630 South Beretania Street, Chemistry Lab		Due Date Requested:		Field Filtered Sample (Yes or No) Perform MS/MSD (Yes or No) SUBCONTRACT - 625 PAH Physic LL (EAL) + TICs 8015B_GRO_LL - (MOD) GRO 8015B_DRO_LL_CS - HNL Ranges: C10-C24/C24-C36/C8-C18 525.2_PREC - (MOD) 525plus PLUS TICs 537.1_DW_PREC - 537.1 Full List 533 - All Analytes						Preservation Codes:			
City: Honolulu		TAT Requested (days): RUSH								A - HCL		M - Hexane	
State, Zip: HI, 96843		Compliance Project: <input type="checkbox"/> No								B - NaOH		N - None	
Phone: 808-748-5840 (tel)		PO #: C20525101 exp 05312023								C - Zn Acetate		O - AsNaO2	
Email: kiwamoto@hbws.org		WO #:								D - Nitric Acid		P - Na2O4S	
Project Name: RED-HILL/HBWS sites Event Desc: RUSH Weekly Red Hill		Project #: 38001111		E - NaHSO4		Q - Na2SO3		R - Na2S2O3					
Site:		SSOW#:		F - MeOH		S - H2SO4		T TSP Dodecahydrate					
Sample Identification		Sample Date		Sample Time		Sample Type (C=Comp, G=grab)		Matrix (W=water, S=solid, O=water/sol)		Other:			
Ka'amilo Wells P2 (331-600-WL085)		13-Apr-2026		1301		G		Water		Special Instructions/Note:			
FB: Ka'amilo Wells P2 (331-600-WL085)		13-Apr-2026		1301				Water		 380-208399 COC			
Possible Hazard Identification <input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological						Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months							
Deliverable Requested: I, II, III, IV, Other (specify)						Special Instructions/QC Requirements							
Empty Kit Relinquished by		Date		Time:		Method of Shipment: 8706 6414 2726							
Relinquished by:		Date/Time: 14/01/2026		Company: HBWS		Received by: [Signature]		Date/Time: 4/15/26 10:10		Company: [Signature]			
Relinquished by:		Date/Time:		Company:		Received by:		Date/Time:		Company:			
Custody Seals Intact: <input type="checkbox"/> Yes <input type="checkbox"/> No		Custody Seal No		Cooler Temperature(s) °C and Other Remarks: 60NA/2.3 = 2.3 602									

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-208399-1
SDG Number: PFAS: Ka'amilo Wells P2

Login Number: 208399
List Number: 1
Creator: Ngo, Theodore

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

