

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 4/30/2026 4:39:22 PM

JOB DESCRIPTION

RED-HILL
PFAS: Moanalua Wells

JOB NUMBER

380-209878-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
4/30/2026 4:39:22 PM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	15
QC Association Summary	26
Lab Chronicle	27
Certification Summary	28
Method Summary	29
Sample Summary	30
Chain of Custody	31
Receipt Checklists	32

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-209878-1

Job ID: 380-209878-1

Eurofins Pomona

Job Narrative 380-209878-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 4/22/2026 10:22 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 3.7°C.

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Client Sample ID: MOANALUA WELLS (331-223-TP202)
PWSID Number: HI0000331

Lab Sample ID: 380-209878-1

No Detections.

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)
PWSID Number: HI0000331

Lab Sample ID: 380-209878-2

No Detections.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

This Detection Summary does not include radiochemical test results.

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-209878-1

Date Collected: 04/20/26 10:33

Matrix: Drinking Water

Date Received: 04/22/26 10:22

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 08:52	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	110		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C6 PFDA	110		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C5 PFHxA	114		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C4 PFHpA	114		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C8 PFOA	113		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C9 PFNA	112		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C7 PFUnA	106		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C2 PFDoA	112		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C4 PFBA	109		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C5 PFPeA	110		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C3 PFBS	113		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C3 PFHxS	112		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C8 PFOS	110		50 - 200	04/27/26 07:29	04/28/26 08:52	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-209878-1

Date Collected: 04/20/26 10:33

Matrix: Drinking Water

Date Received: 04/22/26 10:22

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2-4:2-FTS	102		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C2-6:2-FTS	94		50 - 200	04/27/26 07:29	04/28/26 08:52	1
13C2-8:2-FTS	85		50 - 200	04/27/26 07:29	04/28/26 08:52	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 11:04	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	106		70 - 130			04/24/26 01:00	04/24/26 11:04	1
13C2 PFHxA	112		70 - 130			04/24/26 01:00	04/24/26 11:04	1
13C2 PFDA	103		70 - 130			04/24/26 01:00	04/24/26 11:04	1
13C3-GenX	105		70 - 130			04/24/26 01:00	04/24/26 11:04	1

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-209878-2

Date Collected: 04/20/26 10:33

Matrix: Water

Date Received: 04/22/26 10:22

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-209878-2

Date Collected: 04/20/26 10:33

Matrix: Water

Date Received: 04/22/26 10:22

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		04/27/26 07:29	04/28/26 09:01	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	109		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C6 PFDA	103		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C5 PFHxA	114		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C4 PFHpA	109		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C8 PFOA	112		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C9 PFNA	110		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C7 PFUnA	103		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C2 PFDoA	105		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C4 PFBA	110		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C5 PFPeA	112		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C3 PFBS	114		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C3 PFHxS	112		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C8 PFOS	114		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C2-4:2-FTS	102		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C2-6:2-FTS	91		50 - 200	04/27/26 07:29	04/28/26 09:01	1
13C2-8:2-FTS	86		50 - 200	04/27/26 07:29	04/28/26 09:01	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-209878-2

Date Collected: 04/20/26 10:33

Matrix: Water

Date Received: 04/22/26 10:22

PWSID Number: HI0000331

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		04/24/26 01:00	04/24/26 13:18	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	120		70 - 130			04/24/26 01:00	04/24/26 13:18	1
13C2 PFHxA	122		70 - 130			04/24/26 01:00	04/24/26 13:18	1
13C2 PFDA	110		70 - 130			04/24/26 01:00	04/24/26 13:18	1
13C3-GenX	118		70 - 130			04/24/26 01:00	04/24/26 13:18	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-209878-1

PWSID Number: HI0000331

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-209878-2

PWSID Number: HI0000331

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-209878-1
 SDG: PFAS: Moanalua Wells

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-209878-1	MOANALUA WELLS (331-223-T	106	112	103	105
380-209878-1 MS	MOANALUA WELLS (331-223-TP202)	98	112	111	109
380-209878-1 MSD	MOANALUA WELLS (331-223-TP202)	103	107	102	102

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFHxA = 13C2 PFHxA
 PFDA = 13C2 PFDA
 GenX = 13C3-GenX

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-209878-2	FB: MOANALUA WELLS (331-2	12	122	110	118
LCS 380-222326/21-A	Lab Control Sample	97	105	102	97
MBL 380-222326/19-A	Method Blank	109	122	116	111
MRL 380-222326/20-A	Lab Control Sample	102	105	104	96

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFHxA = 13C2 PFHxA
 PFDA = 13C2 PFDA
 GenX = 13C3-GenX

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDoA (50-200)
380-209878-1	MOANALUA WELLS (331-223-T	110	110	114	114	113	112	106	112

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-209878-1	MOANALUA WELLS (331-223-T	109	110	113	112	110	102	94	85

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDoA = 13C2 PFDoA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDoA (50-200)
380-209591-B-7-A MS	Matrix Spike	112	110	110	116	111	109	109	111
380-209591-C-7-A MSD	Matrix Spike Duplicate	112	111	113	113	110	112	109	114
380-209878-2	FB: MOANALUA WELLS (331-223-TP202)	109	103	114	109	112	110	103	105
LCS 380-223143/22-A	Lab Control Sample	104	110	106	106	109	112	109	110
MBL 380-223143/20-A	Method Blank	88	97	94	99	101	97	101	109
MRL 380-223143/21-A	Lab Control Sample	86	98	90	93	97	97	96	102

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-209591-B-7-A MS	Matrix Spike	109	108	111	110	110	99	94	90
380-209591-C-7-A MSD	Matrix Spike Duplicate	113	113	109	108	111	95	95	92
380-209878-2	FB: MOANALUA WELLS (331-223-TP202)	110	112	114	112	114	102	91	86
LCS 380-223143/22-A	Lab Control Sample	88	104	115	114	113	105	100	98
MBL 380-223143/20-A	Method Blank	97	95	109	109	113	105	102	99
MRL 380-223143/21-A	Lab Control Sample	82	90	110	112	116	107	103	96

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA

Isotope Dilution Summary

Client: City & County of Honolulu

Project/Site: RED-HILL

13C5PHA = 13C5 PFHxA
C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDoA = 13C2 PFDoA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-223143/20-A
Matrix: Water
Analysis Batch: 223382

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 223143

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.25		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		04/27/26 07:29	04/28/26 06:45	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	88		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C6 PFDA	97		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C5 PFHxA	94		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C4 PFHpA	99		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C8 PFOA	101		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C9 PFNA	97		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C7 PFUnA	101		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C2 PFDoA	109		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C4 PFBA	97		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C5 PFPeA	95		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C3 PFBS	109		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C3 PFHxS	109		50 - 200	04/27/26 07:29	04/28/26 06:45	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-223143/20-A
Matrix: Water
Analysis Batch: 223382

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 223143

<i>Isotope Dilution</i>	<i>MBL %Recovery</i>	<i>MBL Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C8 PFOS	113		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C2-4:2-FTS	105		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C2-6:2-FTS	102		50 - 200	04/27/26 07:29	04/28/26 06:45	1
13C2-8:2-FTS	99		50 - 200	04/27/26 07:29	04/28/26 06:45	1

Lab Sample ID: LCS 380-223143/22-A
Matrix: Water
Analysis Batch: 223382

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 223143

<i>Analyte</i>	<i>Spike Added</i>	<i>LCS Result</i>	<i>LCS Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	60.0	56.2		ng/L		94	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	60.0	57.5		ng/L		96	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	60.0	59.0		ng/L		98	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	60.0	58.5		ng/L		97	70 - 130
Perfluorobutanesulfonic acid (PFBS)	60.0	57.1		ng/L		95	70 - 130
Perfluorodecanoic acid (PFDA)	60.0	56.2		ng/L		94	70 - 130
Perfluorododecanoic acid (PFDoA)	60.0	59.2		ng/L		99	70 - 130
Perfluoroheptanoic acid (PFHpA)	60.0	60.0		ng/L		100	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	60.0	57.2		ng/L		95	70 - 130
Perfluorohexanoic acid (PFHxA)	60.0	58.9		ng/L		98	70 - 130
Perfluorononanoic acid (PFNA)	60.0	56.6		ng/L		94	70 - 130
Perfluorooctanesulfonic acid (PFOS)	60.0	59.1		ng/L		99	70 - 130
Perfluorooctanoic acid (PFOA)	60.0	58.4		ng/L		97	70 - 130
Perfluoroundecanoic acid (PFUnA)	60.0	60.4		ng/L		101	70 - 130
Perfluorobutanoic acid (PFBA)	60.0	58.2		ng/L		97	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	60.0	59.9		ng/L		100	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	60.0	59.3		ng/L		99	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	60.0	59.2		ng/L		99	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	60.0	62.9		ng/L		105	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	60.0	55.2		ng/L		92	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	60.0	70.6		ng/L		118	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	60.0	60.3		ng/L		100	70 - 130
Perfluoropentanoic acid (PFPeA)	60.0	60.3		ng/L		101	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	60.0	60.9		ng/L		101	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-223143/22-A
Matrix: Water
Analysis Batch: 223382

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 223143

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	60.0	58.6		ng/L		98	70 - 130
LCS LCS							
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	104		50 - 200				
13C6 PFDA	110		50 - 200				
13C5 PFHxA	106		50 - 200				
13C4 PFHpA	106		50 - 200				
13C8 PFOA	109		50 - 200				
13C9 PFNA	112		50 - 200				
13C7 PFUnA	109		50 - 200				
13C2 PFDoA	110		50 - 200				
13C4 PFBA	88		50 - 200				
13C5 PFPeA	104		50 - 200				
13C3 PFBS	115		50 - 200				
13C3 PFHxS	114		50 - 200				
13C8 PFOS	113		50 - 200				
13C2-4:2-FTS	105		50 - 200				
13C2-6:2-FTS	100		50 - 200				
13C2-8:2-FTS	98		50 - 200				

Lab Sample ID: MRL 380-223143/21-A
Matrix: Water
Analysis Batch: 223382

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 223143

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	1.79	J	ng/L		89	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	1.88	J	ng/L		94	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.13	J	ng/L		106	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	2.11	J	ng/L		105	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	2.07	J	ng/L		103	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	1.91	J	ng/L		96	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.07	J	ng/L		103	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.06	J	ng/L		103	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.05	J	ng/L		102	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.16	J	ng/L		108	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.10	J	ng/L		105	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.04	J	ng/L		102	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.12	J	ng/L		106	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.06	J	ng/L		103	50 - 150
Perfluorobutanoic acid (PFBA)	2.00	2.05	J	ng/L		102	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-223143/21-A
Matrix: Water
Analysis Batch: 223382

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 223143

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.00	2.19	J	ng/L		109	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.00	2.21	J	ng/L		110	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.00	2.06	J	ng/L		103	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.00	2.11	J	ng/L		105	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.00	2.03	J	ng/L		101	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.00	2.37	J	ng/L		118	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.00	2.01	J	ng/L		100	50 - 150
Perfluoropentanoic acid (PFPeA)	2.00	2.16	J	ng/L		108	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.00	1.97	J	ng/L		98	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.00	2.00	J	ng/L		100	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	86		50 - 200
13C6 PFDA	98		50 - 200
13C5 PFHxA	90		50 - 200
13C4 PFHpA	93		50 - 200
13C8 PFOA	97		50 - 200
13C9 PFNA	97		50 - 200
13C7 PFUnA	96		50 - 200
13C2 PFDoA	102		50 - 200
13C4 PFBA	82		50 - 200
13C5 PFPeA	90		50 - 200
13C3 PFBS	110		50 - 200
13C3 PFHxS	112		50 - 200
13C8 PFOS	116		50 - 200
13C2-4:2-FTS	107		50 - 200
13C2-6:2-FTS	103		50 - 200
13C2-8:2-FTS	96		50 - 200

Lab Sample ID: 380-209591-B-7-A MS
Matrix: Water
Analysis Batch: 223382

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 223143

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.4	56.0		ng/L		93	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.4	56.7		ng/L		94	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.4	60.8		ng/L		101	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-209591-B-7-A MS
Matrix: Water
Analysis Batch: 223382

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 223143

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		60.4	60.8		ng/L		101	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	5.9		60.4	66.2		ng/L		100	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		60.4	58.3		ng/L		97	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		60.4	59.6		ng/L		99	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		60.4	56.8		ng/L		94	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	6.4		60.4	65.7		ng/L		98	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		60.4	61.1		ng/L		100	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		60.4	60.1		ng/L		100	70 - 130
Perfluorooctanesulfonic acid (PFOS)	4.2		60.4	65.0		ng/L		101	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		60.4	59.0		ng/L		97	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		60.4	59.9		ng/L		99	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		60.4	60.3		ng/L		100	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.4	60.4		ng/L		100	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.4	60.9		ng/L		101	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.4	63.1		ng/L		105	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.4	65.4		ng/L		108	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.4	57.5		ng/L		95	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.4	59.7		ng/L		99	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.4	61.1		ng/L		101	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		60.4	61.8		ng/L		101	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.4	60.8		ng/L		100	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.4	61.6		ng/L		100	70 - 130

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>
13C3 HFPO-DA	112		50 - 200
13C6 PFDA	110		50 - 200
13C5 PFHxA	110		50 - 200
13C4 PFHpA	116		50 - 200
13C8 PFOA	111		50 - 200
13C9 PFNA	109		50 - 200
13C7 PFUnA	109		50 - 200
13C2 PFDoA	111		50 - 200
13C4 PFBA	109		50 - 200
13C5 PFPeA	108		50 - 200
13C3 PFBS	111		50 - 200
13C3 PFHxS	110		50 - 200
13C8 PFOS	110		50 - 200

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-209591-B-7-A MS
Matrix: Water
Analysis Batch: 223382

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 223143

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>MS MS Qualifier</i>	<i>Limits</i>
13C2-4:2-FTS	99		50 - 200
13C2-6:2-FTS	94		50 - 200
13C2-8:2-FTS	90		50 - 200

Lab Sample ID: 380-209591-C-7-A MSD
Matrix: Water
Analysis Batch: 223382

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 223143

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.2	54.3		ng/L		90	70 - 130	3	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.2	55.1		ng/L		92	70 - 130	3	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.2	58.7		ng/L		98	70 - 130	3	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		60.2	60.4		ng/L		100	70 - 130	1	30
Perfluorobutanesulfonic acid (PFBS)	5.9		60.2	65.7		ng/L		99	70 - 130	1	30
Perfluorodecanoic acid (PFDA)	<2.0		60.2	55.7		ng/L		92	70 - 130	5	30
Perfluorododecanoic acid (PFDoA)	<2.0		60.2	57.0		ng/L		95	70 - 130	4	30
Perfluoroheptanoic acid (PFHpA)	<2.0		60.2	59.6		ng/L		99	70 - 130	5	30
Perfluorohexanesulfonic acid (PFHxS)	6.4		60.2	65.5		ng/L		98	70 - 130	0	30
Perfluorohexanoic acid (PFHxA)	<2.0		60.2	58.7		ng/L		97	70 - 130	4	30
Perfluorononanoic acid (PFNA)	<2.0		60.2	59.7		ng/L		99	70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	4.2		60.2	61.9		ng/L		96	70 - 130	5	30
Perfluorooctanoic acid (PFOA)	<2.0		60.2	59.1		ng/L		97	70 - 130	0	30
Perfluoroundecanoic acid (PFUnA)	<2.0		60.2	59.9		ng/L		99	70 - 130	0	30
Perfluorobutanoic acid (PFBA)	<2.0		60.2	60.2		ng/L		100	70 - 130	0	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.2	59.4		ng/L		99	70 - 130	2	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.2	60.2		ng/L		100	70 - 130	1	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.2	59.5		ng/L		99	70 - 130	6	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.2	59.1		ng/L		98	70 - 130	10	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.2	59.5		ng/L		99	70 - 130	3	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.2	60.0		ng/L		100	70 - 130	1	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.2	60.7		ng/L		101	70 - 130	1	30
Perfluoropentanoic acid (PFPeA)	<2.0		60.2	61.0		ng/L		100	70 - 130	1	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.2	58.4		ng/L		96	70 - 130	4	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.2	61.5		ng/L		99	70 - 130	0	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

<i>Isotope Dilution</i>	<i>MSD</i>	<i>MSD</i>	<i>Limits</i>
<i>%Recovery</i>	<i>Qualifier</i>		
13C3 HFPO-DA	112		50 - 200
13C6 PFDA	111		50 - 200
13C5 PFHxA	113		50 - 200
13C4 PFHpA	113		50 - 200
13C8 PFOA	110		50 - 200
13C9 PFNA	112		50 - 200
13C7 PFUnA	109		50 - 200
13C2 PFDoA	114		50 - 200
13C4 PFBA	113		50 - 200
13C5 PFPeA	113		50 - 200
13C3 PFBS	109		50 - 200
13C3 PFHxS	108		50 - 200
13C8 PFOS	111		50 - 200
13C2-4:2-FTS	95		50 - 200
13C2-6:2-FTS	95		50 - 200
13C2-8:2-FTS	92		50 - 200

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-222326/19-A
Matrix: Water
Analysis Batch: 222689

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 222326

<i>Analyte</i>	<i>MBL</i>	<i>MBL</i>	<i>RL</i>	<i>Unit</i>	<i>D</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
	<i>Result</i>	<i>Qualifier</i>						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
Perfluorotridecanoic acid (PFTrDA)	<0.36		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		04/24/26 01:00	04/24/26 10:34	1
<i>Surrogate</i>	<i>MBL</i>	<i>MBL</i>	<i>Limits</i>			<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
	<i>%Recovery</i>	<i>Qualifier</i>						
d5-NEtFOSAA	109		70 - 130			04/24/26 01:00	04/24/26 10:34	1
13C2 PFHxA	122		70 - 130			04/24/26 01:00	04/24/26 10:34	1
13C2 PFDA	116		70 - 130			04/24/26 01:00	04/24/26 10:34	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MBL 380-222326/19-A
Matrix: Water
Analysis Batch: 222689

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 222326

<i>Surrogate</i>	<i>MBL</i>	<i>MBL</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C3-GenX	111	Qualifier	70 - 130	04/24/26 01:00	04/24/26 10:34	1

Lab Sample ID: LCS 380-222326/21-A
Matrix: Water
Analysis Batch: 222689

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 222326

<i>Analyte</i>	<i>Spike</i>	<i>LCS</i>	<i>LCS</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec</i>	<i>Limits</i>
	<i>Added</i>	<i>Result</i>	<i>Qualifier</i>					
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	50.0	46.4		ng/L		93		70 - 130
Perfluorooctanesulfonic acid (PFOS)	50.0	50.6		ng/L		101		70 - 130
Perfluoroundecanoic acid (PFUnA)	50.0	45.3		ng/L		91		70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	50.0	45.3		ng/L		91		70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	50.0	46.8		ng/L		94		70 - 130
Perfluorohexanoic acid (PFHxA)	50.0	48.4		ng/L		97		70 - 130
Perfluorododecanoic acid (PFDoA)	50.0	47.2		ng/L		94		70 - 130
Perfluorooctanoic acid (PFOA)	50.0	49.1		ng/L		98		70 - 130
Perfluorodecanoic acid (PFDA)	50.0	45.5		ng/L		91		70 - 130
Perfluorohexanesulfonic acid (PFHxS)	50.0	50.0		ng/L		100		70 - 130
Perfluorobutanesulfonic acid (PFBS)	50.0	48.8		ng/L		98		70 - 130
Perfluoroheptanoic acid (PFHpA)	50.0	48.6		ng/L		97		70 - 130
Perfluorononanoic acid (PFNA)	50.0	49.2		ng/L		98		70 - 130
Perfluorotetradecanoic acid (PFTA)	50.0	48.1		ng/L		96		70 - 130
Perfluorotridecanoic acid (PFTrDA)	50.0	50.5		ng/L		101		70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	50.0	49.5		ng/L		99		70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	50.0	50.5		ng/L		101		70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	50.0	46.6		ng/L		93		70 - 130

<i>Surrogate</i>	<i>LCS</i>	<i>LCS</i>	<i>Limits</i>
	<i>%Recovery</i>	<i>Qualifier</i>	
d5-NEtFOSAA	97		70 - 130
13C2 PFHxA	105		70 - 130
13C2 PFDA	102		70 - 130
13C3-GenX	97		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MRL 380-222326/20-A
Matrix: Water
Analysis Batch: 222689

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 222326

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	1.76	J	ng/L		88	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	1.99	J	ng/L		100	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.01	J	ng/L		101	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	1.84	J	ng/L		92	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	1.95	J	ng/L		98	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	1.91	J	ng/L		95	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.01	J	ng/L		100	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	1.94	J	ng/L		97	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.01	J	ng/L		100	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	1.93	J	ng/L		97	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.80	J	ng/L		90	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	1.96	J	ng/L		98	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.04	J	ng/L		102	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.00	2.03	J	ng/L		102	50 - 150
Perfluorotridecanoic acid (PFTTrDA)	2.00	2.10	J	ng/L		105	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	1.85	J	ng/L		93	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	1.81	J	ng/L		90	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	1.72	J	ng/L		86	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	102		70 - 130
13C2 PFHxA	105		70 - 130
13C2 PFDA	104		70 - 130
13C3-GenX	96		70 - 130

Lab Sample ID: 380-209878-1 MS
Matrix: Drinking Water
Analysis Batch: 222689

Client Sample ID: MOANALUA WELLS (331-223-TP202)
Prep Type: Total/NA
Prep Batch: 222326

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		50.1	49.3		ng/L		98	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		50.1	51.9		ng/L		100	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		50.1	49.7		ng/L		99	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		50.1	49.0		ng/L		98	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: 380-209878-1 MS

Matrix: Drinking Water

Analysis Batch: 222689

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Prep Type: Total/NA

Prep Batch: 222326

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		50.1	48.6		ng/L		97	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		50.1	51.5		ng/L		100	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		50.1	50.1		ng/L		100	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		50.1	52.6		ng/L		103	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		50.1	49.8		ng/L		99	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		50.1	50.1		ng/L		97	70 - 130
Perfluorobutanesulfonic acid (PFBS)	<2.0		50.1	53.3		ng/L		105	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		50.1	50.7		ng/L		100	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		50.1	50.2		ng/L		100	70 - 130
Perfluorotetradecanoic acid (PFTA)	<2.0		50.1	48.3		ng/L		96	70 - 130
Perfluorotridecanoic acid (PFTTrDA)	<2.0		50.1	52.2		ng/L		104	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		50.1	50.2		ng/L		100	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		50.1	51.1		ng/L		102	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		50.1	51.8		ng/L		103	70 - 130

Surrogate	MS %Recovery	MS Qualifier	MS Limits
d5-NEtFOSAA	98		70 - 130
13C2 PFHxA	112		70 - 130
13C2 PFDA	111		70 - 130
13C3-GenX	109		70 - 130

Lab Sample ID: 380-209878-1 MSD

Matrix: Drinking Water

Analysis Batch: 222689

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Prep Type: Total/NA

Prep Batch: 222326

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		50.1	45.9		ng/L		92	70 - 130	7	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		50.1	55.5		ng/L		107	70 - 130	7	30
Perfluoroundecanoic acid (PFUnA)	<2.0		50.1	46.0		ng/L		92	70 - 130	8	30
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		50.1	52.0		ng/L		104	70 - 130	6	30
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		50.1	51.1		ng/L		102	70 - 130	5	30
Perfluorohexanoic acid (PFHxA)	<2.0		50.1	50.8		ng/L		99	70 - 130	1	30
Perfluorododecanoic acid (PFDoA)	<2.0		50.1	44.1		ng/L		88	70 - 130	13	30
Perfluorooctanoic acid (PFOA)	<2.0		50.1	48.3		ng/L		95	70 - 130	8	30
Perfluorodecanoic acid (PFDA)	<2.0		50.1	44.1		ng/L		88	70 - 130	12	30

Eurofins Pomona

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

LCMS

Prep Batch: 222326

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-209878-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	537.1 DW	
380-209878-2	FB: MOANALUA WELLS (331-223-TP202)	Total/NA	Water	537.1 DW	
MBL 380-222326/19-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-222326/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-222326/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-209878-1 MS	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	537.1 DW	
380-209878-1 MSD	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	537.1 DW	

Analysis Batch: 222689

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-209878-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	EPA 537.1 V2	222326
380-209878-2	FB: MOANALUA WELLS (331-223-TP202)	Total/NA	Water	EPA 537.1 V2	222326
MBL 380-222326/19-A	Method Blank	Total/NA	Water	EPA 537.1 V2	222326
LCS 380-222326/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	222326
MRL 380-222326/20-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	222326
380-209878-1 MS	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	EPA 537.1 V2	222326
380-209878-1 MSD	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	EPA 537.1 V2	222326

Prep Batch: 223143

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-209878-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	533	
380-209878-2	FB: MOANALUA WELLS (331-223-TP202)	Total/NA	Water	533	
MBL 380-223143/20-A	Method Blank	Total/NA	Water	533	
LCS 380-223143/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-223143/21-A	Lab Control Sample	Total/NA	Water	533	
380-209591-B-7-A MS	Matrix Spike	Total/NA	Water	533	
380-209591-C-7-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Analysis Batch: 223382

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-209878-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	533	223143
380-209878-2	FB: MOANALUA WELLS (331-223-TP202)	Total/NA	Water	533	223143
MBL 380-223143/20-A	Method Blank	Total/NA	Water	533	223143
LCS 380-223143/22-A	Lab Control Sample	Total/NA	Water	533	223143
MRL 380-223143/21-A	Lab Control Sample	Total/NA	Water	533	223143
380-209591-B-7-A MS	Matrix Spike	Total/NA	Water	533	223143
380-209591-C-7-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	223143

Lab Chronicle

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-209878-1
 SDG: PFAS: Moanalua Wells

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-209878-1

Date Collected: 04/20/26 10:33

Matrix: Drinking Water

Date Received: 04/22/26 10:22

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			223143	XTD8	EA POM	04/27/26 07:29
Total/NA	Analysis	533		1	223382	SZ9R	EA POM	04/28/26 08:52
Total/NA	Prep	537.1 DW			222326	G9MN	EA POM	04/24/26 01:00
Total/NA	Analysis	EPA 537.1 V2		1	222689	Y5FM	EA POM	04/24/26 11:04

Client Sample ID: FB: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-209878-2

Date Collected: 04/20/26 10:33

Matrix: Water

Date Received: 04/22/26 10:22

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			223143	XTD8	EA POM	04/27/26 07:29
Total/NA	Analysis	533		1	223382	SZ9R	EA POM	04/28/26 09:01
Total/NA	Prep	537.1 DW			222326	G9MN	EA POM	04/24/26 01:00
Total/NA	Analysis	EPA 537.1 V2		1	222689	Y5FM	EA POM	04/24/26 13:18

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-209878-1
SDG: PFAS: Moanalua Wells

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-209878-1	MOANALUA WELLS (331-223-TP202)	Drinking Water	04/20/26 10:33	04/22/26 10:22	HI0000331
380-209878-2	FB: MOANALUA WELLS (331-223-TP202)	Water	04/20/26 10:33	04/22/26 10:22	HI0000331

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-209878-1
SDG Number: PFAS: Moanalua Wells

Login Number: 209878

List Number: 1

Creator: Del Rosario, Michael

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

