

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 3/4/2026 7:03:22 PM

JOB DESCRIPTION

RED-HILL
Weekly: Aiea Gulch Wells Pump 1
RUSH Weekly Red Hill

JOB NUMBER

380-198978-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
3/4/2026 7:03:22 PM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
QC Sample Results	15
QC Association Summary	31
Lab Chronicle	33
Certification Summary	34
Method Summary	36
Sample Summary	37
Chain of Custody	38
Receipt Checklists	41

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Qualifiers

GC/MS Semi VOA

Qualifier	Qualifier Description
*+	LCS and/or LCSD is outside acceptance limits, high biased.
F1	MS and/or MSD recovery exceeds control limits.
F2	MS/MSD RPD exceeds control limits
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

GC/MS Semi VOA TICs

Qualifier	Qualifier Description
J	Indicates an Estimated Value for TICs
N	Presumptive evidence of material.
T	Result is a tentatively identified compound (TIC) and an estimated value.

GC VOA

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-198978-1

Job ID: 380-198978-1

Eurofins Pomona

Job Narrative 380-198978-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 2/19/2026 9:29 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperatures of the 2 coolers at receipt time were 2.1°C and 3.6°C.

GC/MS Semi VOA

Method 625.1 SIM: The matrix spike / matrix spike duplicate (MS/MSD) recoveries for the following sample associated with preparation batch 570-698456 and analytical batch 570-701134 were outside control limits: (380-198981-C-1-A MS). The associated laboratory control sample (LCS) recovery met acceptance criteria.

Method 625.1 SIM: The matrix spike / matrix spike duplicate (MS/MSD) precision for preparation batch 570-698456 and analytical batch 570-701134 was outside control limits. Sample matrix interference and/or non-homogeneity are suspected because the associated laboratory control sample / laboratory control sample duplicate (LCS/LCSD) precision was within acceptance limits.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Gasoline Range Organics

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Diesel Range Organics

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**
PWSID Number: HI0000331

Lab Sample ID: 380-198978-1

No Detections.

**Client Sample ID: TB: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-198978-2

No Detections.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-198978-1

Date Collected: 02/17/26 10:46

Matrix: Drinking Water

Date Received: 02/19/26 09:29

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
2,4'-DDD	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
2,4'-DDE	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
2,4'-DDT	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
2,4-Dinitrotoluene	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
2,6-Dinitrotoluene	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
2-Methylnaphthalene	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
4,4'-DDD	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
4,4'-DDE	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
4,4'-DDT	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Acenaphthene	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Acenaphthylene	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Acetochlor	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Alachlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
alpha-BHC	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
alpha-Chlordane	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Anthracene	<0.020		0.020	ug/L		02/24/26 14:32	02/25/26 11:52	1
Atrazine	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Benz(a)anthracene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Benzo[a]pyrene	<0.020		0.020	ug/L		02/24/26 14:32	02/25/26 11:52	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		02/24/26 14:32	02/25/26 11:52	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		02/24/26 14:32	02/25/26 11:52	1
beta-BHC	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Bis(2-ethylhexyl) phthalate	<0.59	*+	0.59	ug/L		02/24/26 14:32	02/25/26 11:52	1
Bromacil	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Butachlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Butylbenzylphthalate	<0.49		0.49	ug/L		02/24/26 14:32	02/25/26 11:52	1
Chlorobenzilate	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Chloroneb	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Chlorothalonil (Draconil, Bravo)	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Chlorpyrifos	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Chrysene	<0.020		0.020	ug/L		02/24/26 14:32	02/25/26 11:52	1
delta-BHC	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		02/24/26 14:32	02/25/26 11:52	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Dieldrin	<0.0099		0.0099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Diethylphthalate	<0.49		0.49	ug/L		02/24/26 14:32	02/25/26 11:52	1
Dimethylphthalate	<0.49		0.49	ug/L		02/24/26 14:32	02/25/26 11:52	1
Di-n-butyl phthalate	<0.99		0.99	ug/L		02/24/26 14:32	02/25/26 11:52	1
Di-n-octyl phthalate	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Endosulfan I (Alpha)	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Endosulfan II (Beta)	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Endosulfan sulfate	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Endrin	<0.0099		0.0099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Endrin aldehyde	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
EPTC	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-198978-1

Date Collected: 02/17/26 10:46

Matrix: Drinking Water

Date Received: 02/19/26 09:29

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluoranthene	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Fluorene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
gamma-Chlordane	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Heptachlor	<0.0099		0.0099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Heptachlor epoxide (isomer B)	<0.0099		0.0099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Hexachlorobenzene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Isophorone	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Lindane	<0.0099		0.0099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Malathion	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Methoxychlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Metolachlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Molinate	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Naphthalene	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Parathion	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Pendimethalin (Penoxaline)	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Phenanthrene	<0.039		0.039	ug/L		02/24/26 14:32	02/25/26 11:52	1
Propachlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Pyrene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Simazine	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Terbacil	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Terbutylazine	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Thiobencarb	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		02/24/26 14:32	02/25/26 11:52	1
trans-Nonachlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 11:52	1
Trifluralin	<0.099		0.099	ug/L		02/24/26 14:32	02/25/26 11:52	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	02/24/26 14:32	02/25/26 11:52	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	103		70 - 130	02/24/26 14:32	02/25/26 11:52	1
Perylene-d12	101		70 - 130	02/24/26 14:32	02/25/26 11:52	1
Triphenylphosphate	114		70 - 130	02/24/26 14:32	02/25/26 11:52	1

Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
2-Methylnaphthalene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Acenaphthene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Acenaphthylene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Anthracene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Benzo[a]anthracene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Benzo[a]pyrene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Benzo[b]fluoranthene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Benzo[g,h,i]perylene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Benzo[k]fluoranthene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Chrysene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-198978-1

Date Collected: 02/17/26 10:46

Matrix: Drinking Water

Date Received: 02/19/26 09:29

PWSID Number: HI0000331

Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Dibenz(a,h)anthracene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Fluoranthene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Fluorene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Indeno[1,2,3-cd]pyrene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Naphthalene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Phenanthrene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Pyrene	<0.19		0.19	ug/L		02/20/26 07:00	02/25/26 20:45	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	73		28 - 127			02/20/26 07:00	02/25/26 20:45	1
2-Fluorobiphenyl (Surr)	64		31 - 120			02/20/26 07:00	02/25/26 20:45	1
2-Fluorophenol (Surr)	44		17 - 120			02/20/26 07:00	02/25/26 20:45	1
Nitrobenzene-d5 (Surr)	69		27 - 120			02/20/26 07:00	02/25/26 20:45	1
Phenol-d6 (Surr)	27		10 - 120			02/20/26 07:00	02/25/26 20:45	1
p-Terphenyl-d14 (Surr)	63		45 - 120			02/20/26 07:00	02/25/26 20:45	1

Method: EPA 625.1 - Semivolatile Organic Compounds (GC/MS)

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	02/20/26 07:00	03/02/26 11:05	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	81		33 - 139				02/20/26 07:00	03/02/26 11:05	1
2-Fluorobiphenyl (Surr)	70		33 - 126				02/20/26 07:00	03/02/26 11:05	1
2-Fluorophenol (Surr)	48		12 - 120				02/20/26 07:00	03/02/26 11:05	1
Nitrobenzene-d5 (Surr)	84		36 - 120				02/20/26 07:00	03/02/26 11:05	1
Phenol-d6 (Surr)	33		10 - 120				02/20/26 07:00	03/02/26 11:05	1
p-Terphenyl-d14 (Surr)	72		47 - 131				02/20/26 07:00	03/02/26 11:05	1

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/21/26 00:53	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	93		38 - 134				02/21/26 00:53	1

Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<27		27	ug/L		02/23/26 08:51	02/24/26 03:21	1
Motor Oil Range Organics [C24-C36]	<27		27	ug/L		02/23/26 08:51	02/24/26 03:21	1
C8-C18	<27		27	ug/L		02/23/26 08:51	02/24/26 03:21	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	101		60 - 130			02/23/26 08:51	02/24/26 03:21	1

Client Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-198978-1
 SDG: Weekly: Aiea Gulch Wells Pump 1

**Client Sample ID: TB: AIEA GULCH WELLS PUMP 1
 (331-201-TP071)**

Lab Sample ID: 380-198978-2

Date Collected: 02/17/26 10:46

Matrix: Water

Date Received: 02/19/26 09:29

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			02/20/26 23:20	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	92		38 - 134				02/20/26 23:20	1

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Action Limit Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-198978-1
 SDG: Weekly: Aiea Gulch Wells Pump 1

**Client Sample ID: AIEA GULCH WELLS PUMP 1
 (331-201-TP071)**
PWSID Number: HI0000331

Lab Sample ID: 380-198978-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Alachlor	<0.049		ug/L	2	0.049	525.2	Total/NA
Atrazine	<0.049		ug/L	3	0.049	525.2	Total/NA
Benzo[a]pyrene	<0.020		ug/L	0.2	0.020	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	<0.59	*+	ug/L	6	0.59	525.2	Total/NA
Di(2-ethylhexyl)adipate	<0.59		ug/L	400	0.59	525.2	Total/NA
Endrin	<0.0099		ug/L	2	0.0099	525.2	Total/NA
Heptachlor	<0.0099		ug/L	0.4	0.0099	525.2	Total/NA
Heptachlor epoxide (isomer B)	<0.0099		ug/L	0.2	0.0099	525.2	Total/NA
Hexachlorobenzene	<0.049		ug/L	1	0.049	525.2	Total/NA
Hexachlorocyclopentadiene	<0.049		ug/L	50	0.049	525.2	Total/NA
Lindane	<0.0099		ug/L	0.2	0.0099	525.2	Total/NA
Methoxychlor	<0.049		ug/L	40	0.049	525.2	Total/NA
Simazine	<0.049		ug/L	4	0.049	525.2	Total/NA
Benzo[a]pyrene	<0.19		ug/L	0.2	0.19	625.1 SIM	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	PRY (70-130)	TPP (70-130)
380-198978-1	AIEA GULCH WELLS PUMP 1 (331)	103	101	114

Surrogate Legend

2NMX = 2-Nitro-m-xylene
PRY = Perylene-d12
TPP = Triphenylphosphate

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	PRY (70-130)	TPP (70-130)
380-198936-AA-1-A MS	Matrix Spike	98	108	117
380-198975-I-1-A DU	Duplicate	99	105	121
LCS 380-208072/23-B	Lab Control Sample	98	107	111
MB 380-208072/21-B	Method Blank	97	98	108
MRL 380-208072/22-B	Lab Control Sample	98	97	109

Surrogate Legend

2NMX = 2-Nitro-m-xylene
PRY = Perylene-d12
TPP = Triphenylphosphate

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (33-139)	FBP (33-126)	2FP (12-120)	NBZ (36-120)	PHL6 (10-120)	TPHd14 (47-131)
380-198978-1	AIEA GULCH WELLS PUMP 1 (331)	81	70	48	84	33	72

Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)
FBP = 2-Fluorobiphenyl (Surr)
2FP = 2-Fluorophenol (Surr)
NBZ = Nitrobenzene-d5 (Surr)
PHL6 = Phenol-d6 (Surr)
TPHd14 = p-Terphenyl-d14 (Surr)

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (33-139)	FBP (33-126)	2FP (12-120)	NBZ (36-120)	PHL6 (10-120)	TPHd14 (47-131)
MB 570-698456/1-A	Method Blank	92	83	58	98	39	87

Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)
FBP = 2-Fluorobiphenyl (Surr)
2FP = 2-Fluorophenol (Surr)
NBZ = Nitrobenzene-d5 (Surr)

Surrogate Summary

Client: City & County of Honolulu

Job ID: 380-198978-1

Project/Site: RED-HILL

SDG: Weekly: Aiea Gulch Wells Pump 1

PHL6 = Phenol-d6 (Surr)

TPHd14 = p-Terphenyl-d14 (Surr)

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (28-127)	FBP (31-120)	2FP (17-120)	NBZ (27-120)	PHL6 (10-120)	TPHd14 (45-120)
380-198978-1	AIEA GULCH WELLS PUMP 1 (331)	73	64	44	69	27	63

Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)

FBP = 2-Fluorobiphenyl (Surr)

2FP = 2-Fluorophenol (Surr)

NBZ = Nitrobenzene-d5 (Surr)

PHL6 = Phenol-d6 (Surr)

TPHd14 = p-Terphenyl-d14 (Surr)

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (28-127)	FBP (31-120)	2FP (17-120)	NBZ (27-120)	PHL6 (10-120)	TPHd14 (45-120)
380-198981-C-1-A MS	Matrix Spike	75	67	58	58	38	74
380-198981-C-1-B MSD	Matrix Spike Duplicate	79	70	58	77	36	71
LCS 570-698456/2-A	Lab Control Sample	100	80	71	87	44	83
LCSd 570-698456/3-A	Lab Control Sample Dup	97	87	66	85	39	86
MB 570-698456/1-A	Method Blank	102	87	61	92	37	83

Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)

FBP = 2-Fluorobiphenyl (Surr)

2FP = 2-Fluorophenol (Surr)

NBZ = Nitrobenzene-d5 (Surr)

PHL6 = Phenol-d6 (Surr)

TPHd14 = p-Terphenyl-d14 (Surr)

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)
		BFB1 (38-134)
380-198978-1	AIEA GULCH WELLS PUMP 1 (331)	93

Surrogate Legend

BFB = 4-Bromofluorobenzene (Surr)

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)
		BFB1 (38-134)
380-198978-2	TB: AIEA GULCH WELLS PUMP 1 (92
380-198981-A-1 MS	Matrix Spike	89
380-198981-A-1 MSD	Matrix Spike Duplicate	94

Eurofins Pomona

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 8015B GRO LL - Gasoline Range Organics - (GC) (Continued)

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB1 (38-134)
LCS 570-698930/3	Lab Control Sample	84
LCSD 570-698930/4	Lab Control Sample Dup	92
MB 570-698930/5	Method Blank	92
MRL 570-698930/6	Lab Control Sample	89

Surrogate Legend

BFB = 4-Bromofluorobenzene (Surr)

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Drinking Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
380-198978-1	AIEA GULCH WELLS PUMP 1 (331)	101

Surrogate Legend

OTCSN = n-Octacosane (Surr)

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
380-198981-B-1-A MS	Matrix Spike	102
380-198981-B-1-B MSD	Matrix Spike Duplicate	108
LCS 570-699749/2-A	Lab Control Sample	110
LCSD 570-699749/3-A	Lab Control Sample Dup	114
MB 570-699749/1-A	Method Blank	94
MRL 570-699749/4-A	Lab Control Sample	87

Surrogate Legend

OTCSN = n-Octacosane (Surr)

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-198978-1
 SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 380-208072/21-B
Matrix: Water
Analysis Batch: 208615

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 208072

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
2,4'-DDD	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
2,4'-DDE	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
2,4'-DDT	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
2,4-Dinitrotoluene	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
2,6-Dinitrotoluene	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
2-Methylnaphthalene	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
4,4'-DDD	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
4,4'-DDE	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
4,4'-DDT	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Acenaphthene	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Acenaphthylene	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Acetochlor	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Alachlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
alpha-BHC	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
alpha-Chlordane	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Anthracene	<0.020		0.020	ug/L		02/24/26 14:32	02/25/26 09:51	1
Atrazine	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Benz(a)anthracene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Benzo[a]pyrene	<0.020		0.020	ug/L		02/24/26 14:32	02/25/26 09:51	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		02/24/26 14:32	02/25/26 09:51	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		02/24/26 14:32	02/25/26 09:51	1
beta-BHC	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		02/24/26 14:32	02/25/26 09:51	1
Bromacil	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Butachlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Butylbenzylphthalate	<0.49		0.49	ug/L		02/24/26 14:32	02/25/26 09:51	1
Chlorobenzilate	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Chloroneb	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Chlorothalonil (Draconil, Bravo)	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Chlorpyrifos	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Chrysene	<0.020		0.020	ug/L		02/24/26 14:32	02/25/26 09:51	1
delta-BHC	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		02/24/26 14:32	02/25/26 09:51	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Dieldrin	<0.0098		0.0098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Diethylphthalate	<0.49		0.49	ug/L		02/24/26 14:32	02/25/26 09:51	1
Dimethylphthalate	<0.49		0.49	ug/L		02/24/26 14:32	02/25/26 09:51	1
Di-n-butyl phthalate	<0.98		0.98	ug/L		02/24/26 14:32	02/25/26 09:51	1
Di-n-octyl phthalate	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Endosulfan I (Alpha)	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Endosulfan II (Beta)	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Endosulfan sulfate	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Endrin	<0.0098		0.0098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Endrin aldehyde	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
EPTC	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MB 380-208072/21-B
Matrix: Water
Analysis Batch: 208615

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 208072

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Fluoranthene	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Fluorene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
gamma-Chlordane	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Heptachlor	<0.0098		0.0098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Heptachlor epoxide (isomer B)	<0.0098		0.0098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Hexachlorobenzene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Isophorone	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Lindane	<0.0098		0.0098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Malathion	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Methoxychlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Metolachlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Molinate	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Naphthalene	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Parathion	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Pendimethalin (Penoxaline)	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Phenanthrene	<0.039		0.039	ug/L		02/24/26 14:32	02/25/26 09:51	1
Propachlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Pyrene	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Simazine	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Terbacil	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Terbutylazine	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Thiobencarb	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		02/24/26 14:32	02/25/26 09:51	1
trans-Nonachlor	<0.049		0.049	ug/L		02/24/26 14:32	02/25/26 09:51	1
Trifluralin	<0.098		0.098	ug/L		02/24/26 14:32	02/25/26 09:51	1

Tentatively Identified Compound	MB	MB	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
	Est. Result	Qualifier							
Undecane	5.11	T J N	ug/L		3.18	1120-21-4	02/24/26 14:32	02/25/26 09:51	1
Phenol, m-tert-butyl-	0.574	T J N	ug/L		3.92	585-34-2	02/24/26 14:32	02/25/26 09:51	1
9-Octadecenamide, (Z)-	3.40	T J N	ug/L		7.98	301-02-0	02/24/26 14:32	02/25/26 09:51	1
13-Docosenamide, (Z)-	0.943	T J N	ug/L		10.51	112-84-5	02/24/26 14:32	02/25/26 09:51	1

Surrogate	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
2-Nitro-m-xylene	97		70 - 130	02/24/26 14:32	02/25/26 09:51	1
Perylene-d12	98		70 - 130	02/24/26 14:32	02/25/26 09:51	1
Triphenylphosphate	108		70 - 130	02/24/26 14:32	02/25/26 09:51	1

Lab Sample ID: LCS 380-208072/23-B
Matrix: Water
Analysis Batch: 208615

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 208072

Analyte	Spike Added	LCS	LCS	Unit	D	%Rec	%Rec Limits
		Result	Qualifier				
1-Methylnaphthalene	1.97	2.02		ug/L		103	70 - 130
2,4'-DDD	1.97	2.33		ug/L		118	70 - 130
2,4'-DDE	1.97	2.44		ug/L		124	70 - 130
2,4'-DDT	1.97	2.37		ug/L		120	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-208072/23-B

Matrix: Water

Analysis Batch: 208615

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 208072

Analyte	Spike	LCS	LCS	Unit	D	%Rec	%Rec Limits
	Added	Result	Qualifier				
2,4-Dinitrotoluene	1.97	2.19		ug/L		111	70 - 130
2,6-Dinitrotoluene	1.97	2.07		ug/L		105	70 - 130
2-Methylnaphthalene	1.97	2.03		ug/L		103	70 - 130
4,4'-DDD	1.97	2.55		ug/L		129	70 - 130
4,4'-DDE	1.97	2.20		ug/L		112	70 - 130
4,4'-DDT	1.97	2.31		ug/L		117	70 - 130
Acenaphthene	1.97	2.03		ug/L		103	70 - 130
Acenaphthylene	1.97	2.10		ug/L		107	70 - 130
Acetochlor	1.97	2.36		ug/L		119	70 - 130
Alachlor	1.97	2.32		ug/L		118	70 - 130
alpha-BHC	1.97	2.10		ug/L		106	70 - 130
alpha-Chlordane	1.97	2.34		ug/L		119	70 - 130
Anthracene	1.97	1.87		ug/L		95	70 - 130
Atrazine	1.97	2.42		ug/L		123	70 - 130
Benz(a)anthracene	1.97	2.08		ug/L		106	70 - 130
Benzo[a]pyrene	1.97	2.38		ug/L		121	70 - 130
Benzo[b]fluoranthene	1.97	2.30		ug/L		116	70 - 130
Benzo[g,h,i]perylene	1.97	2.46		ug/L		125	70 - 130
Benzo[k]fluoranthene	1.97	2.42		ug/L		123	70 - 130
beta-BHC	1.97	2.15		ug/L		109	70 - 130
Bis(2-ethylhexyl) phthalate	1.97	2.69	*+	ug/L		136	70 - 130
Bromacil	1.97	2.15		ug/L		109	70 - 130
Butachlor	1.97	2.43		ug/L		123	70 - 130
Butylbenzylphthalate	1.97	2.43		ug/L		123	70 - 130
Chlorobenzilate	1.97	2.52		ug/L		128	70 - 130
Chloroneb	1.97	2.18		ug/L		111	70 - 130
Chlorothalonil (Draconil, Bravo)	1.97	2.25		ug/L		114	70 - 130
Chlorpyrifos	1.97	2.28		ug/L		116	70 - 130
Chrysene	1.97	2.08		ug/L		105	70 - 130
delta-BHC	1.97	2.12		ug/L		107	70 - 130
Di(2-ethylhexyl)adipate	1.97	2.48		ug/L		126	70 - 130
Dibenz(a,h)anthracene	1.97	2.35		ug/L		119	70 - 130
Diclorvos (DDVP)	1.97	2.21		ug/L		112	70 - 130
Dieldrin	1.97	2.40		ug/L		122	70 - 130
Diethylphthalate	1.97	2.30		ug/L		117	70 - 130
Dimethylphthalate	1.97	2.15		ug/L		109	70 - 130
Di-n-butyl phthalate	3.94	4.85		ug/L		123	70 - 130
Di-n-octyl phthalate	1.97	2.45		ug/L		124	70 - 130
Endosulfan I (Alpha)	1.97	2.18		ug/L		111	70 - 130
Endosulfan II (Beta)	1.97	2.07		ug/L		105	70 - 130
Endosulfan sulfate	1.97	2.51		ug/L		127	70 - 130
Endrin	1.97	2.55		ug/L		129	70 - 130
Endrin aldehyde	1.97	2.22		ug/L		113	60 - 130
EPTC	1.97	2.15		ug/L		109	70 - 130
Fluoranthene	1.97	2.17		ug/L		110	70 - 130
Fluorene	1.97	2.00		ug/L		101	70 - 130
gamma-Chlordane	1.97	2.31		ug/L		117	70 - 130
Heptachlor	1.97	2.30		ug/L		116	70 - 130
Heptachlor epoxide (isomer B)	1.97	2.16		ug/L		109	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-208072/23-B

Matrix: Water

Analysis Batch: 208615

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 208072

Analyte	Spike Added	LCS	LCS	Unit	D	%Rec	%Rec Limits
		Result	Qualifier				
Hexachlorobenzene	1.97	2.05		ug/L		104	70 - 130
Hexachlorocyclopentadiene	1.97	2.25		ug/L		114	70 - 130
Indeno[1,2,3-cd]pyrene	1.97	2.34		ug/L		118	70 - 130
Isophorone	1.97	2.02		ug/L		102	70 - 130
Lindane	1.97	2.30		ug/L		117	70 - 130
Malathion	1.97	2.42		ug/L		123	70 - 130
Methoxychlor	1.97	2.31		ug/L		117	70 - 130
Metolachlor	1.97	2.26		ug/L		115	70 - 130
Molinate	1.97	2.20		ug/L		111	70 - 130
Naphthalene	1.97	2.00		ug/L		101	70 - 130
Parathion	1.97	2.43		ug/L		123	70 - 130
Pendimethalin (Penoxaline)	1.97	2.47		ug/L		125	70 - 130
Phenanthrene	1.97	2.05		ug/L		104	70 - 130
Propachlor	1.97	2.27		ug/L		115	70 - 130
Pyrene	1.97	2.19		ug/L		111	70 - 130
Simazine	1.97	2.21		ug/L		112	70 - 130
Terbacil	1.97	2.21		ug/L		112	70 - 130
Terbutylazine	1.97	2.44		ug/L		124	70 - 130
Thiobencarb	1.97	2.35		ug/L		119	70 - 130
trans-Nonachlor	1.97	2.38		ug/L		121	70 - 130
Trifluralin	1.97	2.28		ug/L		116	70 - 130

Surrogate	LCS		Limits
	%Recovery	Qualifier	
2-Nitro-m-xylene	98		70 - 130
Perylene-d12	107		70 - 130
Triphenylphosphate	111		70 - 130

Lab Sample ID: MRL 380-208072/22-B

Matrix: Water

Analysis Batch: 208615

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 208072

Analyte	Spike Added	MRL	MRL	Unit	D	%Rec	%Rec Limits
		Result	Qualifier				
1-Methylnaphthalene	0.0984	0.0995		ug/L		101	50 - 150
2,4'-DDD	0.0984	0.0945	J	ug/L		96	50 - 150
2,4'-DDE	0.0984	0.0956	J	ug/L		97	50 - 150
2,4'-DDT	0.0984	0.110		ug/L		112	50 - 150
2,4-Dinitrotoluene	0.0984	0.0952	J	ug/L		97	50 - 150
2,6-Dinitrotoluene	0.0984	0.104		ug/L		106	50 - 150
2-Methylnaphthalene	0.0984	0.0952	J	ug/L		97	50 - 150
4,4'-DDD	0.0984	0.100		ug/L		102	50 - 150
4,4'-DDE	0.0984	0.0994		ug/L		101	50 - 150
4,4'-DDT	0.0984	0.118		ug/L		120	50 - 150
Acenaphthene	0.0984	0.0834	J	ug/L		85	50 - 150
Acenaphthylene	0.0984	0.0957	J	ug/L		97	50 - 150
Acetochlor	0.0984	0.109		ug/L		111	50 - 150
Alachlor	0.0492	0.0595		ug/L		121	50 - 150
alpha-BHC	0.0984	0.104		ug/L		106	50 - 150
alpha-Chlordane	0.0246	<0.029		ug/L		112	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-198978-1
 SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-208072/22-B

Matrix: Water

Analysis Batch: 208615

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 208072

Analyte	Spike	MRL	MRL	Unit	D	%Rec	%Rec Limits
	Added	Result	Qualifier				
Anthracene	0.0197	0.0206		ug/L		105	50 - 150
Atrazine	0.0492	0.0493		ug/L		100	50 - 150
Benz(a)anthracene	0.0492	0.0528		ug/L		107	50 - 150
Benzo[a]pyrene	0.0197	0.0221		ug/L		112	50 - 150
Benzo[b]fluoranthene	0.0197	0.0245		ug/L		125	50 - 150
Benzo[g,h,i]perylene	0.0492	0.0458	J	ug/L		93	50 - 150
Benzo[k]fluoranthene	0.0197	0.0224		ug/L		114	50 - 150
beta-BHC	0.0984	0.110		ug/L		112	50 - 150
Bis(2-ethylhexyl) phthalate	0.591	0.727		ug/L		123	50 - 150
Bromacil	0.0984	0.128		ug/L		130	50 - 150
Butachlor	0.0492	0.0601		ug/L		122	50 - 150
Butylbenzylphthalate	0.492	0.560		ug/L		114	50 - 150
Chlorobenzilate	0.0984	0.103		ug/L		105	50 - 150
Chloroneb	0.0984	0.0863	J	ug/L		88	50 - 150
Chlorothalonil (Draconil, Bravo)	0.0984	0.0988		ug/L		100	50 - 150
Chlorpyrifos	0.0492	0.0503		ug/L		102	50 - 150
Chrysene	0.0197	0.0205		ug/L		104	50 - 150
delta-BHC	0.0984	0.107		ug/L		108	50 - 150
Di(2-ethylhexyl)adipate	0.591	0.668		ug/L		113	50 - 150
Dibenz(a,h)anthracene	0.0492	0.0469	J	ug/L		95	50 - 150
Diclorvos (DDVP)	0.0492	0.0573		ug/L		116	50 - 150
Dieldrin	0.00984	0.0123		ug/L		125	50 - 150
Diethylphthalate	0.492	0.526		ug/L		107	50 - 150
Dimethylphthalate	0.492	0.482	J	ug/L		98	50 - 150
Di-n-butyl phthalate	0.492	0.527	J	ug/L		107	49 - 243
Di-n-octyl phthalate	0.0984	0.108		ug/L		110	50 - 150
Endosulfan I (Alpha)	0.0984	0.0881	J	ug/L		90	50 - 150
Endosulfan II (Beta)	0.0984	0.123		ug/L		125	50 - 150
Endosulfan sulfate	0.0984	0.108		ug/L		110	50 - 150
Endrin	0.00984	0.0106		ug/L		108	50 - 150
Endrin aldehyde	0.0984	0.108		ug/L		110	50 - 150
EPTC	0.0984	0.0984		ug/L		100	50 - 150
Fluoranthene	0.0984	0.0938	J	ug/L		95	50 - 150
Fluorene	0.0492	0.0507		ug/L		103	50 - 150
gamma-Chlordane	0.0246	0.0253	J	ug/L		103	50 - 150
Heptachlor	0.00984	0.0113		ug/L		114	50 - 150
Heptachlor epoxide (isomer B)	0.00984	0.0102		ug/L		104	50 - 150
Hexachlorobenzene	0.0492	0.0473	J	ug/L		96	50 - 150
Hexachlorocyclopentadiene	0.0492	0.0576		ug/L		117	50 - 150
Indeno[1,2,3-cd]pyrene	0.0492	0.0507		ug/L		103	50 - 150
Isophorone	0.0984	0.110		ug/L		112	50 - 150
Lindane	0.00984	0.0111		ug/L		113	50 - 150
Malathion	0.0984	0.101		ug/L		102	50 - 150
Methoxychlor	0.0492	0.0530		ug/L		108	50 - 150
Metolachlor	0.0492	0.0560		ug/L		114	50 - 150
Molinate	0.0984	0.0993		ug/L		101	50 - 150
Naphthalene	0.0984	0.0981		ug/L		100	50 - 150
Parathion	0.0984	0.0999		ug/L		101	50 - 150
Pendimethalin (Penoxaline)	0.0984	0.108		ug/L		109	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-208072/22-B

Matrix: Water

Analysis Batch: 208615

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 208072

Analyte	Spike	MRL	MRL	Unit	D	%Rec	%Rec Limits
	Added	Result	Qualifier				
Phenanthrene	0.0394	0.0351	J	ug/L		89	50 - 150
Propachlor	0.0492	0.0539		ug/L		109	50 - 150
Pyrene	0.0492	0.0517		ug/L		105	50 - 150
Simazine	0.0492	0.0465	J	ug/L		95	50 - 150
Terbacil	0.0984	0.110		ug/L		112	50 - 150
Terbutylazine	0.0984	0.0965	J	ug/L		98	50 - 150
Thiobencarb	0.0984	0.103		ug/L		104	50 - 150
trans-Nonachlor	0.0246	<0.026		ug/L		102	50 - 150
Trifluralin	0.0984	0.110		ug/L		112	50 - 150

Surrogate	MRL	MRL	Limits
	%Recovery	Qualifier	
2-Nitro-m-xylene	98		70 - 130
Perylene-d12	97		70 - 130
Triphenylphosphate	109		70 - 130

Lab Sample ID: 380-198936-AA-1-A MS

Matrix: Water

Analysis Batch: 208615

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 208072

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
1-Methylnaphthalene	<0.098		1.96	2.01		ug/L		103	70 - 130
2,4'-DDD	<0.098		1.96	1.55		ug/L		79	70 - 130
2,4'-DDE	<0.098		1.96	1.41		ug/L		72	70 - 130
2,4'-DDT	<0.098		1.96	1.69		ug/L		86	70 - 130
2,4-Dinitrotoluene	<0.098		1.96	2.18		ug/L		111	70 - 130
2,6-Dinitrotoluene	<0.098		1.96	2.16		ug/L		110	70 - 130
2-Methylnaphthalene	<0.098		1.96	2.00		ug/L		102	70 - 130
4,4'-DDD	<0.098		1.96	1.82		ug/L		93	70 - 130
4,4'-DDE	<0.098		1.96	1.46		ug/L		75	70 - 130
4,4'-DDT	<0.098		1.96	1.94		ug/L		99	70 - 130
Acenaphthene	<0.098		1.96	2.04		ug/L		104	70 - 130
Acenaphthylene	<0.098		1.96	2.11		ug/L		108	70 - 130
Acetochlor	<0.098		1.96	1.98		ug/L		101	70 - 130
Alachlor	<0.049		1.96	1.87		ug/L		95	70 - 130
alpha-BHC	<0.098		1.96	2.04		ug/L		104	70 - 130
alpha-Chlordane	<0.049		1.96	1.70		ug/L		87	70 - 130
Anthracene	<0.020		1.96	1.78		ug/L		91	70 - 130
Atrazine	<0.049		1.96	2.48		ug/L		126	70 - 130
Benz(a)anthracene	<0.049		1.96	2.25		ug/L		115	70 - 130
Benzo[a]pyrene	<0.020	F1	1.96	2.57	F1	ug/L		131	70 - 130
Benzo[b]fluoranthene	<0.020		1.96	2.43		ug/L		124	70 - 130
Benzo[g,h,i]perylene	<0.049		1.96	2.49		ug/L		127	70 - 130
Benzo[k]fluoranthene	<0.020		1.96	2.46		ug/L		125	70 - 130
beta-BHC	<0.098		1.96	2.07		ug/L		106	70 - 130
Bis(2-ethylhexyl) phthalate	<0.59	*+	1.96	2.26		ug/L		116	70 - 130
Bromacil	1.5		1.96	3.27		ug/L		91	70 - 130
Butachlor	<0.049	F1	1.96	1.35	F1	ug/L		69	70 - 130
Butylbenzylphthalate	<0.49		1.96	1.63		ug/L		83	70 - 130

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-198978-1
 SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-198936-AA-1-A MS

Client Sample ID: Matrix Spike

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 208615

Prep Batch: 208072

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
Chlorobenzilate	<0.098	F1	1.96	1.24	F1	ug/L		63	70 - 130
Chloroneb	<0.098		1.96	2.21		ug/L		113	70 - 130
Chlorothalonil (Draconil, Bravo)	<0.098		1.96	2.03		ug/L		104	70 - 130
Chlorpyrifos	<0.049	F1	1.96	1.35	F1	ug/L		69	70 - 130
Chrysene	<0.020		1.96	2.23		ug/L		114	70 - 130
delta-BHC	<0.098		1.96	1.99		ug/L		102	70 - 130
Di(2-ethylhexyl)adipate	<0.59		1.96	1.66		ug/L		85	70 - 130
Dibenz(a,h)anthracene	<0.049		1.96	2.36		ug/L		120	70 - 130
Diclorvos (DDVP)	<0.049		1.96	2.20		ug/L		112	70 - 130
Dieldrin	<0.0098	F1	1.96	0.909	F1	ug/L		46	70 - 130
Diethylphthalate	<0.49		1.96	2.31		ug/L		118	70 - 130
Dimethylphthalate	<0.49		1.96	2.15		ug/L		110	70 - 130
Di-n-butyl phthalate	<0.98		3.92	5.12		ug/L		110	70 - 130
Di-n-octyl phthalate	<0.098		1.96	2.17		ug/L		111	70 - 130
Endosulfan I (Alpha)	<0.098	F1	1.96	0.428	F1	ug/L		22	70 - 130
Endosulfan II (Beta)	<0.098	F1	1.96	1.26	F1	ug/L		64	70 - 130
Endosulfan sulfate	<0.098		1.96	2.07		ug/L		106	70 - 130
Endrin	<0.0098	F1	1.96	1.19	F1	ug/L		61	70 - 130
Endrin aldehyde	<0.098		1.96	2.46		ug/L		126	60 - 130
EPTC	<0.098		1.96	2.16		ug/L		110	70 - 130
Fluoranthene	<0.098		1.96	1.43		ug/L		73	70 - 130
Fluorene	<0.049		1.96	2.06		ug/L		105	70 - 130
gamma-Chlordane	<0.049		1.96	1.73		ug/L		88	70 - 130
Heptachlor	<0.0098		1.96	2.13		ug/L		109	70 - 130
Heptachlor epoxide (isomer B)	<0.0098		1.96	1.86		ug/L		95	70 - 130
Hexachlorobenzene	<0.049		1.96	2.07		ug/L		106	70 - 130
Hexachlorocyclopentadiene	<0.049		1.96	2.19		ug/L		112	70 - 130
Indeno[1,2,3-cd]pyrene	<0.049		1.96	2.33		ug/L		119	70 - 130
Isophorone	<0.098		1.96	1.98		ug/L		101	70 - 130
Lindane	<0.0098		1.96	2.18		ug/L		111	70 - 130
Malathion	<0.098		1.96	1.86		ug/L		95	70 - 130
Methoxychlor	<0.049		1.96	2.43		ug/L		124	70 - 130
Metolachlor	<0.049		1.96	1.72		ug/L		88	70 - 130
Molinate	<0.098		1.96	2.23		ug/L		114	70 - 130
Naphthalene	<0.098		1.96	2.00		ug/L		102	70 - 130
Parathion	<0.098	F1	1.96	2.58	F1	ug/L		132	70 - 130
Pendimethalin (Penoxaline)	<0.098		1.96	1.88		ug/L		96	70 - 130
Phenanthrene	<0.039		1.96	2.01		ug/L		103	70 - 130
Propachlor	<0.049		1.96	2.29		ug/L		117	70 - 130
Pyrene	<0.049		1.96	1.45		ug/L		74	70 - 130
Simazine	<0.049		1.96	2.24		ug/L		114	70 - 130
Terbacil	<0.098		1.96	2.06		ug/L		105	70 - 130
Terbutylazine	<0.098		1.96	2.45		ug/L		125	70 - 130
Thiobencarb	<0.098		1.96	1.62		ug/L		83	70 - 130
trans-Nonachlor	<0.049		1.96	2.04		ug/L		104	70 - 130
Trifluralin	<0.098		1.96	2.28		ug/L		117	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-198936-AA-1-A MS

Matrix: Water

Analysis Batch: 208615

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 208072

Surrogate	MS MS		Limits
	%Recovery	Qualifier	
2-Nitro-m-xylene	98		70 - 130
Perylene-d12	108		70 - 130
Triphenylphosphate	117		70 - 130

Lab Sample ID: 380-198975-I-1-A DU

Matrix: Water

Analysis Batch: 208615

Client Sample ID: Duplicate

Prep Type: Total/NA

Prep Batch: 208072

Analyte	Sample	Sample	DU	DU	Unit	D	RPD	Limit
	Result	Qualifier	Result	Qualifier				
1-Methylnaphthalene	<0.10		<0.099		ug/L		NC	20
2,4'-DDD	<0.10		<0.099		ug/L		NC	20
2,4'-DDE	<0.10		<0.099		ug/L		NC	20
2,4'-DDT	<0.10		<0.099		ug/L		NC	20
2,4-Dinitrotoluene	<0.10		<0.099		ug/L		NC	20
2,6-Dinitrotoluene	<0.10		<0.099		ug/L		NC	20
2-Methylnaphthalene	<0.10		<0.099		ug/L		NC	20
4,4'-DDD	<0.10		<0.099		ug/L		NC	20
4,4'-DDE	<0.10		<0.099		ug/L		NC	20
4,4'-DDT	<0.10		<0.099		ug/L		NC	20
Acenaphthene	<0.10		<0.099		ug/L		NC	20
Acenaphthylene	<0.10		<0.099		ug/L		NC	20
Acetochlor	<0.10		<0.099		ug/L		NC	20
Alachlor	<0.050		<0.049		ug/L		NC	20
alpha-BHC	<0.10		<0.099		ug/L		NC	20
alpha-Chlordane	<0.050		<0.049		ug/L		NC	20
Anthracene	<0.020		<0.020		ug/L		NC	20
Atrazine	<0.050		<0.049		ug/L		NC	20
Benz(a)anthracene	<0.050		<0.049		ug/L		NC	20
Benzo[a]pyrene	<0.020		<0.020		ug/L		NC	20
Benzo[b]fluoranthene	<0.020		<0.020		ug/L		NC	20
Benzo[g,h,i]perylene	<0.050		<0.049		ug/L		NC	20
Benzo[k]fluoranthene	<0.020		<0.020		ug/L		NC	20
beta-BHC	<0.10		<0.099		ug/L		NC	20
Bis(2-ethylhexyl) phthalate	<0.60	*+	<0.59	*+	ug/L		NC	20
Bromacil	<0.10		<0.099		ug/L		NC	20
Butachlor	<0.050		<0.049		ug/L		NC	20
Butylbenzylphthalate	<0.50		<0.49		ug/L		NC	20
Chlorobenzilate	<0.10		<0.099		ug/L		NC	20
Chloroneb	<0.10		<0.099		ug/L		NC	20
Chlorothalonil (Draconil, Bravo)	<0.10		<0.099		ug/L		NC	20
Chlorpyrifos	<0.050		<0.049		ug/L		NC	20
Chrysene	<0.020		<0.020		ug/L		NC	20
delta-BHC	<0.10		<0.099		ug/L		NC	20
Di(2-ethylhexyl)adipate	<0.60		<0.59		ug/L		NC	20
Dibenz(a,h)anthracene	<0.050		<0.049		ug/L		NC	20
Diclorvos (DDVP)	<0.050		<0.049		ug/L		NC	20
Dieldrin	0.029		0.0247		ug/L		15	20
Diethylphthalate	<0.50		<0.49		ug/L		NC	20

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-198978-1
 SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-198975-I-1-A DU

Matrix: Water

Analysis Batch: 208615

Client Sample ID: Duplicate

Prep Type: Total/NA

Prep Batch: 208072

Analyte	Sample	Sample	DU	DU	Unit	D	RPD	Limit
	Result	Qualifier	Result	Qualifier				
Dimethylphthalate	<0.50		<0.49		ug/L		NC	20
Di-n-butyl phthalate	<1.0		<0.99		ug/L		NC	20
Di-n-octyl phthalate	<0.10		<0.099		ug/L		NC	20
Endosulfan I (Alpha)	<0.10		<0.099		ug/L		NC	20
Endosulfan II (Beta)	<0.10		<0.099		ug/L		NC	20
Endosulfan sulfate	<0.10		<0.099		ug/L		NC	20
Endrin	<0.010		<0.0099		ug/L		NC	20
Endrin aldehyde	<0.10		<0.099		ug/L		NC	20
EPTC	<0.10		<0.099		ug/L		NC	20
Fluoranthene	<0.10		<0.099		ug/L		NC	20
Fluorene	<0.050		<0.049		ug/L		NC	20
gamma-Chlordane	<0.050		<0.049		ug/L		NC	20
Heptachlor	<0.010		<0.0099		ug/L		NC	20
Heptachlor epoxide (isomer B)	<0.010		<0.0099		ug/L		NC	20
Hexachlorobenzene	<0.050		<0.049		ug/L		NC	20
Hexachlorocyclopentadiene	<0.050		<0.049		ug/L		NC	20
Indeno[1,2,3-cd]pyrene	<0.050		<0.049		ug/L		NC	20
Isophorone	<0.10		<0.099		ug/L		NC	20
Lindane	<0.010		<0.0099		ug/L		NC	20
Malathion	<0.10		<0.099		ug/L		NC	20
Methoxychlor	<0.050		<0.049		ug/L		NC	20
Metolachlor	<0.050		<0.049		ug/L		NC	20
Molinate	<0.10		<0.099		ug/L		NC	20
Naphthalene	<0.10		<0.099		ug/L		NC	20
Parathion	<0.10		<0.099		ug/L		NC	20
Pendimethalin (Penoxaline)	<0.10		<0.099		ug/L		NC	20
Phenanthrene	<0.040		<0.039		ug/L		NC	20
Propachlor	<0.050		<0.049		ug/L		NC	20
Pyrene	<0.050		<0.049		ug/L		NC	20
Simazine	<0.050		<0.049		ug/L		NC	20
Terbacil	<0.10		<0.099		ug/L		NC	20
Terbutylazine	<0.10		<0.099		ug/L		NC	20
Thiobencarb	<0.10		<0.099		ug/L		NC	20
Total Permethrin (mixed isomers)	<0.20		<0.20		ug/L		NC	20
trans-Nonachlor	<0.050		<0.049		ug/L		NC	20
Trifluralin	<0.10		<0.099		ug/L		NC	20

Surrogate	DU		Limits
	%Recovery	Qualifier	
2-Nitro-m-xylene	99		70 - 130
Perylene-d12	105		70 - 130
Triphenylphosphate	121		70 - 130

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-198978-1
 SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 570-698456/1-A
Matrix: Water
Analysis Batch: 702941

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 698456

<i>Tentatively Identified Compound</i>	<i>Est. Result</i>	<i>Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>RT</i>	<i>CAS No.</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
<i>Tentatively Identified Compound</i>	<i>None</i>		<i>ug/L</i>			<i>N/A</i>	<i>02/19/26 18:39</i>	<i>03/02/26 13:06</i>	<i>1</i>

<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
<i>2,4,6-Tribromophenol (Surr)</i>	<i>92</i>		<i>33 - 139</i>	<i>02/19/26 18:39</i>	<i>03/02/26 13:06</i>	<i>1</i>
<i>2-Fluorobiphenyl (Surr)</i>	<i>83</i>		<i>33 - 126</i>	<i>02/19/26 18:39</i>	<i>03/02/26 13:06</i>	<i>1</i>
<i>2-Fluorophenol (Surr)</i>	<i>58</i>		<i>12 - 120</i>	<i>02/19/26 18:39</i>	<i>03/02/26 13:06</i>	<i>1</i>
<i>Nitrobenzene-d5 (Surr)</i>	<i>98</i>		<i>36 - 120</i>	<i>02/19/26 18:39</i>	<i>03/02/26 13:06</i>	<i>1</i>
<i>Phenol-d6 (Surr)</i>	<i>39</i>		<i>10 - 120</i>	<i>02/19/26 18:39</i>	<i>03/02/26 13:06</i>	<i>1</i>
<i>p-Terphenyl-d14 (Surr)</i>	<i>87</i>		<i>47 - 131</i>	<i>02/19/26 18:39</i>	<i>03/02/26 13:06</i>	<i>1</i>

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Lab Sample ID: MB 570-698456/1-A
Matrix: Water
Analysis Batch: 700718

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 698456

<i>Analyte</i>	<i>Result</i>	<i>Qualifier</i>	<i>RL</i>	<i>Unit</i>	<i>D</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
<i>1-Methylnaphthalene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>2-Methylnaphthalene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Acenaphthene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Acenaphthylene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Anthracene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Benzo[a]anthracene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Benzo[a]pyrene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Benzo[b]fluoranthene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Benzo[g,h,i]perylene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Benzo[k]fluoranthene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Chrysene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Dibenz(a,h)anthracene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Fluoranthene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Fluorene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Indeno[1,2,3-cd]pyrene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Naphthalene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Phenanthrene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Pyrene</i>	<i><0.20</i>		<i>0.20</i>	<i>ug/L</i>		<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>

<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
<i>2,4,6-Tribromophenol (Surr)</i>	<i>102</i>		<i>28 - 127</i>	<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>2-Fluorobiphenyl (Surr)</i>	<i>87</i>		<i>31 - 120</i>	<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>2-Fluorophenol (Surr)</i>	<i>61</i>		<i>17 - 120</i>	<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Nitrobenzene-d5 (Surr)</i>	<i>92</i>		<i>27 - 120</i>	<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>Phenol-d6 (Surr)</i>	<i>37</i>		<i>10 - 120</i>	<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>
<i>p-Terphenyl-d14 (Surr)</i>	<i>83</i>		<i>45 - 120</i>	<i>02/19/26 18:39</i>	<i>02/25/26 03:53</i>	<i>1</i>

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: LCS 570-698456/2-A

Matrix: Water

Analysis Batch: 700718

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 698456

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits	
1-Methylnaphthalene	20.0	18.8		ug/L		94	47 - 120	
2-Methylnaphthalene	20.0	18.6		ug/L		93	43 - 120	
Acenaphthene	20.0	18.3		ug/L		91	60 - 132	
Acenaphthylene	20.0	17.9		ug/L		89	54 - 126	
Anthracene	20.0	19.4		ug/L		97	43 - 120	
Benzo[a]anthracene	20.0	19.8		ug/L		99	42 - 133	
Benzo[a]pyrene	20.0	21.7		ug/L		109	32 - 148	
Benzo[b]fluoranthene	20.0	20.5		ug/L		102	42 - 140	
Benzo[g,h,i]perylene	20.0	20.6		ug/L		103	1 - 195	
Benzo[k]fluoranthene	20.0	19.3		ug/L		97	25 - 146	
Chrysene	20.0	19.2		ug/L		96	44 - 140	
Dibenz(a,h)anthracene	20.0	21.2		ug/L		106	1 - 200	
Fluoranthene	20.0	20.1		ug/L		101	43 - 121	
Fluorene	20.0	19.1		ug/L		96	70 - 120	
Indeno[1,2,3-cd]pyrene	20.0	22.1		ug/L		111	1 - 151	
Naphthalene	20.0	18.8		ug/L		94	36 - 120	
Phenanthrene	20.0	19.8		ug/L		99	65 - 120	
Pyrene	20.0	20.2		ug/L		101	70 - 120	

Surrogate	LCS LCS		Limits
	%Recovery	Qualifier	
2,4,6-Tribromophenol (Surr)	100		28 - 127
2-Fluorobiphenyl (Surr)	80		31 - 120
2-Fluorophenol (Surr)	71		17 - 120
Nitrobenzene-d5 (Surr)	87		27 - 120
Phenol-d6 (Surr)	44		10 - 120
p-Terphenyl-d14 (Surr)	83		45 - 120

Lab Sample ID: LCSD 570-698456/3-A

Matrix: Water

Analysis Batch: 700718

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 698456

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits		RPD	
									RPD	Limit
1-Methylnaphthalene	20.0	17.8		ug/L		89	47 - 120	6	20	
2-Methylnaphthalene	20.0	17.6		ug/L		88	43 - 120	6	20	
Acenaphthene	20.0	18.1		ug/L		90	60 - 132	1	29	
Acenaphthylene	20.0	18.2		ug/L		91	54 - 126	2	45	
Anthracene	20.0	18.6		ug/L		93	43 - 120	4	40	
Benzo[a]anthracene	20.0	19.4		ug/L		97	42 - 133	2	32	
Benzo[a]pyrene	20.0	21.2		ug/L		106	32 - 148	2	43	
Benzo[b]fluoranthene	20.0	20.3		ug/L		102	42 - 140	1	43	
Benzo[g,h,i]perylene	20.0	19.8		ug/L		99	1 - 195	4	61	
Benzo[k]fluoranthene	20.0	19.0		ug/L		95	25 - 146	2	38	
Chrysene	20.0	18.9		ug/L		94	44 - 140	2	53	
Dibenz(a,h)anthracene	20.0	20.3		ug/L		102	1 - 200	4	75	
Fluoranthene	20.0	19.5		ug/L		98	43 - 121	3	40	
Fluorene	20.0	19.3		ug/L		97	70 - 120	1	23	
Indeno[1,2,3-cd]pyrene	20.0	21.6		ug/L		108	1 - 151	2	60	
Naphthalene	20.0	17.7		ug/L		89	36 - 120	6	39	

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: LCSD 570-698456/3-A

Client Sample ID: Lab Control Sample Dup

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 700718

Prep Batch: 698456

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Phenanthrene	20.0	19.0		ug/L		95	65 - 120	4	24
Pyrene	20.0	19.7		ug/L		98	70 - 120	2	30

Surrogate	LCSD		Limits
	%Recovery	Qualifier	
2,4,6-Tribromophenol (Surr)	97		28 - 127
2-Fluorobiphenyl (Surr)	87		31 - 120
2-Fluorophenol (Surr)	66		17 - 120
Nitrobenzene-d5 (Surr)	85		27 - 120
Phenol-d6 (Surr)	39		10 - 120
p-Terphenyl-d14 (Surr)	86		45 - 120

Lab Sample ID: 380-198981-C-1-A MS

Client Sample ID: Matrix Spike

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 701134

Prep Batch: 698456

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	<0.20	F2	39.0	22.1		ug/L		57	36 - 120
2-Methylnaphthalene	<0.20	F2	39.0	21.8		ug/L		56	32 - 124
Acenaphthene	<0.20	F2	39.0	25.6		ug/L		66	47 - 145
Acenaphthylene	<0.20		39.0	24.6		ug/L		63	33 - 145
Anthracene	<0.20		39.0	25.7		ug/L		66	27 - 133
Benzo[a]anthracene	<0.20	F2 F1	39.0	28.1		ug/L		72	33 - 143
Benzo[a]pyrene	<0.20		39.0	30.0		ug/L		77	17 - 163
Benzo[b]fluoranthene	<0.20		39.0	28.9		ug/L		74	24 - 159
Benzo[g,h,i]perylene	<0.20		39.0	27.6		ug/L		71	1 - 219
Benzo[k]fluoranthene	<0.20		39.0	26.4		ug/L		68	11 - 162
Chrysene	<0.20		39.0	27.2		ug/L		70	17 - 168
Dibenz(a,h)anthracene	<0.20		39.0	28.6		ug/L		73	1 - 227
Fluoranthene	<0.20	F1	39.0	27.6		ug/L		71	26 - 137
Fluorene	<0.20	F2 F1	39.0	27.2		ug/L		70	59 - 121
Indeno[1,2,3-cd]pyrene	<0.20		39.0	29.7		ug/L		76	1 - 171
Naphthalene	<0.20		39.0	20.1		ug/L		52	21 - 133
Phenanthrene	<0.20	F2 F1	39.0	26.4		ug/L		68	54 - 120
Pyrene	<0.20	F2 F1	39.0	28.9		ug/L		74	52 - 120

Surrogate	MS		Limits
	%Recovery	Qualifier	
2,4,6-Tribromophenol (Surr)	75		28 - 127
2-Fluorobiphenyl (Surr)	67		31 - 120
2-Fluorophenol (Surr)	58		17 - 120
Nitrobenzene-d5 (Surr)	58		27 - 120
Phenol-d6 (Surr)	38		10 - 120
p-Terphenyl-d14 (Surr)	74		45 - 120

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: 380-198981-C-1-B MSD

Matrix: Water

Analysis Batch: 701134

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 698456

Analyte	Sample	Sample	Spike	MSD	MSD	Unit	D	%Rec	%Rec	RPD	RPD
	Result	Qualifier	Added	Result	Qualifier				Limits		
1-Methylnaphthalene	<0.20	F2	19.8	15.9	F2	ug/L		81	36 - 120	33	30
2-Methylnaphthalene	<0.20	F2	19.8	15.2	F2	ug/L		77	32 - 124	36	30
Acenaphthene	<0.20	F2	19.8	14.2	F2	ug/L		72	47 - 145	57	48
Acenaphthylene	<0.20		19.8	13.8		ug/L		70	33 - 145	56	74
Anthracene	<0.20		19.8	15.1		ug/L		76	27 - 133	52	66
Benzo[a]anthracene	<0.20	F2 F1	19.8	15.4	F2	ug/L		78	33 - 143	59	53
Benzo[a]pyrene	<0.20		19.8	15.8		ug/L		80	17 - 163	62	72
Benzo[b]fluoranthene	<0.20		19.8	15.2		ug/L		77	24 - 159	62	71
Benzo[g,h,i]perylene	<0.20		19.8	15.8		ug/L		80	1 - 219	54	97
Benzo[k]fluoranthene	<0.20		19.8	14.8		ug/L		75	11 - 162	56	63
Chrysene	<0.20		19.8	14.9		ug/L		75	17 - 168	59	87
Dibenz(a,h)anthracene	<0.20		19.8	16.2		ug/L		82	1 - 227	55	126
Fluoranthene	<0.20	F1	19.8	15.9		ug/L		80	26 - 137	54	66
Fluorene	<0.20	F2 F1	19.8	14.7	F2	ug/L		74	59 - 121	60	38
Indeno[1,2,3-cd]pyrene	<0.20		19.8	16.9		ug/L		85	1 - 171	55	99
Naphthalene	<0.20		19.8	15.2		ug/L		77	21 - 133	28	65
Phenanthrene	<0.20	F2 F1	19.8	15.3	F2	ug/L		77	54 - 120	53	39
Pyrene	<0.20	F2 F1	19.8	15.7	F2	ug/L		80	52 - 120	59	49

Surrogate	MSD	MSD	Limits
	%Recovery	Qualifier	
2,4,6-Tribromophenol (Surr)	79		28 - 127
2-Fluorobiphenyl (Surr)	70		31 - 120
2-Fluorophenol (Surr)	58		17 - 120
Nitrobenzene-d5 (Surr)	77		27 - 120
Phenol-d6 (Surr)	36		10 - 120
p-Terphenyl-d14 (Surr)	71		45 - 120

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Lab Sample ID: MB 570-698930/5

Matrix: Water

Analysis Batch: 698930

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
GRO (C6-C10)	<10		10	ug/L			02/20/26 14:57	1

Surrogate	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
4-Bromofluorobenzene (Surr)	92		38 - 134		02/20/26 14:57	1

Lab Sample ID: LCS 570-698930/3

Matrix: Water

Analysis Batch: 698930

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS	LCS	Unit	D	%Rec	%Rec
		Result	Qualifier				Limits
Gasoline Range Organics (C4-C13)	400	413		ug/L		103	78 - 120

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 8015B GRO LL - Gasoline Range Organics - (GC) (Continued)

Lab Sample ID: LCS 570-698930/3
Matrix: Water
Analysis Batch: 698930

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

	LCS	LCS	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	84		38 - 134

Lab Sample ID: LCSD 570-698930/4
Matrix: Water
Analysis Batch: 698930

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Gasoline Range Organics (C4-C13)	400	395		ug/L		99	78 - 120	5	10

	LCSD	LCSD	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	92		38 - 134

Lab Sample ID: MRL 570-698930/6
Matrix: Water
Analysis Batch: 698930

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	10.0	8.41	J	ug/L		84	50 - 150

	MRL	MRL	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	89		38 - 134

Lab Sample ID: 380-198981-A-1 MS
Matrix: Water
Analysis Batch: 698930

Client Sample ID: Matrix Spike
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	<10		400	383		ug/L		96	68 - 122

	MS	MS	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	89		38 - 134

Lab Sample ID: 380-198981-A-1 MSD
Matrix: Water
Analysis Batch: 698930

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Gasoline Range Organics (C4-C13)	<10		400	401		ug/L		100	68 - 122	5	18

	MSD	MSD	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	94		38 - 134

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Lab Sample ID: MB 570-699749/1-A
Matrix: Water
Analysis Batch: 700123

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 699749

Analyte	MB MB		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Diesel Range Organics (C10-C24)	<25		25	ug/L		02/23/26 08:51	02/24/26 00:27	1
Motor Oil Range Organics [C24-C36]	<25		25	ug/L		02/23/26 08:51	02/24/26 00:27	1
C8-C18	<25		25	ug/L		02/23/26 08:51	02/24/26 00:27	1
Surrogate	MB MB		Limits			Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier						
n-Octacosane (Surr)	94		60 - 130			02/23/26 08:51	02/24/26 00:27	1

Lab Sample ID: LCS 570-699749/2-A
Matrix: Water
Analysis Batch: 700123

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 699749

Analyte	Spike Added	LCS LCS		Unit	D	%Rec	%Rec Limits
		Result	Qualifier				
C10-C28	1600	1600		ug/L		100	56 - 127
Surrogate	LCS LCS		Limits				
	%Recovery	Qualifier					
n-Octacosane (Surr)	110		60 - 130				

Lab Sample ID: LCSD 570-699749/3-A
Matrix: Water
Analysis Batch: 700123

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 699749

Analyte	Spike Added	LCSD LCSD		Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
		Result	Qualifier						
C10-C28	1600	1800		ug/L		113	56 - 127	12	23
Surrogate	LCSD LCSD		Limits						
	%Recovery	Qualifier							
n-Octacosane (Surr)	114		60 - 130						

Lab Sample ID: MRL 570-699749/4-A
Matrix: Water
Analysis Batch: 700123

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 699749

Analyte	Spike Added	MRL MRL		Unit	D	%Rec	%Rec Limits
		Result	Qualifier				
C10-C28	0.0200	0.0267		mg/L		134	50 - 150
Surrogate	MRL MRL		Limits				
	%Recovery	Qualifier					
n-Octacosane (Surr)	87		60 - 130				

Lab Sample ID: 380-198981-B-1-A MS
Matrix: Water
Analysis Batch: 700123

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 699749

Analyte	Sample Result	Sample Qualifier	Spike Added	MS MS		Unit	D	%Rec	%Rec Limits
				Result	Qualifier				
C10-C28	<26		1660	1710		ug/L		103	70 - 130
Surrogate	MS MS		Limits						
	%Recovery	Qualifier							
n-Octacosane (Surr)	102		60 - 130						

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-198978-1
 SDG: Weekly: Aiea Gulch Wells Pump 1

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level (Continued)

Lab Sample ID: 380-198981-B-1-B MSD

Matrix: Water

Analysis Batch: 700123

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 699749

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	<26		1660	1690		ug/L		102	70 - 130	1	20
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
<i>n-Octacosane (Surr)</i>	108		60 - 130								

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

QC Association Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-198978-1
 SDG: Weekly: Aiea Gulch Wells Pump 1

GC/MS Semi VOA

Prep Batch: 208072

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-198978-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	525.2	
MB 380-208072/21-B	Method Blank	Total/NA	Water	525.2	
LCS 380-208072/23-B	Lab Control Sample	Total/NA	Water	525.2	
MRL 380-208072/22-B	Lab Control Sample	Total/NA	Water	525.2	
380-198936-AA-1-A MS	Matrix Spike	Total/NA	Water	525.2	
380-198975-I-1-A DU	Duplicate	Total/NA	Water	525.2	

Analysis Batch: 208615

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-198978-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	525.2	208072
MB 380-208072/21-B	Method Blank	Total/NA	Water	525.2	208072
LCS 380-208072/23-B	Lab Control Sample	Total/NA	Water	525.2	208072
MRL 380-208072/22-B	Lab Control Sample	Total/NA	Water	525.2	208072
380-198936-AA-1-A MS	Matrix Spike	Total/NA	Water	525.2	208072
380-198975-I-1-A DU	Duplicate	Total/NA	Water	525.2	208072

Prep Batch: 698456

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-198978-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	625.1	
MB 570-698456/1-A	Method Blank	Total/NA	Water	625.1	
LCS 570-698456/2-A	Lab Control Sample	Total/NA	Water	625.1	
LCSD 570-698456/3-A	Lab Control Sample Dup	Total/NA	Water	625.1	
380-198981-C-1-A MS	Matrix Spike	Total/NA	Water	625.1	
380-198981-C-1-B MSD	Matrix Spike Duplicate	Total/NA	Water	625.1	

Analysis Batch: 700718

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
MB 570-698456/1-A	Method Blank	Total/NA	Water	625.1 SIM	698456
LCS 570-698456/2-A	Lab Control Sample	Total/NA	Water	625.1 SIM	698456
LCSD 570-698456/3-A	Lab Control Sample Dup	Total/NA	Water	625.1 SIM	698456

Analysis Batch: 701134

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-198978-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	625.1 SIM	698456
380-198981-C-1-A MS	Matrix Spike	Total/NA	Water	625.1 SIM	698456
380-198981-C-1-B MSD	Matrix Spike Duplicate	Total/NA	Water	625.1 SIM	698456

Analysis Batch: 702941

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-198978-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	625.1	698456
MB 570-698456/1-A	Method Blank	Total/NA	Water	625.1	698456

GC VOA

Analysis Batch: 698930

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-198978-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	8015B GRO LL	
380-198978-2	TB: AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Water	8015B GRO LL	
MB 570-698930/5	Method Blank	Total/NA	Water	8015B GRO LL	
LCS 570-698930/3	Lab Control Sample	Total/NA	Water	8015B GRO LL	
LCSD 570-698930/4	Lab Control Sample Dup	Total/NA	Water	8015B GRO LL	

QC Association Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-198978-1
 SDG: Weekly: Aiea Gulch Wells Pump 1

GC VOA (Continued)

Analysis Batch: 698930 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
MRL 570-698930/6	Lab Control Sample	Total/NA	Water	8015B GRO LL	
380-198981-A-1 MS	Matrix Spike	Total/NA	Water	8015B GRO LL	
380-198981-A-1 MSD	Matrix Spike Duplicate	Total/NA	Water	8015B GRO LL	

GC Semi VOA

Prep Batch: 699749

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-198978-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	3510C	
MB 570-699749/1-A	Method Blank	Total/NA	Water	3510C	
LCS 570-699749/2-A	Lab Control Sample	Total/NA	Water	3510C	
LCSD 570-699749/3-A	Lab Control Sample Dup	Total/NA	Water	3510C	
MRL 570-699749/4-A	Lab Control Sample	Total/NA	Water	3510C	
380-198981-B-1-A MS	Matrix Spike	Total/NA	Water	3510C	
380-198981-B-1-B MSD	Matrix Spike Duplicate	Total/NA	Water	3510C	

Analysis Batch: 700123

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-198978-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	8015B	699749
MB 570-699749/1-A	Method Blank	Total/NA	Water	8015B	699749
LCS 570-699749/2-A	Lab Control Sample	Total/NA	Water	8015B	699749
LCSD 570-699749/3-A	Lab Control Sample Dup	Total/NA	Water	8015B	699749
MRL 570-699749/4-A	Lab Control Sample	Total/NA	Water	8015B	699749
380-198981-B-1-A MS	Matrix Spike	Total/NA	Water	8015B	699749
380-198981-B-1-B MSD	Matrix Spike Duplicate	Total/NA	Water	8015B	699749

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-198978-1

Date Collected: 02/17/26 10:46

Matrix: Drinking Water

Date Received: 02/19/26 09:29

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			208072	IQ42	EA POM	02/24/26 14:32
Total/NA	Analysis	525.2		1	208615	Q8LA	EA POM	02/25/26 11:52
Total/NA	Prep	625.1			698456	S4EA	EET CAL 4	02/20/26 07:00
Total/NA	Analysis	625.1		1	702941	PQS1	EET CAL 4	03/02/26 11:05
Total/NA	Prep	625.1			698456	S4EA	EET CAL 4	02/20/26 07:00
Total/NA	Analysis	625.1 SIM		1	701134	J7WE	EET CAL 4	02/25/26 20:45
Total/NA	Analysis	8015B GRO LL		1	698930	A9VE	EET CAL 4	02/21/26 00:53
Total/NA	Prep	3510C			699749	TVD6	EET CAL 4	02/23/26 08:51
Total/NA	Analysis	8015B		1	700123	NR	EET CAL 4	02/24/26 03:21

**Client Sample ID: TB: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-198978-2

Date Collected: 02/17/26 10:46

Matrix: Water

Date Received: 02/19/26 09:29

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	698930	A9VE	EET CAL 4	02/20/26 23:20

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100
EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

Accreditation/Certification Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-198978-1
 SDG: Weekly: Aiea Gulch Wells Pump 1

Laboratory: Eurofins Pomona

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Drinking Water	1-Methylnaphthalene
525.2	525.2	Drinking Water	2,4'-DDD
525.2	525.2	Drinking Water	2,4'-DDE
525.2	525.2	Drinking Water	2,4'-DDT
525.2	525.2	Drinking Water	2,4-Dinitrotoluene
525.2	525.2	Drinking Water	2,6-Dinitrotoluene
525.2	525.2	Drinking Water	2-Methylnaphthalene
525.2	525.2	Drinking Water	4,4'-DDD
525.2	525.2	Drinking Water	4,4'-DDE
525.2	525.2	Drinking Water	4,4' DDT
525.2	525.2	Drinking Water	Acetochlor
525.2	525.2	Drinking Water	alpha-BHC
525.2	525.2	Drinking Water	alpha-Chlordane
525.2	525.2	Drinking Water	beta-BHC
525.2	525.2	Drinking Water	Chlorobenzilate
525.2	525.2	Drinking Water	Chloroneb
525.2	525.2	Drinking Water	Chlorothalonil (Draconil, Bravo)
525.2	525.2	Drinking Water	Chlorpyrifos
525.2	525.2	Drinking Water	delta-BHC
525.2	525.2	Drinking Water	Diclorvos (DDVP)
525.2	525.2	Drinking Water	Endosulfan I (Alpha)
525.2	525.2	Drinking Water	Endosulfan II (Beta)
525.2	525.2	Drinking Water	Endosulfan sulfate
525.2	525.2	Drinking Water	Endrin aldehyde
525.2	525.2	Drinking Water	EPTC
525.2	525.2	Drinking Water	gamma-Chlordane
525.2	525.2	Drinking Water	Isophorone
525.2	525.2	Drinking Water	Malathion
525.2	525.2	Drinking Water	Parathion
525.2	525.2	Drinking Water	Pendimethalin (Penoxaline)
525.2	525.2	Drinking Water	Terbacil
525.2	525.2	Drinking Water	Terbutylazine
525.2	525.2	Drinking Water	Total Permethrin (mixed isomers)
525.2	525.2	Drinking Water	trans-Nonachlor

Laboratory: Eurofins Calscience

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
A2LA	Dept. of Defense ELAP	7296.01	11-30-26
A2LA	ISO/IEC 17025	7296.01	11-30-26
Alaska (UST)	State	25-005	03-02-26
Arizona	State	AZ0830	11-17-26
California	Los Angeles County Sanitation Districts	9257304	07-31-26
California	SCAQMD LAP	17LA0919	11-30-26

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Laboratory: Eurofins Calscience (Continued)

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
California	State	3082	07-31-26
Kansas	NELAP	E-10420	07-31-26
Nevada	State	CA00111	07-31-26
Oregon	NELAP	4175	02-02-27
USDA	US Federal Programs	525-23-159-97150	06-08-26
Utah	NELAP	CA00111	03-01-27
Washington	State	C916	10-12-26

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Method	Method Description	Protocol	Laboratory
525.2	Semivolatile Organic Compounds (GC/MS)	EPA	EA POM
625.1	Semivolatile Organic Compounds (GC/MS)	EPA	EET CAL 4
625.1 SIM	Semivolatile Organic Compounds GC/MS (SIM)	EPA	EET CAL 4
8015B GRO LL	Gasoline Range Organics - (GC)	SW846	EET CAL 4
8015B	Diesel Range Organics (DRO) (GC) Low Level	SW846	EET CAL 4
3510C	Liquid-Liquid Extraction (Separatory Funnel)	SW846	EET CAL 4
5030C	Purge and Trap	SW846	EET CAL 4
525.2	Extraction of Semivolatile Compounds	EPA	EA POM
625.1	Liquid-Liquid Extraction	40CFR136A	EET CAL 4

Protocol References:

40CFR136A = "Methods for Organic Chemical Analysis of Municipal Industrial Wastewater", 40CFR, Part 136, Appendix A, October 26, 1984 and subsequent revisions.

EPA = US Environmental Protection Agency

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-198978-1
SDG: Weekly: Aiea Gulch Wells Pump 1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-198978-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Drinking Water	02/17/26 10:46	02/19/26 09:29	HI0000331
380-198978-2	TB: AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Water	02/17/26 10:46	02/19/26 09:29	

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Chain of Custody Record



Client Information		Sampler: bailey	Lab PM: Lopez, Maria	Carrier Tracking No(s):	COC No:
Client Contact: Mr. Kirk Iwamoto		Phone: +1 808 748 5840	E-Mail: Maria.Lopez@et.eurofinsus.com	State of Origin:	Page: Page 1 of 1
Company: City & County of Honolulu		PWSID:	Analysis Requested		
Address: 630 South Beretania Street Chemistry Lab		Due Date Requested:	Total Number of Containers		
City: Honolulu		TAT Requested (days):	53 - All Analytes		
State, Zip: HI, 96843		Compliance Project: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	537.1_DW_PREC - 637.1 Full List		
Phone: 808-748-5840 (Tel)		PO #: C20525101 exp 05312023	536.2_PREC - (MOD) 526plus Plus TICs		
Email: kiwamoto@hbws.org		WO #:	5015B_PRO_LL_C9 - HNL Ranges: C10-C24/C24-C36/C8-C18		
Project Name: RED-HILL/HBWS Sites Event Desc: RUSH Weekly Red Hill		Project #: 38001111	5015B_GRO_LL_C9 (MOD) GRO		
Site: Hawaii		SSOW#:	525.1_525.1_SIM		
Sample Identification		Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Matrix (Water, Swab, On-surface, Air)
Aiea Gulch Wells Pump 1	17-Feb-2026	1046	G		Water
Aiea Gulch Wells Pump 1 (Matrix Spike)					Water
Aiea Gulch Wells Pump 1 (Matrix Spike Duplicate)					Water
TB: Aiea Gulch Wells Pump 1	17-Feb-2026	1046			Water
Possible Hazard Identification <input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological Deliverable Requested: I, II, III, IV, Other (specify)					
Empty Kit Relinquished by: _____ Date: _____ Relinquished by: _____ Date/Time: _____ Relinquished by: _____ Date/Time: _____					
Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months					
Special Instructions/QC Requirements:					
Method of Shipment: FedEx 1-8888 8200 2594 Date/Time: 7/19/26 9:29 Received by: _____ Date/Time: _____ Received by: _____ Date/Time: _____ Received by: _____ Date/Time: _____					
Cooler Temperature(s) °C and Other Remarks: (63A) 3-4+0 2-3.6 9-1-10-20-21 Custody Seal No.: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					



ORIGIN ID HIKA (808) 746-5840
BWS CHEMILAB
HONOLULU BOARD OF WATER SUPPLY
630 S. BERETANIA ST.
CHEMICAL LABORATORY
HONOLULU HI 96843
UNITED STATES US

SHIP DATE: 18FEB26
ACTWGT 62.00 LB
CAD 258050552/INET4535

BILL RECIPIENT

TO **EUROFINS RECEIVING DEPARTMENT**
EUROFINS DRINKING WATER TESTING
941 CORPORATE CENTER DR

58KJ5/006774848

POMONA CA 91768

REF: (626) 386-1100

PO: DEPT



4261026012001uv

THU - 19 FEB 10:30A
PRIORITY OVERNIGHT

7 of 8


MPS# 8888 8208 2583

Mstr# 8888 8208 2528

WM ONTA

91768

CA-US ONT



(791A) 21100 = 21 get frozen
Mark Harrutin 2/19/26 929

After printing this label
1 Fold the printed page along the horizontal line
2 Place label in shipping pouch and affix it to your shipment

CONSIGNEE COPY - PLEASE PLACE IN FRONT OF POUCH

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Chain of Custody Record



Client Information (Sub Contract Lab)	Sampler: N/A	Lab PM: Lopez, Maria	Carrier Tracking No(s): N/A	CCC No: 380-306862.1									
Client Contact: Shipping/Receiving	Phone: N/A	E-Mail: Maria.Lopez@et.eurofinsus.com	State of Origin: Hawaii	Page: Page 1 of 1									
Company: Eurofins Environment Testing Southwest L	Accreditations Required (See note): State - Hawaii			Job #: 380-198978-1									
Address: 2841 Dow Avenue, Suite 100, City: Tustin State, Zip: CA, 92780 Phone: 714-995-5494(Tel) Email: N/A Project Name: RED-HILL Site: Honolulu BWS Sites	Due Date Requested: 3/4/2028 TAT Requested (days): N/A PO #: N/A WC #: N/A Project #: 38001111 SSOW#: N/A	Analysis Requested											
Sample Identification - Client ID (Lab ID)		Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (W=water, S=solid, O=wastewater, ST=Sludge, Anal)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	825_1_SiMMSD_Prep(MOD) Extended PAH List	8015B_DRO_LL_CS/2510C_LLHNL Ranges: C10-C24/C24-C36/C6-C18	8015B_GRO_LL/5030C(MOD) GRO	Total Number of Containers	Other: N/A	Preservation Codes:
				Preservation Code:									
AIEA GULCH WELLS PUMP 1 (331-201-TP071) (380-198978-1)		2/17/26	10:46 Hawaiian	G	Water			X	X	X	7		MRLs are needed. Confirm any hits >RL.
FB: AIEA GULCH WELLS PUMP 1 (331-201-TP071) (380-198978-2)		2/17/26	10:46 Hawaiian	G	Water					X	2		MRLs are needed.

380-198978 Chain of Custody

Note: Since laboratory accreditations are subject to change, Eurofins Drinking Water and Wastewater West, LLC places the ownership of method, analyte & accreditation compliance upon our subcontract laboratories. This sample shipment is forwarded under chain-of-custody. If the laboratory does not currently maintain accreditation in the State of Origin listed above for analysis/test/matrix being analyzed, the samples must be shipped back to the Eurofins Drinking Water and Wastewater West, LLC laboratory or other instructions will be provided. Any changes to accreditation status should be brought to Eurofins Drinking Water and Wastewater West, LLC attention immediately. If all requested accreditations are current to date, return the signed Chain of Custody attesting to said compliance to Eurofins Drinking Water and Wastewater West, LLC.

Possible Hazard Identification		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)	
Unconfirmed		<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months	
Deliverable Requested: I, II, III, IV, Other (specify)		Special Instructions/QC Requirements:	
Empty Kit Relinquished by:		Date: _____ Time: _____ Method of Shipment:	
Relinquished by: <i>Mack/Kurita</i>	Date/Time: <i>2/19/26 1600</i>	Company: <i>CEAP</i>	Received by: <i>lt</i> Date/Time: <i>2-19-26 1600</i> Company: <i>LJP</i>
Relinquished by: <i>-/1</i>	Date/Time: <i>2-19-26 1706</i>	Company: <i>WP</i>	Received by: <i>[Signature]</i> Date/Time: <i>2-19-26 1706</i> Company: _____
Relinquished by:	Date/Time:	Company:	Received by: Date/Time: Company:
Custody Seals Intact: <input type="checkbox"/> Yes <input type="checkbox"/> No	Custody Seal No.:	Cooler Temperature(s) °C and Other Remarks: <i>1.1/1.2 IRJ</i>	



Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-198978-1
SDG Number: Weekly: Aiea Gulch Wells Pump 1

Login Number: 198978

List Number: 1

Creator: Del Rosario, Michael

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	



Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-198978-1
SDG Number: Weekly: Aiea Gulch Wells Pump 1

Login Number: 198978

List Number: 2

Creator: Khana, Piyush

List Source: Eurofins Calscience

List Creation: 02/19/26 06:35 PM

Question	Answer	Comment
Radioactivity wasn't checked or is <=/ background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	1.2
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	Received project as a subcontract.
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	fgf5
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	